

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080324\
 Data File : BP021364.D
 Acq On : 04 Aug 2024 07:12
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020726

Quant Time: Aug 04 23:27:22 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 29 12:46:18 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.616	152	195809	20.000	ng/u1	-0.02
20) Naphthalene-d8	10.381	136	804133	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.257	164	534667	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.075	188	1190675	20.000	ng/u1	-0.02
79) Chrysene-d12	21.527	240	1230871	20.000	ng/u1	0.00
88) Perylene-d12	24.821	264	900835	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.158	96	33906	7.880	ng/uL	-0.02
4) Pyridine-d5	3.581	84	231343	18.571	ng/u1	-0.01
7) Phenol-d5	6.840	99	294153	18.426	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.969	67	182772	18.892	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.169	132	244563	18.594	ng/u1	-0.01
15) 4-Methylphenol-d8	8.352	113	235377	17.753	ng/u1	-0.01
21) Nitrobenzene-d5	8.781	128	118344	18.948	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.487	143	135510	18.956	ng/u1	-0.02
28) 2,4-Dichlorophenol-d3	10.046	165	244764	18.935	ng/u1	0.00
31) 4-Chloroaniline-d4	10.546	131	333579	19.479	ng/u1	-0.01
46) Dimethylphthalate-d6	13.669	166	728728	18.749	ng/u1	-0.01
49) Acenaphthylene-d8	13.951	160	825512	18.915	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.581	143	137118	24.795	ng/u1	0.00
60) Fluorene-d10	15.275	176	630343	19.768	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.451	200	126313	17.533	ng/u1	-0.01
73) Anthracene-d10	17.181	188	1031852	19.512	ng/u1	-0.01
81) Pyrene-d10	19.575	212	1264735	16.620	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.592	264	1370466	30.846	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.193	88	37434	7.910	ng/uL	97
5) Pyridine	3.599	79	235532	18.331	ng/u1	95
6) Benzaldehyde	6.799	77	183253	22.770	ng/u1	99
8) Phenol	6.863	94	303452	18.785	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.063	93	243636	18.257	ng/u1	99
12) 2-Chlorophenol	7.199	128	249109	18.897	ng/u1	97
13) 2-Methylphenol	8.087	108	237193	18.986	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.128	45	247721	12.849	ng/u1#	93
16) Acetophenone	8.446	105	385831	18.855	ng/u1	93
17) N-Nitroso-di-n-propyla...	8.416	70	190155	17.735	ng/u1	99
18) 4-Methylphenol	8.416	108	248506	18.568	ng/u1	95
19) Hexachloroethane	8.658	117	108767	18.361	ng/u1	98
22) Nitrobenzene	8.822	77	294513	19.684	ng/u1	99
23) Isophorone	9.334	82	542069	18.224	ng/u1	99
25) 2-Nitrophenol	9.522	139	141813	19.021	ng/u1	95
26) 2,4-Dimethylphenol	9.593	107	281140	19.123	ng/u1	97
27) Bis(2-Chloroethoxy)met...	9.805	93	330933	18.304	ng/u1	100
29) 2,4-Dichlorophenol	10.069	162	237828	18.669	ng/u1	98
30) Naphthalene	10.428	128	812927	19.219	ng/u1	100
32) 4-Chloroaniline	10.569	127	319892	19.562	ng/u1	99
33) Hexachlorobutadiene	10.687	225	174413	18.259	ng/u1	100
34) Caprolactam	11.399	113	70947	18.078	ng/u1	98
35) 4-Chloro-3-methylphenol	11.734	107	269986	19.497	ng/u1	100
36) 2-Methylnaphthalene	12.046	142	546980	18.839	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080324\
 Data File : BP021364.D
 Acq On : 04 Aug 2024 07:12
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 31 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SSTD020726


Quant Time: Aug 04 23:27:22 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Jul 29 12:46:18 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.269	142	563390	18.867	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.422	216	323018	18.603	ng/ul	98
40) Hexachlorocyclopentadiene	12.369	237	149789	16.490	ng/ul	99
41) 2,4,6-Trichlorophenol	12.693	196	196940	17.950	ng/ul	98
42) 2,4,5-Trichlorophenol	12.798	196	212248	18.284	ng/ul	98
43) 1,1'-Biphenyl	13.069	154	721402	18.462	ng/ul	98
44) 2-Chloronaphthalene	13.122	162	584878	18.750	ng/ul	99
45) 2-Nitroaniline	13.363	65	176407	20.450	ng/ul	97
47) Dimethylphthalate	13.722	163	734822	18.633	ng/ul	100
48) 2,6-Dinitrotoluene	13.869	165	152244	19.309	ng/ul	100
50) Acenaphthylene	13.975	152	952962	19.286	ng/ul	99
51) 3-Nitroaniline	14.222	138	147210	21.455	ng/ul	99
52) Acenaphthene	14.322	153	627958	19.147	ng/ul	98
53) 2,4-Dinitrophenol	14.463	184	91370	22.066	ng/ul	95
55) 4-Nitrophenol	14.598	109	100796	22.407	ng/ul	96
56) Dibenzofuran	14.675	168	872239	19.421	ng/ul	100
57) 2,4-Dinitrotoluene	14.692	165	226772	20.952	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	14.922	232	190325	19.146	ng/ul	99
59) Diethylphthalate	15.104	149	741778	18.821	ng/ul	99
61) Fluorene	15.328	166	723654	19.745	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.322	204	372923	18.752	ng/ul	99
63) 4-Nitroaniline	15.416	138	131889	23.212	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.469	198	133425	17.464	ng/ul	98
67) N-Nitrosodiphenylamine	15.551	169	611025	17.623	ng/ul	100
68) 4-Bromophenyl-phenylether	16.239	248	239580	17.167	ng/ul	99
69) Hexachlorobenzene	16.351	284	285048	17.813	ng/ul	98
70) Atrazine	16.539	200	241289	19.095	ng/ul	99
71) Pentachlorophenol	16.739	266	191319	22.013	ng/ul	98
72) Phenanthrene	17.122	178	1187472	19.225	ng/ul	99
74) Anthracene	17.222	178	1218082	19.344	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.040	216	318850	16.731	ng/ul	98
76) Pentachlorobenzene	14.581	250	322998	17.057	ng/ul	99
77) Carbazole	17.516	167	1080567	20.992	ng/ul	99
78) Di-n-butylphthalate	18.081	149	1293723	18.217	ng/ul	100
80) Fluoranthene	19.228	202	1486120	16.171	ng/ul	100
82) Pyrene	19.604	202	1560885	16.691	ng/ul	97
83) Butylbenzylphthalate	20.539	149	634731	16.291	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.439	252	551392	19.604	ng/ul	100
85) Benzo(a)anthracene	21.510	228	1620484	18.794	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.410	149	915052	16.319	ng/ul	99
87) Chrysene	21.580	228	1503559	18.766	ng/ul	99
89) Di-n-octyl phthalate	22.645	149	1602768	26.933	ng/ul	100
90) Benzo(b)fluoranthene	23.774	252	1645292	30.097	ng/ul	99
91) Benzo(k)fluoranthene	23.839	252	1668353	30.580	ng/ul	99
93) Benzo(a)pyrene	24.668	252	1556384	30.128	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	28.556	276	1954839	29.366	ng/ul	98
95) Dibenzo(a,h)anthracene	28.609	278	926876	16.814	ng/ul	99
96) Benzo(g,h,i)perylene	29.750	276	1578670	29.119	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 04 23:27:22 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071024.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Jul 29 12:46:18 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.910 ng/uL
RT: 3.193 min Scan# 31
Delta R.T. -0.011 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion: 88 Resp: 37434
Ion Ratio Lower Upper
88 100
43 32.6 25.4 38.0
58 76.9 59.2 88.8



#5
Pyridine
Concen: 18.331 ng/uL
RT: 3.599 min Scan# 104
Delta R.T. -0.011 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion: 79 Resp: 235532
Ion Ratio Lower Upper
79 100
52 87.7 64.9 97.3
51 38.6 30.5 45.7









#6
Benzaldehyde
Concen: 22.770 ng/ul
RT: 6.799 min Scan# 647
Delta R.T. -0.006 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion: 77 Resp: 183253
Ion Ratio Lower Upper
77 100
105 99.2 79.3 118.9
106 91.9 74.6 112.0



#8
Phenol
Concen: 18.785 ng/ul
RT: 6.863 min Scan# 659
Delta R.T. -0.011 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion: 94 Resp: 303452
Ion Ratio Lower Upper
94 100
65 29.0 23.4 35.0
66 40.2 33.9 50.9









#10
Bis(2-Chloroethyl)ether
Concen: 18.257 ng/ul
RT: 7.063 min Scan# 692
Delta R.T. -0.011 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion: 93 Resp: 243636
Ion Ratio Lower Upper
93 100
63 77.3 61.5 92.3
95 32.4 25.1 37.7



#12
2-Chlorophenol
Concen: 18.897 ng/ul
RT: 7.199 min Scan# 716
Delta R.T. -0.011 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion:128 Resp: 249109
Ion Ratio Lower Upper
128 100
64 50.8 37.9 56.9
130 32.2 25.6 38.4





#13

2-Methylphenol

Concen: 18.986 ng/ul

RT: 8.087 min Scan# 8

Delta R.T. -0.011 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12

Instrument :

BNA_P

ClientSampleId :





SSTD020726

Tgt Ion:108 Resp: 237193

Ion Ratio Lower Upper

108 100

107 91.3 73.8 110.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 12.849 ng/ul

RT: 8.128 min Scan# 874

Delta R.T. -0.029 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12




Tgt Ion: 45 Resp: 247721


Ion Ratio Lower Upper

45 100

77 16.6 10.9 16.3#

79 12.1 8.1 12.1









#16
Acetophenone
Concen: 18.855 ng/ul
RT: 8.446 min Scan# 91
Delta R.T. -0.011 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:105 Resp: 385831
Ion Ratio Lower Upper
105 100
77 83.7 61.8 92.8
51 30.8 21.5 32.3



#17
N-Nitroso-di-n-propylamine
Concen: 17.735 ng/ul
RT: 8.416 min Scan# 923
Delta R.T. -0.017 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion: 70 Resp: 190155
Ion Ratio Lower Upper
70 100
42 59.1 46.6 69.8
101 9.7 7.4 11.0
130 20.9 17.2 25.8









#18
4-Methylphenol
Concen: 18.568 ng/ul
RT: 8.416 min Scan# 91
Delta R.T. -0.011 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:108 Resp: 248506
Ion Ratio Lower Upper
108 100
107 114.6 87.8 131.8



#19
Hexachloroethane
Concen: 18.361 ng/ul
RT: 8.658 min Scan# 964
Delta R.T. -0.017 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion:117 Resp: 108767
Ion Ratio Lower Upper
117 100
201 112.5 88.2 132.2
199 71.8 55.8 83.6









#22
Nitrobenzene
Concen: 19.684 ng/ul
RT: 8.822 min Scan# 991
Delta R.T. -0.017 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion: 77 Resp: 294513
Ion Ratio Lower Upper
77 100
123 45.3 36.7 55.1
65 13.3 11.1 16.7



#23
Isophorone
Concen: 18.224 ng/ul
RT: 9.334 min Scan# 1079
Delta R.T. -0.011 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion: 82 Resp: 542069
Ion Ratio Lower Upper
82 100
95 8.0 6.1 9.1
138 17.7 14.4 21.6









#25
2-Nitrophenol
Concen: 19.021 ng/ul
RT: 9.522 min Scan# 1111
Delta R.T. -0.017 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:139 Resp: 141813
Ion Ratio Lower Upper
139 100
65 58.2 42.9 64.3
109 33.7 26.2 39.4



#26
2,4-Dimethylphenol
Concen: 19.123 ng/ul
RT: 9.593 min Scan# 1123
Delta R.T. -0.011 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion:107 Resp: 281140
Ion Ratio Lower Upper
107 100
121 49.6 39.6 59.4
122 79.2 66.3 99.5





#27

Bis(2-Chloroethoxy)methane

Concen: 18.304 ng/ul

RT: 9.805 min Scan# 1157

Delta R.T. -0.017 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion: 93 Resp: 330933

Ion Ratio Lower Upper

93 100

95 32.0 25.5 38.3

123 11.6 9.0 13.4



#29

2,4-Dichlorophenol

Concen: 18.669 ng/ul

RT: 10.069 min Scan# 1204

Delta R.T. -0.011 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12




Tgt Ion:162 Resp: 237828


Ion Ratio Lower Upper

162 100

164 67.6 53.3 79.9

98 37.6 28.7 43.1









#30
Naphthalene
Concen: 19.219 ng/ul
RT: 10.428 min Scan# 1264
Delta R.T. -0.011 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:128 Resp: 812927
Ion Ratio Lower Upper
128 100
129 10.8 8.6 13.0
127 13.4 10.8 16.2



#32
4-Chloroaniline
Concen: 19.562 ng/ul
RT: 10.569 min Scan# 1289
Delta R.T. -0.017 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion:127 Resp: 319892
Ion Ratio Lower Upper
127 100
129 33.1 26.1 39.1





#33

Hexachlorobutadiene

Concen: 18.259 ng/ul

RT: 10.687 min Scan# 11

Delta R.T. -0.023 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:225 Resp: 174413

Ion Ratio Lower Upper

225 100

223 62.5 50.2 75.2

227 63.2 50.6 75.8



#34

Caprolactam

Concen: 18.078 ng/ul

RT: 11.399 min Scan# 1430

Delta R.T. -0.005 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12




Tgt Ion:113 Resp: 70947


Ion Ratio Lower Upper

113 100

55 173.4 140.2 210.4

56 134.8 105.5 158.3









#35
4-Chloro-3-methylphenol
Concen: 19.497 ng/ul
RT: 11.734 min Scan# 1486
Delta R.T. -0.005 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:107 Resp: 269986
Ion Ratio Lower Upper
107 100
144 24.0 19.3 28.9
142 75.5 60.7 91.1



#36
2-Methylnaphthalene
Concen: 18.839 ng/ul
RT: 12.046 min Scan# 1540
Delta R.T. -0.023 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion:142 Resp: 546980
Ion Ratio Lower Upper
142 100
141 89.2 72.6 108.8





#37

1-Methylnaphthalene

Concen: 18.867 ng/ul

RT: 12.269 min Scan# 1

Delta R.T. -0.023 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:142 Resp: 563390

Ion Ratio Lower Upper

142 100

141 93.1 75.0 112.6

116 4.0 3.2 4.8



#39

1,2,4,5-Tetrachlorobenzene

Concen: 18.603 ng/ul

RT: 12.422 min Scan# 1604

Delta R.T. -0.017 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12

Tgt Ion:216 Resp: 323018




Ion Ratio Lower Upper


216 100

214 78.9 64.2 96.4

179 20.6 16.6 24.8

108 15.6 13.4 20.2





#40

Hexachlorocyclopentadiene

Concen: 16.490 ng/ul

RT: 12.369 min Scan# 1

Delta R.T. -0.029 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:237 Resp: 149789

Ion Ratio Lower Upper

237 100

235 63.2 49.7 74.5

272 12.9 10.2 15.2



#41

2,4,6-Trichlorophenol

Concen: 17.950 ng/ul

RT: 12.693 min Scan# 1650

Delta R.T. -0.011 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12




Tgt Ion:196 Resp: 196940


Ion Ratio Lower Upper

196 100

198 98.2 76.6 115.0

200 32.5 26.0 39.0









#42
2,4,5-Trichlorophenol
Concen: 18.284 ng/ul
RT: 12.798 min Scan# 1
Delta R.T. -0.011 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:196 Resp: 212248
Ion Ratio Lower Upper
196 100
198 94.1 77.1 115.7
200 31.0 25.3 37.9



#43
1,1'-Biphenyl
Concen: 18.462 ng/ul
RT: 13.069 min Scan# 1714
Delta R.T. -0.017 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion:154 Resp: 721402
Ion Ratio Lower Upper
154 100
153 40.3 32.7 49.1
76 13.2 9.4 14.2









#44
2-Chloronaphthalene
Concen: 18.750 ng/ul
RT: 13.122 min Scan# 1723
Delta R.T. -0.017 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:162 Resp: 584878
Ion Ratio Lower Upper
162 100
164 32.8 26.3 39.5
127 40.5 31.8 47.8



#45
2-Nitroaniline
Concen: 20.450 ng/ul
RT: 13.363 min Scan# 1764
Delta R.T. -0.005 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion: 65 Resp: 176407
Ion Ratio Lower Upper
65 100
92 68.6 56.7 85.1
138 103.1 85.6 128.4









#47
Dimethylphthalate
Concen: 18.633 ng/ul
RT: 13.722 min Scan# 1825
Delta R.T. -0.011 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:163 Resp: 734822
Ion Ratio Lower Upper
163 100
194 4.0 3.4 5.0
164 10.0 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 19.309 ng/ul
RT: 13.869 min Scan# 1850
Delta R.T. -0.006 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion:165 Resp: 152244
Ion Ratio Lower Upper
165 100
89 55.8 44.7 67.1
121 20.0 16.5 24.7









#50
Acenaphthylene
Concen: 19.286 ng/ul
RT: 13.975 min Scan# 1868
Delta R.T. -0.017 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:152 Resp: 952962
Ion Ratio Lower Upper
152 100
151 19.8 16.2 24.2
153 13.2 10.3 15.5



#51
3-Nitroaniline
Concen: 21.455 ng/ul
RT: 14.222 min Scan# 1910
Delta R.T. 0.000 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion:138 Resp: 147210
Ion Ratio Lower Upper
138 100
108 10.9 9.8 14.8
92 121.3 98.0 147.0









#52
Acenaphthene
Concen: 19.147 ng/ul
RT: 14.322 min Scan# 1927
Delta R.T. -0.017 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:153 Resp: 627958
Ion Ratio Lower Upper
153 100
152 46.3 37.8 56.6
154 87.7 71.8 107.6



#53
2,4-Dinitrophenol
Concen: 22.066 ng/ul
RT: 14.463 min Scan# 1951
Delta R.T. 0.000 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion:184 Resp: 91370
Ion Ratio Lower Upper
184 100
63 64.3 50.3 75.5
154 60.3 53.0 79.6









#55
4-Nitrophenol
Concen: 22.407 ng/ul
RT: 14.598 min Scan# 1974
Delta R.T. 0.006 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:109 Resp: 100796
Ion Ratio Lower Upper
109 100
139 119.0 98.0 147.0
65 126.7 106.0 159.0



#56
Dibenzofuran
Concen: 19.421 ng/ul
RT: 14.675 min Scan# 1987
Delta R.T. -0.011 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion:168 Resp: 872239
Ion Ratio Lower Upper
168 100
139 38.6 31.1 46.7









#57
2,4-Dinitrotoluene
Concen: 20.952 ng/ul
RT: 14.692 min Scan# 1987
Delta R.T. -0.005 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:165 Resp: 226772
Ion Ratio Lower Upper
165 100
63 48.9 41.0 61.4
182 2.9 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 19.146 ng/ul
RT: 14.922 min Scan# 2029
Delta R.T. -0.011 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion:232 Resp: 190325
Ion Ratio Lower Upper
232 100
131 43.9 34.4 51.6
130 2.3 2.0 3.0
166 31.0 24.4 36.6









#59
Diethylphthalate
Concen: 18.821 ng/ul
RT: 15.104 min Scan# 2060
Delta R.T. -0.017 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:149 Resp: 741778
Ion Ratio Lower Upper
149 100
177 21.5 17.7 26.5
150 12.8 10.2 15.2



#61
Fluorene
Concen: 19.745 ng/ul
RT: 15.328 min Scan# 2098
Delta R.T. -0.017 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion:166 Resp: 723654
Ion Ratio Lower Upper
166 100
165 96.7 78.1 117.1
167 13.7 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 18.752 ng/ul
RT: 15.322 min Scan# 2096
Delta R.T. -0.011 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:204 Resp: 372923
Ion Ratio Lower Upper
204 100
206 33.4 26.4 39.6
141 60.2 48.5 72.7



#63
4-Nitroaniline
Concen: 23.212 ng/ul
RT: 15.416 min Scan# 2113
Delta R.T. -0.011 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion:138 Resp: 131889
Ion Ratio Lower Upper
138 100
92 52.1 40.9 61.3
108 86.0 64.7 97.1









#66
 4,6-Dinitro-2-methylphenol
 Concen: 17.464 ng/ul
 RT: 15.469 min Scan# 2122
 Delta R.T. -0.011 min
 Lab File: BP021364.D
 Acq: 04 Aug 2024 07:12




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020726


Tgt Ion:198 Resp: 133425
 Ion Ratio Lower Upper
 198 100
 182 5.3 3.8 5.6
 77 24.1 20.0 30.0



#67
 N-Nitrosodiphenylamine
 Concen: 17.623 ng/ul
 RT: 15.551 min Scan# 2136
 Delta R.T. -0.017 min
 Lab File: BP021364.D
 Acq: 04 Aug 2024 07:12

Tgt Ion:169 Resp: 611025
 Ion Ratio Lower Upper
 169 100
 168 66.6 53.4 80.2
 167 36.0 28.4 42.6









#68
4-Bromophenyl-phenylether
Concen: 17.167 ng/ul
RT: 16.239 min Scan# 211
Delta R.T. -0.023 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:248 Resp: 239580
Ion Ratio Lower Upper
248 100
250 100.4 79.3 118.9
141 72.3 57.4 86.0



#69
Hexachlorobenzene
Concen: 17.813 ng/ul
RT: 16.351 min Scan# 2272
Delta R.T. -0.023 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion:284 Resp: 285048
Ion Ratio Lower Upper
284 100
142 34.2 25.8 38.8
249 32.1 25.1 37.7









#70
Atrazine
Concen: 19.095 ng/ul
RT: 16.539 min Scan# 21
Delta R.T. -0.017 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:200 Resp: 241289
Ion Ratio Lower Upper
200 100
173 25.0 20.8 31.2
215 49.0 38.7 58.1



#71
Pentachlorophenol
Concen: 22.013 ng/ul
RT: 16.739 min Scan# 2338
Delta R.T. -0.011 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion:266 Resp: 191319
Ion Ratio Lower Upper
266 100
264 63.0 52.3 78.5
268 64.9 52.3 78.5





#72

Phenanthrene

Concen: 19.225 ng/ul

RT: 17.122 min Scan# 2402

Delta R.T. -0.011 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:178 Resp: 1187472

Ion Ratio Lower Upper

178 100

179 15.5 12.2 18.4

176 19.0 15.0 22.4



#74

Anthracene

Concen: 19.344 ng/ul

RT: 17.222 min Scan# 2420

Delta R.T. -0.011 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12




Tgt Ion:178 Resp: 1218082


Ion Ratio Lower Upper

178 100

179 15.2 12.5 18.7

176 18.2 14.8 22.2









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 16.731 ng/uL
 RT: 13.040 min Scan# 11
 Delta R.T. -0.017 min
 Lab File: BP021364.D
 Acq: 04 Aug 2024 07:12

Instrument :
 BNA_P
 ClientSampleId :
 SST020726

Tgt Ion: 216 Resp: 318850




Ion	Ratio	Lower	Upper
216	100		
214	76.7	63.2	94.8
218	47.9	38.2	57.4




#76
 Pentachlorobenzene
 Concen: 17.057 ng/uL
 RT: 14.581 min Scan# 1971
 Delta R.T. -0.011 min
 Lab File: BP021364.D
 Acq: 04 Aug 2024 07:12

Tgt Ion: 250 Resp: 322998

Ion	Ratio	Lower	Upper
250	100		
248	62.0	50.8	76.2
252	63.9	51.6	77.4









#77
Carbazole
Concen: 20.992 ng/ul
RT: 17.516 min Scan# 2469
Delta R.T. -0.006 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12




Instrument :
BNA_P
ClientSampleId :
SSTD020726


Tgt Ion:167 Resp: 1080567
Ion Ratio Lower Upper
167 100
166 21.8 17.3 25.9
139 13.6 10.2 15.2



#78
Di-n-butylphthalate
Concen: 18.217 ng/ul
RT: 18.081 min Scan# 2566
Delta R.T. -0.017 min
Lab File: BP021364.D
Acq: 04 Aug 2024 07:12

Tgt Ion:149 Resp: 1293723
Ion Ratio Lower Upper
149 100
150 9.1 7.4 11.0
104 5.1 3.9 5.9





#80

Fluoranthene

Concen: 16.171 ng/ul

RT: 19.228 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:202 Resp: 1486120

Ion Ratio Lower Upper

202 100

101 9.8 8.0 12.0

100 7.8 6.1 9.1



#82

Pyrene

Concen: 16.691 ng/ul

RT: 19.604 min Scan# 2825

Delta R.T. -0.005 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12




Tgt Ion:202 Resp: 1560885


Ion Ratio Lower Upper

202 100

101 12.9 9.4 14.2

100 10.8 7.8 11.6





#83

Butylbenzylphthalate

Concen: 16.291 ng/ul

RT: 20.539 min Scan# 2984

Delta R.T. -0.011 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:149 Resp: 634731

Ion Ratio Lower Upper

149 100

91 69.5 54.1 81.1

206 21.3 17.0 25.6



#84

3,3'-Dichlorobenzidine

Concen: 19.604 ng/ul

RT: 21.439 min Scan# 3137

Delta R.T. 0.000 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12




Tgt Ion:252 Resp: 551392


Ion Ratio Lower Upper

252 100

254 64.0 51.5 77.3

126 10.0 8.0 12.0





#85

Benzo(a)anthracene

Concen: 18.794 ng/ul

RT: 21.510 min Scan# 3149

Delta R.T. 0.000 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:228 Resp: 1620484

Ion Ratio Lower Upper

228 100

229 19.1 15.5 23.3

226 27.0 21.4 32.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 16.319 ng/ul

RT: 21.410 min Scan# 3132

Delta R.T. -0.011 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12




Tgt Ion:149 Resp: 915052


Ion Ratio Lower Upper

149 100

167 26.7 22.1 33.1

279 4.5 3.7 5.5





#87

Chrysene

Concen: 18.766 ng/ul

RT: 21.580 min Scan# 3159

Delta R.T. -0.006 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:228 Resp: 1503559

Ion Ratio Lower Upper

228 100

226 29.8 23.6 35.4

229 19.7 15.5 23.3



#89

Di-n-octyl phthalate

Concen: 26.933 ng/ul




RT: 22.645 min Scan# 3342


Delta R.T. -0.035 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12

Tgt Ion:149 Resp: 1602768





#90

Benzo(b)fluoranthene

Concen: 30.097 ng/ul

RT: 23.774 min Scan# 3533

Delta R.T. -0.011 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:252 Resp: 1645292

Ion Ratio Lower Upper

252 100

253 22.3 17.1 25.7

125 8.7 7.0 10.4



#91

Benzo(k)fluoranthene

Concen: 30.580 ng/ul

RT: 23.839 min Scan# 3545

Delta R.T. -0.011 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12




Tgt Ion:252 Resp: 1668353


Ion Ratio Lower Upper

252 100

253 21.9 17.4 26.2

125 8.4 7.4 11.0





#93

Benzo(a)pyrene

Concen: 30.128 ng/ul

RT: 24.668 min Scan# 3

Delta R.T. -0.011 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:252 Resp: 1556384

Ion Ratio Lower Upper

252 100

253 21.8 17.4 26.0

125 9.8 7.7 11.5



#94

Indeno(1,2,3-cd)pyrene

Concen: 29.366 ng/ul

RT: 28.556 min Scan# 4347

Delta R.T. -0.041 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12




Tgt Ion:276 Resp: 1954839


Ion Ratio Lower Upper

276 100

138 19.5 14.3 21.5

277 24.0 19.2 28.8





#95

Dibenzo(a,h)anthracene

Concen: 16.814 ng/ul

RT: 28.609 min Scan# 41

Delta R.T. -0.029 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12

Instrument :

BNA_P

ClientSampleId :

SSTD020726





Tgt Ion:278 Resp: 926876

Ion Ratio Lower Upper

278 100

139 13.4 10.9 16.3

279 24.5 19.4 29.0



#96

Benzo(g,h,i)perylene

Concen: 29.119 ng/ul

RT: 29.750 min Scan# 4550

Delta R.T. -0.011 min

Lab File: BP021364.D

Acq: 04 Aug 2024 07:12

Tgt Ion:276 Resp: 1578670

Ion Ratio Lower Upper

276 100

138 19.3 15.3 22.9

277 24.4 18.2 27.2

