

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080421\
 Data File : BP006488.D
 Acq On : 04 Aug 2021 10:31
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Aug 04 11:43:15 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 15:41:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.757	152	198143	20.000	ng	0.00
21) Naphthalene-d8	10.540	136	849116	20.000	ng	# 0.00
39) Acenaphthene-d10	14.386	164	493515	20.000	ng	0.00
64) Phenanthrene-d10	17.139	188	964069	20.000	ng	# 0.00
76) Chrysene-d12	21.239	240	931327	20.000	ng	# 0.00
86) Perylene-d12	23.562	264	945877	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.352	112	976927	75.728	ng	0.00
7) Phenol-d6	6.934	99	1442314	78.706	ng	0.00
23) Nitrobenzene-d5	8.916	82	1415638	76.953	ng	0.00
42) 2,4,6-Tribromophenol	15.881	330	400395	77.629	ng	0.00
45) 2-Fluorobiphenyl	13.016	172	2505907	75.968	ng	0.00
79) Terphenyl-d14	19.716	244	3564248	76.686	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.305	88	205807	36.644	ng	91
3) Pyridine	3.693	79	624147	37.810	ng	95
4) n-Nitrosodimethylamine	3.611	42	236307	37.928	ng	# 86
6) Aniline	7.093	93	788386	39.606	ng	94
8) 2-Chlorophenol	7.322	128	540567	38.728	ng	91
9) Benzaldehyde	6.910	77	429380	38.930	ng	92
10) Phenol	6.957	94	723733	39.386	ng	98
11) bis(2-Chloroethyl)ether	7.193	93	549878	39.410	ng	89
12) 1,3-Dichlorobenzene	7.646	146	556503	38.254	ng	97
13) 1,4-Dichlorobenzene	7.793	146	570212	38.621	ng	98
14) 1,2-Dichlorobenzene	8.104	146	543122	38.337	ng	96
15) Benzyl Alcohol	7.999	79	558074	40.609	ng	92
16) 2,2'-oxybis(1-Chloropr...	8.287	45	649059	39.092	ng	92
17) 2-Methylphenol	8.199	107	463725	39.966	ng	98
18) Hexachloroethane	8.828	117	223308	38.213	ng	89
19) n-Nitroso-di-n-propyla...	8.569	70	497538	40.468	ng	97
20) 3+4-Methylphenols	8.528	107	641189	40.599	ng	96
22) Acetophenone	8.581	105	862308	38.294	ng	# 98
24) Nitrobenzene	8.957	77	732450	38.302	ng	90
25) Isophorone	9.481	82	1304290	39.474	ng	94
26) 2-Nitrophenol	9.663	139	276594	39.744	ng	92
27) 2,4-Dimethylphenol	9.728	122	459179	39.402	ng	97
28) bis(2-Chloroethoxy)met...	9.969	93	677068	38.298	ng	98
29) 2,4-Dichlorophenol	10.187	162	457793	38.801	ng	94
30) 1,2,4-Trichlorobenzene	10.404	180	479726	37.839	ng	98
31) Naphthalene	10.593	128	1743081	38.203	ng	99
32) Benzoic acid	9.863	122	344292	38.189	ng	91
33) 4-Chloroaniline	10.704	127	721407	39.072	ng	95
34) Hexachlorobutadiene	10.875	225	276748	37.459	ng	99
35) Caprolactam	11.498	113	173028	40.671	ng	# 69
36) 4-Chloro-3-methylphenol	11.834	107	598635	39.934	ng	91
37) 2-Methylnaphthalene	12.204	142	1197197	38.722	ng	100
38) 1-Methylnaphthalene	12.422	142	1132313	38.538	ng	98
40) 1,2,4,5-Tetrachloroben...	12.569	216	486186	37.640	ng	# 95
41) Hexachlorocyclopentadiene	12.551	237	273375	39.905	ng	96
43) 2,4,6-Trichlorophenol	12.816	196	348144	39.474	ng	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080421\
 Data File : BP006488.D
 Acq On : 04 Aug 2021 10:31
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Aug 04 11:43:15 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 15:41:30 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.887	196	379949	38.919	ng	# 88
46) 1,1'-Biphenyl	13.222	154	1425824	38.170	ng	96
47) 2-Chloronaphthalene	13.263	162	1099068	38.207	ng	95
48) 2-Nitroaniline	13.469	65	406030	39.996	ng	# 85
49) Acenaphthylene	14.110	152	1785729	38.499	ng	99
50) Dimethylphthalate	13.857	163	1376743	38.324	ng	99
51) 2,6-Dinitrotoluene	13.975	165	311869	38.905	ng	# 83
52) Acenaphthene	14.451	154	1082573	38.349	ng	98
53) 3-Nitroaniline	14.298	138	350051	39.390	ng	82
54) 2,4-Dinitrophenol	14.504	184	163984	39.119	ng	# 87
55) Dibenzofuran	14.786	168	1701123	38.311	ng	96
56) 4-Nitrophenol	14.604	139	305758	39.624	ng	85
57) 2,4-Dinitrotoluene	14.757	165	424403	39.478	ng	# 78
58) Fluorene	15.439	166	1352370	38.105	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.010	232	316927	38.737	ng	# 71
60) Diethylphthalate	15.228	149	1438123	38.114	ng	98
61) 4-Chlorophenyl-phenyle...	15.439	204	610038	37.955	ng	91
62) 4-Nitroaniline	15.463	138	377065	39.986	ng	89
63) Azobenzene	15.733	77	1723513	39.179	ng	92
65) 4,6-Dinitro-2-methylph...	15.522	198	226546	38.458	ng	78
66) n-Nitrosodiphenylamine	15.657	169	1174057	38.745	ng	100
67) 4-Bromophenyl-phenylether	16.333	248	355244	38.404	ng	# 88
68) Hexachlorobenzene	16.439	284	401768	38.167	ng	# 86
69) Atrazine	16.616	200	366097	38.717	ng	95
70) Pentachlorophenol	16.786	266	265830	39.802	ng	99
71) Phenanthrene	17.180	178	2094836	38.523	ng	100
72) Anthracene	17.275	178	2037201	38.770	ng	100
73) Carbazole	17.545	167	2083449	38.873	ng	98
74) Di-n-butylphthalate	18.116	149	2384354	39.048	ng	100
75) Fluoranthene	19.157	202	2333685	38.235	ng	94
77) Benzidine	19.345	184	1016000	43.596	ng	98
78) Pyrene	19.510	202	2423325	38.955	ng	99
80) Butylbenzylphthalate	20.398	149	1088160	39.780	ng	86
81) Benzo(a)anthracene	21.221	228	2312133	37.988	ng	100
82) 3,3'-Dichlorobenzidine	21.163	252	624503	39.083	ng	# 97
83) Chrysene	21.274	228	2248973	38.114	ng	99
84) Bis(2-ethylhexyl)phtha...	21.163	149	1578937	38.410	ng	# 96
85) Di-n-octyl phthalate	22.068	149	2622529	38.706	ng	# 93
87) Indeno(1,2,3-cd)pyrene	25.968	276	2639431	38.483	ng	# 88
88) Benzo(b)fluoranthene	22.857	252	2350071	38.228	ng	# 96
89) Benzo(k)fluoranthene	22.904	252	2303169	39.021	ng	# 95
90) Benzo(a)pyrene	23.462	252	2141458	38.837	ng	# 95
91) Dibenzo(a,h)anthracene	25.998	278	2282362	38.488	ng	# 93
92) Benzo(g,h,i)perylene	26.709	276	2180147	38.365	ng	# 88


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080421\
 Data File : BP006488.D
 Acq On : 04 Aug 2021 10:31
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Aug 04 11:43:15 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 15:41:30 2021
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.757 min Scan# 794
Delta R.T. 0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 198143
Ion Ratio Lower Upper
152 100
150 160.2 125.6 188.4
115 66.2 48.5 72.7



#2
1,4-Dioxane
Concen: 36.644 ng
RT: 3.305 min Scan# 37
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion: 88 Resp: 205807
Ion Ratio Lower Upper
88 100
58 67.6 46.7 70.1
43 26.1 20.3 30.5









#3
Pyridine
Concen: 37.810 ng
RT: 3.693 min Scan# 103
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 624147
Ion Ratio Lower Upper
79 100
52 57.4 41.7 62.5
51 29.5 24.1 36.1



#4
n-Nitrosodimethylamine
Concen: 37.928 ng
RT: 3.611 min Scan# 89
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion: 42 Resp: 236307
Ion Ratio Lower Upper
42 100
74 148.6 105.4 158.0
44 8.8 4.2 6.4









#5
2-Fluorophenol
Concen: 75.728 ng
RT: 5.352 min Scan# 385
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:112 Resp: 976927
Ion Ratio Lower Upper
112 100
64 62.8 41.4 62.2#
63 32.8 23.8 35.8



#6
Aniline
Concen: 39.606 ng
RT: 7.093 min Scan# 681
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion: 93 Resp: 788386
Ion Ratio Lower Upper
93 100
66 41.4 29.7 44.5
65 21.1 15.3 22.9









#7
Phenol-d6
Concen: 78.706 ng
RT: 6.934 min Scan# 654
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 99 Resp: 1442314




Ion	Ratio	Lower	Upper
99	100		
42	16.6	13.4	20.0
71	37.3	30.6	46.0




#8
2-Chlorophenol
Concen: 38.728 ng
RT: 7.322 min Scan# 720
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion: 128 Resp: 540567

Ion	Ratio	Lower	Upper
128	100		
130	32.2	12.4	52.4
64	50.6	20.3	60.3









#9
Benzaldehyde
Concen: 38.930 ng
RT: 6.910 min Scan# 650
Delta R.T. 0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 429380
Ion Ratio Lower Upper
77 100
105 92.4 76.7 116.7
106 85.1 76.2 116.2



#10
Phenol
Concen: 39.386 ng
RT: 6.957 min Scan# 658
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion: 94 Resp: 723733
Ion Ratio Lower Upper
94 100
65 29.4 8.8 48.8
66 38.9 20.7 60.7









#11
bis(2-Chloroethyl)ether
Concen: 39.410 ng
RT: 7.193 min Scan# 698
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 93 Resp: 549878
Ion Ratio Lower Upper
93 100
63 78.9 46.2 86.2
95 31.9 12.2 52.2



#12
1,3-Dichlorobenzene
Concen: 38.254 ng
RT: 7.646 min Scan# 775
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion: 146 Resp: 556503
Ion Ratio Lower Upper
146 100
148 63.1 51.5 77.3
75 35.5 25.7 38.5









#13
1,4-Dichlorobenzene
Concen: 38.621 ng
RT: 7.793 min Scan# 800
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:146 Resp: 570212
Ion Ratio Lower Upper
146 100
148 63.3 51.3 76.9
111 43.8 33.0 49.6



#14
1,2-Dichlorobenzene
Concen: 38.337 ng
RT: 8.104 min Scan# 853
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:146 Resp: 543122
Ion Ratio Lower Upper
146 100
148 64.2 50.3 75.5
111 46.8 33.7 50.5









#15
Benzyl Alcohol
Concen: 40.609 ng
RT: 7.999 min Scan# 835
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 558074
Ion Ratio Lower Upper
79 100
108 67.8 62.5 93.7
77 64.0 52.9 79.3



#16
2,2'-oxybis(1-Chloropropane)
Concen: 39.092 ng
RT: 8.287 min Scan# 884
Delta R.T. -0.012 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion: 45 Resp: 649059
Ion Ratio Lower Upper
45 100
77 17.8 2.0 42.0
79 13.4 0.0 36.7









#17
2-Methylphenol
Concen: 39.966 ng
RT: 8.199 min Scan# 869
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:107 Resp: 463725




Ion	Ratio	Lower	Upper
107	100		
108	112.1	87.7	131.5
77	51.0	40.2	60.2
79	49.0	39.0	58.4




#18
Hexachloroethane
Concen: 38.213 ng
RT: 8.828 min Scan# 976
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:117 Resp: 223308

Ion	Ratio	Lower	Upper
117	100		
119	97.3	76.2	114.2
201	85.2	83.4	125.0









#19
n-Nitroso-di-n-propylamine
Concen: 40.468 ng
RT: 8.569 min Scan# 932
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 70 Resp: 497538




Ion	Ratio	Lower	Upper
70	100		
42	50.4	40.8	61.2
101	9.6	8.8	13.2
130	19.3	18.9	28.3




#20
3+4-Methylphenols
Concen: 40.599 ng
RT: 8.528 min Scan# 925
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion: 107 Resp: 641189

Ion	Ratio	Lower	Upper
107	100		
108	85.1	71.0	111.0
77	36.5	16.6	56.6
79	31.0	12.7	52.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.540 min Scan# 1267
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	136	Resp:	849116
Ion Ratio	Lower	Upper	
136	100		
137	10.6	8.6	12.8
54	8.4	4.6	6.8#
68	7.5	3.6	5.4#



#22
Acetophenone
Concen: 38.294 ng
RT: 8.581 min Scan# 934
Delta R.T. 0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:	105	Resp:	862308
Ion Ratio	Lower	Upper	
105	100		
71	5.5	1.9	2.9#
51	24.0	18.4	27.6
120	21.7	17.9	26.9









#23
Nitrobenzene-d5
Concen: 76.953 ng
RT: 8.916 min Scan# 991
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 82 Resp: 1415638




Ion	Ratio	Lower	Upper
82	100		
128	38.6	35.7	53.5
54	42.5	35.0	52.4




#24
Nitrobenzene
Concen: 38.302 ng
RT: 8.957 min Scan# 998
Delta R.T. 0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion: 77 Resp: 732450

Ion	Ratio	Lower	Upper
77	100		
123	39.5	38.2	57.2
65	12.6	11.4	17.2









#25
Isophorone
Concen: 39.474 ng
RT: 9.481 min Scan# 1087
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 1304290
Ion Ratio Lower Upper
82 100
95 7.0 6.5 9.7
138 15.3 15.0 22.4



#26
2-Nitrophenol
Concen: 39.744 ng
RT: 9.663 min Scan# 1118
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion: 139 Resp: 276594
Ion Ratio Lower Upper
139 100
109 33.4 28.8 43.2
65 58.9 41.0 61.6









#27
2,4-Dimethylphenol
Concen: 39.402 ng
RT: 9.728 min Scan# 1129
Delta R.T. 0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:122 Resp: 459179
Ion Ratio Lower Upper
122 100
107 121.8 94.5 141.7
121 61.0 47.4 71.2



#28
bis(2-Chloroethoxy)methane
Concen: 38.298 ng
RT: 9.969 min Scan# 1170
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion: 93 Resp: 677068
Ion Ratio Lower Upper
93 100
95 32.4 25.9 38.9
123 10.6 10.2 15.2









#29
2,4-Dichlorophenol
Concen: 38.801 ng
RT: 10.187 min Scan# 1207
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 457793
Ion Ratio Lower Upper
162 100
164 63.9 44.9 84.9
98 43.5 14.3 54.3



#30
1,2,4-Trichlorobenzene
Concen: 37.839 ng
RT: 10.404 min Scan# 1244
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:180 Resp: 479726
Ion Ratio Lower Upper
180 100
182 96.4 77.9 116.9
145 34.6 24.7 37.1









#31
Naphthalene
Concen: 38.203 ng
RT: 10.593 min Scan# 1276
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:128 Resp: 1743081
Ion Ratio Lower Upper
128 100
129 10.8 9.0 13.6
127 13.2 10.7 16.1



#32
Benzoic acid
Concen: 38.189 ng
RT: 9.863 min Scan# 1152
Delta R.T. 0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:122 Resp: 344292
Ion Ratio Lower Upper
122 100
105 130.5 103.2 143.2
77 108.3 76.2 116.2









#33
4-Chloroaniline
Concen: 39.072 ng
RT: 10.704 min Scan# 1295
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	127	Resp:	721407
Ion	Ratio	Lower	Upper
127	100		
129	32.5	26.1	39.1
65	33.7	22.5	33.7
92	21.2	15.0	22.6



#34
Hexachlorobutadiene
Concen: 37.459 ng
RT: 10.875 min Scan# 1324
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:	225	Resp:	276748
Ion	Ratio	Lower	Upper
225	100		
223	63.2	51.0	76.6
227	65.7	51.4	77.2









#35
Caprolactam
Concen: 40.671 ng
RT: 11.498 min Scan# 1430
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:113 Resp: 173028
Ion Ratio Lower Upper
113 100
55 157.4 98.9 138.9#
56 116.7 71.7 111.7#



#36
4-Chloro-3-methylphenol
Concen: 39.934 ng
RT: 11.834 min Scan# 1487
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:107 Resp: 598635
Ion Ratio Lower Upper
107 100
144 23.4 21.1 31.7
142 73.0 66.0 99.0









#37
2-Methylnaphthalene
Concen: 38.722 ng
RT: 12.204 min Scan# 1550
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:142 Resp: 1197197
Ion Ratio Lower Upper
142 100
141 89.7 71.8 107.6
115 38.8 31.6 47.4



#38
1-Methylnaphthalene
Concen: 38.538 ng
RT: 12.422 min Scan# 1587
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:142 Resp: 1132313
Ion Ratio Lower Upper
142 100
141 93.3 73.3 109.9
116 4.2 3.5 5.3









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.386 min Scan# 1921
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:164 Resp: 493515
Ion Ratio Lower Upper
164 100
162 102.8 80.2 120.2
160 46.2 34.5 51.7



#40
1,2,4,5-Tetrachlorobenzene
Concen: 37.640 ng
RT: 12.569 min Scan# 1612
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:216 Resp: 486186
Ion Ratio Lower Upper
216 100
214 79.6 61.9 92.9
179 23.9 16.8 25.2
108 25.5 14.9 22.3#









#41
Hexachlorocyclopentadiene
Concen: 39.905 ng
RT: 12.551 min Scan# 1609
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:237 Resp: 273375




Ion	Ratio	Lower	Upper
237	100		
235	61.2	44.5	84.5
272	11.8	0.0	32.5




#42
2,4,6-Tribromophenol
Concen: 77.629 ng
RT: 15.881 min Scan# 2175
Delta R.T. 0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:330 Resp: 400395

Ion	Ratio	Lower	Upper
330	100		
332	96.9	77.8	116.8
141	48.2	25.4	38.2









#43
2,4,6-Trichlorophenol
Concen: 39.474 ng
RT: 12.816 min Scan# 1654
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:196 Resp: 348144
Ion Ratio Lower Upper
196 100
198 94.4 79.5 119.3
200 30.3 25.0 37.6



#44
2,4,5-Trichlorophenol
Concen: 38.919 ng
RT: 12.887 min Scan# 1666
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:196 Resp: 379949
Ion Ratio Lower Upper
196 100
198 96.7 74.7 112.1
97 63.5 37.4 56.0#
132 33.4 20.0 30.0#









#45
2-Fluorobiphenyl
Concen: 75.968 ng
RT: 13.016 min Scan# 1688
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:172 Resp: 2505907
Ion Ratio Lower Upper
172 100
171 35.4 27.9 41.9
170 23.3 18.6 27.8



#46
1,1'-Biphenyl
Concen: 38.170 ng
RT: 13.222 min Scan# 1723
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:154 Resp: 1425824
Ion Ratio Lower Upper
154 100
153 41.1 20.4 60.4
76 16.4 0.0 30.8









#47
2-Chloronaphthalene
Concen: 38.207 ng
RT: 13.263 min Scan# 1730
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 1099068
Ion Ratio Lower Upper
162 100
127 44.1 31.4 47.0
164 32.8 26.6 39.8



#48
2-Nitroaniline
Concen: 39.996 ng
RT: 13.469 min Scan# 1765
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion: 65 Resp: 406030
Ion Ratio Lower Upper
65 100
92 68.1 57.7 86.5
138 91.1 92.1 138.1#









#49
Acenaphthylene
Concen: 38.499 ng
RT: 14.110 min Scan# 1874
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 1785729
Ion Ratio Lower Upper
152 100
151 20.8 16.2 24.2
153 13.1 10.6 16.0



#50
Dimethylphthalate
Concen: 38.324 ng
RT: 13.857 min Scan# 1831
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:163 Resp: 1376743
Ion Ratio Lower Upper
163 100
194 3.7 3.4 5.2
164 10.3 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 38.905 ng
RT: 13.975 min Scan# 1851
Delta R.T. 0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:165 Resp: 311869
Ion Ratio Lower Upper
165 100
63 64.7 41.3 61.9#
89 66.6 44.8 67.2



#52
Acenaphthene
Concen: 38.349 ng
RT: 14.451 min Scan# 1932
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:154 Resp: 1082573
Ion Ratio Lower Upper
154 100
153 115.0 89.4 134.2
152 53.9 43.0 64.4









#53
3-Nitroaniline
Concen: 39.390 ng
RT: 14.298 min Scan# 1906
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:138 Resp: 350051
Ion Ratio Lower Upper
138 100
108 11.9 8.4 12.6
92 130.6 87.8 131.8



#54
2,4-Dinitrophenol
Concen: 39.119 ng
RT: 14.504 min Scan# 1941
Delta R.T. 0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:184 Resp: 163984
Ion Ratio Lower Upper
184 100
63 76.6 47.0 70.6#
154 68.1 52.3 78.5









#55
Dibenzenofuran
Concen: 38.311 ng
RT: 14.786 min Scan# 1989
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:168 Resp: 1701123
Ion Ratio Lower Upper
168 100
139 42.0 31.0 46.6
169 13.5 10.8 16.2



#56
4-Nitrophenol
Concen: 39.624 ng
RT: 14.604 min Scan# 1958
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31


Tgt Ion:139 Resp: 305758
Ion Ratio Lower Upper
139 100
109 82.3 72.3 112.3
65 111.5 72.5 112.5








#57
2,4-Dinitrotoluene
Concen: 39.478 ng
RT: 14.757 min Scan# 1984
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040







Tgt Ion:165 Resp: 424403
Ion Ratio Lower Upper
165 100
63 53.2 32.3 48.5#
89 89.6 55.9 83.9#
182 2.9 2.2 3.4



#58
Fluorene
Concen: 38.105 ng
RT: 15.439 min Scan# 2100
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:166 Resp: 1352370
Ion Ratio Lower Upper
166 100
165 99.6 78.0 117.0
167 13.3 10.5 15.7









#59
2,3,4,6-Tetrachlorophenol
Concen: 38.737 ng
RT: 15.010 min Scan# 2027
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:232 Resp: 316927
Ion Ratio Lower Upper
232 100
131 60.6 32.2 48.4#
130 3.3 2.0 3.0#
166 36.9 19.4 29.2#



#60
Diethylphthalate
Concen: 38.114 ng
RT: 15.228 min Scan# 2064
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:149 Resp: 1438123
Ion Ratio Lower Upper
149 100
177 20.7 17.5 26.3
150 12.6 9.9 14.9









#61
4-Chlorophenyl-phenylether
Concen: 37.955 ng
RT: 15.439 min Scan# 2100
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:204 Resp: 610038
Ion Ratio Lower Upper
204 100
206 32.6 26.6 40.0
141 71.5 48.6 73.0



#62
4-Nitroaniline
Concen: 39.986 ng
RT: 15.463 min Scan# 2104
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:138 Resp: 377065
Ion Ratio Lower Upper
138 100
92 59.9 31.2 71.2
108 88.7 78.6 118.6









#63
Azobenzene
Concen: 39.179 ng
RT: 15.733 min Scan# 2150
Delta R.T. 0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 1723513




Ion	Ratio	Lower	Upper
77	100		
182	19.3	6.0	46.0
105	18.7	0.0	38.8
51	24.5	8.8	48.8




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.139 min Scan# 2389
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion: 188 Resp: 964069

Ion	Ratio	Lower	Upper
188	100		
94	12.1	5.6	8.4#
80	13.8	6.3	9.5#









#65
4,6-Dinitro-2-methylphenol
Concen: 38.458 ng
RT: 15.522 min Scan# 2114
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:198 Resp: 226546
Ion Ratio Lower Upper
198 100
51 46.9 14.0 54.0
105 52.9 19.5 59.5



#66
n-Nitrosodiphenylamine
Concen: 38.745 ng
RT: 15.657 min Scan# 2137
Delta R.T. 0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:169 Resp: 1174057
Ion Ratio Lower Upper
169 100
168 68.2 54.4 81.6
167 36.8 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 38.404 ng
RT: 16.333 min Scan# 2252
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:248 Resp: 355244
Ion Ratio Lower Upper
248 100
250 97.9 79.8 119.8
141 92.9 57.8 86.6#



#68
Hexachlorobenzene
Concen: 38.167 ng
RT: 16.439 min Scan# 2270
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:284 Resp: 401768
Ion Ratio Lower Upper
284 100
142 45.6 26.9 40.3#
249 33.5 24.5 36.7









#69
Atrazine
Concen: 38.717 ng
RT: 16.616 min Scan# 2300
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:200 Resp: 366097
Ion Ratio Lower Upper
200 100
173 27.4 4.1 44.1
215 44.8 27.7 67.7



#70
Pentachlorophenol
Concen: 39.802 ng
RT: 16.786 min Scan# 2329
Delta R.T. 0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:266 Resp: 265830
Ion Ratio Lower Upper
266 100
268 65.8 51.6 77.4
264 62.2 50.3 75.5









#71
Phenanthrene
Concen: 38.523 ng
RT: 17.180 min Scan# 2396
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:178 Resp: 2094836
Ion Ratio Lower Upper
178 100
176 19.3 15.1 22.7
179 15.5 12.4 18.6



#72
Anthracene
Concen: 38.770 ng
RT: 17.275 min Scan# 2412
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:178 Resp: 2037201
Ion Ratio Lower Upper
178 100
176 18.6 14.9 22.3
179 15.4 12.2 18.2









#73
Carbazole
Concen: 38.873 ng
RT: 17.545 min Scan# 2458
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:167 Resp: 2083449
Ion Ratio Lower Upper
167 100
166 21.9 17.1 25.7
139 14.6 10.7 16.1



#74
Di-n-butylphthalate
Concen: 39.048 ng
RT: 18.116 min Scan# 2555
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:149 Resp: 2384354
Ion Ratio Lower Upper
149 100
150 9.1 7.3 10.9
104 5.8 4.5 6.7









#75
Fluoranthene
Concen: 38.235 ng
RT: 19.157 min Scan# 2732
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:202 Resp: 2333685
Ion Ratio Lower Upper
202 100
101 13.9 0.0 27.6
203 17.6 0.0 37.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.239 min Scan# 3086
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:240 Resp: 931327
Ion Ratio Lower Upper
240 100
120 13.3 6.5 9.7#
236 26.3 20.6 31.0









#77
Benzidine
Concen: 43.596 ng
RT: 19.345 min Scan# 2764
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:184 Resp: 1016000
Ion Ratio Lower Upper
184 100
185 14.6 11.0 16.4
183 12.3 9.5 14.3



#78
Pyrene
Concen: 38.955 ng
RT: 19.510 min Scan# 2792
Delta R.T. 0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:202 Resp: 2423325
Ion Ratio Lower Upper
202 100
200 21.2 16.3 24.5
203 17.2 14.2 21.2









#79
Terphenyl-d14
Concen: 76.686 ng
RT: 19.716 min Scan# 2827
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:244 Resp: 3564248
Ion Ratio Lower Upper
244 100
212 7.7 6.0 9.0
122 13.2 7.1 10.7#



#80
Butylbenzylphthalate
Concen: 39.780 ng
RT: 20.398 min Scan# 2943
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:149 Resp: 1088160
Ion Ratio Lower Upper
149 100
91 81.3 54.6 81.8
206 17.4 16.7 25.1









#81
Benzo(a)anthracene
Concen: 37.988 ng
RT: 21.221 min Scan# 3083
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 2312133
Ion Ratio Lower Upper
228 100
226 26.9 21.6 32.4
229 19.9 15.8 23.6



#82
3,3'-Dichlorobenzidine
Concen: 39.083 ng
RT: 21.163 min Scan# 3073
Delta R.T. -0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:252 Resp: 624503
Ion Ratio Lower Upper
252 100
254 64.3 50.6 76.0
126 13.9 7.0 10.6#









#83
Chrysene
Concen: 38.114 ng
RT: 21.274 min Scan# 3092
Delta R.T. 0.000 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 2248973
Ion Ratio Lower Upper
228 100
226 30.2 23.8 35.6
229 19.6 15.4 23.2



#84
Bis(2-ethylhexyl)phthalate
Concen: 38.410 ng
RT: 21.163 min Scan# 3073
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:149 Resp: 1578937
Ion Ratio Lower Upper
149 100
167 25.6 21.9 32.9
279 3.4 4.3 6.5#









#85
Di-n-octyl phthalate
Concen: 38.706 ng
RT: 22.068 min Scan# 3227
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:149 Resp: 2622529
Ion Ratio Lower Upper
149 100
167 1.4 1.4 2.0
43 9.9 5.6 8.4#



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.562 min Scan# 3481
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:264 Resp: 945877
Ion Ratio Lower Upper
264 100
260 23.6 18.6 28.0
265 22.1 17.6 26.4









#87
Indeno(1,2,3-cd)pyrene
Concen: 38.483 ng
RT: 25.968 min Scan# 3890
Delta R.T. -0.012 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:276 Resp: 2639431
Ion Ratio Lower Upper
276 100
138 31.3 15.3 22.9#
277 25.5 20.4 30.6



#88
Benzo(b)fluoranthene
Concen: 38.228 ng
RT: 22.857 min Scan# 3361
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:252 Resp: 2350071
Ion Ratio Lower Upper
252 100
253 22.1 17.5 26.3
125 13.4 6.2 9.4#









#89
Benzo(k)fluoranthene
Concen: 39.021 ng
RT: 22.904 min Scan# 3369
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:252 Resp: 2303169
Ion Ratio Lower Upper
252 100
253 22.2 17.6 26.4
125 13.5 6.0 9.0#



#90
Benzo(a)pyrene
Concen: 38.837 ng
RT: 23.462 min Scan# 3464
Delta R.T. -0.006 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:252 Resp: 2141458
Ion Ratio Lower Upper
252 100
253 22.1 17.9 26.9
125 14.2 6.5 9.7#









#91
Dibenzo(a,h)anthracene
Concen: 38.488 ng
RT: 25.998 min Scan# 3895
Delta R.T. -0.012 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 2282362
Ion Ratio Lower Upper
278 100
139 18.7 9.3 13.9#
279 23.1 18.8 28.2



#92
Benzo(g,h,i)perylene
Concen: 38.365 ng
RT: 26.709 min Scan# 4016
Delta R.T. -0.018 min
Lab File: BP006488.D
Acq: 04 Aug 2021 10:31

Tgt Ion:276 Resp: 2180147
Ion Ratio Lower Upper
276 100
277 23.9 18.9 28.3
138 27.4 12.6 18.8#

