

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080521\
 Data File : BP006528.D
 Acq On : 06 Aug 2021 03:43
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC

Quant Time: Aug 06 05:45:50 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 15:41:30 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.758	152	269693	20.000	ng	0.00
21) Naphthalene-d8	10.540	136	1108949	20.000	ng	# 0.00
39) Acenaphthene-d10	14.387	164	626545	20.000	ng	0.00
64) Phenanthrene-d10	17.134	188	1203453	20.000	ng	# 0.00
76) Chrysene-d12	21.233	240	1140202	20.000	ng	# 0.00
86) Perylene-d12	23.557	264	1171077	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.358	112	1361591	77.545	ng	0.00
7) Phenol-d6	6.940	99	1971978	79.061	ng	0.00
23) Nitrobenzene-d5	8.916	82	1919199	79.882	ng	0.00
42) 2,4,6-Tribromophenol	15.881	330	527875	80.615	ng	0.00
45) 2-Fluorobiphenyl	13.010	172	3274539	78.193	ng	0.00
79) Terphenyl-d14	19.710	244	4558155	80.105	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.305	88	280496	36.693	ng	91
3) Pyridine	3.699	79	843754	37.553	ng	95
4) n-Nitrosodimethylamine	3.611	42	315330	37.184	ng	# 82
6) Aniline	7.093	93	1073703	39.629	ng	94
8) 2-Chlorophenol	7.322	128	747993	39.371	ng	90
9) Benzaldehyde	6.905	77	593761	39.552	ng	92
10) Phenol	6.964	94	978400	39.119	ng	98
11) bis(2-Chloroethyl)ether	7.193	93	747888	39.381	ng	90
12) 1,3-Dichlorobenzene	7.646	146	765534	38.662	ng	96
13) 1,4-Dichlorobenzene	7.793	146	778830	38.756	ng	97
14) 1,2-Dichlorobenzene	8.105	146	742776	38.520	ng	95
15) Benzyl Alcohol	7.999	79	750068	40.100	ng	94
16) 2,2'-oxybis(1-Chloropr...	8.293	45	897992	39.736	ng	92
17) 2-Methylphenol	8.205	107	635222	40.222	ng	100
18) Hexachloroethane	8.822	117	313259	39.384	ng	# 88
19) n-Nitroso-di-n-propyla...	8.569	70	667977	39.917	ng	# 96
20) 3+4-Methylphenols	8.534	107	863086	40.151	ng	99
22) Acetophenone	8.581	105	1146614	38.989	ng	# 97
24) Nitrobenzene	8.958	77	982588	39.343	ng	92
25) Isophorone	9.481	82	1755370	40.678	ng	94
26) 2-Nitrophenol	9.658	139	394157	43.367	ng	91
27) 2,4-Dimethylphenol	9.728	122	602829	39.608	ng	97
28) bis(2-Chloroethoxy)met...	9.969	93	912977	39.542	ng	# 96
29) 2,4-Dichlorophenol	10.193	162	621387	40.326	ng	94
30) 1,2,4-Trichlorobenzene	10.405	180	647844	39.126	ng	97
31) Naphthalene	10.587	128	2306878	38.713	ng	100
32) Benzoic acid	9.881	122	498326	42.324	ng	92
33) 4-Chloroaniline	10.705	127	939983	38.981	ng	# 94
34) Hexachlorobutadiene	10.869	225	379127	39.293	ng	99
35) Caprolactam	11.516	113	229901	41.377	ng	# 72
36) 4-Chloro-3-methylphenol	11.834	107	787461	40.222	ng	89
37) 2-Methylnaphthalene	12.204	142	1569520	38.870	ng	99
38) 1-Methylnaphthalene	12.422	142	1497137	39.016	ng	99
40) 1,2,4,5-Tetrachloroben...	12.569	216	645366	39.355	ng	# 96
41) Hexachlorocyclopentadiene	12.546	237	323894	37.241	ng	100
43) 2,4,6-Trichlorophenol	12.816	196	468899	41.877	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080521\
 Data File : BP006528.D
 Acq On : 06 Aug 2021 03:43
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC

Quant Time: Aug 06 05:45:50 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 15:41:30 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.887	196	501102	40.430	ng	# 89
46) 1,1'-Biphenyl	13.222	154	1854092	39.096	ng	97
47) 2-Chloronaphthalene	13.257	162	1426794	39.069	ng	94
48) 2-Nitroaniline	13.469	65	536189	41.603	ng	86
49) Acenaphthylene	14.104	152	2314394	39.302	ng	100
50) Dimethylphthalate	13.857	163	1764464	38.688	ng	99
51) 2,6-Dinitrotoluene	13.969	165	406204	39.914	ng	# 77
52) Acenaphthene	14.451	154	1400889	39.089	ng	98
53) 3-Nitroaniline	14.298	138	454540	40.288	ng	81
54) 2,4-Dinitrophenol	14.504	184	204286	38.386	ng	# 77
55) Dibenzofuran	14.787	168	2159718	38.312	ng	94
56) 4-Nitrophenol	14.610	139	395173	40.338	ng	85
57) 2,4-Dinitrotoluene	14.757	165	559224	40.974	ng	# 80
58) Fluorene	15.434	166	1747615	38.786	ng	99
59) 2,3,4,6-Tetrachlorophenol	15.010	232	413978	39.856	ng	# 70
60) Diethylphthalate	15.222	149	1885253	39.356	ng	97
61) 4-Chlorophenyl-phenyle...	15.434	204	789088	38.671	ng	# 89
62) 4-Nitroaniline	15.469	138	483411	40.379	ng	89
63) Azobenzene	15.728	77	2225711	39.852	ng	92
65) 4,6-Dinitro-2-methylph...	15.522	198	294905	40.105	ng	75
66) n-Nitrosodiphenylamine	15.651	169	1510496	39.932	ng	99
67) 4-Bromophenyl-phenylether	16.328	248	468670	40.588	ng	# 85
68) Hexachlorobenzene	16.434	284	515834	39.256	ng	# 84
69) Atrazine	16.616	200	481595	40.801	ng	95
70) Pentachlorophenol	16.781	266	350727	42.068	ng	98
71) Phenanthrene	17.181	178	2651701	39.064	ng	99
72) Anthracene	17.269	178	2604614	39.709	ng	100
73) Carbazole	17.545	167	2630799	39.321	ng	98
74) Di-n-butylphthalate	18.110	149	3227853	42.346	ng	99
75) Fluoranthene	19.151	202	2949260	38.709	ng	93
77) Benzidine	19.339	184	1056802	37.040	ng	99
78) Pyrene	19.504	202	3056390	40.131	ng	99
80) Butylbenzylphthalate	20.392	149	1496510	44.686	ng	# 84
81) Benzo(a)anthracene	21.216	228	2917226	39.149	ng	100
82) 3,3'-Dichlorobenzidine	21.157	252	823147	42.078	ng	# 97
83) Chrysene	21.269	228	2780670	38.492	ng	98
84) Bis(2-ethylhexyl)phtha...	21.157	149	2218820	44.089	ng	# 96
85) Di-n-octyl phthalate	22.063	149	3751627	45.228	ng	# 94
87) Indeno(1,2,3-cd)pyrene	25.963	276	3337478	39.304	ng	# 89
88) Benzo(b)fluoranthene	22.851	252	2935950	38.575	ng	# 95
89) Benzo(k)fluoranthene	22.898	252	2859601	39.132	ng	# 96
90) Benzo(a)pyrene	23.457	252	2725473	39.924	ng	# 94
91) Dibenzo(a,h)anthracene	25.980	278	2870174	39.093	ng	# 92
92) Benzo(g,h,i)perylene	26.704	276	2739856	38.943	ng	# 90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080521\
 Data File : BP006528.D
 Acq On : 06 Aug 2021 03:43
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC

Quant Time: Aug 06 05:45:50 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 15:41:30 2021
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.758 min Scan# 794
Delta R.T. 0.001 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:152 Resp: 269693
Ion Ratio Lower Upper
152 100
150 156.7 125.6 188.4
115 65.4 48.5 72.7



#2
1,4-Dioxane
Concen: 36.693 ng
RT: 3.305 min Scan# 37
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion: 88 Resp: 280496
Ion Ratio Lower Upper
88 100
58 67.5 46.7 70.1
43 25.0 20.3 30.5









#3
Pyridine
Concen: 37.553 ng
RT: 3.699 min Scan# 104
Delta R.T. 0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 79 Resp: 843754
Ion Ratio Lower Upper
79 100
52 56.6 41.7 62.5
51 28.6 24.1 36.1



#4
n-Nitrosodimethylamine
Concen: 37.184 ng
RT: 3.611 min Scan# 89
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion: 42 Resp: 315330
Ion Ratio Lower Upper
42 100
74 153.5 105.4 158.0
44 10.4 4.2 6.4









#5
2-Fluorophenol
Concen: 77.545 ng
RT: 5.358 min Scan# 386
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:112 Resp: 1361591
Ion Ratio Lower Upper
112 100
64 61.7 41.4 62.2
63 31.8 23.8 35.8



#6
Aniline
Concen: 39.629 ng
RT: 7.093 min Scan# 681
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion: 93 Resp: 1073703
Ion Ratio Lower Upper
93 100
66 41.2 29.7 44.5
65 20.9 15.3 22.9









#7
Phenol-d6
Concen: 79.061 ng
RT: 6.940 min Scan# 655
Delta R.T. 0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 99 Resp: 1971978
Ion Ratio Lower Upper
99 100
42 16.8 13.4 20.0
71 36.4 30.6 46.0



#8
2-Chlorophenol
Concen: 39.371 ng
RT: 7.322 min Scan# 720
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion: 128 Resp: 747993
Ion Ratio Lower Upper
128 100
130 32.4 12.4 52.4
64 51.6 20.3 60.3









#9
Benzaldehyde
Concen: 39.552 ng
RT: 6.905 min Scan# 649
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 77 Resp: 593761
Ion Ratio Lower Upper
77 100
105 91.4 76.7 116.7
106 85.2 76.2 116.2



#10
Phenol
Concen: 39.119 ng
RT: 6.964 min Scan# 659
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion: 94 Resp: 978400
Ion Ratio Lower Upper
94 100
65 29.4 8.8 48.8
66 38.7 20.7 60.7









#11
bis(2-Chloroethyl)ether
Concen: 39.381 ng
RT: 7.193 min Scan# 698
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 93 Resp: 747888
Ion Ratio Lower Upper
93 100
63 77.8 46.2 86.2
95 32.0 12.2 52.2



#12
1,3-Dichlorobenzene
Concen: 38.662 ng
RT: 7.646 min Scan# 775
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:146 Resp: 765534
Ion Ratio Lower Upper
146 100
148 66.0 51.5 77.3
75 36.8 25.7 38.5









#13
1,4-Dichlorobenzene
Concen: 38.756 ng
RT: 7.793 min Scan# 800
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:146 Resp: 778830
Ion Ratio Lower Upper
146 100
148 64.4 51.3 76.9
111 45.1 33.0 49.6



#14
1,2-Dichlorobenzene
Concen: 38.520 ng
RT: 8.105 min Scan# 853
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:146 Resp: 742776
Ion Ratio Lower Upper
146 100
148 65.2 50.3 75.5
111 46.9 33.7 50.5









#15
Benzyl Alcohol
Concen: 40.100 ng
RT: 7.999 min Scan# 835
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 79 Resp: 750068
Ion Ratio Lower Upper
79 100
108 68.4 62.5 93.7
77 65.6 52.9 79.3



#16
2,2'-oxybis(1-Chloropropane)
Concen: 39.736 ng
RT: 8.293 min Scan# 885
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion: 45 Resp: 897992
Ion Ratio Lower Upper
45 100
77 18.0 2.0 42.0
79 13.7 0.0 36.7









#17
2-Methylphenol
Concen: 40.222 ng
RT: 8.205 min Scan# 870
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	107	Resp:	635222
Ion	Ratio	Lower	Upper
107	100		
108	109.5	87.7	131.5
77	50.3	40.2	60.2
79	48.5	39.0	58.4



#18
Hexachloroethane
Concen: 39.384 ng
RT: 8.822 min Scan# 975
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:	117	Resp:	313259
Ion	Ratio	Lower	Upper
117	100		
119	94.5	76.2	114.2
201	80.7	83.4	125.0









#19
n-Nitroso-di-n-propylamine
Concen: 39.917 ng
RT: 8.569 min Scan# 932
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	70	Resp:	667977
Ion	Ratio	Lower	Upper
70	100		
42	50.5	40.8	61.2
101	9.1	8.8	13.2
130	18.5	18.9	28.3



#20
3+4-Methylphenols
Concen: 40.151 ng
RT: 8.534 min Scan# 926
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:	107	Resp:	863086
Ion	Ratio	Lower	Upper
107	100		
108	89.8	71.0	111.0
77	37.8	16.6	56.6
79	33.0	12.7	52.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.540 min Scan# 1267
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:136 Resp: 1108949




Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.6	12.8
54	8.1	4.6	6.8#
68	7.2	3.6	5.4#




#22
Acetophenone
Concen: 38.989 ng
RT: 8.581 min Scan# 934
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:105 Resp: 1146614

Ion	Ratio	Lower	Upper
105	100		
71	5.5	1.9	2.9#
51	24.2	18.4	27.6
120	21.6	17.9	26.9









#23
Nitrobenzene-d5
Concen: 79.882 ng
RT: 8.916 min Scan# 991
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 82 Resp: 1919199
Ion Ratio Lower Upper
82 100
128 39.8 35.7 53.5
54 43.2 35.0 52.4



#24
Nitrobenzene
Concen: 39.343 ng
RT: 8.958 min Scan# 998
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion: 77 Resp: 982588
Ion Ratio Lower Upper
77 100
123 41.4 38.2 57.2
65 13.4 11.4 17.2









#25
Isophorone
Concen: 40.678 ng
RT: 9.481 min Scan# 1087
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion: 82 Resp: 1755370
Ion Ratio Lower Upper
82 100
95 7.1 6.5 9.7
138 15.5 15.0 22.4



#26
2-Nitrophenol
Concen: 43.367 ng
RT: 9.658 min Scan# 1117
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:139 Resp: 394157
Ion Ratio Lower Upper
139 100
109 34.1 28.8 43.2
65 59.8 41.0 61.6









#27
2,4-Dimethylphenol
Concen: 39.608 ng
RT: 9.728 min Scan# 1129
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:122 Resp: 602829
Ion Ratio Lower Upper
122 100
107 122.2 94.5 141.7
121 59.9 47.4 71.2



#28
bis(2-Chloroethoxy)methane
Concen: 39.542 ng
RT: 9.969 min Scan# 1170
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion: 93 Resp: 912977
Ion Ratio Lower Upper
93 100
95 31.1 25.9 38.9
123 10.1 10.2 15.2#









#29
2,4-Dichlorophenol
Concen: 40.326 ng
RT: 10.193 min Scan# 1208
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:162 Resp: 621387
Ion Ratio Lower Upper
162 100
164 63.7 44.9 84.9
98 43.3 14.3 54.3



#30
1,2,4-Trichlorobenzene
Concen: 39.126 ng
RT: 10.405 min Scan# 1244
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:180 Resp: 647844
Ion Ratio Lower Upper
180 100
182 96.1 77.9 116.9
145 34.7 24.7 37.1









#31
Naphthalene
Concen: 38.713 ng
RT: 10.587 min Scan# 1275
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:128 Resp: 2306878
Ion Ratio Lower Upper
128 100
129 10.9 9.0 13.6
127 13.4 10.7 16.1



#32
Benzoic acid
Concen: 42.324 ng
RT: 9.881 min Scan# 1155
Delta R.T. 0.024 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:122 Resp: 498326
Ion Ratio Lower Upper
122 100
105 129.8 103.2 143.2
77 105.9 76.2 116.2









#33
4-Chloroaniline
Concen: 38.981 ng
RT: 10.705 min Scan# 1295
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:	127	Resp:	939983
Ion	Ratio	Lower	Upper
127	100		
129	32.0	26.1	39.1
65	34.5	22.5	33.7#
92	21.0	15.0	22.6



#34
Hexachlorobutadiene
Concen: 39.293 ng
RT: 10.869 min Scan# 1323
Delta R.T. -0.012 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:	225	Resp:	379127
Ion	Ratio	Lower	Upper
225	100		
223	62.5	51.0	76.6
227	64.3	51.4	77.2









#35
Caprolactam
Concen: 41.377 ng
RT: 11.516 min Scan# 1433
Delta R.T. 0.012 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:113 Resp: 229901
Ion Ratio Lower Upper
113 100
55 151.3 98.9 138.9#
56 116.1 71.7 111.7#



#36
4-Chloro-3-methylphenol
Concen: 40.222 ng
RT: 11.834 min Scan# 1487
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:107 Resp: 787461
Ion Ratio Lower Upper
107 100
144 23.1 21.1 31.7
142 71.1 66.0 99.0









#37
2-Methylnaphthalene
Concen: 38.870 ng
RT: 12.204 min Scan# 1550
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:142 Resp: 1569520
Ion Ratio Lower Upper
142 100
141 91.4 71.8 107.6
115 39.0 31.6 47.4



#38
1-Methylnaphthalene
Concen: 39.016 ng
RT: 12.422 min Scan# 1587
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:142 Resp: 1497137
Ion Ratio Lower Upper
142 100
141 92.4 73.3 109.9
116 4.2 3.5 5.3









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.387 min Scan# 1921
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:164 Resp: 626545
Ion Ratio Lower Upper
164 100
162 102.9 80.2 120.2
160 45.0 34.5 51.7



#40
1,2,4,5-Tetrachlorobenzene
Concen: 39.355 ng
RT: 12.569 min Scan# 1612
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:216 Resp: 645366
Ion Ratio Lower Upper
216 100
214 78.5 61.9 92.9
179 23.8 16.8 25.2
108 25.0 14.9 22.3#









#41
Hexachlorocyclopentadiene
Concen: 37.241 ng
RT: 12.546 min Scan# 1608
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:237 Resp: 323894
Ion Ratio Lower Upper
237 100
235 64.3 44.5 84.5
272 12.0 0.0 32.5



#42
2,4,6-Tribromophenol
Concen: 80.615 ng
RT: 15.881 min Scan# 2175
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:330 Resp: 527875
Ion Ratio Lower Upper
330 100
332 97.6 77.8 116.8
141 48.8 25.4 38.2#









#43
2,4,6-Trichlorophenol
Concen: 41.877 ng
RT: 12.816 min Scan# 1654
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:196 Resp: 468899
Ion Ratio Lower Upper
196 100
198 95.5 79.5 119.3
200 30.2 25.0 37.6



#44
2,4,5-Trichlorophenol
Concen: 40.430 ng
RT: 12.887 min Scan# 1666
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:196 Resp: 501102
Ion Ratio Lower Upper
196 100
198 95.4 74.7 112.1
97 63.5 37.4 56.0#
132 32.9 20.0 30.0#









#45
2-Fluorobiphenyl
Concen: 78.193 ng
RT: 13.010 min Scan# 1687
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:172 Resp: 3274539
Ion Ratio Lower Upper
172 100
171 35.9 27.9 41.9
170 23.8 18.6 27.8



#46
1,1'-Biphenyl
Concen: 39.096 ng
RT: 13.222 min Scan# 1723
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:154 Resp: 1854092
Ion Ratio Lower Upper
154 100
153 40.6 20.4 60.4
76 15.9 0.0 30.8









#47
2-Chloronaphthalene
Concen: 39.069 ng
RT: 13.257 min Scan# 1729
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:162 Resp: 1426794
Ion Ratio Lower Upper
162 100
127 44.7 31.4 47.0
164 32.2 26.6 39.8



#48
2-Nitroaniline
Concen: 41.603 ng
RT: 13.469 min Scan# 1765
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion: 65 Resp: 536189
Ion Ratio Lower Upper
65 100
92 68.7 57.7 86.5
138 92.7 92.1 138.1









#49
Acenaphthylene
Concen: 39.302 ng
RT: 14.104 min Scan# 1873
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:152 Resp: 2314394
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.2
153 13.3 10.6 16.0



#50
Dimethylphthalate
Concen: 38.688 ng
RT: 13.857 min Scan# 1831
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:163 Resp: 1764464
Ion Ratio Lower Upper
163 100
194 3.7 3.4 5.2
164 10.1 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 39.914 ng
RT: 13.969 min Scan# 1850
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:165 Resp: 406204
Ion Ratio Lower Upper
165 100
63 70.7 41.3 61.9#
89 70.4 44.8 67.2#



#52
Acenaphthene
Concen: 39.089 ng
RT: 14.451 min Scan# 1932
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:154 Resp: 1400889
Ion Ratio Lower Upper
154 100
153 114.2 89.4 134.2
152 54.0 43.0 64.4









#53
3-Nitroaniline
Concen: 40.288 ng
RT: 14.298 min Scan# 1906
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:138 Resp: 454540
Ion Ratio Lower Upper
138 100
108 11.3 8.4 12.6
92 131.0 87.8 131.8



#54
2,4-Dinitrophenol
Concen: 38.386 ng
RT: 14.504 min Scan# 1941
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:184 Resp: 204286
Ion Ratio Lower Upper
184 100
63 86.9 47.0 70.6#
154 73.1 52.3 78.5









#55
Dibenzenofuran
Concen: 38.312 ng
RT: 14.787 min Scan# 1989
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:168 Resp: 2159718
Ion Ratio Lower Upper
168 100
139 43.6 31.0 46.6
169 13.1 10.8 16.2



#56
4-Nitrophenol
Concen: 40.338 ng
RT: 14.610 min Scan# 1959
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43


Tgt Ion:139 Resp: 395173
Ion Ratio Lower Upper
139 100
109 83.6 72.3 112.3
65 112.4 72.5 112.5








#57
2,4-Dinitrotoluene
Concen: 40.974 ng
RT: 14.757 min Scan# 1984
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC







Tgt Ion:165 Resp: 559224
Ion Ratio Lower Upper
165 100
63 51.9 32.3 48.5#
89 88.1 55.9 83.9#
182 2.8 2.2 3.4



#58
Fluorene
Concen: 38.786 ng
RT: 15.434 min Scan# 2099
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:166 Resp: 1747615
Ion Ratio Lower Upper
166 100
165 98.6 78.0 117.0
167 13.7 10.5 15.7









#59
2,3,4,6-Tetrachlorophenol
Concen: 39.856 ng
RT: 15.010 min Scan# 2027
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:232 Resp: 413978
Ion Ratio Lower Upper
232 100
131 61.1 32.2 48.4#
130 3.2 2.0 3.0#
166 37.0 19.4 29.2#



#60
Diethylphthalate
Concen: 39.356 ng
RT: 15.222 min Scan# 2063
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:149 Resp: 1885253
Ion Ratio Lower Upper
149 100
177 19.7 17.5 26.3
150 12.8 9.9 14.9









#61
4-Chlorophenyl-phenylether
Concen: 38.671 ng
RT: 15.434 min Scan# 2099
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:204 Resp: 789088
Ion Ratio Lower Upper
204 100
206 32.6 26.6 40.0
141 73.2 48.6 73.0#



#62
4-Nitroaniline
Concen: 40.379 ng
RT: 15.469 min Scan# 2105
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43


Tgt Ion:138 Resp: 483411
Ion Ratio Lower Upper
138 100
92 57.9 31.2 71.2
108 87.4 78.6 118.6








#63
Azobenzene
Concen: 39.852 ng
RT: 15.728 min Scan# 2149
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC







Tgt Ion:	77	Resp:	2225711
Ion	Ratio	Lower	Upper
77	100		
182	19.3	6.0	46.0
105	18.6	0.0	38.8
51	24.9	8.8	48.8



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.134 min Scan# 2388
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:	188	Resp:	1203453
Ion	Ratio	Lower	Upper
188	100		
94	12.3	5.6	8.4#
80	14.8	6.3	9.5#









#65
4,6-Dinitro-2-methylphenol
Concen: 40.105 ng
RT: 15.522 min Scan# 2114
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:198 Resp: 294905
Ion Ratio Lower Upper
198 100
51 47.9 14.0 54.0
105 54.8 19.5 59.5



#66
n-Nitrosodiphenylamine
Concen: 39.932 ng
RT: 15.651 min Scan# 2136
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:169 Resp: 1510496
Ion Ratio Lower Upper
169 100
168 67.6 54.4 81.6
167 36.6 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 40.588 ng
RT: 16.328 min Scan# 2251
Delta R.T. -0.012 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:248 Resp: 468670
Ion Ratio Lower Upper
248 100
250 96.3 79.8 119.8
141 97.0 57.8 86.6#



#68
Hexachlorobenzene
Concen: 39.256 ng
RT: 16.434 min Scan# 2269
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:284 Resp: 515834
Ion Ratio Lower Upper
284 100
142 48.0 26.9 40.3#
249 33.9 24.5 36.7









#69
Atrazine
Concen: 40.801 ng
RT: 16.616 min Scan# 2300
Delta R.T. 0.000 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:200 Resp: 481595
Ion Ratio Lower Upper
200 100
173 27.2 4.1 44.1
215 44.9 27.7 67.7



#70
Pentachlorophenol
Concen: 42.068 ng
RT: 16.781 min Scan# 2328
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:266 Resp: 350727
Ion Ratio Lower Upper
266 100
268 63.6 51.6 77.4
264 60.9 50.3 75.5









#71
Phenanthrene
Concen: 39.064 ng
RT: 17.181 min Scan# 2396
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:178 Resp: 2651701
Ion Ratio Lower Upper
178 100
176 19.3 15.1 22.7
179 15.4 12.4 18.6



#72
Anthracene
Concen: 39.709 ng
RT: 17.269 min Scan# 2411
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:178 Resp: 2604614
Ion Ratio Lower Upper
178 100
176 18.7 14.9 22.3
179 15.3 12.2 18.2









#73
 Carbazole
 Concen: 39.321 ng
 RT: 17.545 min Scan# 2458
 Delta R.T. -0.006 min
 Lab File: BP006528.D
 Acq: 06 Aug 2021 03:43




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040EC


Tgt Ion:167 Resp: 2630799
 Ion Ratio Lower Upper
 167 100
 166 22.1 17.1 25.7
 139 14.9 10.7 16.1



#74
 Di-n-butylphthalate
 Concen: 42.346 ng
 RT: 18.110 min Scan# 2554
 Delta R.T. -0.006 min
 Lab File: BP006528.D
 Acq: 06 Aug 2021 03:43

Tgt Ion:149 Resp: 3227853
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.3 10.9
 104 5.8 4.5 6.7









#75
Fluoranthene
Concen: 38.709 ng
RT: 19.151 min Scan# 2731
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:202 Resp: 2949260
Ion Ratio Lower Upper
202 100
101 15.2 0.0 27.6
203 17.8 0.0 37.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.233 min Scan# 3085
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:240 Resp: 1140202
Ion Ratio Lower Upper
240 100
120 14.2 6.5 9.7#
236 25.8 20.6 31.0









#77
Benzidine
Concen: 37.040 ng
RT: 19.339 min Scan# 2763
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:184 Resp: 1056802
Ion Ratio Lower Upper
184 100
185 14.4 11.0 16.4
183 12.2 9.5 14.3



#78
Pyrene
Concen: 40.131 ng
RT: 19.504 min Scan# 2791
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:202 Resp: 3056390
Ion Ratio Lower Upper
202 100
200 21.1 16.3 24.5
203 17.4 14.2 21.2









#79
Terphenyl-d14
Concen: 80.105 ng
RT: 19.710 min Scan# 2826
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:244 Resp: 4558155
Ion Ratio Lower Upper
244 100
212 7.9 6.0 9.0
122 14.0 7.1 10.7#



#80
Butylbenzylphthalate
Concen: 44.686 ng
RT: 20.392 min Scan# 2942
Delta R.T. -0.012 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:149 Resp: 1496510
Ion Ratio Lower Upper
149 100
91 83.3 54.6 81.8#
206 17.4 16.7 25.1









#81
Benzo(a)anthracene
Concen: 39.149 ng
RT: 21.216 min Scan# 3082
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:228 Resp: 2917226
Ion Ratio Lower Upper
228 100
226 27.2 21.6 32.4
229 19.8 15.8 23.6



#82
3,3'-Dichlorobenzidine
Concen: 42.078 ng
RT: 21.157 min Scan# 3072
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:252 Resp: 823147
Ion Ratio Lower Upper
252 100
254 64.3 50.6 76.0
126 14.1 7.0 10.6#









#83
Chrysene
Concen: 38.492 ng
RT: 21.269 min Scan# 3091
Delta R.T. -0.006 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:228 Resp: 2780670
Ion Ratio Lower Upper
228 100
226 31.0 23.8 35.6
229 19.9 15.4 23.2



#84
Bis(2-ethylhexyl)phthalate
Concen: 44.089 ng
RT: 21.157 min Scan# 3072
Delta R.T. -0.012 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:149 Resp: 2218820
Ion Ratio Lower Upper
149 100
167 25.6 21.9 32.9
279 3.6 4.3 6.5#









#85
Di-n-octyl phthalate
Concen: 45.228 ng
RT: 22.063 min Scan# 3226
Delta R.T. -0.012 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:149 Resp: 3751627
Ion Ratio Lower Upper
149 100
167 1.5 1.4 2.0
43 9.6 5.6 8.4#



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.557 min Scan# 3480
Delta R.T. -0.012 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:264 Resp: 1171077
Ion Ratio Lower Upper
264 100
260 24.1 18.6 28.0
265 21.6 17.6 26.4









#87
Indeno(1,2,3-cd)pyrene
Concen: 39.304 ng
RT: 25.963 min Scan# 3889
Delta R.T. -0.017 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:276 Resp: 3337478
Ion Ratio Lower Upper
276 100
138 31.0 15.3 22.9#
277 25.3 20.4 30.6



#88
Benzo(b)fluoranthene
Concen: 38.575 ng
RT: 22.851 min Scan# 3360
Delta R.T. -0.012 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:252 Resp: 2935950
Ion Ratio Lower Upper
252 100
253 22.4 17.5 26.3
125 13.6 6.2 9.4#









#89
Benzo(k)fluoranthene
Concen: 39.132 ng
RT: 22.898 min Scan# 3368
Delta R.T. -0.012 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC


Tgt Ion:252 Resp: 2859601
Ion Ratio Lower Upper
252 100
253 21.9 17.6 26.4
125 13.2 6.0 9.0#



#90
Benzo(a)pyrene
Concen: 39.924 ng
RT: 23.457 min Scan# 3463
Delta R.T. -0.012 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:252 Resp: 2725473
Ion Ratio Lower Upper
252 100
253 21.7 17.9 26.9
125 14.6 6.5 9.7#









#91
Dibenzo(a,h)anthracene
Concen: 39.093 ng
RT: 25.980 min Scan# 3892
Delta R.T. -0.029 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040EC

Tgt Ion:278 Resp: 2870174
Ion Ratio Lower Upper
278 100
139 19.8 9.3 13.9#
279 24.0 18.8 28.2



#92
Benzo(g,h,i)perylene
Concen: 38.943 ng
RT: 26.704 min Scan# 4015
Delta R.T. -0.023 min
Lab File: BP006528.D
Acq: 06 Aug 2021 03:43

Tgt Ion:276 Resp: 2739856
Ion Ratio Lower Upper
276 100
277 23.5 18.9 28.3
138 26.0 12.6 18.8#

