

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080521\
 Data File : BP006508.D
 Acq On : 05 Aug 2021 16:21
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 05 17:49:36 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 15:41:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	135	0.00
2	1,4-Dioxane	0.567	0.522	7.9	125	0.00
3	Pyridine	1.666	1.573	5.6	121	0.00
4	n-Nitrosodimethylamine	0.629	0.576	8.4	118	0.00
5 S	2-Fluorophenol	1.302	1.267	2.7	127	0.00
6	Aniline	2.009	1.963	2.3	120	0.00
7 S	Phenol-d6	1.850	1.805	2.4	123	0.00
8	2-Chlorophenol	1.409	1.390	1.3	126	0.00
9	Benzaldehyde	1.113	1.066	4.2	126	0.00
10 C	Phenol	1.855	1.789	3.6	121	0.00
11	bis(2-Chloroethyl)ether	1.408	1.341	4.8	121	0.00
12	1,3-Dichlorobenzene	1.468	1.409	4.0	125	0.00
13 C	1,4-Dichlorobenzene	1.490	1.417	4.9	124	0.00
14	1,2-Dichlorobenzene	1.430	1.362	4.8	122	0.00
15	Benzyl Alcohol	1.387	1.358	2.1	120	0.00
16	2,2'-oxybis(1-Chloropropane	1.676	1.545	7.8	115	0.00
17	2-Methylphenol	1.171	1.155	1.4	122	0.00
18	Hexachloroethane	0.590	0.553	6.3	123	0.00
19 P	n-Nitroso-di-n-propylamine	1.241	1.185	4.5	115	0.00
20	3+4-Methylphenols	1.594	1.586	0.5	121	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	127	0.00
22	Acetophenone	0.530	0.506	4.5	116	0.00
23 S	Nitrobenzene-d5	0.433	0.411	5.1	116	0.00
24	Nitrobenzene	0.450	0.424	5.8	116	0.00
25	Isophorone	0.778	0.767	1.4	118	0.00
26 C	2-Nitrophenol	0.164	0.179	-9.1	130	0.00
27	2,4-Dimethylphenol	0.274	0.271	1.1	120	0.00
28	bis(2-Chloroethoxy)methane	0.416	0.393	5.5	116	0.00
29 C	2,4-Dichlorophenol	0.278	0.278	0.0	119	0.00
30	1,2,4-Trichlorobenzene	0.299	0.284	5.0	120	0.00
31	Naphthalene	1.075	1.018	5.3	117	0.00
32	Benzoic acid	0.212	0.243	-14.6	139	0.01
33	4-Chloroaniline	0.435	0.421	3.2	116	0.00
34 C	Hexachlorobutadiene	0.174	0.167	4.0	122	0.00
35	Caprolactam	0.100	0.107	-7.0	119	0.00
36 C	4-Chloro-3-methylphenol	0.353	0.349	1.1	115	0.00
37	2-Methylnaphthalene	0.728	0.693	4.8	115	0.00
38	1-Methylnaphthalene	0.692	0.654	5.5	115	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	121	0.00
40	1,2,4,5-Tetrachlorobenzene	0.523	0.503	3.8	116	0.00
41 P	Hexachlorocyclopentadiene	0.278	0.278	0.0	118	0.00
42 S	2,4,6-Tribromophenol	0.209	0.217	-3.8	118	0.00
43 C	2,4,6-Trichlorophenol	0.357	0.371	-3.9	120	0.00
44	2,4,5-Trichlorophenol	0.396	0.401	-1.3	117	0.00
45 S	2-Fluorobiphenyl	1.337	1.277	4.5	115	0.00
46	1,1'-Biphenyl	1.514	1.444	4.6	114	0.00
47	2-Chloronaphthalene	1.166	1.107	5.1	113	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080521\
 Data File : BP006508.D
 Acq On : 05 Aug 2021 16:21
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 05 17:49:36 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 15:41:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.411	0.418	-1.7	113	0.00
49	Acenaphthylene	1.880	1.829	2.7	114	0.00
50	Dimethylphthalate	1.456	1.406	3.4	113	0.00
51	2,6-Dinitrotoluene	0.325	0.323	0.6	113	0.00
52 C	Acenaphthene	1.144	1.104	3.5	113	0.00
53	3-Nitroaniline	0.360	0.366	-1.7	113	0.00
54 P	2,4-Dinitrophenol	0.170	0.187	-10.0	127	0.00
55	Dibenzofuran	1.799	1.695	5.8	111	0.00
56 P	4-Nitrophenol	0.313	0.329	-5.1	115	0.00
57	2,4-Dinitrotoluene	0.436	0.450	-3.2	115	0.00
58	Fluorene	1.438	1.386	3.6	112	0.00
59	2,3,4,6-Tetrachlorophenol	0.332	0.335	-0.9	116	0.00
60	Diethylphthalate	1.529	1.496	2.2	113	0.00
61	4-Chlorophenyl-phenylether	0.651	0.621	4.6	112	0.00
62	4-Nitroaniline	0.382	0.397	-3.9	113	0.00
63	Azobenzene	1.783	1.704	4.4	109	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	120	0.00
65	4,6-Dinitro-2-methylphenol	0.122	0.128	-4.9	121	0.00
66 c	n-Nitrosodiphenylamine	0.629	0.601	4.5	112	0.00
67	4-Bromophenyl-phenylether	0.192	0.186	3.1	115	0.00
68	Hexachlorobenzene	0.218	0.208	4.6	114	0.00
69	Atrazine	0.196	0.201	-2.6	116	0.00
70 C	Pentachlorophenol	0.139	0.148	-6.5	120	0.00
71	Phenanthrene	1.128	1.066	5.5	111	0.00
72	Anthracene	1.090	1.057	3.0	113	0.00
73	Carbazole	1.112	1.082	2.7	111	0.00
74	Di-n-butylphthalate	1.267	1.329	-4.9	120	0.00
75 C	Fluoranthene	1.266	1.224	3.3	111	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	117	0.00
77	Benzydine	0.500	0.472	5.6	94	0.00
78	Pyrene	1.336	1.298	2.8	110	0.00
79 S	Terphenyl-d14	0.998	0.968	3.0	112	0.00
80	Butylbenzylphthalate	0.587	0.649	-10.6	121	0.00
81	Benzo(a)anthracene	1.307	1.245	4.7	108	0.00
82	3,3'-Dichlorobenzidine	0.343	0.366	-6.7	118	0.00
83	Chrysene	1.267	1.208	4.7	109	0.00
84	Bis(2-ethylhexyl)phthalate	0.883	0.947	-7.2	120	0.00
85 c	Di-n-octyl phthalate	1.455	1.659	-14.0	127	0.00
86 I	Perylene-d12	1.000	1.000	0.0	122	0.00
87	Indeno(1,2,3-cd)pyrene	1.450	1.380	4.8	111	0.00
88	Benzo(b)fluoranthene	1.300	1.205	7.3	109	0.00
89	Benzo(k)fluoranthene	1.248	1.177	5.7	109	0.00
90 C	Benzo(a)pyrene	1.166	1.129	3.2	112	0.00
91	Dibenzo(a,h)anthracene	1.254	1.190	5.1	111	0.00
92	Benzo(g,h,i)perylene	1.202	1.146	4.7	111	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080521\
Data File : BP006508.D
Acq On : 05 Aug 2021 16:21
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Aug 05 17:49:36 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 03 15:41:30 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0