

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080521\
 Data File : BP006508.D
 Acq On : 05 Aug 2021 16:21
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 05 17:49:36 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 15:41:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	135	0.00
2	1,4-Dioxane	40.000	36.860	7.9	125	0.00
3	Pyridine	40.000	37.768	5.6	121	0.00
4	n-Nitrosodimethylamine	40.000	36.613	8.5	118	0.00
5 S	2-Fluorophenol	80.000	77.821	2.7	127	0.00
6	Aniline	40.000	39.078	2.3	120	0.00
7 S	Phenol-d6	80.000	78.052	2.4	123	0.00
8	2-Chlorophenol	40.000	39.476	1.3	126	0.00
9	Benzaldehyde	40.000	38.291	4.3	126	0.00
10 C	Phenol	40.000	38.577	3.6	121	0.00
11	bis(2-Chloroethyl)ether	40.000	38.101	4.7	121	0.00
12	1,3-Dichlorobenzene	40.000	38.392	4.0	125	0.00
13 C	1,4-Dichlorobenzene	40.000	38.033	4.9	124	0.00
14	1,2-Dichlorobenzene	40.000	38.099	4.8	122	0.00
15	Benzyl Alcohol	40.000	39.168	2.1	120	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.875	7.8	115	0.00
17	2-Methylphenol	40.000	39.454	1.4	122	0.00
18	Hexachloroethane	40.000	37.472	6.3	123	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.200	4.5	115	0.00
20	3+4-Methylphenols	40.000	39.800	0.5	121	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	127	0.00
22	Acetophenone	40.000	38.131	4.7	116	0.00
23 S	Nitrobenzene-d5	80.000	75.896	5.1	116	0.00
24	Nitrobenzene	40.000	37.645	5.9	116	0.00
25	Isophorone	40.000	39.415	1.5	118	0.00
26 C	2-Nitrophenol	40.000	43.667	-9.2	130	0.00
27	2,4-Dimethylphenol	40.000	39.519	1.2	120	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.769	5.6	116	0.00
29 C	2,4-Dichlorophenol	40.000	39.992	0.0	119	0.00
30	1,2,4-Trichlorobenzene	40.000	38.101	4.7	120	0.00
31	Naphthalene	40.000	37.903	5.2	117	0.00
32	Benzoic acid	40.000	45.831	-14.6	139	0.01
33	4-Chloroaniline	40.000	38.745	3.1	116	0.00
34 C	Hexachlorobutadiene	40.000	38.397	4.0	122	0.00
35	Caprolactam	40.000	42.866	-7.2	119	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.503	1.2	115	0.00
37	2-Methylnaphthalene	40.000	38.073	4.8	115	0.00
38	1-Methylnaphthalene	40.000	37.798	5.5	115	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	121	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.404	4.0	116	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.014	-0.0	118	0.00
42 S	2,4,6-Tribromophenol	80.000	83.034	-3.8	118	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.542	-3.9	120	0.00
44	2,4,5-Trichlorophenol	40.000	40.580	-1.4	117	0.00
45 S	2-Fluorobiphenyl	80.000	76.413	4.5	115	0.00
46	1,1'-Biphenyl	40.000	38.159	4.6	114	0.00
47	2-Chloronaphthalene	40.000	37.982	5.0	113	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080521\
 Data File : BP006508.D
 Acq On : 05 Aug 2021 16:21
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 05 17:49:36 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 15:41:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.659	-1.6	113	0.00
49	Acenaphthylene	40.000	38.916	2.7	114	0.00
50	Dimethylphthalate	40.000	38.626	3.4	113	0.00
51	2,6-Dinitrotoluene	40.000	39.830	0.4	113	0.00
52 C	Acenaphthene	40.000	38.599	3.5	113	0.00
53	3-Nitroaniline	40.000	40.625	-1.6	113	0.00
54 P	2,4-Dinitrophenol	40.000	43.923	-9.8	127	0.00
55	Dibenzofuran	40.000	37.682	5.8	111	0.00
56 P	4-Nitrophenol	40.000	42.143	-5.4	115	0.00
57	2,4-Dinitrotoluene	40.000	41.314	-3.3	115	0.00
58	Fluorene	40.000	38.537	3.7	112	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.381	-1.0	116	0.00
60	Diethylphthalate	40.000	39.135	2.2	113	0.00
61	4-Chlorophenyl-phenylether	40.000	38.109	4.7	112	0.00
62	4-Nitroaniline	40.000	41.524	-3.8	113	0.00
63	Azobenzene	40.000	38.223	4.4	109	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	120	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.881	-4.7	121	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.221	4.4	112	0.00
67	4-Bromophenyl-phenylether	40.000	38.767	3.1	115	0.00
68	Hexachlorobenzene	40.000	38.043	4.9	114	0.00
69	Atrazine	40.000	40.956	-2.4	116	0.00
70 C	Pentachlorophenol	40.000	42.771	-6.9	120	0.00
71	Phenanthrene	40.000	37.799	5.5	111	0.00
72	Anthracene	40.000	38.789	3.0	113	0.00
73	Carbazole	40.000	38.909	2.7	111	0.00
74	Di-n-butylphthalate	40.000	41.969	-4.9	120	0.00
75 C	Fluoranthene	40.000	38.666	3.3	111	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	117	0.00
77	Benzydine	40.000	37.710	5.7	94	0.00
78	Pyrene	40.000	38.851	2.9	110	0.00
79 S	Terphenyl-d14	80.000	77.565	3.0	112	0.00
80	Butylbenzylphthalate	40.000	44.222	-10.6	121	0.00
81	Benzo(a)anthracene	40.000	38.103	4.7	108	0.00
82	3,3'-Dichlorobenzidine	40.000	42.614	-6.5	118	0.00
83	Chrysene	40.000	38.125	4.7	109	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.910	-7.3	120	0.00
85 c	Di-n-octyl phthalate	40.000	45.599	-14.0	127	0.00
86 I	Perylene-d12	20.000	20.000	0.0	122	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	38.063	4.8	111	0.00
88	Benzo(b)fluoranthene	40.000	37.094	7.3	109	0.00
89	Benzo(k)fluoranthene	40.000	37.734	5.7	109	0.00
90 C	Benzo(a)pyrene	40.000	38.741	3.1	112	0.00
91	Dibenzo(a,h)anthracene	40.000	37.949	5.1	111	0.00
92	Benzo(g,h,i)perylene	40.000	38.137	4.7	111	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080521\
Data File : BP006508.D
Acq On : 05 Aug 2021 16:21
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Aug 05 17:49:36 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 03 15:41:30 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0