

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080521\
 Data File : BP006510.D
 Acq On : 05 Aug 2021 17:30
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 05 18:02:46 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 15:41:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	145	0.00
2	1,4-Dioxane	40.000	35.482	11.3	130	0.00
3	Pyridine	40.000	36.683	8.3	127	0.00
4	n-Nitrosodimethylamine	40.000	36.036	9.9	126	0.00
5 S	2-Fluorophenol	80.000	76.753	4.1	135	0.00
6	Aniline	40.000	38.925	2.7	129	0.00
7 S	Phenol-d6	80.000	77.548	3.1	131	0.00
8	2-Chlorophenol	40.000	39.222	1.9	135	0.00
9	Benzaldehyde	40.000	38.539	3.7	137	0.00
10 C	Phenol	40.000	38.504	3.7	130	0.00
11	bis(2-Chloroethyl)ether	40.000	38.455	3.9	132	0.00
12	1,3-Dichlorobenzene	40.000	38.623	3.4	136	0.00
13 C	1,4-Dichlorobenzene	40.000	38.131	4.7	134	0.00
14	1,2-Dichlorobenzene	40.000	38.329	4.2	133	0.00
15	Benzyl Alcohol	40.000	39.555	1.1	130	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	37.959	5.1	128	-0.01
17	2-Methylphenol	40.000	39.553	1.1	131	0.00
18	Hexachloroethane	40.000	38.446	3.9	136	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.098	2.3	127	0.00
20	3+4-Methylphenols	40.000	39.880	0.3	130	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	139	0.00
22	Acetophenone	40.000	38.056	4.9	127	0.00
23 S	Nitrobenzene-d5	80.000	75.847	5.2	127	0.00
24	Nitrobenzene	40.000	37.710	5.7	127	0.00
25	Isophorone	40.000	39.735	0.7	130	0.00
26 C	2-Nitrophenol	40.000	42.820	-7.1	140	0.00
27	2,4-Dimethylphenol	40.000	39.322	1.7	130	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.445	3.9	129	0.00
29 C	2,4-Dichlorophenol	40.000	40.576	-1.4	132	0.00
30	1,2,4-Trichlorobenzene	40.000	38.399	4.0	133	0.00
31	Naphthalene	40.000	38.097	4.8	129	0.00
32	Benzoic acid	40.000	43.624	-9.1	145	0.02
33	4-Chloroaniline	40.000	38.663	3.3	126	0.00
34 C	Hexachlorobutadiene	40.000	38.788	3.0	135	0.00
35	Caprolactam	40.000	41.469	-3.7	126	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.677	0.8	127	0.00
37	2-Methylnaphthalene	40.000	38.353	4.1	127	0.00
38	1-Methylnaphthalene	40.000	38.184	4.5	127	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	133	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.708	3.2	129	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.887	2.8	127	0.00
42 S	2,4,6-Tribromophenol	80.000	81.848	-2.3	128	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.368	-3.4	132	0.00
44	2,4,5-Trichlorophenol	40.000	40.250	-0.6	128	0.00
45 S	2-Fluorobiphenyl	80.000	76.085	4.9	126	0.00
46	1,1'-Biphenyl	40.000	38.096	4.8	126	0.00
47	2-Chloronaphthalene	40.000	38.139	4.7	125	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080521\
 Data File : BP006510.D
 Acq On : 05 Aug 2021 17:30
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 05 18:02:46 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 15:41:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.140	-0.4	123	0.00
49	Acenaphthylene	40.000	38.641	3.4	124	0.00
50	Dimethylphthalate	40.000	38.182	4.5	123	0.00
51	2,6-Dinitrotoluene	40.000	38.765	3.1	121	0.00
52 C	Acenaphthene	40.000	38.167	4.6	123	0.00
53	3-Nitroaniline	40.000	39.510	1.2	122	0.00
54 P	2,4-Dinitrophenol	40.000	38.185	4.5	121	0.00
55	Dibenzofuran	40.000	37.345	6.6	121	0.00
56 P	4-Nitrophenol	40.000	39.926	0.2	120	0.00
57	2,4-Dinitrotoluene	40.000	40.381	-1.0	124	0.00
58	Fluorene	40.000	37.761	5.6	121	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.279	-0.7	127	0.00
60	Diethylphthalate	40.000	38.456	3.9	123	0.00
61	4-Chlorophenyl-phenylether	40.000	37.816	5.5	123	0.00
62	4-Nitroaniline	40.000	39.430	1.4	119	0.00
63	Azobenzene	40.000	38.393	4.0	120	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	128	0.00
65	4,6-Dinitro-2-methylphenol	40.000	38.406	4.0	118	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.782	3.0	121	0.00
67	4-Bromophenyl-phenylether	40.000	39.728	0.7	126	0.00
68	Hexachlorobenzene	40.000	38.274	4.3	123	0.00
69	Atrazine	40.000	40.430	-1.1	122	0.00
70 C	Pentachlorophenol	40.000	41.692	-4.2	125	0.00
71	Phenanthrene	40.000	37.801	5.5	118	0.00
72	Anthracene	40.000	38.529	3.7	119	0.00
73	Carbazole	40.000	37.868	5.3	115	0.00
74	Di-n-butylphthalate	40.000	41.575	-3.9	127	0.00
75 C	Fluoranthene	40.000	37.612	6.0	115	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	118	0.00
77	Benzydine	40.000	30.715	23.2	78	0.00
78	Pyrene	40.000	39.975	0.1	115	0.00
79 S	Terphenyl-d14	80.000	79.682	0.4	116	0.00
80	Butylbenzylphthalate	40.000	45.105	-12.8	125	0.00
81	Benzo(a)anthracene	40.000	38.190	4.5	110	0.00
82	3,3'-Dichlorobenzidine	40.000	40.836	-2.1	115	0.00
83	Chrysene	40.000	38.126	4.7	110	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.623	-9.1	123	0.00
85 c	Di-n-octyl phthalate	40.000	45.367	-13.4	127	0.00
86 I	Perylene-d12	20.000	20.000	0.0	120	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.834	5.4	109	0.00
88	Benzo(b)fluoranthene	40.000	38.565	3.6	112	0.00
89	Benzo(k)fluoranthene	40.000	37.041	7.4	106	0.00
90 C	Benzo(a)pyrene	40.000	38.956	2.6	111	0.00
91	Dibenzo(a,h)anthracene	40.000	37.710	5.7	109	0.00
92	Benzo(g,h,i)perylene	40.000	38.107	4.7	110	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080521\
Data File : BP006510.D
Acq On : 05 Aug 2021 17:30
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Aug 05 18:02:46 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 03 15:41:30 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0