

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080521\
 Data File : BP006528.D
 Acq On : 06 Aug 2021 03:43
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 06 05:45:50 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 15:41:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	153	0.00
2	1,4-Dioxane	40.000	36.693	8.3	142	0.00
3	Pyridine	40.000	37.553	6.1	137	0.00
4	n-Nitrosodimethylamine	40.000	37.184	7.0	137	0.00
5 S	2-Fluorophenol	80.000	77.545	3.1	144	0.00
6	Aniline	40.000	39.629	0.9	139	0.00
7 S	Phenol-d6	80.000	79.061	1.2	141	0.00
8	2-Chlorophenol	40.000	39.371	1.6	143	0.00
9	Benzaldehyde	40.000	39.552	1.1	148	0.00
10 C	Phenol	40.000	39.119	2.2	139	0.00
11	bis(2-Chloroethyl)ether	40.000	39.381	1.5	143	0.00
12	1,3-Dichlorobenzene	40.000	38.662	3.3	143	0.00
13 C	1,4-Dichlorobenzene	40.000	38.756	3.1	144	0.00
14	1,2-Dichlorobenzene	40.000	38.520	3.7	141	0.00
15	Benzyl Alcohol	40.000	40.100	-0.3	140	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.736	0.7	141	0.00
17	2-Methylphenol	40.000	40.222	-0.6	141	0.00
18	Hexachloroethane	40.000	39.384	1.5	147	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.917	0.2	137	0.00
20	3+4-Methylphenols	40.000	40.151	-0.4	138	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	146	0.00
22	Acetophenone	40.000	38.989	2.5	136	0.00
23 S	Nitrobenzene-d5	80.000	79.882	0.1	140	0.00
24	Nitrobenzene	40.000	39.343	1.6	139	0.00
25	Isophorone	40.000	40.678	-1.7	139	0.00
26 C	2-Nitrophenol	40.000	43.367	-8.4	148	0.00
27	2,4-Dimethylphenol	40.000	39.608	1.0	137	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.542	1.1	139	0.00
29 C	2,4-Dichlorophenol	40.000	40.326	-0.8	138	0.00
30	1,2,4-Trichlorobenzene	40.000	39.126	2.2	142	0.00
31	Naphthalene	40.000	38.713	3.2	137	0.00
32	Benzoic acid	40.000	42.324	-5.8	147	0.02
33	4-Chloroaniline	40.000	38.981	2.5	133	0.00
34 C	Hexachlorobutadiene	40.000	39.293	1.8	144	-0.01
35	Caprolactam	40.000	41.377	-3.4	132	0.01
36 C	4-Chloro-3-methylphenol	40.000	40.222	-0.6	134	0.00
37	2-Methylnaphthalene	40.000	38.870	2.8	135	0.00
38	1-Methylnaphthalene	40.000	39.016	2.5	136	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	138	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.355	1.6	136	0.00
41 P	Hexachlorocyclopentadiene	40.000	37.241	6.9	125	0.00
42 S	2,4,6-Tribromophenol	80.000	80.615	-0.8	131	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.877	-4.7	138	0.00
44	2,4,5-Trichlorophenol	40.000	40.430	-1.1	133	0.00
45 S	2-Fluorobiphenyl	80.000	78.193	2.3	134	0.00
46	1,1'-Biphenyl	40.000	39.096	2.3	133	0.00
47	2-Chloronaphthalene	40.000	39.069	2.3	133	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080521\
 Data File : BP006528.D
 Acq On : 06 Aug 2021 03:43
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 06 05:45:50 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 15:41:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.603	-4.0	132	0.00
49	Acenaphthylene	40.000	39.302	1.7	131	0.00
50	Dimethylphthalate	40.000	38.688	3.3	128	0.00
51	2,6-Dinitrotoluene	40.000	39.914	0.2	129	0.00
52 C	Acenaphthene	40.000	39.089	2.3	130	0.00
53	3-Nitroaniline	40.000	40.288	-0.7	128	0.00
54 P	2,4-Dinitrophenol	40.000	38.386	4.0	126	0.00
55	Dibenzofuran	40.000	38.312	4.2	128	0.00
56 P	4-Nitrophenol	40.000	40.338	-0.8	125	0.00
57	2,4-Dinitrotoluene	40.000	40.974	-2.4	130	0.00
58	Fluorene	40.000	38.786	3.0	128	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.856	0.4	130	0.00
60	Diethylphthalate	40.000	39.356	1.6	130	0.00
61	4-Chlorophenyl-phenylether	40.000	38.671	3.3	130	0.00
62	4-Nitroaniline	40.000	40.379	-0.9	125	0.00
63	Azobenzene	40.000	39.852	0.4	129	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	131	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.105	-0.3	126	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.932	0.2	127	0.00
67	4-Bromophenyl-phenylether	40.000	40.588	-1.5	131	-0.01
68	Hexachlorobenzene	40.000	39.256	1.9	128	0.00
69	Atrazine	40.000	40.801	-2.0	126	0.00
70 C	Pentachlorophenol	40.000	42.068	-5.2	129	0.00
71	Phenanthrene	40.000	39.064	2.3	125	0.00
72	Anthracene	40.000	39.709	0.7	126	0.00
73	Carbazole	40.000	39.321	1.7	122	0.00
74	Di-n-butylphthalate	40.000	42.346	-5.9	132	0.00
75 C	Fluoranthene	40.000	38.709	3.2	121	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	124	0.00
77	Benzidine	40.000	37.040	7.4	98	0.00
78	Pyrene	40.000	40.131	-0.3	121	0.00
79 S	Terphenyl-d14	80.000	80.105	-0.1	123	0.00
80	Butylbenzylphthalate	40.000	44.686	-11.7	130	-0.01
81	Benzo(a)anthracene	40.000	39.149	2.1	118	0.00
82	3,3'-Dichlorobenzidine	40.000	42.078	-5.2	124	0.00
83	Chrysene	40.000	38.492	3.8	116	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.089	-10.2	131	-0.01
85 c	Di-n-octyl phthalate	40.000	45.228	-13.1	133	-0.01
86 I	Perylene-d12	20.000	20.000	0.0	125	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	39.304	1.7	118	-0.02
88	Benzo(b)fluoranthene	40.000	38.575	3.6	116	-0.01
89	Benzo(k)fluoranthene	40.000	39.132	2.2	116	-0.01
90 C	Benzo(a)pyrene	40.000	39.924	0.2	118	-0.01
91	Dibenzo(a,h)anthracene	40.000	39.093	2.3	117	-0.03
92	Benzo(g,h,i)perylene	40.000	38.943	2.6	117	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080521\
Data File : BP006528.D
Acq On : 06 Aug 2021 03:43
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Aug 06 05:45:50 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 03 15:41:30 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0