

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080525\
 Data File : BP025298.D
 Acq On : 05 Aug 2025 12:03
 Operator : CG/JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDICC2.5

Quant Time: Aug 06 05:58:26 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080525.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 06 05:56:10 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.796	152	172992	20.000	ng	0.00
21) Naphthalene-d8	10.566	136	722061	20.000	ng	0.00
39) Acenaphthene-d10	14.413	164	469671	20.000	ng	0.00
64) Phenanthrene-d10	17.213	188	927131	20.000	ng	-0.01
76) Chrysene-d12	21.660	240	1039870	20.000	ng	0.00
86) Perylene-d12	25.036	264	1150516	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.396	112	51334	4.742	ng	-0.01
7) Phenol-d6	6.949	99	66396	4.649	ng	-0.02
23) Nitrobenzene-d5	8.931	82	54242	4.064	ng	-0.01
42) 2,4,6-Tribromophenol	15.913	330	17694	3.649	ng	-0.01
45) 2-Fluorobiphenyl	13.025	172	176352	5.214	ng	0.00
79) Terphenyl-d14	19.942	244	275667	5.008	ng	0.00
Target Compounds						
19) n-Nitroso-di-n-propyla...	8.584	70	21391	2.300	ng	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080525\
Data File : BP025298.D
Acq On : 05 Aug 2025 12:03
Operator : CG/JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SSTDICC2.5

Quant Time: Aug 06 05:58:26 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080525.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 06 05:56:10 2025
Response via : Initial Calibration

