

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080621\
 Data File : BP006530.D
 Acq On : 06 Aug 2021 08:56
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 06 12:14:04 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 15:41:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	133	-0.01
2	1,4-Dioxane	40.000	41.026	-2.6	138	0.00
3	Pyridine	40.000	41.080	-2.7	130	0.00
4	n-Nitrosodimethylamine	40.000	41.470	-3.7	132	0.00
5 S	2-Fluorophenol	80.000	83.146	-3.9	134	-0.01
6	Aniline	40.000	41.472	-3.7	126	-0.01
7 S	Phenol-d6	80.000	82.440	-3.0	128	0.00
8	2-Chlorophenol	40.000	41.232	-3.1	130	0.00
9	Benzaldehyde	40.000	40.935	-2.3	133	-0.01
10 C	Phenol	40.000	40.927	-2.3	126	0.00
11	bis(2-Chloroethyl)ether	40.000	41.357	-3.4	130	-0.01
12	1,3-Dichlorobenzene	40.000	40.341	-0.9	130	-0.01
13 C	1,4-Dichlorobenzene	40.000	40.627	-1.6	131	-0.01
14	1,2-Dichlorobenzene	40.000	40.561	-1.4	129	-0.01
15	Benzyl Alcohol	40.000	41.863	-4.7	126	-0.01
16	2,2'-oxybis(1-Chloropropane	40.000	42.105	-5.3	130	-0.01
17	2-Methylphenol	40.000	41.806	-4.5	127	-0.01
18	Hexachloroethane	40.000	41.608	-4.0	135	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	41.918	-4.8	125	0.00
20	3+4-Methylphenols	40.000	41.804	-4.5	125	-0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	127	-0.01
22	Acetophenone	40.000	40.688	-1.7	124	-0.01
23 S	Nitrobenzene-d5	80.000	83.160	-3.9	127	0.00
24	Nitrobenzene	40.000	41.031	-2.6	126	0.00
25	Isophorone	40.000	41.812	-4.5	125	-0.01
26 C	2-Nitrophenol	40.000	43.701	-9.3	130	-0.01
27	2,4-Dimethylphenol	40.000	41.277	-3.2	125	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.738	-1.8	125	0.00
29 C	2,4-Dichlorophenol	40.000	41.688	-4.2	124	0.00
30	1,2,4-Trichlorobenzene	40.000	40.201	-0.5	127	-0.01
31	Naphthalene	40.000	39.920	0.2	123	-0.01
32	Benzoic acid	40.000	42.205	-5.5	128	0.01
33	4-Chloroaniline	40.000	40.195	-0.5	120	-0.01
34 C	Hexachlorobutadiene	40.000	40.691	-1.7	130	-0.01
35	Caprolactam	40.000	41.384	-3.5	115	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.835	-2.1	119	0.00
37	2-Methylnaphthalene	40.000	40.006	-0.0	121	0.00
38	1-Methylnaphthalene	40.000	40.017	-0.0	122	-0.01
39 I	Acenaphthene-d10	20.000	20.000	0.0	120	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.110	-0.3	121	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.466	1.3	116	0.00
42 S	2,4,6-Tribromophenol	80.000	81.281	-1.6	115	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.451	-6.1	122	0.00
44	2,4,5-Trichlorophenol	40.000	41.352	-3.4	119	0.00
45 S	2-Fluorobiphenyl	80.000	81.028	-1.3	121	0.00
46	1,1'-Biphenyl	40.000	40.401	-1.0	120	0.00
47	2-Chloronaphthalene	40.000	40.340	-0.9	119	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080621\
 Data File : BP006530.D
 Acq On : 06 Aug 2021 08:56
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 06 12:14:04 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 03 15:41:30 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.159	-5.4	117	0.00
49	Acenaphthylene	40.000	40.650	-1.6	118	0.00
50	Dimethylphthalate	40.000	39.435	1.4	114	0.00
51	2,6-Dinitrotoluene	40.000	40.288	-0.7	113	0.00
52 C	Acenaphthene	40.000	40.162	-0.4	117	0.00
53	3-Nitroaniline	40.000	40.599	-1.5	112	0.00
54 P	2,4-Dinitrophenol	40.000	37.269	6.8	107	0.00
55	Dibenzofuran	40.000	39.876	0.3	117	0.00
56 P	4-Nitrophenol	40.000	39.683	0.8	107	0.00
57	2,4-Dinitrotoluene	40.000	40.955	-2.4	113	0.00
58	Fluorene	40.000	39.769	0.6	115	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.545	-1.4	115	0.00
60	Diethylphthalate	40.000	40.242	-0.6	115	0.00
61	4-Chlorophenyl-phenylether	40.000	39.851	0.4	116	0.00
62	4-Nitroaniline	40.000	40.188	-0.5	109	0.00
63	Azobenzene	40.000	41.503	-3.8	117	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	112	0.00
65	4,6-Dinitro-2-methylphenol	40.000	40.240	-0.6	108	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.371	-3.4	113	0.00
67	4-Bromophenyl-phenylether	40.000	41.808	-4.5	116	0.00
68	Hexachlorobenzene	40.000	40.893	-2.2	115	0.00
69	Atrazine	40.000	41.793	-4.5	111	0.00
70 C	Pentachlorophenol	40.000	42.936	-7.3	113	0.00
71	Phenanthrene	40.000	40.381	-1.0	111	0.00
72	Anthracene	40.000	41.108	-2.8	111	0.00
73	Carbazole	40.000	40.380	-1.0	107	0.00
74	Di-n-butylphthalate	40.000	42.721	-6.8	114	0.00
75 C	Fluoranthene	40.000	39.859	0.4	107	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	107	0.00
77	Benzydine	40.000	44.871	-12.2	103	0.00
78	Pyrene	40.000	41.071	-2.7	106	0.00
79 S	Terphenyl-d14	80.000	82.359	-2.9	109	0.00
80	Butylbenzylphthalate	40.000	45.568	-13.9	114	0.00
81	Benzo(a)anthracene	40.000	39.719	0.7	103	0.00
82	3,3'-Dichlorobenzidine	40.000	41.632	-4.1	106	0.00
83	Chrysene	40.000	39.742	0.6	104	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.469	-11.2	114	0.00
85 c	Di-n-octyl phthalate	40.000	45.568	-13.9	116	0.00
86 I	Perylene-d12	20.000	20.000	0.0	105	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.405	-1.0	102	0.00
88	Benzo(b)fluoranthene	40.000	40.901	-2.3	104	0.00
89	Benzo(k)fluoranthene	40.000	40.030	-0.1	101	0.00
90 C	Benzo(a)pyrene	40.000	41.143	-2.9	103	0.00
91	Dibenzo(a,h)anthracene	40.000	40.160	-0.4	102	-0.01
92	Benzo(g,h,i)perylene	40.000	40.190	-0.5	102	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080621\
Data File : BP006530.D
Acq On : 06 Aug 2021 08:56
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Aug 06 12:14:04 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Aug 03 15:41:30 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0