

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080824\
 Data File : BP021465.D
 Acq On : 09 Aug 2024 09:59
 Operator : RC/JU
 Sample : P3480-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 E0AA3

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 08/09/2024
 Supervised By :mohammad ahmed 08/12/2024

Quant Time: Aug 09 11:03:33 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071024.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 08 18:23:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.605	152	141486	20.000	ng/u1	0.00
20) Naphthalene-d8	10.375	136	638000	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.245	164	429789	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.069	188	938362	20.000	ng/u1	0.00
79) Chrysene-d12	21.522	240	942904	20.000	ng/u1	0.01
88) Perylene-d12	24.786	264	1091352	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.146	96	16630	5.349	ng/uL	0.00
4) Pyridine-d5	3.599	84	54243m	6.026	ng/u1	0.02
7) Phenol-d5	6.887	99	53647	4.651	ng/u1	0.06
9) Bis-(2-Chloroethyl)eth...	6.963	67	151810	21.717	ng/u1	0.00
11) 2-Chlorophenol-d4	7.175	132	189049	19.892	ng/u1	0.02
15) 4-Methylphenol-d8	8.369	113	137063	14.307	ng/u1	0.02
21) Nitrobenzene-d5	8.787	128	114190	23.043	ng/u1	0.01
24) 2-Nitrophenol-d4	9.499	143	126197m	22.250	ng/u1	0.01
28) 2,4-Dichlorophenol-d3	10.075	165	217214m	21.179	ng/u1	0.03
31) 4-Chloroaniline-d4	10.575	131	213956	15.747	ng/u1	0.04
46) Dimethylphthalate-d6	13.681	166	836807	26.783	ng/u1	0.02
49) Acenaphthylene-d8	13.945	160	846167	24.120	ng/u1	0.01
54) 4-Nitrophenol-d4	14.757	143	28226m	6.350	ng/u1	0.18
60) Fluorene-d10	15.275	176	696246	27.163	ng/u1	0.01
65) 4,6-Dinitro-2-methylph...	15.475	200	111019m	19.554	ng/u1	0.03
73) Anthracene-d10	17.181	188	1201630	28.832	ng/u1	0.00
81) Pyrene-d10	19.563	212	1514275	25.976	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.574	264	1668632	31.001	ng/u1	0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP080824\
Data File : BP021465.D
Acq On : 09 Aug 2024 09:59
Operator : RC/JU
Sample : P3480-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
E0AA3

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/09/2024
Supervised By :mohammad ahmed 08/12/2024

Quant Time: Aug 09 11:03:33 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP071024.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Aug 08 18:23:21 2024
Response via : Initial Calibration

