

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081122\
 Data File : BPCCC428.D
 Acq On : 12 Aug 2022 17:32
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Aug 12 23:19:36 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 11 23:05:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.575	152	113508	20.000	ng/u1	0.00
20) Naphthalene-d8	10.363	136	507808	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.251	164	305805	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.022	188	631718	20.000	ng/u1	0.00
79) Chrysene-d12	21.157	240	618357	20.000	ng/u1	0.00
88) Perylene-d12	23.462	264	619548	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.081	96	28221	7.622	ng/uL	0.00
4) Pyridine-d5	3.487	84	185287	19.838	ng/u1	0.00
7) Phenol-d5	6.757	99	193487	19.262	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.922	67	139459	19.580	ng/u1	0.00
11) 2-Chlorophenol-d4	7.110	132	166059	21.069	ng/u1	0.00
15) 4-Methylphenol-d8	8.299	113	165869	21.030	ng/u1	0.00
21) Nitrobenzene-d5	8.740	128	80302	23.236	ng/u1	0.00
24) 2-Nitrophenol-d4	9.457	143	82708	26.139	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.993	165	140926	22.055	ng/u1	0.00
31) 4-Chloroaniline-d4	10.516	131	234977	21.034	ng/u1	0.00
46) Dimethylphthalate-d6	13.675	166	466145	21.969	ng/u1	0.00
49) Acenaphthylene-d8	13.939	160	539447	21.120	ng/u1	0.00
54) 4-Nitrophenol-d4	14.481	143	78715	20.311	ng/u1	0.00
60) Fluorene-d10	15.257	176	377916	20.827	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.392	200	63875	21.769	ng/u1	0.00
73) Anthracene-d10	17.122	188	580242	20.435	ng/u1	0.00
81) Pyrene-d10	19.386	212	665021	19.521	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.315	264	623529	19.957	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.111	88	29495	7.500	ng/uL#	81
5) Pyridine	3.511	79	186576	19.178	ng/u1	88
6) Benzaldehyde	6.728	77	42422	9.690	ng/u1	96
8) Phenol	6.787	94	208658	19.190	ng/u1#	89
10) Bis(2-Chloroethyl)ether	7.016	93	172812	19.735	ng/u1	93
12) 2-Chlorophenol	7.140	128	174939	21.301	ng/u1	98
13) 2-Methylphenol	8.028	108	154858	19.776	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.116	45	249994	19.034	ng/u1	95
16) Acetophenone	8.404	105	273184	21.408	ng/u1	96
17) N-Nitroso-di-n-propyla...	8.393	70	147826	23.515	ng/u1	97
18) 4-Methylphenol	8.363	108	169592	20.205	ng/u1	100
19) Hexachloroethane	8.640	117	80363	23.174	ng/u1	98
22) Nitrobenzene	8.781	77	212942	22.010	ng/u1	96
23) Isophorone	9.310	82	391868	22.081	ng/u1#	97
25) 2-Nitrophenol	9.493	139	92741	25.150	ng/u1#	90
26) 2,4-Dimethylphenol	9.557	107	206372	22.149	ng/u1	94
27) Bis(2-Chloroethoxy)met...	9.798	93	230218	20.299	ng/u1	100
29) 2,4-Dichlorophenol	10.022	162	145050	21.707	ng/u1	98
30) Naphthalene	10.416	128	552368	19.993	ng/u1	99
32) 4-Chloroaniline	10.540	127	237436	21.205	ng/u1	98
33) Hexachlorobutadiene	10.693	225	94909	23.152	ng/u1	97
34) Caprolactam	11.340	113	50809	22.738	ng/u1	85
35) 4-Chloro-3-methylphenol	11.681	107	180736	23.005	ng/u1	99
36) 2-Methylnaphthalene	12.045	142	358915	20.958	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081122\
 Data File : BPCCC428.D
 Acq On : 12 Aug 2022 17:32
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Aug 12 23:19:36 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 11 23:05:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.263	142	367202	21.139	ng/ul#	98
39) 1,2,4,5-Tetrachloroben...	12.416	216	161338	20.040	ng/ul	95
40) Hexachlorocyclopentadiene	12.392	237	92470	21.434	ng/ul	97
41) 2,4,6-Trichlorophenol	12.669	196	106280	21.818	ng/ul	99
42) 2,4,5-Trichlorophenol	12.740	196	113932	21.914	ng/ul	99
43) 1,1'-Biphenyl	13.075	154	467535	19.470	ng/ul	99
44) 2-Chloronaphthalene	13.116	162	351907	19.532	ng/ul	96
45) 2-Nitroaniline	13.339	65	127526	27.869	ng/ul	97
47) Dimethylphthalate	13.722	163	468939	21.820	ng/ul	100
48) 2,6-Dinitrotoluene	13.845	165	88341	26.290	ng/ul	98
50) Acenaphthylene	13.969	152	601584	20.433	ng/ul	99
51) 3-Nitroaniline	14.175	138	61231	16.565	ng/ul	97
52) Acenaphthene	14.316	153	397483	20.242	ng/ul	98
53) 2,4-Dinitrophenol	14.386	184	40122	20.124	ng/ul#	88
55) 4-Nitrophenol	14.492	109	92565	25.747	ng/ul#	71
56) Dibenzofuran	14.657	168	532878	20.017	ng/ul	92
57) 2,4-Dinitrotoluene	14.639	165	131787	25.587	ng/ul	90
58) 2,3,4,6-Tetrachlorophenol	14.886	232	94803	23.440	ng/ul#	94
59) Diethylphthalate	15.104	149	523106	23.496	ng/ul	98
61) Fluorene	15.310	166	444208	20.547	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.310	204	200230	21.188	ng/ul	89
63) 4-Nitroaniline	15.351	138	36978	10.281	ng/ul#	74
66) 4,6-Dinitro-2-methylph...	15.404	198	65378	21.638	ng/ul#	91
67) N-Nitrosodiphenylamine	15.533	169	379079	20.364	ng/ul	99
68) 4-Bromophenyl-phenylether	16.216	248	116424	20.826	ng/ul	90
69) Hexachlorobenzene	16.316	284	134581	20.607	ng/ul	93
70) Atrazine	16.504	200	138014	22.243	ng/ul	97
71) Pentachlorophenol	16.669	266	79306	20.918	ng/ul	99
72) Phenanthrene	17.069	178	701812	19.813	ng/ul	99
74) Anthracene	17.157	178	716958	20.331	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.034	216	168105	18.536	ng/ul	99
76) Pentachlorobenzene	14.569	250	158743	20.099	ng/ul	99
77) Carbazole	17.439	167	655180	19.883	ng/ul	98
78) Di-n-butylphthalate	18.016	149	880254	24.430	ng/ul	100
80) Fluoranthene	19.063	202	798478	19.256	ng/ul#	93
82) Pyrene	19.416	202	828958	18.940	ng/ul#	93
83) Butylbenzylphthalate	20.316	149	409227	29.341	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.086	252	210563	18.711	ng/ul	99
85) Benzo(a)anthracene	21.139	228	785231	19.793	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.086	149	603341	26.947	ng/ul	98
87) Chrysene	21.192	228	780424	19.873	ng/ul	99
89) Di-n-octyl phthalate	21.986	149	1001974	23.309	ng/ul	100
90) Benzo(b)fluoranthene	22.763	252	798200	19.575	ng/ul	98
91) Benzo(k)fluoranthene	22.810	252	762764	19.119	ng/ul	97
93) Benzo(a)pyrene	23.362	252	781375	19.868	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	25.839	276	850852	19.570	ng/ul#	93
95) Dibenzo(a,h)anthracene	25.862	278	736180	19.909	ng/ul#	94
96) Benzo(g,h,i)perylene	26.574	276	729518	19.885	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :

BNA P

ClientSampleId :


Quant Time: Aug 12 23:19:36 2022


Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080822.M

Quant Title : SVOA CALIBRATION

QLast Update : Thu Aug 11 23:05:53 2022


Response via : Initial Calibration








#2
1,4-Dioxane
Concen: 7.500 ng/uL
RT: 3.111 min Scan# 21
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Instrument :
BNA_P
ClientSampleId :







Tgt Ion: 88 Resp: 29495
Ion Ratio Lower Upper
88 100
43 34.5 29.6 44.4
58 82.5 49.1 73.7#



#5
Pyridine
Concen: 19.178 ng/uL
RT: 3.511 min Scan# 89
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion: 79 Resp: 186576
Ion Ratio Lower Upper
79 100
52 84.5 57.8 86.6
51 44.0 31.5 47.3









#6
Benzaldehyde
Concen: 9.690 ng/ul
RT: 6.728 min Scan# 61
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 42422
Ion Ratio Lower Upper
77 100
105 86.2 72.0 108.0
106 83.8 69.4 104.0



#8
Phenol
Concen: 19.190 ng/ul
RT: 6.787 min Scan# 646
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion: 94 Resp: 208658
Ion Ratio Lower Upper
94 100
65 35.1 24.7 37.1
66 49.0 32.6 49.0#









#10
Bis(2-Chloroethyl)ether
Concen: 19.735 ng/ul
RT: 7.016 min Scan# 686
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 93 Resp: 172812
Ion Ratio Lower Upper
93 100
63 80.8 71.0 106.6
95 31.9 26.0 39.0



#12
2-Chlorophenol
Concen: 21.301 ng/ul
RT: 7.140 min Scan# 706
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:128 Resp: 174939
Ion Ratio Lower Upper
128 100
64 53.6 43.8 65.6
130 33.4 25.4 38.2





#13

2-Methylphenol

Concen: 19.776 ng/ul

RT: 8.028 min Scan# 8

Delta R.T. -0.006 min

Lab File: BPCCC428.D

Acq: 12 Aug 2022 17:32

Instrument :

BNA_P





ClientSampleId :

Tgt Ion:108 Resp: 154858

Ion Ratio Lower Upper

108 100

107 91.7 71.8 107.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.034 ng/ul

RT: 8.116 min Scan# 872

Delta R.T. -0.006 min

Lab File: BPCCC428.D

Acq: 12 Aug 2022 17:32




Tgt Ion: 45 Resp: 249994


Ion Ratio Lower Upper

45 100

77 14.8 9.9 14.9

79 10.7 7.2 10.8









#16
Acetophenone
Concen: 21.408 ng/ul
RT: 8.404 min Scan# 919
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:105 Resp: 273184
Ion Ratio Lower Upper
105 100
77 85.0 65.4 98.0
51 35.5 26.5 39.7



#17
N-Nitroso-di-n-propylamine
Concen: 23.515 ng/ul
RT: 8.393 min Scan# 919
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion: 70 Resp: 147826
Ion Ratio Lower Upper
70 100
42 66.5 55.5 83.3
101 9.7 7.1 10.7
130 17.9 14.8 22.2









#18
4-Methylphenol
Concen: 20.205 ng/ul
RT: 8.363 min Scan# 914
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 169592
Ion Ratio Lower Upper
108 100
107 110.5 88.2 132.4



#19
Hexachloroethane
Concen: 23.174 ng/ul
RT: 8.640 min Scan# 961
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:117 Resp: 80363
Ion Ratio Lower Upper
117 100
201 76.7 62.2 93.2
199 46.9 39.8 59.6









#22
Nitrobenzene
Concen: 22.010 ng/ul
RT: 8.781 min Scan# 985
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 212942
Ion Ratio Lower Upper
77 100
123 41.7 31.4 47.0
65 14.5 10.6 15.8



#23
Isophorone
Concen: 22.081 ng/ul
RT: 9.310 min Scan# 1075
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion: 82 Resp: 391868
Ion Ratio Lower Upper
82 100
95 9.0 5.8 8.8#
138 16.4 12.6 18.8









#25
2-Nitrophenol
Concen: 25.150 ng/ul
RT: 9.493 min Scan# 1106
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:139 Resp: 92741
Ion Ratio Lower Upper
139 100
65 62.7 46.7 70.1
109 40.9 24.2 36.2#



#26
2,4-Dimethylphenol
Concen: 22.149 ng/ul
RT: 9.557 min Scan# 1117
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:107 Resp: 206372
Ion Ratio Lower Upper
107 100
121 46.0 40.3 60.5
122 79.4 67.7 101.5





#27

Bis(2-Chloroethoxy)methane

Concen: 20.299 ng/ul

RT: 9.798 min Scan# 1158

Delta R.T. -0.006 min

Lab File: BPCCC428.D

Acq: 12 Aug 2022 17:32

Instrument :

BNA_P

ClientSampleId :





Tgt Ion: 93 Resp: 230218

Ion Ratio Lower Upper

93 100

95 31.6 25.2 37.8

123 11.0 8.4 12.6



#29

2,4-Dichlorophenol

Concen: 21.707 ng/ul

RT: 10.022 min Scan# 1196

Delta R.T. -0.000 min

Lab File: BPCCC428.D

Acq: 12 Aug 2022 17:32




Tgt Ion:162 Resp: 145050


Ion Ratio Lower Upper

162 100

164 63.9 52.6 79.0

98 43.1 34.2 51.4









#30
Naphthalene
Concen: 19.993 ng/ul
RT: 10.416 min Scan# 1263
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:128 Resp: 552368
Ion Ratio Lower Upper
128 100
129 10.9 9.0 13.4
127 13.3 10.4 15.6



#32
4-Chloroaniline
Concen: 21.205 ng/ul
RT: 10.540 min Scan# 1284
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:127 Resp: 237436
Ion Ratio Lower Upper
127 100
129 32.1 26.6 40.0









#33
Hexachlorobutadiene
Concen: 23.152 ng/ul
RT: 10.693 min Scan# 11
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:225 Resp: 94909
Ion Ratio Lower Upper
225 100
223 60.1 51.2 76.8
227 62.5 50.6 75.8



#34
Caprolactam
Concen: 22.738 ng/ul
RT: 11.340 min Scan# 1420
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:113 Resp: 50809
Ion Ratio Lower Upper
113 100
55 183.2 164.2 246.4
56 139.1 127.0 190.6









#35
4-Chloro-3-methylphenol
Concen: 23.005 ng/ul
RT: 11.681 min Scan# 1478
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:107 Resp: 180736
Ion Ratio Lower Upper
107 100
144 23.8 19.0 28.4
142 73.6 59.8 89.8



#36
2-Methylnaphthalene
Concen: 20.958 ng/ul
RT: 12.045 min Scan# 1540
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:142 Resp: 358915
Ion Ratio Lower Upper
142 100
141 90.7 71.8 107.8









#37
1-Methylnaphthalene
Concen: 21.139 ng/ul
RT: 12.263 min Scan# 11
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:142 Resp: 367202
Ion Ratio Lower Upper
142 100
141 94.3 74.1 111.1
116 5.2 3.4 5.0#



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.040 ng/ul
RT: 12.416 min Scan# 1603
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:216 Resp: 161338
Ion Ratio Lower Upper
216 100
214 79.3 61.1 91.7
179 26.4 18.6 27.8
108 27.0 19.1 28.7





#40

Hexachlorocyclopentadiene

Concen: 21.434 ng/ul

RT: 12.392 min Scan# 1599

Delta R.T. -0.000 min

Lab File: BPCCC428.D





Acq: 12 Aug 2022 17:32

Instrument :

BNA_P

ClientSampleId :

Tgt Ion:237	Resp:	92470
Ion Ratio	Lower	Upper
237	100	
235	60.8	50.4 75.6
272	12.0	9.9 14.9



#41

2,4,6-Trichlorophenol

Concen: 21.818 ng/ul




RT: 12.669 min Scan# 1646


Delta R.T. -0.006 min

Lab File: BPCCC428.D

Acq: 12 Aug 2022 17:32

Tgt Ion:196	Resp:	106280
Ion Ratio	Lower	Upper
196	100	
198	94.1	75.5 113.3
200	29.6	24.4 36.6









#42
2,4,5-Trichlorophenol
Concen: 21.914 ng/ul
RT: 12.740 min Scan# 1658
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:196 Resp: 113932
Ion Ratio Lower Upper
196 100
198 95.0 76.8 115.2
200 31.4 25.0 37.6



#43
1,1'-Biphenyl
Concen: 19.470 ng/ul
RT: 13.075 min Scan# 1715
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:154 Resp: 467535
Ion Ratio Lower Upper
154 100
153 39.9 31.9 47.9
76 13.8 12.2 18.2





#44

2-Chloronaphthalene

Concen: 19.532 ng/ul

RT: 13.116 min Scan# 1

Delta R.T. -0.000 min

Lab File: BPCCC428.D

Acq: 12 Aug 2022 17:32

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:162 Resp: 351907

Ion Ratio Lower Upper

162 100

164 32.9 25.6 38.4

127 47.4 34.6 51.8



#45

2-Nitroaniline

Concen: 27.869 ng/ul

RT: 13.339 min Scan# 1760

Delta R.T. -0.000 min

Lab File: BPCCC428.D

Acq: 12 Aug 2022 17:32




Tgt Ion: 65 Resp: 127526


Ion Ratio Lower Upper

65 100

92 65.5 51.4 77.2

138 88.5 67.4 101.0









#47
Dimethylphthalate
Concen: 21.820 ng/ul
RT: 13.722 min Scan# 1825
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:163 Resp: 468939
Ion Ratio Lower Upper
163 100
194 3.6 2.9 4.3
164 10.1 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 26.290 ng/ul
RT: 13.845 min Scan# 1846
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:165 Resp: 88341
Ion Ratio Lower Upper
165 100
89 82.7 65.8 98.6
121 27.2 24.2 36.2









#50
Acenaphthylene
Concen: 20.433 ng/ul
RT: 13.969 min Scan# 1867
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:152 Resp: 601584
Ion Ratio Lower Upper
152 100
151 20.5 15.8 23.8
153 13.1 10.5 15.7



#51
3-Nitroaniline
Concen: 16.565 ng/ul
RT: 14.175 min Scan# 1902
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:138 Resp: 61231
Ion Ratio Lower Upper
138 100
108 12.6 9.9 14.9
92 130.0 106.8 160.2





#52

Acenaphthene

Concen: 20.242 ng/ul

RT: 14.316 min Scan# 1926

Delta R.T. -0.000 min

Lab File: BPCCC428.D

Acq: 12 Aug 2022 17:32

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:153 Resp: 397483

Ion Ratio Lower Upper

153 100

152 49.1 39.5 59.3

154 88.5 68.6 103.0



#53

2,4-Dinitrophenol

Concen: 20.124 ng/ul

RT: 14.386 min Scan# 1938

Delta R.T. -0.000 min

Lab File: BPCCC428.D

Acq: 12 Aug 2022 17:32




Tgt Ion:184 Resp: 40122


Ion Ratio Lower Upper

184 100

63 98.2 75.0 112.6

154 83.9 53.1 79.7#









#55
4-Nitrophenol
Concen: 25.747 ng/ul
RT: 14.492 min Scan# 1957
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:109 Resp: 92565
Ion Ratio Lower Upper
109 100
139 90.9 100.6 150.8#
65 117.9 122.6 184.0#



#56
Dibenzofuran
Concen: 20.017 ng/ul
RT: 14.657 min Scan# 1984
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:168 Resp: 532878
Ion Ratio Lower Upper
168 100
139 47.6 34.2 51.2









#57
2,4-Dinitrotoluene
Concen: 25.587 ng/ul
RT: 14.639 min Scan# 1981
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:165 Resp: 131787
Ion Ratio Lower Upper
165 100
63 71.1 50.6 75.8
182 2.7 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 23.440 ng/ul
RT: 14.886 min Scan# 2023
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:232 Resp: 94803
Ion Ratio Lower Upper
232 100
131 69.7 50.4 75.6
130 3.9 2.6 3.8#
166 38.6 31.9 47.9









#59
Diethylphthalate
Concen: 23.496 ng/ul
RT: 15.104 min Scan# 2060
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:149 Resp: 523106
Ion Ratio Lower Upper
149 100
177 18.9 16.1 24.1
150 12.5 9.4 14.0



#61
Fluorene
Concen: 20.547 ng/ul
RT: 15.310 min Scan# 2095
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:166 Resp: 444208
Ion Ratio Lower Upper
166 100
165 98.1 77.7 116.5
167 13.0 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 21.188 ng/ul
RT: 15.310 min Scan# 2095
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:204 Resp: 200230
Ion Ratio Lower Upper
204 100
206 32.4 27.3 40.9
141 84.7 57.7 86.5



#63
4-Nitroaniline
Concen: 10.281 ng/ul
RT: 15.351 min Scan# 2102
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:138 Resp: 36978
Ion Ratio Lower Upper
138 100
92 55.8 55.9 83.9#
108 113.8 67.1 100.7#









#66
 4,6-Dinitro-2-methylphenol
 Concen: 21.638 ng/ul
 RT: 15.404 min Scan# 2111
 Delta R.T. -0.000 min
 Lab File: BPCCC428.D
 Acq: 12 Aug 2022 17:32




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion:198 Resp: 65378
 Ion Ratio Lower Upper
 198 100
 182 4.6 7.4 11.0#
 77 40.6 28.6 42.8



#67
 N-Nitrosodiphenylamine
 Concen: 20.364 ng/ul
 RT: 15.533 min Scan# 2133
 Delta R.T. -0.000 min
 Lab File: BPCCC428.D
 Acq: 12 Aug 2022 17:32

Tgt Ion:169 Resp: 379079
 Ion Ratio Lower Upper
 169 100
 168 67.7 53.6 80.4
 167 36.9 30.0 45.0









#68
4-Bromophenyl-phenylether
Concen: 20.826 ng/ul
RT: 16.216 min Scan# 21
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:248 Resp: 116424
Ion Ratio Lower Upper
248 100
250 97.4 77.3 115.9
141 113.5 76.4 114.6



#69
Hexachlorobenzene
Concen: 20.607 ng/ul
RT: 16.316 min Scan# 2266
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:284 Resp: 134581
Ion Ratio Lower Upper
284 100
142 54.9 39.5 59.3
249 34.1 25.1 37.7









#70
Atrazine
Concen: 22.243 ng/ul
RT: 16.504 min Scan# 21
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:200 Resp: 138014
Ion Ratio Lower Upper
200 100
173 27.6 22.6 34.0
215 42.8 36.5 54.7



#71
Pentachlorophenol
Concen: 20.918 ng/ul
RT: 16.669 min Scan# 2326
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:266 Resp: 79306
Ion Ratio Lower Upper
266 100
264 64.4 52.2 78.2
268 62.7 50.6 75.8









#72
Phenanthrene
Concen: 19.813 ng/u1
RT: 17.069 min Scan# 2109
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:178 Resp: 701812
Ion Ratio Lower Upper
178 100
179 14.5 12.1 18.1
176 19.2 15.4 23.2



#74
Anthracene
Concen: 20.331 ng/u1
RT: 17.157 min Scan# 2409
Delta R.T. -0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:178 Resp: 716958
Ion Ratio Lower Upper
178 100
179 15.1 12.3 18.5
176 19.2 14.7 22.1









#75
1,2,3,4-Tetrachlorobenzene
Concen: 18.536 ng/uL
RT: 13.034 min Scan# 1708
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:216 Resp: 168105
Ion Ratio Lower Upper
216 100
214 78.3 62.0 93.0
218 47.2 38.2 57.4



#76
Pentachlorobenzene
Concen: 20.099 ng/uL
RT: 14.569 min Scan# 1969
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32


Tgt Ion:250 Resp: 158743
Ion Ratio Lower Upper
250 100
248 63.3 49.2 73.8
252 63.2 50.8 76.2








#77
 Carbazole
 Concen: 19.883 ng/ul
 RT: 17.439 min Scan# 2458
 Delta R.T. -0.006 min
 Lab File: BPCCC428.D
 Acq: 12 Aug 2022 17:32

Instrument :
 BNA_P
 ClientSampleId :







Tgt Ion:167 Resp: 655180
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.1 25.7
 139 15.9 11.5 17.3



#78
 Di-n-butylphthalate
 Concen: 24.430 ng/ul
 RT: 18.016 min Scan# 2555
 Delta R.T. -0.000 min
 Lab File: BPCCC428.D
 Acq: 12 Aug 2022 17:32

Tgt Ion:149 Resp: 880254
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.4 11.0
 104 6.6 5.1 7.7









#80
Fluoranthene
Concen: 19.256 ng/ul
RT: 19.063 min Scan# 2733
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:202 Resp: 798478
Ion Ratio Lower Upper
202 100
101 11.5 11.9 17.9#
100 9.1 9.3 13.9#



#82
Pyrene
Concen: 18.940 ng/ul
RT: 19.416 min Scan# 2793
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:202 Resp: 828958
Ion Ratio Lower Upper
202 100
101 14.1 14.3 21.5#
100 12.3 11.8 17.6





#83

Butylbenzylphthalate

Concen: 29.341 ng/ul

RT: 20.316 min Scan# 2946

Delta R.T. -0.000 min

Lab File: BPCCC428.D

Acq: 12 Aug 2022 17:32

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:149 Resp: 409227

Ion Ratio Lower Upper

149 100

91 82.1 64.2 96.4

206 16.2 14.0 21.0



#84

3,3'-Dichlorobenzidine

Concen: 18.711 ng/ul

RT: 21.086 min Scan# 3077

Delta R.T. -0.000 min

Lab File: BPCCC428.D

Acq: 12 Aug 2022 17:32




Tgt Ion:252 Resp: 210563


Ion Ratio Lower Upper

252 100

254 64.4 51.8 77.6

126 13.3 11.4 17.0









#85
Benzo(a)anthracene
Concen: 19.793 ng/ul
RT: 21.139 min Scan# 3086
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:228 Resp: 785231
Ion Ratio Lower Upper
228 100
229 19.3 15.8 23.6
226 27.2 21.8 32.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 26.947 ng/ul
RT: 21.086 min Scan# 3077
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:149 Resp: 603341
Ion Ratio Lower Upper
149 100
167 24.6 20.4 30.6
279 3.6 2.9 4.3









#87
Chrysene
Concen: 19.873 ng/ul
RT: 21.192 min Scan# 3095
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:228 Resp: 780424
Ion Ratio Lower Upper
228 100
226 30.6 23.8 35.6
229 19.7 15.7 23.5



#89
Di-n-octyl phthalate
Concen: 23.309 ng/ul
RT: 21.986 min Scan# 3230
Delta R.T. 0.006 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:149 Resp: 1001974









#90
Benzo(b)fluoranthene
Concen: 19.575 ng/ul
RT: 22.763 min Scan# 3362
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:252 Resp: 798200
Ion Ratio Lower Upper
252 100
253 22.1 17.4 26.0
125 11.6 10.8 16.2



#91
Benzo(k)fluoranthene
Concen: 19.119 ng/ul
RT: 22.810 min Scan# 3370
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:252 Resp: 762764
Ion Ratio Lower Upper
252 100
253 21.5 17.6 26.4
125 11.8 11.0 16.6









#93
Benzo(a)pyrene
Concen: 19.868 ng/ul
RT: 23.362 min Scan# 3464
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:252 Resp: 781375
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.2
125 12.6 12.5 18.7



#94
Indeno(1,2,3-cd)pyrene
Concen: 19.570 ng/ul
RT: 25.839 min Scan# 3885
Delta R.T. -0.000 min
Lab File: BPCCC428.D
Acq: 12 Aug 2022 17:32

Tgt Ion:276 Resp: 850852
Ion Ratio Lower Upper
276 100
138 24.9 25.5 38.3#
277 24.6 19.8 29.8





#95

Dibenzo(a,h)anthracene

Concen: 19.909 ng/ul

RT: 25.862 min Scan# 3

Delta R.T. -0.000 min

Lab File: BPCCC428.D

Acq: 12 Aug 2022 17:32

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:278 Resp: 736180

Ion Ratio Lower Upper

278 100

139 17.6 18.5 27.7#

279 24.0 19.6 29.4



#96

Benzo(g,h,i)perylene

Concen: 19.885 ng/ul

RT: 26.574 min Scan# 4010

Delta R.T. 0.006 min

Lab File: BPCCC428.D

Acq: 12 Aug 2022 17:32

Tgt Ion:276 Resp: 729518

Ion Ratio Lower Upper

276 100

138 22.9 22.8 34.2

277 23.5 18.5 27.7

