

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081122\
 Data File : BP011407.D
 Acq On : 12 Aug 2022 03:08
 Operator : CG/JU
 Sample : N4128-08
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BGBX4

Quant Time: Aug 12 04:01:21 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 11 23:05:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.575	152	94843	20.000	ng/u1	0.00
20) Naphthalene-d8	10.369	136	399259	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.251	164	228877	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.022	188	473258	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.151	240	463122	20.000	ng/u1	0.00
88) Perylene-d12	23.457	264	488987	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.081	96	12166	3.932	ng/uL	0.00
4) Pyridine-d5	3.499	84	50779	6.507	ng/u1	0.01
7) Phenol-d5	6.764	99	49710	5.923	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.922	67	157214	26.417	ng/u1	0.00
11) 2-Chlorophenol-d4	7.111	132	147627	22.417	ng/u1	0.00
15) 4-Methylphenol-d8	8.299	113	92392	14.019	ng/u1	0.00
21) Nitrobenzene-d5	8.740	128	88987	32.749	ng/u1	0.00
24) 2-Nitrophenol-d4	9.457	143	84782	34.079	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.993	165	132005	26.275	ng/u1	0.00
31) 4-Chloroaniline-d4	10.516	131	193913	22.078	ng/u1	0.00
46) Dimethylphthalate-d6	13.675	166	534702	33.670	ng/u1	0.00
49) Acenaphthylene-d8	13.940	160	603892	31.590	ng/u1	0.00
54) 4-Nitrophenol-d4	14.493	143	9441	3.255	ng/u1	0.01
60) Fluorene-d10	15.257	176	447823	32.974	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.387	200	66575	30.286	ng/u1	0.00
73) Anthracene-d10	17.122	188	788191	37.053	ng/u1	0.00
81) Pyrene-d10	19.386	212	918613	36.003	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.310	264	908178	36.828	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081122\
Data File : BP011407.D
Acq On : 12 Aug 2022 03:08
Operator : CG/JU
Sample : N4128-08
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BGBX4

Quant Time: Aug 12 04:01:21 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080822.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Aug 11 23:05:53 2022
Response via : Initial Calibration

