

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081122\
 Data File : BP011392.D
 Acq On : 11 Aug 2022 17:55
 Operator : CG/JU
 Sample : PB146869BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS869

Quant Time: Aug 11 18:53:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 08 18:47:19 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 08/12/2022
 Supervised By :Sohil Jodhani 08/13/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.581	152	83101	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.369	136	372380	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.257	164	227734	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.028	188	470498	20.000	ng/u1	-0.01
79) Chrysene-d12	21.163	240	467249	20.000	ng/u1	0.00
88) Perylene-d12	23.469	264	461112	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.081	96	15680	5.784	ng/uL	0.00
4) Pyridine-d5	3.487	84	197577	28.894	ng/u1	-0.01
7) Phenol-d5	6.764	99	236537	32.163	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	6.928	67	155903	29.898	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.116	132	192220	33.312	ng/u1	0.00
15) 4-Methylphenol-d8	8.305	113	188698	32.678	ng/u1	-0.01
21) Nitrobenzene-d5	8.746	128	90456	35.692	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.463	143	93949	40.490	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	9.999	165	170007	36.282	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.522	131	242195	29.565	ng/u1	-0.01
46) Dimethylphthalate-d6	13.681	166	555463	35.153	ng/u1	0.00
49) Acenaphthylene-d8	13.946	160	650873	34.219	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.487	143	99327	34.416	ng/u1	0.00
60) Fluorene-d10	15.257	176	456161	33.757	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.393	200	81829	37.443	ng/u1	0.00
73) Anthracene-d10	17.128	188	698589	33.034	ng/u1	-0.01
81) Pyrene-d10	19.392	212	817777	31.768	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.322	264	817459	35.153	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.117	88	33099	11.495	ng/uL#	79
5) Pyridine	3.511	79	196893	27.644	ng/u1	85
6) Benzaldehyde	6.734	77	140883	43.957	ng/u1	99
8) Phenol	6.793	94	240990	30.273	ng/u1	90
10) Bis(2-Chloroethyl)ether	7.022	93	192645	30.049	ng/u1	95
12) 2-Chlorophenol	7.146	128	198003	32.931	ng/u1	99
13) 2-Methylphenol	8.034	108	177184	30.906	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.122	45	288983m	30.054	ng/u1	
16) Acetophenone	8.411	105	309720	33.152	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.405	70	163555	35.537	ng/u1	98
18) 4-Methylphenol	8.369	108	193647	31.513	ng/u1	99
19) Hexachloroethane	8.646	117	89064	35.081	ng/u1	99
22) Nitrobenzene	8.787	77	243852	34.372	ng/u1	99
23) Isophorone	9.316	82	444147	34.129	ng/u1	99
25) 2-Nitrophenol	9.493	139	103087	38.122	ng/u1#	82
26) 2,4-Dimethylphenol	9.563	107	215621	31.557	ng/u1	92
27) Bis(2-Chloroethoxy)met...	9.805	93	261008	31.384	ng/u1	98
29) 2,4-Dichlorophenol	10.028	162	165993	33.876	ng/u1	98
30) Naphthalene	10.422	128	627005	30.948	ng/u1	99
32) 4-Chloroaniline	10.546	127	225842	27.504	ng/u1	98
33) Hexachlorobutadiene	10.699	225	108294	36.025	ng/u1	98
34) Caprolactam	11.346	113	59573m	36.356	ng/u1	
35) 4-Chloro-3-methylphenol	11.687	107	211866	36.774	ng/u1	96
36) 2-Methylnaphthalene	12.046	142	410413	32.681	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081122\
 Data File : BP011392.D
 Acq On : 11 Aug 2022 17:55
 Operator : CG/JU
 Sample : PB146869BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS869


Quant Time: Aug 11 18:53:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Aug 08 18:47:19 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 08/12/2022
 Supervised By :Sohil Jodhani 08/13/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.269	142	416755	32.717	ng/ul#	99
39) 1,2,4,5-Tetrachloroben...	12.422	216	183489	30.605	ng/ul	98
40) Hexachlorocyclopentadiene	12.393	237	87495	27.234	ng/ul	98
41) 2,4,6-Trichlorophenol	12.675	196	121502	33.494	ng/ul	96
42) 2,4,5-Trichlorophenol	12.746	196	134738	34.801	ng/ul	97
43) 1,1'-Biphenyl	13.081	154	530335	29.656	ng/ul	97
44) 2-Chloronaphthalene	13.116	162	402218	29.977	ng/ul	95
45) 2-Nitroaniline	13.340	65	144601	42.434	ng/ul	96
47) Dimethylphthalate	13.728	163	552063	34.495	ng/ul	100
48) 2,6-Dinitrotoluene	13.851	165	104885	41.914	ng/ul	97
50) Acenaphthylene	13.975	152	695861	31.737	ng/ul	99
51) 3-Nitroaniline	14.181	138	111124	40.368	ng/ul	99
52) Acenaphthene	14.322	153	460041	31.460	ng/ul	98
53) 2,4-Dinitrophenol	14.387	184	53201	35.832	ng/ul#	86
55) 4-Nitrophenol	14.498	109	113684	42.462	ng/ul#	71
56) Dibenzofuran	14.657	168	631034	31.831	ng/ul	90
57) 2,4-Dinitrotoluene	14.645	165	158249	41.258	ng/ul	88
58) 2,3,4,6-Tetrachlorophenol	14.893	232	110811	36.791	ng/ul#	94
59) Diethylphthalate	15.104	149	624647	37.675	ng/ul	97
61) Fluorene	15.316	166	530618	32.958	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.316	204	237220	33.707	ng/ul	89
63) 4-Nitroaniline	15.357	138	127104m	47.452	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.410	198	82153	36.507	ng/ul#	90
67) N-Nitrosodiphenylamine	15.540	169	454990	32.817	ng/ul	99
68) 4-Bromophenyl-phenylether	16.222	248	139066	33.401	ng/ul	94
69) Hexachlorobenzene	16.322	284	160128	32.920	ng/ul	92
70) Atrazine	16.510	200	135903	29.408	ng/ul	96
71) Pentachlorophenol	16.675	266	93041	32.949	ng/ul	99
72) Phenanthrene	17.069	178	849099	32.185	ng/ul	100
74) Anthracene	17.163	178	846923	32.246	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.040	216	185474	27.459	ng/uL	100
76) Pentachlorobenzene	14.575	250	181080	30.783	ng/uL	98
77) Carbazole	17.445	167	803170	32.726	ng/ul	99
78) Di-n-butylphthalate	18.016	149	1088792	40.571	ng/ul	99
80) Fluoranthene	19.063	202	993674	31.713	ng/ul	95
82) Pyrene	19.422	202	1027915	31.081	ng/ul#	92
83) Butylbenzylphthalate	20.322	149	502026	47.635	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.092	252	311095	36.584	ng/ul	100
85) Benzo(a)anthracene	21.145	228	977456	32.607	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.086	149	749030	44.273	ng/ul#	97
87) Chrysene	21.198	228	960274	32.360	ng/ul	100
89) Di-n-octyl phthalate	21.986	149	1273058	39.791	ng/ul	100
90) Benzo(b)fluoranthene	22.769	252	1010938	33.311	ng/ul	98
91) Benzo(k)fluoranthene	22.816	252	959009	32.297	ng/ul	98
93) Benzo(a)pyrene	23.369	252	983152	33.588	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	25.857	276	1102552	34.073	ng/ul	95
95) Dibenzo(a,h)anthracene	25.868	278	942822	34.259	ng/ul	95
96) Benzo(g,h,i)perylene	26.580	276	927607	33.972	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 11.495 ng/uL
RT: 3.117 min Scan# 21
Delta R.T. -0.005 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869



Tgt Ion: 88 Resp: 33099
Ion Ratio Lower Upper
88 100
43 37.4 29.6 44.4
58 86.5 49.1 73.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022



#5
Pyridine
Concen: 27.644 ng/uL
RT: 3.511 min Scan# 89
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55


Tgt Ion: 79 Resp: 196893
Ion Ratio Lower Upper
79 100
52 85.8 57.8 86.6
51 46.3 31.5 47.3





#6
Benzaldehyde
Concen: 43.957 ng/ul
RT: 6.734 min Scan# 61
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869




Tgt Ion: 77 Resp: 14088
Ion Ratio Lower Upper
77 100
105 91.6 72.0 108.0
106 87.5 69.4 104.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022




#8
Phenol
Concen: 30.273 ng/ul
RT: 6.793 min Scan# 647
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Tgt Ion: 94 Resp: 240990
Ion Ratio Lower Upper
94 100
65 35.0 24.7 37.1
66 47.9 32.6 49.0





#10
Bis(2-Chloroethyl)ether
Concen: 30.049 ng/ul
RT: 7.022 min Scan# 686
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869




Tgt Ion: 93 Resp: 192649
Ion Ratio Lower Upper
93 100
63 82.1 71.0 106.6
95 32.5 26.0 39.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022




#12
2-Chlorophenol
Concen: 32.931 ng/ul
RT: 7.146 min Scan# 707
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55



Tgt Ion:128 Resp: 198003
Ion Ratio Lower Upper
128 100
64 53.5 43.8 65.6
130 32.5 25.4 38.2





#13

2-Methylphenol

Concen: 30.906 ng/ul

RT: 8.034 min Scan# 8

Delta R.T. -0.012 min

Lab File: BP011392.D

Acq: 11 Aug 2022 17:55

Instrument :

BNA_P

ClientSampleId :

SLCS869

Tgt Ion:108 Resp: 177184

Ion Ratio Lower Upper

108 100





107 92.2 71.8 107.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/12/2022

Supervised By :Sohil Jodhani 08/13/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 30.054 ng/ul m

RT: 8.122 min Scan# 873

Delta R.T. -0.012 min

Lab File: BP011392.D

Acq: 11 Aug 2022 17:55




Tgt Ion: 45 Resp: 288983


Ion Ratio Lower Upper

45 100

77 14.9 9.9 14.9#

79 11.4 7.2 10.8#









#16
Acetophenone
Concen: 33.152 ng/ul
RT: 8.411 min Scan# 911
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55

Instrument :
BNA_P
ClientSampleId :
SLCS869

Tgt Ion:105 Resp: 309720
Ion Ratio Lower Upper
105 100
77 84.5 65.4 98.0
51 34.5 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022



#17
N-Nitroso-di-n-propylamine
Concen: 35.537 ng/ul
RT: 8.405 min Scan# 921
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55


Tgt Ion: 70 Resp: 163555
Ion Ratio Lower Upper
70 100
42 68.6 55.5 83.3
101 9.1 7.1 10.7
130 20.2 14.8 22.2





#18
4-Methylphenol
Concen: 31.513 ng/ul
RT: 8.369 min Scan# 915
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869




Tgt Ion:108 Resp: 19364
Ion Ratio Lower Upper
108 100
107 111.3 88.2 132.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022




#19
Hexachloroethane
Concen: 35.081 ng/ul
RT: 8.646 min Scan# 962
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Tgt Ion:117 Resp: 89064
Ion Ratio Lower Upper
117 100
201 76.5 62.2 93.2
199 49.4 39.8 59.6





#22
Nitrobenzene
Concen: 34.372 ng/ul
RT: 8.787 min Scan# 986
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869



Tgt Ion: 77 Resp: 243852
Ion Ratio Lower Upper
77 100
123 39.6 31.4 47.0
65 14.6 10.6 15.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022



#23
Isophorone
Concen: 34.129 ng/ul
RT: 9.316 min Scan# 1076
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55


Tgt Ion: 82 Resp: 444147
Ion Ratio Lower Upper
82 100
95 8.4 5.8 8.8
138 15.7 12.6 18.8





#25
2-Nitrophenol
Concen: 38.122 ng/ul
RT: 9.493 min Scan# 1106
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869




Tgt Ion:139 Resp: 10308
Ion Ratio Lower Upper
139 100
65 68.7 46.7 70.1
109 44.6 24.2 36.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022




#26
2,4-Dimethylphenol
Concen: 31.557 ng/ul
RT: 9.563 min Scan# 1118
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55



Tgt Ion:107 Resp: 215621
Ion Ratio Lower Upper
107 100
121 46.2 40.3 60.5
122 76.1 67.7 101.5





#27

Bis(2-Chloroethoxy)methane

Concen: 31.384 ng/ul

RT: 9.805 min Scan# 1159

Delta R.T. -0.012 min

Lab File: BP011392.D

Acq: 11 Aug 2022 17:55

Instrument :

BNA_P

ClientSampleId :

SLCS869

Tgt Ion: 93 Resp: 261000

Ion Ratio Lower Upper

93 100

95 32.3 25.2 37.8





123 11.1 8.4 12.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/12/2022

Supervised By :Sohil Jodhani 08/13/2022



#29

2,4-Dichlorophenol

Concen: 33.876 ng/ul

RT: 10.028 min Scan# 1197

Delta R.T. -0.006 min

Lab File: BP011392.D

Acq: 11 Aug 2022 17:55




Tgt Ion:162 Resp: 165993


Ion Ratio Lower Upper

162 100

164 64.5 52.6 79.0


98 41.9 34.2 51.4





#30
Naphthalene
Concen: 30.948 ng/ul
RT: 10.422 min Scan# 1264
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869



Tgt Ion:128 Resp: 627000
Ion Ratio Lower Upper
128 100
129 11.0 9.0 13.4
127 13.6 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022



#32
4-Chloroaniline
Concen: 27.504 ng/ul
RT: 10.546 min Scan# 1285
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55


Tgt Ion:127 Resp: 225842
Ion Ratio Lower Upper
127 100
129 32.4 26.6 40.0





#33
Hexachlorobutadiene
Concen: 36.025 ng/ul
RT: 10.699 min Scan# 1311
Delta R.T. -0.018 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869




Tgt Ion:225 Resp: 108294
Ion Ratio Lower Upper
225 100
223 62.6 51.2 76.8
227 61.2 50.6 75.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022




#34
Caprolactam
Concen: 36.356 ng/ul m
RT: 11.346 min Scan# 1421
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Tgt Ion:113 Resp: 59573
Ion Ratio Lower Upper
113 100
55 188.5 164.2 246.4
56 152.7 127.0 190.6





#35
4-Chloro-3-methylphenol
Concen: 36.774 ng/ul
RT: 11.687 min Scan# 1480
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869



Tgt Ion:107 Resp: 211860
Ion Ratio Lower Upper
107 100
144 22.7 19.0 28.4
142 71.2 59.8 89.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022



#36
2-Methylnaphthalene
Concen: 32.681 ng/ul
RT: 12.046 min Scan# 1540
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55

Tgt Ion:142 Resp: 410413
Ion Ratio Lower Upper
142 100
141 90.9 71.8 107.8









#37
1-Methylnaphthalene
Concen: 32.717 ng/ul
RT: 12.269 min Scan# 1578
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55

Instrument :
BNA_P
ClientSampleId :
SLCS869

Tgt Ion:142 Resp: 41675
Ion Ratio Lower Upper
142 100
141 93.7 74.1 111.1
116 5.1 3.4 5.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022



#39
1,2,4,5-Tetrachlorobenzene
Concen: 30.605 ng/ul
RT: 12.422 min Scan# 1604
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55


Tgt Ion:216 Resp: 183489
Ion Ratio Lower Upper
216 100
214 77.1 61.1 91.7
179 25.6 18.6 27.8
108 26.1 19.1 28.7





#40
Hexachlorocyclopentadiene
Concen: 27.234 ng/ul
RT: 12.393 min Scan# 1599
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869



Tgt Ion: 237 Resp: 8749
Ion Ratio Lower Upper
237 100
235 64.6 50.4 75.6
272 11.3 9.9 14.9




Manual Integrations
APPROVED


Reviewed By : Jagrut Upadhyay 08/12/2022
Supervised By : Sohil Jodhani 08/13/2022



#41
2,4,6-Trichlorophenol
Concen: 33.494 ng/ul
RT: 12.675 min Scan# 1647
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55


Tgt Ion: 196 Resp: 121502
Ion Ratio Lower Upper
196 100
198 99.3 75.5 113.3
200 31.1 24.4 36.6





#42
2,4,5-Trichlorophenol
Concen: 34.801 ng/ul
RT: 12.746 min Scan# 1659
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869




Tgt Ion:196 Resp: 134738
Ion Ratio Lower Upper
196 100
198 94.2 76.8 115.2
200 28.6 25.0 37.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022




#43
1,1'-Biphenyl
Concen: 29.656 ng/ul
RT: 13.081 min Scan# 1716
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Tgt Ion:154 Resp: 530335
Ion Ratio Lower Upper
154 100
153 41.6 31.9 47.9
76 14.3 12.2 18.2





#44
2-Chloronaphthalene
Concen: 29.977 ng/ul
RT: 13.116 min Scan# 1722
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869




Tgt Ion:162 Resp: 402218
Ion Ratio Lower Upper
162 100
164 32.9 25.6 38.4
127 47.9 34.6 51.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022




#45
2-Nitroaniline
Concen: 42.434 ng/ul
RT: 13.340 min Scan# 1760
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Tgt Ion: 65 Resp: 144601
Ion Ratio Lower Upper
65 100
92 60.8 51.4 77.2
138 87.8 67.4 101.0





#47
Dimethylphthalate
Concen: 34.495 ng/ul
RT: 13.728 min Scan# 1826
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869




Tgt Ion:163 Resp: 55206
Ion Ratio Lower Upper
163 100
194 3.6 2.9 4.3
164 10.3 8.2 12.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022




#48
2,6-Dinitrotoluene
Concen: 41.914 ng/ul
RT: 13.851 min Scan# 1847
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Tgt Ion:165 Resp: 104885
Ion Ratio Lower Upper
165 100
89 79.9 65.8 98.6
121 27.2 24.2 36.2





#50
Acenaphthylene
Concen: 31.737 ng/ul
RT: 13.975 min Scan# 1868
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument : BNA_P
ClientSampleId : SLCS869




Tgt Ion:152 Resp: 69586
Ion Ratio Lower Upper
152 100
151 20.1 15.8 23.8
153 13.0 10.5 15.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022




#51
3-Nitroaniline
Concen: 40.368 ng/ul
RT: 14.181 min Scan# 1903
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Tgt Ion:138 Resp: 111124
Ion Ratio Lower Upper
138 100
108 12.0 9.9 14.9
92 132.5 106.8 160.2





#52
Acenaphthene
Concen: 31.460 ng/ul
RT: 14.322 min Scan# 1927
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869




Tgt Ion:153 Resp: 46004
Ion Ratio Lower Upper
153 100
152 49.4 39.5 59.3
154 88.0 68.6 103.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022




#53
2,4-Dinitrophenol
Concen: 35.832 ng/ul
RT: 14.387 min Scan# 1938
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Tgt Ion:184 Resp: 53201
Ion Ratio Lower Upper
184 100
63 105.4 75.0 112.6
154 80.3 53.1 79.7#





#55
4-Nitrophenol
Concen: 42.462 ng/ul
RT: 14.498 min Scan# 1957
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869



Tgt Ion:109 Resp: 113684
Ion Ratio Lower Upper
109 100
139 87.6 100.6 150.8
65 119.6 122.6 184.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022



#56
Dibenzofuran
Concen: 31.831 ng/ul
RT: 14.657 min Scan# 1984
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55


Tgt Ion:168 Resp: 631034
Ion Ratio Lower Upper
168 100
139 48.9 34.2 51.2





#57
2,4-Dinitrotoluene
Concen: 41.258 ng/ul
RT: 14.645 min Scan# 1982
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869




Tgt Ion:165 Resp: 158249
Ion Ratio Lower Upper
165 100
63 72.6 50.6 75.8
182 2.5 2.2 3.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 36.791 ng/ul
RT: 14.893 min Scan# 2024
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55



Tgt Ion:232 Resp: 110811
Ion Ratio Lower Upper
232 100
131 69.7 50.4 75.6
130 4.0 2.6 3.8#
166 40.8 31.9 47.9









#59
Diethylphthalate
Concen: 37.675 ng/ul
RT: 15.104 min Scan# 2060
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55

Instrument :
BNA_P
ClientSampleId :
SLCS869

Tgt Ion:149 Resp: 62464
Ion Ratio Lower Upper
149 100
177 18.6 16.1 24.1
150 12.6 9.4 14.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022



#61
Fluorene
Concen: 32.958 ng/ul
RT: 15.316 min Scan# 2096
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55


Tgt Ion:166 Resp: 530618
Ion Ratio Lower Upper
166 100
165 98.4 77.7 116.5
167 12.9 11.0 16.4





#62
4-Chlorophenyl-phenylether
Concen: 33.707 ng/ul
RT: 15.316 min Scan# 2096
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869



Tgt Ion:204 Resp: 237220
Ion Ratio Lower Upper
204 100
206 32.3 27.3 40.9
141 84.7 57.7 86.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022



#63
4-Nitroaniline
Concen: 47.452 ng/ul m
RT: 15.357 min Scan# 2103
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55


Tgt Ion:138 Resp: 127104
Ion Ratio Lower Upper
138 100
92 61.0 55.9 83.9
108 115.6 67.1 100.7#





#66
4,6-Dinitro-2-methylphenol
Concen: 36.507 ng/ul
RT: 15.410 min Scan# 2112
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869




Tgt Ion:198 Resp: 8215
Ion Ratio Lower Upper
198 100
182 5.0 7.4 11.0
77 41.6 28.6 42.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022




#67
N-Nitrosodiphenylamine
Concen: 32.817 ng/ul
RT: 15.540 min Scan# 2134
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Tgt Ion:169 Resp: 454990
Ion Ratio Lower Upper
169 100
168 68.0 53.6 80.4
167 37.2 30.0 45.0





#68
4-Bromophenyl-phenylether
Concen: 33.401 ng/ul
RT: 16.222 min Scan# 211
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869




Tgt Ion:248 Resp: 139060
Ion Ratio Lower Upper
248 100
250 97.0 77.3 115.9
141 106.6 76.4 114.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022




#69
Hexachlorobenzene
Concen: 32.920 ng/ul
RT: 16.322 min Scan# 2267
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55



Tgt Ion:284 Resp: 160128
Ion Ratio Lower Upper
284 100
142 54.9 39.5 59.3
249 35.7 25.1 37.7





#70

Atrazine

Concen: 29.408 ng/ul

RT: 16.510 min Scan# 2299

Delta R.T. -0.006 min

Lab File: BP011392.D

Acq: 11 Aug 2022 17:55

Instrument :

BNA_P

ClientSampleId :





SLCS869

Tgt Ion:	200	Resp:	13590
Ion Ratio	Lower	Upper	
200	100		
173	27.3	22.6	34.0
215	42.4	36.5	54.7

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022



#71

Pentachlorophenol

Concen: 32.949 ng/ul




RT: 16.675 min Scan# 2327


Delta R.T. -0.012 min

Lab File: BP011392.D

Acq: 11 Aug 2022 17:55


Tgt Ion:	266	Resp:	93041
Ion Ratio	Lower	Upper	
266	100		
264	64.5	52.2	78.2
268	64.0	50.6	75.8





#72
Phenanthrene
Concen: 32.185 ng/u1
RT: 17.069 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869




Tgt Ion:178 Resp: 849099
Ion Ratio Lower Upper
178 100
179 15.1 12.1 18.1
176 19.4 15.4 23.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022




#74
Anthracene
Concen: 32.246 ng/u1
RT: 17.163 min Scan# 2410
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Tgt Ion:178 Resp: 846923
Ion Ratio Lower Upper
178 100
179 15.5 12.3 18.5
176 19.1 14.7 22.1





#75
1,2,3,4-Tetrachlorobenzene
Concen: 27.459 ng/uL
RT: 13.040 min Scan# 1709
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869




Tgt Ion:216 Resp: 185474
Ion Ratio Lower Upper
216 100
214 77.6 62.0 93.0
218 48.0 38.2 57.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022




#76
Pentachlorobenzene
Concen: 30.783 ng/uL
RT: 14.575 min Scan# 1970
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Tgt Ion:250 Resp: 181080
Ion Ratio Lower Upper
250 100
248 63.2 49.2 73.8
252 64.5 50.8 76.2





#77
 Carbazole
 Concen: 32.726 ng/ul
 RT: 17.445 min Scan# 2458
 Delta R.T. -0.006 min
 Lab File: BP011392.D
 Acq: 11 Aug 2022 17:55




Instrument : BNA_P
 ClientSampleId : SLCS869




Tgt Ion:167 Resp: 803170
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.1 25.7
 139 15.5 11.5 17.3

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 08/12/2022
 Supervised By :Sohil Jodhani 08/13/2022




#78
 Di-n-butylphthalate
 Concen: 40.571 ng/ul
 RT: 18.016 min Scan# 2555
 Delta R.T. -0.012 min
 Lab File: BP011392.D
 Acq: 11 Aug 2022 17:55



Tgt Ion:149 Resp: 1088792
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.4 11.0
 104 7.0 5.1 7.7









#80
Fluoranthene
Concen: 31.713 ng/u1
RT: 19.063 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55

Instrument :
BNA_P
ClientSampleId :
SLCS869

Tgt Ion:202 Resp: 993674
Ion Ratio Lower Upper
202 100
101 12.8 11.9 17.9
100 9.8 9.3 13.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022



#82
Pyrene
Concen: 31.081 ng/u1
RT: 19.422 min Scan# 2794
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55

Tgt Ion:202 Resp: 1027915
Ion Ratio Lower Upper
202 100
101 13.9 14.3 21.5#
100 11.7 11.8 17.6#





#83

Butylbenzylphthalate

Concen: 47.635 ng/ul

RT: 20.322 min Scan# 2947

Delta R.T. -0.006 min

Lab File: BP011392.D

Acq: 11 Aug 2022 17:55

Instrument :

BNA_P

ClientSampleId :





SLCS869

Tgt Ion:	149	Resp:	502020
Ion Ratio	Lower	Upper	
149	100		
91	81.0	64.2	96.4
206	16.6	14.0	21.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022



#84

3,3'-Dichlorobenzidine

Concen: 36.584 ng/ul




RT: 21.092 min Scan# 3078


Delta R.T. -0.006 min

Lab File: BP011392.D

Acq: 11 Aug 2022 17:55


Tgt Ion:	252	Resp:	311095
Ion Ratio	Lower	Upper	
252	100		
254	64.5	51.8	77.6
126	14.0	11.4	17.0





#85
Benzo(a)anthracene
Concen: 32.607 ng/u1
RT: 21.145 min Scan# 3087
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869




Tgt Ion:228 Resp: 977450
Ion Ratio Lower Upper
228 100
229 19.5 15.8 23.6
226 26.9 21.8 32.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022




#86
Bis(2-ethylhexyl)phthalate
Concen: 44.273 ng/u1
RT: 21.086 min Scan# 3077
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55



Tgt Ion:149 Resp: 749030
Ion Ratio Lower Upper
149 100
167 23.8 20.4 30.6
279 2.9 2.9 4.3#





#87

Chrysene

Concen: 32.360 ng/u1

RT: 21.198 min Scan# 3096

Delta R.T. -0.006 min

Lab File: BP011392.D

Acq: 11 Aug 2022 17:55

Instrument :

BNA_P

ClientSampleId :





SLCS869

Tgt Ion:228 Resp: 960274
 Ion Ratio Lower Upper
 228 100
 226 29.6 23.8 35.6
 229 19.5 15.7 23.5

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/12/2022
 Supervised By :Sohil Jodhani 08/13/2022



#89

Di-n-octyl phthalate

Concen: 39.791 ng/u1




RT: 21.986 min Scan# 3230


Delta R.T. -0.012 min

Lab File: BP011392.D

Acq: 11 Aug 2022 17:55

Tgt Ion:149 Resp: 1273058









#90
Benzo(b)fluoranthene
Concen: 33.311 ng/ul
RT: 22.769 min Scan# 3363
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55

Instrument :
BNA_P
ClientSampleId :
SLCS869

Tgt Ion:252 Resp: 1010938
Ion Ratio Lower Upper
252 100
253 21.5 17.4 26.0
125 11.8 10.8 16.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022



#91
Benzo(k)fluoranthene
Concen: 32.297 ng/ul
RT: 22.816 min Scan# 3371
Delta R.T. -0.006 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55

Tgt Ion:252 Resp: 959009
Ion Ratio Lower Upper
252 100
253 21.9 17.6 26.4
125 12.0 11.0 16.6





#93

Benzo(a)pyrene

Concen: 33.588 ng/ul

RT: 23.369 min Scan# 3465

Delta R.T. -0.006 min

Lab File: BP011392.D

Acq: 11 Aug 2022 17:55

Instrument :

BNA_P

ClientSampleId :





SLCS869

Tgt Ion:	252	Resp:	983152
Ion Ratio	Lower	Upper	
252	100		
253	21.7	17.4	26.2
125	13.1	12.5	18.7

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022



#94

Indeno(1,2,3-cd)pyrene

Concen: 34.073 ng/ul




RT: 25.857 min Scan# 3888


Delta R.T. 0.000 min

Lab File: BP011392.D

Acq: 11 Aug 2022 17:55


Tgt Ion:	276	Resp:	1102552
Ion Ratio	Lower	Upper	
276	100		
138	27.3	25.5	38.3
277	24.7	19.8	29.8





#95
Dibenzo(a,h)anthracene
Concen: 34.259 ng/ul
RT: 25.868 min Scan# 3891
Delta R.T. -0.018 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55




Instrument :
BNA_P
ClientSampleId :
SLCS869



Tgt Ion: 278 Resp: 94282
Ion Ratio Lower Upper
278 100
139 18.5 18.5 27.7
279 24.0 19.6 29.4

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022



#96
Benzo(g,h,i)perylene
Concen: 33.972 ng/ul
RT: 26.580 min Scan# 4011
Delta R.T. -0.012 min
Lab File: BP011392.D
Acq: 11 Aug 2022 17:55

Tgt Ion: 276 Resp: 927607
Ion Ratio Lower Upper
276 100
138 25.1 22.8 34.2
277 23.9 18.5 27.7

