

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081122\
 Data File : BP011421.D
 Acq On : 12 Aug 2022 12:56
 Operator : CG/JU
 Sample : N4129-09
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BGBW9

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 08/12/2022
 Supervised By :Sohil Jodhani 08/13/2022

Quant Time: Aug 12 13:20:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 11 23:05:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.575	152	68988	20.000	ng/u1	0.00
20) Naphthalene-d8	10.363	136	296725m	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.251	164	168351	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.022	188	349399	20.000	ng/u1	0.00
79) Chrysene-d12	21.157	240	341660	20.000	ng/u1	# 0.00
88) Perylene-d12	23.457	264	363183	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.081	96	10676	4.744	ng/uL	0.00
4) Pyridine-d5	3.493	84	49837	8.779	ng/u1	0.00
7) Phenol-d5	6.757	99	38655	6.331	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.922	67	134263	31.015	ng/u1	0.00
11) 2-Chlorophenol-d4	7.105	132	120913	25.241	ng/u1	0.00
15) 4-Methylphenol-d8	8.299	113	75265	15.701	ng/u1	0.00
21) Nitrobenzene-d5	8.740	128	76121	37.694	ng/u1	0.00
24) 2-Nitrophenol-d4	9.457	143	70358	38.054	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.993	165	109539	29.338	ng/u1	0.00
31) 4-Chloroaniline-d4	10.516	131	159912	24.498	ng/u1	0.00
46) Dimethylphthalate-d6	13.675	166	476499	40.793	ng/u1	0.00
49) Acenaphthylene-d8	13.939	160	524507	37.302	ng/u1	0.00
54) 4-Nitrophenol-d4	14.498	143	6899m	3.234	ng/u1	0.02
60) Fluorene-d10	15.257	176	385452	38.585	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.386	200	54550	33.612	ng/u1	0.00
73) Anthracene-d10	17.122	188	672805	42.841	ng/u1	0.00
81) Pyrene-d10	19.386	212	782386	41.565	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.310	264	766718	41.862	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081122\
Data File : BP011421.D
Acq On : 12 Aug 2022 12:56
Operator : CG/JU
Sample : N4129-09
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
BGBW9

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 08/12/2022
Supervised By :Sohil Jodhani 08/13/2022

Quant Time: Aug 12 13:20:52 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080822.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Aug 11 23:05:53 2022
Response via : Initial Calibration

