

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081122\
 Data File : BP011423.D
 Acq On : 12 Aug 2022 14:05
 Operator : CG/JU
 Sample : N4129-14
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 BGBW4

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 08/12/2022
 Supervised By :Sohil Jodhani 08/13/2022

Quant Time: Aug 12 14:54:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 11 23:05:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.575	152	88249	20.000	ng/u1	0.00
20) Naphthalene-d8	10.363	136	366324	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.251	164	211032	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.022	188	436213m	20.000	ng/u1	0.00
79) Chrysene-d12	21.157	240	427961	20.000	ng/u1	# 0.00
88) Perylene-d12	23.462	264	454215	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.081	96	11391m	3.957	ng/uL	0.00
4) Pyridine-d5	3.499	84	52312	7.204	ng/u1	0.01
7) Phenol-d5	6.757	99	41977	5.375	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.922	67	154662	27.930	ng/u1	0.00
11) 2-Chlorophenol-d4	7.104	132	138359	22.579	ng/u1	0.00
15) 4-Methylphenol-d8	8.293	113	82192	13.404	ng/u1	-0.01
21) Nitrobenzene-d5	8.740	128	85616	34.341	ng/u1	0.00
24) 2-Nitrophenol-d4	9.457	143	81664	35.777	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.993	165	124219	26.948	ng/u1	0.00
31) 4-Chloroaniline-d4	10.516	131	175490	21.777	ng/u1	0.00
46) Dimethylphthalate-d6	13.675	166	526608	35.965	ng/u1	0.00
49) Acenaphthylene-d8	13.939	160	582531	33.050	ng/u1	0.00
54) 4-Nitrophenol-d4	14.492	143	7227m	2.702	ng/u1	0.01
60) Fluorene-d10	15.257	176	433197	34.594	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.386	200	61998m	30.599	ng/u1	0.00
73) Anthracene-d10	17.122	188	734604	37.467	ng/u1	0.00
81) Pyrene-d10	19.386	212	852028	36.137	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.315	264	843008	36.802	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081122\
 Data File : BP011423.D
 Acq On : 12 Aug 2022 14:05
 Operator : CG/JU
 Sample : N4129-14
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 BNA_P
ClientSampleId :
 BGBW4

Quant Time: Aug 12 14:54:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 11 23:05:53 2022
 Response via : Initial Calibration

Manual Integrations
APPROVED
 Reviewed By :Jagrut Upadhyay 08/12/2022
 Supervised By :Sohil Jodhani 08/13/2022

