

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081122\
 Data File : BP011425.D
 Acq On : 12 Aug 2022 15:14
 Operator : CG/JU
 Sample : PB146917BS
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS917

Quant Time: Aug 12 17:40:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 11 23:05:53 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/13/2022
 Supervised By :Sohil Jodhani 08/13/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.575	152	97125	20.000	ng/u1	0.00
20) Naphthalene-d8	10.363	136	429926	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.251	164	264823	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.028	188	562912m	20.000	ng/u1	0.00
79) Chrysene-d12	21.157	240	557370	20.000	ng/u1	0.00
88) Perylene-d12	23.468	264	555265	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.081	96	17372	5.483	ng/uL	0.00
4) Pyridine-d5	3.487	84	214755	26.872	ng/u1	0.00
7) Phenol-d5	6.758	99	264559	30.779	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.922	67	175489	28.795	ng/u1	0.00
11) 2-Chlorophenol-d4	7.110	132	214613	31.822	ng/u1	0.00
15) 4-Methylphenol-d8	8.299	113	213375	31.616	ng/u1	0.00
21) Nitrobenzene-d5	8.740	128	102286	34.958	ng/u1	0.00
24) 2-Nitrophenol-d4	9.457	143	105514	39.388	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.993	165	190712	35.253	ng/u1	0.00
31) 4-Chloroaniline-d4	10.516	131	273679	28.937	ng/u1	0.00
46) Dimethylphthalate-d6	13.675	166	630367	34.306	ng/u1	0.00
49) Acenaphthylene-d8	13.940	160	742735	33.580	ng/u1	0.00
54) 4-Nitrophenol-d4	14.481	143	116688	34.769	ng/u1	0.00
60) Fluorene-d10	15.257	176	520053	33.095	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.392	200	94513m	36.147	ng/u1	0.00
73) Anthracene-d10	17.122	188	813107	32.136	ng/u1	0.00
81) Pyrene-d10	19.386	212	942906	30.706	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.316	264	924775	33.025	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.111	88	36960	10.983	ng/uL#	81
5) Pyridine	3.505	79	218774	26.281	ng/u1#	81
6) Benzaldehyde	6.728	77	159860	42.676	ng/u1	98
8) Phenol	6.787	94	268658	28.876	ng/u1#	88
10) Bis(2-Chloroethyl)ether	7.016	93	213592	28.506	ng/u1	97
12) 2-Chlorophenol	7.140	128	220091	31.319	ng/u1	98
13) 2-Methylphenol	8.028	108	199155	29.722	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.116	45	323530	28.788	ng/u1	96
16) Acetophenone	8.405	105	349132	31.974	ng/u1	95
17) N-Nitroso-di-n-propyla...	8.399	70	184342	34.270	ng/u1	96
18) 4-Methylphenol	8.363	108	216533	30.149	ng/u1	97
19) Hexachloroethane	8.640	117	100522	33.877	ng/u1	100
22) Nitrobenzene	8.781	77	273693	33.414	ng/u1	98
23) Isophorone	9.310	82	503365	33.502	ng/u1	98
25) 2-Nitrophenol	9.487	139	115561	37.015	ng/u1#	80
26) 2,4-Dimethylphenol	9.557	107	259729	32.925	ng/u1	92
27) Bis(2-Chloroethoxy)met...	9.799	93	292228	30.435	ng/u1	98
29) 2,4-Dichlorophenol	10.022	162	188912	33.393	ng/u1	98
30) Naphthalene	10.416	128	706774	30.216	ng/u1	100
32) 4-Chloroaniline	10.540	127	254115	26.805	ng/u1	99
33) Hexachlorobutadiene	10.699	225	120594	34.747	ng/u1	98
34) Caprolactam	11.346	113	67088m	35.462	ng/u1	
35) 4-Chloro-3-methylphenol	11.687	107	242880	36.515	ng/u1	97
36) 2-Methylnaphthalene	12.046	142	467412	32.238	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081122\
 Data File : BP011425.D
 Acq On : 12 Aug 2022 15:14
 Operator : CG/JU
 Sample : PB146917BS
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS917


Quant Time: Aug 12 17:40:52 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 11 23:05:53 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/13/2022
 Supervised By :Sohil Jodhani 08/13/2022


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.263	142	472420	32.123	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.416	216	209853	30.100	ng/ul	98
40) Hexachlorocyclopentadiene	12.387	237	113989	30.511	ng/ul	99
41) 2,4,6-Trichlorophenol	12.669	196	141576	33.562	ng/ul	98
42) 2,4,5-Trichlorophenol	12.740	196	149278	33.156	ng/ul	99
43) 1,1'-Biphenyl	13.075	154	605233	29.104	ng/ul	98
44) 2-Chloronaphthalene	13.116	162	462019	29.612	ng/ul	95
45) 2-Nitroaniline	13.340	65	174138	43.944	ng/ul	97
47) Dimethylphthalate	13.722	163	623671	33.511	ng/ul	100
48) 2,6-Dinitrotoluene	13.845	165	120225	41.315	ng/ul	98
50) Acenaphthylene	13.969	152	797564	31.281	ng/ul	100
51) 3-Nitroaniline	14.175	138	132145	41.282	ng/ul	100
52) Acenaphthene	14.316	153	525940	30.929	ng/ul	98
53) 2,4-Dinitrophenol	14.387	184	57826	33.493	ng/ul#	81
55) 4-Nitrophenol	14.498	109	129771	41.682	ng/ul#	71
56) Dibenzofuran	14.657	168	718747	31.178	ng/ul	91
57) 2,4-Dinitrotoluene	14.640	165	182189	40.847	ng/ul	91
58) 2,3,4,6-Tetrachlorophenol	14.887	232	128064	36.564	ng/ul#	95
59) Diethylphthalate	15.104	149	715459	37.109	ng/ul	97
61) Fluorene	15.310	166	605473	32.340	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.316	204	269591	32.942	ng/ul	91
63) 4-Nitroaniline	15.351	138	147936m	47.494	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.404	198	94543m	35.116	ng/ul	
67) N-Nitrosodiphenylamine	15.534	169	519695	31.330	ng/ul	100
68) 4-Bromophenyl-phenylether	16.216	248	158973	31.914	ng/ul	91
69) Hexachlorobenzene	16.316	284	185865	31.938	ng/ul	94
70) Atrazine	16.510	200	156541	28.313	ng/ul	97
71) Pentachlorophenol	16.675	266	111778	33.086	ng/ul	98
72) Phenanthrene	17.069	178	964861m	30.568	ng/ul	
74) Anthracene	17.163	178	979115	31.159	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.034	216	209673	25.946	ng/uL	99
76) Pentachlorobenzene	14.569	250	208402	29.612	ng/uL	99
77) Carbazole	17.445	167	911631	31.047	ng/ul	98
78) Di-n-butylphthalate	18.016	149	1261124	39.278	ng/ul	99
80) Fluoranthene	19.063	202	1138175	30.452	ng/ul	94
82) Pyrene	19.416	202	1172868	29.730	ng/ul	94
83) Butylbenzylphthalate	20.316	149	580738	46.194	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.086	252	361371	35.625	ng/ul	99
85) Benzo(a)anthracene	21.145	228	1117927	31.263	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.086	149	872760	43.245	ng/ul	98
87) Chrysene	21.198	228	1105241	31.223	ng/ul	99
89) Di-n-octyl phthalate	21.986	149	1469964	38.155	ng/ul	100
90) Benzo(b)fluoranthene	22.768	252	1131597	30.964	ng/ul	98
91) Benzo(k)fluoranthene	22.816	252	1114609	31.172	ng/ul	97
93) Benzo(a)pyrene	23.368	252	1122528	31.847	ng/ul#	96
94) Indeno(1,2,3-cd)pyrene	25.851	276	1242155	31.878	ng/ul	94
95) Dibenzo(a,h)anthracene	25.868	278	1061207	32.022	ng/ul#	94
96) Benzo(g,h,i)perylene	26.580	276	1048988	31.903	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 10.983 ng/uL
RT: 3.111 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917




Tgt Ion: 88 Resp: 36960
Ion Ratio Lower Upper
88 100
43 39.3 29.6 44.4
58 83.4 49.1 73.7

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022




#5
Pyridine
Concen: 26.281 ng/uL
RT: 3.505 min Scan# 88
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Tgt Ion: 79 Resp: 218774
Ion Ratio Lower Upper
79 100
52 91.8 57.8 86.6
51 46.3 31.5 47.3





#6
Benzaldehyde
Concen: 42.676 ng/ul
RT: 6.728 min Scan# 636
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
Client Sample Id :
SLCS917




Tgt Ion: 77 Resp: 159860
Ion Ratio Lower Upper
77 100
105 87.9 72.0 108.0
106 85.4 69.4 104.0

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022




#8
Phenol
Concen: 28.876 ng/ul
RT: 6.787 min Scan# 646
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Tgt Ion: 94 Resp: 268658
Ion Ratio Lower Upper
94 100
65 35.8 24.7 37.1
66 49.5 32.6 49.0#





#10
Bis(2-Chloroethyl)ether
Concen: 28.506 ng/ul
RT: 7.016 min Scan# 685
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917



Tgt Ion: 93 Resp: 213592
Ion Ratio Lower Upper
93 100
63 85.5 71.0 106.6
95 32.0 26.0 39.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#12
2-Chlorophenol
Concen: 31.319 ng/ul
RT: 7.140 min Scan# 706
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14

Tgt Ion:128 Resp: 220091
Ion Ratio Lower Upper
128 100
64 52.9 43.8 65.6
130 31.7 25.4 38.2





#13

2-Methylphenol

Concen: 29.722 ng/ul

RT: 8.028 min Scan# 857

Delta R.T. -0.006 min

Lab File: BP011425.D

Acq: 12 Aug 2022 15:14

Instrument :

BNA_P

ClientSampleId :

SLCS917

Tgt Ion:108 Resp: 19915

Ion Ratio Lower Upper

108 100





107 93.0 71.8 107.6

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/13/2022

Supervised By :Sohil Jodhani 08/13/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 28.788 ng/ul

RT: 8.116 min Scan# 872

Delta R.T. -0.006 min

Lab File: BP011425.D

Acq: 12 Aug 2022 15:14




Tgt Ion: 45 Resp: 323530


Ion Ratio Lower Upper

45 100

77 13.9 9.9 14.9


79 10.5 7.2 10.8





#16
Acetophenone
Concen: 31.974 ng/ul
RT: 8.405 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917




Tgt Ion: 105 Resp: 349132
Ion Ratio Lower Upper
105 100
77 86.0 65.4 98.0
51 36.4 26.5 39.7

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022




#17
N-Nitroso-di-n-propylamine
Concen: 34.270 ng/ul
RT: 8.399 min Scan# 920
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Tgt Ion: 70 Resp: 184342
Ion Ratio Lower Upper
70 100
42 72.8 55.5 83.3
101 10.1 7.1 10.7
130 19.2 14.8 22.2





#18
4-Methylphenol
Concen: 30.149 ng/ul
RT: 8.363 min Scan# 914
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917




Tgt Ion:108 Resp: 216533
Ion Ratio Lower Upper
108 100
107 113.2 88.2 132.4

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022




#19
Hexachloroethane
Concen: 33.877 ng/ul
RT: 8.640 min Scan# 961
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Tgt Ion:117 Resp: 100522
Ion Ratio Lower Upper
117 100
201 78.3 62.2 93.2
199 49.7 39.8 59.6





#22
Nitrobenzene
Concen: 33.414 ng/ul
RT: 8.781 min Scan# 985
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
Client Sample Id :
SLCS917




Tgt Ion: 77 Resp: 273693
Ion Ratio Lower Upper
77 100
123 38.6 31.4 47.0
65 15.2 10.6 15.8

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022




#23
Isophorone
Concen: 33.502 ng/ul
RT: 9.310 min Scan# 1075
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Tgt Ion: 82 Resp: 503365
Ion Ratio Lower Upper
82 100
95 8.7 5.8 8.8
138 16.0 12.6 18.8





#25
2-Nitrophenol
Concen: 37.015 ng/ul
RT: 9.487 min Scan# 1105
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917




Tgt Ion:139 Resp: 115563
Ion Ratio Lower Upper
139 100
65 70.1 46.7 70.1
109 45.3 24.2 36.2

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022




#26
2,4-Dimethylphenol
Concen: 32.925 ng/ul
RT: 9.557 min Scan# 1117
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14



Tgt Ion:107 Resp: 259729
Ion Ratio Lower Upper
107 100
121 46.0 40.3 60.5
122 77.0 67.7 101.5





#27

Bis(2-Chloroethoxy)methane

Concen: 30.435 ng/ul

RT: 9.799 min Scan# 1158

Delta R.T. -0.006 min

Lab File: BP011425.D

Acq: 12 Aug 2022 15:14

Instrument :

BNA_P

ClientSampleId :

SLCS917

Tgt Ion: 93 Resp: 292228

Ion Ratio Lower Upper

93 100

95 32.5 25.2 37.8





123 11.2 8.4 12.6

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/13/2022

Supervised By :Sohil Jodhani 08/13/2022



#29

2,4-Dichlorophenol

Concen: 33.393 ng/ul

RT: 10.022 min Scan# 1196

Delta R.T. 0.000 min

Lab File: BP011425.D

Acq: 12 Aug 2022 15:14




Tgt Ion:162 Resp: 188912


Ion Ratio Lower Upper

162 100

164 64.6 52.6 79.0


98 41.4 34.2 51.4





#30
Naphthalene
Concen: 30.216 ng/ul
RT: 10.416 min Scan# 1263
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917



Tgt Ion:128 Resp: 706774
Ion Ratio Lower Upper
128 100
129 11.0 9.0 13.4
127 13.1 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#32
4-Chloroaniline
Concen: 26.805 ng/ul
RT: 10.540 min Scan# 1284
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14


Tgt Ion:127 Resp: 254115
Ion Ratio Lower Upper
127 100
129 32.8 26.6 40.0





#33
Hexachlorobutadiene
Concen: 34.747 ng/ul
RT: 10.699 min Scan# 1311
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917



Tgt Ion:225 Resp: 120594
Ion Ratio Lower Upper
225 100
223 62.4 51.2 76.8
227 64.4 50.6 75.8




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#34
Caprolactam
Concen: 35.462 ng/ul m
RT: 11.346 min Scan# 1421
Delta R.T. 0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14


Tgt Ion:113 Resp: 67088
Ion Ratio Lower Upper
113 100
55 188.7 164.2 246.4
56 146.4 127.0 190.6





#35
4-Chloro-3-methylphenol
Concen: 36.515 ng/ul
RT: 11.687 min Scan# 1478
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917




Tgt Ion:107 Resp: 242880
Ion Ratio Lower Upper
107 100
144 22.9 19.0 28.4
142 72.4 59.8 89.8

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022




#36
2-Methylnaphthalene
Concen: 32.238 ng/ul
RT: 12.046 min Scan# 1540
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Tgt Ion:142 Resp: 467412
Ion Ratio Lower Upper
142 100
141 90.1 71.8 107.8





#37
1-Methylnaphthalene
Concen: 32.123 ng/ul
RT: 12.263 min Scan# 1577
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917




Tgt Ion:142 Resp: 472420
Ion Ratio Lower Upper
142 100
141 92.8 74.1 111.1
116 4.9 3.4 5.0

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 30.100 ng/ul
RT: 12.416 min Scan# 1603
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Tgt Ion:216 Resp: 209853
Ion Ratio Lower Upper
216 100
214 77.6 61.1 91.7
179 25.5 18.6 27.8
108 25.5 19.1 28.7





#40
Hexachlorocyclopentadiene
Concen: 30.511 ng/ul
RT: 12.387 min Scan# 1598
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917



Tgt Ion: 237 Resp: 113989
Ion Ratio Lower Upper
237 100
235 62.5 50.4 75.6
272 11.0 9.9 14.9




Manual Integrations
APPROVED


Reviewed By : Yogesh Patel 08/13/2022
Supervised By : Sohil Jodhani 08/13/2022



#41
2,4,6-Trichlorophenol
Concen: 33.562 ng/ul
RT: 12.669 min Scan# 1646
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14


Tgt Ion: 196 Resp: 141576
Ion Ratio Lower Upper
196 100
198 97.0 75.5 113.3
200 31.0 24.4 36.6





#42
2,4,5-Trichlorophenol
Concen: 33.156 ng/ul
RT: 12.740 min Scan# 1658
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917



Tgt Ion:196 Resp: 149278
Ion Ratio Lower Upper
196 100
198 97.7 76.8 115.2
200 30.9 25.0 37.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#43
1,1'-Biphenyl
Concen: 29.104 ng/ul
RT: 13.075 min Scan# 1715
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14


Tgt Ion:154 Resp: 605233
Ion Ratio Lower Upper
154 100
153 41.2 31.9 47.9
76 14.2 12.2 18.2





#44
2-Chloronaphthalene
Concen: 29.612 ng/ul
RT: 13.116 min Scan# 1722
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917




Tgt Ion:162 Resp: 462019
Ion Ratio Lower Upper
162 100
164 32.9 25.6 38.4
127 48.0 34.6 51.8

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022




#45
2-Nitroaniline
Concen: 43.944 ng/ul
RT: 13.340 min Scan# 1760
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14



Tgt Ion: 65 Resp: 174138
Ion Ratio Lower Upper
65 100
92 64.5 51.4 77.2
138 88.3 67.4 101.0





#47

Dimethylphthalate

Concen: 33.511 ng/ul

RT: 13.722 min Scan# 1825

Delta R.T. -0.006 min

Lab File: BP011425.D

Acq: 12 Aug 2022 15:14

Instrument :

BNA_P

ClientSampleId :

SLCS917





Tgt Ion:	163	Resp:	62367
Ion Ratio	Lower	Upper	
163	100		
194	3.6	2.9	4.3
164	10.4	8.2	12.4

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/13/2022

Supervised By :Sohil Jodhani 08/13/2022



#48

2,6-Dinitrotoluene

Concen: 41.315 ng/ul




RT: 13.845 min Scan# 1846


Delta R.T. 0.000 min

Lab File: BP011425.D

Acq: 12 Aug 2022 15:14

Tgt Ion:	165	Resp:	120225
Ion Ratio	Lower	Upper	
165	100		
89	81.7	65.8	98.6
121	27.6	24.2	36.2






#50
Acenaphthylene
Concen: 31.281 ng/ul
RT: 13.969 min Scan# 1867
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14

Instrument :

BNA_P

ClientSampleId :

SLCS917






Tgt Ion:152 Resp: 797564
Ion Ratio Lower Upper
152 100
151 20.1 15.8 23.8
153 13.1 10.5 15.7

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#51
3-Nitroaniline
Concen: 41.282 ng/ul
RT: 14.175 min Scan# 1902
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14


Tgt Ion:138 Resp: 132145
Ion Ratio Lower Upper
138 100
108 12.0 9.9 14.9
92 133.7 106.8 160.2





#52
Acenaphthene
Concen: 30.929 ng/ul
RT: 14.316 min Scan# 1926
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917



Tgt Ion:153 Resp: 525940
Ion Ratio Lower Upper
153 100
152 47.6 39.5 59.3
154 86.7 68.6 103.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#53
2,4-Dinitrophenol
Concen: 33.493 ng/ul
RT: 14.387 min Scan# 1938
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14


Tgt Ion:184 Resp: 57826
Ion Ratio Lower Upper
184 100
63 104.9 75.0 112.6
154 90.0 53.1 79.7#





#55
4-Nitrophenol
Concen: 41.682 ng/ul
RT: 14.498 min Scan# 1957
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917



Tgt Ion:109 Resp: 12977
Ion Ratio Lower Upper
109 100
139 90.3 100.6 150.8
65 117.1 122.6 184.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#56
Dibenzofuran
Concen: 31.178 ng/ul
RT: 14.657 min Scan# 1984
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14


Tgt Ion:168 Resp: 718747
Ion Ratio Lower Upper
168 100
139 48.4 34.2 51.2





#57
2,4-Dinitrotoluene
Concen: 40.847 ng/ul
RT: 14.640 min Scan# 1981
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917




Tgt Ion:165 Resp: 182189
Ion Ratio Lower Upper
165 100
63 70.8 50.6 75.8
182 2.3 2.2 3.2

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022




#58
2,3,4,6-Tetrachlorophenol
Concen: 36.564 ng/ul
RT: 14.887 min Scan# 2023
Delta R.T. -0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14



Tgt Ion:232 Resp: 128064
Ion Ratio Lower Upper
232 100
131 69.0 50.4 75.6
130 4.1 2.6 3.8#
166 39.0 31.9 47.9





#59

Diethylphthalate

Concen: 37.109 ng/ul

RT: 15.104 min Scan# 2060

Delta R.T. 0.000 min

Lab File: BP011425.D

Acq: 12 Aug 2022 15:14

Instrument :

BNA_P

ClientSampleId :

SLCS917

Tgt Ion:149 Resp: 715459





Ion	Ratio	Lower	Upper
149	100		
177	18.7	16.1	24.1
150	12.6	9.4	14.0

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/13/2022

Supervised By :Sohil Jodhani 08/13/2022



#61

Fluorene

Concen: 32.340 ng/ul

RT: 15.310 min Scan# 2095




Delta R.T. -0.006 min


Lab File: BP011425.D

Acq: 12 Aug 2022 15:14

Tgt Ion:166 Resp: 605473


Ion	Ratio	Lower	Upper
166	100		
165	99.6	77.7	116.5
167	13.5	11.0	16.4





#62
4-Chlorophenyl-phenylether
Concen: 32.942 ng/ul
RT: 15.316 min Scan# 2096
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917




Tgt Ion:204 Resp: 269591
Ion Ratio Lower Upper
204 100
206 33.0 27.3 40.9
141 82.0 57.7 86.5

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022




#63
4-Nitroaniline
Concen: 47.494 ng/ul m
RT: 15.351 min Scan# 2102
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Tgt Ion:138 Resp: 147936
Ion Ratio Lower Upper
138 100
92 63.8 55.9 83.9
108 117.1 67.1 100.7#





#66
4,6-Dinitro-2-methylphenol
Concen: 35.116 ng/ul m
RT: 15.404 min Scan# 2111
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917



Tgt Ion:198 Resp: 94543
Ion Ratio Lower Upper
198 100
182 5.0 7.4 11.0
77 43.9 28.6 42.8




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#67
N-Nitrosodiphenylamine
Concen: 31.330 ng/ul
RT: 15.534 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14


Tgt Ion:169 Resp: 519695
Ion Ratio Lower Upper
169 100
168 67.2 53.6 80.4
167 37.3 30.0 45.0





#68
4-Bromophenyl-phenylether
Concen: 31.914 ng/ul
RT: 16.216 min Scan# 2249
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917



Tgt Ion:248 Resp: 158973
Ion Ratio Lower Upper
248 100
250 97.2 77.3 115.9
141 112.8 76.4 114.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#69
Hexachlorobenzene
Concen: 31.938 ng/ul
RT: 16.316 min Scan# 2266
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14

Tgt Ion:284 Resp: 185865
Ion Ratio Lower Upper
284 100
142 53.9 39.5 59.3
249 34.5 25.1 37.7





#70

Atrazine

Concen: 28.313 ng/ul

RT: 16.510 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP011425.D

Acq: 12 Aug 2022 15:14

Instrument :

BNA_P

ClientSampleId :

SLCS917

Tgt Ion:200 Resp: 15654

Ion Ratio Lower Upper

200 100

173 26.9 22.6 34.0





215 43.6 36.5 54.7

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/13/2022

Supervised By :Sohil Jodhani 08/13/2022



#71

Pentachlorophenol

Concen: 33.086 ng/ul

RT: 16.675 min Scan# 2327

Delta R.T. 0.000 min

Lab File: BP011425.D

Acq: 12 Aug 2022 15:14




Tgt Ion:266 Resp: 111778


Ion Ratio Lower Upper

266 100

264 62.1 52.2 78.2

268 63.8 50.6 75.8









#72
Phenanthrene
Concen: 30.568 ng/ul m
RT: 17.069 min Scan# 2394
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14

Instrument :
BNA_P
Client SampleId :
SLCS917

Tgt Ion:178 Resp: 96486
Ion Ratio Lower Upper
178 100
179 15.1 12.1 18.1
176 19.4 15.4 23.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#74
Anthracene
Concen: 31.159 ng/ul
RT: 17.163 min Scan# 2410
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14


Tgt Ion:178 Resp: 979115
Ion Ratio Lower Upper
178 100
179 15.1 12.3 18.5
176 18.9 14.7 22.1





#75
1,2,3,4-Tetrachlorobenzene
Concen: 25.946 ng/uL
RT: 13.034 min Scan# 1708
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917




Tgt Ion:216 Resp: 209673
Ion Ratio Lower Upper
216 100
214 79.3 62.0 93.0
218 47.9 38.2 57.4

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022




#76
Pentachlorobenzene
Concen: 29.612 ng/uL
RT: 14.569 min Scan# 1969
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Tgt Ion:250 Resp: 208402
Ion Ratio Lower Upper
250 100
248 62.2 49.2 73.8
252 63.4 50.8 76.2





#77
 Carbazole
 Concen: 31.047 ng/ul
 RT: 17.445 min Scan# 2457
 Delta R.T. 0.000 min
 Lab File: BP011425.D
 Acq: 12 Aug 2022 15:14




Instrument :
 BNA_P
 ClientSampleId :
 SLCS917




Tgt Ion:167 Resp: 911631
 Ion Ratio Lower Upper
 167 100
 166 22.2 17.1 25.7
 139 15.4 11.5 17.3

Manual Integrations
 APPROVED



Reviewed By :Yogesh Patel 08/13/2022
 Supervised By :Sohil Jodhani 08/13/2022




#78
 Di-n-butylphthalate
 Concen: 39.278 ng/ul
 RT: 18.016 min Scan# 2555
 Delta R.T. 0.000 min
 Lab File: BP011425.D
 Acq: 12 Aug 2022 15:14




Tgt Ion:149 Resp: 1261124
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.4 11.0
 104 6.6 5.1 7.7





#80
Fluoranthene
Concen: 30.452 ng/u1
RT: 19.063 min Scan# 2732
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917




Tgt Ion:202 Resp: 1138175
Ion Ratio Lower Upper
202 100
101 12.1 11.9 17.9
100 9.4 9.3 13.9

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022




#82
Pyrene
Concen: 29.730 ng/u1
RT: 19.416 min Scan# 2793
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14



Tgt Ion:202 Resp: 1172868
Ion Ratio Lower Upper
202 100
101 14.9 14.3 21.5
100 12.6 11.8 17.6





#83

Butylbenzylphthalate

Concen: 46.194 ng/ul

RT: 20.316 min Scan# 2946

Delta R.T. 0.000 min

Lab File: BP011425.D

Acq: 12 Aug 2022 15:14

Instrument :

BNA_P

ClientSampleId :

SLCS917





Tgt Ion:	149	Resp:	580738
Ion Ratio	Lower	Upper	
149	100		
91	84.5	64.2	96.4
206	16.2	14.0	21.0

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/13/2022

Supervised By :Sohil Jodhani 08/13/2022



#84

3,3'-Dichlorobenzidine

Concen: 35.625 ng/ul




RT: 21.086 min Scan# 3077


Delta R.T. 0.000 min

Lab File: BP011425.D

Acq: 12 Aug 2022 15:14


Tgt Ion:	252	Resp:	361371
Ion Ratio	Lower	Upper	
252	100		
254	64.9	51.8	77.6
126	15.2	11.4	17.0





#85
Benzo(a)anthracene
Concen: 31.263 ng/ul
RT: 21.145 min Scan# 3086
Delta R.T. 0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917



Tgt Ion:228 Resp: 111792
Ion Ratio Lower Upper
228 100
229 19.2 15.8 23.6
226 27.5 21.8 32.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 43.245 ng/ul
RT: 21.086 min Scan# 3077
Delta R.T. 0.000 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14


Tgt Ion:149 Resp: 872760
Ion Ratio Lower Upper
149 100
167 24.4 20.4 30.6
279 3.2 2.9 4.3





#87
Chrysene
Concen: 31.223 ng/ul
RT: 21.198 min Scan# 3095
Delta R.T. 0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917



Tgt Ion:228 Resp: 110524
Ion Ratio Lower Upper
228 100
226 29.9 23.8 35.6
229 19.0 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#89
Di-n-octyl phthalate
Concen: 38.155 ng/ul
RT: 21.986 min Scan# 3230
Delta R.T. 0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14

Tgt Ion:149 Resp: 1469964









#90
Benzo(b)fluoranthene
Concen: 30.964 ng/ul
RT: 22.768 min Scan# 3362
Delta R.T. 0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14

Instrument :
BNA_P
ClientSampleId :
SLCS917

Tgt Ion:252 Resp: 1131597
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.0
125 11.5 10.8 16.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#91
Benzo(k)fluoranthene
Concen: 31.172 ng/ul
RT: 22.816 min Scan# 3371
Delta R.T. 0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14


Tgt Ion:252 Resp: 1114609
Ion Ratio Lower Upper
252 100
253 21.8 17.6 26.4
125 11.4 11.0 16.6





#93
Benzo(a)pyrene
Concen: 31.847 ng/ul
RT: 23.368 min Scan# 3464
Delta R.T. 0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14




Instrument :
BNA_P
ClientSampleId :
SLCS917




Tgt Ion:252 Resp: 1122528
Ion Ratio Lower Upper
252 100
253 21.4 17.4 26.2
125 12.3 12.5 18.7

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022




#94
Indeno(1,2,3-cd)pyrene
Concen: 31.878 ng/ul
RT: 25.851 min Scan# 3887
Delta R.T. 0.012 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14



Tgt Ion:276 Resp: 1242155
Ion Ratio Lower Upper
276 100
138 26.2 25.5 38.3
277 25.2 19.8 29.8





#95
Dibenzo(a,h)anthracene
Concen: 32.022 ng/ul
RT: 25.868 min Scan# 3888
Delta R.T. 0.006 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14

Instrument :

BNA_P

ClientSampleId :





SLCS917

Tgt Ion:278 Resp: 1061207
Ion Ratio Lower Upper
278 100
139 18.3 18.5 27.7
279 23.4 19.6 29.4

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#96
Benzo(g,h,i)perylene
Concen: 31.903 ng/ul
RT: 26.580 min Scan# 4011
Delta R.T. 0.012 min
Lab File: BP011425.D
Acq: 12 Aug 2022 15:14

Tgt Ion:276 Resp: 1048988
Ion Ratio Lower Upper
276 100
138 24.1 22.8 34.2
277 24.1 18.5 27.7

