

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081122\
 Data File : BP011441.D
 Acq On : 13 Aug 2022 00:59
 Operator : CG/JU
 Sample : PB146958BS
 Misc :
 ALS Vial : 66 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS958

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 08/13/2022
 Supervised By :Sohil Jodhani 08/13/2022

Quant Time: Aug 13 01:26:41 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 11 23:05:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.569	152	88351	20.000	ng/u1	0.00
20) Naphthalene-d8	10.363	136	389999	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.245	164	232292	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.022	188	476261	20.000	ng/u1	0.00
79) Chrysene-d12	21.151	240	461696	20.000	ng/u1	0.00
88) Perylene-d12	23.457	264	462190	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.081	96	17844m	6.191	ng/uL	0.00
4) Pyridine-d5	3.493	84	130407	17.938	ng/u1	0.00
7) Phenol-d5	6.758	99	263112	33.651	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.922	67	174993	31.565	ng/u1	0.00
11) 2-Chlorophenol-d4	7.105	132	209499	34.149	ng/u1	0.00
15) 4-Methylphenol-d8	8.299	113	210587	34.302	ng/u1	0.00
21) Nitrobenzene-d5	8.740	128	101375	38.194	ng/u1	0.00
24) 2-Nitrophenol-d4	9.457	143	104280	42.912	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.993	165	186613	38.027	ng/u1	0.00
31) 4-Chloroaniline-d4	10.516	131	151220	17.626	ng/u1	0.00
46) Dimethylphthalate-d6	13.675	166	591690	36.711	ng/u1	0.00
49) Acenaphthylene-d8	13.939	160	673988	34.739	ng/u1	0.00
54) 4-Nitrophenol-d4	14.481	143	106726	36.254	ng/u1	0.00
60) Fluorene-d10	15.251	176	491897	35.687	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.386	200	87011	39.333	ng/u1	0.00
73) Anthracene-d10	17.122	188	779697	36.423	ng/u1	0.00
81) Pyrene-d10	19.386	212	868770	34.154	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.310	264	840049	36.040	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.117	88	35336	11.543	ng/uL#	82
5) Pyridine	3.511	79	218655	28.875	ng/u1	85
6) Benzaldehyde	6.728	77	142604	41.850	ng/u1	98
8) Phenol	6.787	94	258767	30.575	ng/u1#	88
10) Bis(2-Chloroethyl)ether	7.016	93	206367	30.277	ng/u1	96
12) 2-Chlorophenol	7.140	128	214056	33.485	ng/u1	98
13) 2-Methylphenol	8.028	108	194004	31.829	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.116	45	308497	30.177	ng/u1#	95
16) Acetophenone	8.404	105	330915	33.316	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.393	70	178405	36.460	ng/u1	98
18) 4-Methylphenol	8.357	108	209663	32.092	ng/u1	98
19) Hexachloroethane	8.634	117	98443	36.471	ng/u1	95
22) Nitrobenzene	8.781	77	265109	35.680	ng/u1	97
23) Isophorone	9.310	82	483551	35.479	ng/u1	99
25) 2-Nitrophenol	9.487	139	112778	39.822	ng/u1#	86
26) 2,4-Dimethylphenol	9.557	107	256845	35.892	ng/u1	94
27) Bis(2-Chloroethoxy)met...	9.799	93	278980	32.029	ng/u1	98
29) 2,4-Dichlorophenol	10.016	162	180369	35.147	ng/u1	100
30) Naphthalene	10.410	128	667907	31.478	ng/u1	98
32) 4-Chloroaniline	10.540	127	232332	27.016	ng/u1	100
33) Hexachlorobutadiene	10.693	225	114363	36.325	ng/u1	99
34) Caprolactam	11.340	113	61572m	35.878	ng/u1	
35) 4-Chloro-3-methylphenol	11.681	107	227626	37.725	ng/u1	97
36) 2-Methylnaphthalene	12.040	142	444000	33.758	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081122\
 Data File : BP011441.D
 Acq On : 13 Aug 2022 00:59
 Operator : CG/JU
 Sample : PB146958BS
 Misc :
 ALS Vial : 66 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS958


Quant Time: Aug 13 01:26:41 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP080822.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 11 23:05:53 2022
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Yogesh Patel 08/13/2022
 Supervised By :Sohil Jodhani 08/13/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.263	142	442441	33.164	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.410	216	196767	32.176	ng/ul	95
40) Hexachlorocyclopentadiene	12.387	237	114521	34.947	ng/ul	99
41) 2,4,6-Trichlorophenol	12.669	196	132326	35.762	ng/ul	100
42) 2,4,5-Trichlorophenol	12.740	196	143724	36.393	ng/ul	99
43) 1,1'-Biphenyl	13.075	154	569441	31.218	ng/ul	98
44) 2-Chloronaphthalene	13.110	162	434170	31.724	ng/ul	96
45) 2-Nitroaniline	13.334	65	159785	45.969	ng/ul	98
47) Dimethylphthalate	13.722	163	575425	35.249	ng/ul	99
48) 2,6-Dinitrotoluene	13.845	165	111150	43.546	ng/ul	97
50) Acenaphthylene	13.969	152	741028	33.134	ng/ul	99
51) 3-Nitroaniline	14.175	138	116302	41.420	ng/ul	97
52) Acenaphthene	14.310	153	491406	32.945	ng/ul	98
53) 2,4-Dinitrophenol	14.381	184	53387	35.252	ng/ul#	77
55) 4-Nitrophenol	14.492	109	119427	43.732	ng/ul#	74
56) Dibenzofuran	14.651	168	660703	32.673	ng/ul	90
57) 2,4-Dinitrotoluene	14.634	165	167257	42.751	ng/ul	89
58) 2,3,4,6-Tetrachlorophenol	14.886	232	115587	37.623	ng/ul#	95
59) Diethylphthalate	15.098	149	661019	39.086	ng/ul	97
61) Fluorene	15.310	166	555367	33.818	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.310	204	248981	34.684	ng/ul	90
63) 4-Nitroaniline	15.351	138	131738m	48.217	ng/ul	
66) 4,6-Dinitro-2-methylph...	15.404	198	87497	38.411	ng/ul#	93
67) N-Nitrosodiphenylamine	15.528	169	473046	33.706	ng/ul	100
68) 4-Bromophenyl-phenylether	16.210	248	146887	34.853	ng/ul#	88
69) Hexachlorobenzene	16.310	284	170319	34.592	ng/ul	91
70) Atrazine	16.504	200	78820	16.849	ng/ul	96
71) Pentachlorophenol	16.669	266	97356	34.060	ng/ul	98
72) Phenanthrene	17.063	178	881393	33.004	ng/ul	99
74) Anthracene	17.157	178	904661	34.028	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.028	216	196564	28.749	ng/uL	97
76) Pentachlorobenzene	14.563	250	193487	32.494	ng/uL	99
77) Carbazole	17.439	167	816338	32.860	ng/ul	98
78) Di-n-butylphthalate	18.010	149	1150074	42.336	ng/ul	99
80) Fluoranthene	19.057	202	1021464	32.992	ng/ul#	93
82) Pyrene	19.410	202	1065250	32.598	ng/ul	94
83) Butylbenzylphthalate	20.310	149	525551	50.467	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.080	252	258908	30.813	ng/ul	99
85) Benzo(a)anthracene	21.139	228	990735	33.448	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.080	149	786325	47.036	ng/ul	98
87) Chrysene	21.186	228	984198	33.565	ng/ul	99
89) Di-n-octyl phthalate	21.980	149	1320736	41.185	ng/ul	100
90) Benzo(b)fluoranthene	22.757	252	1021585	33.583	ng/ul	97
91) Benzo(k)fluoranthene	22.804	252	966683	32.480	ng/ul	97
93) Benzo(a)pyrene	23.357	252	992713	33.835	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	25.827	276	1083059	33.392	ng/ul	94
95) Dibenzo(a,h)anthracene	25.851	278	934744	33.886	ng/ul	95
96) Benzo(g,h,i)perylene	26.556	276	917822	33.535	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#2
1,4-Dioxane
Concen: 11.543 ng/uL
RT: 3.117 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS958

Tgt Ion: 88 Resp: 35336
Ion Ratio Lower Upper
88 100
43 37.6 29.6 44.4
58 82.7 49.1 73.7




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#5
Pyridine
Concen: 28.875 ng/uL
RT: 3.511 min Scan# 89
Delta R.T. 0.000 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59


Tgt Ion: 79 Resp: 218655
Ion Ratio Lower Upper
79 100
52 86.0 57.8 86.6
51 46.2 31.5 47.3





#6
Benzaldehyde
Concen: 41.850 ng/ul
RT: 6.728 min Scan# 636
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS958




Tgt Ion: 77 Resp: 142604
Ion Ratio Lower Upper
77 100
105 92.8 72.0 108.0
106 87.1 69.4 104.0

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022




#8
Phenol
Concen: 30.575 ng/ul
RT: 6.787 min Scan# 646
Delta R.T. 0.000 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59



Tgt Ion: 94 Resp: 258767
Ion Ratio Lower Upper
94 100
65 36.1 24.7 37.1
66 49.0 32.6 49.0#









#10
Bis(2-Chloroethyl)ether
Concen: 30.277 ng/ul
RT: 7.016 min Scan# 685
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS958

Tgt Ion: 93 Resp: 206362
Ion Ratio Lower Upper
93 100
63 83.5 71.0 106.6
95 32.8 26.0 39.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#12
2-Chlorophenol
Concen: 33.485 ng/ul
RT: 7.140 min Scan# 706
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Tgt Ion:128 Resp: 214056
Ion Ratio Lower Upper
128 100
64 53.0 43.8 65.6
130 32.4 25.4 38.2





#13

2-Methylphenol

Concen: 31.829 ng/ul

RT: 8.028 min Scan# 857

Delta R.T. -0.006 min

Lab File: BP011441.D

Acq: 13 Aug 2022 00:59

Instrument :

BNA_P

ClientSampleId :

SLCS958

Tgt Ion:108 Resp: 194004

Ion Ratio Lower Upper

108 100





107 88.9 71.8 107.6

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/13/2022

Supervised By :Sohil Jodhani 08/13/2022



#14

2,2'-oxybis(1-Chloropropane)

Concen: 30.177 ng/ul

RT: 8.116 min Scan# 872

Delta R.T. -0.006 min

Lab File: BP011441.D

Acq: 13 Aug 2022 00:59




Tgt Ion: 45 Resp: 308497


Ion Ratio Lower Upper

45 100

77 14.0 9.9 14.9

79 11.1 7.2 10.8









#16
Acetophenone
Concen: 33.316 ng/ul
RT: 8.404 min Scan# 919
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS958

Tgt Ion: 105 Resp: 330915
Ion Ratio Lower Upper
105 100
77 85.1 65.4 98.0
51 34.4 26.5 39.7




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#17
N-Nitroso-di-n-propylamine
Concen: 36.460 ng/ul
RT: 8.393 min Scan# 919
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Tgt Ion: 70 Resp: 178405
Ion Ratio Lower Upper
70 100
42 67.7 55.5 83.3
101 9.0 7.1 10.7
130 18.1 14.8 22.2









#18
4-Methylphenol
Concen: 32.092 ng/ul
RT: 8.357 min Scan# 914
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS958

Tgt Ion:108 Resp: 209663
Ion Ratio Lower Upper
108 100
107 112.4 88.2 132.4




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#19
Hexachloroethane
Concen: 36.471 ng/ul
RT: 8.634 min Scan# 960
Delta R.T. -0.012 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59


Tgt Ion:117 Resp: 98443
Ion Ratio Lower Upper
117 100
201 72.9 62.2 93.2
199 47.7 39.8 59.6





#22
Nitrobenzene
Concen: 35.680 ng/ul
RT: 8.781 min Scan# 985
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS958




Tgt Ion: 77 Resp: 265109
Ion Ratio Lower Upper
77 100
123 41.2 31.4 47.0
65 14.3 10.6 15.8

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022




#23
Isophorone
Concen: 35.479 ng/ul
RT: 9.310 min Scan# 1075
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59



Tgt Ion: 82 Resp: 483551
Ion Ratio Lower Upper
82 100
95 8.2 5.8 8.8
138 15.9 12.6 18.8









#25
2-Nitrophenol
Concen: 39.822 ng/ul
RT: 9.487 min Scan# 1105
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS958

Tgt Ion:139 Resp: 112778
Ion Ratio Lower Upper
139 100
65 65.7 46.7 70.1
109 42.2 24.2 36.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#26
2,4-Dimethylphenol
Concen: 35.892 ng/ul
RT: 9.557 min Scan# 1117
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59


Tgt Ion:107 Resp: 256845
Ion Ratio Lower Upper
107 100
121 48.1 40.3 60.5
122 77.9 67.7 101.5





#27
Bis(2-Chloroethoxy)methane
Concen: 32.029 ng/ul
RT: 9.799 min Scan# 1158
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS958




Tgt Ion: 93 Resp: 278980
Ion Ratio Lower Upper
93 100
95 32.8 25.2 37.8
123 10.8 8.4 12.6

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022




#29
2,4-Dichlorophenol
Concen: 35.147 ng/ul
RT: 10.016 min Scan# 1195
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59



Tgt Ion:162 Resp: 180369
Ion Ratio Lower Upper
162 100
164 66.0 52.6 79.0
98 42.6 34.2 51.4









#30
Naphthalene
Concen: 31.478 ng/ul
RT: 10.410 min Scan# 1263
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS958

Tgt Ion:128 Resp: 667907
Ion Ratio Lower Upper
128 100
129 10.9 9.0 13.4
127 13.8 10.4 15.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#32
4-Chloroaniline
Concen: 27.016 ng/ul
RT: 10.540 min Scan# 1284
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59


Tgt Ion:127 Resp: 232332
Ion Ratio Lower Upper
127 100
129 33.3 26.6 40.0





#33
Hexachlorobutadiene
Concen: 36.325 ng/ul
RT: 10.693 min Scan# 1310
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS958



Tgt Ion:225 Resp: 114363
Ion Ratio Lower Upper
225 100
223 62.8 51.2 76.8
227 63.5 50.6 75.8




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#34
Caprolactam
Concen: 35.878 ng/ul m
RT: 11.340 min Scan# 1420
Delta R.T. 0.000 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59


Tgt Ion:113 Resp: 61572
Ion Ratio Lower Upper
113 100
55 191.3 164.2 246.4
56 149.2 127.0 190.6





#35
4-Chloro-3-methylphenol
Concen: 37.725 ng/ul
RT: 11.681 min Scan# 1478
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS958



Tgt Ion:107 Resp: 227626
Ion Ratio Lower Upper
107 100
144 22.7 19.0 28.4
142 72.1 59.8 89.8




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#36
2-Methylnaphthalene
Concen: 33.758 ng/ul
RT: 12.040 min Scan# 1539
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59


Tgt Ion:142 Resp: 444000
Ion Ratio Lower Upper
142 100
141 90.3 71.8 107.8





#37
1-Methylnaphthalene
Concen: 33.164 ng/ul
RT: 12.263 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS958




Tgt Ion:142 Resp: 442441
Ion Ratio Lower Upper
142 100
141 92.2 74.1 111.1
116 4.9 3.4 5.0

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022




#39
1,2,4,5-Tetrachlorobenzene
Concen: 32.176 ng/ul
RT: 12.410 min Scan# 1602
Delta R.T. -0.012 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59



Tgt Ion:216 Resp: 196767
Ion Ratio Lower Upper
216 100
214 79.1 61.1 91.7
179 25.9 18.6 27.8
108 28.0 19.1 28.7









#40
Hexachlorocyclopentadiene
Concen: 34.947 ng/ul
RT: 12.387 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS958

Tgt Ion:237 Resp: 11452
Ion Ratio Lower Upper
237 100
235 61.9 50.4 75.6
272 11.9 9.9 14.9




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#41
2,4,6-Trichlorophenol
Concen: 35.762 ng/ul
RT: 12.669 min Scan# 1646
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Tgt Ion:196 Resp: 132326
Ion Ratio Lower Upper
196 100
198 94.3 75.5 113.3
200 31.2 24.4 36.6





#42

2,4,5-Trichlorophenol

Concen: 36.393 ng/ul

RT: 12.740 min Scan# 1658

Delta R.T. -0.006 min

Lab File: BP011441.D

Acq: 13 Aug 2022 00:59

Instrument :

BNA_P





ClientSampleId :

SLCS958

Tgt Ion:	196	Resp:	143724
Ion Ratio	Lower	Upper	
196	100		
198	96.1	76.8	115.2
200	30.2	25.0	37.6

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#43

1,1'-Biphenyl

Concen: 31.218 ng/ul




RT: 13.075 min Scan# 1715


Delta R.T. -0.006 min

Lab File: BP011441.D

Acq: 13 Aug 2022 00:59

Tgt Ion:	154	Resp:	569441
Ion Ratio	Lower	Upper	
154	100		
153	41.0	31.9	47.9
76	13.7	12.2	18.2





#44
2-Chloronaphthalene
Concen: 31.724 ng/ul
RT: 13.110 min Scan# 1722
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Instrument :

BNA_P

ClientSampleId :





SLCS958

Tgt Ion:162 Resp: 434170
Ion Ratio Lower Upper
162 100
164 32.6 25.6 38.4
127 47.4 34.6 51.8

Manual Integrations




APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#45
2-Nitroaniline
Concen: 45.969 ng/ul
RT: 13.334 min Scan# 1759
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Tgt Ion: 65 Resp: 159785
Ion Ratio Lower Upper
65 100
92 62.6 51.4 77.2
138 85.4 67.4 101.0









#47
Dimethylphthalate
Concen: 35.249 ng/ul
RT: 13.722 min Scan# 1825
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS958

Tgt Ion:163 Resp: 575425
Ion Ratio Lower Upper
163 100
194 3.5 2.9 4.3
164 10.0 8.2 12.4




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#48
2,6-Dinitrotoluene
Concen: 43.546 ng/ul
RT: 13.845 min Scan# 1846
Delta R.T. 0.000 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Tgt Ion:165 Resp: 111150
Ion Ratio Lower Upper
165 100
89 80.5 65.8 98.6
121 26.4 24.2 36.2





#50

Acenaphthylene

Concen: 33.134 ng/ul

RT: 13.969 min Scan# 1867

Delta R.T. -0.006 min

Lab File: BP011441.D

Acq: 13 Aug 2022 00:59

Instrument :

BNA_P

ClientSampleId :

SLCS958

Tgt Ion:152 Resp: 741028

Ion Ratio Lower Upper

152 100

151 20.1 15.8 23.8





153 12.9 10.5 15.7

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/13/2022

Supervised By :Sohil Jodhani 08/13/2022



#51

3-Nitroaniline

Concen: 41.420 ng/ul

RT: 14.175 min Scan# 1902

Delta R.T. -0.006 min

Lab File: BP011441.D

Acq: 13 Aug 2022 00:59




Tgt Ion:138 Resp: 116302


Ion Ratio Lower Upper

138 100

108 12.4 9.9 14.9


92 129.4 106.8 160.2





#52
Acenaphthene
Concen: 32.945 ng/ul
RT: 14.310 min Scan# 1926
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS958




Tgt Ion:153 Resp: 491406
Ion Ratio Lower Upper
153 100
152 48.4 39.5 59.3
154 87.5 68.6 103.0

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022




#53
2,4-Dinitrophenol
Concen: 35.252 ng/ul
RT: 14.381 min Scan# 1937
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59



Tgt Ion:184 Resp: 53387
Ion Ratio Lower Upper
184 100
63 112.8 75.0 112.6#
154 89.2 53.1 79.7#









#55
4-Nitrophenol
Concen: 43.732 ng/ul
RT: 14.492 min Scan# 1956
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS958

Tgt Ion:109 Resp: 11942
Ion Ratio Lower Upper
109 100
139 92.8 100.6 150.8
65 122.9 122.6 184.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#56
Dibenzofuran
Concen: 32.673 ng/ul
RT: 14.651 min Scan# 1983
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Tgt Ion:168 Resp: 660703
Ion Ratio Lower Upper
168 100
139 49.1 34.2 51.2









#57
2,4-Dinitrotoluene
Concen: 42.751 ng/ul
RT: 14.634 min Scan# 1981
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS958

Tgt Ion:165 Resp: 16725
Ion Ratio Lower Upper
165 100
63 72.3 50.6 75.8
182 2.7 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#58
2,3,4,6-Tetrachlorophenol
Concen: 37.623 ng/ul
RT: 14.886 min Scan# 2023
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Tgt Ion:232 Resp: 115587
Ion Ratio Lower Upper
232 100
131 68.8 50.4 75.6
130 4.4 2.6 3.8#
166 39.0 31.9 47.9





#59
Diethylphthalate
Concen: 39.086 ng/ul
RT: 15.098 min Scan# 2059
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Instrument :

BNA_P





ClientSampleId :

SLCS958

Tgt Ion:149 Resp: 661019
Ion Ratio Lower Upper
149 100
177 18.8 16.1 24.1
150 12.5 9.4 14.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#61
Fluorene
Concen: 33.818 ng/ul
RT: 15.310 min Scan# 2095
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Tgt Ion:166 Resp: 555367
Ion Ratio Lower Upper
166 100
165 99.5 77.7 116.5
167 13.5 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 34.684 ng/ul
RT: 15.310 min Scan# 2095
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS958

Tgt Ion:204 Resp: 24898
Ion Ratio Lower Upper
204 100
206 33.4 27.3 40.9
141 84.2 57.7 86.5




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#63
4-Nitroaniline
Concen: 48.217 ng/ul m
RT: 15.351 min Scan# 2102
Delta R.T. 0.000 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Tgt Ion:138 Resp: 131738
Ion Ratio Lower Upper
138 100
92 61.7 55.9 83.9
108 119.3 67.1 100.7#









#66
4,6-Dinitro-2-methylphenol
Concen: 38.411 ng/ul
RT: 15.404 min Scan# 2111
Delta R.T. -0.000 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS958

Tgt Ion:198 Resp: 87497
Ion Ratio Lower Upper
198 100
182 4.2 7.4 11.0
77 39.1 28.6 42.8




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#67
N-Nitrosodiphenylamine
Concen: 33.706 ng/ul
RT: 15.528 min Scan# 2132
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59


Tgt Ion:169 Resp: 473046
Ion Ratio Lower Upper
169 100
168 66.8 53.6 80.4
167 37.6 30.0 45.0





#68
4-Bromophenyl-phenylether
Concen: 34.853 ng/ul
RT: 16.210 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS958



Tgt Ion:248 Resp: 14688
Ion Ratio Lower Upper
248 100
250 98.3 77.3 115.9
141 117.6 76.4 114.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#69
Hexachlorobenzene
Concen: 34.592 ng/ul
RT: 16.310 min Scan# 2265
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Tgt Ion:284 Resp: 170319
Ion Ratio Lower Upper
284 100
142 57.6 39.5 59.3
249 34.4 25.1 37.7





#70

Atrazine

Concen: 16.849 ng/ul

RT: 16.504 min Scan# 2298

Delta R.T. -0.006 min

Lab File: BP011441.D

Acq: 13 Aug 2022 00:59

Instrument :

BNA_P

ClientSampleId :

SLCS958





Tgt Ion:	200	Resp:	78820
Ion Ratio	Lower	Upper	
200	100		
173	26.3	22.6	34.0
215	43.1	36.5	54.7

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/13/2022

Supervised By :Sohil Jodhani 08/13/2022



#71

Pentachlorophenol

Concen: 34.060 ng/ul




RT: 16.669 min Scan# 2326


Delta R.T. -0.006 min

Lab File: BP011441.D

Acq: 13 Aug 2022 00:59

Tgt Ion:	266	Resp:	97356
Ion Ratio	Lower	Upper	
266	100		
264	63.1	52.2	78.2
268	64.2	50.6	75.8









#72
Phenanthrene
Concen: 33.004 ng/ul
RT: 17.063 min Scan# 2393
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS958

Tgt Ion:178 Resp: 881393
Ion Ratio Lower Upper
178 100
179 15.3 12.1 18.1
176 20.0 15.4 23.2




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#74
Anthracene
Concen: 34.028 ng/ul
RT: 17.157 min Scan# 2409
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Tgt Ion:178 Resp: 904661
Ion Ratio Lower Upper
178 100
179 15.0 12.3 18.5
176 18.9 14.7 22.1





#75

1,2,3,4-Tetrachlorobenzene

Concen: 28.749 ng/uL

RT: 13.028 min Scan# 1707

Delta R.T. -0.006 min

Lab File: BP011441.D

Acq: 13 Aug 2022 00:59

Instrument :

BNA_P

ClientSampleId :

SLCS958





Tgt Ion:	216	Resp:	196564
Ion Ratio	Lower	Upper	
216	100		
214	80.9	62.0	93.0
218	47.2	38.2	57.4

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/13/2022

Supervised By :Sohil Jodhani 08/13/2022



#76

Pentachlorobenzene

Concen: 32.494 ng/uL




RT: 14.563 min Scan# 1968


Delta R.T. -0.006 min

Lab File: BP011441.D

Acq: 13 Aug 2022 00:59

Tgt Ion:	250	Resp:	193487
Ion Ratio	Lower	Upper	
250	100		
248	62.6	49.2	73.8
252	63.9	50.8	76.2









#77
 Carbazole
 Concen: 32.860 ng/ul
 RT: 17.439 min Scan# 2457
 Delta R.T. -0.006 min
 Lab File: BP011441.D
 Acq: 13 Aug 2022 00:59

Instrument :
 BNA_P
 ClientSampleId :
 SLCS958

Tgt Ion:167 Resp: 816338
 Ion Ratio Lower Upper
 167 100
 166 22.2 17.1 25.7
 139 15.9 11.5 17.3




Manual Integrations
 APPROVED


Reviewed By :Yogesh Patel 08/13/2022
 Supervised By :Sohil Jodhani 08/13/2022



#78
 Di-n-butylphthalate
 Concen: 42.336 ng/ul
 RT: 18.010 min Scan# 2554
 Delta R.T. -0.006 min
 Lab File: BP011441.D
 Acq: 13 Aug 2022 00:59

Tgt Ion:149 Resp: 1150074
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.4 11.0
 104 6.7 5.1 7.7









#80
Fluoranthene
Concen: 32.992 ng/u1
RT: 19.057 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS958

Tgt Ion:202 Resp: 1021464
Ion Ratio Lower Upper
202 100
101 12.0 11.9 17.9
100 9.2 9.3 13.9




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#82
Pyrene
Concen: 32.598 ng/u1
RT: 19.410 min Scan# 2792
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Tgt Ion:202 Resp: 1065250
Ion Ratio Lower Upper
202 100
101 14.7 14.3 21.5
100 12.6 11.8 17.6









#83
Butylbenzylphthalate
Concen: 50.467 ng/ul
RT: 20.310 min Scan# 2946
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS958

Tgt Ion:149 Resp: 52555
Ion Ratio Lower Upper
149 100
91 83.8 64.2 96.4
206 16.2 14.0 21.0




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#84
3,3'-Dichlorobenzidine
Concen: 30.813 ng/ul
RT: 21.080 min Scan# 3076
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59


Tgt Ion:252 Resp: 258908
Ion Ratio Lower Upper
252 100
254 65.4 51.8 77.6
126 14.0 11.4 17.0





#85
Benzo(a)anthracene
Concen: 33.448 ng/ul
RT: 21.139 min Scan# 3086
Delta R.T. 0.000 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS958



Tgt Ion:228 Resp: 990735
Ion Ratio Lower Upper
228 100
229 19.2 15.8 23.6
226 27.2 21.8 32.6




Manual Integrations
APPROVED


Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#86
Bis(2-ethylhexyl)phthalate
Concen: 47.036 ng/ul
RT: 21.080 min Scan# 3076
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Tgt Ion:149 Resp: 786325
Ion Ratio Lower Upper
149 100
167 24.3 20.4 30.6
279 3.2 2.9 4.3





#87

Chrysene

Concen: 33.565 ng/ul

RT: 21.186 min Scan# 3095

Delta R.T. -0.006 min

Lab File: BP011441.D

Acq: 13 Aug 2022 00:59

Instrument :

BNA_P

ClientSampleId :





SLCS958

Tgt Ion:228 Resp: 984198
 Ion Ratio Lower Upper
 228 100
 226 30.7 23.8 35.6
 229 19.9 15.7 23.5

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/13/2022
 Supervised By :Sohil Jodhani 08/13/2022



#89

Di-n-octyl phthalate

Concen: 41.185 ng/ul




RT: 21.980 min Scan# 3229


Delta R.T. -0.000 min

Lab File: BP011441.D

Acq: 13 Aug 2022 00:59


Tgt Ion:149 Resp: 1320736





#90
Benzo(b)fluoranthene
Concen: 33.583 ng/ul
RT: 22.757 min Scan# 3361
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59




Instrument :
BNA_P
ClientSampleId :
SLCS958




Tgt Ion:252 Resp: 1021585
Ion Ratio Lower Upper
252 100
253 22.5 17.4 26.0
125 11.8 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022




#91
Benzo(k)fluoranthene
Concen: 32.480 ng/ul
RT: 22.804 min Scan# 3369
Delta R.T. -0.006 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59



Tgt Ion:252 Resp: 966683
Ion Ratio Lower Upper
252 100
253 22.0 17.6 26.4
125 11.2 11.0 16.6





#93

Benzo(a)pyrene

Concen: 33.835 ng/ul

RT: 23.357 min Scan# 3464

Delta R.T. -0.006 min

Lab File: BP011441.D

Acq: 13 Aug 2022 00:59

Instrument :

BNA_P

ClientSampleId :

SLCS958





Tgt Ion:	252	Resp:	992713
Ion Ratio	Lower	Upper	
252	100		
253	21.8	17.4	26.2
125	13.3	12.5	18.7

Manual Integrations

APPROVED

Reviewed By :Yogesh Patel 08/13/2022

Supervised By :Sohil Jodhani 08/13/2022



#94

Indeno(1,2,3-cd)pyrene

Concen: 33.392 ng/ul




RT: 25.827 min Scan# 3883


Delta R.T. -0.012 min

Lab File: BP011441.D

Acq: 13 Aug 2022 00:59

Tgt Ion:	276	Resp:	1083059
Ion Ratio	Lower	Upper	
276	100		
138	26.3	25.5	38.3
277	25.0	19.8	29.8









#95
Dibenzo(a,h)anthracene
Concen: 33.886 ng/ul
RT: 25.851 min Scan# 3889
Delta R.T. -0.012 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Instrument :
BNA_P
ClientSampleId :
SLCS958

Tgt Ion:278 Resp: 934744
Ion Ratio Lower Upper
278 100
139 18.5 18.5 27.7
279 24.0 19.6 29.4

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 08/13/2022
Supervised By :Sohil Jodhani 08/13/2022



#96
Benzo(g,h,i)perylene
Concen: 33.535 ng/ul
RT: 26.556 min Scan# 4007
Delta R.T. -0.012 min
Lab File: BP011441.D
Acq: 13 Aug 2022 00:59

Tgt Ion:276 Resp: 917822
Ion Ratio Lower Upper
276 100
138 24.6 22.8 34.2
277 24.4 18.5 27.7

