

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP081221\  
 Data File : BP006660.D  
 Acq On : 12 Aug 2021 11:02  
 Operator : CG/JU  
 Sample : PB138110BS  
 Misc :  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SLCS110

Manual Integrations  
 APPROVED

mohammad  
 8/13/2021 8:47:41 AM

Quant Time: Aug 12 11:07:56 2021  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_P\METHODS\SFAM-EPA-BP081021.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Thu Aug 12 11:07:37 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units  | Dev(Min) |
|-------------------------------|--------|------|----------|--------|--------|----------|
| -----                         |        |      |          |        |        |          |
| Internal Standards            |        |      |          |        |        |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.710  | 152  | 137070   | 20.000 | ng/ul  | 0.00     |
| 20) Naphthalene-d8            | 10.498 | 136  | 589694   | 20.000 | ng/ul  | 0.00     |
| 38) Acenaphthene-d10          | 14.357 | 164  | 350592   | 20.000 | ng/ul  | 0.00     |
| 64) Phenanthrene-d10          | 17.116 | 188  | 707456   | 20.000 | ng/ul  | 0.00     |
| 79) Chrysene-d12              | 21.221 | 240  | 717854   | 20.000 | ng/ul  | 0.00     |
| 88) Perylene-d12              | 23.533 | 264  | 750050   | 20.000 | ng/ul  | 0.00     |
| System Monitoring Compounds   |        |      |          |        |        |          |
| 3) 1,4-Dioxane-d8             | 3.216  | 96   | 30126    | 7.434  | ng/uL  | 0.00     |
| 4) Pyridine-d5                | 3.622  | 84   | 421788   | 36.206 | ng/ul  | 0.00     |
| 7) Phenol-d5                  | 6.893  | 99   | 505147   | 36.659 | ng/ul  | 0.00     |
| 9) Bis-(2-Chloroethyl)eth...  | 7.057  | 67   | 312296   | 36.878 | ng/ul  | 0.00     |
| 11) 2-Chlorophenol-d4         | 7.246  | 132  | 391067   | 38.106 | ng/ul  | 0.00     |
| 15) 4-Methylphenol-d8         | 8.428  | 113  | 403522   | 36.656 | ng/ul  | 0.00     |
| 21) Nitrobenzene-d5           | 8.869  | 128  | 197451   | 39.446 | ng/ul  | 0.00     |
| 24) 2-Nitrophenol-d4          | 9.587  | 143  | 211679   | 40.414 | ng/ul  | 0.00     |
| 28) 2,4-Dichlorophenol-d3     | 10.122 | 165  | 346646   | 39.171 | ng/ul  | 0.00     |
| 31) 4-Chloroaniline-d4        | 10.645 | 131  | 532967   | 37.054 | ng/ul  | 0.00     |
| 46) Dimethylphthalate-d6      | 13.780 | 166  | 1110958  | 41.320 | ng/ul  | 0.00     |
| 49) Acenaphthylene-d8         | 14.045 | 160  | 1348790  | 40.898 | ng/ul  | 0.00     |
| 54) 4-Nitrophenol-d4          | 14.575 | 143  | 236935   | 42.189 | ng/ul  | 0.00     |
| 60) Fluorene-d10              | 15.351 | 176  | 899321   | 41.127 | ng/ul  | 0.00     |
| 65) 4,6-Dinitro-2-methylph... | 15.486 | 200  | 185233   | 39.612 | ng/ul  | 0.00     |
| 73) Anthracene-d10            | 17.216 | 188  | 1372093  | 40.991 | ng/ul  | 0.00     |
| 81) Pyrene-d10                | 19.463 | 212  | 1582763  | 41.986 | ng/ul  | 0.00     |
| 92) Benzo(a)pyrene-d12        | 23.386 | 264  | 1630908  | 40.322 | ng/ul  | 0.00     |
| Target Compounds              |        |      |          |        |        |          |
|                               |        |      |          |        | Qvalue |          |
| 2) 1,4-Dioxane                | 3.246  | 88   | 64070    | 15.560 | ng/uL  | 98       |
| 5) Pyridine                   | 3.640  | 79   | 434983   | 36.068 | ng/ul  | 99       |
| 6) Benzaldehyde               | 6.863  | 77   | 342764   | 50.591 | ng/ul  | 96       |
| 8) Phenol                     | 6.916  | 94   | 534431   | 37.976 | ng/ul  | 97       |
| 10) Bis(2-Chloroethyl)ether   | 7.146  | 93   | 407455   | 37.564 | ng/ul  | 97       |
| 12) 2-Chlorophenol            | 7.275  | 128  | 403296   | 38.972 | ng/ul  | 97       |
| 13) 2-Methylphenol            | 8.157  | 108  | 382778   | 36.969 | ng/ul  | 100      |
| 14) 2,2'-oxybis(1-Chloropr... | 8.251  | 45   | 467797   | 36.910 | ng/ul  | 99       |
| 16) Acetophenone              | 8.540  | 105  | 644656   | 37.200 | ng/ul  | 97       |
| 17) N-Nitroso-di-n-propyla... | 8.528  | 70   | 359814   | 38.600 | ng/ul  | 99       |
| 18) 4-Methylphenol            | 8.493  | 108  | 417805   | 37.116 | ng/ul  | 98       |
| 19) Hexachloroethane          | 8.775  | 117  | 179871   | 39.942 | ng/ul  | 95       |
| 22) Nitrobenzene              | 8.910  | 77   | 529421   | 40.398 | ng/ul  | 99       |
| 23) Isophorone                | 9.445  | 82   | 976286   | 39.950 | ng/ul  | 100      |
| 25) 2-Nitrophenol             | 9.616  | 139  | 224518   | 40.641 | ng/ul  | 95       |
| 26) 2,4-Dimethylphenol        | 9.687  | 107  | 407582   | 34.343 | ng/ul  | 99       |
| 27) Bis(2-Chloroethoxy)met... | 9.928  | 93   | 560368   | 39.557 | ng/ul  | 100      |
| 29) 2,4-Dichlorophenol        | 10.151 | 162  | 350686   | 40.847 | ng/ul  | 99       |
| 30) Naphthalene               | 10.545 | 128  | 1280657  | 40.016 | ng/ul  | 99       |
| 32) 4-Chloroaniline           | 10.663 | 127  | 529642   | 37.526 | ng/ul  | 100      |
| 33) Hexachlorobutadiene       | 10.828 | 225  | 201669   | 39.840 | ng/ul  | 98       |
| 34) Caprolactam               | 11.463 | 113  | 148194m  | 43.115 | ng/ul  |          |
| 35) 4-Chloro-3-methylphenol   | 11.804 | 107  | 460981   | 41.974 | ng/ul  | 99       |
| 36) 2-Methylnaphthalene       | 12.163 | 142  | 887964   | 40.377 | ng/ul  | 98       |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP081221\  
 Data File : BP006660.D  
 Acq On : 12 Aug 2021 11:02  
 Operator : CG/JU  
 Sample : PB138110BS  
 Misc :  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SLCS110

Manual Integrations  
 APPROVED


mohammad  
 8/13/2021 8:47:41 AM

Quant Time: Aug 12 11:07:56 2021  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_P\METHODS\SFAM-EPA-BP081021.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Thu Aug 12 11:07:37 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| 37) 1-Methylnaphthalene       | 12.386 | 142  | 865390   | 39.541 | ng/ul | 99       |
| 39) 1,2,4,5-Tetrachloroben... | 12.533 | 216  | 382236   | 40.456 | ng/ul | 98       |
| 40) Hexachlorocyclopentadiene | 12.510 | 237  | 171519   | 27.671 | ng/ul | 96       |
| 41) 2,4,6-Trichlorophenol     | 12.781 | 196  | 271282   | 41.809 | ng/ul | 100      |
| 42) 2,4,5-Trichlorophenol     | 12.851 | 196  | 294080   | 42.570 | ng/ul | 98       |
| 43) 1,1'-Biphenyl             | 13.186 | 154  | 1096986  | 40.249 | ng/ul | 100      |
| 44) 2-Chloronaphthalene       | 13.228 | 162  | 831124   | 40.194 | ng/ul | 99       |
| 45) 2-Nitroaniline            | 13.439 | 65   | 324895   | 43.572 | ng/ul | 98       |
| 47) Dimethylphthalate         | 13.828 | 163  | 1111897  | 41.811 | ng/ul | 99       |
| 48) 2,6-Dinitrotoluene        | 13.945 | 165  | 225414   | 43.916 | ng/ul | 100      |
| 50) Acenaphthylene            | 14.075 | 152  | 1302853  | 40.412 | ng/ul | 99       |
| 51) 3-Nitroaniline            | 14.275 | 138  | 269165   | 45.791 | ng/ul | 95       |
| 52) Acenaphthene              | 14.416 | 153  | 940852   | 40.929 | ng/ul | 98       |
| 53) 2,4-Dinitrophenol         | 14.480 | 184  | 133556   | 40.229 | ng/ul | 96       |
| 55) 4-Nitrophenol             | 14.592 | 109  | 225560   | 44.788 | ng/ul | 97       |
| 56) Dibenzofuran              | 14.757 | 168  | 1261605  | 41.046 | ng/ul | 98       |
| 57) 2,4-Dinitrotoluene        | 14.733 | 165  | 339865   | 45.008 | ng/ul | 98       |
| 58) 2,3,4,6-Tetrachlorophenol | 14.986 | 232  | 246093   | 42.913 | ng/ul | 99       |
| 59) Diethylphthalate          | 15.198 | 149  | 1221131  | 43.200 | ng/ul | 99       |
| 61) Fluorene                  | 15.410 | 166  | 1064283  | 41.527 | ng/ul | 100      |
| 62) 4-Chlorophenyl-phenyle... | 15.410 | 204  | 484276   | 41.741 | ng/ul | 100      |
| 63) 4-Nitroaniline            | 15.439 | 138  | 281970m  | 51.227 | ng/ul |          |
| 66) 4,6-Dinitro-2-methylph... | 15.498 | 198  | 189648   | 40.570 | ng/ul | 93       |
| 67) N-Nitrosodiphenylamine    | 15.627 | 169  | 904126   | 41.606 | ng/ul | 99       |
| 68) 4-Bromophenyl-phenylether | 16.304 | 248  | 287261   | 41.171 | ng/ul | 100      |
| 69) Hexachlorobenzene         | 16.410 | 284  | 324575   | 41.162 | ng/ul | 98       |
| 70) Atrazine                  | 16.592 | 200  | 319455   | 41.440 | ng/ul | 99       |
| 71) Pentachlorophenol         | 16.763 | 266  | 215560   | 41.736 | ng/ul | 97       |
| 72) Phenanthrene              | 17.157 | 178  | 1675783  | 41.930 | ng/ul | 99       |
| 74) Anthracene                | 17.251 | 178  | 1664211  | 41.006 | ng/ul | 100      |
| 75) 1,2,3,4-Tetrachloroben... | 13.145 | 216  | 387169   | 40.246 | ng/ul | 99       |
| 76) Pentachlorobenzene        | 14.675 | 250  | 387445   | 40.877 | ng/ul | 99       |
| 77) Carbazole                 | 17.527 | 167  | 1597069  | 42.132 | ng/ul | 99       |
| 78) Di-n-butylphthalate       | 18.092 | 149  | 2150031  | 43.868 | ng/ul | 100      |
| 80) Fluoranthene              | 19.133 | 202  | 1957669  | 41.318 | ng/ul | 100      |
| 82) Pyrene                    | 19.486 | 202  | 2009544  | 41.063 | ng/ul | 98       |
| 83) Butylbenzylphthalate      | 20.380 | 149  | 1031300  | 42.626 | ng/ul | 99       |
| 84) 3,3'-Dichlorobenzidine    | 21.145 | 252  | 663778   | 40.521 | ng/ul | 99       |
| 85) Benzo(a)anthracene        | 21.204 | 228  | 1948882  | 41.681 | ng/ul | 100      |
| 86) Bis(2-ethylhexyl)phtha... | 21.145 | 149  | 1592604  | 42.718 | ng/ul | 100      |
| 87) Chrysene                  | 21.257 | 228  | 1903976  | 42.234 | ng/ul | 99       |
| 89) Di-n-octyl phthalate      | 22.045 | 149  | 2774947  | 42.800 | ng/ul | 100      |
| 90) Benzo(b)fluoranthene      | 22.833 | 252  | 2095373  | 43.005 | ng/ul | 99       |
| 91) Benzo(k)fluoranthene      | 22.880 | 252  | 2009275  | 42.594 | ng/ul | 100      |
| 93) Benzo(a)pyrene            | 23.433 | 252  | 1857808  | 42.090 | ng/ul | 100      |
| 94) Indeno(1,2,3-cd)pyrene    | 25.933 | 276  | 2448797  | 42.431 | ng/ul | 99       |
| 95) Dibenzo(a,h)anthracene    | 25.950 | 278  | 2055392  | 42.399 | ng/ul | 99       |
| 96) Benzo(g,h,i)perylene      | 26.668 | 276  | 2038476  | 42.541 | ng/ul | 98       |


(#) = qualifier out of range (m) = manual integration (+) = signals summed





#2  
1,4-Dioxane  
Concen: 15.560 ng/uL  
RT: 3.246 min Scan# 27  
Delta R.T. -0.012 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




Tgt Ion: 88 Resp: 64070  
Ion Ratio Lower Upper  
88 100  
43 25.5 20.8 31.2  
58 64.0 53.0 79.4

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#5  
Pyridine  
Concen: 36.068 ng/uL  
RT: 3.640 min Scan# 94  
Delta R.T. -0.012 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Tgt Ion: 79 Resp: 434983  
Ion Ratio Lower Upper  
79 100  
52 60.3 48.9 73.3  
51 28.7 22.7 34.1





#6  
Benzaldehyde  
Concen: 50.591 ng/ul  
RT: 6.863 min Scan# 642  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




Tgt Ion: 77 Resp: 342764  
Ion Ratio Lower Upper  
77 100  
105 91.7 76.5 114.7  
106 85.1 70.7 106.1

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#8  
Phenol  
Concen: 37.976 ng/ul  
RT: 6.916 min Scan# 651  
Delta R.T. -0.006 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Tgt Ion: 94 Resp: 534431  
Ion Ratio Lower Upper  
94 100  
65 30.7 23.1 34.7  
66 40.4 31.1 46.7





#10  
Bis(2-Chloroethyl)ether  
Concen: 37.564 ng/ul  
RT: 7.146 min Scan# 690  
Delta R.T. -0.006 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




Tgt Ion: 93 Resp: 407455  
Ion Ratio Lower Upper  
93 100  
63 77.1 59.9 89.9  
95 33.2 24.8 37.2

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#12  
2-Chlorophenol  
Concen: 38.972 ng/ul  
RT: 7.275 min Scan# 712  
Delta R.T. -0.006 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02



Tgt Ion: 128 Resp: 403296  
Ion Ratio Lower Upper  
128 100  
64 49.6 38.2 57.2  
130 33.4 25.6 38.4





#13

2-Methylphenol

Concen: 36.969 ng/ul

RT: 8.157 min Scan# 862

Delta R.T. -0.006 min

Lab File: BP006660.D





Acq: 12 Aug 2021 11:02

Tgt Ion:108 Resp: 382778

Ion Ratio Lower Upper

108 100

107 91.1 72.6 108.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 36.910 ng/ul

RT: 8.251 min Scan# 878

Delta R.T. -0.000 min

Lab File: BP006660.D

Acq: 12 Aug 2021 11:02




Tgt Ion: 45 Resp: 467797

Ion Ratio Lower Upper

45 100

77 18.5 14.4 21.6

79 14.4 11.4 17.2



Instrument :

BNA\_P

ClientSampled :


SLCS110

Manual Integrations

APPROVED


mohammad

8/13/2021 8:47:41 AM



#16  
Acetophenone  
Concen: 37.200 ng/ul  
RT: 8.540 min Scan# 927  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampleId :  
SLCS110




Tgt Ion: 105 Resp: 644656  
Ion Ratio Lower Upper  
105 100  
77 81.4 67.5 101.3  
51 23.3 19.5 29.3

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#17  
N-Nitroso-di-n-propylamine  
Concen: 38.600 ng/ul  
RT: 8.528 min Scan# 925  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Tgt Ion: 70 Resp: 359814  
Ion Ratio Lower Upper  
70 100  
42 50.9 40.5 60.7  
101 9.1 7.7 11.5  
130 18.2 15.6 23.4





#18  
4-Methylphenol  
Concen: 37.116 ng/ul  
RT: 8.493 min Scan# 919  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampleId :  
SLCS110




Tgt Ion:108 Resp: 417805  
Ion Ratio Lower Upper  
108 100  
107 114.7 90.1 135.1

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#19  
Hexachloroethane  
Concen: 39.942 ng/ul  
RT: 8.775 min Scan# 967  
Delta R.T. -0.006 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Tgt Ion:117 Resp: 179871  
Ion Ratio Lower Upper  
117 100  
201 81.7 68.3 102.5  
199 50.1 43.2 64.8





#22  
Nitrobenzene  
Concen: 40.398 ng/ul  
RT: 8.910 min Scan# 990  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampleId :  
SLCS110




Tgt Ion: 77 Resp: 529421  
Ion Ratio Lower Upper  
77 100  
123 40.3 32.6 48.8  
65 13.8 10.6 15.8

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#23  
Isophorone  
Concen: 39.950 ng/ul  
RT: 9.445 min Scan# 1081  
Delta R.T. 0.005 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02



Tgt Ion: 82 Resp: 976286  
Ion Ratio Lower Upper  
82 100  
95 7.4 6.0 9.0  
138 15.6 12.3 18.5









#25  
2-Nitrophenol  
Concen: 40.641 ng/ul  
RT: 9.616 min Scan# 1110  
Delta R.T. -0.006 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02

Tgt Ion:139 Resp: 224518  
Ion Ratio Lower Upper  
139 100  
65 61.8 45.8 68.6  
109 35.7 26.8 40.2

Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




Manual Integrations  
APPROVED


mohammad  
8/13/2021 8:47:41 AM



#26  
2,4-Dimethylphenol  
Concen: 34.343 ng/ul  
RT: 9.687 min Scan# 1122  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02

Tgt Ion:107 Resp: 407582  
Ion Ratio Lower Upper  
107 100  
121 48.0 39.0 58.4  
122 81.5 64.6 97.0






#27

Bis(2-Chloroethoxy)methane  
Concen: 39.557 ng/ul  
RT: 9.928 min Scan# 1163  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110



| Tgt Ion:  | 93   | Resp: | 560368 |
|-----------|------|-------|--------|
| Ion Ratio | 100  | Lower | Upper  |
| 93        | 100  |       |        |
| 95        | 31.9 | 25.4  | 38.2   |
| 123       | 10.3 | 8.2   | 12.4   |


Manual Integrations  
APPROVED

mohammad  
8/13/2021 8:47:41 AM






#29

2,4-Dichlorophenol  
Concen: 40.847 ng/ul  
RT: 10.151 min Scan# 1201  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




| Tgt Ion:  | 162  | Resp: | 350686 |
|-----------|------|-------|--------|
| Ion Ratio | 100  | Lower | Upper  |
| 162       | 100  |       |        |
| 164       | 66.0 | 51.4  | 77.2   |
| 98        | 42.1 | 33.5  | 50.3   |





#30  
Naphthalene  
Concen: 40.016 ng/ul  
RT: 10.545 min Scan# 1268  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110



Tgt Ion:128 Resp: 1280657  
Ion Ratio Lower Upper  
128 100  
129 10.8 8.8 13.2  
127 13.1 10.3 15.5




Manual Integrations  
APPROVED


mohammad  
8/13/2021 8:47:41 AM



#32  
4-Chloroaniline  
Concen: 37.526 ng/ul  
RT: 10.663 min Scan# 1288  
Delta R.T. -0.006 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02


Tgt Ion:127 Resp: 529642  
Ion Ratio Lower Upper  
127 100  
129 31.9 25.6 38.4





#33  
Hexachlorobutadiene  
Concen: 39.840 ng/ul  
RT: 10.828 min Scan# 1316  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




Tgt Ion:225 Resp: 201669  
Ion Ratio Lower Upper  
225 100  
223 64.0 49.8 74.6  
227 63.1 51.4 77.2

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#34  
Caprolactam  
Concen: 43.115 ng/ul m  
RT: 11.463 min Scan# 1424  
Delta R.T. 0.005 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Tgt Ion:113 Resp: 148194  
Ion Ratio Lower Upper  
113 100  
55 161.9 127.7 191.5  
56 123.0 96.2 144.4





#35  
4-Chloro-3-methylphenol  
Concen: 41.974 ng/ul  
RT: 11.804 min Scan# 1482  
Delta R.T. 0.005 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110



Tgt Ion:107 Resp: 460981  
Ion Ratio Lower Upper  
107 100  
144 23.9 18.4 27.6  
142 72.9 57.6 86.4




Manual Integrations  
APPROVED


mohammad  
8/13/2021 8:47:41 AM



#36  
2-Methylnaphthalene  
Concen: 40.377 ng/ul  
RT: 12.163 min Scan# 1543  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02


Tgt Ion:142 Resp: 887964  
Ion Ratio Lower Upper  
142 100  
141 89.0 72.7 109.1





#37  
1-Methylnaphthalene  
Concen: 39.541 ng/ul  
RT: 12.386 min Scan# 1581  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




Tgt Ion:142 Resp: 865390  
Ion Ratio Lower Upper  
142 100  
141 93.7 75.7 113.5  
116 4.4 3.5 5.3

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#39  
1,2,4,5-Tetrachlorobenzene  
Concen: 40.456 ng/ul  
RT: 12.533 min Scan# 1606  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Tgt Ion:216 Resp: 382236  
Ion Ratio Lower Upper  
216 100  
214 79.2 64.6 97.0  
179 24.0 19.7 29.5  
108 23.9 18.7 28.1





#40  
Hexachlorocyclopentadiene  
Concen: 27.671 ng/ul  
RT: 12.510 min Scan# 1602  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 237     | 100   |       |       |
| 235     | 66.5  | 50.6  | 76.0  |
| 272     | 12.6  | 9.8   | 14.6  |

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#41  
2,4,6-Trichlorophenol  
Concen: 41.809 ng/ul  
RT: 12.781 min Scan# 1648  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 196     | 100   |       |       |
| 198     | 94.9  | 75.7  | 113.5 |
| 200     | 30.6  | 25.0  | 37.4  |





#42  
2,4,5-Trichlorophenol  
Concen: 42.570 ng/ul  
RT: 12.851 min Scan# 1660  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




Tgt Ion:196 Resp: 294080  
Ion Ratio Lower Upper  
196 100  
198 94.6 77.8 116.6  
200 29.8 24.7 37.1

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#43  
1,1'-Biphenyl  
Concen: 40.249 ng/ul  
RT: 13.186 min Scan# 1717  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Tgt Ion:154 Resp: 1096986  
Ion Ratio Lower Upper  
154 100  
153 41.1 33.2 49.8  
76 15.6 12.5 18.7





#44  
2-Chloronaphthalene  
Concen: 40.194 ng/ul  
RT: 13.228 min Scan# 1724  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampleId :  
SLCS110




Tgt Ion:162 Resp: 831124  
Ion Ratio Lower Upper  
162 100  
164 32.8 26.2 39.2  
127 45.1 35.4 53.2

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#45  
2-Nitroaniline  
Concen: 43.572 ng/ul  
RT: 13.439 min Scan# 1760  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Tgt Ion: 65 Resp: 324895  
Ion Ratio Lower Upper  
65 100  
92 68.5 55.8 83.6  
138 91.9 75.6 113.4





#47  
Dimethylphthalate  
Concen: 41.811 ng/ul  
RT: 13.828 min Scan# 1826  
Delta R.T. 0.005 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110



Tgt Ion:163 Resp: 1111897  
Ion Ratio Lower Upper  
163 100  
194 3.7 2.8 4.2  
164 9.8 8.2 12.2




Manual Integrations  
APPROVED


mohammad  
8/13/2021 8:47:41 AM



#48  
2,6-Dinitrotoluene  
Concen: 43.916 ng/ul  
RT: 13.945 min Scan# 1846  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02


Tgt Ion:165 Resp: 225414  
Ion Ratio Lower Upper  
165 100  
89 73.3 58.3 87.5  
121 25.0 20.1 30.1





#50  
Acenaphthylene  
Concen: 40.412 ng/ul  
RT: 14.075 min Scan# 1868  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




Tgt Ion:152 Resp: 1302853  
Ion Ratio Lower Upper  
152 100  
151 20.4 16.0 24.0  
153 13.3 10.3 15.5

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#51  
3-Nitroaniline  
Concen: 45.791 ng/ul  
RT: 14.275 min Scan# 1902  
Delta R.T. 0.005 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Tgt Ion:138 Resp: 269165  
Ion Ratio Lower Upper  
138 100  
108 12.0 9.7 14.5  
92 126.5 106.6 160.0





#52  
Acenaphthene  
Concen: 40.929 ng/ul  
RT: 14.416 min Scan# 1926  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampleId :  
SLCS110




Tgt Ion:153 Resp: 940852  
Ion Ratio Lower Upper  
153 100  
152 47.3 38.2 57.2  
154 86.4 70.7 106.1

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#53  
2,4-Dinitrophenol  
Concen: 40.229 ng/ul  
RT: 14.480 min Scan# 1937  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02



Tgt Ion:184 Resp: 133556  
Ion Ratio Lower Upper  
184 100  
63 87.9 66.6 100.0  
154 74.5 58.2 87.2









#55  
4-Nitrophenol  
Concen: 44.788 ng/ul  
RT: 14.592 min Scan# 1956  
Delta R.T. 0.005 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02

Tgt Ion:109 Resp: 225560  
Ion Ratio Lower Upper  
109 100  
139 110.7 92.3 138.5  
65 128.7 105.4 158.2

Instrument :  
BNA\_P  
ClientSampleId :  
SLCS110




Manual Integrations  
APPROVED


mohammad  
8/13/2021 8:47:41 AM



#56  
Dibenzofuran  
Concen: 41.046 ng/ul  
RT: 14.757 min Scan# 1984  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02


Tgt Ion:168 Resp: 1261605  
Ion Ratio Lower Upper  
168 100  
139 43.7 33.9 50.9





#57  
2,4-Dinitrotoluene  
Concen: 45.008 ng/ul  
RT: 14.733 min Scan# 1980  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




Tgt Ion:165 Resp: 339865  
Ion Ratio Lower Upper  
165 100  
63 54.7 42.7 64.1  
182 2.5 2.2 3.2

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#58  
2,3,4,6-Tetrachlorophenol  
Concen: 42.913 ng/ul  
RT: 14.986 min Scan# 2023  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Tgt Ion:232 Resp: 246093  
Ion Ratio Lower Upper  
232 100  
131 60.5 48.2 72.4  
130 3.4 2.7 4.1  
166 36.8 28.3 42.5





#59  
Diethylphthalate  
Concen: 43.200 ng/ul  
RT: 15.198 min Scan# 2059  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




Tgt Ion:149 Resp: 1221131  
Ion Ratio Lower Upper  
149 100  
177 19.4 16.3 24.5  
150 12.4 10.0 15.0

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#61  
Fluorene  
Concen: 41.527 ng/ul  
RT: 15.410 min Scan# 2095  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Tgt Ion:166 Resp: 1064283  
Ion Ratio Lower Upper  
166 100  
165 99.3 79.6 119.4  
167 13.5 10.6 15.8





#62  
4-Chlorophenyl-phenylether  
Concen: 41.741 ng/ul  
RT: 15.410 min Scan# 2095  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




Tgt Ion:204 Resp: 484276  
Ion Ratio Lower Upper  
204 100  
206 32.7 26.1 39.1  
141 71.6 57.3 85.9

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#63  
4-Nitroaniline  
Concen: 51.227 ng/ul m  
RT: 15.439 min Scan# 2100  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Tgt Ion:138 Resp: 281970  
Ion Ratio Lower Upper  
138 100  
92 61.1 48.0 72.0  
108 96.0 74.4 111.6





#66  
4,6-Dinitro-2-methylphenol  
Concen: 40.570 ng/ul  
RT: 15.498 min Scan# 2110  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




Tgt Ion:198 Resp: 189648  
Ion Ratio Lower Upper  
198 100  
182 5.4 4.3 6.5  
77 38.5 27.2 40.8

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#67  
N-Nitrosodiphenylamine  
Concen: 41.606 ng/ul  
RT: 15.627 min Scan# 2132  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Tgt Ion:169 Resp: 904126  
Ion Ratio Lower Upper  
169 100  
168 68.5 54.2 81.4  
167 37.4 30.0 45.0





#68  
4-Bromophenyl-phenylether  
Concen: 41.171 ng/ul  
RT: 16.304 min Scan# 2247  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110



Tgt Ion:248 Resp: 287261  
Ion Ratio Lower Upper  
248 100  
250 98.1 78.6 118.0  
141 102.0 82.0 123.0




Manual Integrations  
APPROVED


mohammad  
8/13/2021 8:47:41 AM



#69  
Hexachlorobenzene  
Concen: 41.162 ng/ul  
RT: 16.410 min Scan# 2265  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02


Tgt Ion:284 Resp: 324575  
Ion Ratio Lower Upper  
284 100  
142 48.3 37.0 55.6  
249 34.0 26.7 40.1





#70  
Atrazine  
Concen: 41.440 ng/ul  
RT: 16.592 min Scan# 2296  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




Tgt Ion:200 Resp: 319455  
Ion Ratio Lower Upper  
200 100  
173 27.5 21.7 32.5  
215 44.3 36.3 54.5

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#71  
Pentachlorophenol  
Concen: 41.736 ng/ul  
RT: 16.763 min Scan# 2325  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02



Tgt Ion:266 Resp: 215560  
Ion Ratio Lower Upper  
266 100  
264 64.3 49.4 74.2  
268 65.2 50.5 75.7









#72  
Phenanthrene  
Concen: 41.930 ng/u1  
RT: 17.157 min Scan# 2392  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02

Tgt Ion:178 Resp: 1675783  
Ion Ratio Lower Upper  
178 100  
179 15.3 12.1 18.1  
176 19.5 15.2 22.8

Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




Manual Integrations  
APPROVED


mohammad  
8/13/2021 8:47:41 AM



#74  
Anthracene  
Concen: 41.006 ng/u1  
RT: 17.251 min Scan# 2408  
Delta R.T. 0.005 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02

Tgt Ion:178 Resp: 1664211  
Ion Ratio Lower Upper  
178 100  
179 15.1 12.2 18.4  
176 18.9 15.0 22.6









#75  
1,2,3,4-Tetrachlorobenzene  
Concen: 40.246 ng/uL  
RT: 13.145 min Scan# 1710  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02

Instrument :  
BNA\_P  
ClientSampled :  
SLCS110

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 216     | 100   |       |       |
| 214     | 77.7  | 61.8  | 92.6  |
| 218     | 47.3  | 38.6  | 58.0  |




Manual Integrations  
APPROVED


mohammad  
8/13/2021 8:47:41 AM



#76  
Pentachlorobenzene  
Concen: 40.877 ng/uL  
RT: 14.675 min Scan# 1970  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 250     | 100   |       |       |
| 248     | 61.5  | 50.1  | 75.1  |
| 252     | 63.2  | 50.0  | 75.0  |





#77  
Carbazole  
Concen: 42.132 ng/ul  
RT: 17.527 min Scan# 2455  
Delta R.T. 0.006 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110



Tgt Ion:167 Resp: 1597069  
Ion Ratio Lower Upper  
167 100  
166 21.7 17.0 25.4  
139 14.5 11.5 17.3




Manual Integrations  
APPROVED


mohammad  
8/13/2021 8:47:41 AM



#78  
Di-n-butylphthalate  
Concen: 43.868 ng/ul  
RT: 18.092 min Scan# 2551  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02


Tgt Ion:149 Resp: 2150031  
Ion Ratio Lower Upper  
149 100  
150 9.1 7.3 10.9  
104 6.2 4.8 7.2





#80  
Fluoranthene  
Concen: 41.318 ng/ul  
RT: 19.133 min Scan# 2728  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampleId :  
SLCS110




Tgt Ion:202 Resp: 1957669  
Ion Ratio Lower Upper  
202 100  
101 13.6 10.8 16.2  
100 10.5 8.2 12.4

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#82  
Pyrene  
Concen: 41.063 ng/ul  
RT: 19.486 min Scan# 2788  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Tgt Ion:202 Resp: 2009544  
Ion Ratio Lower Upper  
202 100  
101 16.6 12.6 18.8  
100 13.7 10.5 15.7





#83  
Butylbenzylphthalate  
Concen: 42.626 ng/ul  
RT: 20.380 min Scan# 2940  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




Tgt Ion:149 Resp: 1031300  
Ion Ratio Lower Upper  
149 100  
91 79.0 62.1 93.1  
206 17.9 14.4 21.6

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#84  
3,3'-Dichlorobenzidine  
Concen: 40.521 ng/ul  
RT: 21.145 min Scan# 3070  
Delta R.T. 0.005 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Tgt Ion:252 Resp: 663778  
Ion Ratio Lower Upper  
252 100  
254 63.8 50.8 76.2  
126 13.4 12.3 18.5





#85  
Benzo(a)anthracene  
Concen: 41.681 ng/ul  
RT: 21.204 min Scan# 3080  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




Tgt Ion:228 Resp: 1948882  
Ion Ratio Lower Upper  
228 100  
229 19.7 15.7 23.5  
226 27.2 21.6 32.4

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#86  
Bis(2-ethylhexyl)phthalate  
Concen: 42.718 ng/ul  
RT: 21.145 min Scan# 3070  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02



Tgt Ion:149 Resp: 1592604  
Ion Ratio Lower Upper  
149 100  
167 25.5 20.2 30.4  
279 3.5 2.8 4.2









#87  
Chrysene  
Concen: 42.234 ng/ul  
RT: 21.257 min Scan# 3089  
Delta R.T. 0.006 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02

Tgt Ion:228 Resp: 1903976  
Ion Ratio Lower Upper  
228 100  
226 30.4 23.9 35.9  
229 19.7 15.9 23.9

Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




Manual Integrations  
APPROVED


mohammad  
8/13/2021 8:47:41 AM



#89  
Di-n-octyl phthalate  
Concen: 42.800 ng/ul  
RT: 22.045 min Scan# 3223  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02


Tgt Ion:149 Resp: 2774947





#90  
Benzo(b)fluoranthene  
Concen: 43.005 ng/ul  
RT: 22.833 min Scan# 3357  
Delta R.T. 0.006 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




Tgt Ion:252 Resp: 2095373  
Ion Ratio Lower Upper  
252 100  
253 21.9 17.4 26.0  
125 12.4 10.2 15.2

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#91  
Benzo(k)fluoranthene  
Concen: 42.594 ng/ul  
RT: 22.880 min Scan# 3365  
Delta R.T. 0.005 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Tgt Ion:252 Resp: 2009275  
Ion Ratio Lower Upper  
252 100  
253 21.9 17.4 26.0  
125 12.4 9.8 14.8





#93  
Benzo(a)pyrene  
Concen: 42.090 ng/ul  
RT: 23.433 min Scan# 3459  
Delta R.T. -0.000 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




Tgt Ion:252 Resp: 1857808  
Ion Ratio Lower Upper  
252 100  
253 21.9 17.4 26.2  
125 13.8 11.2 16.8

Manual Integrations  
APPROVED



mohammad  
8/13/2021 8:47:41 AM




#94  
Indeno(1,2,3-cd)pyrene  
Concen: 42.431 ng/ul  
RT: 25.933 min Scan# 3884  
Delta R.T. 0.005 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Tgt Ion:276 Resp: 2448797  
Ion Ratio Lower Upper  
276 100  
138 27.8 22.3 33.5  
277 25.4 19.7 29.5





#95  
Dibenzo(a,h)anthracene  
Concen: 42.399 ng/ul  
RT: 25.950 min Scan# 3887  
Delta R.T. 0.005 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02




Instrument :  
BNA\_P  
ClientSampled :  
SLCS110




Tgt Ion:278 Resp: 2055392  
Ion Ratio Lower Upper  
278 100  
139 19.4 15.0 22.6  
279 23.5 19.1 28.7

Manual Integrations  
APPROVED

mohammad  
8/13/2021 8:47:41 AM



#96  
Benzo(g,h,i)perylene  
Concen: 42.541 ng/ul  
RT: 26.668 min Scan# 4009  
Delta R.T. 0.005 min  
Lab File: BP006660.D  
Acq: 12 Aug 2021 11:02



Tgt Ion:276 Resp: 2038476  
Ion Ratio Lower Upper  
276 100  
138 25.8 19.9 29.9  
277 23.8 18.5 27.7

