

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081424\
 Data File : BP021504.D
 Acq On : 14 Aug 2024 18:04
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 14 18:31:51 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	105	0.00
2	1,4-Dioxane	40.000	39.492	1.3	103	0.00
3	Pyridine	40.000	39.386	1.5	102	0.00
4	n-Nitrosodimethylamine	40.000	40.011	-0.0	105	0.00
5 S	2-Fluorophenol	80.000	81.446	-1.8	107	0.00
6	Aniline	40.000	38.565	3.6	100	0.00
7 S	Phenol-d6	80.000	79.825	0.2	104	0.00
8	2-Chlorophenol	40.000	40.483	-1.2	106	0.00
9	Benzaldehyde	40.000	37.993	5.0	104	0.00
10 C	Phenol	40.000	39.271	1.8	102	0.00
11	bis(2-Chloroethyl)ether	40.000	38.642	3.4	102	0.00
12	1,3-Dichlorobenzene	40.000	39.356	1.6	105	0.00
13 C	1,4-Dichlorobenzene	40.000	39.490	1.3	105	0.00
14	1,2-Dichlorobenzene	40.000	39.108	2.2	104	0.00
15	Benzyl Alcohol	40.000	39.275	1.8	102	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.809	3.0	104	0.00
17	2-Methylphenol	40.000	39.461	1.3	103	0.00
18	Hexachloroethane	40.000	40.320	-0.8	106	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.101	7.2	98	0.00
20	3+4-Methylphenols	40.000	39.529	1.2	102	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	102	0.00
22	Acetophenone	40.000	39.428	1.4	101	0.00
23 S	Nitrobenzene-d5	80.000	90.013	-12.5	109	0.00
24	Nitrobenzene	40.000	43.882	-9.7	107	0.00
25	Isophorone	40.000	38.228	4.4	98	0.00
26 C	2-Nitrophenol	40.000	44.559	-11.4	122	0.00
27	2,4-Dimethylphenol	40.000	40.106	-0.3	103	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.031	2.4	101	0.00
29 C	2,4-Dichlorophenol	40.000	42.507	-6.3	106	0.00
30	1,2,4-Trichlorobenzene	40.000	39.851	0.4	103	0.00
31	Naphthalene	40.000	39.409	1.5	101	0.00
32	Benzoic acid	40.000	45.032	-12.6	128	0.00
33	4-Chloroaniline	40.000	39.235	1.9	100	0.00
34 C	Hexachlorobutadiene	40.000	40.217	-0.5	103	0.00
35	Caprolactam	40.000	40.643	-1.6	101	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.873	-2.2	101	0.00
37	2-Methylnaphthalene	40.000	38.553	3.6	100	0.00
38	1-Methylnaphthalene	40.000	38.877	2.8	100	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	101	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.724	-1.8	101	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.150	-2.9	101	0.00
42 S	2,4,6-Tribromophenol	80.000	76.290	4.6	91	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.203	-8.0	104	-0.01
44	2,4,5-Trichlorophenol	40.000	43.930	-9.8	105	0.00
45 S	2-Fluorobiphenyl	80.000	79.179	1.0	99	0.00
46	1,1'-Biphenyl	40.000	39.873	0.3	100	0.00
47	2-Chloronaphthalene	40.000	39.991	0.0	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081424\
 Data File : BP021504.D
 Acq On : 14 Aug 2024 18:04
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 14 18:31:51 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.959	-7.4	116	0.00
49	Acenaphthylene	40.000	39.759	0.6	99	0.00
50	Dimethylphthalate	40.000	40.444	-1.1	101	-0.01
51	2,6-Dinitrotoluene	40.000	41.792	-4.5	110	0.00
52 C	Acenaphthene	40.000	39.980	0.1	100	0.00
53	3-Nitroaniline	40.000	42.740	-6.9	111	0.00
54 P	2,4-Dinitrophenol	40.000	44.280	-10.7	121	0.00
55	Dibenzofuran	40.000	39.824	0.4	100	0.00
56 P	4-Nitrophenol	40.000	42.575	-6.4	111	0.00
57	2,4-Dinitrotoluene	40.000	42.556	-6.4	113	0.00
58	Fluorene	40.000	39.276	1.8	98	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.969	-7.4	101	0.00
60	Diethylphthalate	40.000	40.156	-0.4	100	-0.01
61	4-Chlorophenyl-phenylether	40.000	39.238	1.9	99	0.00
62	4-Nitroaniline	40.000	43.073	-7.7	108	-0.01
63	Azobenzene	40.000	39.282	1.8	98	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.988	-7.5	115	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.005	-0.0	98	-0.01
67	4-Bromophenyl-phenylether	40.000	36.332	9.2	92	0.00
68	Hexachlorobenzene	40.000	39.526	1.2	97	0.00
69	Atrazine	40.000	37.921	5.2	90	0.00
70 C	Pentachlorophenol	40.000	42.988	-7.5	106	0.00
71	Phenanthrene	40.000	39.701	0.7	98	0.00
72	Anthracene	40.000	39.443	1.4	97	0.00
73	Carbazole	40.000	40.125	-0.3	97	0.00
74	Di-n-butylphthalate	40.000	40.058	-0.1	96	0.00
75 C	Fluoranthene	40.000	40.206	-0.5	98	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	98	0.00
77	Benzydine	40.000	30.396	24.0	73	0.00
78	Pyrene	40.000	40.042	-0.1	100	0.00
79 S	Terphenyl-d14	80.000	78.725	1.6	99	0.00
80	Butylbenzylphthalate	40.000	38.605	3.5	98	0.00
81	Benzo(a)anthracene	40.000	40.011	-0.0	97	0.00
82	3,3'-Dichlorobenzidine	40.000	41.730	-4.3	96	-0.01
83	Chrysene	40.000	40.417	-1.0	98	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.166	-5.4	99	-0.01
85 c	Di-n-octyl phthalate	40.000	38.689	3.3	97	-0.01
86 I	Perylene-d12	20.000	20.000	0.0	98	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.391	-1.0	96	-0.02
88	Benzo(b)fluoranthene	40.000	40.475	-1.2	98	0.00
89	Benzo(k)fluoranthene	40.000	40.367	-0.9	97	0.00
90 C	Benzo(a)pyrene	40.000	40.685	-1.7	98	-0.02
91	Dibenzo(a,h)anthracene	40.000	40.493	-1.2	97	0.00
92	Benzo(g,h,i)perylene	40.000	40.739	-1.8	97	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081424\
Data File : BP021504.D
Acq On : 14 Aug 2024 18:04
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Aug 14 18:31:51 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 14 00:00:30 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0