

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081524\
 Data File : BP021506.D
 Acq On : 15 Aug 2024 09:10
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 15 10:37:47 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	93	0.00
2	1,4-Dioxane	40.000	39.164	2.1	90	0.00
3	Pyridine	40.000	39.867	0.3	91	0.00
4	n-Nitrosodimethylamine	40.000	41.290	-3.2	96	0.00
5 S	2-Fluorophenol	80.000	80.780	-1.0	93	0.00
6	Aniline	40.000	38.758	3.1	89	0.00
7 S	Phenol-d6	80.000	79.544	0.6	92	0.00
8	2-Chlorophenol	40.000	40.972	-2.4	95	0.00
9	Benzaldehyde	40.000	34.741	13.1	84	0.00
10 C	Phenol	40.000	39.442	1.4	91	0.00
11	bis(2-Chloroethyl)ether	40.000	39.133	2.2	92	0.00
12	1,3-Dichlorobenzene	40.000	39.531	1.2	93	0.00
13 C	1,4-Dichlorobenzene	40.000	39.675	0.8	93	0.00
14	1,2-Dichlorobenzene	40.000	39.246	1.9	93	0.00
15	Benzyl Alcohol	40.000	39.650	0.9	91	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.678	0.8	94	0.00
17	2-Methylphenol	40.000	39.809	0.5	92	0.00
18	Hexachloroethane	40.000	41.451	-3.6	97	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.055	4.9	89	0.00
20	3+4-Methylphenols	40.000	40.185	-0.5	92	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	94	0.00
22	Acetophenone	40.000	38.733	3.2	91	0.00
23 S	Nitrobenzene-d5	80.000	89.238	-11.5	99	0.00
24	Nitrobenzene	40.000	43.269	-8.2	96	0.00
25	Isophorone	40.000	38.409	4.0	90	0.00
26 C	2-Nitrophenol	40.000	43.591	-9.0	108	0.00
27	2,4-Dimethylphenol	40.000	39.818	0.5	93	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.443	1.4	94	0.00
29 C	2,4-Dichlorophenol	40.000	42.364	-5.9	97	-0.01
30	1,2,4-Trichlorobenzene	40.000	39.605	1.0	93	0.00
31	Naphthalene	40.000	39.287	1.8	92	0.00
32	Benzoic acid	40.000	45.077	-12.7	117	0.00
33	4-Chloroaniline	40.000	40.192	-0.5	93	0.00
34 C	Hexachlorobutadiene	40.000	39.837	0.4	93	0.00
35	Caprolactam	40.000	42.095	-5.2	96	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.750	-4.4	95	0.00
37	2-Methylnaphthalene	40.000	39.198	2.0	93	0.00
38	1-Methylnaphthalene	40.000	39.281	1.8	93	-0.01
39 I	Acenaphthene-d10	20.000	20.000	0.0	97	-0.02
40	1,2,4,5-Tetrachlorobenzene	40.000	39.373	1.6	94	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.987	2.5	92	-0.01
42 S	2,4,6-Tribromophenol	80.000	79.183	1.0	91	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.983	-5.0	97	-0.01
44	2,4,5-Trichlorophenol	40.000	43.637	-9.1	100	0.00
45 S	2-Fluorobiphenyl	80.000	78.169	2.3	94	0.00
46	1,1'-Biphenyl	40.000	39.341	1.6	95	-0.01
47	2-Chloronaphthalene	40.000	38.815	3.0	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081524\
 Data File : BP021506.D
 Acq On : 15 Aug 2024 09:10
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 15 10:37:47 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.755	-6.9	111	0.00
49	Acenaphthylene	40.000	39.493	1.3	94	0.00
50	Dimethylphthalate	40.000	40.372	-0.9	97	-0.01
51	2,6-Dinitrotoluene	40.000	42.337	-5.8	107	-0.02
52 C	Acenaphthene	40.000	39.732	0.7	95	-0.01
53	3-Nitroaniline	40.000	43.965	-9.9	110	-0.01
54 P	2,4-Dinitrophenol	40.000	45.092	-12.7	119	-0.01
55	Dibenzofuran	40.000	39.990	0.0	96	0.00
56 P	4-Nitrophenol	40.000	44.552	-11.4	113	-0.01
57	2,4-Dinitrotoluene	40.000	43.860	-9.6	113	-0.01
58	Fluorene	40.000	39.865	0.3	96	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	44.878	-12.2	102	-0.01
60	Diethylphthalate	40.000	41.373	-3.4	99	-0.01
61	4-Chlorophenyl-phenylether	40.000	39.582	1.0	96	-0.01
62	4-Nitroaniline	40.000	44.971	-12.4	109	-0.01
63	Azobenzene	40.000	40.456	-1.1	97	-0.02
64 I	Phenanthrene-d10	20.000	20.000	0.0	99	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.473	-8.7	117	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.790	3.0	96	-0.01
67	4-Bromophenyl-phenylether	40.000	35.671	10.8	90	-0.01
68	Hexachlorobenzene	40.000	38.710	3.2	95	-0.01
69	Atrazine	40.000	37.903	5.2	91	-0.01
70 C	Pentachlorophenol	40.000	42.247	-5.6	104	0.00
71	Phenanthrene	40.000	39.931	0.2	98	0.00
72	Anthracene	40.000	39.892	0.3	98	0.00
73	Carbazole	40.000	40.657	-1.6	98	-0.01
74	Di-n-butylphthalate	40.000	40.929	-2.3	98	0.00
75 C	Fluoranthene	40.000	41.211	-3.0	100	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	103	0.00
77	Benzydine	40.000	41.070	-2.7	103	0.00
78	Pyrene	40.000	38.768	3.1	101	0.00
79 S	Terphenyl-d14	80.000	77.369	3.3	102	0.00
80	Butylbenzylphthalate	40.000	38.293	4.3	102	0.00
81	Benzo(a)anthracene	40.000	40.207	-0.5	103	0.00
82	3,3'-Dichlorobenzidine	40.000	41.687	-4.2	101	-0.02
83	Chrysene	40.000	40.158	-0.4	102	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.219	-5.5	103	-0.01
85 c	Di-n-octyl phthalate	40.000	38.361	4.1	100	-0.01
86 I	Perylene-d12	20.000	20.000	0.0	102	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.108	-2.8	102	-0.02
88	Benzo(b)fluoranthene	40.000	40.653	-1.6	103	0.00
89	Benzo(k)fluoranthene	40.000	41.122	-2.8	103	-0.01
90 C	Benzo(a)pyrene	40.000	41.143	-2.9	103	-0.02
91	Dibenzo(a,h)anthracene	40.000	41.268	-3.2	103	-0.02
92	Benzo(g,h,i)perylene	40.000	41.441	-3.6	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081524\
Data File : BP021506.D
Acq On : 15 Aug 2024 09:10
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Aug 15 10:37:47 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 14 00:00:30 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0