

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081622\
 Data File : BP011447.D
 Acq On : 16 Aug 2022 16:52
 Operator : CG/JU
 Sample : SST02045
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST020647

Quant Time: Aug 17 01:07:47 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 17 01:01:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.558	152	78093	20.000	ng/u1	0.00
20) Naphthalene-d8	10.352	136	359473	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.234	164	220870	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.004	188	451935	20.000	ng/u1	0.00
79) Chrysene-d12	21.133	240	446566	20.000	ng/u1	0.00
88) Perylene-d12	23.427	264	440370	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.070	96	20426	8.018	ng/uL	0.00
4) Pyridine-d5	3.481	84	118483	18.439	ng/u1	0.00
7) Phenol-d5	6.746	99	131337	19.004	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.911	67	89048	18.172	ng/u1	0.00
11) 2-Chlorophenol-d4	7.093	132	103043	19.003	ng/u1	0.00
15) 4-Methylphenol-d8	8.281	113	106322	19.593	ng/u1	0.00
21) Nitrobenzene-d5	8.728	128	49967	20.424	ng/u1	0.00
24) 2-Nitrophenol-d4	9.440	143	50211	22.417	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.975	165	94924	20.985	ng/u1	0.00
31) 4-Chloroaniline-d4	10.499	131	150965	19.090	ng/u1	0.00
46) Dimethylphthalate-d6	13.657	166	300423	19.603	ng/u1	0.00
49) Acenaphthylene-d8	13.922	160	336047	18.216	ng/u1	0.00
54) 4-Nitrophenol-d4	14.463	143	53869	19.245	ng/u1	0.00
60) Fluorene-d10	15.234	176	251315	19.176	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.369	200	42172	20.090	ng/u1	0.00
73) Anthracene-d10	17.104	188	385244	18.965	ng/u1	0.00
81) Pyrene-d10	19.363	212	430081	17.481	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.280	264	410395	18.479	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.105	88	20508	7.579	ng/uL	100
5) Pyridine	3.499	79	120909	18.064	ng/u1	100
6) Benzaldehyde	6.717	77	65564	21.769	ng/u1	100
8) Phenol	6.769	94	136080	18.191	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.005	93	109039	18.099	ng/u1	100
12) 2-Chlorophenol	7.122	128	113522	20.091	ng/u1	100
13) 2-Methylphenol	8.016	108	100461	18.647	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.152	45	1835	0.203	ng/u1#	35
16) Acetophenone	8.393	105	178606	20.344	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.381	70	92435	21.372	ng/u1	100
18) 4-Methylphenol	8.346	108	112344	19.454	ng/u1	100
19) Hexachloroethane	8.628	117	49634	20.804	ng/u1	100
22) Nitrobenzene	8.763	77	142728	20.840	ng/u1	100
23) Isophorone	9.293	82	248054	19.745	ng/u1	100
25) 2-Nitrophenol	9.475	139	58223	22.304	ng/u1	100
26) 2,4-Dimethylphenol	9.540	107	137787	20.890	ng/u1	100
27) Bis(2-Chloroethoxy)met...	9.781	93	149640	18.639	ng/u1	100
29) 2,4-Dichlorophenol	10.005	162	96368	20.373	ng/u1	100
30) Naphthalene	10.399	128	360247	18.420	ng/u1	100
32) 4-Chloroaniline	10.528	127	155688	19.641	ng/u1	100
33) Hexachlorobutadiene	10.681	225	60489	20.845	ng/u1	100
34) Caprolactam	11.322	113	31685	20.031	ng/u1	100
35) 4-Chloro-3-methylphenol	11.663	107	122655	22.054	ng/u1	100
36) 2-Methylnaphthalene	12.028	142	235965	19.464	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081622\
 Data File : BP011447.D
 Acq On : 16 Aug 2022 16:52
 Operator : CG/JU
 Sample : SSTD02045
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020647


Quant Time: Aug 17 01:07:47 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 17 01:01:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.246	142	240453	19.554	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.399	216	107083	18.416	ng/ul	100
40) Hexachlorocyclopentadiene	12.375	237	58637	18.819	ng/ul	100
41) 2,4,6-Trichlorophenol	12.652	196	69907	19.870	ng/ul	100
42) 2,4,5-Trichlorophenol	12.722	196	77681	20.687	ng/ul	100
43) 1,1'-Biphenyl	13.057	154	314283	18.121	ng/ul	100
44) 2-Chloronaphthalene	13.099	162	236360	18.163	ng/ul	100
45) 2-Nitroaniline	13.322	65	83127	25.152	ng/ul	100
47) Dimethylphthalate	13.704	163	307636	19.819	ng/ul	100
48) 2,6-Dinitrotoluene	13.828	165	55416	22.833	ng/ul	100
50) Acenaphthylene	13.951	152	397649	18.700	ng/ul	100
51) 3-Nitroaniline	14.157	138	54407	20.379	ng/ul	100
52) Acenaphthene	14.298	153	262674	18.521	ng/ul	100
53) 2,4-Dinitrophenol	14.369	184	27830	19.327	ng/ul	100
55) 4-Nitrophenol	14.475	109	65730	25.314	ng/ul	100
56) Dibenzofuran	14.640	168	361563	18.805	ng/ul	100
57) 2,4-Dinitrotoluene	14.622	165	84310	22.664	ng/ul	100
58) 2,3,4,6-Tetrachlorophenol	14.869	232	61733	21.133	ng/ul	100
59) Diethylphthalate	15.087	149	342433	21.295	ng/ul	100
61) Fluorene	15.293	166	299012	19.149	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.293	204	134554	19.713	ng/ul	100
63) 4-Nitroaniline	15.328	138	48098	18.515	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.387	198	43275	20.020	ng/ul	100
67) N-Nitrosodiphenylamine	15.516	169	256493	19.260	ng/ul	100
68) 4-Bromophenyl-phenylether	16.198	248	77300	19.329	ng/ul	100
69) Hexachlorobenzene	16.298	284	89280	19.109	ng/ul	100
70) Atrazine	16.487	200	92314	20.796	ng/ul	100
71) Pentachlorophenol	16.651	266	48704	17.956	ng/ul	100
72) Phenanthrene	17.045	178	461951	18.229	ng/ul	100
74) Anthracene	17.140	178	476810	18.900	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.016	216	109736	16.914	ng/uL	100
76) Pentachlorobenzene	14.551	250	105080	18.597	ng/uL	100
77) Carbazole	17.422	167	433780	18.401	ng/ul	100
78) Di-n-butylphthalate	17.992	149	557842	21.641	ng/ul	100
80) Fluoranthene	19.039	202	523266	17.474	ng/ul	100
82) Pyrene	19.392	202	552650	17.485	ng/ul	100
83) Butylbenzylphthalate	20.292	149	251796	24.999	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.063	252	150125	18.472	ng/ul	100
85) Benzo(a)anthracene	21.116	228	521292	18.195	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.063	149	367358	22.719	ng/ul	100
87) Chrysene	21.169	228	504665	17.794	ng/ul	100
89) Di-n-octyl phthalate	21.957	149	615881	20.157	ng/ul	100
90) Benzo(b)fluoranthene	22.727	252	521991	18.010	ng/ul	100
91) Benzo(k)fluoranthene	22.775	252	508867	17.945	ng/ul	100
93) Benzo(a)pyrene	23.327	252	516913	18.491	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	25.786	276	567399	18.360	ng/ul	100
95) Dibenzo(a,h)anthracene	25.804	278	484323	18.428	ng/ul	100
96) Benzo(g,h,i)perylene	26.510	276	487180	18.683	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 17 01:07:47 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081622.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Aug 17 01:01:09 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 7.579 ng/uL
 RT: 3.105 min Scan# 20
 Delta R.T. 0.000 min
 Lab File: BP011447.D
 Acq: 16 Aug 2022 16:52




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020647


Tgt Ion: 88	Resp: 20508
Ion Ratio	Lower Upper
88	100
43	36.2 29.0 43.4
58	80.9 64.7 97.1



#5
 Pyridine
 Concen: 18.064 ng/uL
 RT: 3.499 min Scan# 87
 Delta R.T. 0.000 min
 Lab File: BP011447.D
 Acq: 16 Aug 2022 16:52

Tgt Ion: 79	Resp: 120909
Ion Ratio	Lower Upper
79	100
52	87.7 70.2 105.2
51	45.1 36.1 54.1









#6
Benzaldehyde
Concen: 21.769 ng/ul
RT: 6.717 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion: 77 Resp: 65564
Ion Ratio Lower Upper
77 100
105 90.0 72.0 108.0
106 84.2 67.4 101.0



#8
Phenol
Concen: 18.191 ng/ul
RT: 6.769 min Scan# 643
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion: 94 Resp: 136080
Ion Ratio Lower Upper
94 100
65 36.1 28.9 43.3
66 49.2 39.4 59.0









#10
 Bis(2-Chloroethyl)ether
 Concen: 18.099 ng/ul
 RT: 7.005 min Scan# 683
 Delta R.T. 0.000 min
 Lab File: BP011447.D
 Acq: 16 Aug 2022 16:52




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020647


Tgt Ion: 93 Resp: 109039
 Ion Ratio Lower Upper
 93 100
 63 84.7 67.8 101.6
 95 31.3 25.0 37.6



#12
 2-Chlorophenol
 Concen: 20.091 ng/ul
 RT: 7.122 min Scan# 703
 Delta R.T. 0.000 min
 Lab File: BP011447.D
 Acq: 16 Aug 2022 16:52

Tgt Ion: 128 Resp: 113522
 Ion Ratio Lower Upper
 128 100
 64 50.2 40.2 60.2
 130 30.8 24.6 37.0









#13
2-Methylphenol
Concen: 18.647 ng/ul
RT: 8.016 min Scan# 855
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:108 Resp: 100461
Ion Ratio Lower Upper
108 100
107 91.5 73.2 109.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 0.203 ng/ul
RT: 8.152 min Scan# 878
Delta R.T. 0.047 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion: 45 Resp: 1835
Ion Ratio Lower Upper
45 100
77 51.3 11.4 17.0#
79 22.7 8.6 13.0#









#16
Acetophenone
Concen: 20.344 ng/ul
RT: 8.393 min Scan# 919
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:105 Resp: 178606
Ion Ratio Lower Upper
105 100
77 83.0 66.4 99.6
51 35.3 28.2 42.4



#17
N-Nitroso-di-n-propylamine
Concen: 21.372 ng/ul
RT: 8.381 min Scan# 917
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion: 70 Resp: 92435
Ion Ratio Lower Upper
70 100
42 71.2 57.0 85.4
101 9.5 7.6 11.4
130 19.0 15.2 22.8









#18
4-Methylphenol
Concen: 19.454 ng/ul
RT: 8.346 min Scan# 911
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:108 Resp: 112344
Ion Ratio Lower Upper
108 100
107 109.6 87.7 131.5



#19
Hexachloroethane
Concen: 20.804 ng/ul
RT: 8.628 min Scan# 959
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:117 Resp: 49634
Ion Ratio Lower Upper
117 100
201 80.3 64.2 96.4
199 51.3 41.0 61.6









#22
 Nitrobenzene
 Concen: 20.840 ng/ul
 RT: 8.763 min Scan# 982
 Delta R.T. 0.000 min
 Lab File: BP011447.D
 Acq: 16 Aug 2022 16:52




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020647


Tgt Ion: 77 Resp: 142728
 Ion Ratio Lower Upper
 77 100
 123 38.0 30.4 45.6
 65 14.4 11.5 17.3



#23
 Isophorone
 Concen: 19.745 ng/ul
 RT: 9.293 min Scan# 1072
 Delta R.T. 0.000 min
 Lab File: BP011447.D
 Acq: 16 Aug 2022 16:52

Tgt Ion: 82 Resp: 248054
 Ion Ratio Lower Upper
 82 100
 95 8.9 7.1 10.7
 138 15.8 12.6 19.0









#25
2-Nitrophenol
Concen: 22.304 ng/ul
RT: 9.475 min Scan# 1103
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:	139	Resp:	58223
Ion Ratio	Lower	Upper	
139	100		
65	63.0	50.4	75.6
109	44.4	35.5	53.3



#26
2,4-Dimethylphenol
Concen: 20.890 ng/ul
RT: 9.540 min Scan# 1114
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:	107	Resp:	137787
Ion Ratio	Lower	Upper	
107	100		
121	45.8	36.6	55.0
122	73.4	58.7	88.1





#27

Bis(2-Chloroethoxy)methane

Concen: 18.639 ng/ul

RT: 9.781 min Scan# 1155

Delta R.T. 0.000 min

Lab File: BP011447.D

Acq: 16 Aug 2022 16:52

Instrument :

BNA_P

ClientSampleId :

SSTD020647





Tgt Ion: 93 Resp: 149640

Ion Ratio Lower Upper

93 100

95 31.0 24.8 37.2

123 10.3 8.2 12.4



#29

2,4-Dichlorophenol

Concen: 20.373 ng/ul

RT: 10.005 min Scan# 1193

Delta R.T. 0.000 min

Lab File: BP011447.D

Acq: 16 Aug 2022 16:52




Tgt Ion:162 Resp: 96368


Ion Ratio Lower Upper

162 100

164 62.7 50.2 75.2

98 41.1 32.9 49.3









#30
Naphthalene
Concen: 18.420 ng/ul
RT: 10.399 min Scan# 1260
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:128 Resp: 360247
Ion Ratio Lower Upper
128 100
129 11.3 9.0 13.6
127 12.9 10.3 15.5



#32
4-Chloroaniline
Concen: 19.641 ng/ul
RT: 10.528 min Scan# 1282
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:127 Resp: 155688
Ion Ratio Lower Upper
127 100
129 31.4 25.1 37.7





#33

Hexachlorobutadiene

Concen: 20.845 ng/ul

RT: 10.681 min Scan# 1308

Delta R.T. 0.000 min

Lab File: BP011447.D

Acq: 16 Aug 2022 16:52

Instrument :

BNA_P

ClientSampleId :

SSTD020647





Tgt Ion:225 Resp: 60489

Ion Ratio Lower Upper

225 100

223 59.8 47.8 71.8

227 62.5 50.0 75.0



#34

Caprolactam

Concen: 20.031 ng/ul

RT: 11.322 min Scan# 1417

Delta R.T. 0.000 min

Lab File: BP011447.D

Acq: 16 Aug 2022 16:52




Tgt Ion:113 Resp: 31685


Ion Ratio Lower Upper

113 100

55 191.8 153.4 230.2

56 149.5 119.6 179.4









#35
4-Chloro-3-methylphenol
Concen: 22.054 ng/ul
RT: 11.663 min Scan# 1475
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:107 Resp: 122655
Ion Ratio Lower Upper
107 100
144 22.9 18.3 27.5
142 69.8 55.8 83.8



#36
2-Methylnaphthalene
Concen: 19.464 ng/ul
RT: 12.028 min Scan# 1537
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:142 Resp: 235965
Ion Ratio Lower Upper
142 100
141 91.7 73.4 110.0









#37
1-Methylnaphthalene
Concen: 19.554 ng/ul
RT: 12.246 min Scan# 1574
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:142 Resp: 240453
Ion Ratio Lower Upper
142 100
141 91.3 73.0 109.6
116 5.0 4.0 6.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 18.416 ng/ul
RT: 12.399 min Scan# 1600
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:216 Resp: 107083
Ion Ratio Lower Upper
216 100
214 75.5 60.4 90.6
179 25.0 20.0 30.0
108 25.3 20.2 30.4









#40
Hexachlorocyclopentadiene
Concen: 18.819 ng/ul
RT: 12.375 min Scan# 1596
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:	237	Resp:	58637
Ion Ratio	Lower	Upper	
237	100		
235	63.6	50.9	76.3
272	13.7	11.0	16.4



#41
2,4,6-Trichlorophenol
Concen: 19.870 ng/ul
RT: 12.652 min Scan# 1643
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:	196	Resp:	69907
Ion Ratio	Lower	Upper	
196	100		
198	94.5	75.6	113.4
200	28.9	23.1	34.7









#42
2,4,5-Trichlorophenol
Concen: 20.687 ng/ul
RT: 12.722 min Scan# 1655
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:196 Resp: 77681
Ion Ratio Lower Upper
196 100
198 94.5 75.6 113.4
200 29.7 23.8 35.6



#43
1,1'-Biphenyl
Concen: 18.121 ng/ul
RT: 13.057 min Scan# 1712
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:154 Resp: 314283
Ion Ratio Lower Upper
154 100
153 40.3 32.2 48.4
76 14.5 11.6 17.4









#44
2-Chloronaphthalene
Concen: 18.163 ng/ul
RT: 13.099 min Scan# 1719
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:162 Resp: 236360
Ion Ratio Lower Upper
162 100
164 33.0 26.4 39.6
127 46.8 37.4 56.2



#45
2-Nitroaniline
Concen: 25.152 ng/ul
RT: 13.322 min Scan# 1757
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion: 65 Resp: 83127
Ion Ratio Lower Upper
65 100
92 63.5 50.8 76.2
138 83.4 66.7 100.1









#47
Dimethylphthalate
Concen: 19.819 ng/ul
RT: 13.704 min Scan# 1822
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:163 Resp: 307636
Ion Ratio Lower Upper
163 100
194 3.2 2.6 3.8
164 10.4 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 22.833 ng/ul
RT: 13.828 min Scan# 1843
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:165 Resp: 55416
Ion Ratio Lower Upper
165 100
89 83.2 66.6 99.8
121 26.3 21.0 31.6









#50
Acenaphthylene
Concen: 18.700 ng/ul
RT: 13.951 min Scan# 1864
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:152 Resp: 397649
Ion Ratio Lower Upper
152 100
151 21.1 16.9 25.3
153 13.2 10.6 15.8



#51
3-Nitroaniline
Concen: 20.379 ng/ul
RT: 14.157 min Scan# 1899
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:138 Resp: 54407
Ion Ratio Lower Upper
138 100
108 10.8 8.6 13.0
92 130.2 104.2 156.2









#52
Acenaphthene
Concen: 18.521 ng/ul
RT: 14.298 min Scan# 1923
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:153 Resp: 262674
Ion Ratio Lower Upper
153 100
152 47.1 37.7 56.5
154 86.7 69.4 104.0



#53
2,4-Dinitrophenol
Concen: 19.327 ng/ul
RT: 14.369 min Scan# 1935
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:184 Resp: 27830
Ion Ratio Lower Upper
184 100
63 103.1 82.5 123.7
154 83.3 66.6 100.0









#55
4-Nitrophenol
Concen: 25.314 ng/ul
RT: 14.475 min Scan# 1953
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:109 Resp: 65730
Ion Ratio Lower Upper
109 100
139 95.0 76.0 114.0
65 123.5 98.8 148.2



#56
Dibenzofuran
Concen: 18.805 ng/ul
RT: 14.640 min Scan# 1981
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:168 Resp: 361563
Ion Ratio Lower Upper
168 100
139 48.0 38.4 57.6









#57
2,4-Dinitrotoluene
Concen: 22.664 ng/ul
RT: 14.622 min Scan# 1978
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:165 Resp: 84310
Ion Ratio Lower Upper
165 100
63 75.2 60.2 90.2
182 2.3 1.8 2.8



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.133 ng/ul
RT: 14.869 min Scan# 2020
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:232 Resp: 61733
Ion Ratio Lower Upper
232 100
131 69.2 55.4 83.0
130 4.5 3.6 5.4
166 38.7 31.0 46.4









#59
Diethylphthalate
Concen: 21.295 ng/ul
RT: 15.087 min Scan# 2057
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:149 Resp: 342433
Ion Ratio Lower Upper
149 100
177 19.6 15.7 23.5
150 12.6 10.1 15.1



#61
Fluorene
Concen: 19.149 ng/ul
RT: 15.293 min Scan# 2092
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:166 Resp: 299012
Ion Ratio Lower Upper
166 100
165 97.7 78.2 117.2
167 12.9 10.3 15.5









#62
4-Chlorophenyl-phenylether
Concen: 19.713 ng/ul
RT: 15.293 min Scan# 2092
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:204 Resp: 134554
Ion Ratio Lower Upper
204 100
206 31.7 25.4 38.0
141 83.1 66.5 99.7



#63
4-Nitroaniline
Concen: 18.515 ng/ul
RT: 15.328 min Scan# 2098
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:138 Resp: 48098
Ion Ratio Lower Upper
138 100
92 61.9 49.5 74.3
108 117.1 93.7 140.5









#66
4,6-Dinitro-2-methylphenol
Concen: 20.020 ng/ul
RT: 15.387 min Scan# 2108
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:198 Resp: 43275
Ion Ratio Lower Upper
198 100
182 4.9 3.9 5.9
77 37.6 30.1 45.1



#67
N-Nitrosodiphenylamine
Concen: 19.260 ng/ul
RT: 15.516 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:169 Resp: 256493
Ion Ratio Lower Upper
169 100
168 68.3 54.6 82.0
167 37.8 30.2 45.4









#68
4-Bromophenyl-phenylether
Concen: 19.329 ng/ul
RT: 16.198 min Scan# 211
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:248 Resp: 77300
Ion Ratio Lower Upper
248 100
250 103.5 82.8 124.2
141 113.5 90.8 136.2



#69
Hexachlorobenzene
Concen: 19.109 ng/ul
RT: 16.298 min Scan# 2263
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:284 Resp: 89280
Ion Ratio Lower Upper
284 100
142 54.4 43.5 65.3
249 35.5 28.4 42.6





#70

Atrazine

Concen: 20.796 ng/ul

RT: 16.487 min Scan# 2295

Delta R.T. 0.000 min

Lab File: BP011447.D

Acq: 16 Aug 2022 16:52

Instrument :

BNA_P

ClientSampleId :

SSTD020647





Tgt Ion:200 Resp: 92314

Ion Ratio Lower Upper

200 100

173 27.5 22.0 33.0

215 45.4 36.3 54.5



#71

Pentachlorophenol

Concen: 17.956 ng/ul

RT: 16.651 min Scan# 2323

Delta R.T. 0.000 min

Lab File: BP011447.D

Acq: 16 Aug 2022 16:52




Tgt Ion:266 Resp: 48704


Ion Ratio Lower Upper

266 100

264 63.3 50.6 76.0

268 64.1 51.3 76.9





#72

Phenanthrene

Concen: 18.229 ng/ul

RT: 17.045 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP011447.D

Acq: 16 Aug 2022 16:52

Instrument :

BNA_P

ClientSampleId :

SSTD020647





Tgt Ion:178 Resp: 461951

Ion Ratio Lower Upper

178 100

179 15.7 12.6 18.8

176 19.8 15.8 23.8



#74

Anthracene

Concen: 18.900 ng/ul

RT: 17.140 min Scan# 2406

Delta R.T. 0.000 min

Lab File: BP011447.D

Acq: 16 Aug 2022 16:52




Tgt Ion:178 Resp: 476810


Ion Ratio Lower Upper

178 100

179 15.2 12.2 18.2

176 19.2 15.4 23.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 16.914 ng/uL
RT: 13.016 min Scan# 1705
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:216 Resp: 109736
Ion Ratio Lower Upper
216 100
214 77.2 61.8 92.6
218 46.7 37.4 56.0



#76
Pentachlorobenzene
Concen: 18.597 ng/uL
RT: 14.551 min Scan# 1966
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:250 Resp: 105080
Ion Ratio Lower Upper
250 100
248 64.6 51.7 77.5
252 62.5 50.0 75.0









#77
 Carbazole
 Concen: 18.401 ng/ul
 RT: 17.422 min Scan# 2454
 Delta R.T. 0.000 min
 Lab File: BP011447.D
 Acq: 16 Aug 2022 16:52




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020647


Tgt Ion:167 Resp: 433780
 Ion Ratio Lower Upper
 167 100
 166 21.9 17.5 26.3
 139 15.1 12.1 18.1



#78
 Di-n-butylphthalate
 Concen: 21.641 ng/ul
 RT: 17.992 min Scan# 2551
 Delta R.T. 0.000 min
 Lab File: BP011447.D
 Acq: 16 Aug 2022 16:52

Tgt Ion:149 Resp: 557842
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.4 11.2
 104 6.9 5.5 8.3









#80
Fluoranthene
Concen: 17.474 ng/u1
RT: 19.039 min Scan# 2729
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:202 Resp: 523266
Ion Ratio Lower Upper
202 100
101 11.9 9.5 14.3
100 9.9 7.9 11.9



#82
Pyrene
Concen: 17.485 ng/u1
RT: 19.392 min Scan# 2789
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:202 Resp: 552650
Ion Ratio Lower Upper
202 100
101 14.3 11.4 17.2
100 11.7 9.4 14.0









#83
Butylbenzylphthalate
Concen: 24.999 ng/ul
RT: 20.292 min Scan# 2942
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:149 Resp: 251796
Ion Ratio Lower Upper
149 100
91 85.6 68.5 102.7
206 15.7 12.6 18.8



#84
3,3'-Dichlorobenzidine
Concen: 18.472 ng/ul
RT: 21.063 min Scan# 3073
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:252 Resp: 150125
Ion Ratio Lower Upper
252 100
254 64.0 51.2 76.8
126 15.2 12.2 18.2





#85

Benzo(a)anthracene

Concen: 18.195 ng/ul

RT: 21.116 min Scan# 3082

Delta R.T. 0.000 min

Lab File: BP011447.D

Acq: 16 Aug 2022 16:52

Instrument :

BNA_P

ClientSampleId :

SSTD020647





Tgt Ion:228 Resp: 521292

Ion Ratio Lower Upper

228 100

229 20.0 16.0 24.0

226 28.2 22.6 33.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.719 ng/ul

RT: 21.063 min Scan# 3073

Delta R.T. 0.000 min

Lab File: BP011447.D

Acq: 16 Aug 2022 16:52




Tgt Ion:149 Resp: 367358


Ion Ratio Lower Upper

149 100

167 24.8 19.8 29.8

279 3.4 2.7 4.1









#87
Chrysene
Concen: 17.794 ng/ul
RT: 21.169 min Scan# 3091
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:228 Resp: 504665
Ion Ratio Lower Upper
228 100
226 29.9 23.9 35.9
229 20.0 16.0 24.0



#89
Di-n-octyl phthalate
Concen: 20.157 ng/ul
RT: 21.957 min Scan# 3225
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:149 Resp: 615881









#90
Benzo(b)fluoranthene
Concen: 18.010 ng/ul
RT: 22.727 min Scan# 3356
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:252 Resp: 521991
Ion Ratio Lower Upper
252 100
253 21.9 17.5 26.3
125 11.2 9.0 13.4



#91
Benzo(k)fluoranthene
Concen: 17.945 ng/ul
RT: 22.775 min Scan# 3364
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:252 Resp: 508867
Ion Ratio Lower Upper
252 100
253 22.4 17.9 26.9
125 12.3 9.8 14.8









#93
Benzo(a)pyrene
Concen: 18.491 ng/ul
RT: 23.327 min Scan# 3458
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52




Instrument :
BNA_P
ClientSampleId :
SSTD020647


Tgt Ion:252 Resp: 516913
Ion Ratio Lower Upper
252 100
253 21.0 16.8 25.2
125 12.4 9.9 14.9



#94
Indeno(1,2,3-cd)pyrene
Concen: 18.360 ng/ul
RT: 25.786 min Scan# 3876
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:276 Resp: 567399
Ion Ratio Lower Upper
276 100
138 25.3 20.2 30.4
277 25.1 20.1 30.1









#95
Dibenzo(a,h)anthracene
Concen: 18.428 ng/ul
RT: 25.804 min Scan# 3879
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Instrument :
BNA_P
ClientSampleId :
SSTD020647

Tgt Ion:278 Resp: 484323

Ion	Ratio	Lower	Upper
278	100		
139	19.2	15.4	23.0
279	24.1	19.3	28.9



#96
Benzo(g,h,i)perylene
Concen: 18.683 ng/ul
RT: 26.510 min Scan# 3999
Delta R.T. 0.000 min
Lab File: BP011447.D
Acq: 16 Aug 2022 16:52

Tgt Ion:276 Resp: 487180

Ion	Ratio	Lower	Upper
276	100		
138	24.3	19.4	29.2
277	23.9	19.1	28.7

