




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081721\
Data File : BP006707.D
Acq On : 17 Aug 2021 08:40
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 17 11:27:24 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 16 18:17:59 2021
Response via : Initial Calibration



TIC: BP006707.D\data.ms

(70) Pentachlorophenol (C)




16.757min (+ 0.006) 16962.56 ng

response 288611

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	64.50	62.25
264.00	62.90	63.09
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081721\
 Data File : BP006707.D
 Acq On : 17 Aug 2021 08:40
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 17 11:27:24 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 16 18:17:59 2021
 Response via : Initial Calibration



TIC: BP006707.D\data.ms

(77) Benzidine

19.321min (+ 0.006) 3727.76 ng

response 1652751

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.70	15.54
183.00	11.90	11.87
0.00	0.00	0.00