

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081721\
 Data File : BP006717.D
 Acq On : 17 Aug 2021 14:39
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 17 17:38:36 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 16 18:17:59 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	100	0.00
2	1,4-Dioxane	40.000	37.202	7.0	94	0.00
3	Pyridine	40.000	36.272	9.3	86	0.00
4	n-Nitrosodimethylamine	40.000	35.295	11.8	85	0.00
5 S	2-Fluorophenol	80.000	74.585	6.8	90	0.00
6	Aniline	40.000	35.655	10.9	81	0.00
7 S	Phenol-d6	80.000	72.923	8.8	85	0.00
8	2-Chlorophenol	40.000	37.536	6.2	89	0.00
9	Benzaldehyde	40.000	35.072	12.3	86	0.00
10 C	Phenol	40.000	35.960	10.1	83	0.00
11	bis(2-Chloroethyl)ether	40.000	35.875	10.3	85	0.00
12	1,3-Dichlorobenzene	40.000	37.044	7.4	89	0.00
13 C	1,4-Dichlorobenzene	40.000	36.921	7.7	89	0.00
14	1,2-Dichlorobenzene	40.000	37.284	6.8	89	0.00
15	Benzyl Alcohol	40.000	36.142	9.6	82	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	34.983	12.5	81	-0.01
17	2-Methylphenol	40.000	37.450	6.4	86	0.00
18	Hexachloroethane	40.000	38.661	3.3	94	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.592	8.5	82	0.00
20	3+4-Methylphenols	40.000	37.297	6.8	84	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	100	0.00
22	Acetophenone	40.000	35.419	11.5	84	0.00
23 S	Nitrobenzene-d5	80.000	72.825	9.0	87	0.00
24	Nitrobenzene	40.000	35.965	10.1	87	0.00
25	Isophorone	40.000	35.743	10.6	83	0.00
26 C	2-Nitrophenol	40.000	37.968	5.1	89	0.00
27	2,4-Dimethylphenol	40.000	35.351	11.6	84	0.00
28	bis(2-Chloroethoxy)methane	40.000	34.849	12.9	84	0.00
29 C	2,4-Dichlorophenol	40.000	37.258	6.9	87	0.00
30	1,2,4-Trichlorobenzene	40.000	36.073	9.8	89	0.00
31	Naphthalene	40.000	35.664	10.8	86	0.00
32	Benzoic acid	40.000	37.018	7.5	88	0.00
33	4-Chloroaniline	40.000	35.033	12.4	82	0.00
34 C	Hexachlorobutadiene	40.000	37.375	6.6	93	0.00
35	Caprolactam	40.000	35.635	10.9	77	0.00
36 C	4-Chloro-3-methylphenol	40.000	36.587	8.5	84	0.00
37	2-Methylnaphthalene	40.000	35.597	11.0	84	0.00
38	1-Methylnaphthalene	40.000	35.540	11.2	85	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	96	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	35.865	10.3	86	0.00
41 P	Hexachlorocyclopentadiene	40.000	32.334	19.2	75	0.00
42 S	2,4,6-Tribromophenol	80.000	73.819	7.7	83	0.00
43 C	2,4,6-Trichlorophenol	40.000	37.346	6.6	85	0.00
44	2,4,5-Trichlorophenol	40.000	36.769	8.1	84	0.00
45 S	2-Fluorobiphenyl	80.000	70.760	11.5	84	0.00
46	1,1'-Biphenyl	40.000	35.473	11.3	84	0.00
47	2-Chloronaphthalene	40.000	35.766	10.6	84	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081721\
 Data File : BP006717.D
 Acq On : 17 Aug 2021 14:39
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 17 17:38:36 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 16 18:17:59 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	36.578	8.6	80	0.00
49	Acenaphthylene	40.000	35.931	10.2	83	0.00
50	Dimethylphthalate	40.000	35.218	12.0	81	0.00
51	2,6-Dinitrotoluene	40.000	35.530	11.2	79	0.00
52 C	Acenaphthene	40.000	35.666	10.8	82	0.00
53	3-Nitroaniline	40.000	35.285	11.8	78	0.00
54 P	2,4-Dinitrophenol	40.000	32.114	19.7	73	0.00
55	Dibenzofuran	40.000	35.563	11.1	83	0.00
56 P	4-Nitrophenol	40.000	35.266	11.8	76	0.00
57	2,4-Dinitrotoluene	40.000	37.153	7.1	82	0.00
58	Fluorene	40.000	35.685	10.8	82	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	37.294	6.8	84	0.00
60	Diethylphthalate	40.000	36.120	9.7	82	0.00
61	4-Chlorophenyl-phenylether	40.000	35.627	10.9	83	0.00
62	4-Nitroaniline	40.000	35.626	10.9	77	0.00
63	Azobenzene	40.000	37.235	6.9	84	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	93	0.00
65	4,6-Dinitro-2-methylphenol	40.000	33.599	16.0	75	0.00
66 c	n-Nitrosodiphenylamine	40.000	35.086	12.3	80	0.00
67	4-Bromophenyl-phenylether	40.000	36.194	9.5	83	0.00
68	Hexachlorobenzene	40.000	35.439	11.4	83	0.00
69	Atrazine	40.000	35.776	10.6	79	0.00
70 C	Pentachlorophenol	40.000	37.183	7.0	81	0.00
71	Phenanthrene	40.000	35.346	11.6	81	0.00
72	Anthracene	40.000	35.859	10.4	81	0.00
73	Carbazole	40.000	35.298	11.8	78	0.00
74	Di-n-butylphthalate	40.000	37.361	6.6	83	0.00
75 C	Fluoranthene	40.000	35.327	11.7	79	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	89	0.00
77	Benzydine	40.000	33.219	17.0	63	0.00
78	Pyrene	40.000	36.169	9.6	78	0.00
79 S	Terphenyl-d14	80.000	73.470	8.2	81	0.00
80	Butylbenzylphthalate	40.000	39.335	1.7	82	0.00
81	Benzo(a)anthracene	40.000	35.840	10.4	78	0.00
82	3,3'-Dichlorobenzidine	40.000	35.440	11.4	75	0.00
83	Chrysene	40.000	36.198	9.5	79	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	38.679	3.3	82	0.00
85 c	Di-n-octyl phthalate	40.000	39.707	0.7	84	0.00
86 I	Perylene-d12	20.000	20.000	0.0	90	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	36.421	8.9	79	0.00
88	Benzo(b)fluoranthene	40.000	35.874	10.3	78	0.00
89	Benzo(k)fluoranthene	40.000	35.679	10.8	77	0.00
90 C	Benzo(a)pyrene	40.000	36.400	9.0	78	0.00
91	Dibenzo(a,h)anthracene	40.000	36.381	9.0	79	0.00
92	Benzo(g,h,i)perylene	40.000	36.652	8.4	79	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081721\
Data File : BP006717.D
Acq On : 17 Aug 2021 14:39
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Aug 17 17:38:36 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 16 18:17:59 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0