

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081922\
 Data File : BPCCC514.D
 Acq On : 19 Aug 2022 19:32
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Aug 20 04:44:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 17 01:58:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.557	152	102476	20.000	ng/u1	0.00
20) Naphthalene-d8	10.340	136	423838	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.228	164	237297	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.998	188	474277	20.000	ng/u1	0.00
79) Chrysene-d12	21.127	240	454849	20.000	ng/u1	0.00
88) Perylene-d12	23.421	264	466035	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.069	96	23909	7.240	ng/uL	0.00
4) Pyridine-d5	3.481	84	149213	17.648	ng/u1	0.00
7) Phenol-d5	6.740	99	168036	17.844	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.904	67	108031	17.658	ng/u1	0.00
11) 2-Chlorophenol-d4	7.087	132	134770	19.458	ng/u1	0.00
15) 4-Methylphenol-d8	8.281	113	129829	17.225	ng/u1	0.00
21) Nitrobenzene-d5	8.722	128	63875	20.933	ng/u1	0.00
24) 2-Nitrophenol-d4	9.434	143	64988	20.927	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	9.969	165	115428	20.137	ng/u1	0.00
31) 4-Chloroaniline-d4	10.498	131	173416	18.174	ng/u1	0.00
46) Dimethylphthalate-d6	13.651	166	325933	19.420	ng/u1	0.00
49) Acenaphthylene-d8	13.916	160	377139	20.218	ng/u1	0.00
54) 4-Nitrophenol-d4	14.457	143	60464	18.698	ng/u1	0.00
60) Fluorene-d10	15.233	176	269400	19.079	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.369	200	45772	17.860	ng/u1	0.00
73) Anthracene-d10	17.098	188	414791	19.747	ng/u1	0.00
81) Pyrene-d10	19.363	212	453591	19.710	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.268	264	444962	19.628	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.105	88	26043	7.563	ng/uL	94
5) Pyridine	3.499	79	153935	18.240	ng/u1	98
6) Benzaldehyde	6.710	77	86644	20.049	ng/u1	99
8) Phenol	6.769	94	172007	17.774	ng/u1	96
10) Bis(2-Chloroethyl)ether	6.999	93	134147	17.770	ng/u1	97
12) 2-Chlorophenol	7.122	128	142497	19.119	ng/u1	99
13) 2-Methylphenol	8.010	108	129415	18.153	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.081	45	86054	7.695	ng/u1	98
16) Acetophenone	8.387	105	207252	16.529	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.375	70	110120	17.656	ng/u1	93
18) 4-Methylphenol	8.340	108	136474	17.291	ng/u1	95
19) Hexachloroethane	8.616	117	63918	18.907	ng/u1	96
22) Nitrobenzene	8.763	77	169261	19.680	ng/u1	97
23) Isophorone	9.287	82	292853	19.397	ng/u1	98
25) 2-Nitrophenol	9.469	139	73583	20.768	ng/u1	96
26) 2,4-Dimethylphenol	9.540	107	163994	19.494	ng/u1	97
27) Bis(2-Chloroethoxy)met...	9.781	93	173516	19.014	ng/u1	97
29) 2,4-Dichlorophenol	9.998	162	117537	20.036	ng/u1	98
30) Naphthalene	10.393	128	428954	19.168	ng/u1	98
32) 4-Chloroaniline	10.522	127	176404	17.843	ng/u1	99
33) Hexachlorobutadiene	10.669	225	75498	20.216	ng/u1	95
34) Caprolactam	11.316	113	38456	18.079	ng/u1	88
35) 4-Chloro-3-methylphenol	11.657	107	141766	19.114	ng/u1	98
36) 2-Methylnaphthalene	12.022	142	270014	18.545	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081922\
 Data File : BPCCC514.D
 Acq On : 19 Aug 2022 19:32
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 17 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :


Quant Time: Aug 20 04:44:49 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Aug 17 01:58:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.240	142	275287	18.642	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.392	216	123348	20.480	ng/ul	96
40) Hexachlorocyclopentadiene	12.363	237	59959	16.046	ng/ul	97
41) 2,4,6-Trichlorophenol	12.645	196	83382	21.510	ng/ul	97
42) 2,4,5-Trichlorophenol	12.716	196	88289	20.460	ng/ul	97
43) 1,1'-Biphenyl	13.057	154	350393	19.817	ng/ul	98
44) 2-Chloronaphthalene	13.092	162	265620	19.935	ng/ul	100
45) 2-Nitroaniline	13.316	65	95282	20.953	ng/ul	97
47) Dimethylphthalate	13.698	163	334067	19.258	ng/ul	99
48) 2,6-Dinitrotoluene	13.822	165	65194	20.904	ng/ul	98
50) Acenaphthylene	13.945	152	441521	19.923	ng/ul	98
51) 3-Nitroaniline	14.151	138	63762	18.770	ng/ul	97
52) Acenaphthene	14.292	153	289099	19.477	ng/ul	100
53) 2,4-Dinitrophenol	14.369	184	31099	16.292	ng/ul	96
55) 4-Nitrophenol	14.475	109	69610	18.067	ng/ul	97
56) Dibenzofuran	14.634	168	392095	19.156	ng/ul	100
57) 2,4-Dinitrotoluene	14.616	165	95638	20.485	ng/ul	95
58) 2,3,4,6-Tetrachlorophenol	14.863	232	69919	20.638	ng/ul	97
59) Diethylphthalate	15.081	149	366979	19.315	ng/ul	99
61) Fluorene	15.286	166	317106	18.937	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.292	204	143028	18.907	ng/ul	98
63) 4-Nitroaniline	15.328	138	59031	17.990	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.386	198	47163	17.791	ng/ul#	98
67) N-Nitrosodiphenylamine	15.510	169	270155	19.415	ng/ul	99
68) 4-Bromophenyl-phenylether	16.192	248	83594	19.745	ng/ul	97
69) Hexachlorobenzene	16.292	284	96792	19.837	ng/ul	97
70) Atrazine	16.480	200	101185	19.486	ng/ul	96
71) Pentachlorophenol	16.645	266	59487	19.662	ng/ul	96
72) Phenanthrene	17.039	178	494060	19.246	ng/ul	99
74) Anthracene	17.133	178	506284	19.549	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.010	216	127118	20.850	ng/uL	99
76) Pentachlorobenzene	14.545	250	116130	20.074	ng/uL	99
77) Carbazole	17.416	167	464763	19.079	ng/ul	99
78) Di-n-butylphthalate	17.986	149	626161	20.873	ng/ul	99
80) Fluoranthene	19.033	202	560704	19.812	ng/ul	99
82) Pyrene	19.392	202	582482	19.588	ng/ul	98
83) Butylbenzylphthalate	20.286	149	298898	22.407	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.057	252	155922	18.108	ng/ul	99
85) Benzo(a)anthracene	21.116	228	544343	19.643	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.057	149	442422	22.412	ng/ul	99
87) Chrysene	21.163	228	534556	19.637	ng/ul	99
89) Di-n-octyl phthalate	21.951	149	750317	20.028	ng/ul	100
90) Benzo(b)fluoranthene	22.721	252	560665	19.340	ng/ul	99
91) Benzo(k)fluoranthene	22.768	252	530961	19.193	ng/ul	99
93) Benzo(a)pyrene	23.315	252	546243	19.300	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	25.774	276	622726	19.726	ng/ul	99
95) Dibenzo(a,h)anthracene	25.798	278	534749	19.673	ng/ul	99
96) Benzo(g,h,i)perylene	26.498	276	525535	19.571	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Aug 20 04:44:49 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081622.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Aug 17 01:58:37 2022
Response via : Initial Calibration








#2
1,4-Dioxane
Concen: 7.563 ng/uL
RT: 3.105 min Scan# 20
Delta R.T. -0.000 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Instrument :
BNA_P
ClientSampleId :







Tgt Ion: 88 Resp: 26043
Ion Ratio Lower Upper
88 100
43 33.1 29.0 43.4
58 74.7 64.7 97.1



#5
Pyridine
Concen: 18.240 ng/uL
RT: 3.499 min Scan# 87
Delta R.T. -0.000 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion: 79 Resp: 153935
Ion Ratio Lower Upper
79 100
52 88.5 70.2 105.2
51 43.0 36.1 54.1









#6
Benzaldehyde
Concen: 20.049 ng/ul
RT: 6.710 min Scan# 61
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 86644
Ion Ratio Lower Upper
77 100
105 89.1 72.0 108.0
106 82.5 67.4 101.0



#8
Phenol
Concen: 17.774 ng/ul
RT: 6.769 min Scan# 643
Delta R.T. -0.000 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32


Tgt Ion: 94 Resp: 172007
Ion Ratio Lower Upper
94 100
65 34.0 28.9 43.3
66 46.5 39.4 59.0








#10
 Bis(2-Chloroethyl)ether
 Concen: 17.770 ng/ul
 RT: 6.999 min Scan# 682
 Delta R.T. -0.006 min
 Lab File: BPCCC514.D
 Acq: 19 Aug 2022 19:32

Instrument :
 BNA_P
 ClientSampleId :







Tgt Ion: 93 Resp: 134147
 Ion Ratio Lower Upper
 93 100
 63 82.5 67.8 101.6
 95 33.7 25.0 37.6



#12
 2-Chlorophenol
 Concen: 19.119 ng/ul
 RT: 7.122 min Scan# 703
 Delta R.T. -0.000 min
 Lab File: BPCCC514.D
 Acq: 19 Aug 2022 19:32

Tgt Ion:128 Resp: 142497
 Ion Ratio Lower Upper
 128 100
 64 49.3 40.2 60.2
 130 31.7 24.6 37.0





#13

2-Methylphenol

Concen: 18.153 ng/ul

RT: 8.010 min Scan# 855

Delta R.T. -0.006 min

Lab File: BPCCC514.D

Acq: 19 Aug 2022 19:32

Instrument :

BNA_P





ClientSampleId :

Tgt Ion:108 Resp: 129415

Ion Ratio Lower Upper

108 100

107 90.3 73.2 109.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 7.695 ng/ul

RT: 8.081 min Scan# 866

Delta R.T. -0.024 min

Lab File: BPCCC514.D

Acq: 19 Aug 2022 19:32




Tgt Ion: 45 Resp: 86054


Ion Ratio Lower Upper

45 100

77 15.2 11.4 17.0

79 10.2 8.6 13.0









#16
Acetophenone
Concen: 16.529 ng/ul
RT: 8.387 min Scan# 918
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:105 Resp: 207252
Ion Ratio Lower Upper
105 100
77 84.9 66.4 99.6
51 33.9 28.2 42.4



#17
N-Nitroso-di-n-propylamine
Concen: 17.656 ng/ul
RT: 8.375 min Scan# 916
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion: 70 Resp: 110120
Ion Ratio Lower Upper
70 100
42 64.2 57.0 85.4
101 9.8 7.6 11.4
130 20.9 15.2 22.8









#18
4-Methylphenol
Concen: 17.291 ng/ul
RT: 8.340 min Scan# 912
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:108 Resp: 136474
Ion Ratio Lower Upper
108 100
107 114.9 87.7 131.5



#19
Hexachloroethane
Concen: 18.907 ng/ul
RT: 8.616 min Scan# 957
Delta R.T. -0.012 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion:117 Resp: 63918
Ion Ratio Lower Upper
117 100
201 84.4 64.2 96.4
199 48.6 41.0 61.6









#22
Nitrobenzene
Concen: 19.680 ng/ul
RT: 8.763 min Scan# 91
Delta R.T. -0.000 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 169261
Ion Ratio Lower Upper
77 100
123 40.0 30.4 45.6
65 14.2 11.5 17.3



#23
Isophorone
Concen: 19.397 ng/ul
RT: 9.287 min Scan# 1071
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion: 82 Resp: 292853
Ion Ratio Lower Upper
82 100
95 8.1 7.1 10.7
138 16.9 12.6 19.0









#25
2-Nitrophenol
Concen: 20.768 ng/ul
RT: 9.469 min Scan# 1102
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:139 Resp: 73583
Ion Ratio Lower Upper
139 100
65 62.0 50.4 75.6
109 39.6 35.5 53.3



#26
2,4-Dimethylphenol
Concen: 19.494 ng/ul
RT: 9.540 min Scan# 1114
Delta R.T. -0.000 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion:107 Resp: 163994
Ion Ratio Lower Upper
107 100
121 45.4 36.6 55.0
122 77.7 58.7 88.1









#27
Bis(2-Chloroethoxy)methane
Concen: 19.014 ng/ul
RT: 9.781 min Scan# 1155
Delta R.T. -0.000 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 93 Resp: 173516
Ion Ratio Lower Upper
93 100
95 32.8 24.8 37.2
123 11.5 8.2 12.4



#29
2,4-Dichlorophenol
Concen: 20.036 ng/ul
RT: 9.998 min Scan# 1192
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion:162 Resp: 117537
Ion Ratio Lower Upper
162 100
164 63.8 50.2 75.2
98 42.2 32.9 49.3









#30
Naphthalene
Concen: 19.168 ng/ul
RT: 10.393 min Scan# 1259
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:128 Resp: 428954
Ion Ratio Lower Upper
128 100
129 10.6 9.0 13.6
127 13.4 10.3 15.5



#32
4-Chloroaniline
Concen: 17.843 ng/ul
RT: 10.522 min Scan# 1281
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32


Tgt Ion:127 Resp: 176404
Ion Ratio Lower Upper
127 100
129 32.0 25.1 37.7








#33
Hexachlorobutadiene
Concen: 20.216 ng/ul
RT: 10.669 min Scan# 11
Delta R.T. -0.012 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:225 Resp: 75498
Ion Ratio Lower Upper
225 100
223 65.1 47.8 71.8
227 65.0 50.0 75.0



#34
Caprolactam
Concen: 18.079 ng/ul
RT: 11.316 min Scan# 1416
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion:113 Resp: 38456
Ion Ratio Lower Upper
113 100
55 174.6 153.4 230.2
56 133.0 119.6 179.4









#35
4-Chloro-3-methylphenol
Concen: 19.114 ng/ul
RT: 11.657 min Scan# 1476
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:107 Resp: 141766
Ion Ratio Lower Upper
107 100
144 22.0 18.3 27.5
142 68.3 55.8 83.8



#36
2-Methylnaphthalene
Concen: 18.545 ng/ul
RT: 12.022 min Scan# 1536
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion:142 Resp: 270014
Ion Ratio Lower Upper
142 100
141 90.9 73.4 110.0









#37
1-Methylnaphthalene
Concen: 18.642 ng/ul
RT: 12.240 min Scan# 1575
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:142 Resp: 275287
Ion Ratio Lower Upper
142 100
141 91.8 73.0 109.6
116 4.6 4.0 6.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.480 ng/ul
RT: 12.392 min Scan# 1599
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion:216 Resp: 123348
Ion Ratio Lower Upper
216 100
214 80.2 60.4 90.6
179 24.7 20.0 30.0
108 25.5 20.2 30.4





#40

Hexachlorocyclopentadiene

Concen: 16.046 ng/ul

RT: 12.363 min Scan# 1594

Delta R.T. -0.012 min

Lab File: BPCCC514.D





Acq: 19 Aug 2022 19:32

Instrument :

BNA_P

ClientSampleId :

Tgt Ion:237	Resp:	59959
Ion Ratio	Lower	Upper
237	100	
235	60.9	50.9 76.3
272	12.7	11.0 16.4



#41

2,4,6-Trichlorophenol

Concen: 21.510 ng/ul




RT: 12.645 min Scan# 1642


Delta R.T. -0.006 min

Lab File: BPCCC514.D

Acq: 19 Aug 2022 19:32

Tgt Ion:196	Resp:	83382
Ion Ratio	Lower	Upper
196	100	
198	97.8	75.6 113.4
200	30.5	23.1 34.7









#42
2,4,5-Trichlorophenol
Concen: 20.460 ng/ul
RT: 12.716 min Scan# 1654
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:196 Resp: 88289
Ion Ratio Lower Upper
196 100
198 97.8 75.6 113.4
200 30.4 23.8 35.6



#43
1,1'-Biphenyl
Concen: 19.817 ng/ul
RT: 13.057 min Scan# 1712
Delta R.T. -0.000 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion:154 Resp: 350393
Ion Ratio Lower Upper
154 100
153 40.9 32.2 48.4
76 13.1 11.6 17.4









#44
2-Chloronaphthalene
Concen: 19.935 ng/ul
RT: 13.092 min Scan# 1718
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:162 Resp: 265620
Ion Ratio Lower Upper
162 100
164 33.1 26.4 39.6
127 46.8 37.4 56.2



#45
2-Nitroaniline
Concen: 20.953 ng/ul
RT: 13.316 min Scan# 1756
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion: 65 Resp: 95282
Ion Ratio Lower Upper
65 100
92 64.6 50.8 76.2
138 86.5 66.7 100.1









#47
Dimethylphthalate
Concen: 19.258 ng/ul
RT: 13.698 min Scan# 1821
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:163 Resp: 334067
Ion Ratio Lower Upper
163 100
194 3.7 2.6 3.8
164 10.5 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 20.904 ng/ul
RT: 13.822 min Scan# 1842
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion:165 Resp: 65194
Ion Ratio Lower Upper
165 100
89 81.3 66.6 99.8
121 27.9 21.0 31.6





#50

Acenaphthylene

Concen: 19.923 ng/ul

RT: 13.945 min Scan# 1863

Delta R.T. -0.006 min

Lab File: BPCCC514.D

Acq: 19 Aug 2022 19:32

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:152 Resp: 441521

Ion Ratio Lower Upper

152 100

151 19.8 16.9 25.3

153 13.1 10.6 15.8



#51

3-Nitroaniline

Concen: 18.770 ng/ul

RT: 14.151 min Scan# 1898

Delta R.T. -0.006 min

Lab File: BPCCC514.D

Acq: 19 Aug 2022 19:32




Tgt Ion:138 Resp: 63762


Ion Ratio Lower Upper

138 100

108 12.5 8.6 13.0

92 134.2 104.2 156.2









#52
Acenaphthene
Concen: 19.477 ng/ul
RT: 14.292 min Scan# 1922
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:153 Resp: 289099
Ion Ratio Lower Upper
153 100
152 47.3 37.7 56.5
154 87.2 69.4 104.0



#53
2,4-Dinitrophenol
Concen: 16.292 ng/ul
RT: 14.369 min Scan# 1935
Delta R.T. -0.000 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion:184 Resp: 31099
Ion Ratio Lower Upper
184 100
63 104.4 82.5 123.7
154 89.0 66.6 100.0









#55
4-Nitrophenol
Concen: 18.067 ng/ul
RT: 14.475 min Scan# 1953
Delta R.T. -0.000 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion	Ratio	Lower	Upper
109	100		
139	91.7	76.0	114.0
65	119.6	98.8	148.2



#56
Dibenzofuran
Concen: 19.156 ng/ul
RT: 14.634 min Scan# 1980
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion	Ratio	Lower	Upper
168	100		
139	48.2	38.4	57.6









#57
2,4-Dinitrotoluene
Concen: 20.485 ng/ul
RT: 14.616 min Scan# 1979
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:165 Resp: 95638
Ion Ratio Lower Upper
165 100
63 70.9 60.2 90.2
182 2.6 1.8 2.8



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.638 ng/ul
RT: 14.863 min Scan# 2019
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion:232 Resp: 69919
Ion Ratio Lower Upper
232 100
131 72.3 55.4 83.0
130 4.1 3.6 5.4
166 38.3 31.0 46.4





#59

Diethylphthalate

Concen: 19.315 ng/ul

RT: 15.081 min Scan# 2057

Delta R.T. -0.006 min

Lab File: BPCCC514.D

Acq: 19 Aug 2022 19:32

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:149 Resp: 366979

Ion Ratio Lower Upper

149 100

177 19.8 15.7 23.5

150 12.2 10.1 15.1



#61

Fluorene

Concen: 18.937 ng/ul

RT: 15.286 min Scan# 2091

Delta R.T. -0.006 min

Lab File: BPCCC514.D

Acq: 19 Aug 2022 19:32




Tgt Ion:166 Resp: 317106


Ion Ratio Lower Upper

166 100

165 97.0 78.2 117.2

167 13.4 10.3 15.5









#62
4-Chlorophenyl-phenylether
Concen: 18.907 ng/ul
RT: 15.292 min Scan# 2092
Delta R.T. -0.000 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:204 Resp: 143028
Ion Ratio Lower Upper
204 100
206 32.9 25.4 38.0
141 81.7 66.5 99.7



#63
4-Nitroaniline
Concen: 17.990 ng/ul
RT: 15.328 min Scan# 2098
Delta R.T. -0.000 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion:138 Resp: 59031
Ion Ratio Lower Upper
138 100
92 58.0 49.5 74.3
108 116.0 93.7 140.5









#66
4,6-Dinitro-2-methylphenol
Concen: 17.791 ng/ul
RT: 15.386 min Scan# 2108
Delta R.T. -0.000 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:198 Resp: 47163
Ion Ratio Lower Upper
198 100
182 5.9 3.9 5.9#
77 38.8 30.1 45.1



#67
N-Nitrosodiphenylamine
Concen: 19.415 ng/ul
RT: 15.510 min Scan# 2129
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion:169 Resp: 270155
Ion Ratio Lower Upper
169 100
168 68.6 54.6 82.0
167 36.8 30.2 45.4









#68
4-Bromophenyl-phenylether
Concen: 19.745 ng/ul
RT: 16.192 min Scan# 21
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:248 Resp: 83594
Ion Ratio Lower Upper
248 100
250 98.2 82.8 124.2
141 114.3 90.8 136.2



#69
Hexachlorobenzene
Concen: 19.837 ng/ul
RT: 16.292 min Scan# 2262
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion:284 Resp: 96792
Ion Ratio Lower Upper
284 100
142 52.5 43.5 65.3
249 33.7 28.4 42.6









#70
 Atrazine
 Concen: 19.486 ng/ul
 RT: 16.480 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BPCCC514.D
 Acq: 19 Aug 2022 19:32




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion:200 Resp: 101185
 Ion Ratio Lower Upper
 200 100
 173 27.0 22.0 33.0
 215 41.7 36.3 54.5



#71
 Pentachlorophenol
 Concen: 19.662 ng/ul
 RT: 16.645 min Scan# 2322
 Delta R.T. -0.006 min
 Lab File: BPCCC514.D
 Acq: 19 Aug 2022 19:32

Tgt Ion:266 Resp: 59487
 Ion Ratio Lower Upper
 266 100
 264 58.2 50.6 76.0
 268 62.9 51.3 76.9









#72
Phenanthrene
Concen: 19.246 ng/ul
RT: 17.039 min Scan# 21
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:178 Resp: 494060
Ion Ratio Lower Upper
178 100
179 15.9 12.6 18.8
176 19.2 15.8 23.8



#74
Anthracene
Concen: 19.549 ng/ul
RT: 17.133 min Scan# 2405
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion:178 Resp: 506284
Ion Ratio Lower Upper
178 100
179 15.0 12.2 18.2
176 18.7 15.4 23.0









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.850 ng/uL
RT: 13.010 min Scan# 1704
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:216 Resp: 127118
Ion Ratio Lower Upper
216 100
214 78.3 61.8 92.6
218 47.0 37.4 56.0



#76
Pentachlorobenzene
Concen: 20.074 ng/uL
RT: 14.545 min Scan# 1965
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32


Tgt Ion:250 Resp: 116130
Ion Ratio Lower Upper
250 100
248 63.4 51.7 77.5
252 62.4 50.0 75.0








#77
 Carbazole
 Concen: 19.079 ng/ul
 RT: 17.416 min Scan# 2453
 Delta R.T. -0.006 min
 Lab File: BPCCC514.D
 Acq: 19 Aug 2022 19:32

Instrument :
 BNA_P
 ClientSampleId :







Tgt Ion:167 Resp: 464763
 Ion Ratio Lower Upper
 167 100
 166 21.9 17.5 26.3
 139 16.3 12.1 18.1



#78
 Di-n-butylphthalate
 Concen: 20.873 ng/ul
 RT: 17.986 min Scan# 2550
 Delta R.T. -0.006 min
 Lab File: BPCCC514.D
 Acq: 19 Aug 2022 19:32

Tgt Ion:149 Resp: 626161
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.4 11.2
 104 6.6 5.5 8.3





#80

Fluoranthene

Concen: 19.812 ng/ul

RT: 19.033 min Scan# 21

Delta R.T. -0.006 min

Lab File: BPCCC514.D

Acq: 19 Aug 2022 19:32

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:202 Resp: 560704

Ion Ratio Lower Upper

202 100

101 12.2 9.5 14.3

100 9.7 7.9 11.9



#82

Pyrene

Concen: 19.588 ng/ul

RT: 19.392 min Scan# 2789

Delta R.T. -0.000 min

Lab File: BPCCC514.D

Acq: 19 Aug 2022 19:32




Tgt Ion:202 Resp: 582482


Ion Ratio Lower Upper

202 100

101 13.4 11.4 17.2

100 11.0 9.4 14.0









#83
Butylbenzylphthalate
Concen: 22.407 ng/ul
RT: 20.286 min Scan# 2941
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:149 Resp: 298898
Ion Ratio Lower Upper
149 100
91 82.1 68.5 102.7
206 16.5 12.6 18.8



#84
3,3'-Dichlorobenzidine
Concen: 18.108 ng/ul
RT: 21.057 min Scan# 3072
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion:252 Resp: 155922
Ion Ratio Lower Upper
252 100
254 64.3 51.2 76.8
126 14.0 12.2 18.2





#85

Benzo(a)anthracene

Concen: 19.643 ng/ul

RT: 21.116 min Scan# 3082

Delta R.T. -0.000 min

Lab File: BPCCC514.D

Acq: 19 Aug 2022 19:32

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 544343

Ion Ratio Lower Upper

228 100

229 19.4 16.0 24.0

226 27.6 22.6 33.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 22.412 ng/ul

RT: 21.057 min Scan# 3072

Delta R.T. -0.006 min

Lab File: BPCCC514.D

Acq: 19 Aug 2022 19:32




Tgt Ion:149 Resp: 442422


Ion Ratio Lower Upper

149 100

167 24.2 19.8 29.8

279 3.2 2.7 4.1





#87

Chrysene

Concen: 19.637 ng/ul

RT: 21.163 min Scan# 3093

Delta R.T. -0.006 min

Lab File: BPCCC514.D

Acq: 19 Aug 2022 19:32

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 534556

Ion Ratio Lower Upper

228 100

226 30.5 23.9 35.9

229 19.6 16.0 24.0



#89

Di-n-octyl phthalate

Concen: 20.028 ng/ul




RT: 21.951 min Scan# 3224


Delta R.T. -0.006 min

Lab File: BPCCC514.D

Acq: 19 Aug 2022 19:32

Tgt Ion:149 Resp: 750317









#90
Benzo(b)fluoranthene
Concen: 19.340 ng/ul
RT: 22.721 min Scan# 3355
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:252 Resp: 560665
Ion Ratio Lower Upper
252 100
253 22.5 17.5 26.3
125 11.1 9.0 13.4



#91
Benzo(k)fluoranthene
Concen: 19.193 ng/ul
RT: 22.768 min Scan# 3363
Delta R.T. -0.006 min
Lab File: BPCCC514.D
Acq: 19 Aug 2022 19:32

Tgt Ion:252 Resp: 530961
Ion Ratio Lower Upper
252 100
253 22.0 17.9 26.9
125 11.7 9.8 14.8





#93

Benzo(a)pyrene

Concen: 19.300 ng/ul

RT: 23.315 min Scan# 3461

Delta R.T. -0.012 min

Lab File: BPCCC514.D

Acq: 19 Aug 2022 19:32

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:252 Resp: 546243

Ion Ratio Lower Upper

252 100

253 22.1 16.8 25.2

125 12.5 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.726 ng/ul

RT: 25.774 min Scan# 3874

Delta R.T. -0.012 min

Lab File: BPCCC514.D

Acq: 19 Aug 2022 19:32




Tgt Ion:276 Resp: 622726


Ion Ratio Lower Upper

276 100

138 25.7 20.2 30.4

277 24.9 20.1 30.1





#95

Dibenzo(a,h)anthracene

Concen: 19.673 ng/ul

RT: 25.798 min Scan# 31

Delta R.T. -0.006 min

Lab File: BPCCC514.D

Acq: 19 Aug 2022 19:32

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:278 Resp: 534749

Ion Ratio Lower Upper

278 100

139 18.6 15.4 23.0

279 23.9 19.3 28.9



#96

Benzo(g,h,i)perylene

Concen: 19.571 ng/ul

RT: 26.498 min Scan# 3997

Delta R.T. -0.012 min

Lab File: BPCCC514.D

Acq: 19 Aug 2022 19:32

Tgt Ion:276 Resp: 525535

Ion Ratio Lower Upper

276 100

138 23.8 19.4 29.2

277 24.0 19.1 28.7

