

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081925\
 Data File : BP025487.D
 Acq On : 19 Aug 2025 23:46
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 20 02:26:31 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 14 18:14:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.472	0.461	2.3	105	0.00
3	Pyridine	1.291	1.256	2.7	104	0.00
4	n-Nitrosodimethylamine	0.532	0.578	-8.6	114	0.00
5 S	2-Fluorophenol	1.204	1.204	0.0	105	0.00
6	Aniline	2.080	2.116	-1.7	106	0.00
7 S	Phenol-d6	1.544	1.570	-1.7	105	0.00
8	2-Chlorophenol	1.359	1.389	-2.2	107	0.00
9	Benzaldehyde	1.082	0.876	19.0	65	0.00
10 C	Phenol	1.620	1.661	-2.5	106	0.00
11	bis(2-Chloroethyl)ether	1.263	1.268	-0.4	105	0.00
12	1,3-Dichlorobenzene	1.499	1.501	-0.1	105	0.00
13 C	1,4-Dichlorobenzene	1.532	1.525	0.5	105	0.00
14	1,2-Dichlorobenzene	1.483	1.477	0.4	104	0.00
15	Benzyl Alcohol	1.227	1.247	-1.6	105	0.00
16	2,2'-oxybis(1-Chloropropane	1.692	1.784	-5.4	111	0.00
17	2-Methylphenol	1.125	1.152	-2.4	105	0.00
18	Hexachloroethane	0.563	0.566	-0.5	105	0.00
19 P	n-Nitroso-di-n-propylamine	1.023	1.046	-2.2	106	0.00
20	3+4-Methylphenols	1.560	1.571	-0.7	104	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
22	Acetophenone	0.502	0.503	-0.2	106	0.00
23 S	Nitrobenzene-d5	0.395	0.401	-1.5	107	0.00
24	Nitrobenzene	0.353	0.362	-2.5	107	0.00
25	Isophorone	0.700	0.705	-0.7	107	0.00
26 C	2-Nitrophenol	0.181	0.185	-2.2	105	0.00
27	2,4-Dimethylphenol	0.312	0.315	-1.0	105	0.00
28	bis(2-Chloroethoxy)methane	0.423	0.427	-0.9	107	0.00
29 C	2,4-Dichlorophenol	0.310	0.313	-1.0	105	0.00
30	1,2,4-Trichlorobenzene	0.340	0.331	2.6	104	0.00
31	Naphthalene	1.040	1.035	0.5	106	0.00
32	Benzoic acid	0.230	0.300	-30.4#	135	0.02
33	4-Chloroaniline	0.459	0.467	-1.7	106	0.00
34 C	Hexachlorobutadiene	0.211	0.207	1.9	103	0.00
35	Caprolactam	0.114	0.117	-2.6	109	0.00
36 C	4-Chloro-3-methylphenol	0.355	0.367	-3.4	108	0.00
37	2-Methylnaphthalene	0.676	0.673	0.4	106	0.00
38	1-Methylnaphthalene	0.719	0.710	1.3	106	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
40	1,2,4,5-Tetrachlorobenzene	0.578	0.572	1.0	104	0.00
41 P	Hexachlorocyclopentadiene	0.349	0.402	-15.2	118	0.00
42 S	2,4,6-Tribromophenol	0.269	0.278	-3.3	108	0.00
43 C	2,4,6-Trichlorophenol	0.390	0.400	-2.6	107	0.00
44	2,4,5-Trichlorophenol	0.427	0.443	-3.7	107	0.00
45 S	2-Fluorobiphenyl	1.463	1.408	3.8	103	0.00
46	1,1'-Biphenyl	1.453	1.462	-0.6	106	0.00
47	2-Chloronaphthalene	1.131	1.130	0.1	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081925\
 Data File : BP025487.D
 Acq On : 19 Aug 2025 23:46
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 20 02:26:31 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 14 18:14:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.329	0.359	-9.1	111	0.00
49	Acenaphthylene	1.871	1.889	-1.0	107	0.00
50	Dimethylphthalate	1.465	1.503	-2.6	110	0.00
51	2,6-Dinitrotoluene	0.309	0.327	-5.8	108	0.00
52 C	Acenaphthene	1.098	1.071	2.5	105	0.00
53	3-Nitroaniline	0.347	0.367	-5.8	110	0.00
54 P	2,4-Dinitrophenol	0.165	0.332	-101.2#	207#	0.00
55	Dibenzofuran	1.736	1.710	1.5	104	0.00
56 P	4-Nitrophenol	0.285	0.301	-5.6	110	0.01
57	2,4-Dinitrotoluene	0.440	0.475	-8.0	111	0.00
58	Fluorene	1.370	1.332	2.8	104	0.01
59	2,3,4,6-Tetrachlorophenol	0.377	0.388	-2.9	108	0.00
60	Diethylphthalate	1.488	1.537	-3.3	110	0.00
61	4-Chlorophenyl-phenylether	0.681	0.664	2.5	104	0.00
62	4-Nitroaniline	0.356	0.368	-3.4	107	0.01
63	Azobenzene	1.274	1.376	-8.0	114	0.01
64 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
65	4,6-Dinitro-2-methylphenol	0.125	0.178	-42.4#	148	0.00
66 c	n-Nitrosodiphenylamine	0.610	0.625	-2.5	108	0.00
67	4-Bromophenyl-phenylether	0.220	0.222	-0.9	107	0.01
68	Hexachlorobenzene	0.255	0.255	0.0	108	0.01
69	Atrazine	0.228	0.232	-1.8	109	0.00
70 C	Pentachlorophenol	0.161	0.170	-5.6	111	0.00
71	Phenanthrene	1.099	1.084	1.4	107	0.00
72	Anthracene	1.122	1.144	-2.0	109	0.00
73	Carbazole	1.057	1.052	0.5	107	0.00
74	Di-n-butylphthalate	1.300	1.333	-2.5	110	0.00
75 C	Fluoranthene	1.311	1.283	2.1	107	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	104	0.00
77	Benzidine	0.683	0.661	3.2	71	0.00
78	Pyrene	1.300	1.285	1.2	105	0.00
79 S	Terphenyl-d14	1.093	1.061	2.9	105	0.01
80	Butylbenzylphthalate	0.589	0.605	-2.7	108	0.00
81	Benzo(a)anthracene	1.319	1.302	1.3	106	0.00
82	3,3'-Dichlorobenzidine	0.497	0.485	2.4	103	0.00
83	Chrysene	1.235	1.231	0.3	108	0.00
84	Bis(2-ethylhexyl)phthalate	0.867	0.890	-2.7	111	-0.01
85 c	Di-n-octyl phthalate	1.492	1.525	-2.2	109	-0.02
86 I	Perylene-d12	1.000	1.000	0.0	103	-0.02
87	Indeno(1,2,3-cd)pyrene	1.467	1.434	2.2	105	-0.02
88	Benzo(b)fluoranthene	1.179	1.172	0.6	107	0.00
89	Benzo(k)fluoranthene	1.180	1.163	1.4	106	0.00
90 C	Benzo(a)pyrene	1.148	1.141	0.6	106	0.01
91	Dibenzo(a,h)anthracene	1.199	1.180	1.6	105	-0.01
92	Benzo(g,h,i)perylene	1.178	1.150	2.4	104	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081925\
Data File : BP025487.D
Acq On : 19 Aug 2025 23:46
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Aug 20 02:26:31 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 14 18:14:31 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0