




Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP082020\
Data File : BP003055.D
Acq On : 20 Aug 2020 08:36
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 20 16:19:22 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP073020.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Aug 17 14:07:35 2020
Response via : Initial Calibration



(70) Pentachlorophenol (C)




16.734min (+0.000) 55206.26ng

response 269906

Ion	Exp%	Act%
265.70	100	100
268.00	63.60	63.02
264.00	63.00	62.15
0.00	0.00	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP082020\
 Data File : BP003055.D
 Acq On : 20 Aug 2020 08:36
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 20 16:19:22 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP073020.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 17 14:07:35 2020
 Response via : Initial Calibration



(77) Benzidine

19.269min (-0.006) 71636.44ng

response 2077410

Ion	Exp%	Act%
184.00	100	100
185.00	14.40	14.69
183.00	11.70	12.00
0.00	0.00	0.00