

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082321\
 Data File : BP006824.D
 Acq On : 23 Aug 2021 10:27
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Aug 23 11:40:04 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 23 11:39:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.699	152	181929	20.000	ng	0.00
21) Naphthalene-d8	10.481	136	754572	20.000	ng	# 0.00
39) Acenaphthene-d10	14.346	164	426350	20.000	ng	0.00
64) Phenanthrene-d10	17.104	188	835799	20.000	ng	# 0.00
76) Chrysene-d12	21.216	240	798630	20.000	ng	# 0.00
86) Perylene-d12	23.533	264	854338	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.299	112	849811	71.746	ng	0.00
7) Phenol-d6	6.887	99	1150013	68.348	ng	0.00
23) Nitrobenzene-d5	8.858	82	1103995	67.532	ng	0.00
42) 2,4,6-Tribromophenol	15.845	330	347872	78.071	ng	0.00
45) 2-Fluorobiphenyl	12.963	172	2063549	72.413	ng	0.00
79) Terphenyl-d14	19.686	244	2955425	74.152	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.246	88	172089	33.371	ng	96
3) Pyridine	3.640	79	485246	32.016	ng	98
4) n-Nitrosodimethylamine	3.558	42	162182	28.351	ng	# 71
6) Aniline	7.034	93	618052	33.816	ng	97
8) 2-Chlorophenol	7.264	128	466027	36.363	ng	95
9) Benzaldehyde	6.852	77	323308	31.925	ng	98
10) Phenol	6.917	94	566146	33.556	ng	97
11) bis(2-Chloroethyl)ether	7.140	93	431317	33.668	ng	97
12) 1,3-Dichlorobenzene	7.581	146	488682	36.586	ng	98
13) 1,4-Dichlorobenzene	7.734	146	497933	36.731	ng	99
14) 1,2-Dichlorobenzene	8.046	146	477585	36.715	ng	97
15) Benzyl Alcohol	7.952	79	423262	33.544	ng	96
16) 2,2'-oxybis(1-Chloropr...	8.234	45	439828	28.851	ng	98
17) 2-Methylphenol	8.152	107	378666	35.544	ng	99
18) Hexachloroethane	8.758	117	195945	36.519	ng	90
19) n-Nitroso-di-n-propyla...	8.511	70	371159	32.880	ng	94
20) 3+4-Methylphenols	8.481	107	517547	35.691	ng	97
22) Acetophenone	8.522	105	685691	34.266	ng	# 98
24) Nitrobenzene	8.899	77	566967	33.363	ng	94
25) Isophorone	9.428	82	1015541	34.586	ng	96
26) 2-Nitrophenol	9.605	139	250090	40.439	ng	95
27) 2,4-Dimethylphenol	9.675	122	368519	35.584	ng	95
28) bis(2-Chloroethoxy)met...	9.911	93	525298	33.436	ng	99
29) 2,4-Dichlorophenol	10.140	162	395795	37.749	ng	96
30) 1,2,4-Trichlorobenzene	10.346	180	413185	36.674	ng	97
31) Naphthalene	10.534	128	1424153	35.124	ng	100
32) Benzoic acid	9.840	122	295419	36.874	ng	95
33) 4-Chloroaniline	10.658	127	577762	35.212	ng	97
34) Hexachlorobutadiene	10.810	225	252016	38.386	ng	99
35) Caprolactam	11.469	113	139712	36.954	ng	82
36) 4-Chloro-3-methylphenol	11.793	107	478601	35.927	ng	91
37) 2-Methylnaphthalene	12.152	142	972568	35.398	ng	99
38) 1-Methylnaphthalene	12.369	142	920130	35.240	ng	99
40) 1,2,4,5-Tetrachloroben...	12.522	216	418875	37.538	ng	# 97
41) Hexachlorocyclopentadiene	12.493	237	175546	29.661	ng	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082321\
 Data File : BP006824.D
 Acq On : 23 Aug 2021 10:27
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Aug 23 11:40:04 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 23 11:39:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) 2,4,6-Trichlorophenol	12.775	196	298256	39.144	ng	98
44) 2,4,5-Trichlorophenol	12.852	196	320242	37.970	ng #	91
46) 1,1'-Biphenyl	13.175	154	1153046	35.730	ng	98
47) 2-Chloronaphthalene	13.216	162	892483	35.913	ng	96
48) 2-Nitroaniline	13.434	65	306850	34.988	ng	93
49) Acenaphthylene	14.063	152	1439483	35.923	ng	99
50) Dimethylphthalate	13.816	163	1097924	35.377	ng	99
51) 2,6-Dinitrotoluene	13.940	165	251650	36.338	ng	87
52) Acenaphthene	14.410	154	856521	35.121	ng	96
53) 3-Nitroaniline	14.269	138	278044	36.217	ng	88
54) 2,4-Dinitrophenol	14.481	184	132337	36.543	ng #	87
55) Dibenzofuran	14.746	168	1338456	34.892	ng	95
56) 4-Nitrophenol	14.593	139	231945	34.794	ng	88
57) 2,4-Dinitrotoluene	14.728	165	348645	37.540	ng #	79
58) Fluorene	15.398	166	1070155	34.903	ng	99
59) 2,3,4,6-Tetrachlorophenol	14.981	232	264012	37.353	ng #	73
60) Diethylphthalate	15.187	149	1165261	35.748	ng	98
61) 4-Chlorophenyl-phenyle...	15.398	204	492431	35.465	ng	91
62) 4-Nitroaniline	15.434	138	288039	35.357	ng	93
63) Azobenzene	15.693	77	1270842	33.440	ng	93
65) 4,6-Dinitro-2-methylph...	15.498	198	191731	37.543	ng	81
66) n-Nitrosodiphenylamine	15.616	169	927699	35.313	ng	99
67) 4-Bromophenyl-phenylether	16.298	248	298293	37.196	ng #	89
68) Hexachlorobenzene	16.404	284	332562	36.441	ng #	90
69) Atrazine	16.587	200	305470	37.263	ng	96
70) Pentachlorophenol	16.757	266	211777	36.575	ng	99
71) Phenanthrene	17.145	178	1649297	34.984	ng	99
72) Anthracene	17.240	178	1639813	35.997	ng	100
73) Carbazole	17.522	167	1631047	35.102	ng	98
74) Di-n-butylphthalate	18.081	149	2026877	38.287	ng	99
75) Fluoranthene	19.128	202	1904699	35.996	ng	95
77) Benzidine	19.316	184	746973	37.378	ng	99
78) Pyrene	19.481	202	1961108	36.763	ng	99
80) Butylbenzylphthalate	20.369	149	964064	41.099	ng	90
81) Benzo(a)anthracene	21.198	228	1877953	35.981	ng	100
82) 3,3'-Dichlorobenzidine	21.139	252	521894	38.088	ng #	98
83) Chrysene	21.251	228	1794302	35.461	ng	99
84) Bis(2-ethylhexyl)phtha...	21.133	149	1397750	39.652	ng #	96
85) Di-n-octyl phthalate	22.039	149	2447169	42.120	ng	98
87) Indeno(1,2,3-cd)pyrene	25.939	276	2238645	36.137	ng #	91
88) Benzo(b)fluoranthene	22.833	252	1908672	34.375	ng #	97
89) Benzo(k)fluoranthene	22.880	252	1879202	35.250	ng #	97
90) Benzo(a)pyrene	23.433	252	1793386	36.010	ng #	95
91) Dibenzo(a,h)anthracene	25.951	278	1923784	35.917	ng #	94
92) Benzo(g,h,i)perylene	26.668	276	1855294	36.147	ng #	91





(#) = qualifier out of range (m) = manual integration (+) = signals summed



Scan 794 (7.757 min): BP006476.D\data.ms (-786) (#1)
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.699 min Scan# 784
 Delta R.T. -0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	152	Resp:	181929
Ion	Ratio	Lower	Upper
152	100		
150	154.2	125.6	188.4
115	65.2	48.5	72.7



Scan 37 (3.305 min): BP006476.D\data.ms (-32) (#2)
 1,4-Dioxane
 Concen: 33.371 ng
 RT: 3.246 min Scan# 27
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27

Tgt Ion:	88	Resp:	172089
Ion	Ratio	Lower	Upper
88	100		
58	60.4	46.7	70.1
43	22.3	20.3	30.5









Pyridine
Concen: 32.016 ng
RT: 3.640 min Scan# 94
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 485246
Ion Ratio Lower Upper
79 100
52 52.0 41.7 62.5
51 26.6 24.1 36.1



n-Nitrosodimethylamine
Concen: 28.351 ng
RT: 3.558 min Scan# 80
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Tgt Ion: 42 Resp: 162182
Ion Ratio Lower Upper
42 100
74 165.7 105.4 158.0#
44 18.6 4.2 6.4#









Scan 386 (5.358 min): BP006476.D\data.ms (-375) #5
 2-Fluorophenol
 Concen: 71.746 ng
 RT: 5.299 min Scan# 376
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	112	Resp:	849811
Ion	Ratio	Lower	Upper
112	100		
64	55.2	41.4	62.2
63	28.6	23.8	35.8



Scan 682 (7.099 min): BP006476.D\data.ms (-672) #6
 Aniline
 Concen: 33.816 ng
 RT: 7.034 min Scan# 671
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27

Tgt Ion:	93	Resp:	618052
Ion	Ratio	Lower	Upper
93	100		
66	39.7	29.7	44.5
65	19.3	15.3	22.9









Phenol-d6
 Concen: 68.348 ng
 RT: 6.887 min Scan# 646
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	99	Resp:	1150013
Ion	Ratio	Lower	Upper
99	100		
42	15.1	13.4	20.0
71	36.8	30.6	46.0



2-Chlorophenol
 Concen: 36.363 ng
 RT: 7.264 min Scan# 710
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27

Tgt Ion:	128	Resp:	466027
Ion	Ratio	Lower	Upper
128	100		
130	32.0	12.4	52.4
64	45.5	20.3	60.3









Benzaldehyde
 Concen: 31.925 ng
 RT: 6.852 min Scan# 640
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion: 77	Resp: 323308
Ion Ratio	Lower Upper
77 100	
105 97.5	76.7 116.7
106 92.4	76.2 116.2



Phenol
 Concen: 33.556 ng
 RT: 6.917 min Scan# 651
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27

Tgt Ion: 94	Resp: 566146
Ion Ratio	Lower Upper
94 100	
65 28.1	8.8 48.8
66 38.1	20.7 60.7









bis(2-Chloroethyl)ether
 Concen: 33.668 ng
 RT: 7.140 min Scan# 689
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Tgt Ion:	93	Resp:	431317
Ion	Ratio	Lower	Upper
93	100		
63	69.7	46.2	86.2
95	32.1	12.2	52.2




1,3-Dichlorobenzene
 Concen: 36.586 ng
 RT: 7.581 min Scan# 764
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27

Tgt Ion:	146	Resp:	488682
Ion	Ratio	Lower	Upper
146	100		
148	63.4	51.5	77.3
75	34.7	25.7	38.5






Abundance Scan 800 (7.793 min): BP006476.D\data.ms (-792) #13




1,4-Dichlorobenzene
Concen: 36.731 ng
RT: 7.734 min Scan# 790
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:	146	Resp:	497933
Ion	Ratio	Lower	Upper
146	100		
148	63.3	51.3	76.9
111	41.8	33.0	49.6







Abundance Scan 854 (8.110 min): BP006476.D\data.ms (-842) #14



1,2-Dichlorobenzene
Concen: 36.715 ng
RT: 8.046 min Scan# 843
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Tgt Ion:	146	Resp:	477585
Ion	Ratio	Lower	Upper
146	100		
148	64.8	50.3	75.5
111	45.3	33.7	50.5









Benzyl Alcohol
 Concen: 33.544 ng
 RT: 7.952 min Scan# 827
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Tgt Ion:	79	Resp:	423262
Ion	Ratio	Lower	Upper
79	100		
108	72.8	62.5	93.7
77	65.4	52.9	79.3



2,2'-oxybis(1-Chloropropane)
 Concen: 28.851 ng
 RT: 8.234 min Scan# 875
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27


Tgt Ion:	45	Resp:	439828
Ion	Ratio	Lower	Upper
45	100		
77	21.3	2.0	42.0
79	16.1	0.0	36.7






Abundance Scan 870 (8.204 min): BP006476.D\data.ms (-861) #17

2-Methylphenol
Concen: 35.544 ng
RT: 8.152 min Scan# 861
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040






Tgt Ion:	107	Resp:	378666
Ion	Ratio	Lower	Upper
107	100		
108	110.2	87.7	131.5
77	50.2	40.2	60.2
79	47.8	39.0	58.4




Abundance Scan 976 (8.828 min): BP006476.D\data.ms (-964) #18

Hexachloroethane
Concen: 36.519 ng
RT: 8.758 min Scan# 964
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Tgt Ion:	117	Resp:	195945
Ion	Ratio	Lower	Upper
117	100		
119	97.1	76.2	114.2
201	87.4	83.4	125.0









#19
 n-Nitroso-di-n-propylamine
 Concen: 32.880 ng
 RT: 8.511 min Scan# 922
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27




Instrument :
 BNA_P
 ClientSampleId :
 SSTCCCC040


Tgt Ion:	70	Resp:	371159
Ion	Ratio	Lower	Upper
70	100		
42	46.0	40.8	61.2
101	9.9	8.8	13.2
130	20.6	18.9	28.3



#20
 3+4-Methylphenols
 Concen: 35.691 ng
 RT: 8.481 min Scan# 917
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27

Tgt Ion:	107	Resp:	517547
Ion	Ratio	Lower	Upper
107	100		
108	87.7	71.0	111.0
77	35.3	16.6	56.6
79	30.6	12.7	52.7









Naphthalene-d8
 Concen: 20.000 ng
 RT: 10.481 min Scan# 1257
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	136	Resp:	754572
Ion	Ratio	Lower	Upper
136	100		
137	11.1	8.6	12.8
54	7.3	4.6	6.8#
68	6.2	3.6	5.4#



Acetophenone
 Concen: 34.266 ng
 RT: 8.522 min Scan# 924
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27

Tgt Ion:	105	Resp:	685691
Ion	Ratio	Lower	Upper
105	100		
71	5.6	1.9	2.9#
51	22.1	18.4	27.6
120	21.7	17.9	26.9









Nitrobenzene-d5
 Concen: 67.532 ng
 RT: 8.858 min Scan# 981
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Tgt Ion:	82	Resp:	1103995
Ion	Ratio	Lower	Upper
82	100		
128	41.3	35.7	53.5
54	41.4	35.0	52.4




Nitrobenzene
 Concen: 33.363 ng
 RT: 8.899 min Scan# 988
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27

Tgt Ion:	77	Resp:	566967
Ion	Ratio	Lower	Upper
77	100		
123	42.8	38.2	57.2
65	13.2	11.4	17.2






Abundance Scan 1088 (9.487 min): BP006476.D\data.ms (-10725)




Isophorone
Concen: 34.586 ng
RT: 9.428 min Scan# 1078
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:	82	Resp:	1015541
Ion	Ratio	Lower	Upper
82	100		
95	7.5	6.5	9.7
138	16.5	15.0	22.4






Abundance Scan 1118 (9.663 min): BP006476.D\data.ms (-11026)




2-Nitrophenol
Concen: 40.439 ng
RT: 9.605 min Scan# 1108
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Tgt Ion:	139	Resp:	250090
Ion	Ratio	Lower	Upper
139	100		
109	33.2	28.8	43.2
65	54.8	41.0	61.6






Abundance Scan 1129 (9.728 min): BP006476.D\data.ms (-1129) (-)




2,4-Dimethylphenol
Concen: 35.584 ng
RT: 9.675 min Scan# 1120
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:	122	Resp:	368519
Ion	Ratio	Lower	Upper
122	100		
107	125.3	94.5	141.7
121	61.8	47.4	71.2







Abundance Scan 1170 (9.969 min): BP006476.D\data.ms (-1169) (-)



bis(2-Chloroethoxy)methane
Concen: 33.436 ng
RT: 9.911 min Scan# 1160
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Tgt Ion:	93	Resp:	525298
Ion	Ratio	Lower	Upper
93	100		
95	32.5	25.9	38.9
123	10.9	10.2	15.2









Scan 1208 (10.193 min): BP006476.D\data.ms (-1159) (-)
 2,4-Dichlorophenol
 Concen: 37.749 ng
 RT: 10.140 min Scan# 1199
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	162	Resp:	395795
Ion	Ratio	Lower	Upper
162	100		
164	64.7	44.9	84.9
98	40.4	14.3	54.3



Scan 1245 (10.410 min): BP006476.D\data.ms (-1234) (-)
 1,2,4-Trichlorobenzene
 Concen: 36.674 ng
 RT: 10.346 min Scan# 1234
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27

Tgt Ion:	180	Resp:	413185
Ion	Ratio	Lower	Upper
180	100		
182	95.1	77.9	116.9
145	34.0	24.7	37.1









#30 (-)
Naphthalene
Concen: 35.124 ng
RT: 10.534 min Scan# 1266
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:	128	Resp:	1424153
Ion	Ratio	Lower	Upper
128	100		
129	11.0	9.0	13.6
127	13.4	10.7	16.1




#32 (-)
Benzoic acid
Concen: 36.874 ng
RT: 9.840 min Scan# 1148
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Tgt Ion:	122	Resp:	295419
Ion	Ratio	Lower	Upper
122	100		
105	127.5	103.2	143.2
77	101.6	76.2	116.2






Abundance Scan 1296 (10.710 min): BP006476.D\data.ms (-1283)




4-Chloroaniline
Concen: 35.212 ng
RT: 10.658 min Scan# 1287
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion	Ratio	Lower	Upper
127	100		
129	32.6	26.1	39.1
65	31.7	22.5	33.7
92	20.2	15.0	22.6






Abundance Scan 1325 (10.881 min): BP006476.D\data.ms (-1304)



Hexachlorobutadiene
Concen: 38.386 ng
RT: 10.810 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27


Tgt Ion	Ratio	Lower	Upper
225	100		
223	64.1	51.0	76.6
227	65.6	51.4	77.2





Abundance Scan 1431 (11.504 min): BP006476.D\data.ms (-1435) (-)

Caprolactam
Concen: 36.954 ng
RT: 11.469 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27


Instrument :
BNA_P
ClientSampleId :
SSTDCCC040



Tgt Ion:	113	Resp:	139712
Ion	Ratio	Lower	Upper
113	100		
55	138.9	98.9	138.9
56	108.8	71.7	111.7




Abundance Scan 1425 (11.469 min): BP006824.D\data.ms (-1374) (-)





Abundance Scan 1487 (11.834 min): BP006476.D\data.ms (-1436) (-)


4-Chloro-3-methylphenol
Concen: 35.927 ng
RT: 11.793 min Scan# 1480
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27




Tgt Ion:	107	Resp:	478601
Ion	Ratio	Lower	Upper
107	100		
144	24.1	21.1	31.7
142	72.8	66.0	99.0



Abundance Scan 1480 (11.793 min): BP006824.D\data.ms (-1429) (-)






Abundance Scan 1550 (12.204 min): BP006476.D\data.ms (-1537) (-)




2-Methylnaphthalene
Concen: 35.398 ng
RT: 12.152 min Scan# 1541
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion	Ratio	Lower	Upper
142	100		
141	91.0	71.8	107.6
115	38.9	31.6	47.4






Abundance Scan 1588 (12.428 min): BP006476.D\data.ms (-1538) (-)




1-Methylnaphthalene
Concen: 35.240 ng
RT: 12.369 min Scan# 1578
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Tgt Ion	Ratio	Lower	Upper
142	100		
141	92.5	73.3	109.9
116	4.2	3.5	5.3





Abundance Scan 1922 (14.392 min): BP006476.D\data.ms (-1939) (-)




Acenaphthene-d10
Concen: 20.000 ng
RT: 14.346 min Scan# 1914
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	164	Resp:	426350
Ion Ratio	Lower	Upper	
164	100		
162	103.0	80.2	120.2
160	46.7	34.5	51.7



Abundance Scan 1914 (14.346 min): BP006824.D\data.ms (-1863) (-)





Abundance Scan 1613 (12.575 min): BP006476.D\data.ms (-1604) (-)




1,2,4,5-Tetrachlorobenzene
Concen: 37.538 ng
RT: 12.522 min Scan# 1604
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27


Tgt Ion:	216	Resp:	418875
Ion Ratio	Lower	Upper	
216	100		
214	78.8	61.9	92.9
179	23.1	16.8	25.2
108	22.6	14.9	22.3#



Abundance Scan 1604 (12.522 min): BP006824.D\data.ms (-1553) (-)





Abundance Scan 1609 (12.551 min): BP006476.D\data.ms (-1594) (-)




Hexachlorocyclopentadiene
Concen: 29.661 ng
RT: 12.493 min Scan# 1599
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
237	100		
235	63.7	44.5	84.5
272	12.7	0.0	32.5



Abundance Scan 1599 (12.493 min): BP006824.D\data.ms (-1548) (-)





Abundance Scan 2175 (15.881 min): BP006476.D\data.ms (-2162) (-)




2,4,6-Tribromophenol
Concen: 78.071 ng
RT: 15.845 min Scan# 2169
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27


Tgt Ion	Ratio	Lower	Upper
330	100		
332	96.3	77.8	116.8
141	46.1	25.4	38.2#



Abundance Scan 2169 (15.845 min): BP006824.D\data.ms (-2118) (-)



Abundance Scan 1654 (12.816 min): BP006476.D\data.ms (-1643) (-)





2,4,6-Trichlorophenol
Concen: 39.144 ng
RT: 12.775 min Scan# 1647
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27


Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:	196	Resp:	298256
Ion	Ratio	Lower	Upper
196	100		
198	96.9	79.5	119.3
200	31.3	25.0	37.6


Abundance Scan 1647 (12.775 min): BP006824.D\data.ms



Abundance Scan 1647 (12.775 min): BP006824.D\data.ms (-1596) (-)





Abundance Scan 1666 (12.887 min): BP006476.D\data.ms (-1604) (-)




2,4,5-Trichlorophenol
Concen: 37.970 ng
RT: 12.852 min Scan# 1660
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Tgt Ion:	196	Resp:	320242
Ion	Ratio	Lower	Upper
196	100		
198	96.2	74.7	112.1
97	59.4	37.4	56.0#
132	33.0	20.0	30.0#


Abundance Scan 1660 (12.852 min): BP006824.D\data.ms



Abundance Scan 1660 (12.852 min): BP006824.D\data.ms (-1609) (-)





Abundance Scan 1688 (13.016 min): BP006476.D\data.ms (-1674) (-)




2-Fluorobiphenyl
Concen: 72.413 ng
RT: 12.963 min Scan# 1679
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	172	Resp:	2063549
Ion	Ratio	Lower	Upper
172	100		
171	35.6	27.9	41.9
170	23.5	18.6	27.8



Abundance Scan 1679 (12.963 min): BP006824.D\data.ms (-1628) (-)





Abundance Scan 1724 (13.228 min): BP006476.D\data.ms (-1716) (-)




1,1'-Biphenyl
Concen: 35.730 ng
RT: 13.175 min Scan# 1715
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27


Tgt Ion:	154	Resp:	1153046
Ion	Ratio	Lower	Upper
154	100		
153	40.3	20.4	60.4
76	14.4	0.0	30.8



Abundance Scan 1715 (13.175 min): BP006824.D\data.ms (-1664) (-)





Abundance Scan 1730 (13.263 min): BP006476.D\data.ms (-1718) (-)




2-Chloronaphthalene
Concen: 35.913 ng
RT: 13.216 min Scan# 1722
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	162	Resp:	892483
Ion	Ratio	Lower	Upper
162	100		
127	43.7	31.4	47.0
164	32.6	26.6	39.8



Abundance Scan 1722 (13.216 min): BP006824.D\data.ms (-1671) (-)





Abundance Scan 1766 (13.475 min): BP006476.D\data.ms (-1758) (-)





2-Nitroaniline
Concen: 34.988 ng
RT: 13.434 min Scan# 1759
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Tgt Ion:	65	Resp:	306850
Ion	Ratio	Lower	Upper
65	100		
92	71.1	57.7	86.5
138	103.4	92.1	138.1



Abundance Scan 1759 (13.434 min): BP006824.D\data.ms (-1708) (-)









Acenaphthylene
Concen: 35.923 ng
RT: 14.063 min Scan# 1866
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:	152	Resp:	1439483
Ion	Ratio	Lower	Upper
152	100		
151	20.5	16.2	24.2
153	13.0	10.6	16.0



Dimethylphthalate
Concen: 35.377 ng
RT: 13.816 min Scan# 1824
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27


Tgt Ion:	163	Resp:	1097924
Ion	Ratio	Lower	Upper
163	100		
194	3.7	3.4	5.2
164	10.1	8.2	12.2






Abundance Scan 1851 (13.975 min): BP006476.D\data.ms (-1851) (-)

2,6-Dinitrotoluene
Concen: 36.338 ng
RT: 13.940 min Scan# 1845
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion:	165	Resp:	251650
Ion	Ratio	Lower	Upper
165	100		
63	60.3	41.3	61.9
89	65.4	44.8	67.2







Abundance Scan 1933 (14.457 min): BP006476.D\data.ms (-1933) (-)

Acenaphthene
Concen: 35.121 ng
RT: 14.410 min Scan# 1925
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27



Tgt Ion:	154	Resp:	856521
Ion	Ratio	Lower	Upper
154	100		
153	116.7	89.4	134.2
152	55.0	43.0	64.4









3-Nitroaniline
 Concen: 36.217 ng
 RT: 14.269 min Scan# 1901
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Tgt Ion:	138	Resp:	278044
Ion	Ratio	Lower	Upper
138	100		
108	10.9	8.4	12.6
92	123.8	87.8	131.8




2,4-Dinitrophenol
 Concen: 36.543 ng
 RT: 14.481 min Scan# 1937
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27

Tgt Ion:	184	Resp:	132337
Ion	Ratio	Lower	Upper
184	100		
63	73.5	47.0	70.6#
154	70.3	52.3	78.5






Abundance Scan 1990 (14.792 min): BP006476.D\data.ms (-1955) (-)




Dibenzofuran
Concen: 34.892 ng
RT: 14.746 min Scan# 1982
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:168 Resp: 1338456
Ion Ratio Lower Upper
168 100
139 42.6 31.0 46.6
169 13.5 10.8 16.2






Abundance Scan 1959 (14.610 min): BP006476.D\data.ms (-1956) (-)



4-Nitrophenol
Concen: 34.794 ng
RT: 14.593 min Scan# 1956
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27


Tgt Ion:139 Resp: 231945
Ion Ratio Lower Upper
139 100
109 83.8 72.3 112.3
65 106.6 72.5 112.5






Abundance Scan 1985 (14.763 min): BP006476.D\data.ms (-1957) (-)

2,4-Dinitrotoluene
Concen: 37.540 ng
RT: 14.728 min Scan# 1979
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion	Ratio	Lower	Upper
165	100		
63	52.7	32.3	48.5#
89	87.9	55.9	83.9#
182	2.7	2.2	3.4






Abundance Scan 2100 (15.439 min): BP006476.D\data.ms (-2058) (-)

Fluorene
Concen: 34.903 ng
RT: 15.398 min Scan# 2093
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27




Tgt Ion	Ratio	Lower	Upper
166	100		
165	98.8	78.0	117.0
167	13.5	10.5	15.7






Abundance Scan 2028 (15.016 min): BP006476.D\data.ms (-2059) (-)

2,3,4,6-Tetrachlorophenol
Concen: 37.353 ng
RT: 14.981 min Scan# 2022
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion	Ratio	Lower	Upper
232	100		
131	58.5	32.2	48.4#
130	3.1	2.0	3.0#
166	36.7	19.4	29.2#






Abundance Scan 2064 (15.228 min): BP006476.D\data.ms (-2060) (-)

Diethylphthalate
Concen: 35.748 ng
RT: 15.187 min Scan# 2057
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27




Tgt Ion	Ratio	Lower	Upper
149	100		
177	20.5	17.5	26.3
150	12.4	9.9	14.9






Abundance Scan 2100 (15.439 min): BP006476.D\data.ms (-2050) (-)

4-Chlorophenyl-phenylether
Concen: 35.465 ng
RT: 15.398 min Scan# 2093
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040




Tgt Ion	Ratio	Lower	Upper
204	100		
206	33.2	26.6	40.0
141	71.4	48.6	73.0






Abundance Scan 2105 (15.469 min): BP006476.D\data.ms (-2052) (-)


4-Nitroaniline
Concen: 35.357 ng
RT: 15.434 min Scan# 2099
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27



Tgt Ion	Ratio	Lower	Upper
138	100		
92	57.5	31.2	71.2
108	92.5	78.6	118.6






Abundance Scan 2150 (15.733 min): BP006476.D\data.ms (-2150) (-)




Azobenzene
Concen: 33.440 ng
RT: 15.693 min Scan# 2143
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 1270842
Ion Ratio Lower Upper
77 100
182 20.9 6.0 46.0
105 19.4 0.0 38.8
51 24.3 8.8 48.8






Abundance Scan 2389 (17.139 min): BP006476.D\data.ms (-2389) (-)




Phenanthrene-d10
Concen: 20.000 ng
RT: 17.104 min Scan# 2383
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Tgt Ion: 188 Resp: 835799
Ion Ratio Lower Upper
188 100
94 10.9 5.6 8.4#
80 12.4 6.3 9.5#





Abundance Scan 2114 (15.522 min): BP006476.D\data.ms (-2165) (-)




4,6-Dinitro-2-methylphenol
Concen: 37.543 ng
RT: 15.498 min Scan# 2110
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion	Ratio	Lower	Upper
198	100		
51	43.9	14.0	54.0
105	52.5	19.5	59.5



Abundance Scan 2110 (15.498 min): BP006824.D\data.ms (-2059) (-)





Abundance Scan 2137 (15.657 min): BP006476.D\data.ms (-2166) (-)




n-Nitrosodiphenylamine
Concen: 35.313 ng
RT: 15.616 min Scan# 2130
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27


Tgt Ion	Ratio	Lower	Upper
169	100		
168	66.8	54.4	81.6
167	36.4	29.0	43.6



Abundance Scan 2130 (15.616 min): BP006824.D\data.ms (-2079) (-)






Abundance Scan 2253 (16.339 min): BP006476.D\data.ms (-2253) (-)




4-Bromophenyl-phenylether
Concen: 37.196 ng
RT: 16.298 min Scan# 2246
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion	Ratio	Lower	Upper
248	100		
250	97.8	79.8	119.8
141	91.1	57.8	86.6#






Abundance Scan 2270 (16.439 min): BP006476.D\data.ms (-2268) (-)




Hexachlorobenzene
Concen: 36.441 ng
RT: 16.404 min Scan# 2264
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Tgt Ion	Ratio	Lower	Upper
284	100		
142	42.5	26.9	40.3#
249	33.3	24.5	36.7



Abundance Scan 2300 (16.616 min): BP006476.D\data.ms (-2259) (-)





Atrazine
Concen: 37.263 ng
RT: 16.587 min Scan# 2295
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27


Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion	Ratio	Lower	Upper
200	100		
173	26.7	4.1	44.1
215	45.0	27.7	67.7


Abundance Scan 2295 (16.587 min): BP006824.D\data.ms



Abundance Scan 2295 (16.587 min): BP006824.D\data.ms (-2244) (-)





Abundance Scan 2329 (16.786 min): BP006476.D\data.ms (-2378) (-)




Pentachlorophenol
Concen: 36.575 ng
RT: 16.757 min Scan# 2324
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Tgt Ion	Ratio	Lower	Upper
266	100		
268	65.5	51.6	77.4
264	62.7	50.3	75.5


Abundance Scan 2324 (16.757 min): BP006824.D\data.ms



Abundance Scan 2324 (16.757 min): BP006824.D\data.ms (-2273) (-)





Abundance Scan 2397 (17.186 min): BP006476.D\data.ms (-2397) (-)




Phenanthrene
Concen: 34.984 ng
RT: 17.145 min Scan# 2390
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:178 Resp: 1649297
Ion Ratio Lower Upper
178 100
176 19.3 15.1 22.7
179 15.3 12.4 18.6



Abundance Scan 2390 (17.145 min): BP006824.D\data.ms (-2339) (-)





Abundance Scan 2412 (17.275 min): BP006476.D\data.ms (-2406) (-)




Anthracene
Concen: 35.997 ng
RT: 17.240 min Scan# 2406
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27


Tgt Ion:178 Resp: 1639813
Ion Ratio Lower Upper
178 100
176 18.6 14.9 22.3
179 15.0 12.2 18.2



Abundance Scan 2406 (17.240 min): BP006824.D\data.ms (-2355) (-)






Abundance Scan 2459 (17.551 min): BP006476.D\data.ms (-2403) (-)




Carbazole
Concen: 35.102 ng
RT: 17.522 min Scan# 2454
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:	167	Resp:	1631047
Ion	Ratio	Lower	Upper
167	100		
166	21.9	17.1	25.7
139	14.8	10.7	16.1






Abundance Scan 2555 (18.116 min): BP006476.D\data.ms (-2476) (-)




Di-n-butylphthalate
Concen: 38.287 ng
RT: 18.081 min Scan# 2549
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Tgt Ion:	149	Resp:	2026877
Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.3	10.9
104	5.9	4.5	6.7






Abundance Scan 2732 (19.157 min): BP006476.D\data.ms (-2732) (-)




Fluoranthene
Concen: 35.996 ng
RT: 19.128 min Scan# 2727
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:	202	Resp:	1904699
Ion Ratio	Lower	Upper	
202	100		
101	12.7	0.0	27.6
203	17.7	0.0	37.8







Abundance Scan 3086 (21.239 min): BP006476.D\data.ms (-3079) (-)



Chrysene-d12
Concen: 20.000 ng
RT: 21.216 min Scan# 3082
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Tgt Ion:	240	Resp:	798630
Ion Ratio	Lower	Upper	
240	100		
120	12.3	6.5	9.7#
236	25.4	20.6	31.0









Benzidine
 Concen: 37.378 ng
 RT: 19.316 min Scan# 2759
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Tgt Ion:	184	Resp:	746973
Ion	Ratio	Lower	Upper
184	100		
185	13.9	11.0	16.4
183	12.1	9.5	14.3




Pyrene
 Concen: 36.763 ng
 RT: 19.481 min Scan# 2787
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27

Tgt Ion:	202	Resp:	1961108
Ion	Ratio	Lower	Upper
202	100		
200	21.1	16.3	24.5
203	17.3	14.2	21.2






Abundance Scan 2827 (19.716 min): BP006476.D\data.ms (-2770) (-)




Terphenyl-d14
Concen: 74.152 ng
RT: 19.686 min Scan# 2822
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:	244	Resp:	2955425
Ion	Ratio	Lower	Upper
244	100		
212	7.6	6.0	9.0
122	12.4	7.1	10.7#







Abundance Scan 2944 (20.404 min): BP006476.D\data.ms (-2880) (-)



Butylbenzylphthalate
Concen: 41.099 ng
RT: 20.369 min Scan# 2938
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Tgt Ion:	149	Resp:	964064
Ion	Ratio	Lower	Upper
149	100		
91	76.9	54.6	81.8
206	17.9	16.7	25.1









Benzo(a)anthracene
Concen: 35.981 ng
RT: 21.198 min Scan# 3079
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	228	Resp:	1877953
Ion	Ratio	Lower	Upper
228	100		
226	27.1	21.6	32.4
229	19.8	15.8	23.6



3,3'-Dichlorobenzidine
Concen: 38.088 ng
RT: 21.139 min Scan# 3069
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Tgt Ion:	252	Resp:	521894
Ion	Ratio	Lower	Upper
252	100		
254	64.2	50.6	76.0
126	12.6	7.0	10.6#









Scan 3092 (21.274 min): BP006476.D\data.ms (-3088) (-)
 Chrysene
 Concen: 35.461 ng
 RT: 21.251 min Scan# 3088
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Tgt Ion:	228	Resp:	1794302
Ion	Ratio	Lower	Upper
228	100		
226	30.3	23.8	35.6
229	19.6	15.4	23.2




Scan 3074 (21.169 min): BP006476.D\data.ms (-3087) (-)
 Bis(2-ethylhexyl)phthalate
 Concen: 39.652 ng
 RT: 21.133 min Scan# 3068
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27

Tgt Ion:	149	Resp:	1397750
Ion	Ratio	Lower	Upper
149	100		
167	25.4	21.9	32.9
279	3.5	4.3	6.5#






Abundance Scan 3228 (22.074 min): BP006476.D\data.ms (-3228) (-)




Di-n-octyl phthalate
Concen: 42.120 ng
RT: 22.039 min Scan# 3222
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:149 Resp: 2447169
Ion Ratio Lower Upper
149 100
167 1.5 1.4 2.0
43 7.9 5.6 8.4






Abundance Scan 3482 (23.568 min): BP006476.D\data.ms (-3482) (-)




Perylene-d12
Concen: 20.000 ng
RT: 23.533 min Scan# 3476
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Tgt Ion:264 Resp: 854338
Ion Ratio Lower Upper
264 100
260 23.8 18.6 28.0
265 22.1 17.6 26.4






Abundance Scan 3892 (25.980 min): BP006476.D\data.ms (-3877) (-)




Indeno(1,2,3-cd)pyrene
Concen: 36.137 ng
RT: 25.939 min Scan# 3885
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion	Ratio	Lower	Upper
276	100		
138	28.5	15.3	22.9#
277	25.3	20.4	30.6







Abundance Scan 3362 (22.863 min): BP006476.D\data.ms (-3358) (-)



Benzo(b)fluoranthene
Concen: 34.375 ng
RT: 22.833 min Scan# 3357
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Tgt Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.5	26.3
125	11.4	6.2	9.4#









Scan 3370 (22.910 min): BP006476.D\data.ms (-3369) (-)
 Benzo(k)fluoranthene
 Concen: 35.250 ng
 RT: 22.880 min Scan# 3365
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27




Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040


Tgt Ion:	252	Resp:	1879202
Ion	Ratio	Lower	Upper
252	100		
253	22.0	17.6	26.4
125	11.4	6.0	9.0#



Scan 3465 (23.468 min): BP006476.D\data.ms (-3459) (-)
 Benzo(a)pyrene
 Concen: 36.010 ng
 RT: 23.433 min Scan# 3459
 Delta R.T. 0.000 min
 Lab File: BP006824.D
 Acq: 23 Aug 2021 10:27

Tgt Ion:	252	Resp:	1793386
Ion	Ratio	Lower	Upper
252	100		
253	21.5	17.9	26.9
125	12.6	6.5	9.7#









Dibenzo(a,h)anthracene
Concen: 35.917 ng
RT: 25.951 min Scan# 3887
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 1923784
Ion Ratio Lower Upper
278 100
139 18.2 9.3 13.9#
279 23.3 18.8 28.2



Benzo(g,h,i)perylene
Concen: 36.147 ng
RT: 26.668 min Scan# 4009
Delta R.T. 0.000 min
Lab File: BP006824.D
Acq: 23 Aug 2021 10:27

Tgt Ion:276 Resp: 1855294
Ion Ratio Lower Upper
276 100
277 24.0 18.9 28.3
138 25.2 12.6 18.8#

