

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP082321\  
 Data File : BP006826.D  
 Acq On : 23 Aug 2021 12:12  
 Operator : CG/JU  
 Sample : PB138596BS  
 Misc :  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 PB138596BS

Quant Time: Aug 23 12:59:16 2021  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP080321.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Aug 23 11:39:38 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-------------------------------|--------|------|----------|---------|-------|----------|
| -----                         |        |      |          |         |       |          |
| Internal Standards            |        |      |          |         |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 7.693  | 152  | 160795   | 20.000  | ng    | 0.00     |
| 21) Naphthalene-d8            | 10.481 | 136  | 678256   | 20.000  | ng    | # 0.00   |
| 39) Acenaphthene-d10          | 14.345 | 164  | 383200   | 20.000  | ng    | 0.00     |
| 64) Phenanthrene-d10          | 17.104 | 188  | 755677   | 20.000  | ng    | # 0.00   |
| 76) Chrysene-d12              | 21.216 | 240  | 724596   | 20.000  | ng    | # 0.00   |
| 86) Perylene-d12              | 23.539 | 264  | 761662   | 20.000  | ng    | 0.00     |
| System Monitoring Compounds   |        |      |          |         |       |          |
| 5) 2-Fluorophenol             | 5.305  | 112  | 1474536  | 140.850 | ng    | 0.00     |
| 7) Phenol-d6                  | 6.893  | 99   | 1814072  | 121.986 | ng    | 0.00     |
| 23) Nitrobenzene-d5           | 8.858  | 82   | 1212356  | 82.505  | ng    | 0.00     |
| 42) 2,4,6-Tribromophenol      | 15.845 | 330  | 592964   | 148.061 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl          | 12.963 | 172  | 2249795  | 87.839  | ng    | 0.00     |
| 79) Terphenyl-d14             | 19.686 | 244  | 3340510  | 92.378  | ng    | 0.00     |
| Target Compounds              |        |      |          |         |       |          |
|                               |        |      |          |         |       | Qvalue   |
| 2) 1,4-Dioxane                | 3.240  | 88   | 159268   | 34.945  | ng    | 97       |
| 3) Pyridine                   | 3.634  | 79   | 415540   | 31.020  | ng    | 96       |
| 4) n-Nitrosodimethylamine     | 3.552  | 42   | 184209   | 36.433  | ng    | # 70     |
| 6) Aniline                    | 7.034  | 93   | 720721   | 44.616  | ng    | 98       |
| 8) 2-Chlorophenol             | 7.263  | 128  | 548885   | 48.458  | ng    | 95       |
| 9) Benzaldehyde               | 6.846  | 77   | 383245   | 42.818  | ng    | 98       |
| 10) Phenol                    | 6.916  | 94   | 665589   | 44.635  | ng    | 98       |
| 11) bis(2-Chloroethyl)ether   | 7.134  | 93   | 497841   | 43.968  | ng    | 95       |
| 12) 1,3-Dichlorobenzene       | 7.581  | 146  | 549984   | 46.587  | ng    | 98       |
| 13) 1,4-Dichlorobenzene       | 7.728  | 146  | 564191   | 47.089  | ng    | 98       |
| 14) 1,2-Dichlorobenzene       | 8.040  | 146  | 541057   | 47.062  | ng    | 97       |
| 15) Benzyl Alcohol            | 7.946  | 79   | 504603   | 45.247  | ng    | 95       |
| 16) 2,2'-oxybis(1-Chloropr... | 8.234  | 45   | 513676   | 38.124  | ng    | 96       |
| 17) 2-Methylphenol            | 8.152  | 107  | 447573   | 47.534  | ng    | 99       |
| 18) Hexachloroethane          | 8.758  | 117  | 223200   | 47.066  | ng    | 91       |
| 19) n-Nitroso-di-n-propyla... | 8.510  | 70   | 410413   | 41.135  | ng    | 94       |
| 20) 3+4-Methylphenols         | 8.481  | 107  | 604412   | 47.160  | ng    | 96       |
| 22) Acetophenone              | 8.528  | 105  | 815314   | 45.328  | ng    | 97       |
| 24) Nitrobenzene              | 8.899  | 77   | 630101   | 41.251  | ng    | 93       |
| 25) Isophorone                | 9.428  | 82   | 1084684  | 41.097  | ng    | 96       |
| 26) 2-Nitrophenol             | 9.605  | 139  | 285162   | 51.298  | ng    | 95       |
| 27) 2,4-Dimethylphenol        | 9.675  | 122  | 487299   | 52.348  | ng    | 96       |
| 28) bis(2-Chloroethoxy)met... | 9.910  | 93   | 657770   | 46.579  | ng    | 99       |
| 29) 2,4-Dichlorophenol        | 10.140 | 162  | 458944   | 48.697  | ng    | 95       |
| 30) 1,2,4-Trichlorobenzene    | 10.346 | 180  | 474325   | 46.837  | ng    | 97       |
| 31) Naphthalene               | 10.534 | 128  | 1624299  | 44.567  | ng    | 99       |
| 32) Benzoic acid              | 9.852  | 122  | 349917   | 48.591  | ng    | 94       |
| 33) 4-Chloroaniline           | 10.652 | 127  | 503287   | 34.125  | ng    | 96       |
| 34) Hexachlorobutadiene       | 10.810 | 225  | 282388   | 47.851  | ng    | 99       |
| 35) Caprolactam               | 11.475 | 113  | 169244   | 49.802  | ng    | # 76     |
| 36) 4-Chloro-3-methylphenol   | 11.799 | 107  | 561221   | 46.870  | ng    | 92       |
| 37) 2-Methylnaphthalene       | 12.151 | 142  | 1118508  | 45.290  | ng    | 100      |
| 38) 1-Methylnaphthalene       | 12.369 | 142  | 1028753  | 43.834  | ng    | 99       |
| 40) 1,2,4,5-Tetrachloroben... | 12.522 | 216  | 488217   | 48.678  | ng    | # 96     |
| 41) Hexachlorocyclopentadiene | 12.493 | 237  | 603821   | 113.514 | ng    | 99       |

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP082321\  
 Data File : BP006826.D  
 Acq On : 23 Aug 2021 12:12  
 Operator : CG/JU  
 Sample : PB138596BS  
 Misc :  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 PB138596BS

Quant Time: Aug 23 12:59:16 2021  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP080321.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Aug 23 11:39:38 2021  
 Response via : Initial Calibration


| Compound                      | R.T.   | QIon | Response | Conc    | Units | Dev(Min) |
|-------------------------------|--------|------|----------|---------|-------|----------|
| 43) 2,4,6-Trichlorophenol     | 12.775 | 196  | 338823   | 49.476  | ng    | 97       |
| 44) 2,4,5-Trichlorophenol     | 12.846 | 196  | 371695   | 49.034  | ng    | # 90     |
| 46) 1,1'-Biphenyl             | 13.175 | 154  | 1299454  | 44.802  | ng    | 98       |
| 47) 2-Chloronaphthalene       | 13.216 | 162  | 1022589  | 45.782  | ng    | 95       |
| 48) 2-Nitroaniline            | 13.434 | 65   | 367249   | 46.590  | ng    | 92       |
| 49) Acenaphthylene            | 14.063 | 152  | 1676548  | 46.550  | ng    | 100      |
| 50) Dimethylphthalate         | 13.816 | 163  | 1296276  | 46.472  | ng    | 99       |
| 51) 2,6-Dinitrotoluene        | 13.940 | 165  | 287764   | 46.232  | ng    | # 85     |
| 52) Acenaphthene              | 14.410 | 154  | 972198   | 44.354  | ng    | 99       |
| 53) 3-Nitroaniline            | 14.269 | 138  | 278365   | 40.341  | ng    | 86       |
| 54) 2,4-Dinitrophenol         | 14.481 | 184  | 329176   | 101.131 | ng    | # 84     |
| 55) Dibenzofuran              | 14.745 | 168  | 1508500  | 43.753  | ng    | 95       |
| 56) 4-Nitrophenol             | 14.598 | 139  | 568846   | 94.941  | ng    | 90       |
| 57) 2,4-Dinitrotoluene        | 14.728 | 165  | 407547   | 48.823  | ng    | # 78     |
| 58) Fluorene                  | 15.398 | 166  | 1236461  | 44.868  | ng    | 99       |
| 59) 2,3,4,6-Tetrachlorophenol | 14.981 | 232  | 306565   | 48.258  | ng    | # 72     |
| 60) Diethylphthalate          | 15.187 | 149  | 1388003  | 47.376  | ng    | 97       |
| 61) 4-Chlorophenyl-phenyle... | 15.398 | 204  | 573271   | 45.936  | ng    | 90       |
| 62) 4-Nitroaniline            | 15.440 | 138  | 319957   | 43.697  | ng    | 91       |
| 63) Azobenzene                | 15.692 | 77   | 1454941  | 42.595  | ng    | 93       |
| 65) 4,6-Dinitro-2-methylph... | 15.498 | 198  | 203863   | 44.152  | ng    | 80       |
| 66) n-Nitrosodiphenylamine    | 15.622 | 169  | 1070180  | 45.056  | ng    | 99       |
| 67) 4-Bromophenyl-phenylether | 16.298 | 248  | 349663   | 48.225  | ng    | # 89     |
| 68) Hexachlorobenzene         | 16.404 | 284  | 391414   | 47.437  | ng    | # 89     |
| 69) Atrazine                  | 16.587 | 200  | 388681   | 52.441  | ng    | 96       |
| 70) Pentachlorophenol         | 16.757 | 266  | 526162   | 100.506 | ng    | 100      |
| 71) Phenanthrene              | 17.151 | 178  | 1911935  | 44.855  | ng    | 99       |
| 72) Anthracene                | 17.239 | 178  | 1961519  | 47.624  | ng    | 99       |
| 73) Carbazole                 | 17.522 | 167  | 1844413  | 43.903  | ng    | 98       |
| 74) Di-n-butylphthalate       | 18.081 | 149  | 2436544  | 50.906  | ng    | 99       |
| 75) Fluoranthene              | 19.128 | 202  | 2203324  | 46.054  | ng    | 95       |
| 77) Benzidine                 | 19.322 | 184  | 1669982  | 92.102  | ng    | 99       |
| 78) Pyrene                    | 19.486 | 202  | 2244523  | 46.375  | ng    | 99       |
| 80) Butylbenzylphthalate      | 20.375 | 149  | 1140810  | 53.603  | ng    | 92       |
| 81) Benzo(a)anthracene        | 21.204 | 228  | 2163104  | 45.679  | ng    | 99       |
| 82) 3,3'-Dichlorobenzidine    | 21.139 | 252  | 744124   | 59.856  | ng    | # 97     |
| 83) Chrysene                  | 21.257 | 228  | 2079294  | 45.292  | ng    | 100      |
| 84) Bis(2-ethylhexyl)phtha... | 21.139 | 149  | 1679160  | 52.503  | ng    | # 97     |
| 85) Di-n-octyl phthalate      | 22.039 | 149  | 2915232  | 55.302  | ng    | 98       |
| 87) Indeno(1,2,3-cd)pyrene    | 25.945 | 276  | 2643402  | 47.863  | ng    | # 92     |
| 88) Benzo(b)fluoranthene      | 22.833 | 252  | 2257078  | 45.595  | ng    | # 97     |
| 89) Benzo(k)fluoranthene      | 22.880 | 252  | 2153727  | 45.315  | ng    | # 97     |
| 90) Benzo(a)pyrene            | 23.433 | 252  | 2182727  | 49.160  | ng    | # 96     |
| 91) Dibenzo(a,h)anthracene    | 25.957 | 278  | 2264447  | 47.422  | ng    | # 95     |
| 92) Benzo(g,h,i)perylene      | 26.680 | 276  | 2210512  | 48.308  | ng    | # 92     |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP082321\  
 Data File : BP006826.D  
 Acq On : 23 Aug 2021 12:12  
 Operator : CG/JU  
 Sample : PB138596BS  
 Misc :  
 ALS Vial : 4 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 PB138596BS

Quant Time: Aug 23 12:59:16 2021  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\8270E-BP080321.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Mon Aug 23 11:39:38 2021  
 Response via : Initial Calibration









Scan 794 (7.757 min): BP006476.D\data.ms (-786) (#1)  
 1,4-Dichlorobenzene-d4  
 Concen: 20.000 ng  
 RT: 7.693 min Scan# 783  
 Delta R.T. -0.006 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12




Instrument :  
 BNA\_P  
 ClientSampleId :  
 PB138596BS


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 152   | Resp: | 160795 |
| Ion      | Ratio | Lower | Upper  |
| 152      | 100   |       |        |
| 150      | 155.3 | 125.6 | 188.4  |
| 115      | 66.9  | 48.5  | 72.7   |



Scan 37 (3.305 min): BP006476.D\data.ms (-32) (#2)  
 1,4-Dioxane  
 Concen: 34.945 ng  
 RT: 3.240 min Scan# 26  
 Delta R.T. -0.006 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 88    | Resp: | 159268 |
| Ion      | Ratio | Lower | Upper  |
| 88       | 100   |       |        |
| 58       | 60.3  | 46.7  | 70.1   |
| 43       | 22.8  | 20.3  | 30.5   |









Pyridine  
Concen: 31.020 ng  
RT: 3.634 min Scan# 93  
Delta R.T. -0.006 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12




Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 79    | Resp: | 415540 |
| Ion      | Ratio | Lower | Upper  |
| 79       | 100   |       |        |
| 52       | 50.6  | 41.7  | 62.5   |
| 51       | 26.4  | 24.1  | 36.1   |



n-Nitrosodimethylamine  
Concen: 36.433 ng  
RT: 3.552 min Scan# 79  
Delta R.T. -0.006 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 42    | Resp: | 184209 |
| Ion      | Ratio | Lower | Upper  |
| 42       | 100   |       |        |
| 74       | 166.5 | 105.4 | 158.0# |
| 44       | 16.6  | 4.2   | 6.4#   |









2-Fluorophenol  
 Concen: 140.850 ng  
 RT: 5.305 min Scan# 377  
 Delta R.T. 0.006 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12




Instrument :  
 BNA\_P  
 ClientSampleId :  
 PB138596BS


|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 112   | Resp: | 1474536 |
| Ion      | Ratio | Lower | Upper   |
| 112      | 100   |       |         |
| 64       | 54.2  | 41.4  | 62.2    |
| 63       | 28.7  | 23.8  | 35.8    |



Aniline  
 Concen: 44.616 ng  
 RT: 7.034 min Scan# 671  
 Delta R.T. -0.000 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 93    | Resp: | 720721 |
| Ion      | Ratio | Lower | Upper  |
| 93       | 100   |       |        |
| 66       | 39.2  | 29.7  | 44.5   |
| 65       | 19.3  | 15.3  | 22.9   |









Phenol-d6  
 Concen: 121.986 ng  
 RT: 6.893 min Scan# 647  
 Delta R.T. 0.006 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12

Instrument :  
 BNA\_P  
 ClientSampleId :  
 PB138596BS

Tgt Ion: 99 Resp: 1814072




| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 99  | 100   |       |       |
| 42  | 15.0  | 13.4  | 20.0  |
| 71  | 36.5  | 30.6  | 46.0  |




2-Chlorophenol  
 Concen: 48.458 ng  
 RT: 7.263 min Scan# 710  
 Delta R.T. -0.000 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12

Tgt Ion: 128 Resp: 548885

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 128 | 100   |       |       |
| 130 | 32.6  | 12.4  | 52.4  |
| 64  | 45.8  | 20.3  | 60.3  |









Benzaldehyde  
 Concen: 42.818 ng  
 RT: 6.846 min Scan# 639  
 Delta R.T. -0.006 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12




Instrument :  
 BNA\_P  
 ClientSampleID :  
 PB138596BS


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 77    | Resp: | 383245 |
| Ion      | Ratio | Lower | Upper  |
| 77       | 100   |       |        |
| 105      | 96.5  | 76.7  | 116.7  |
| 106      | 92.1  | 76.2  | 116.2  |



Phenol  
 Concen: 44.635 ng  
 RT: 6.916 min Scan# 651  
 Delta R.T. -0.000 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 94    | Resp: | 665589 |
| Ion      | Ratio | Lower | Upper  |
| 94       | 100   |       |        |
| 65       | 29.0  | 8.8   | 48.8   |
| 66       | 39.2  | 20.7  | 60.7   |









Scan 699 (7.199 min): BP006476.D\data.ms (-692) #11  
 bis(2-Chloroethyl)ether  
 Concen: 43.968 ng  
 RT: 7.134 min Scan# 688  
 Delta R.T. -0.006 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12




Instrument :  
 BNA\_P  
 ClientSampleID :  
 PB138596BS

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 93    | Resp: | 497841 |
| Ion      | Ratio | Lower | Upper  |
| 93       | 100   |       |        |
| 63       | 71.9  | 46.2  | 86.2   |
| 95       | 32.3  | 12.2  | 52.2   |




Scan 775 (7.646 min): BP006476.D\data.ms (-763) #12  
 1,3-Dichlorobenzene  
 Concen: 46.587 ng  
 RT: 7.581 min Scan# 764  
 Delta R.T. -0.000 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 146   | Resp: | 549984 |
| Ion      | Ratio | Lower | Upper  |
| 146      | 100   |       |        |
| 148      | 63.4  | 51.5  | 77.3   |
| 75       | 33.7  | 25.7  | 38.5   |






Abundance Scan 800 (7.793 min): BP006476.D\data.ms (-792) #13




1,4-Dichlorobenzene  
Concen: 47.089 ng  
RT: 7.728 min Scan# 789  
Delta R.T. -0.006 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 146     | 100   |       |       |
| 148     | 64.0  | 51.3  | 76.9  |
| 111     | 44.4  | 33.0  | 49.6  |







Abundance Scan 854 (8.110 min): BP006476.D\data.ms (-842) #14



1,2-Dichlorobenzene  
Concen: 47.062 ng  
RT: 8.040 min Scan# 842  
Delta R.T. -0.006 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 146     | 100   |       |       |
| 148     | 63.9  | 50.3  | 75.5  |
| 111     | 46.4  | 33.7  | 50.5  |









Benzyl Alcohol  
 Concen: 45.247 ng  
 RT: 7.946 min Scan# 826  
 Delta R.T. -0.006 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12




Instrument :  
 BNA\_P  
 ClientSampleID :  
 PB138596BS


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 79    | Resp: | 504603 |
| Ion      | Ratio | Lower | Upper  |
| 79       | 100   |       |        |
| 108      | 71.3  | 62.5  | 93.7   |
| 77       | 65.0  | 52.9  | 79.3   |



2,2'-oxybis(1-Chloropropane)  
 Concen: 38.124 ng  
 RT: 8.234 min Scan# 875  
 Delta R.T. -0.000 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 45    | Resp: | 513676 |
| Ion      | Ratio | Lower | Upper  |
| 45       | 100   |       |        |
| 77       | 20.0  | 2.0   | 42.0   |
| 79       | 15.1  | 0.0   | 36.7   |









2-Methylphenol  
Concen: 47.534 ng  
RT: 8.152 min Scan# 861  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12




Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 107   | Resp: | 447573 |
| Ion      | Ratio | Lower | Upper  |
| 107      | 100   |       |        |
| 108      | 110.9 | 87.7  | 131.5  |
| 77       | 49.9  | 40.2  | 60.2   |
| 79       | 49.1  | 39.0  | 58.4   |



Hexachloroethane  
Concen: 47.066 ng  
RT: 8.758 min Scan# 964  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 117   | Resp: | 223200 |
| Ion      | Ratio | Lower | Upper  |
| 117      | 100   |       |        |
| 119      | 96.2  | 76.2  | 114.2  |
| 201      | 87.6  | 83.4  | 125.0  |









n-Nitroso-di-n-propylamine  
Concen: 41.135 ng  
RT: 8.510 min Scan# 922  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12




Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 70    | Resp: | 410413 |
| Ion      | Ratio | Lower | Upper  |
| 70       | 100   |       |        |
| 42       | 46.6  | 40.8  | 61.2   |
| 101      | 9.6   | 8.8   | 13.2   |
| 130      | 19.8  | 18.9  | 28.3   |



3+4-Methylphenols  
Concen: 47.160 ng  
RT: 8.481 min Scan# 917  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 107   | Resp: | 604412 |
| Ion      | Ratio | Lower | Upper  |
| 107      | 100   |       |        |
| 108      | 86.6  | 71.0  | 111.0  |
| 77       | 36.0  | 16.6  | 56.6   |
| 79       | 30.6  | 12.7  | 52.7   |





Abundance Scan 1268 (10.546 min): BP006476.D\data.ms (-122) (-)

Naphthalene-d8  
Concen: 20.000 ng  
RT: 10.481 min Scan# 1257  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12


Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS



| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 136     | 100   |       |       |
| 137     | 10.4  | 8.6   | 12.8  |
| 54      | 6.9   | 4.6   | 6.8#  |
| 68      | 6.2   | 3.6   | 5.4#  |







Abundance Scan 1257 (10.481 min): BP006826.D\data.ms (-1206) (-)




Abundance Scan 934 (8.581 min): BP006476.D\data.ms (-918) (-)

Acetophenone  
Concen: 45.328 ng  
RT: 8.528 min Scan# 925  
Delta R.T. 0.006 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 105     | 100   |       |       |
| 71      | 2.6   | 1.9   | 2.9   |
| 51      | 21.3  | 18.4  | 27.6  |
| 120     | 21.5  | 17.9  | 26.9  |






Abundance Scan 991 (8.916 min): BP006476.D\data.ms (-983) #23




Nitrobenzene-d5  
Concen: 82.505 ng  
RT: 8.858 min Scan# 981  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 82    | Resp: | 1212356 |
| Ion      | Ratio | Lower | Upper   |
| 82       | 100   |       |         |
| 128      | 41.5  | 35.7  | 53.5    |
| 54       | 41.4  | 35.0  | 52.4    |






Abundance Scan 998 (8.957 min): BP006476.D\data.ms (-983) #24




Nitrobenzene  
Concen: 41.251 ng  
RT: 8.899 min Scan# 988  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 77    | Resp: | 630101 |
| Ion      | Ratio | Lower | Upper  |
| 77       | 100   |       |        |
| 123      | 42.3  | 38.2  | 57.2   |
| 65       | 13.3  | 11.4  | 17.2   |






Abundance Scan 1088 (9.487 min): BP006476.D\data.ms (-1075) (-)




Isophorone  
Concen: 41.097 ng  
RT: 9.428 min Scan# 1078  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 82    | Resp: | 1084684 |
| Ion      | Ratio | Lower | Upper   |
| 82       | 100   |       |         |
| 95       | 7.6   | 6.5   | 9.7     |
| 138      | 16.4  | 15.0  | 22.4    |






Abundance Scan 1118 (9.663 min): BP006476.D\data.ms (-1102) (-)




2-Nitrophenol  
Concen: 51.298 ng  
RT: 9.605 min Scan# 1108  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 139   | Resp: | 285162 |
| Ion      | Ratio | Lower | Upper  |
| 139      | 100   |       |        |
| 109      | 33.7  | 28.8  | 43.2   |
| 65       | 55.6  | 41.0  | 61.6   |






Abundance Scan 1129 (9.728 min): BP006476.D\data.ms (-1129)




2,4-Dimethylphenol  
Concen: 52.348 ng  
RT: 9.675 min Scan# 1120  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 122   | Resp: | 487299 |
| Ion      | Ratio | Lower | Upper  |
| 122      | 100   |       |        |
| 107      | 124.0 | 94.5  | 141.7  |
| 121      | 61.3  | 47.4  | 71.2   |







Abundance Scan 1170 (9.969 min): BP006476.D\data.ms (-1169)



bis(2-Chloroethoxy)methane  
Concen: 46.579 ng  
RT: 9.910 min Scan# 1160  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 93    | Resp: | 657770 |
| Ion      | Ratio | Lower | Upper  |
| 93       | 100   |       |        |
| 95       | 32.5  | 25.9  | 38.9   |
| 123      | 10.9  | 10.2  | 15.2   |









Scan 1208 (10.193 min): BP006476.D\data.ms (-1159) (-)  
 2,4-Dichlorophenol  
 Concen: 48.697 ng  
 RT: 10.140 min Scan# 1199  
 Delta R.T. -0.000 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12




Instrument :  
 BNA\_P  
 ClientSampleId :  
 PB138596BS


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 162   | Resp: | 458944 |
| Ion      | Ratio | Lower | Upper  |
| 162      | 100   |       |        |
| 164      | 63.4  | 44.9  | 84.9   |
| 98       | 40.6  | 14.3  | 54.3   |



Scan 1245 (10.410 min): BP006476.D\data.ms (-1234) (-)  
 1,2,4-Trichlorobenzene  
 Concen: 46.837 ng  
 RT: 10.346 min Scan# 1234  
 Delta R.T. -0.000 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 180   | Resp: | 474325 |
| Ion      | Ratio | Lower | Upper  |
| 180      | 100   |       |        |
| 182      | 95.6  | 77.9  | 116.9  |
| 145      | 35.0  | 24.7  | 37.1   |









#34 (-)  
Naphthalene  
Concen: 44.567 ng  
RT: 10.534 min Scan# 1266  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12




Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS


|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 128   | Resp: | 1624299 |
| Ion      | Ratio | Lower | Upper   |
| 128      | 100   |       |         |
| 129      | 10.8  | 9.0   | 13.6    |
| 127      | 13.3  | 10.7  | 16.1    |



#32 (-)  
Benzoic acid  
Concen: 48.591 ng  
RT: 9.852 min Scan# 1150  
Delta R.T. 0.012 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 122   | Resp: | 349917 |
| Ion      | Ratio | Lower | Upper  |
| 122      | 100   |       |        |
| 105      | 129.1 | 103.2 | 143.2  |
| 77       | 103.0 | 76.2  | 116.2  |









4-Chloroaniline  
 Concen: 34.125 ng  
 RT: 10.652 min Scan# 1286  
 Delta R.T. -0.006 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12




Instrument :  
 BNA\_P  
 ClientSampleID :  
 PB138596BS

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 127   | Resp: | 503287 |
| Ion      | Ratio | Lower | Upper  |
| 127      | 100   |       |        |
| 129      | 32.0  | 26.1  | 39.1   |
| 65       | 32.2  | 22.5  | 33.7   |
| 92       | 21.1  | 15.0  | 22.6   |



Hexachlorobutadiene  
 Concen: 47.851 ng  
 RT: 10.810 min Scan# 1313  
 Delta R.T. -0.000 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 225   | Resp: | 282388 |
| Ion      | Ratio | Lower | Upper  |
| 225      | 100   |       |        |
| 223      | 62.2  | 51.0  | 76.6   |
| 227      | 64.6  | 51.4  | 77.2   |






Abundance Scan 1431 (11.504 min): BP006476.D\data.ms (-1435) (-)

Caprolactam  
Concen: 49.802 ng  
RT: 11.475 min Scan# 1426  
Delta R.T. 0.006 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS




|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 113   | Resp: | 169244 |
| Ion      | Ratio | Lower | Upper  |
| 113      | 100   |       |        |
| 55       | 145.9 | 98.9  | 138.9# |
| 56       | 113.3 | 71.7  | 111.7# |






Abundance Scan 1487 (11.834 min): BP006476.D\data.ms (-1486) (-)


4-Chloro-3-methylphenol  
Concen: 46.870 ng  
RT: 11.799 min Scan# 1481  
Delta R.T. 0.006 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12



|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 107   | Resp: | 561221 |
| Ion      | Ratio | Lower | Upper  |
| 107      | 100   |       |        |
| 144      | 23.8  | 21.1  | 31.7   |
| 142      | 74.5  | 66.0  | 99.0   |






Abundance Scan 1550 (12.204 min): BP006476.D\data.ms (-1537) (-)




2-Methylnaphthalene  
Concen: 45.290 ng  
RT: 12.151 min Scan# 1541  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 142   | Resp: | 1118508 |
| Ion      | Ratio | Lower | Upper   |
| 142      | 100   |       |         |
| 141      | 90.1  | 71.8  | 107.6   |
| 115      | 39.4  | 31.6  | 47.4    |






Abundance Scan 1588 (12.428 min): BP006476.D\data.ms (-1538) (-)




1-Methylnaphthalene  
Concen: 43.834 ng  
RT: 12.369 min Scan# 1578  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 142   | Resp: | 1028753 |
| Ion      | Ratio | Lower | Upper   |
| 142      | 100   |       |         |
| 141      | 92.9  | 73.3  | 109.9   |
| 116      | 4.4   | 3.5   | 5.3     |






Abundance Scan 1922 (14.392 min): BP006476.D\data.ms (-1863) (-)




Acenaphthene-d10  
Concen: 20.000 ng  
RT: 14.345 min Scan# 1914  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 164     | 100   |       |       |
| 162     | 103.0 | 80.2  | 120.2 |
| 160     | 46.3  | 34.5  | 51.7  |






Abundance Scan 1613 (12.575 min): BP006476.D\data.ms (-1604) (-)



1,2,4,5-Tetrachlorobenzene  
Concen: 48.678 ng  
RT: 12.522 min Scan# 1604  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 216     | 100   |       |       |
| 214     | 78.2  | 61.9  | 92.9  |
| 179     | 23.5  | 16.8  | 25.2  |
| 108     | 26.4  | 14.9  | 22.3# |





Abundance Scan 1609 (12.551 min): BP006476.D\data.ms (-1544) (-)

Hexachlorocyclopentadiene  
Concen: 113.514 ng  
RT: 12.493 min Scan# 1599  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12


Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS



| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 237     | 100   |       |       |
| 235     | 63.9  | 44.5  | 84.5  |
| 272     | 12.1  | 0.0   | 32.5  |




Abundance Scan 1599 (12.493 min): BP006826.D\data.ms (-1548) (-)





Abundance Scan 2175 (15.881 min): BP006476.D\data.ms (-2162) (-)


2,4,6-Tribromophenol  
Concen: 148.061 ng  
RT: 15.845 min Scan# 2169  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12




| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 330     | 100   |       |       |
| 332     | 96.0  | 77.8  | 116.8 |
| 141     | 45.0  | 25.4  | 38.2# |



Abundance Scan 2169 (15.845 min): BP006826.D\data.ms (-2118) (-)



Abundance Scan 1654 (12.816 min): BP006476.D\data.ms (-1643)





2,4,6-Trichlorophenol  
Concen: 49.476 ng  
RT: 12.775 min Scan# 1647  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12


Instrument :  
BNA\_P  
ClientSampleID :  
PB138596BS

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 196   | Resp: | 338823 |
| Ion      | Ratio | Lower | Upper  |
| 196      | 100   |       |        |
| 198      | 96.3  | 79.5  | 119.3  |
| 200      | 30.5  | 25.0  | 37.6   |


Abundance Scan 1647 (12.775 min): BP006826.D\data.ms



Abundance Scan 1647 (12.775 min): BP006826.D\data.ms (-15961)





Abundance Scan 1666 (12.887 min): BP006476.D\data.ms (-1660)




2,4,5-Trichlorophenol  
Concen: 49.034 ng  
RT: 12.846 min Scan# 1659  
Delta R.T. -0.006 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 196   | Resp: | 371695 |
| Ion      | Ratio | Lower | Upper  |
| 196      | 100   |       |        |
| 198      | 93.9  | 74.7  | 112.1  |
| 97       | 63.2  | 37.4  | 56.0#  |
| 132      | 33.1  | 20.0  | 30.0#  |


Abundance Scan 1659 (12.846 min): BP006826.D\data.ms



Abundance Scan 1659 (12.846 min): BP006826.D\data.ms (-16091)






Abundance Scan 1688 (13.016 min): BP006476.D\data.ms (-16745) (-)




2-Fluorobiphenyl  
Concen: 87.839 ng  
RT: 12.963 min Scan# 1679  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 172   | Resp: | 2249795 |
| Ion      | Ratio | Lower | Upper   |
| 172      | 100   |       |         |
| 171      | 35.8  | 27.9  | 41.9    |
| 170      | 24.1  | 18.6  | 27.8    |






Abundance Scan 1724 (13.228 min): BP006476.D\data.ms (-17146) (-)




1,1'-Biphenyl  
Concen: 44.802 ng  
RT: 13.175 min Scan# 1715  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 154   | Resp: | 1299454 |
| Ion      | Ratio | Lower | Upper   |
| 154      | 100   |       |         |
| 153      | 40.6  | 20.4  | 60.4    |
| 76       | 14.7  | 0.0   | 30.8    |






Abundance Scan 1730 (13.263 min): BP006476.D\data.ms (-1718)




2-Chloronaphthalene  
Concen: 45.782 ng  
RT: 13.216 min Scan# 1722  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 162   | Resp: | 1022589 |
| Ion      | Ratio | Lower | Upper   |
| 162      | 100   |       |         |
| 127      | 44.1  | 31.4  | 47.0    |
| 164      | 32.6  | 26.6  | 39.8    |







Abundance Scan 1766 (13.475 min): BP006476.D\data.ms (-1758)



2-Nitroaniline  
Concen: 46.590 ng  
RT: 13.434 min Scan# 1759  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 65    | Resp: | 367249 |
| Ion      | Ratio | Lower | Upper  |
| 65       | 100   |       |        |
| 92       | 71.9  | 57.7  | 86.5   |
| 138      | 100.7 | 92.1  | 138.1  |









Acenaphthylene  
Concen: 46.550 ng  
RT: 14.063 min Scan# 1866  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12




Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 152   | Resp: | 1676548 |
| Ion      | Ratio | Lower | Upper   |
| 152      | 100   |       |         |
| 151      | 20.3  | 16.2  | 24.2    |
| 153      | 13.0  | 10.6  | 16.0    |



Dimethylphthalate  
Concen: 46.472 ng  
RT: 13.816 min Scan# 1824  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12


|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 163   | Resp: | 1296276 |
| Ion      | Ratio | Lower | Upper   |
| 163      | 100   |       |         |
| 194      | 3.8   | 3.4   | 5.2     |
| 164      | 10.4  | 8.2   | 12.2    |





Abundance Scan 1851 (13.975 min): BP006476.D\data.ms (-1851) (-)

2,6-Dinitrotoluene  
Concen: 46.232 ng  
RT: 13.940 min Scan# 1845  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12


Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS



|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 165   | Resp: | 287764 |
| Ion      | Ratio | Lower | Upper  |
| 165      | 100   |       |        |
| 63       | 62.9  | 41.3  | 61.9#  |
| 89       | 66.8  | 44.8  | 67.2   |




Abundance Scan 1845 (13.940 min): BP006826.D\data.ms (-1794) (-)





Abundance Scan 1933 (14.457 min): BP006476.D\data.ms (-1953) (-)


Acenaphthene  
Concen: 44.354 ng  
RT: 14.410 min Scan# 1925  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12




|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 154   | Resp: | 972198 |
| Ion      | Ratio | Lower | Upper  |
| 154      | 100   |       |        |
| 153      | 113.4 | 89.4  | 134.2  |
| 152      | 54.5  | 43.0  | 64.4   |



Abundance Scan 1925 (14.410 min): BP006826.D\data.ms (-1874) (-)









3-Nitroaniline  
Concen: 40.341 ng  
RT: 14.269 min Scan# 1901  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12




Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 138   | Resp: | 278365 |
| Ion      | Ratio | Lower | Upper  |
| 138      | 100   |       |        |
| 108      | 10.9  | 8.4   | 12.6   |
| 92       | 125.5 | 87.8  | 131.8  |




2,4-Dinitrophenol  
Concen: 101.131 ng  
RT: 14.481 min Scan# 1937  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 184   | Resp: | 329176 |
| Ion      | Ratio | Lower | Upper  |
| 184      | 100   |       |        |
| 63       | 78.4  | 47.0  | 70.6#  |
| 154      | 71.3  | 52.3  | 78.5   |






Abundance Scan 1990 (14.792 min): BP006476.D\data.ms (-1955) (-)




Dibenzenofuran  
Concen: 43.753 ng  
RT: 14.745 min Scan# 1982  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 168   | Resp: | 1508500 |
| Ion      | Ratio | Lower | Upper   |
| 168      | 100   |       |         |
| 139      | 43.1  | 31.0  | 46.6    |
| 169      | 13.4  | 10.8  | 16.2    |






Abundance Scan 1959 (14.610 min): BP006476.D\data.ms (-1956) (-)




4-Nitrophenol  
Concen: 94.941 ng  
RT: 14.598 min Scan# 1957  
Delta R.T. 0.006 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 139   | Resp: | 568846 |
| Ion      | Ratio | Lower | Upper  |
| 139      | 100   |       |        |
| 109      | 85.0  | 72.3  | 112.3  |
| 65       | 103.6 | 72.5  | 112.5  |






Abundance Scan 1985 (14.763 min): BP006476.D\data.ms (-1957) (-)




2,4-Dinitrotoluene  
Concen: 48.823 ng  
RT: 14.728 min Scan# 1979  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 165     | 100   |       |       |
| 63      | 55.1  | 32.3  | 48.5# |
| 89      | 88.1  | 55.9  | 83.9# |
| 182     | 2.6   | 2.2   | 3.4   |






Abundance Scan 2100 (15.439 min): BP006476.D\data.ms (-2058) (-)




Fluorene  
Concen: 44.868 ng  
RT: 15.398 min Scan# 2093  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 166     | 100   |       |       |
| 165     | 98.4  | 78.0  | 117.0 |
| 167     | 13.5  | 10.5  | 15.7  |





Abundance Scan 2028 (15.016 min): BP006476.D\data.ms (-2059) (-)




2,3,4,6-Tetrachlorophenol  
Concen: 48.258 ng  
RT: 14.981 min Scan# 2022  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 232     | 100   |       |       |
| 131     | 60.2  | 32.2  | 48.4# |
| 130     | 3.2   | 2.0   | 3.0#  |
| 166     | 36.2  | 19.4  | 29.2# |



Abundance Scan 2022 (14.981 min): BP006826.D\data.ms (-1971) (-)





Abundance Scan 2064 (15.228 min): BP006476.D\data.ms (-2060) (-)




Diethylphthalate  
Concen: 47.376 ng  
RT: 15.187 min Scan# 2057  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 149     | 100   |       |       |
| 177     | 20.0  | 17.5  | 26.3  |
| 150     | 12.5  | 9.9   | 14.9  |



Abundance Scan 2057 (15.187 min): BP006826.D\data.ms (-2006) (-)






Abundance Scan 2100 (15.439 min): BP006476.D\data.ms (-2050) (-)




4-Chlorophenyl-phenylether  
Concen: 45.936 ng  
RT: 15.398 min Scan# 2093  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 204     | 100   |       |       |
| 206     | 33.4  | 26.6  | 40.0  |
| 141     | 72.7  | 48.6  | 73.0  |






Abundance Scan 2105 (15.469 min): BP006476.D\data.ms (-2052) (-)




4-Nitroaniline  
Concen: 43.697 ng  
RT: 15.440 min Scan# 2100  
Delta R.T. 0.006 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 138     | 100   |       |       |
| 92      | 57.2  | 31.2  | 71.2  |
| 108     | 89.9  | 78.6  | 118.6 |






Abundance Scan 2150 (15.733 min): BP006476.D\data.ms (-2150) (-)




Azobenzene  
Concen: 42.595 ng  
RT: 15.692 min Scan# 2143  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleID :  
PB138596BS

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 77    | Resp: | 1454941 |
| Ion      | Ratio | Lower | Upper   |
| 77       | 100   |       |         |
| 182      | 20.5  | 6.0   | 46.0    |
| 105      | 18.9  | 0.0   | 38.8    |
| 51       | 24.8  | 8.8   | 48.8    |






Abundance Scan 2389 (17.139 min): BP006476.D\data.ms (-2389) (-)




Phenanthrene-d10  
Concen: 20.000 ng  
RT: 17.104 min Scan# 2383  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 188   | Resp: | 755677 |
| Ion      | Ratio | Lower | Upper  |
| 188      | 100   |       |        |
| 94       | 10.8  | 5.6   | 8.4#   |
| 80       | 12.2  | 6.3   | 9.5#   |



Abundance Scan 2114 (15.522 min): BP006476.D\data.ms (-2165) (-)





4,6-Dinitro-2-methylphenol  
Concen: 44.152 ng  
RT: 15.498 min Scan# 2110  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12


Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 198   | Resp: | 203863 |
| Ion      | Ratio | Lower | Upper  |
| 198      | 100   |       |        |
| 51       | 44.4  | 14.0  | 54.0   |
| 105      | 52.5  | 19.5  | 59.5   |


Abundance Scan 2110 (15.498 min): BP006826.D\data.ms



Abundance Scan 2110 (15.498 min): BP006826.D\data.ms (-2059) (-)





Abundance Scan 2137 (15.657 min): BP006476.D\data.ms (-2166) (-)




n-Nitrosodiphenylamine  
Concen: 45.056 ng  
RT: 15.622 min Scan# 2131  
Delta R.T. 0.006 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 169   | Resp: | 1070180 |
| Ion      | Ratio | Lower | Upper   |
| 169      | 100   |       |         |
| 168      | 67.7  | 54.4  | 81.6    |
| 167      | 36.8  | 29.0  | 43.6    |


Abundance Scan 2131 (15.622 min): BP006826.D\data.ms



Abundance Scan 2131 (15.622 min): BP006826.D\data.ms (-2079) (-)






Abundance Scan 2253 (16.339 min): BP006476.D\data.ms (-2195) (-)




4-Bromophenyl-phenylether  
Concen: 48.225 ng  
RT: 16.298 min Scan# 2246  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 248     | 100   |       |       |
| 250     | 98.1  | 79.8  | 119.8 |
| 141     | 91.8  | 57.8  | 86.6# |






Abundance Scan 2270 (16.439 min): BP006476.D\data.ms (-2195) (-)




Hexachlorobenzene  
Concen: 47.437 ng  
RT: 16.404 min Scan# 2264  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 284     | 100   |       |       |
| 142     | 43.8  | 26.9  | 40.3# |
| 249     | 32.4  | 24.5  | 36.7  |





Abundance Scan 2300 (16.616 min): BP006476.D\data.ms (-2299) (-)




Atrazine  
Concen: 52.441 ng  
RT: 16.587 min Scan# 2295  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS


| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 200     | 100   |       |       |
| 173     | 27.4  | 4.1   | 44.1  |
| 215     | 46.0  | 27.7  | 67.7  |



Abundance Scan 2295 (16.587 min): BP006826.D\data.ms (-2244) (-)





Abundance Scan 2329 (16.786 min): BP006476.D\data.ms (-2378) (-)





Pentachlorophenol  
Concen: 100.506 ng  
RT: 16.757 min Scan# 2324  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 266     | 100   |       |       |
| 268     | 64.2  | 51.6  | 77.4  |
| 264     | 63.2  | 50.3  | 75.5  |



Abundance Scan 2324 (16.757 min): BP006826.D\data.ms (-2273) (-)









Phenanthrene  
Concen: 44.855 ng  
RT: 17.151 min Scan# 2391  
Delta R.T. 0.006 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12




Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 178   | Resp: | 1911935 |
| Ion      | Ratio | Lower | Upper   |
| 178      | 100   |       |         |
| 176      | 19.5  | 15.1  | 22.7    |
| 179      | 15.5  | 12.4  | 18.6    |




Anthracene  
Concen: 47.624 ng  
RT: 17.239 min Scan# 2406  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 178   | Resp: | 1961519 |
| Ion      | Ratio | Lower | Upper   |
| 178      | 100   |       |         |
| 176      | 19.1  | 14.9  | 22.3    |
| 179      | 15.6  | 12.2  | 18.2    |





Abundance Scan 2459 (17.551 min): BP006476.D\data.ms (-2403) (-)




Carbazole  
Concen: 43.903 ng  
RT: 17.522 min Scan# 2454  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS


|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 167   | Resp: | 1844413 |
| Ion      | Ratio | Lower | Upper   |
| 167      | 100   |       |         |
| 166      | 21.9  | 17.1  | 25.7    |
| 139      | 14.6  | 10.7  | 16.1    |



Abundance Scan 2454 (17.522 min): BP006826.D\data.ms (-2403) (-)





Abundance Scan 2555 (18.116 min): BP006476.D\data.ms (-2476) (-)




Di-n-butylphthalate  
Concen: 50.906 ng  
RT: 18.081 min Scan# 2549  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12


|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 149   | Resp: | 2436544 |
| Ion      | Ratio | Lower | Upper   |
| 149      | 100   |       |         |
| 150      | 9.2   | 7.3   | 10.9    |
| 104      | 6.0   | 4.5   | 6.7     |



Abundance Scan 2549 (18.081 min): BP006826.D\data.ms (-2498) (-)






Abundance Scan 2732 (19.157 min): BP006476.D\data.ms (-2725)




Fluoranthene  
Concen: 46.054 ng  
RT: 19.128 min Scan# 2727  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 202   | Resp: | 2203324 |
| Ion      | Ratio | Lower | Upper   |
| 202      | 100   |       |         |
| 101      | 13.2  | 0.0   | 27.6    |
| 203      | 17.7  | 0.0   | 37.8    |







Abundance Scan 3086 (21.239 min): BP006476.D\data.ms (-3079)



Chrysene-d12  
Concen: 20.000 ng  
RT: 21.216 min Scan# 3082  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 240   | Resp: | 724596 |
| Ion      | Ratio | Lower | Upper  |
| 240      | 100   |       |        |
| 120      | 12.5  | 6.5   | 9.7#   |
| 236      | 25.6  | 20.6  | 31.0   |









Benzidine  
 Concen: 92.102 ng  
 RT: 19.322 min Scan# 2760  
 Delta R.T. 0.006 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12




Instrument :  
 BNA\_P  
 ClientSampleId :  
 PB138596BS

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 184   | Resp: | 1669982 |
| Ion      | Ratio | Lower | Upper   |
| 184      | 100   |       |         |
| 185      | 14.2  | 11.0  | 16.4    |
| 183      | 12.2  | 9.5   | 14.3    |




Pyrene  
 Concen: 46.375 ng  
 RT: 19.486 min Scan# 2788  
 Delta R.T. 0.006 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 202   | Resp: | 2244523 |
| Ion      | Ratio | Lower | Upper   |
| 202      | 100   |       |         |
| 200      | 20.9  | 16.3  | 24.5    |
| 203      | 17.3  | 14.2  | 21.2    |






Abundance Scan 2827 (19.716 min): BP006476.D\data.ms (-2770) (-)




Terphenyl-d14  
Concen: 92.378 ng  
RT: 19.686 min Scan# 2822  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS

Tgt Ion:244 Resp: 3340510  
Ion Ratio Lower Upper  
244 100  
212 7.7 6.0 9.0  
122 13.0 7.1 10.7#







Abundance Scan 2944 (20.404 min): BP006476.D\data.ms (-2880) (-)



Butylbenzylphthalate  
Concen: 53.603 ng  
RT: 20.375 min Scan# 2939  
Delta R.T. 0.006 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Tgt Ion:149 Resp: 1140810  
Ion Ratio Lower Upper  
149 100  
91 75.4 54.6 81.8  
206 18.6 16.7 25.1









Scan 3083 (21.221 min): BP006476.D\data.ms (-3083) (-)  
 Benzo(a)anthracene  
 Concen: 45.679 ng  
 RT: 21.204 min Scan# 3080  
 Delta R.T. 0.006 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12




Instrument :  
 BNA\_P  
 ClientSampleId :  
 PB138596BS


|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 228   | Resp: | 2163104 |
| Ion      | Ratio | Lower | Upper   |
| 228      | 100   |       |         |
| 226      | 27.6  | 21.6  | 32.4    |
| 229      | 19.7  | 15.8  | 23.6    |



Scan 3073 (21.163 min): BP006476.D\data.ms (-3073) (-)  
 3,3'-Dichlorobenzidine  
 Concen: 59.856 ng  
 RT: 21.139 min Scan# 3069  
 Delta R.T. -0.000 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 252   | Resp: | 744124 |
| Ion      | Ratio | Lower | Upper  |
| 252      | 100   |       |        |
| 254      | 64.9  | 50.6  | 76.0   |
| 126      | 12.8  | 7.0   | 10.6#  |









Chrysene  
 Concen: 45.292 ng  
 RT: 21.257 min Scan# 3089  
 Delta R.T. 0.006 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12




Instrument :  
 BNA\_P  
 ClientSampleId :  
 PB138596BS

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 228   | Resp: | 2079294 |
| Ion      | Ratio | Lower | Upper   |
| 228      | 100   |       |         |
| 226      | 29.9  | 23.8  | 35.6    |
| 229      | 19.5  | 15.4  | 23.2    |




Bis(2-ethylhexyl)phthalate  
 Concen: 52.503 ng  
 RT: 21.139 min Scan# 3069  
 Delta R.T. 0.006 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 149   | Resp: | 1679160 |
| Ion      | Ratio | Lower | Upper   |
| 149      | 100   |       |         |
| 167      | 26.2  | 21.9  | 32.9    |
| 279      | 4.0   | 4.3   | 6.5#    |





Abundance Scan 3228 (22.074 min): BP006476.D\data.ms (-3228) (-)




Di-n-octyl phthalate  
Concen: 55.302 ng  
RT: 22.039 min Scan# 3222  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS


|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 149   | Resp: | 2915232 |
| Ion      | Ratio | Lower | Upper   |
| 149      | 100   |       |         |
| 167      | 1.5   | 1.4   | 2.0     |
| 43       | 7.9   | 5.6   | 8.4     |



Abundance Scan 3222 (22.039 min): BP006826.D\data.ms (-3222) (-)





Abundance Scan 3482 (23.568 min): BP006476.D\data.ms (-3482) (-)




Perylene-d12  
Concen: 20.000 ng  
RT: 23.539 min Scan# 3477  
Delta R.T. 0.006 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12


|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 264   | Resp: | 761662 |
| Ion      | Ratio | Lower | Upper  |
| 264      | 100   |       |        |
| 260      | 23.6  | 18.6  | 28.0   |
| 265      | 21.8  | 17.6  | 26.4   |



Abundance Scan 3477 (23.539 min): BP006826.D\data.ms (-3477) (-)





Abundance Scan 3892 (25.980 min): BP006476.D\data.ms (-3877) (-)




Indeno(1,2,3-cd)pyrene  
Concen: 47.863 ng  
RT: 25.945 min Scan# 3886  
Delta R.T. 0.006 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS


Tgt Ion:276 Resp: 2643402  
Ion Ratio Lower Upper  
276 100  
138 28.0 15.3 22.9#  
277 25.5 20.4 30.6



Abundance Scan 3886 (25.945 min): BP006826.D\data.ms (-3834) (-)





Abundance Scan 3362 (22.863 min): BP006476.D\data.ms (-3358) (-)





Benzo(b)fluoranthene  
Concen: 45.595 ng  
RT: 22.833 min Scan# 3357  
Delta R.T. -0.000 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Tgt Ion:252 Resp: 2257078  
Ion Ratio Lower Upper  
252 100  
253 22.1 17.5 26.3  
125 11.8 6.2 9.4#



Abundance Scan 3357 (22.833 min): BP006826.D\data.ms (-3306) (-)









Benzo(k)fluoranthene  
 Concen: 45.315 ng  
 RT: 22.880 min Scan# 3365  
 Delta R.T. -0.000 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12




Instrument :  
 BNA\_P  
 ClientSampleId :  
 PB138596BS

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 252   | Resp: | 2153727 |
| Ion      | Ratio | Lower | Upper   |
| 252      | 100   |       |         |
| 253      | 22.2  | 17.6  | 26.4    |
| 125      | 11.4  | 6.0   | 9.0#    |




Benzo(a)pyrene  
 Concen: 49.160 ng  
 RT: 23.433 min Scan# 3459  
 Delta R.T. -0.000 min  
 Lab File: BP006826.D  
 Acq: 23 Aug 2021 12:12

|          |       |       |         |
|----------|-------|-------|---------|
| Tgt Ion: | 252   | Resp: | 2182727 |
| Ion      | Ratio | Lower | Upper   |
| 252      | 100   |       |         |
| 253      | 22.1  | 17.9  | 26.9    |
| 125      | 12.9  | 6.5   | 9.7#    |






Abundance Scan 3897 (26.009 min): BP006476.D\data.ms (-3897) (-)




Dibenzo(a,h)anthracene  
Concen: 47.422 ng  
RT: 25.957 min Scan# 3888  
Delta R.T. 0.006 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Instrument :  
BNA\_P  
ClientSampleId :  
PB138596BS

Tgt Ion:278 Resp: 2264447  
Ion Ratio Lower Upper  
278 100  
139 17.2 9.3 13.9#  
279 23.8 18.8 28.2



Abundance Scan 4019 (26.727 min): BP006476.D\data.ms (-4019) (-)



Benzo(g,h,i)perylene  
Concen: 48.308 ng  
RT: 26.680 min Scan# 4011  
Delta R.T. 0.012 min  
Lab File: BP006826.D  
Acq: 23 Aug 2021 12:12

Tgt Ion:276 Resp: 2210512  
Ion Ratio Lower Upper  
276 100  
277 23.5 18.9 28.3  
138 24.4 12.6 18.8#

