

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082421\
 Data File : BP006854.D
 Acq On : 24 Aug 2021 19:22
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Aug 25 02:34:02 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 24 13:40:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.687	152	180858	20.000	ng	0.00
21) Naphthalene-d8	10.475	136	735359	20.000	ng	# 0.00
39) Acenaphthene-d10	14.334	164	410806	20.000	ng	0.00
64) Phenanthrene-d10	17.092	188	790865	20.000	ng	# 0.00
76) Chrysene-d12	21.198	240	743494	20.000	ng	#-0.01
86) Perylene-d12	23.504	264	792977	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.293	112	979688	83.200	ng	0.00
7) Phenol-d6	6.881	99	1343550	80.324	ng	0.00
23) Nitrobenzene-d5	8.852	82	1306541	82.010	ng	0.00
42) 2,4,6-Tribromophenol	15.834	330	407847	94.994	ng	-0.01
45) 2-Fluorobiphenyl	12.957	172	2426078	88.356	ng	0.00
79) Terphenyl-d14	19.675	244	3378520	91.054	ng	-0.01
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.240	88	192637	37.577	ng	96
3) Pyridine	3.640	79	556855	36.958	ng	95
4) n-Nitrosodimethylamine	3.552	42	190845	33.559	ng	# 72
6) Aniline	7.028	93	717540	39.492	ng	97
8) 2-Chlorophenol	7.258	128	544007	42.699	ng	96
9) Benzaldehyde	6.846	77	375030	37.252	ng	98
10) Phenol	6.911	94	661600	39.446	ng	99
11) bis(2-Chloroethyl)ether	7.134	93	499536	39.224	ng	95
12) 1,3-Dichlorobenzene	7.575	146	568282	42.797	ng	98
13) 1,4-Dichlorobenzene	7.722	146	576382	42.770	ng	98
14) 1,2-Dichlorobenzene	8.034	146	551618	42.658	ng	97
15) Benzyl Alcohol	7.940	79	508588	40.545	ng	95
16) 2,2'-oxybis(1-Chloropr...	8.216	45	503567	33.228	ng	97
17) 2-Methylphenol	8.146	107	443088	41.837	ng	99
18) Hexachloroethane	8.752	117	227676	42.684	ng	91
19) n-Nitroso-di-n-propyla...	8.505	70	433935	38.668	ng	93
20) 3+4-Methylphenols	8.475	107	607081	42.114	ng	97
22) Acetophenone	8.516	105	811270	41.601	ng	# 97
24) Nitrobenzene	8.893	77	666534	40.247	ng	94
25) Isophorone	9.422	82	1188514	41.535	ng	96
26) 2-Nitrophenol	9.599	139	298776	49.573	ng	96
27) 2,4-Dimethylphenol	9.669	122	433108	42.914	ng	96
28) bis(2-Chloroethoxy)met...	9.905	93	623356	40.714	ng	98
29) 2,4-Dichlorophenol	10.134	162	469058	45.905	ng	95
30) 1,2,4-Trichlorobenzene	10.340	180	486989	44.354	ng	96
31) Naphthalene	10.522	128	1678406	42.476	ng	100
32) Benzoic acid	9.846	122	362688	46.453	ng	94
33) 4-Chloroaniline	10.646	127	680403	42.551	ng	96
34) Hexachlorobutadiene	10.799	225	298805	46.701	ng	99
35) Caprolactam	11.469	113	162257	44.039	ng	82
36) 4-Chloro-3-methylphenol	11.787	107	569744	43.886	ng	91
37) 2-Methylnaphthalene	12.140	142	1145375	42.777	ng	100
38) 1-Methylnaphthalene	12.363	142	1083433	42.579	ng	99
40) 1,2,4,5-Tetrachloroben...	12.510	216	490489	45.618	ng	# 96
41) Hexachlorocyclopentadiene	12.481	237	232948	40.850	ng	99
43) 2,4,6-Trichlorophenol	12.763	196	352731	48.046	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082421\
 Data File : BP006854.D
 Acq On : 24 Aug 2021 19:22
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Aug 25 02:34:02 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 24 13:40:38 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	12.840	196	376914	46.381	ng	# 89
46) 1,1'-Biphenyl	13.163	154	1357423	43.655	ng	97
47) 2-Chloronaphthalene	13.204	162	1047476	43.745	ng	94
48) 2-Nitroaniline	13.422	65	367068	43.438	ng	90
49) Acenaphthylene	14.051	152	1687329	43.701	ng	100
50) Dimethylphthalate	13.810	163	1297301	43.383	ng	100
51) 2,6-Dinitrotoluene	13.928	165	297737	44.620	ng	# 80
52) Acenaphthene	14.398	154	1007399	42.871	ng	97
53) 3-Nitroaniline	14.257	138	323705	43.759	ng	85
54) 2,4-Dinitrophenol	14.469	184	164984	47.281	ng	# 82
55) Dibenzofuran	14.734	168	1570597	42.493	ng	93
56) 4-Nitrophenol	14.581	139	278078	43.293	ng	90
57) 2,4-Dinitrotoluene	14.722	165	405533	45.317	ng	# 79
58) Fluorene	15.387	166	1259680	42.639	ng	99
59) 2,3,4,6-Tetrachlorophenol	14.969	232	310359	45.572	ng	# 71
60) Diethylphthalate	15.175	149	1370546	43.637	ng	97
61) 4-Chlorophenyl-phenyle...	15.387	204	590011	44.100	ng	91
62) 4-Nitroaniline	15.428	138	316514	40.322	ng	94
63) Azobenzene	15.681	77	1507487	41.167	ng	93
65) 4,6-Dinitro-2-methylph...	15.487	198	229816	47.558	ng	78
66) n-Nitrosodiphenylamine	15.610	169	1078798	43.398	ng	99
67) 4-Bromophenyl-phenylether	16.287	248	349525	46.061	ng	# 89
68) Hexachlorobenzene	16.392	284	389896	45.151	ng	# 88
69) Atrazine	16.575	200	332077	42.810	ng	96
70) Pentachlorophenol	16.745	266	251872	45.971	ng	100
71) Phenanthrene	17.134	178	1886773	42.295	ng	99
72) Anthracene	17.228	178	1891766	43.887	ng	99
73) Carbazole	17.504	167	1854331	42.175	ng	98
74) Di-n-butylphthalate	18.069	149	2319079	46.296	ng	99
75) Fluoranthene	19.116	202	2144505	42.830	ng	95
77) Benzidine	19.304	184	821063	44.132	ng	99
78) Pyrene	19.469	202	2207873	44.458	ng	99
80) Butylbenzylphthalate	20.357	149	1093800	50.088	ng	90
81) Benzo(a)anthracene	21.180	228	2134141	43.921	ng	99
82) 3,3'-Dichlorobenzidine	21.122	252	610665	47.872	ng	# 98
83) Chrysene	21.233	228	2047868	43.474	ng	98
84) Bis(2-ethylhexyl)phtha...	21.122	149	1586758	48.352	ng	# 98
85) Di-n-octyl phthalate	22.016	149	2777016	51.341	ng	98
87) Indeno(1,2,3-cd)pyrene	25.892	276	2519748	43.822	ng	# 92
88) Benzo(b)fluoranthene	22.804	252	2171261	42.130	ng	# 97
89) Benzo(k)fluoranthene	22.851	252	2123848	42.922	ng	# 96
90) Benzo(a)pyrene	23.404	252	2021100	43.722	ng	# 95
91) Dibenzo(a,h)anthracene	25.910	278	2172982	43.709	ng	# 94
92) Benzo(g,h,i)perylene	26.621	276	2085859	43.784	ng	# 92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082421\
 Data File : BP006854.D
 Acq On : 24 Aug 2021 19:22
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTDCCC040

Quant Time: Aug 25 02:34:02 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Aug 24 13:40:38 2021
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.687 min Scan# 782
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 180858
Ion Ratio Lower Upper
152 100
150 152.3 125.6 188.4
115 65.9 48.5 72.7



#2
1,4-Dioxane
Concen: 37.577 ng
RT: 3.240 min Scan# 26
Delta R.T. 0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion: 88 Resp: 192637
Ion Ratio Lower Upper
88 100
58 60.0 46.7 70.1
43 22.0 20.3 30.5









#3
Pyridine
Concen: 36.958 ng
RT: 3.640 min Scan# 94
Delta R.T. 0.012 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 556855
Ion Ratio Lower Upper
79 100
52 50.1 41.7 62.5
51 25.9 24.1 36.1



#4
n-Nitrosodimethylamine
Concen: 33.559 ng
RT: 3.552 min Scan# 79
Delta R.T. 0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion: 42 Resp: 190845
Ion Ratio Lower Upper
42 100
74 164.2 105.4 158.0#
44 15.4 4.2 6.4#









#5
2-Fluorophenol
Concen: 83.200 ng
RT: 5.293 min Scan# 375
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:112 Resp: 979688
Ion Ratio Lower Upper
112 100
64 54.9 41.4 62.2
63 29.4 23.8 35.8



#6
Aniline
Concen: 39.492 ng
RT: 7.028 min Scan# 670
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion: 93 Resp: 717540
Ion Ratio Lower Upper
93 100
66 39.2 29.7 44.5
65 19.6 15.3 22.9









#7
Phenol-d6
Concen: 80.324 ng
RT: 6.881 min Scan# 645
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 99 Resp: 1343550
Ion Ratio Lower Upper
99 100
42 15.3 13.4 20.0
71 37.2 30.6 46.0



#8
2-Chlorophenol
Concen: 42.699 ng
RT: 7.258 min Scan# 709
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion: 128 Resp: 544007
Ion Ratio Lower Upper
128 100
130 32.2 12.4 52.4
64 44.3 20.3 60.3









#9
Benzaldehyde
Concen: 37.252 ng
RT: 6.846 min Scan# 639
Delta R.T. 0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 77 Resp: 375030
Ion Ratio Lower Upper
77 100
105 95.8 76.7 116.7
106 93.2 76.2 116.2



#10
Phenol
Concen: 39.446 ng
RT: 6.911 min Scan# 650
Delta R.T. 0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion: 94 Resp: 661600
Ion Ratio Lower Upper
94 100
65 28.8 8.8 48.8
66 39.4 20.7 60.7









#11
bis(2-Chloroethyl)ether
Concen: 39.224 ng
RT: 7.134 min Scan# 688
Delta R.T. 0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 93 Resp: 499536
Ion Ratio Lower Upper
93 100
63 71.2 46.2 86.2
95 33.3 12.2 52.2



#12
1,3-Dichlorobenzene
Concen: 42.797 ng
RT: 7.575 min Scan# 763
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:146 Resp: 568282
Ion Ratio Lower Upper
146 100
148 63.7 51.5 77.3
75 34.2 25.7 38.5









#13
1,4-Dichlorobenzene
Concen: 42.770 ng
RT: 7.722 min Scan# 788
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:146 Resp: 576382
Ion Ratio Lower Upper
146 100
148 63.1 51.3 76.9
111 43.4 33.0 49.6



#14
1,2-Dichlorobenzene
Concen: 42.658 ng
RT: 8.034 min Scan# 841
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:146 Resp: 551618
Ion Ratio Lower Upper
146 100
148 63.6 50.3 75.5
111 46.4 33.7 50.5









#15
Benzyl Alcohol
Concen: 40.545 ng
RT: 7.940 min Scan# 825
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 79 Resp: 508588
Ion Ratio Lower Upper
79 100
108 71.0 62.5 93.7
77 65.6 52.9 79.3



#16
2,2'-oxybis(1-Chloropropane)
Concen: 33.228 ng
RT: 8.216 min Scan# 872
Delta R.T. -0.012 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion: 45 Resp: 503567
Ion Ratio Lower Upper
45 100
77 20.9 2.0 42.0
79 15.5 0.0 36.7









#17
2-Methylphenol
Concen: 41.837 ng
RT: 8.146 min Scan# 860
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	107	Resp:	443088
Ion	Ratio	Lower	Upper
107	100		
108	111.7	87.7	131.5
77	51.0	40.2	60.2
79	48.7	39.0	58.4



#18
Hexachloroethane
Concen: 42.684 ng
RT: 8.752 min Scan# 963
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:	117	Resp:	227676
Ion	Ratio	Lower	Upper
117	100		
119	98.0	76.2	114.2
201	88.6	83.4	125.0









#19
n-Nitroso-di-n-propylamine
Concen: 38.668 ng
RT: 8.505 min Scan# 921
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	70	Resp:	433935
Ion	Ratio	Lower	Upper
70	100		
42	45.9	40.8	61.2
101	9.6	8.8	13.2
130	20.5	18.9	28.3



#20
3+4-Methylphenols
Concen: 42.114 ng
RT: 8.475 min Scan# 916
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:	107	Resp:	607081
Ion	Ratio	Lower	Upper
107	100		
108	86.8	71.0	111.0
77	36.0	16.6	56.6
79	31.3	12.7	52.7









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.475 min Scan# 1256
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	136	Resp:	735359
Ion Ratio	Lower	Upper	
136	100		
137	10.8	8.6	12.8
54	6.5	4.6	6.8
68	5.8	3.6	5.4#



#22
Acetophenone
Concen: 41.601 ng
RT: 8.516 min Scan# 923
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:	105	Resp:	811270
Ion Ratio	Lower	Upper	
105	100		
71	5.7	1.9	2.9#
51	21.5	18.4	27.6
120	21.4	17.9	26.9









#23
Nitrobenzene-d5
Concen: 82.010 ng
RT: 8.852 min Scan# 980
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 1306541
Ion Ratio Lower Upper
82 100
128 41.2 35.7 53.5
54 41.4 35.0 52.4



#24
Nitrobenzene
Concen: 40.247 ng
RT: 8.893 min Scan# 987
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion: 77 Resp: 666534
Ion Ratio Lower Upper
77 100
123 42.6 38.2 57.2
65 13.5 11.4 17.2









#25
Isophorone
Concen: 41.535 ng
RT: 9.422 min Scan# 1077
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion: 82 Resp: 1188514
Ion Ratio Lower Upper
82 100
95 7.7 6.5 9.7
138 16.6 15.0 22.4



#26
2-Nitrophenol
Concen: 49.573 ng
RT: 9.599 min Scan# 1107
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:139 Resp: 298776
Ion Ratio Lower Upper
139 100
109 33.8 28.8 43.2
65 54.6 41.0 61.6









#27
2,4-Dimethylphenol
Concen: 42.914 ng
RT: 9.669 min Scan# 1119
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:122 Resp: 433108
Ion Ratio Lower Upper
122 100
107 123.2 94.5 141.7
121 61.2 47.4 71.2



#28
bis(2-Chloroethoxy)methane
Concen: 40.714 ng
RT: 9.905 min Scan# 1159
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion: 93 Resp: 623356
Ion Ratio Lower Upper
93 100
95 32.8 25.9 38.9
123 11.1 10.2 15.2









#29
2,4-Dichlorophenol
Concen: 45.905 ng
RT: 10.134 min Scan# 1198
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 469058
Ion Ratio Lower Upper
162 100
164 64.2 44.9 84.9
98 41.1 14.3 54.3



#30
1,2,4-Trichlorobenzene
Concen: 44.354 ng
RT: 10.340 min Scan# 1233
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:180 Resp: 486989
Ion Ratio Lower Upper
180 100
182 94.2 77.9 116.9
145 33.7 24.7 37.1









#31
Naphthalene
Concen: 42.476 ng
RT: 10.522 min Scan# 1264
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:128 Resp: 1678406
Ion Ratio Lower Upper
128 100
129 11.1 9.0 13.6
127 13.4 10.7 16.1



#32
Benzoic acid
Concen: 46.453 ng
RT: 9.846 min Scan# 1149
Delta R.T. 0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:122 Resp: 362688
Ion Ratio Lower Upper
122 100
105 130.0 103.2 143.2
77 102.8 76.2 116.2









#33
4-Chloroaniline
Concen: 42.551 ng
RT: 10.646 min Scan# 1285
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	127	Resp:	680403
Ion	Ratio	Lower	Upper
127	100		
129	32.1	26.1	39.1
65	31.7	22.5	33.7
92	20.7	15.0	22.6



#34
Hexachlorobutadiene
Concen: 46.701 ng
RT: 10.799 min Scan# 1311
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:	225	Resp:	298805
Ion	Ratio	Lower	Upper
225	100		
223	62.8	51.0	76.6
227	63.4	51.4	77.2









#35
Caprolactam
Concen: 44.039 ng
RT: 11.469 min Scan# 1425
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:113 Resp: 162257
Ion Ratio Lower Upper
113 100
55 138.2 98.9 138.9
56 109.2 71.7 111.7



#36
4-Chloro-3-methylphenol
Concen: 43.886 ng
RT: 11.787 min Scan# 1479
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:107 Resp: 569744
Ion Ratio Lower Upper
107 100
144 23.7 21.1 31.7
142 73.2 66.0 99.0









#37
2-Methylnaphthalene
Concen: 42.777 ng
RT: 12.140 min Scan# 1539
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:142 Resp: 1145375
Ion Ratio Lower Upper
142 100
141 89.8 71.8 107.6
115 39.8 31.6 47.4



#38
1-Methylnaphthalene
Concen: 42.579 ng
RT: 12.363 min Scan# 1577
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:142 Resp: 1083433
Ion Ratio Lower Upper
142 100
141 92.9 73.3 109.9
116 4.2 3.5 5.3









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.334 min Scan# 1912
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:164 Resp: 410806
Ion Ratio Lower Upper
164 100
162 101.3 80.2 120.2
160 46.1 34.5 51.7



#40
1,2,4,5-Tetrachlorobenzene
Concen: 45.618 ng
RT: 12.510 min Scan# 1602
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:216 Resp: 490489
Ion Ratio Lower Upper
216 100
214 78.7 61.9 92.9
179 23.6 16.8 25.2
108 23.7 14.9 22.3#









#41
Hexachlorocyclopentadiene
Concen: 40.850 ng
RT: 12.481 min Scan# 1597
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:	237	Resp:	232948
Ion Ratio	Lower	Upper	
237	100		
235	63.5	44.5	84.5
272	12.2	0.0	32.5



#42
2,4,6-Tribromophenol
Concen: 94.994 ng
RT: 15.834 min Scan# 2167
Delta R.T. -0.012 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:	330	Resp:	407847
Ion Ratio	Lower	Upper	
330	100		
332	97.1	77.8	116.8
141	47.0	25.4	38.2









#43
2,4,6-Trichlorophenol
Concen: 48.046 ng
RT: 12.763 min Scan# 1645
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:196 Resp: 352731
Ion Ratio Lower Upper
196 100
198 95.5 79.5 119.3
200 31.5 25.0 37.6



#44
2,4,5-Trichlorophenol
Concen: 46.381 ng
RT: 12.840 min Scan# 1658
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:196 Resp: 376914
Ion Ratio Lower Upper
196 100
198 97.4 74.7 112.1
97 61.3 37.4 56.0#
132 32.7 20.0 30.0#









#45
2-Fluorobiphenyl
Concen: 88.356 ng
RT: 12.957 min Scan# 1678
Delta R.T. 0.000 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:172 Resp: 2426078
Ion Ratio Lower Upper
172 100
171 35.9 27.9 41.9
170 23.8 18.6 27.8



#46
1,1'-Biphenyl
Concen: 43.655 ng
RT: 13.163 min Scan# 1713
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:154 Resp: 1357423
Ion Ratio Lower Upper
154 100
153 40.8 20.4 60.4
76 14.5 0.0 30.8









#47
2-Chloronaphthalene
Concen: 43.745 ng
RT: 13.204 min Scan# 1720
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:162 Resp: 1047476
Ion Ratio Lower Upper
162 100
127 45.0 31.4 47.0
164 32.7 26.6 39.8



#48
2-Nitroaniline
Concen: 43.438 ng
RT: 13.422 min Scan# 1757
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion: 65 Resp: 367068
Ion Ratio Lower Upper
65 100
92 70.3 57.7 86.5
138 98.9 92.1 138.1









#49
Acenaphthylene
Concen: 43.701 ng
RT: 14.051 min Scan# 1864
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:152 Resp: 1687329
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.2
153 13.0 10.6 16.0



#50
Dimethylphthalate
Concen: 43.383 ng
RT: 13.810 min Scan# 1823
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:163 Resp: 1297301
Ion Ratio Lower Upper
163 100
194 3.9 3.4 5.2
164 10.2 8.2 12.2









#51
2,6-Dinitrotoluene
Concen: 44.620 ng
RT: 13.928 min Scan# 1843
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:165 Resp: 297737
Ion Ratio Lower Upper
165 100
63 65.6 41.3 61.9#
89 70.0 44.8 67.2#



#52
Acenaphthene
Concen: 42.871 ng
RT: 14.398 min Scan# 1923
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:154 Resp: 1007399
Ion Ratio Lower Upper
154 100
153 115.9 89.4 134.2
152 54.9 43.0 64.4









#53
3-Nitroaniline
Concen: 43.759 ng
RT: 14.257 min Scan# 1899
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:138 Resp: 323705
Ion Ratio Lower Upper
138 100
108 11.3 8.4 12.6
92 127.0 87.8 131.8



#54
2,4-Dinitrophenol
Concen: 47.281 ng
RT: 14.469 min Scan# 1935
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:184 Resp: 164984
Ion Ratio Lower Upper
184 100
63 79.2 47.0 70.6#
154 73.3 52.3 78.5









#55
Dibenzofuran
Concen: 42.493 ng
RT: 14.734 min Scan# 1980
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:168 Resp: 1570597
Ion Ratio Lower Upper
168 100
139 44.2 31.0 46.6
169 13.4 10.8 16.2



#56
4-Nitrophenol
Concen: 43.293 ng
RT: 14.581 min Scan# 1954
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22


Tgt Ion:139 Resp: 278078
Ion Ratio Lower Upper
139 100
109 87.4 72.3 112.3
65 106.2 72.5 112.5








#57
2,4-Dinitrotoluene
Concen: 45.317 ng
RT: 14.722 min Scan# 1978
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040







Tgt Ion:165 Resp: 405533
Ion Ratio Lower Upper
165 100
63 56.4 32.3 48.5#
89 85.7 55.9 83.9#
182 2.7 2.2 3.4



#58
Fluorene
Concen: 42.639 ng
RT: 15.387 min Scan# 2091
Delta R.T. -0.012 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:166 Resp: 1259680
Ion Ratio Lower Upper
166 100
165 98.8 78.0 117.0
167 13.7 10.5 15.7









#59
2,3,4,6-Tetrachlorophenol
Concen: 45.572 ng
RT: 14.969 min Scan# 2020
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:232	Resp:	310359
Ion Ratio	Lower	Upper
232	100	
131	60.7	32.2 48.4#
130	3.3	2.0 3.0#
166	36.8	19.4 29.2#



#60
Diethylphthalate
Concen: 43.637 ng
RT: 15.175 min Scan# 2055
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:149	Resp:	1370546
Ion Ratio	Lower	Upper
149	100	
177	20.1	17.5 26.3
150	12.5	9.9 14.9









#61
4-Chlorophenyl-phenylether
Concen: 44.100 ng
RT: 15.387 min Scan# 2091
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:204 Resp: 590011
Ion Ratio Lower Upper
204 100
206 32.7 26.6 40.0
141 71.1 48.6 73.0



#62
4-Nitroaniline
Concen: 40.322 ng
RT: 15.428 min Scan# 2098
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:138 Resp: 316514
Ion Ratio Lower Upper
138 100
92 57.8 31.2 71.2
108 94.1 78.6 118.6









#63
Azobenzene
Concen: 41.167 ng
RT: 15.681 min Scan# 2141
Delta R.T. -0.012 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion: 77 Resp: 1507487




Ion	Ratio	Lower	Upper
77	100		
182	20.1	6.0	46.0
105	19.3	0.0	38.8
51	25.3	8.8	48.8




#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.092 min Scan# 2381
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion: 188 Resp: 790865

Ion	Ratio	Lower	Upper
188	100		
94	10.5	5.6	8.4#
80	12.2	6.3	9.5#









#65
4,6-Dinitro-2-methylphenol
Concen: 47.558 ng
RT: 15.487 min Scan# 2108
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:198 Resp: 229816
Ion Ratio Lower Upper
198 100
51 45.8 14.0 54.0
105 54.2 19.5 59.5



#66
n-Nitrosodiphenylamine
Concen: 43.398 ng
RT: 15.610 min Scan# 2129
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:169 Resp: 1078798
Ion Ratio Lower Upper
169 100
168 67.5 54.4 81.6
167 37.1 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 46.061 ng
RT: 16.287 min Scan# 2244
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:248 Resp: 349525
Ion Ratio Lower Upper
248 100
250 98.7 79.8 119.8
141 92.8 57.8 86.6#



#68
Hexachlorobenzene
Concen: 45.151 ng
RT: 16.392 min Scan# 2262
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:284 Resp: 389896
Ion Ratio Lower Upper
284 100
142 44.2 26.9 40.3#
249 33.4 24.5 36.7









#69
Atrazine
Concen: 42.810 ng
RT: 16.575 min Scan# 2293
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:200 Resp: 332077
Ion Ratio Lower Upper
200 100
173 26.7 4.1 44.1
215 45.4 27.7 67.7



#70
Pentachlorophenol
Concen: 45.971 ng
RT: 16.745 min Scan# 2322
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:266 Resp: 251872
Ion Ratio Lower Upper
266 100
268 64.0 51.6 77.4
264 63.1 50.3 75.5









#71
Phenanthrene
Concen: 42.295 ng
RT: 17.134 min Scan# 2388
Delta R.T. -0.012 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:178 Resp: 1886773
Ion Ratio Lower Upper
178 100
176 19.6 15.1 22.7
179 15.6 12.4 18.6



#72
Anthracene
Concen: 43.887 ng
RT: 17.228 min Scan# 2404
Delta R.T. -0.006 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:178 Resp: 1891766
Ion Ratio Lower Upper
178 100
176 18.8 14.9 22.3
179 15.9 12.2 18.2









#73
Carbazole
Concen: 42.175 ng
RT: 17.504 min Scan# 2451
Delta R.T. -0.012 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:167 Resp: 1854331
Ion Ratio Lower Upper
167 100
166 21.9 17.1 25.7
139 15.2 10.7 16.1



#74
Di-n-butylphthalate
Concen: 46.296 ng
RT: 18.069 min Scan# 2547
Delta R.T. -0.012 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:149 Resp: 2319079
Ion Ratio Lower Upper
149 100
150 9.1 7.3 10.9
104 6.1 4.5 6.7









#75
Fluoranthene
Concen: 42.830 ng
RT: 19.116 min Scan# 2725
Delta R.T. -0.012 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:202 Resp: 2144505
Ion Ratio Lower Upper
202 100
101 12.1 0.0 27.6
203 17.3 0.0 37.8



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.198 min Scan# 3079
Delta R.T. -0.012 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:240 Resp: 743494
Ion Ratio Lower Upper
240 100
120 12.2 6.5 9.7#
236 26.1 20.6 31.0









#77
Benzidine
Concen: 44.132 ng
RT: 19.304 min Scan# 2757
Delta R.T. -0.012 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:184 Resp: 821063
Ion Ratio Lower Upper
184 100
185 14.2 11.0 16.4
183 11.8 9.5 14.3



#78
Pyrene
Concen: 44.458 ng
RT: 19.469 min Scan# 2785
Delta R.T. -0.012 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:202 Resp: 2207873
Ion Ratio Lower Upper
202 100
200 21.4 16.3 24.5
203 17.3 14.2 21.2









#79
Terphenyl-d14
Concen: 91.054 ng
RT: 19.675 min Scan# 2820
Delta R.T. -0.012 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:244 Resp: 3378520
Ion Ratio Lower Upper
244 100
212 7.8 6.0 9.0
122 11.9 7.1 10.7#



#80
Butylbenzylphthalate
Concen: 50.088 ng
RT: 20.357 min Scan# 2936
Delta R.T. -0.012 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:149 Resp: 1093800
Ion Ratio Lower Upper
149 100
91 77.3 54.6 81.8
206 18.4 16.7 25.1









#81
Benzo(a)anthracene
Concen: 43.921 ng
RT: 21.180 min Scan# 3076
Delta R.T. -0.018 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:228 Resp: 2134141
Ion Ratio Lower Upper
228 100
226 27.6 21.6 32.4
229 19.8 15.8 23.6



#82
3,3'-Dichlorobenzidine
Concen: 47.872 ng
RT: 21.122 min Scan# 3066
Delta R.T. -0.018 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:252 Resp: 610665
Ion Ratio Lower Upper
252 100
254 64.1 50.6 76.0
126 13.0 7.0 10.6#





#83

Chrysene

Concen: 43.474 ng

RT: 21.233 min Scan# 3085

Delta R.T. -0.018 min

Lab File: BP006854.D

Acq: 24 Aug 2021 19:22

Tgt Ion:228 Resp: 2047868





Ion	Ratio	Lower	Upper
228	100		
226	30.6	23.8	35.6
229	20.1	15.4	23.2

Instrument :

BNA_P

ClientSampleId :

SSTDCCC040



#84

Bis(2-ethylhexyl)phthalate

Concen: 48.352 ng

RT: 21.122 min Scan# 3066




Delta R.T. -0.012 min


Lab File: BP006854.D

Acq: 24 Aug 2021 19:22

Tgt Ion:149 Resp: 1586758

Ion	Ratio	Lower	Upper
149	100		
167	26.4	21.9	32.9
279	4.0	4.3	6.5









#85
Di-n-octyl phthalate
Concen: 51.341 ng
RT: 22.016 min Scan# 3218
Delta R.T. -0.018 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:149 Resp: 2777016
Ion Ratio Lower Upper
149 100
167 1.5 1.4 2.0
43 7.9 5.6 8.4



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.504 min Scan# 3471
Delta R.T. -0.023 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:264 Resp: 792977
Ion Ratio Lower Upper
264 100
260 24.2 18.6 28.0
265 21.8 17.6 26.4









#87
Indeno(1,2,3-cd)pyrene
Concen: 43.822 ng
RT: 25.892 min Scan# 3877
Delta R.T. -0.029 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:276 Resp: 2519748
Ion Ratio Lower Upper
276 100
138 27.8 15.3 22.9#
277 25.5 20.4 30.6



#88
Benzo(b)fluoranthene
Concen: 42.130 ng
RT: 22.804 min Scan# 3352
Delta R.T. -0.023 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:252 Resp: 2171261
Ion Ratio Lower Upper
252 100
253 22.2 17.5 26.3
125 11.5 6.2 9.4#









#89
Benzo(k)fluoranthene
Concen: 42.922 ng
RT: 22.851 min Scan# 3360
Delta R.T. -0.023 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22




Instrument :
BNA_P
ClientSampleId :
SSTDCCC040


Tgt Ion:252 Resp: 2123848
Ion Ratio Lower Upper
252 100
253 22.3 17.6 26.4
125 11.7 6.0 9.0#



#90
Benzo(a)pyrene
Concen: 43.722 ng
RT: 23.404 min Scan# 3454
Delta R.T. -0.023 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:252 Resp: 2021100
Ion Ratio Lower Upper
252 100
253 21.6 17.9 26.9
125 12.5 6.5 9.7#









#91
Dibenzo(a,h)anthracene
Concen: 43.709 ng
RT: 25.910 min Scan# 3880
Delta R.T. -0.029 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Instrument :
BNA_P
ClientSampleId :
SSTDCCC040

Tgt Ion:278 Resp: 2172982
Ion Ratio Lower Upper
278 100
139 17.4 9.3 13.9#
279 24.2 18.8 28.2



#92
Benzo(g,h,i)perylene
Concen: 43.784 ng
RT: 26.621 min Scan# 4001
Delta R.T. -0.035 min
Lab File: BP006854.D
Acq: 24 Aug 2021 19:22

Tgt Ion:276 Resp: 2085859
Ion Ratio Lower Upper
276 100
277 23.8 18.9 28.3
138 24.1 12.6 18.8#

