

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082422\
 Data File : BP011562.D
 Acq On : 25 Aug 2022 13:03
 Operator : CG/JU
 Sample : N4320-09
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Aug 26 01:04:26 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 25 02:14:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.546	152	57334	20.000	ng/ul	0.00
20) Naphthalene-d8	10.334	136	247365	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.222	164	147361	20.000	ng/ul	0.00
64) Phenanthrene-d10	16.992	188	292778	20.000	ng/ul	0.00
79) Chrysene-d12	21.122	240	270268	20.000	ng/ul	0.00
88) Perylene-d12	23.416	264	281591	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.064	96	8187	4.431	ng/uL	0.00
4) Pyridine-d5	3.481	84	25272	5.343	ng/ul	0.00
7) Phenol-d5	6.734	99	36691	6.964	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.899	67	98319	28.723	ng/ul	0.00
11) 2-Chlorophenol-d4	7.081	132	96493	24.900	ng/ul	0.00
15) 4-Methylphenol-d8	8.269	113	65124	15.443	ng/ul	0.00
21) Nitrobenzene-d5	8.716	128	57703	32.402	ng/ul	0.00
24) 2-Nitrophenol-d4	9.428	143	58125	32.070	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	9.963	165	93927	28.075	ng/ul	0.00
31) 4-Chloroaniline-d4	10.493	131	106167	19.064	ng/ul	0.00
46) Dimethylphthalate-d6	13.645	166	366615	35.175	ng/ul	0.00
49) Acenaphthylene-d8	13.910	160	389737	33.645	ng/ul	0.00
54) 4-Nitrophenol-d4	14.463	143	8207	4.087	ng/ul	0.00
60) Fluorene-d10	15.228	176	310565	35.419	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.363	200	50278	31.781	ng/ul	0.00
73) Anthracene-d10	17.092	188	533508	41.144	ng/ul	0.00
81) Pyrene-d10	19.357	212	575422	42.081	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.268	264	555676	40.567	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082422\
Data File : BP011562.D
Acq On : 25 Aug 2022 13:03
Operator : CG/JU
Sample : N4320-09
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :

Quant Time: Aug 26 01:04:26 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081622.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Aug 25 02:14:54 2022
Response via : Initial Calibration

