

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082422\
 Data File : BP011571.D
 Acq On : 25 Aug 2022 20:12
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Aug 26 01:08:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 25 02:14:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.546	152	68499	20.000	ng/u1	0.00
20) Naphthalene-d8	10.334	136	297792	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.216	164	181589	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.986	188	365432	20.000	ng/u1	0.00
79) Chrysene-d12	21.115	240	358066	20.000	ng/u1	-0.01
88) Perylene-d12	23.404	264	366301	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.063	96	17033	7.717	ng/uL	0.00
4) Pyridine-d5	3.469	84	103642	18.339	ng/u1	0.00
7) Phenol-d5	6.734	99	113728	18.067	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.899	67	76292	18.655	ng/u1	0.00
11) 2-Chlorophenol-d4	7.081	132	90904	19.635	ng/u1	0.00
15) 4-Methylphenol-d8	8.269	113	92234	18.306	ng/u1	0.00
21) Nitrobenzene-d5	8.710	128	45113	21.043	ng/u1	0.00
24) 2-Nitrophenol-d4	9.422	143	45163	20.699	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	9.957	165	82969	20.600	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.487	131	127064	18.952	ng/u1	0.00
46) Dimethylphthalate-d6	13.645	166	257725	20.067	ng/u1	0.00
49) Acenaphthylene-d8	13.904	160	292552	20.495	ng/u1	0.00
54) 4-Nitrophenol-d4	14.451	143	44608	18.027	ng/u1	0.00
60) Fluorene-d10	15.222	176	217450	20.125	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.357	200	37894	19.191	ng/u1	-0.01
73) Anthracene-d10	17.086	188	331999	20.513	ng/u1	0.00
81) Pyrene-d10	19.351	212	363601	20.070	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.251	264	360857	20.252	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.099	88	17411	7.565	ng/uL	92
5) Pyridine	3.493	79	104689	18.558	ng/u1	98
6) Benzaldehyde	6.704	77	64755	22.416	ng/u1	99
8) Phenol	6.757	94	117481	18.161	ng/u1	95
10) Bis(2-Chloroethyl)ether	6.987	93	94203	18.669	ng/u1	99
12) 2-Chlorophenol	7.110	128	96859	19.442	ng/u1	96
13) 2-Methylphenol	7.998	108	85037	17.844	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.075	45	76777	10.271	ng/u1	98
16) Acetophenone	8.375	105	156231	18.640	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.363	70	79571	19.086	ng/u1	97
18) 4-Methylphenol	8.334	108	96073	18.210	ng/u1	92
19) Hexachloroethane	8.604	117	45367	20.076	ng/u1	93
22) Nitrobenzene	8.751	77	124993	20.684	ng/u1	98
23) Isophorone	9.281	82	214968	20.265	ng/u1	99
25) 2-Nitrophenol	9.457	139	53190	21.366	ng/u1	97
26) 2,4-Dimethylphenol	9.528	107	119972	20.297	ng/u1	97
27) Bis(2-Chloroethoxy)met...	9.769	93	124543	19.424	ng/u1	96
29) 2,4-Dichlorophenol	9.987	162	83273	20.203	ng/u1	96
30) Naphthalene	10.381	128	312337	19.864	ng/u1	99
32) 4-Chloroaniline	10.510	127	134687	19.390	ng/u1	98
33) Hexachlorobutadiene	10.657	225	56072	21.369	ng/u1	99
34) Caprolactam	11.310	113	26026	17.414	ng/u1	97
35) 4-Chloro-3-methylphenol	11.651	107	104189	19.993	ng/u1	97
36) 2-Methylnaphthalene	12.010	142	202256	19.771	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082422\
 Data File : BP011571.D
 Acq On : 25 Aug 2022 20:12
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Aug 26 01:08:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 25 02:14:54 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.234	142	204962	19.754	ng/ul	94
39) 1,2,4,5-Tetrachloroben...	12.381	216	93655	20.320	ng/ul	94
40) Hexachlorocyclopentadiene	12.357	237	48328	16.901	ng/ul	97
41) 2,4,6-Trichlorophenol	12.639	196	62328	21.011	ng/ul	96
42) 2,4,5-Trichlorophenol	12.710	196	67187	20.347	ng/ul	100
43) 1,1'-Biphenyl	13.045	154	269841	19.943	ng/ul	98
44) 2-Chloronaphthalene	13.081	162	204080	20.016	ng/ul	99
45) 2-Nitroaniline	13.310	65	71758	20.621	ng/ul	97
47) Dimethylphthalate	13.692	163	265935	20.034	ng/ul	99
48) 2,6-Dinitrotoluene	13.816	165	48925	20.500	ng/ul	98
50) Acenaphthylene	13.933	152	344842	20.335	ng/ul	98
51) 3-Nitroaniline	14.145	138	47400	18.234	ng/ul	99
52) Acenaphthene	14.281	153	228238	20.094	ng/ul	99
53) 2,4-Dinitrophenol	14.357	184	24469	16.751	ng/ul	94
55) 4-Nitrophenol	14.463	109	57610	19.540	ng/ul	88
56) Dibenzofuran	14.622	168	314889	20.103	ng/ul	99
57) 2,4-Dinitrotoluene	14.610	165	74420	20.830	ng/ul	91
58) 2,3,4,6-Tetrachlorophenol	14.851	232	55752	21.505	ng/ul	95
59) Diethylphthalate	15.069	149	296911	20.422	ng/ul	99
61) Fluorene	15.275	166	262090	20.453	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.280	204	120176	20.760	ng/ul	99
63) 4-Nitroaniline	15.316	138	42270	16.834	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.375	198	39463	19.320	ng/ul#	88
67) N-Nitrosodiphenylamine	15.498	169	220430	20.560	ng/ul	99
68) 4-Bromophenyl-phenylether	16.180	248	68042	20.858	ng/ul	95
69) Hexachlorobenzene	16.280	284	77969	20.739	ng/ul	97
70) Atrazine	16.475	200	79169	19.787	ng/ul	98
71) Pentachlorophenol	16.639	266	45035	19.319	ng/ul	99
72) Phenanthrene	17.033	178	397139	20.078	ng/ul	99
74) Anthracene	17.122	178	407021	20.397	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	12.998	216	97121	20.675	ng/uL	100
76) Pentachlorobenzene	14.533	250	93640	21.008	ng/uL	96
77) Carbazole	17.404	167	369495	19.686	ng/ul	99
78) Di-n-butylphthalate	17.980	149	490967	21.241	ng/ul	99
80) Fluoranthene	19.021	202	447144	20.070	ng/ul	99
82) Pyrene	19.380	202	466570	19.931	ng/ul	98
83) Butylbenzylphthalate	20.280	149	219810	20.932	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.045	252	135590	20.003	ng/ul	97
85) Benzo(a)anthracene	21.104	228	445366	20.415	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.045	149	325471	20.944	ng/ul	99
87) Chrysene	21.157	228	438853	20.479	ng/ul	99
89) Di-n-octyl phthalate	21.939	149	562889	19.116	ng/ul	100
90) Benzo(b)fluoranthene	22.709	252	464328	20.378	ng/ul	99
91) Benzo(k)fluoranthene	22.757	252	435466	20.027	ng/ul	97
93) Benzo(a)pyrene	23.304	252	451579	20.300	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	25.751	276	503601	20.296	ng/ul	98
95) Dibenzo(a,h)anthracene	25.774	278	429966	20.125	ng/ul	97
96) Benzo(g,h,i)perylene	26.474	276	419846	19.892	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082422\
 Data File : BP011571.D
 Acq On : 25 Aug 2022 20:12
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Aug 26 01:08:37 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP081622.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 25 02:14:54 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.565 ng/uL
RT: 3.099 min Scan# 19
Delta R.T. -0.000 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 88 Resp: 17411
Ion Ratio Lower Upper
88 100
43 36.6 29.0 43.4
58 71.2 64.7 97.1



#5
Pyridine
Concen: 18.558 ng/uL
RT: 3.493 min Scan# 86
Delta R.T. -0.000 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Tgt Ion: 79 Resp: 104689
Ion Ratio Lower Upper
79 100
52 89.0 70.2 105.2
51 47.9 36.1 54.1









#6
Benzaldehyde
Concen: 22.416 ng/ul
RT: 6.704 min Scan# 61
Delta R.T. -0.000 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 64755
Ion Ratio Lower Upper
77 100
105 88.4 72.0 108.0
106 85.2 67.4 101.0



#8
Phenol
Concen: 18.161 ng/ul
RT: 6.757 min Scan# 641
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Tgt Ion: 94 Resp: 117481
Ion Ratio Lower Upper
94 100
65 37.5 28.9 43.3
66 54.1 39.4 59.0





#10

Bis(2-Chloroethyl)ether

Concen: 18.669 ng/ul

RT: 6.987 min Scan# 61

Delta R.T. -0.006 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12

Instrument :

BNA_P

ClientSampleId :





Tgt Ion: 93 Resp: 94203

Ion Ratio Lower Upper

93 100

63 84.2 67.8 101.6

95 31.9 25.0 37.6



#12

2-Chlorophenol

Concen: 19.442 ng/ul

RT: 7.110 min Scan# 701

Delta R.T. -0.006 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12




Tgt Ion:128 Resp: 96859


Ion Ratio Lower Upper

128 100

64 53.3 40.2 60.2

130 33.1 24.6 37.0





#13

2-Methylphenol

Concen: 17.844 ng/ul

RT: 7.998 min Scan# 81

Delta R.T. -0.006 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12

Instrument :

BNA_P





ClientSampleId :

Tgt Ion:108 Resp: 85037

Ion Ratio Lower Upper

108 100

107 91.3 73.2 109.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 10.271 ng/ul

RT: 8.075 min Scan# 865

Delta R.T. -0.006 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12




Tgt Ion: 45 Resp: 76777


Ion Ratio Lower Upper

45 100

77 14.6 11.4 17.0

79 9.4 8.6 13.0









#16
Acetophenone
Concen: 18.640 ng/ul
RT: 8.375 min Scan# 918
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:105 Resp: 156231
Ion Ratio Lower Upper
105 100
77 82.4 66.4 99.6
51 38.6 28.2 42.4



#17
N-Nitroso-di-n-propylamine
Concen: 19.086 ng/ul
RT: 8.363 min Scan# 914
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12


Tgt Ion: 70 Resp: 79571
Ion Ratio Lower Upper
70 100
42 74.4 57.0 85.4
101 9.8 7.6 11.4
130 18.9 15.2 22.8








#18
4-Methylphenol
Concen: 18.210 ng/ul
RT: 8.334 min Scan# 909
Delta R.T. -0.000 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:108 Resp: 96073
Ion Ratio Lower Upper
108 100
107 117.5 87.7 131.5



#19
Hexachloroethane
Concen: 20.076 ng/ul
RT: 8.604 min Scan# 955
Delta R.T. -0.012 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Tgt Ion:117 Resp: 45367
Ion Ratio Lower Upper
117 100
201 75.3 64.2 96.4
199 45.1 41.0 61.6









#22
Nitrobenzene
Concen: 20.684 ng/ul
RT: 8.751 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12




Instrument :
BNA_P
ClientSampleId :


Tgt Ion: 77 Resp: 124993
Ion Ratio Lower Upper
77 100
123 38.6 30.4 45.6
65 15.8 11.5 17.3



#23
Isophorone
Concen: 20.265 ng/ul
RT: 9.281 min Scan# 1070
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Tgt Ion: 82 Resp: 214968
Ion Ratio Lower Upper
82 100
95 8.6 7.1 10.7
138 16.4 12.6 19.0









#25
2-Nitrophenol
Concen: 21.366 ng/ul
RT: 9.457 min Scan# 1100
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:139	Resp:	53190
Ion Ratio	Lower	Upper
139	100	
65	65.9	50.4 75.6
109	42.5	35.5 53.3



#26
2,4-Dimethylphenol
Concen: 20.297 ng/ul
RT: 9.528 min Scan# 1112
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Tgt Ion:107	Resp:	119972
Ion Ratio	Lower	Upper
107	100	
121	44.0	36.6 55.0
122	75.5	58.7 88.1









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.424 ng/ul
 RT: 9.769 min Scan# 1155
 Delta R.T. -0.006 min
 Lab File: BP011571.D
 Acq: 25 Aug 2022 20:12




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion: 93 Resp: 124543
 Ion Ratio Lower Upper
 93 100
 95 33.5 24.8 37.2
 123 11.2 8.2 12.4



#29
 2,4-Dichlorophenol
 Concen: 20.203 ng/ul
 RT: 9.987 min Scan# 1190
 Delta R.T. -0.006 min
 Lab File: BP011571.D
 Acq: 25 Aug 2022 20:12

Tgt Ion: 162 Resp: 83273
 Ion Ratio Lower Upper
 162 100
 164 66.6 50.2 75.2
 98 42.5 32.9 49.3









#30
Naphthalene
Concen: 19.864 ng/ul
RT: 10.381 min Scan# 1257
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:128 Resp: 312337
Ion Ratio Lower Upper
128 100
129 10.9 9.0 13.6
127 13.4 10.3 15.5



#32
4-Chloroaniline
Concen: 19.390 ng/ul
RT: 10.510 min Scan# 1279
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Tgt Ion:127 Resp: 134687
Ion Ratio Lower Upper
127 100
129 32.3 25.1 37.7





#33

Hexachlorobutadiene

Concen: 21.369 ng/ul

RT: 10.657 min Scan# 11

Delta R.T. -0.006 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:225 Resp: 56072

Ion Ratio Lower Upper

225 100

223 60.3 47.8 71.8

227 63.7 50.0 75.0



#34

Caprolactam

Concen: 17.414 ng/ul

RT: 11.310 min Scan# 1415

Delta R.T. -0.000 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12




Tgt Ion:113 Resp: 26026


Ion Ratio Lower Upper

113 100

55 186.9 153.4 230.2

56 152.7 119.6 179.4









#35
4-Chloro-3-methylphenol
Concen: 19.993 ng/ul
RT: 11.651 min Scan# 1475
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:107 Resp: 104189
Ion Ratio Lower Upper
107 100
144 23.8 18.3 27.5
142 72.7 55.8 83.8



#36
2-Methylnaphthalene
Concen: 19.771 ng/ul
RT: 12.010 min Scan# 1534
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Tgt Ion:142 Resp: 202256
Ion Ratio Lower Upper
142 100
141 91.2 73.4 110.0









#37
1-Methylnaphthalene
Concen: 19.754 ng/ul
RT: 12.234 min Scan# 1574
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:142 Resp: 204962
Ion Ratio Lower Upper
142 100
141 96.8 73.0 109.6
116 5.4 4.0 6.0



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.320 ng/ul
RT: 12.381 min Scan# 1597
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Tgt Ion:216 Resp: 93655
Ion Ratio Lower Upper
216 100
214 81.8 60.4 90.6
179 26.7 20.0 30.0
108 27.8 20.2 30.4





#40

Hexachlorocyclopentadiene

Concen: 16.901 ng/ul

RT: 12.357 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:237 Resp: 48328

Ion Ratio Lower Upper

237 100

235 61.9 50.9 76.3

272 12.0 11.0 16.4



#41

2,4,6-Trichlorophenol

Concen: 21.011 ng/ul

RT: 12.639 min Scan# 1641

Delta R.T. -0.000 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12




Tgt Ion:196 Resp: 62328


Ion Ratio Lower Upper

196 100

198 96.6 75.6 113.4

200 33.9 23.1 34.7





#42

2,4,5-Trichlorophenol

Concen: 20.347 ng/ul

RT: 12.710 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:196 Resp: 67187

Ion Ratio Lower Upper

196 100

198 94.0 75.6 113.4

200 29.9 23.8 35.6



#43

1,1'-Biphenyl

Concen: 19.943 ng/ul

RT: 13.045 min Scan# 1710

Delta R.T. -0.006 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12




Tgt Ion:154 Resp: 269841


Ion Ratio Lower Upper

154 100

153 39.9 32.2 48.4

76 12.6 11.6 17.4









#44
 2-Chloronaphthalene
 Concen: 20.016 ng/ul
 RT: 13.081 min Scan# 1716
 Delta R.T. -0.006 min
 Lab File: BP011571.D
 Acq: 25 Aug 2022 20:12

Instrument :
 BNA_P
 ClientSampleId :

Tgt Ion: 162 Resp: 204080




Ion	Ratio	Lower	Upper
162	100		
164	33.1	26.4	39.6
127	47.9	37.4	56.2




#45
 2-Nitroaniline
 Concen: 20.621 ng/ul
 RT: 13.310 min Scan# 1755
 Delta R.T. -0.000 min
 Lab File: BP011571.D
 Acq: 25 Aug 2022 20:12

Tgt Ion: 65 Resp: 71758

Ion	Ratio	Lower	Upper
65	100		
92	60.4	50.8	76.2
138	85.5	66.7	100.1









#47
Dimethylphthalate
Concen: 20.034 ng/ul
RT: 13.692 min Scan# 1820
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:163 Resp: 265935
Ion Ratio Lower Upper
163 100
194 3.7 2.6 3.8
164 10.2 8.3 12.5



#48
2,6-Dinitrotoluene
Concen: 20.500 ng/ul
RT: 13.816 min Scan# 1841
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12


Tgt Ion:165 Resp: 48925
Ion Ratio Lower Upper
165 100
89 85.3 66.6 99.8
121 27.6 21.0 31.6








#50
Acenaphthylene
Concen: 20.335 ng/ul
RT: 13.933 min Scan# 1861
Delta R.T. -0.012 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:152 Resp: 344842
Ion Ratio Lower Upper
152 100
151 19.9 16.9 25.3
153 12.9 10.6 15.8



#51
3-Nitroaniline
Concen: 18.234 ng/ul
RT: 14.145 min Scan# 1897
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Tgt Ion:138 Resp: 47400
Ion Ratio Lower Upper
138 100
108 12.8 8.6 13.0
92 131.1 104.2 156.2









#52
Acenaphthene
Concen: 20.094 ng/ul
RT: 14.281 min Scan# 1920
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:153 Resp: 228238
Ion Ratio Lower Upper
153 100
152 49.1 37.7 56.5
154 86.2 69.4 104.0



#53
2,4-Dinitrophenol
Concen: 16.751 ng/ul
RT: 14.357 min Scan# 1933
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Tgt Ion:184 Resp: 24469
Ion Ratio Lower Upper
184 100
63 111.5 82.5 123.7
154 86.0 66.6 100.0









#55
4-Nitrophenol
Concen: 19.540 ng/ul
RT: 14.463 min Scan# 1951
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:109 Resp: 57610
Ion Ratio Lower Upper
109 100
139 81.9 76.0 114.0
65 112.1 98.8 148.2



#56
Dibenzofuran
Concen: 20.103 ng/ul
RT: 14.622 min Scan# 1978
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Tgt Ion:168 Resp: 314889
Ion Ratio Lower Upper
168 100
139 48.7 38.4 57.6









#57
2,4-Dinitrotoluene
Concen: 20.830 ng/ul
RT: 14.610 min Scan# 1978
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:165 Resp: 74420
Ion Ratio Lower Upper
165 100
63 83.2 60.2 90.2
182 2.6 1.8 2.8



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.505 ng/ul
RT: 14.851 min Scan# 2017
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12


Tgt Ion:232 Resp: 55752
Ion Ratio Lower Upper
232 100
131 73.7 55.4 83.0
130 3.8 3.6 5.4
166 40.9 31.0 46.4








#59
Diethylphthalate
Concen: 20.422 ng/ul
RT: 15.069 min Scan# 2054
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Instrument :
BNA_P
ClientSampleId :







Tgt Ion:149 Resp: 296911
Ion Ratio Lower Upper
149 100
177 19.1 15.7 23.5
150 12.8 10.1 15.1



#61
Fluorene
Concen: 20.453 ng/ul
RT: 15.275 min Scan# 2089
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Tgt Ion:166 Resp: 262090
Ion Ratio Lower Upper
166 100
165 100.0 78.2 117.2
167 13.1 10.3 15.5









#62
4-Chlorophenyl-phenylether
Concen: 20.760 ng/ul
RT: 15.280 min Scan# 2092
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:204 Resp: 120176
Ion Ratio Lower Upper
204 100
206 32.1 25.4 38.0
141 83.6 66.5 99.7



#63
4-Nitroaniline
Concen: 16.834 ng/ul
RT: 15.316 min Scan# 2096
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Tgt Ion:138 Resp: 42270
Ion Ratio Lower Upper
138 100
92 63.9 49.5 74.3
108 128.0 93.7 140.5









#66
4,6-Dinitro-2-methylphenol
Concen: 19.320 ng/ul
RT: 15.375 min Scan# 2106
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:198 Resp: 39463
Ion Ratio Lower Upper
198 100
182 6.0 3.9 5.9#
77 45.7 30.1 45.1#



#67
N-Nitrosodiphenylamine
Concen: 20.560 ng/ul
RT: 15.498 min Scan# 2127
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Tgt Ion:169 Resp: 220430
Ion Ratio Lower Upper
169 100
168 67.3 54.6 82.0
167 36.7 30.2 45.4









#68
4-Bromophenyl-phenylether
Concen: 20.858 ng/ul
RT: 16.180 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:248 Resp: 68042
Ion Ratio Lower Upper
248 100
250 99.5 82.8 124.2
141 119.5 90.8 136.2



#69
Hexachlorobenzene
Concen: 20.739 ng/ul
RT: 16.280 min Scan# 2260
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Tgt Ion:284 Resp: 77969
Ion Ratio Lower Upper
284 100
142 57.0 43.5 65.3
249 34.4 28.4 42.6









#70
Atrazine
Concen: 19.787 ng/ul
RT: 16.475 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:200 Resp: 79169
Ion Ratio Lower Upper
200 100
173 26.4 22.0 33.0
215 44.5 36.3 54.5



#71
Pentachlorophenol
Concen: 19.319 ng/ul
RT: 16.639 min Scan# 2321
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Tgt Ion:266 Resp: 45035
Ion Ratio Lower Upper
266 100
264 62.2 50.6 76.0
268 65.1 51.3 76.9





#72

Phenanthrene

Concen: 20.078 ng/ul

RT: 17.033 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:178 Resp: 397139

Ion Ratio Lower Upper

178 100

179 15.5 12.6 18.8

176 19.5 15.8 23.8



#74

Anthracene

Concen: 20.397 ng/ul

RT: 17.122 min Scan# 2403

Delta R.T. -0.006 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12




Tgt Ion:178 Resp: 407021


Ion Ratio Lower Upper

178 100

179 15.2 12.2 18.2

176 19.1 15.4 23.0





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.675 ng/uL

RT: 12.998 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP011571.D





Acq: 25 Aug 2022 20:12

Instrument :

BNA_P

ClientSampleId :

Tgt Ion:	216	Resp:	97121
Ion Ratio	Lower	Upper	
216	100		
214	77.3	61.8	92.6
218	47.0	37.4	56.0



#76

Pentachlorobenzene

Concen: 21.008 ng/uL




RT: 14.533 min Scan# 1963


Delta R.T. -0.006 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12

Tgt Ion:	250	Resp:	93640
Ion Ratio	Lower	Upper	
250	100		
248	61.3	51.7	77.5
252	65.2	50.0	75.0









#77
 Carbazole
 Concen: 19.686 ng/ul
 RT: 17.404 min Scan# 2454
 Delta R.T. -0.006 min
 Lab File: BP011571.D
 Acq: 25 Aug 2022 20:12




Instrument :
 BNA_P
 ClientSampleId :


Tgt Ion:167 Resp: 369495
 Ion Ratio Lower Upper
 167 100
 166 22.3 17.5 26.3
 139 15.9 12.1 18.1



#78
 Di-n-butylphthalate
 Concen: 21.241 ng/ul
 RT: 17.980 min Scan# 2549
 Delta R.T. -0.006 min
 Lab File: BP011571.D
 Acq: 25 Aug 2022 20:12

Tgt Ion:149 Resp: 490967
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.4 11.2
 104 7.3 5.5 8.3









#80
Fluoranthene
Concen: 20.070 ng/ul
RT: 19.021 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:202 Resp: 447144
Ion Ratio Lower Upper
202 100
101 11.9 9.5 14.3
100 9.0 7.9 11.9



#82
Pyrene
Concen: 19.931 ng/ul
RT: 19.380 min Scan# 2787
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Tgt Ion:202 Resp: 466570
Ion Ratio Lower Upper
202 100
101 13.1 11.4 17.2
100 11.5 9.4 14.0





#83

Butylbenzylphthalate

Concen: 20.932 ng/ul

RT: 20.280 min Scan# 2943

Delta R.T. -0.006 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:149 Resp: 219810

Ion Ratio Lower Upper

149 100

91 81.5 68.5 102.7

206 16.1 12.6 18.8



#84

3,3'-Dichlorobenzidine

Concen: 20.003 ng/ul

RT: 21.045 min Scan# 3070

Delta R.T. -0.006 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12




Tgt Ion:252 Resp: 135590


Ion Ratio Lower Upper

252 100

254 61.4 51.2 76.8

126 15.0 12.2 18.2





#85

Benzo(a)anthracene

Concen: 20.415 ng/ul

RT: 21.104 min Scan# 3080

Delta R.T. -0.006 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:228 Resp: 445366

Ion Ratio Lower Upper

228 100

229 19.3 16.0 24.0

226 27.7 22.6 33.8



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.944 ng/ul

RT: 21.045 min Scan# 3070

Delta R.T. -0.006 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12




Tgt Ion:149 Resp: 325471


Ion Ratio Lower Upper

149 100

167 24.4 19.8 29.8

279 3.2 2.7 4.1









#87
Chrysene
Concen: 20.479 ng/ul
RT: 21.157 min Scan# 3092
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12




Instrument :
BNA_P
ClientSampleId :


Tgt Ion:228 Resp: 438853
Ion Ratio Lower Upper
228 100
226 29.9 23.9 35.9
229 18.9 16.0 24.0



#89
Di-n-octyl phthalate
Concen: 19.116 ng/ul
RT: 21.939 min Scan# 3222
Delta R.T. -0.006 min
Lab File: BP011571.D
Acq: 25 Aug 2022 20:12

Tgt Ion:149 Resp: 562889





#90

Benzo(b)fluoranthene

Concen: 20.378 ng/ul

RT: 22.709 min Scan# 3358

Delta R.T. -0.006 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:252 Resp: 464328

Ion Ratio Lower Upper

252 100

253 21.8 17.5 26.3

125 10.7 9.0 13.4



#91

Benzo(k)fluoranthene

Concen: 20.027 ng/ul

RT: 22.757 min Scan# 3361

Delta R.T. -0.006 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12




Tgt Ion:252 Resp: 435466


Ion Ratio Lower Upper

252 100

253 21.2 17.9 26.9

125 11.0 9.8 14.8





#93

Benzo(a)pyrene

Concen: 20.300 ng/ul

RT: 23.304 min Scan# 34

Delta R.T. -0.006 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12

Instrument :

BNA_P

ClientSampleId :





Tgt Ion:252 Resp: 451579

Ion Ratio Lower Upper

252 100

253 21.8 16.8 25.2

125 11.7 9.9 14.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.296 ng/ul

RT: 25.751 min Scan# 3870

Delta R.T. -0.024 min

Lab File: BP011571.D

Acq: 25 Aug 2022 20:12




Tgt Ion:276 Resp: 503601


Ion Ratio Lower Upper

276 100

138 25.8 20.2 30.4

277 24.1 20.1 30.1









#95
 Dibenzo(a,h)anthracene
 Concen: 20.125 ng/ul
 RT: 25.774 min Scan# 3911
 Delta R.T. -0.018 min
 Lab File: BP011571.D
 Acq: 25 Aug 2022 20:12

Instrument :
 BNA_P
 ClientSampleId :

Tgt Ion:278 Resp: 429966
 Ion Ratio Lower Upper
 278 100
 139 16.4 15.4 23.0
 279 23.6 19.3 28.9



#96
 Benzo(g,h,i)perylene
 Concen: 19.892 ng/ul
 RT: 26.474 min Scan# 3993
 Delta R.T. -0.012 min
 Lab File: BP011571.D
 Acq: 25 Aug 2022 20:12

Tgt Ion:276 Resp: 419846
 Ion Ratio Lower Upper
 276 100
 138 23.4 19.4 29.2
 277 22.7 19.1 28.7

