

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082521\
 Data File : BP006874.D
 Acq On : 25 Aug 2021 10:00
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 25 11:28:02 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 25 11:27:41 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	70	0.00
2	1,4-Dioxane	40.000	39.324	1.7	70	0.00
3	Pyridine	40.000	37.148	7.1	62	0.00
4	n-Nitrosodimethylamine	40.000	34.957	12.6	59	0.00
5 S	2-Fluorophenol	80.000	81.833	-2.3	69	0.00
6	Aniline	40.000	37.779	5.6	60	0.00
7 S	Phenol-d6	80.000	76.985	3.8	63	0.00
8	2-Chlorophenol	40.000	40.407	-1.0	67	0.00
9	Benzaldehyde	40.000	36.024	9.9	62	0.00
10 C	Phenol	40.000	37.809	5.5	61	0.00
11	bis(2-Chloroethyl)ether	40.000	37.710	5.7	62	0.00
12	1,3-Dichlorobenzene	40.000	40.828	-2.1	69	0.00
13 C	1,4-Dichlorobenzene	40.000	40.506	-1.3	69	0.00
14	1,2-Dichlorobenzene	40.000	40.362	-0.9	67	0.00
15	Benzyl Alcohol	40.000	36.133	9.7	57	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	32.897	17.8	53	0.00
17	2-Methylphenol	40.000	39.184	2.0	63	0.00
18	Hexachloroethane	40.000	41.884	-4.7	71	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	36.747	8.1	58	0.00
20	3+4-Methylphenols	40.000	39.560	1.1	62	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	65	0.00
22	Acetophenone	40.000	40.268	-0.7	62	0.00
23 S	Nitrobenzene-d5	80.000	81.008	-1.3	63	0.00
24	Nitrobenzene	40.000	39.870	0.3	62	0.00
25	Isophorone	40.000	39.875	0.3	61	0.00
26 C	2-Nitrophenol	40.000	44.915	-12.3	68	0.00
27	2,4-Dimethylphenol	40.000	40.594	-1.5	63	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.025	2.4	61	0.00
29 C	2,4-Dichlorophenol	40.000	43.089	-7.7	65	0.00
30	1,2,4-Trichlorobenzene	40.000	41.525	-3.8	67	0.00
31	Naphthalene	40.000	40.525	-1.3	64	0.00
32	Benzoic acid	40.000	42.121	-5.3	65	0.00
33	4-Chloroaniline	40.000	39.832	0.4	61	0.00
34 C	Hexachlorobutadiene	40.000	44.417	-11.0	72	0.00
35	Caprolactam	40.000	40.529	-1.3	57	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.771	-4.4	62	0.00
37	2-Methylnaphthalene	40.000	40.819	-2.0	63	0.00
38	1-Methylnaphthalene	40.000	40.609	-1.5	63	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	61	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	42.471	-6.2	65	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.007	10.0	54	0.00
42 S	2,4,6-Tribromophenol	80.000	88.605	-10.8	64	0.00
43 C	2,4,6-Trichlorophenol	40.000	44.307	-10.8	65	0.00
44	2,4,5-Trichlorophenol	40.000	43.179	-7.9	63	0.00
45 S	2-Fluorobiphenyl	80.000	82.880	-3.6	63	0.00
46	1,1'-Biphenyl	40.000	41.339	-3.3	62	0.00
47	2-Chloronaphthalene	40.000	41.447	-3.6	62	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082521\
 Data File : BP006874.D
 Acq On : 25 Aug 2021 10:00
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 25 11:28:02 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 25 11:27:41 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	40.800	-2.0	57	0.00
49	Acenaphthylene	40.000	41.719	-4.3	61	0.00
50	Dimethylphthalate	40.000	40.999	-2.5	60	0.00
51	2,6-Dinitrotoluene	40.000	41.390	-3.5	59	0.00
52 C	Acenaphthene	40.000	41.130	-2.8	61	0.00
53	3-Nitroaniline	40.000	41.266	-3.2	58	0.00
54 P	2,4-Dinitrophenol	40.000	39.415	1.5	57	0.00
55	Dibenzofuran	40.000	41.068	-2.7	61	0.00
56 P	4-Nitrophenol	40.000	40.112	-0.3	55	0.00
57	2,4-Dinitrotoluene	40.000	43.198	-8.0	61	0.00
58	Fluorene	40.000	41.338	-3.3	61	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.412	-8.5	63	0.00
60	Diethylphthalate	40.000	41.953	-4.9	61	0.00
61	4-Chlorophenyl-phenylether	40.000	41.727	-4.3	62	0.00
62	4-Nitroaniline	40.000	37.695	5.8	52	0.00
63	Azobenzene	40.000	41.852	-4.6	60	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	59	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.814	-4.5	60	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.014	-2.5	59	0.00
67	4-Bromophenyl-phenylether	40.000	42.417	-6.0	62	0.00
68	Hexachlorobenzene	40.000	41.977	-4.9	62	0.00
69	Atrazine	40.000	41.895	-4.7	59	0.00
70 C	Pentachlorophenol	40.000	43.269	-8.2	60	0.00
71	Phenanthrene	40.000	41.163	-2.9	60	0.00
72	Anthracene	40.000	42.029	-5.1	60	0.00
73	Carbazole	40.000	41.343	-3.4	58	0.00
74	Di-n-butylphthalate	40.000	43.955	-9.9	62	0.00
75 C	Fluoranthene	40.000	42.055	-5.1	60	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	61	0.00
77	Benzydine	40.000	41.922	-4.8	55	0.00
78	Pyrene	40.000	40.339	-0.8	59	0.00
79 S	Terphenyl-d14	80.000	82.498	-3.1	62	0.00
80	Butylbenzylphthalate	40.000	43.812	-9.5	63	0.00
81	Benzo(a)anthracene	40.000	41.303	-3.3	61	0.00
82	3,3'-Dichlorobenzidine	40.000	41.580	-3.9	60	0.00
83	Chrysene	40.000	41.250	-3.1	61	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.208	-10.5	64	0.00
85 c	Di-n-octyl phthalate	40.000	47.085	-17.7	68	0.00
86 I	Perylene-d12	20.000	20.000	0.0	64	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.970	-4.9	65	0.00
88	Benzo(b)fluoranthene	40.000	40.672	-1.7	63	0.00
89	Benzo(k)fluoranthene	40.000	40.181	-0.5	62	0.00
90 C	Benzo(a)pyrene	40.000	41.702	-4.3	64	0.00
91	Dibenzo(a,h)anthracene	40.000	41.820	-4.6	65	0.00
92	Benzo(g,h,i)perylene	40.000	42.396	-6.0	66	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082521\
Data File : BP006874.D
Acq On : 25 Aug 2021 10:00
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Aug 25 11:28:02 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 25 11:27:41 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0