

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082521\
 Data File : BP006879.D
 Acq On : 25 Aug 2021 13:45
 Operator : CG/JU
 Sample : M3463-15
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SWCS-16-0204

Manual Integrations
 APPROVED

mohammad
 8/26/2021 11:40:35 AM

Quant Time: Aug 25 14:29:09 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 25 11:27:41 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.681	152	164550	20.000	ng	0.00
21) Naphthalene-d8	10.463	136	694454	20.000	ng	# 0.00
39) Acenaphthene-d10	14.328	164	401660	20.000	ng	0.00
64) Phenanthrene-d10	17.086	188	804634	20.000	ng	# 0.00
76) Chrysene-d12	21.198	240	784610	20.000	ng	# 0.00
86) Perylene-d12	23.504	264	849842	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.287	112	808176	75.437	ng	0.00
7) Phenol-d6	6.875	99	1042933	68.531	ng	0.00
23) Nitrobenzene-d5	8.840	82	706605	46.965	ng	0.00
42) 2,4,6-Tribromophenol	15.828	330	349220	83.191	ng	0.00
45) 2-Fluorobiphenyl	12.945	172	1163675	43.345	ng	0.00
79) Terphenyl-d14	19.669	244	1534483	39.189	ng	0.00
Target Compounds						Qvalue
49) Acenaphthylene	14.045	152	131363	3.480	ng	97
58) Fluorene	15.381	166	71418	2.472	ng	# 98
71) Phenanthrene	17.128	178	553845	12.203	ng	100
72) Anthracene	17.222	178	188595	4.300	ng	99
75) Fluoranthene	19.110	202	1384706	27.182	ng	96
78) Pyrene	19.463	202	1195650	22.814	ng	99
81) Benzo(a)anthracene	21.180	228	535181	10.437	ng	95
83) Chrysene	21.233	228	547895	11.022	ng	97
84) Bis(2-ethylhexyl)phtha...	21.121	149	347715	10.040	ng	# 98
87) Indeno(1,2,3-cd)pyrene	25.880	276	328170m	5.325	ng	
88) Benzo(b)fluoranthene	22.804	252	678480	12.284	ng	# 88
89) Benzo(k)fluoranthene	22.851	252	206894m	3.901	ng	
90) Benzo(a)pyrene	23.404	252	482566	9.741	ng	# 90
92) Benzo(g,h,i)perylene	26.615	276	340842	6.676	ng	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082521\
Data File : BP006879.D
Acq On : 25 Aug 2021 13:45
Operator : CG/JU
Sample : M3463-15
Misc :
ALS Vial : 7 Sample Multiplier: 1


Instrument :
BNA_P
ClientSampleId :
SWCS-16-0204

Manual Integrations
APPROVED

mohammad
8/26/2021 11:40:35 AM

Quant Time: Aug 25 14:29:09 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP080321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 25 11:27:41 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.681 min Scan# 781
Delta R.T. -0.000 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45

Tgt Ion:152 Resp: 164550
Ion Ratio Lower Upper
152 100
150 155.2 125.6 188.4
115 66.4 48.5 72.7

Instrument :
BNA_P
ClientSampleId :
SWCS-16-0204




Manual Integrations
APPROVED

mohammad
8/26/2021 11:40:35 AM



#5
2-Fluorophenol
Concen: 75.437 ng
RT: 5.287 min Scan# 374
Delta R.T. 0.000 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45


Tgt Ion:112 Resp: 808176
Ion Ratio Lower Upper
112 100
64 55.9 41.4 62.2
63 28.7 23.8 35.8





#7
Phenol-d6
Concen: 68.531 ng
RT: 6.875 min Scan# 644
Delta R.T. 0.006 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45

Instrument :
BNA_P
ClientSampleId :
SWCS-16-0204






Tgt Ion: 99 Resp: 1042933


Ion	Ratio	Lower	Upper
99	100		
42	15.5	13.4	20.0
71	36.8	30.6	46.0

Manual Integrations
APPROVED

mohammad
8/26/2021 11:40:35 AM





#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.463 min Scan# 1254
Delta R.T. 0.000 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45



Tgt Ion: 136 Resp: 694454


Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.6	12.8
54	7.2	4.6	6.8#
68	6.8	3.6	5.4#





#23
Nitrobenzene-d5
Concen: 46.965 ng
RT: 8.840 min Scan# 978
Delta R.T. -0.006 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45




Instrument :
BNA_P
ClientSampleId :
SWCS-16-0204




Tgt Ion: 82 Resp: 706605
Ion Ratio Lower Upper
82 100
128 39.6 35.7 53.5
54 41.2 35.0 52.4

Manual Integrations
APPROVED



mohammad
8/26/2021 11:40:35 AM




#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.328 min Scan# 1911
Delta R.T. 0.000 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45




Tgt Ion: 164 Resp: 401660
Ion Ratio Lower Upper
164 100
162 102.5 80.2 120.2
160 46.1 34.5 51.7





#42
2,4,6-Tribromophenol
Concen: 83.191 ng
RT: 15.828 min Scan# 2166
Delta R.T. 0.000 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45




Instrument :
BNA_P
ClientSampleId :
SWCS-16-0204




Tgt Ion:330 Resp: 349220
Ion Ratio Lower Upper
330 100
332 97.1 77.8 116.8
141 47.8 25.4 38.2#

Manual Integrations
APPROVED



mohammad
8/26/2021 11:40:35 AM




#45
2-Fluorobiphenyl
Concen: 43.345 ng
RT: 12.945 min Scan# 1676
Delta R.T. 0.000 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45




Tgt Ion:172 Resp: 1163675
Ion Ratio Lower Upper
172 100
171 35.3 27.9 41.9
170 23.4 18.6 27.8





#49
Acenaphthylene
Concen: 3.480 ng
RT: 14.045 min Scan# 1863
Delta R.T. 0.000 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45




Instrument :
BNA_P
ClientSampleId :
SWCS-16-0204



Tgt Ion:152 Resp: 131363
Ion Ratio Lower Upper
152 100
151 18.8 16.2 24.2
153 14.1 10.6 16.0




Manual Integrations
APPROVED


mohammad
8/26/2021 11:40:35 AM



#58
Fluorene
Concen: 2.472 ng
RT: 15.381 min Scan# 2090
Delta R.T. 0.000 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45


Tgt Ion:166 Resp: 71418
Ion Ratio Lower Upper
166 100
165 98.3 78.0 117.0
167 15.8 10.5 15.7#





#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.086 min Scan# 2380
Delta R.T. 0.000 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45




Instrument :
BNA_P
ClientSampleId :
SWCS-16-0204




Tgt Ion:188 Resp: 804634
Ion Ratio Lower Upper
188 100
94 10.7 5.6 8.4#
80 12.5 6.3 9.5#

Manual Integrations
APPROVED



mohammad
8/26/2021 11:40:35 AM




#71
Phenanthrene
Concen: 12.203 ng
RT: 17.128 min Scan# 2387
Delta R.T. 0.000 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45




Tgt Ion:178 Resp: 553845
Ion Ratio Lower Upper
178 100
176 19.2 15.1 22.7
179 15.5 12.4 18.6





#72
 Anthracene
 Concen: 4.300 ng
 RT: 17.222 min Scan# 2403
 Delta R.T. 0.000 min
 Lab File: BP006879.D
 Acq: 25 Aug 2021 13:45




Instrument :
 BNA_P
 ClientSampleId :
 SWCS-16-0204




Tgt Ion:178 Resp: 188595
 Ion Ratio Lower Upper
 178 100
 176 18.8 14.9 22.3
 179 16.1 12.2 18.2

Manual Integrations
 APPROVED



mohammad
 8/26/2021 11:40:35 AM




#75
 Fluoranthene
 Concen: 27.182 ng
 RT: 19.110 min Scan# 2724
 Delta R.T. 0.000 min
 Lab File: BP006879.D
 Acq: 25 Aug 2021 13:45



Tgt Ion:202 Resp: 1384706
 Ion Ratio Lower Upper
 202 100
 101 12.7 0.0 27.6
 203 17.9 0.0 37.8









#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.198 min Scan# 3079
Delta R.T. 0.006 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45

Tgt Ion:240 Resp: 784610
Ion Ratio Lower Upper
240 100
120 12.0 6.5 9.7#
236 25.9 20.6 31.0

Instrument :
BNA_P
ClientSampleId :
SWCS-16-0204




Manual Integrations
APPROVED


mohammad
8/26/2021 11:40:35 AM



#78
Pyrene
Concen: 22.814 ng
RT: 19.463 min Scan# 2784
Delta R.T. 0.000 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45


Tgt Ion:202 Resp: 1195650
Ion Ratio Lower Upper
202 100
200 21.1 16.3 24.5
203 18.3 14.2 21.2





#79
Terphenyl-d14
Concen: 39.189 ng
RT: 19.669 min Scan# 2819
Delta R.T. 0.000 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45




Instrument :
BNA_P
ClientSampleId :
SWCS-16-0204




Tgt Ion:244 Resp: 1534483
Ion Ratio Lower Upper
244 100
212 7.9 6.0 9.0
122 13.0 7.1 10.7#

Manual Integrations
APPROVED



mohammad
8/26/2021 11:40:35 AM




#81
Benzo(a)anthracene
Concen: 10.437 ng
RT: 21.180 min Scan# 3076
Delta R.T. 0.000 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45



Tgt Ion:228 Resp: 535181
Ion Ratio Lower Upper
228 100
226 30.3 21.6 32.4
229 21.6 15.8 23.6









#83
Chrysene
Concen: 11.022 ng
RT: 21.233 min Scan# 3085
Delta R.T. 0.006 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45

Tgt Ion:	228	Resp:	547895
Ion Ratio	Lower	Upper	
228	100		
226	30.1	23.8	35.6
229	21.9	15.4	23.2

Instrument :
BNA_P
ClientSampleId :
SWCS-16-0204




Manual Integrations
APPROVED


mohammad
8/26/2021 11:40:35 AM



#84
Bis(2-ethylhexyl)phthalate
Concen: 10.040 ng
RT: 21.121 min Scan# 3066
Delta R.T. 0.006 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45

Tgt Ion:	149	Resp:	347715
Ion Ratio	Lower	Upper	
149	100		
167	26.3	21.9	32.9
279	4.0	4.3	6.5









#86
Perylene-d12
Concen: 20.000 ng
RT: 23.504 min Scan# 3471
Delta R.T. 0.006 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45

Tgt Ion: 264 Resp: 849842
Ion Ratio Lower Upper
264 100
260 24.7 18.6 28.0
265 23.1 17.6 26.4

Instrument :
BNA_P
ClientSampleId :
SWCS-16-0204




Manual Integrations
APPROVED


mohammad
8/26/2021 11:40:35 AM



#87
Indeno(1,2,3-cd)pyrene
Concen: 5.325 ng m
RT: 25.880 min Scan# 3875
Delta R.T. 0.006 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45


Tgt Ion: 276 Resp: 328170
Ion Ratio Lower Upper
276 100
138 1.9 15.3 22.9#
277 2.8 20.4 30.6#





#88
Benzo(b)fluoranthene
Concen: 12.284 ng
RT: 22.804 min Scan# 3352
Delta R.T. 0.006 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45




Instrument :
BNA_P
ClientSampled :
SWCS-16-0204




Tgt Ion:252 Resp: 678480
Ion Ratio Lower Upper
252 100
253 24.4 17.5 26.3
125 18.2 6.2 9.4#

Manual Integrations
APPROVED



mohammad
8/26/2021 11:40:35 AM




#89
Benzo(k)fluoranthene
Concen: 3.901 ng m
RT: 22.851 min Scan# 3360
Delta R.T. 0.006 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45




Tgt Ion:252 Resp: 206894
Ion Ratio Lower Upper
252 100
253 25.0 17.6 26.4
125 21.3 6.0 9.0#





#90
Benzo(a)pyrene
Concen: 9.741 ng
RT: 23.404 min Scan# 3454
Delta R.T. 0.006 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45




Instrument :
BNA_P
ClientSampled :
SWCS-16-0204




Tgt Ion:252 Resp: 482566
Ion Ratio Lower Upper
252 100
253 24.1 17.9 26.9
125 17.4 6.5 9.7#

Manual Integrations
APPROVED

mohammad
8/26/2021 11:40:35 AM



#92
Benzo(g,h,i)perylene
Concen: 6.676 ng
RT: 26.615 min Scan# 4000
Delta R.T. 0.012 min
Lab File: BP006879.D
Acq: 25 Aug 2021 13:45



Tgt Ion:276 Resp: 340842
Ion Ratio Lower Upper
276 100
277 23.9 18.9 28.3
138 29.9 12.6 18.8#

