

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082622\
 Data File : BP011580.D
 Acq On : 26 Aug 2022 15:32
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP082622

Quant Time: Aug 27 00:39:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP082622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 27 00:36:07 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	133	0.00
2	1,4-Dioxane	40.000	37.652	5.9	128	0.00
3	Pyridine	40.000	39.055	2.4	133	0.00
4	n-Nitrosodimethylamine	40.000	36.651	8.4	121	0.00
5 S	2-Fluorophenol	80.000	77.824	2.7	131	0.00
6	Aniline	40.000	39.342	1.6	128	0.00
7 S	Phenol-d6	80.000	78.486	1.9	128	0.00
8	2-Chlorophenol	40.000	38.924	2.7	132	0.00
9	Benzaldehyde	40.000	38.032	4.9	128	0.00
10 C	Phenol	40.000	38.975	2.6	127	0.00
11	bis(2-Chloroethyl)ether	40.000	39.260	1.9	128	0.00
12	1,3-Dichlorobenzene	40.000	38.833	2.9	133	0.00
13 C	1,4-Dichlorobenzene	40.000	39.067	2.3	133	0.00
14	1,2-Dichlorobenzene	40.000	38.830	2.9	132	0.00
15	Benzyl Alcohol	40.000	39.743	0.6	124	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.909	7.7	120	0.00
17	2-Methylphenol	40.000	39.369	1.6	127	0.00
18	Hexachloroethane	40.000	38.488	3.8	127	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.640	3.4	123	0.00
20	3+4-Methylphenols	40.000	39.778	0.6	128	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	128	0.00
22	Acetophenone	40.000	38.627	3.4	126	0.00
23 S	Nitrobenzene-d5	80.000	76.166	4.8	123	0.00
24	Nitrobenzene	40.000	38.094	4.8	123	0.00
25	Isophorone	40.000	38.730	3.2	124	0.00
26 C	2-Nitrophenol	40.000	41.947	-4.9	133	0.00
27	2,4-Dimethylphenol	40.000	39.474	1.3	129	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.901	5.2	123	0.00
29 C	2,4-Dichlorophenol	40.000	40.619	-1.5	133	0.00
30	1,2,4-Trichlorobenzene	40.000	39.226	1.9	135	0.00
31	Naphthalene	40.000	38.338	4.2	128	0.00
32	Benzoic acid	40.000	41.961	-4.9	134	0.01
33	4-Chloroaniline	40.000	39.211	2.0	127	0.00
34 C	Hexachlorobutadiene	40.000	39.881	0.3	138	0.00
35	Caprolactam	40.000	39.984	0.0	125	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.340	1.6	125	0.00
37	2-Methylnaphthalene	40.000	38.749	3.1	128	0.00
38	1-Methylnaphthalene	40.000	38.929	2.7	129	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	129	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.078	-2.7	140	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.702	-4.3	137	0.00
42 S	2,4,6-Tribromophenol	80.000	89.260	-11.6	162	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.493	-3.7	136	0.00
44	2,4,5-Trichlorophenol	40.000	41.096	-2.7	135	0.00
45 S	2-Fluorobiphenyl	80.000	78.793	1.5	132	0.00
46	1,1'-Biphenyl	40.000	39.039	2.4	129	0.00
47	2-Chloronaphthalene	40.000	38.912	2.7	129	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082622\
 Data File : BP011580.D
 Acq On : 26 Aug 2022 15:32
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP082622

Quant Time: Aug 27 00:39:06 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP082622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Aug 27 00:36:07 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.786	0.5	119	0.00
49	Acenaphthylene	40.000	39.172	2.1	128	0.00
50	Dimethylphthalate	40.000	38.845	2.9	129	0.00
51	2,6-Dinitrotoluene	40.000	40.563	-1.4	130	0.00
52 C	Acenaphthene	40.000	38.886	2.8	128	0.00
53	3-Nitroaniline	40.000	41.371	-3.4	127	0.00
54 P	2,4-Dinitrophenol	40.000	42.978	-7.4	133	0.00
55	Dibenzofuran	40.000	38.848	2.9	129	0.00
56 P	4-Nitrophenol	40.000	42.016	-5.0	127	0.00
57	2,4-Dinitrotoluene	40.000	40.737	-1.8	129	0.00
58	Fluorene	40.000	39.064	2.3	130	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.780	-4.5	141	0.00
60	Diethylphthalate	40.000	38.234	4.4	125	0.00
61	4-Chlorophenyl-phenylether	40.000	39.825	0.4	136	0.00
62	4-Nitroaniline	40.000	42.129	-5.3	127	0.00
63	Azobenzene	40.000	37.097	7.3	116	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	131	0.00
65	4,6-Dinitro-2-methylphenol	40.000	42.223	-5.6	137	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.343	4.1	128	0.00
67	4-Bromophenyl-phenylether	40.000	40.628	-1.6	144	0.01
68	Hexachlorobenzene	40.000	41.579	-3.9	150	0.01
69	Atrazine	40.000	39.796	0.5	132	0.00
70 C	Pentachlorophenol	40.000	43.580	-8.9	149	0.00
71	Phenanthrene	40.000	38.250	4.4	129	0.00
72	Anthracene	40.000	38.608	3.5	129	0.00
73	Carbazole	40.000	38.122	4.7	127	0.00
74	Di-n-butylphthalate	40.000	38.476	3.8	124	0.00
75 C	Fluoranthene	40.000	38.947	2.6	133	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	138	0.00
77	Benzydine	40.000	41.997	-5.0	124	0.01
78	Pyrene	40.000	38.624	3.4	133	0.01
79 S	Terphenyl-d14	80.000	78.688	1.6	137	0.00
80	Butylbenzylphthalate	40.000	38.629	3.4	124	0.00
81	Benzo(a)anthracene	40.000	38.748	3.1	136	0.01
82	3,3'-Dichlorobenzidine	40.000	40.237	-0.6	139	0.00
83	Chrysene	40.000	38.202	4.5	135	0.01
84	Bis(2-ethylhexyl)phthalate	40.000	38.049	4.9	122	0.01
85 c	Di-n-octyl phthalate	40.000	37.658	5.9	122	0.01
86 I	Perylene-d12	20.000	20.000	0.0	145	0.01
87	Indeno(1,2,3-cd)pyrene	40.000	41.410	-3.5	155	0.01
88	Benzo(b)fluoranthene	40.000	38.749	3.1	142	0.01
89	Benzo(k)fluoranthene	40.000	39.062	2.3	140	0.01
90 C	Benzo(a)pyrene	40.000	39.418	1.5	143	0.02
91	Dibenzo(a,h)anthracene	40.000	41.190	-3.0	152	0.02
92	Benzo(g,h,i)perylene	40.000	40.013	-0.0	157	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082622\
Data File : BP011580.D
Acq On : 26 Aug 2022 15:32
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP082622

Quant Time: Aug 27 00:39:06 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP082622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Aug 27 00:36:07 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0