

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082624\
 Data File : BP021671.D
 Acq On : 27 Aug 2024 14:38
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020743

Quant Time: Aug 27 22:46:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 27 05:11:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.793	152	306629	20.000	ng/u1	0.00
20) Naphthalene-d8	10.581	136	1269254	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.434	164	831179	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.239	188	1850014	20.000	ng/u1	-0.01
79) Chrysene-d12	21.698	240	1921848	20.000	ng/u1	0.00
88) Perylene-d12	25.104	264	2307693	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	72592	7.860	ng/uL	0.00
4) Pyridine-d5	3.693	84	536527	20.084	ng/u1	0.00
7) Phenol-d5	6.958	99	625759	19.123	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.122	67	351547	19.740	ng/u1	0.00
11) 2-Chlorophenol-d4	7.328	132	424046	20.072	ng/u1	0.00
15) 4-Methylphenol-d8	8.493	113	484055	19.143	ng/u1	0.00
21) Nitrobenzene-d5	8.940	128	203707	19.948	ng/u1	0.00
24) 2-Nitrophenol-d4	9.663	143	204742	19.752	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.199	165	410975	20.120	ng/u1	0.00
31) 4-Chloroaniline-d4	10.710	131	589893	19.850	ng/u1	0.00
46) Dimethylphthalate-d6	13.834	166	1270743	19.980	ng/u1	0.00
49) Acenaphthylene-d8	14.122	160	1453808	20.360	ng/u1	0.00
54) 4-Nitrophenol-d4	14.628	143	247553	20.459	ng/u1	0.00
60) Fluorene-d10	15.445	176	1094565	20.471	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.557	200	180365	17.726	ng/u1	-0.01
73) Anthracene-d10	17.345	188	1777738	20.248	ng/u1	0.00
81) Pyrene-d10	19.716	212	2165935	19.715	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.886	264	2454250	19.944	ng/u1	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.317	88	79060	8.054	ng/uL	98
5) Pyridine	3.711	79	535108	20.045	ng/u1	99
6) Benzaldehyde	6.940	77	406682	24.297	ng/u1	99
8) Phenol	6.987	94	656093	19.228	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.216	93	518353	19.544	ng/u1	99
12) 2-Chlorophenol	7.363	128	444453	20.044	ng/u1	99
13) 2-Methylphenol	8.228	108	470154	19.100	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.310	45	508416	19.860	ng/u1	98
16) Acetophenone	8.610	105	805564	19.871	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.593	70	376652	19.688	ng/u1	100
18) 4-Methylphenol	8.558	108	523387	19.422	ng/u1	98
19) Hexachloroethane	8.875	117	200379	20.210	ng/u1	97
22) Nitrobenzene	8.981	77	579812	20.362	ng/u1	99
23) Isophorone	9.510	82	1106104	19.358	ng/u1	99
25) 2-Nitrophenol	9.693	139	233421	20.325	ng/u1	97
26) 2,4-Dimethylphenol	9.752	107	569791	20.215	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.987	93	694705	19.387	ng/u1	98
29) 2,4-Dichlorophenol	10.228	162	417022	20.317	ng/u1	99
30) Naphthalene	10.634	128	1421113	19.866	ng/u1	100
32) 4-Chloroaniline	10.734	127	604187	20.138	ng/u1	97
33) Hexachlorobutadiene	10.922	225	274433	19.927	ng/u1	99
34) Caprolactam	11.504	113	163147	18.796	ng/u1	97
35) 4-Chloro-3-methylphenol	11.863	107	549869	20.240	ng/u1	97
36) 2-Methylnaphthalene	12.246	142	962936	19.901	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082624\
 Data File : BP021671.D
 Acq On : 27 Aug 2024 14:38
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020743

Quant Time: Aug 27 22:46:49 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 27 05:11:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.463	142	973010	19.993	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.610	216	526486	20.097	ng/ul	99
40) Hexachlorocyclopentadiene	12.599	237	277868	18.485	ng/ul	98
41) 2,4,6-Trichlorophenol	12.851	196	342330	20.296	ng/ul	97
42) 2,4,5-Trichlorophenol	12.922	196	373187	20.490	ng/ul	98
43) 1,1'-Biphenyl	13.257	154	1266633	20.142	ng/ul	99
44) 2-Chloronaphthalene	13.293	162	996420	20.209	ng/ul	100
45) 2-Nitroaniline	13.493	65	323732	20.958	ng/ul	96
47) Dimethylphthalate	13.881	163	1267715	19.907	ng/ul	99
48) 2,6-Dinitrotoluene	13.993	165	242840	20.676	ng/ul	99
50) Acenaphthylene	14.151	152	1568990	20.431	ng/ul	99
51) 3-Nitroaniline	14.328	138	281826	21.959	ng/ul	100
52) Acenaphthene	14.498	153	1087099	20.129	ng/ul	99
53) 2,4-Dinitrophenol	14.540	184	117219	16.671	ng/ul	93
55) 4-Nitrophenol	14.640	109	245939	22.198	ng/ul	97
56) Dibenzofuran	14.840	168	1536980	20.571	ng/ul	98
57) 2,4-Dinitrotoluene	14.798	165	376535	21.335	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.069	232	321862	20.512	ng/ul	95
59) Diethylphthalate	15.275	149	1302196	20.363	ng/ul	99
61) Fluorene	15.498	166	1244522	20.594	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.498	204	613805	20.214	ng/ul	100
63) 4-Nitroaniline	15.510	138	306821	24.736	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.575	198	205049	18.015	ng/ul	97
67) N-Nitrosodiphenylamine	15.716	169	1084994	19.936	ng/ul	99
68) 4-Bromophenyl-phenylether	16.416	248	398523	19.513	ng/ul	96
69) Hexachlorobenzene	16.534	284	467740	19.292	ng/ul	97
70) Atrazine	16.692	200	411182	19.918	ng/ul	98
71) Pentachlorophenol	16.881	266	299684	19.313	ng/ul	100
72) Phenanthrene	17.286	178	2055655	20.240	ng/ul	99
74) Anthracene	17.381	178	2097328	20.383	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.222	216	533379	19.392	ng/ul	97
76) Pentachlorobenzene	14.763	250	545342	19.129	ng/ul	99
77) Carbazole	17.657	167	1961274	21.193	ng/ul	100
78) Di-n-butylphthalate	18.263	149	2235621	19.846	ng/ul	100
80) Fluoranthene	19.369	202	2545771	19.858	ng/ul	100
82) Pyrene	19.745	202	2708345	19.947	ng/ul	100
83) Butylbenzylphthalate	20.698	149	1061652	19.097	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.598	252	889167	18.359	ng/ul	98
85) Benzo(a)anthracene	21.674	228	2748296	20.103	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.622	149	1580826	19.266	ng/ul	98
87) Chrysene	21.745	228	2599189	20.202	ng/ul	100
89) Di-n-octyl phthalate	22.933	149	2642963	18.093	ng/ul	100
90) Benzo(b)fluoranthene	24.027	252	2855503	19.776	ng/ul	98
91) Benzo(k)fluoranthene	24.104	252	2850361	20.164	ng/ul	99
93) Benzo(a)pyrene	24.951	252	2712628	20.391	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	28.986	276	3463960	20.195	ng/ul	98
95) Dibenzo(a,h)anthracene	29.056	278	2794088	20.272	ng/ul	99
96) Benzo(g,h,i)perylene	30.156	276	2698762	20.364	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 8.054 ng/uL

RT: 3.317 min Scan# 50

Delta R.T. 0.000 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38

Instrument :

BNA_P

ClientSampleId :

SSTD020743





Tgt Ion: 88 Resp: 79060

Ion Ratio Lower Upper

88 100

43 20.8 15.4 23.2

58 58.9 48.2 72.2



#5

Pyridine

Concen: 20.045 ng/uL

RT: 3.711 min Scan# 123

Delta R.T. 0.000 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38




Tgt Ion: 79 Resp: 535108


Ion Ratio Lower Upper

79 100

52 57.6 46.8 70.2

51 27.4 21.7 32.5









#6
Benzaldehyde
Concen: 24.297 ng/ul
RT: 6.940 min Scan# 61
Delta R.T. 0.000 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38




Instrument :
BNA_P
ClientSampleId :
SSTD020743


Tgt Ion: 77 Resp: 406682
Ion Ratio Lower Upper
77 100
105 102.2 80.2 120.4
106 94.6 76.2 114.4



#8
Phenol
Concen: 19.228 ng/ul
RT: 6.987 min Scan# 680
Delta R.T. 0.000 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38

Tgt Ion: 94 Resp: 656093
Ion Ratio Lower Upper
94 100
65 25.1 20.0 30.0
66 34.4 27.0 40.4









#10
 Bis(2-Chloroethyl)ether
 Concen: 19.544 ng/ul
 RT: 7.216 min Scan# 719
 Delta R.T. 0.000 min
 Lab File: BP021671.D
 Acq: 27 Aug 2024 14:38




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020743


Tgt Ion: 93 Resp: 518353
 Ion Ratio Lower Upper
 93 100
 63 69.1 54.9 82.3
 95 32.2 26.0 39.0



#12
 2-Chlorophenol
 Concen: 20.044 ng/ul
 RT: 7.363 min Scan# 744
 Delta R.T. 0.000 min
 Lab File: BP021671.D
 Acq: 27 Aug 2024 14:38

Tgt Ion: 128 Resp: 444453
 Ion Ratio Lower Upper
 128 100
 64 48.6 39.5 59.3
 130 32.1 25.9 38.9





#13

2-Methylphenol

Concen: 19.100 ng/ul

RT: 8.228 min Scan# 891

Delta R.T. 0.000 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38

Instrument :

BNA_P

ClientSampleId :





SSTD020743

Tgt Ion:108 Resp: 470154

Ion Ratio Lower Upper

108 100

107 90.2 73.2 109.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 19.860 ng/ul

RT: 8.310 min Scan# 905

Delta R.T. 0.000 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38




Tgt Ion: 45 Resp: 508416


Ion Ratio Lower Upper

45 100

77 19.6 17.0 25.4

79 15.0 12.0 18.0









#16
Acetophenone
Concen: 19.871 ng/ul
RT: 8.610 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38




Instrument :
BNA_P
ClientSampleId :
SSTD020743


Tgt Ion:105 Resp: 805564
Ion Ratio Lower Upper
105 100
77 74.2 59.4 89.2
51 20.2 16.4 24.6



#17
N-Nitroso-di-n-propylamine
Concen: 19.688 ng/ul
RT: 8.593 min Scan# 953
Delta R.T. 0.000 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38

Tgt Ion: 70 Resp: 376652
Ion Ratio Lower Upper
70 100
42 44.8 35.8 53.6
101 10.6 8.2 12.4
130 18.7 15.1 22.7





#18

4-Methylphenol

Concen: 19.422 ng/ul

RT: 8.558 min Scan# 947

Delta R.T. 0.000 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38

Instrument :

BNA_P

ClientSampleId :





SSTD020743

Tgt Ion:108 Resp: 523387

Ion Ratio Lower Upper

108 100

107 110.1 90.1 135.1



#19

Hexachloroethane

Concen: 20.210 ng/ul

RT: 8.875 min Scan# 1001

Delta R.T. 0.006 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38




Tgt Ion:117 Resp: 200379


Ion Ratio Lower Upper

117 100

201 91.5 75.3 112.9

199 56.5 47.4 71.0









#22
Nitrobenzene
Concen: 20.362 ng/ul
RT: 8.981 min Scan# 1019
Delta R.T. 0.000 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38




Instrument :
BNA_P
ClientSampleId :
SSTD020743


Tgt Ion: 77 Resp: 579812
Ion Ratio Lower Upper
77 100
123 40.8 33.2 49.8
65 13.1 9.8 14.6



#23
Isophorone
Concen: 19.358 ng/ul
RT: 9.510 min Scan# 1109
Delta R.T. 0.000 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38

Tgt Ion: 82 Resp: 1106104
Ion Ratio Lower Upper
82 100
95 7.3 5.8 8.8
138 14.5 11.8 17.8









#25
2-Nitrophenol
Concen: 20.325 ng/ul
RT: 9.693 min Scan# 1140
Delta R.T. 0.000 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38




Instrument :
BNA_P
ClientSampleId :
SSTD020743


Tgt Ion:139 Resp: 233421
Ion Ratio Lower Upper
139 100
65 59.3 46.2 69.2
109 37.7 28.0 42.0



#26
2,4-Dimethylphenol
Concen: 20.215 ng/ul
RT: 9.752 min Scan# 1150
Delta R.T. 0.000 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38

Tgt Ion:107 Resp: 569791
Ion Ratio Lower Upper
107 100
121 45.3 36.9 55.3
122 76.7 61.0 91.4





#27

Bis(2-Chloroethoxy)methane

Concen: 19.387 ng/ul

RT: 9.987 min Scan# 1190

Delta R.T. 0.000 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38

Instrument :

BNA_P

ClientSampleId :

SSTD020743





Tgt Ion: 93 Resp: 694705

Ion Ratio Lower Upper

93 100

95 32.4 25.3 37.9

123 9.8 7.4 11.0



#29

2,4-Dichlorophenol

Concen: 20.317 ng/ul

RT: 10.228 min Scan# 1231

Delta R.T. 0.006 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38




Tgt Ion:162 Resp: 417022


Ion Ratio Lower Upper

162 100

164 64.6 52.3 78.5

98 43.4 35.3 52.9









#30
Naphthalene
Concen: 19.866 ng/ul
RT: 10.634 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38




Instrument :
BNA_P
ClientSampleId :
SSTD020743


Tgt Ion:128 Resp: 1421113
Ion Ratio Lower Upper
128 100
129 10.8 8.6 13.0
127 12.8 10.5 15.7



#32
4-Chloroaniline
Concen: 20.138 ng/ul
RT: 10.734 min Scan# 1317
Delta R.T. 0.006 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38

Tgt Ion:127 Resp: 604187
Ion Ratio Lower Upper
127 100
129 30.9 25.8 38.8





#33

Hexachlorobutadiene

Concen: 19.927 ng/ul

RT: 10.922 min Scan# 1

Delta R.T. 0.000 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38

Instrument :

BNA_P

ClientSampleId :

SSTD020743





Tgt Ion:225 Resp: 274433

Ion Ratio Lower Upper

225 100

223 63.9 51.5 77.3

227 63.7 51.3 76.9



#34

Caprolactam

Concen: 18.796 ng/ul

RT: 11.504 min Scan# 1448

Delta R.T. 0.000 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38




Tgt Ion:113 Resp: 163147


Ion Ratio Lower Upper

113 100

55 143.1 110.2 165.2

56 108.3 84.9 127.3









#35
4-Chloro-3-methylphenol
Concen: 20.240 ng/ul
RT: 11.863 min Scan# 1509
Delta R.T. 0.000 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38




Instrument :
BNA_P
ClientSampleId :
SSTD020743


Tgt Ion:107 Resp: 549869
Ion Ratio Lower Upper
107 100
144 21.1 17.5 26.3
142 64.0 53.3 79.9



#36
2-Methylnaphthalene
Concen: 19.901 ng/ul
RT: 12.246 min Scan# 1574
Delta R.T. 0.000 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38

Tgt Ion:142 Resp: 962936
Ion Ratio Lower Upper
142 100
141 89.4 71.1 106.7









#37
1-Methylnaphthalene
Concen: 19.993 ng/ul
RT: 12.463 min Scan# 1611
Delta R.T. 0.000 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38




Instrument :
BNA_P
ClientSampleId :
SSTD020743


Tgt Ion:142 Resp: 973010
Ion Ratio Lower Upper
142 100
141 92.3 74.6 111.8
116 4.3 3.5 5.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.097 ng/ul
RT: 12.610 min Scan# 1636
Delta R.T. 0.000 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38

Tgt Ion:216 Resp: 526486
Ion Ratio Lower Upper
216 100
214 78.0 63.2 94.8
179 21.1 16.6 25.0
108 21.5 16.2 24.4









#40
Hexachlorocyclopentadiene
Concen: 18.485 ng/ul
RT: 12.599 min Scan# 1634
Delta R.T. 0.000 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38




Instrument :
BNA_P
ClientSampleId :
SSTD020743


Tgt Ion:237 Resp: 277868
Ion Ratio Lower Upper
237 100
235 62.1 50.7 76.1
272 13.0 11.0 16.6



#41
2,4,6-Trichlorophenol
Concen: 20.296 ng/ul
RT: 12.851 min Scan# 1677
Delta R.T. 0.000 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38

Tgt Ion:196 Resp: 342330
Ion Ratio Lower Upper
196 100
198 99.0 76.5 114.7
200 31.2 25.3 37.9





#42

2,4,5-Trichlorophenol

Concen: 20.490 ng/ul

RT: 12.922 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38

Instrument :

BNA_P

ClientSampleId :

SSTD020743





Tgt Ion:196 Resp: 373187

Ion Ratio Lower Upper

196 100

198 95.0 78.0 117.0

200 30.9 24.3 36.5



#43

1,1'-Biphenyl

Concen: 20.142 ng/ul

RT: 13.257 min Scan# 1746

Delta R.T. -0.006 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38




Tgt Ion:154 Resp: 1266633


Ion Ratio Lower Upper

154 100

153 40.1 31.6 47.4

76 14.7 12.6 18.8





#44

2-Chloronaphthalene

Concen: 20.209 ng/ul

RT: 13.293 min Scan# 1753

Delta R.T. -0.006 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38

Instrument :

BNA_P

ClientSampleId :

SSTD020743





Tgt Ion:162 Resp: 996420

Ion Ratio Lower Upper

162 100

164 32.6 26.0 39.0

127 41.5 33.1 49.7



#45

2-Nitroaniline

Concen: 20.958 ng/ul

RT: 13.493 min Scan# 1786

Delta R.T. -0.012 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38




Tgt Ion: 65 Resp: 323732


Ion Ratio Lower Upper

65 100

92 77.5 65.4 98.2

138 94.1 77.5 116.3









#47
Dimethylphthalate
Concen: 19.907 ng/ul
RT: 13.881 min Scan# 1852
Delta R.T. -0.006 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38




Instrument :
BNA_P
ClientSampleId :
SSTD020743


Tgt Ion:163 Resp: 1267715
Ion Ratio Lower Upper
163 100
194 3.7 3.5 5.3
164 10.2 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 20.676 ng/ul
RT: 13.993 min Scan# 1871
Delta R.T. -0.006 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38

Tgt Ion:165 Resp: 242840
Ion Ratio Lower Upper
165 100
89 70.9 55.7 83.5
121 23.2 18.6 28.0





#50

Acenaphthylene

Concen: 20.431 ng/ul

RT: 14.151 min Scan# 1899

Delta R.T. -0.006 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38

Instrument :

BNA_P

ClientSampleId :

SSTD020743





Tgt Ion:152 Resp: 1568990

Ion Ratio Lower Upper

152 100

151 19.9 15.6 23.4

153 12.9 10.2 15.4



#51

3-Nitroaniline

Concen: 21.959 ng/ul

RT: 14.328 min Scan# 1928

Delta R.T. 0.000 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38




Tgt Ion:138 Resp: 281826


Ion Ratio Lower Upper

138 100

108 14.7 12.0 18.0

92 142.3 114.1 171.1









#52
Acenaphthene
Concen: 20.129 ng/ul
RT: 14.498 min Scan# 1957
Delta R.T. -0.006 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38




Instrument :
BNA_P
ClientSampleId :
SSTD020743


Tgt Ion:153 Resp: 1087099
Ion Ratio Lower Upper
153 100
152 47.7 38.0 57.0
154 88.6 70.2 105.4



#53
2,4-Dinitrophenol
Concen: 16.671 ng/ul
RT: 14.540 min Scan# 1964
Delta R.T. -0.006 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38

Tgt Ion:184 Resp: 117219
Ion Ratio Lower Upper
184 100
63 75.3 55.5 83.3
154 69.1 51.3 76.9









#55
4-Nitrophenol
Concen: 22.198 ng/ul
RT: 14.640 min Scan# 1982
Delta R.T. -0.006 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38




Instrument :
BNA_P
ClientSampleId :
SSTD020743


Tgt Ion:109 Resp: 245939
Ion Ratio Lower Upper
109 100
139 117.2 96.2 144.2
65 126.3 98.5 147.7



#56
Dibenzofuran
Concen: 20.571 ng/ul
RT: 14.840 min Scan# 2015
Delta R.T. -0.006 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38

Tgt Ion:168 Resp: 1536980
Ion Ratio Lower Upper
168 100
139 38.6 29.8 44.8





#57

2,4-Dinitrotoluene

Concen: 21.335 ng/ul

RT: 14.798 min Scan# 2008

Delta R.T. -0.006 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38

Instrument :

BNA_P

ClientSampleId :

SSTD020743





Tgt Ion:165 Resp: 376535

Ion Ratio Lower Upper

165 100

63 49.3 37.8 56.8

182 2.7 2.2 3.4



#58

2,3,4,6-Tetrachlorophenol

Concen: 20.512 ng/ul

RT: 15.069 min Scan# 2054

Delta R.T. -0.006 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38

Tgt Ion:232 Resp: 321862




Ion Ratio Lower Upper


232 100

131 47.2 33.9 50.9

130 2.1 2.0 3.0

166 31.4 24.3 36.5









#59
Diethylphthalate
Concen: 20.363 ng/ul
RT: 15.275 min Scan# 2090
Delta R.T. -0.006 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38




Instrument :
BNA_P
ClientSampleId :
SSTD020743


Tgt Ion:149 Resp: 1302196
Ion Ratio Lower Upper
149 100
177 21.5 18.0 27.0
150 12.2 9.8 14.8



#61
Fluorene
Concen: 20.594 ng/ul
RT: 15.498 min Scan# 2127
Delta R.T. -0.012 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38

Tgt Ion:166 Resp: 1244522
Ion Ratio Lower Upper
166 100
165 96.9 78.2 117.2
167 13.8 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 20.214 ng/ul
RT: 15.498 min Scan# 2127
Delta R.T. -0.012 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38




Instrument :
BNA_P
ClientSampleId :
SSTD020743


Tgt Ion:204 Resp: 613805
Ion Ratio Lower Upper
204 100
206 33.1 26.2 39.4
141 60.8 48.6 73.0



#63
4-Nitroaniline
Concen: 24.736 ng/ul
RT: 15.510 min Scan# 2129
Delta R.T. -0.006 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38

Tgt Ion:138 Resp: 306821
Ion Ratio Lower Upper
138 100
92 61.7 50.7 76.1
108 93.3 72.2 108.2









#66
4,6-Dinitro-2-methylphenol
Concen: 18.015 ng/ul
RT: 15.575 min Scan# 2141
Delta R.T. -0.006 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38




Instrument :
BNA_P
ClientSampleId :
SSTD020743


Tgt Ion:198 Resp: 205049
Ion Ratio Lower Upper
198 100
182 4.8 3.3 4.9
77 30.6 23.3 34.9



#67
N-Nitrosodiphenylamine
Concen: 19.936 ng/ul
RT: 15.716 min Scan# 2164
Delta R.T. -0.006 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38

Tgt Ion:169 Resp: 1084994
Ion Ratio Lower Upper
169 100
168 66.3 52.6 79.0
167 36.2 28.2 42.4









#68
4-Bromophenyl-phenylether
Concen: 19.513 ng/ul
RT: 16.416 min Scan# 2283
Delta R.T. 0.000 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38




Instrument :
BNA_P
ClientSampleId :
SSTD020743


Tgt Ion:248 Resp: 398523
Ion Ratio Lower Upper
248 100
250 97.6 79.6 119.4
141 76.2 57.0 85.4



#69
Hexachlorobenzene
Concen: 19.292 ng/ul
RT: 16.534 min Scan# 2303
Delta R.T. 0.000 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38

Tgt Ion:284 Resp: 467740
Ion Ratio Lower Upper
284 100
142 34.3 26.5 39.7
249 32.1 24.2 36.4





#70

Atrazine

Concen: 19.918 ng/ul

RT: 16.692 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38

Instrument :

BNA_P

ClientSampleId :

SSTD020743





Tgt Ion:200 Resp: 411182

Ion Ratio Lower Upper

200 100

173 26.5 21.4 32.0

215 47.3 39.0 58.6



#71

Pentachlorophenol

Concen: 19.313 ng/ul

RT: 16.881 min Scan# 2362

Delta R.T. -0.012 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38




Tgt Ion:266 Resp: 299684


Ion Ratio Lower Upper

266 100

264 64.5 51.4 77.0

268 63.6 50.6 75.8









#72
Phenanthrene
Concen: 20.240 ng/ul
RT: 17.286 min Scan# 2431
Delta R.T. -0.006 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38




Instrument :
BNA_P
ClientSampleId :
SSTD020743


Tgt Ion:178 Resp: 2055655
Ion Ratio Lower Upper
178 100
179 15.4 12.0 18.0
176 18.9 14.8 22.2



#74
Anthracene
Concen: 20.383 ng/ul
RT: 17.381 min Scan# 2447
Delta R.T. -0.006 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38

Tgt Ion:178 Resp: 2097328
Ion Ratio Lower Upper
178 100
179 15.3 12.3 18.5
176 18.1 14.6 21.8









#75
1,2,3,4-Tetrachlorobenzene
Concen: 19.392 ng/uL
RT: 13.222 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38




Instrument :
BNA_P
ClientSampleId :
SSTD020743


Tgt Ion:216 Resp: 533379
Ion Ratio Lower Upper
216 100
214 76.1 63.9 95.9
218 48.0 39.5 59.3



#76
Pentachlorobenzene
Concen: 19.129 ng/uL
RT: 14.763 min Scan# 2002
Delta R.T. -0.006 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38

Tgt Ion:250 Resp: 545342
Ion Ratio Lower Upper
250 100
248 63.0 51.0 76.6
252 64.3 51.4 77.0





#77

Carbazole

Concen: 21.193 ng/ul

RT: 17.657 min Scan# 2494

Delta R.T. 0.000 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38

Instrument :

BNA_P

ClientSampleId :

SSTD020743





Tgt Ion:167 Resp: 1961274

Ion Ratio Lower Upper

167 100

166 21.3 17.1 25.7

139 12.6 10.4 15.6



#78

Di-n-butylphthalate

Concen: 19.846 ng/ul

RT: 18.263 min Scan# 2597

Delta R.T. -0.006 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38




Tgt Ion:149 Resp: 2235621


Ion Ratio Lower Upper

149 100

150 9.1 7.4 11.0

104 6.1 4.8 7.2









#80
Fluoranthene
Concen: 19.858 ng/ul
RT: 19.369 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38




Instrument :
BNA_P
ClientSampleId :
SSTD020743


Tgt Ion:202 Resp: 2545771
Ion Ratio Lower Upper
202 100
101 13.9 11.2 16.8
100 10.9 8.5 12.7



#82
Pyrene
Concen: 19.947 ng/ul
RT: 19.745 min Scan# 2849
Delta R.T. 0.000 min
Lab File: BP021671.D
Acq: 27 Aug 2024 14:38

Tgt Ion:202 Resp: 2708345
Ion Ratio Lower Upper
202 100
101 16.2 12.8 19.2
100 13.6 10.8 16.2





#83

Butylbenzylphthalate

Concen: 19.097 ng/ul

RT: 20.698 min Scan# 3011

Delta R.T. 0.000 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38

Instrument :

BNA_P

ClientSampleId :

SSTD020743





Tgt Ion:149 Resp: 1061652

Ion Ratio Lower Upper

149 100

91 83.3 70.1 105.1

206 21.5 16.8 25.2



#84

3,3'-Dichlorobenzidine

Concen: 18.359 ng/ul

RT: 21.598 min Scan# 3164

Delta R.T. 0.006 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38




Tgt Ion:252 Resp: 889167


Ion Ratio Lower Upper

252 100

254 64.0 52.2 78.4

126 10.5 8.9 13.3





#85

Benzo(a)anthracene

Concen: 20.103 ng/ul

RT: 21.674 min Scan# 3177

Delta R.T. 0.000 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38

Instrument :

BNA_P

ClientSampleId :

SSTD020743





Tgt Ion:228 Resp: 2748296

Ion Ratio Lower Upper

228 100

229 19.6 15.7 23.5

226 26.4 21.7 32.5



#86

Bis(2-ethylhexyl)phthalate

Concen: 19.266 ng/ul

RT: 21.622 min Scan# 3168

Delta R.T. -0.006 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38




Tgt Ion:149 Resp: 1580826


Ion Ratio Lower Upper

149 100

167 26.3 21.8 32.6

279 4.0 3.7 5.5





#87

Chrysene

Concen: 20.202 ng/ul

RT: 21.745 min Scan# 3189

Delta R.T. 0.000 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38

Instrument :

BNA_P

ClientSampleId :

SSTD020743





Tgt Ion:228 Resp: 2599189

Ion Ratio Lower Upper

228 100

226 29.2 23.5 35.3

229 19.5 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 18.093 ng/ul




RT: 22.933 min Scan# 3391


Delta R.T. 0.000 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38

Tgt Ion:149 Resp: 2642963





#90

Benzo(b)fluoranthene

Concen: 19.776 ng/ul

RT: 24.027 min Scan# 3576

Delta R.T. 0.006 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38

Instrument :

BNA_P

ClientSampleId :

SSTD020743





Tgt Ion:252 Resp: 2855503

Ion Ratio Lower Upper

252 100

253 21.8 18.1 27.1

125 11.3 8.6 12.8



#91

Benzo(k)fluoranthene

Concen: 20.164 ng/ul

RT: 24.104 min Scan# 3590

Delta R.T. 0.006 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38




Tgt Ion:252 Resp: 2850361


Ion Ratio Lower Upper

252 100

253 21.5 17.8 26.8

125 11.1 8.6 13.0





#93

Benzo(a)pyrene

Concen: 20.391 ng/ul

RT: 24.951 min Scan# 31

Delta R.T. 0.006 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38

Instrument :

BNA_P

ClientSampleId :

SSTD020743





Tgt Ion:252 Resp: 2712628

Ion Ratio Lower Upper

252 100

253 22.1 17.4 26.0

125 12.1 9.6 14.4



#94

Indeno(1,2,3-cd)pyrene

Concen: 20.195 ng/ul

RT: 28.986 min Scan# 4420

Delta R.T. 0.012 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38




Tgt Ion:276 Resp: 3463960


Ion Ratio Lower Upper

276 100

138 19.9 14.9 22.3

277 23.6 19.0 28.6





#95

Dibenzo(a,h)anthracene

Concen: 20.272 ng/ul

RT: 29.056 min Scan# 4433

Delta R.T. -0.006 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38

Instrument :

BNA_P

ClientSampleId :

SSTD020743





Tgt Ion:278 Resp: 2794088

Ion Ratio Lower Upper

278 100

139 14.8 12.7 19.1

279 23.8 18.9 28.3



#96

Benzo(g,h,i)perylene

Concen: 20.364 ng/ul

RT: 30.156 min Scan# 4619

Delta R.T. 0.006 min

Lab File: BP021671.D

Acq: 27 Aug 2024 14:38

Tgt Ion:276 Resp: 2698762

Ion Ratio Lower Upper

276 100

138 21.2 17.3 25.9

277 23.5 19.4 29.0

