

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082624\
 Data File : BP021661.D
 Acq On : 27 Aug 2024 07:35
 Operator : RC/JU
 Sample : P3711-03
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 E2495

Quant Time: Aug 27 08:03:21 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Aug 27 05:11:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.793	152	329997	20.000	ng/u1	0.00
20) Naphthalene-d8	10.581	136	1341988	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.440	164	871838	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.245	188	1926086	20.000	ng/u1	0.00
79) Chrysene-d12	21.692	240	2001242	20.000	ng/u1	0.00
88) Perylene-d12	25.116	264	2368625	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.281	96	40579	4.082	ng/uL	0.00
4) Pyridine-d5	3.693	84	116678	4.058	ng/u1	0.00
7) Phenol-d5	6.958	99	263152	7.472	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.122	67	532310	27.774	ng/u1	0.00
11) 2-Chlorophenol-d4	7.328	132	564629	24.833	ng/u1	0.00
15) 4-Methylphenol-d8	8.487	113	450911	16.569	ng/u1	0.00
21) Nitrobenzene-d5	8.940	128	304679	28.219	ng/u1	0.00
24) 2-Nitrophenol-d4	9.658	143	309532	28.243	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.199	165	563357	26.085	ng/u1	0.00
31) 4-Chloroaniline-d4	10.710	131	612888	19.506	ng/u1	0.00
46) Dimethylphthalate-d6	13.834	166	2050735	30.741	ng/u1	0.00
49) Acenaphthylene-d8	14.122	160	2085245	27.841	ng/u1	0.00
54) 4-Nitrophenol-d4	14.634	143	85748	6.756	ng/u1	0.00
60) Fluorene-d10	15.445	176	1682338	29.996	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.563	200	285220	26.924	ng/u1	0.00
73) Anthracene-d10	17.345	188	3017015	33.005	ng/u1	0.00
81) Pyrene-d10	19.710	212	3692729	32.279	ng/u1	-0.01
92) Benzo(a)pyrene-d12	24.874	264	4347328	34.420	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP082624\
Data File : BP021661.D
Acq On : 27 Aug 2024 07:35
Operator : RC/JU
Sample : P3711-03
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
E2495

Quant Time: Aug 27 08:03:21 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082324.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Aug 27 05:11:38 2024
Response via : Initial Calibration

