

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP083024\
 Data File : BP021777.D
 Acq On : 31 Aug 2024 08:59
 Operator : RC/JU
 Sample : P3733-12
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 A44A4

Quant Time: Sep 02 01:35:02 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 29 23:56:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.781	152	310151	20.000	ng/u1	0.00
20) Naphthalene-d8	10.569	136	1205078	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.428	164	705841	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.228	188	1424824	20.000	ng/u1	0.00
79) Chrysene-d12	21.675	240	1351611	20.000	ng/u1	0.00
88) Perylene-d12	25.074	264	1639769	20.000	ng/u1	0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.276	96	41898	4.485	ng/uL	0.00
4) Pyridine-d5	3.687	84	564321	20.884	ng/u1	0.00
7) Phenol-d5	6.952	99	793763	23.981	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.111	67	435226	24.162	ng/u1	0.00
11) 2-Chlorophenol-d4	7.316	132	532450	24.917	ng/u1	0.00
15) 4-Methylphenol-d8	8.481	113	606403	23.709	ng/u1	0.00
21) Nitrobenzene-d5	8.928	128	252396	26.033	ng/u1	0.00
24) 2-Nitrophenol-d4	9.646	143	255974	26.010	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.193	165	474663	24.475	ng/u1	0.00
31) 4-Chloroaniline-d4	10.699	131	710974	25.199	ng/u1	0.00
46) Dimethylphthalate-d6	13.834	166	1547987	28.662	ng/u1	0.00
49) Acenaphthylene-d8	14.116	160	1752699	28.904	ng/u1	0.00
54) 4-Nitrophenol-d4	14.640	143	228667	22.254	ng/u1	0.01
60) Fluorene-d10	15.440	176	1328044	29.248	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.557	200	157750	20.130	ng/u1	0.00
73) Anthracene-d10	17.328	188	2032680	30.060	ng/u1	-0.01
81) Pyrene-d10	19.692	212	2339433	30.279	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.845	264	2701944	30.901	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP083024\
Data File : BP021777.D
Acq On : 31 Aug 2024 08:59
Operator : RC/JU
Sample : P3733-12
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
A44A4

Quant Time: Sep 02 01:35:02 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082324.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Aug 29 23:56:59 2024
Response via : Initial Calibration

