

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090122\
 Data File : BP011601.D
 Acq On : 01 Sep 2022 11:23
 Operator : CG/JU
 Sample : SSTD01050
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010611

Quant Time: Sep 01 13:45:05 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 01 13:42:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	459447	20.000	ng/u1	0.00
20) Naphthalene-d8	10.787	136	1980983	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.616	164	1221856	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.369	188	2495603	20.000	ng/u1	0.00
79) Chrysene-d12	21.451	240	2365703	20.000	ng/u1	0.00
88) Perylene-d12	23.927	264	2299773	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.364	96	42214	3.896	ng/uL	0.00
4) Pyridine-d5	3.787	84	261053	8.967	ng/u1	0.00
7) Phenol-d5	7.116	99	351732	9.289	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.299	67	226353	9.919	ng/u1	0.00
11) 2-Chlorophenol-d4	7.499	132	280862	9.819	ng/u1	0.00
15) 4-Methylphenol-d8	8.675	113	269904	9.500	ng/u1	0.00
21) Nitrobenzene-d5	9.140	128	116425	9.978	ng/u1	0.00
24) 2-Nitrophenol-d4	9.863	143	87653	9.626	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.398	165	266013	9.754	ng/u1	0.00
31) 4-Chloroaniline-d4	10.922	131	417783	9.632	ng/u1	0.00
46) Dimethylphthalate-d6	14.022	166	816727	9.885	ng/u1	0.00
49) Acenaphthylene-d8	14.310	160	961943	9.718	ng/u1	0.00
54) 4-Nitrophenol-d4	14.786	143	111047	8.011	ng/u1	0.00
60) Fluorene-d10	15.610	176	732621	9.901	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.716	200	55108	6.753	ng/u1	0.00
73) Anthracene-d10	17.469	188	1095512	9.804	ng/u1	0.00
81) Pyrene-d10	19.698	212	1195330	10.063	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.768	264	1052695	9.465	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.399	88	44072	3.854	ng/uL	96
5) Pyridine	3.805	79	257828	8.965	ng/u1	96
6) Benzaldehyde	7.105	77	176223	9.991	ng/u1	97
8) Phenol	7.146	94	368696	9.470	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.393	93	304035	9.940	ng/u1	100
12) 2-Chlorophenol	7.534	128	303937	10.065	ng/u1	96
13) 2-Methylphenol	8.410	108	275688	9.329	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.510	45	401750	9.878	ng/u1	99
16) Acetophenone	8.799	105	472735	10.161	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.781	70	215140	9.977	ng/u1	98
18) 4-Methylphenol	8.740	108	300743	9.623	ng/u1	97
19) Hexachloroethane	9.063	117	121837	10.023	ng/u1	97
22) Nitrobenzene	9.181	77	293073	10.024	ng/u1	100
23) Isophorone	9.710	82	599067	9.517	ng/u1	100
25) 2-Nitrophenol	9.898	139	110758	9.743	ng/u1	97
26) 2,4-Dimethylphenol	9.951	107	324950	9.731	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.198	93	395894	9.991	ng/u1	100
29) 2,4-Dichlorophenol	10.428	162	274350	9.785	ng/u1	97
30) Naphthalene	10.840	128	1026484	10.049	ng/u1	100
32) 4-Chloroaniline	10.945	127	434785	9.787	ng/u1	98
33) Hexachlorobutadiene	11.128	225	187297	9.931	ng/u1	99
34) Caprolactam	11.704	113	66526	7.870	ng/u1	97
35) 4-Chloro-3-methylphenol	12.057	107	275121	9.727	ng/u1	98
36) 2-Methylnaphthalene	12.445	142	679432	9.957	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090122\
 Data File : BP011601.D
 Acq On : 01 Sep 2022 11:23
 Operator : CG/JU
 Sample : SSTD01050
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010611

Quant Time: Sep 01 13:45:05 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 01 13:42:41 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.663	142	682823	9.932	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.810	216	359468	9.771	ng/ul	98
40) Hexachlorocyclopentadiene	12.792	237	179031	8.559	ng/ul	100
41) 2,4,6-Trichlorophenol	13.045	196	191700	9.685	ng/ul	99
42) 2,4,5-Trichlorophenol	13.110	196	214171	9.780	ng/ul	99
43) 1,1'-Biphenyl	13.451	154	923122	9.994	ng/ul	98
44) 2-Chloronaphthalene	13.492	162	706365	10.012	ng/ul	99
45) 2-Nitroaniline	13.687	65	121531	9.239	ng/ul	98
47) Dimethylphthalate	14.069	163	841911	9.968	ng/ul	99
48) 2,6-Dinitrotoluene	14.187	165	113077	9.272	ng/ul	97
50) Acenaphthylene	14.339	152	1127186	9.916	ng/ul	99
51) 3-Nitroaniline	14.510	138	120696	7.984	ng/ul	97
52) Acenaphthene	14.681	153	737988	9.976	ng/ul	99
53) 2,4-Dinitrophenol	14.710	184	36642	7.091	ng/ul	97
55) 4-Nitrophenol	14.804	109	96843	8.702	ng/ul	95
56) Dibenzofuran	15.016	168	1049440	10.006	ng/ul	99
57) 2,4-Dinitrotoluene	14.969	165	159827	9.371	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.233	232	158633	9.816	ng/ul	98
59) Diethylphthalate	15.439	149	813024	9.806	ng/ul	100
61) Fluorene	15.663	166	851568	9.898	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.663	204	427474	9.992	ng/ul	99
63) 4-Nitroaniline	15.669	138	123393	8.453	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.728	198	62206	6.983	ng/ul#	94
67) N-Nitrosodiphenylamine	15.869	169	710318	9.955	ng/ul	100
68) 4-Bromophenyl-phenylether	16.557	248	249447	9.946	ng/ul	99
69) Hexachlorobenzene	16.669	284	297745	9.893	ng/ul	99
70) Atrazine	16.828	200	229060	8.872	ng/ul	99
71) Pentachlorophenol	17.010	266	115326	7.528	ng/ul	100
72) Phenanthrene	17.410	178	1309252	9.924	ng/ul	99
74) Anthracene	17.504	178	1315930	9.927	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.416	216	373007	10.053	ng/uL	100
76) Pentachlorobenzene	14.934	250	355718	10.122	ng/uL	99
77) Carbazole	17.769	167	1169493	9.855	ng/ul	100
78) Di-n-butylphthalate	18.327	149	1231442	9.568	ng/ul	99
80) Fluoranthene	19.374	202	1454683	10.135	ng/ul	97
82) Pyrene	19.721	202	1533008	10.258	ng/ul	100
83) Butylbenzylphthalate	20.604	149	406722	9.758	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.363	252	330569	7.767	ng/ul	99
85) Benzo(a)anthracene	21.433	228	1430377	9.909	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.368	149	675025	9.539	ng/ul	99
87) Chrysene	21.486	228	1394324	9.958	ng/ul	99
89) Di-n-octyl phthalate	22.327	149	1061279	9.023	ng/ul	100
90) Benzo(b)fluoranthene	23.168	252	1432601	10.138	ng/ul	99
91) Benzo(k)fluoranthene	23.221	252	1327828	9.666	ng/ul	99
93) Benzo(a)pyrene	23.821	252	1355762	9.783	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.533	276	1519171	9.430	ng/ul	98
95) Dibenzo(a,h)anthracene	26.551	278	1320227	9.527	ng/ul	98
96) Benzo(g,h,i)perylene	27.321	276	1283503	9.471	ng/ul	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090122\
 Data File : BP011601.D
 Acq On : 01 Sep 2022 11:23
 Operator : CG/JU
 Sample : SST01050
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST010611

Quant Time: Sep 01 13:45:05 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 01 13:42:41 2022
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 3.854 ng/uL
RT: 3.399 min Scan# 70
Delta R.T. -0.006 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion: 88 Resp: 44072
Ion Ratio Lower Upper
88 100
43 27.3 22.2 33.2
58 72.0 61.4 92.0



#5
Pyridine
Concen: 8.965 ng/uL
RT: 3.805 min Scan# 139
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion: 79 Resp: 257828
Ion Ratio Lower Upper
79 100
52 80.0 67.2 100.8
51 35.7 30.6 45.8









#6
Benzaldehyde
Concen: 9.991 ng/ul
RT: 7.105 min Scan# 700
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion: 77 Resp: 176223
Ion Ratio Lower Upper
77 100
105 101.0 82.4 123.6
106 93.9 79.1 118.7



#8
Phenol
Concen: 9.470 ng/ul
RT: 7.146 min Scan# 707
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion: 94 Resp: 368696
Ion Ratio Lower Upper
94 100
65 27.7 22.4 33.6
66 34.4 27.9 41.9









#10
Bis(2-Chloroethyl)ether
Concen: 9.940 ng/ul
RT: 7.393 min Scan# 749
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion: 93 Resp: 304035
Ion Ratio Lower Upper
93 100
63 78.2 62.2 93.2
95 32.2 25.9 38.9



#12
2-Chlorophenol
Concen: 10.065 ng/ul
RT: 7.534 min Scan# 773
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion:128 Resp: 303937
Ion Ratio Lower Upper
128 100
64 42.3 36.4 54.6
130 33.8 25.7 38.5









#13
2-Methylphenol
Concen: 9.329 ng/ul
RT: 8.410 min Scan# 911
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:108 Resp: 275688
Ion Ratio Lower Upper
108 100
107 85.8 70.3 105.5



#14
2,2'-oxybis(1-Chloropropane)
Concen: 9.878 ng/ul
RT: 8.510 min Scan# 939
Delta R.T. 0.006 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion: 45 Resp: 401750
Ion Ratio Lower Upper
45 100
77 14.3 11.4 17.0
79 11.6 8.4 12.6









#16
Acetophenone
Concen: 10.161 ng/ul
RT: 8.799 min Scan# 911
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:105 Resp: 472735
Ion Ratio Lower Upper
105 100
77 74.7 60.0 90.0
51 26.4 21.0 31.6



#17
N-Nitroso-di-n-propylamine
Concen: 9.977 ng/ul
RT: 8.781 min Scan# 985
Delta R.T. -0.006 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion: 70 Resp: 215140
Ion Ratio Lower Upper
70 100
42 55.9 46.3 69.5
101 8.7 8.1 12.1
130 21.4 17.3 25.9









#18
4-Methylphenol
Concen: 9.623 ng/ul
RT: 8.740 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:108 Resp: 300743
Ion Ratio Lower Upper
108 100
107 108.1 89.0 133.4



#19
Hexachloroethane
Concen: 10.023 ng/ul
RT: 9.063 min Scan# 1033
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion:117 Resp: 121837
Ion Ratio Lower Upper
117 100
201 100.9 82.7 124.1
199 61.6 51.4 77.2









#22
Nitrobenzene
Concen: 10.024 ng/ul
RT: 9.181 min Scan# 1053
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion: 77 Resp: 293073
Ion Ratio Lower Upper
77 100
123 46.9 37.6 56.4
65 13.5 11.0 16.4



#23
Isophorone
Concen: 9.517 ng/ul
RT: 9.710 min Scan# 1143
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion: 82 Resp: 599067
Ion Ratio Lower Upper
82 100
95 7.2 5.6 8.4
138 19.2 15.2 22.8









#25
2-Nitrophenol
Concen: 9.743 ng/ul
RT: 9.898 min Scan# 1175
Delta R.T. 0.006 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:139 Resp: 110758
Ion Ratio Lower Upper
139 100
65 46.0 38.5 57.7
109 28.7 22.3 33.5



#26
2,4-Dimethylphenol
Concen: 9.731 ng/ul
RT: 9.951 min Scan# 1184
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion:107 Resp: 324950
Ion Ratio Lower Upper
107 100
121 53.5 42.9 64.3
122 90.5 71.8 107.6





#27

Bis(2-Chloroethoxy)methane

Concen: 9.991 ng/ul

RT: 10.198 min Scan# 1226

Delta R.T. -0.000 min

Lab File: BP011601.D

Acq: 01 Sep 2022 11:23

Instrument :

BNA_P

ClientSampleId :

SSTD010611





Tgt Ion: 93 Resp: 395894

Ion Ratio Lower Upper

93 100

95 32.9 26.3 39.5

123 11.4 8.7 13.1



#29

2,4-Dichlorophenol

Concen: 9.785 ng/ul

RT: 10.428 min Scan# 1265

Delta R.T. -0.000 min

Lab File: BP011601.D

Acq: 01 Sep 2022 11:23




Tgt Ion:162 Resp: 274350


Ion Ratio Lower Upper

162 100

164 64.1 53.4 80.2

98 33.5 28.3 42.5









#30
Naphthalene
Concen: 10.049 ng/ul
RT: 10.840 min Scan# 1335
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:128 Resp: 1026484
Ion Ratio Lower Upper
128 100
129 11.1 8.8 13.2
127 12.9 10.2 15.4



#32
4-Chloroaniline
Concen: 9.787 ng/ul
RT: 10.945 min Scan# 1353
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion:127 Resp: 434785
Ion Ratio Lower Upper
127 100
129 31.9 26.3 39.5









#33
Hexachlorobutadiene
Concen: 9.931 ng/ul
RT: 11.128 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:225 Resp: 187297
Ion Ratio Lower Upper
225 100
223 64.0 50.5 75.7
227 63.6 51.1 76.7



#34
Caprolactam
Concen: 7.870 ng/ul
RT: 11.704 min Scan# 1482
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion:113 Resp: 66526
Ion Ratio Lower Upper
113 100
55 165.7 137.1 205.7
56 126.0 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 9.727 ng/ul
RT: 12.057 min Scan# 1542
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:107 Resp: 275121
Ion Ratio Lower Upper
107 100
144 27.4 21.4 32.0
142 85.0 66.5 99.7



#36
2-Methylnaphthalene
Concen: 9.957 ng/ul
RT: 12.445 min Scan# 1608
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion:142 Resp: 679432
Ion Ratio Lower Upper
142 100
141 88.1 71.5 107.3









#37
1-Methylnaphthalene
Concen: 9.932 ng/ul
RT: 12.663 min Scan# 1645
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:142 Resp: 682823
Ion Ratio Lower Upper
142 100
141 93.4 74.2 111.4
116 3.7 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 9.771 ng/ul
RT: 12.810 min Scan# 1670
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion:216 Resp: 359468
Ion Ratio Lower Upper
216 100
214 80.7 62.6 94.0
179 20.2 16.1 24.1
108 16.7 13.2 19.8





#40

Hexachlorocyclopentadiene

Concen: 8.559 ng/ul

RT: 12.792 min Scan# 1667

Delta R.T. -0.000 min

Lab File: BP011601.D

Acq: 01 Sep 2022 11:23

Instrument :

BNA_P

ClientSampleId :

SSTD010611





Tgt Ion:237 Resp: 179031

Ion Ratio Lower Upper

237 100

235 61.9 49.4 74.0

272 13.4 10.5 15.7



#41

2,4,6-Trichlorophenol

Concen: 9.685 ng/ul

RT: 13.045 min Scan# 1710

Delta R.T. -0.000 min

Lab File: BP011601.D

Acq: 01 Sep 2022 11:23




Tgt Ion:196 Resp: 191700


Ion Ratio Lower Upper

196 100

198 94.9 75.1 112.7

200 30.6 24.9 37.3









#42
2,4,5-Trichlorophenol
Concen: 9.780 ng/ul
RT: 13.110 min Scan# 1721
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:196 Resp: 214171
Ion Ratio Lower Upper
196 100
198 94.9 77.0 115.6
200 30.7 24.1 36.1



#43
1,1'-Biphenyl
Concen: 9.994 ng/ul
RT: 13.451 min Scan# 1779
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion:154 Resp: 923122
Ion Ratio Lower Upper
154 100
153 40.1 31.4 47.0
76 11.7 9.8 14.8









#44
2-Chloronaphthalene
Concen: 10.012 ng/ul
RT: 13.492 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:162 Resp: 706365
Ion Ratio Lower Upper
162 100
164 32.5 26.5 39.7
127 37.8 30.3 45.5



#45
2-Nitroaniline
Concen: 9.239 ng/ul
RT: 13.687 min Scan# 1819
Delta R.T. -0.006 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion: 65 Resp: 121531
Ion Ratio Lower Upper
65 100
92 70.5 55.4 83.2
138 108.4 89.1 133.7









#47
Dimethylphthalate
Concen: 9.968 ng/ul
RT: 14.069 min Scan# 1884
Delta R.T. -0.006 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:163 Resp: 841911
Ion Ratio Lower Upper
163 100
194 3.8 3.3 4.9
164 10.4 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 9.272 ng/ul
RT: 14.187 min Scan# 1904
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion:165 Resp: 113077
Ion Ratio Lower Upper
165 100
89 58.5 49.1 73.7
121 26.1 19.8 29.8









#50
Acenaphthylene
Concen: 9.916 ng/ul
RT: 14.339 min Scan# 1930
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:152 Resp: 1127186
Ion Ratio Lower Upper
152 100
151 20.0 15.8 23.6
153 12.9 10.5 15.7



#51
3-Nitroaniline
Concen: 7.984 ng/ul
RT: 14.510 min Scan# 1959
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion:138 Resp: 120696
Ion Ratio Lower Upper
138 100
108 21.9 17.0 25.4
92 114.4 88.2 132.4









#52
Acenaphthene
Concen: 9.976 ng/ul
RT: 14.681 min Scan# 1988
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:153 Resp: 737988
Ion Ratio Lower Upper
153 100
152 48.4 39.0 58.6
154 87.5 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 7.091 ng/ul
RT: 14.710 min Scan# 1993
Delta R.T. -0.006 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion:184 Resp: 36642
Ion Ratio Lower Upper
184 100
63 63.7 49.8 74.6
154 68.4 52.3 78.5









#55
4-Nitrophenol
Concen: 8.702 ng/ul
RT: 14.804 min Scan# 2009
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:109 Resp: 96843
Ion Ratio Lower Upper
109 100
139 142.9 108.4 162.6
65 135.8 104.7 157.1



#56
Dibenzofuran
Concen: 10.006 ng/ul
RT: 15.016 min Scan# 2045
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion:168 Resp: 1049440
Ion Ratio Lower Upper
168 100
139 35.9 29.1 43.7









#57
2,4-Dinitrotoluene
Concen: 9.371 ng/ul
RT: 14.969 min Scan# 2037
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:165 Resp: 159827
Ion Ratio Lower Upper
165 100
63 44.2 37.3 55.9
182 2.7 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 9.816 ng/ul
RT: 15.233 min Scan# 2082
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion:232 Resp: 158633
Ion Ratio Lower Upper
232 100
131 44.8 37.7 56.5
130 2.5 2.0 3.0
166 31.2 25.5 38.3









#59
Diethylphthalate
Concen: 9.806 ng/ul
RT: 15.439 min Scan# 2117
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:149 Resp: 813024
Ion Ratio Lower Upper
149 100
177 22.1 17.7 26.5
150 12.4 9.9 14.9



#61
Fluorene
Concen: 9.898 ng/ul
RT: 15.663 min Scan# 2155
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion:166 Resp: 851568
Ion Ratio Lower Upper
166 100
165 97.3 77.7 116.5
167 13.3 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 9.992 ng/ul
RT: 15.663 min Scan# 2155
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:204 Resp: 427474
Ion Ratio Lower Upper
204 100
206 33.4 26.5 39.7
141 55.6 45.0 67.6



#63
4-Nitroaniline
Concen: 8.453 ng/ul
RT: 15.669 min Scan# 2156
Delta R.T. -0.006 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion:138 Resp: 123393
Ion Ratio Lower Upper
138 100
92 49.5 39.1 58.7
108 81.1 65.5 98.3









#66
 4,6-Dinitro-2-methylphenol
 Concen: 6.983 ng/ul
 RT: 15.728 min Scan# 2166
 Delta R.T. -0.000 min
 Lab File: BP011601.D
 Acq: 01 Sep 2022 11:23




Instrument :
 BNA_P
 ClientSampleId :
 SSTD010611


Tgt Ion:198 Resp: 62206
 Ion Ratio Lower Upper
 198 100
 182 5.3 3.2 4.8#
 77 27.5 19.4 29.0



#67
 N-Nitrosodiphenylamine
 Concen: 9.955 ng/ul
 RT: 15.869 min Scan# 2190
 Delta R.T. -0.000 min
 Lab File: BP011601.D
 Acq: 01 Sep 2022 11:23

Tgt Ion:169 Resp: 710318
 Ion Ratio Lower Upper
 169 100
 168 66.7 53.8 80.6
 167 36.6 29.2 43.8









#68
4-Bromophenyl-phenylether
Concen: 9.946 ng/ul
RT: 16.557 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:248 Resp: 249447
Ion Ratio Lower Upper
248 100
250 97.1 79.0 118.4
141 70.3 57.1 85.7



#69
Hexachlorobenzene
Concen: 9.893 ng/ul
RT: 16.669 min Scan# 2326
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion:284 Resp: 297745
Ion Ratio Lower Upper
284 100
142 34.1 27.5 41.3
249 29.1 23.9 35.9





#70

Atrazine

Concen: 8.872 ng/ul

RT: 16.828 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP011601.D

Acq: 01 Sep 2022 11:23

Instrument :

BNA_P

ClientSampleId :

SSTD010611





Tgt Ion:200 Resp: 229060

Ion Ratio Lower Upper

200 100

173 25.8 21.4 32.2

215 49.4 40.0 60.0



#71

Pentachlorophenol

Concen: 7.528 ng/ul

RT: 17.010 min Scan# 2384

Delta R.T. -0.000 min

Lab File: BP011601.D

Acq: 01 Sep 2022 11:23




Tgt Ion:266 Resp: 115326


Ion Ratio Lower Upper

266 100

264 63.5 50.8 76.2

268 63.1 51.0 76.4









#72
Phenanthrene
Concen: 9.924 ng/ul
RT: 17.410 min Scan# 2452
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:178 Resp: 1309252
Ion Ratio Lower Upper
178 100
179 15.5 12.6 18.8
176 18.9 15.5 23.3



#74
Anthracene
Concen: 9.927 ng/ul
RT: 17.504 min Scan# 2468
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion:178 Resp: 1315930
Ion Ratio Lower Upper
178 100
179 15.5 12.0 18.0
176 18.6 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 10.053 ng/uL

RT: 13.416 min Scan# 1773

Delta R.T. -0.000 min

Lab File: BP011601.D

Acq: 01 Sep 2022 11:23

Instrument :

BNA_P

ClientSampleId :

SSTD010611





Tgt Ion:216 Resp: 373007

Ion Ratio Lower Upper

216 100

214 77.2 61.7 92.5

218 48.0 38.6 58.0



#76

Pentachlorobenzene

Concen: 10.122 ng/uL

RT: 14.934 min Scan# 2031

Delta R.T. 0.006 min

Lab File: BP011601.D

Acq: 01 Sep 2022 11:23




Tgt Ion:250 Resp: 355718


Ion Ratio Lower Upper

250 100

248 63.0 50.0 75.0

252 63.1 51.6 77.4









#77
 Carbazole
 Concen: 9.855 ng/ul
 RT: 17.769 min Scan# 2513
 Delta R.T. -0.000 min
 Lab File: BP011601.D
 Acq: 01 Sep 2022 11:23




Instrument : BNA_P
 ClientSampleId : SSTD010611


Tgt Ion:167 Resp: 1169493
 Ion Ratio Lower Upper
 167 100
 166 21.4 17.1 25.7
 139 12.9 9.9 14.9



#78
 Di-n-butylphthalate
 Concen: 9.568 ng/ul
 RT: 18.327 min Scan# 2608
 Delta R.T. -0.000 min
 Lab File: BP011601.D
 Acq: 01 Sep 2022 11:23

Tgt Ion:149 Resp: 1231442
 Ion Ratio Lower Upper
 149 100
 150 9.5 7.2 10.8
 104 5.3 4.1 6.1









#80
Fluoranthene
Concen: 10.135 ng/ul
RT: 19.374 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:202 Resp: 1454683
Ion Ratio Lower Upper
202 100
101 10.6 9.4 14.2
100 7.7 7.3 10.9



#82
Pyrene
Concen: 10.258 ng/ul
RT: 19.721 min Scan# 2845
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion:202 Resp: 1533008
Ion Ratio Lower Upper
202 100
101 13.2 10.6 16.0
100 10.7 8.6 12.8





#83

Butylbenzylphthalate

Concen: 9.758 ng/ul

RT: 20.604 min Scan# 2995

Delta R.T. -0.000 min

Lab File: BP011601.D

Acq: 01 Sep 2022 11:23

Instrument :

BNA_P

ClientSampleId :

SSTD010611





Tgt Ion:149 Resp: 406722

Ion Ratio Lower Upper

149 100

91 68.4 54.4 81.6

206 20.9 16.7 25.1



#84

3,3'-Dichlorobenzidine

Concen: 7.767 ng/ul

RT: 21.363 min Scan# 3124

Delta R.T. -0.000 min

Lab File: BP011601.D

Acq: 01 Sep 2022 11:23




Tgt Ion:252 Resp: 330569


Ion Ratio Lower Upper

252 100

254 63.9 52.0 78.0

126 11.2 8.7 13.1





#85

Benzo(a)anthracene

Concen: 9.909 ng/ul

RT: 21.433 min Scan# 3136

Delta R.T. -0.000 min

Lab File: BP011601.D

Acq: 01 Sep 2022 11:23

Instrument :

BNA_P

ClientSampleId :

SSTD010611





Tgt Ion:228 Resp: 1430377

Ion Ratio Lower Upper

228 100

229 20.1 15.6 23.4

226 27.4 22.2 33.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 9.539 ng/ul

RT: 21.368 min Scan# 3125

Delta R.T. -0.000 min

Lab File: BP011601.D

Acq: 01 Sep 2022 11:23




Tgt Ion:149 Resp: 675025


Ion Ratio Lower Upper

149 100

167 26.3 21.6 32.4

279 3.6 3.2 4.8





#87

Chrysene

Concen: 9.958 ng/ul

RT: 21.486 min Scan# 3145

Delta R.T. -0.000 min

Lab File: BP011601.D

Acq: 01 Sep 2022 11:23

Instrument :

BNA_P

ClientSampleId :

SSTD010611





Tgt Ion:228 Resp: 1394324

Ion Ratio Lower Upper

228 100

226 30.9 24.3 36.5

229 19.9 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 9.023 ng/ul




RT: 22.327 min Scan# 3288


Delta R.T. -0.000 min

Lab File: BP011601.D

Acq: 01 Sep 2022 11:23

Tgt Ion:149 Resp: 1061279









#90
Benzo(b)fluoranthene
Concen: 10.138 ng/ul
RT: 23.168 min Scan# 3431
Delta R.T. -0.000 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23




Instrument :
BNA_P
ClientSampleId :
SSTD010611


Tgt Ion:252 Resp: 1432601
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.8
125 9.9 8.2 12.4



#91
Benzo(k)fluoranthene
Concen: 9.666 ng/ul
RT: 23.221 min Scan# 3440
Delta R.T. 0.006 min
Lab File: BP011601.D
Acq: 01 Sep 2022 11:23

Tgt Ion:252 Resp: 1327828
Ion Ratio Lower Upper
252 100
253 22.1 18.0 27.0
125 9.7 8.2 12.4





#93

Benzo(a)pyrene

Concen: 9.783 ng/ul

RT: 23.821 min Scan# 3

Delta R.T. 0.006 min

Lab File: BP011601.D

Acq: 01 Sep 2022 11:23

Instrument :

BNA_P

ClientSampleId :

SSTD010611





Tgt Ion:252 Resp: 1355762

Ion Ratio Lower Upper

252 100

253 22.1 17.8 26.6

125 10.7 8.9 13.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 9.430 ng/ul

RT: 26.533 min Scan# 4003

Delta R.T. 0.006 min

Lab File: BP011601.D

Acq: 01 Sep 2022 11:23




Tgt Ion:276 Resp: 1519171


Ion Ratio Lower Upper

276 100

138 23.0 17.2 25.8

277 25.0 19.4 29.2









#95
 Dibenzo(a,h)anthracene
 Concen: 9.527 ng/ul
 RT: 26.551 min Scan# 4005
 Delta R.T. 0.006 min
 Lab File: BP011601.D
 Acq: 01 Sep 2022 11:23

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010611

Tgt Ion: 278 Resp: 1320227

Ion	Ratio	Lower	Upper
278	100		
139	15.0	12.4	18.6
279	23.2	19.3	28.9



#96
 Benzo(g,h,i)perylene
 Concen: 9.471 ng/ul
 RT: 27.321 min Scan# 4137
 Delta R.T. 0.006 min
 Lab File: BP011601.D
 Acq: 01 Sep 2022 11:23

Tgt Ion: 276 Resp: 1283503

Ion	Ratio	Lower	Upper
276	100		
138	22.3	16.6	24.8
277	24.2	19.0	28.6

