

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090122\
 Data File : BP011602.D
 Acq On : 01 Sep 2022 11:58
 Operator : CG/JU
 Sample : SSTD02051
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020612

Quant Time: Sep 01 13:45:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 01 13:42:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	500491	20.000	ng/u1	0.00
20) Naphthalene-d8	10.793	136	2118505	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.616	164	1290779	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.369	188	2619321	20.000	ng/u1	0.00
79) Chrysene-d12	21.451	240	2481070	20.000	ng/u1	0.00
88) Perylene-d12	23.927	264	2446958	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.364	96	87733	7.433	ng/uL	0.00
4) Pyridine-d5	3.781	84	597465	18.839	ng/u1	0.00
7) Phenol-d5	7.122	99	780017	18.910	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.299	67	483522	19.451	ng/u1	0.00
11) 2-Chlorophenol-d4	7.499	132	620441	19.912	ng/u1	0.00
15) 4-Methylphenol-d8	8.675	113	608622	19.665	ng/u1	0.00
21) Nitrobenzene-d5	9.140	128	272405	21.831	ng/u1	0.00
24) 2-Nitrophenol-d4	9.863	143	222152	22.812	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.398	165	598727	20.528	ng/u1	0.00
31) 4-Chloroaniline-d4	10.922	131	901632	19.438	ng/u1	0.00
46) Dimethylphthalate-d6	14.028	166	1751421	20.065	ng/u1	0.00
49) Acenaphthylene-d8	14.310	160	2117489	20.249	ng/u1	0.00
54) 4-Nitrophenol-d4	14.792	143	276478	18.881	ng/u1	0.00
60) Fluorene-d10	15.610	176	1552814	19.865	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.716	200	145184	16.950	ng/u1	0.00
73) Anthracene-d10	17.469	188	2335400	19.912	ng/u1	0.00
81) Pyrene-d10	19.698	212	2569534	20.627	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.768	264	2378290	20.098	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	92493	7.425	ng/uL	97
5) Pyridine	3.805	79	588726	18.792	ng/u1	92
6) Benzaldehyde	7.105	77	423711	22.053	ng/u1	99
8) Phenol	7.146	94	810449	19.110	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.393	93	646779	19.411	ng/u1	100
12) 2-Chlorophenol	7.534	128	659190	20.039	ng/u1	98
13) 2-Methylphenol	8.410	108	610318	18.959	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.504	45	844362	19.058	ng/u1	99
16) Acetophenone	8.799	105	1018689	20.101	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.787	70	479680	20.421	ng/u1	97
18) 4-Methylphenol	8.740	108	666888	19.589	ng/u1	99
19) Hexachloroethane	9.063	117	257371	19.436	ng/u1	99
22) Nitrobenzene	9.181	77	655130	20.952	ng/u1	99
23) Isophorone	9.710	82	1344560	19.973	ng/u1	100
25) 2-Nitrophenol	9.893	139	276536	22.746	ng/u1	98
26) 2,4-Dimethylphenol	9.951	107	717660	20.095	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.198	93	862002	20.342	ng/u1	99
29) 2,4-Dichlorophenol	10.428	162	618195	20.618	ng/u1	98
30) Naphthalene	10.840	128	2173802	19.900	ng/u1	99
32) 4-Chloroaniline	10.946	127	928899	19.552	ng/u1	100
33) Hexachlorobutadiene	11.134	225	401479	19.905	ng/u1	98
34) Caprolactam	11.704	113	170603	18.873	ng/u1	97
35) 4-Chloro-3-methylphenol	12.057	107	621989	20.564	ng/u1	99
36) 2-Methylnaphthalene	12.445	142	1467480	20.109	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090122\
 Data File : BP011602.D
 Acq On : 01 Sep 2022 11:58
 Operator : CG/JU
 Sample : SST02051
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SST020612


Quant Time: Sep 01 13:45:48 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 01 13:42:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.663	142	1476003	20.076	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.810	216	781501	20.109	ng/ul	100
40) Hexachlorocyclopentadiene	12.792	237	414946	18.778	ng/ul	98
41) 2,4,6-Trichlorophenol	13.045	196	444699	21.266	ng/ul	98
42) 2,4,5-Trichlorophenol	13.110	196	486063	21.011	ng/ul	99
43) 1,1'-Biphenyl	13.451	154	1974316	20.233	ng/ul	99
44) 2-Chloronaphthalene	13.492	162	1499295	20.117	ng/ul	100
45) 2-Nitroaniline	13.692	65	313189	22.539	ng/ul	98
47) Dimethylphthalate	14.075	163	1798199	20.154	ng/ul	100
48) 2,6-Dinitrotoluene	14.187	165	283519	22.007	ng/ul	95
50) Acenaphthylene	14.339	152	2433072	20.261	ng/ul	99
51) 3-Nitroaniline	14.510	138	323952	20.286	ng/ul#	97
52) Acenaphthene	14.681	153	1562598	19.995	ng/ul	99
53) 2,4-Dinitrophenol	14.716	184	90237	16.530	ng/ul	98
55) 4-Nitrophenol	14.804	109	224614	19.106	ng/ul	96
56) Dibenzofuran	15.016	168	2221549	20.050	ng/ul	99
57) 2,4-Dinitrotoluene	14.969	165	405101	22.484	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.234	232	369041	21.617	ng/ul	98
59) Diethylphthalate	15.439	149	1753905	20.025	ng/ul	99
61) Fluorene	15.663	166	1822775	20.055	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.663	204	909768	20.129	ng/ul	98
63) 4-Nitroaniline	15.675	138	324843	21.065	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.728	198	162997	17.432	ng/ul#	96
67) N-Nitrosodiphenylamine	15.869	169	1520163	20.298	ng/ul	100
68) 4-Bromophenyl-phenylether	16.557	248	531093	20.177	ng/ul	98
69) Hexachlorobenzene	16.669	284	631175	19.982	ng/ul	99
70) Atrazine	16.828	200	520912	19.222	ng/ul	99
71) Pentachlorophenol	17.010	266	294418	18.311	ng/ul	98
72) Phenanthrene	17.410	178	2764280	19.964	ng/ul	100
74) Anthracene	17.504	178	2777709	19.965	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.416	216	797769	20.485	ng/ul	99
76) Pentachlorobenzene	14.928	250	753904	20.439	ng/ul	99
77) Carbazole	17.769	167	2484202	19.945	ng/ul	99
78) Di-n-butylphthalate	18.327	149	2842797	21.044	ng/ul	99
80) Fluoranthene	19.369	202	3123662	20.751	ng/ul	99
82) Pyrene	19.722	202	3236279	20.649	ng/ul	99
83) Butylbenzylphthalate	20.604	149	1032342	23.615	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.363	252	792675	17.759	ng/ul	99
85) Benzo(a)anthracene	21.433	228	3069124	20.273	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.369	149	1684151	22.693	ng/ul	99
87) Chrysene	21.486	228	2939973	20.020	ng/ul	99
89) Di-n-octyl phthalate	22.327	149	2677164	21.393	ng/ul	100
90) Benzo(b)fluoranthene	23.168	252	3030637	20.157	ng/ul	99
91) Benzo(k)fluoranthene	23.215	252	2959316	20.247	ng/ul	99
93) Benzo(a)pyrene	23.815	252	3004402	20.376	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.527	276	3404977	19.865	ng/ul	98
95) Dibenzo(a,h)anthracene	26.545	278	2944802	19.971	ng/ul	99
96) Benzo(g,h,i)perylene	27.315	276	2831447	19.636	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 01 13:45:48 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090122.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Sep 01 13:42:41 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 7.425 ng/uL
 RT: 3.405 min Scan# 71
 Delta R.T. 0.000 min
 Lab File: BP011602.D
 Acq: 01 Sep 2022 11:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020612


Tgt Ion: 88 Resp: 92493
 Ion Ratio Lower Upper
 88 100
 43 27.1 22.2 33.2
 58 73.4 61.4 92.0



#5
 Pyridine
 Concen: 18.792 ng/uL
 RT: 3.805 min Scan# 139
 Delta R.T. 0.000 min
 Lab File: BP011602.D
 Acq: 01 Sep 2022 11:58

Tgt Ion: 79 Resp: 588726
 Ion Ratio Lower Upper
 79 100
 52 75.5 67.2 100.8
 51 36.1 30.6 45.8









#6
Benzaldehyde
Concen: 22.053 ng/ul
RT: 7.105 min Scan# 700
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion: 77 Resp: 423711
Ion Ratio Lower Upper
77 100
105 103.1 82.4 123.6
106 97.4 79.1 118.7



#8
Phenol
Concen: 19.110 ng/ul
RT: 7.146 min Scan# 707
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion: 94 Resp: 810449
Ion Ratio Lower Upper
94 100
65 28.1 22.4 33.6
66 35.3 27.9 41.9









#10
Bis(2-Chloroethyl)ether
Concen: 19.411 ng/ul
RT: 7.393 min Scan# 749
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion: 93 Resp: 646779
Ion Ratio Lower Upper
93 100
63 77.5 62.2 93.2
95 32.6 25.9 38.9



#12
2-Chlorophenol
Concen: 20.039 ng/ul
RT: 7.534 min Scan# 773
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:128 Resp: 659190
Ion Ratio Lower Upper
128 100
64 43.2 36.4 54.6
130 32.3 25.7 38.5









#13
2-Methylphenol
Concen: 18.959 ng/ul
RT: 8.410 min Scan# 911
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:108 Resp: 610318
Ion Ratio Lower Upper
108 100
107 89.5 70.3 105.5



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.058 ng/ul
RT: 8.504 min Scan# 938
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion: 45 Resp: 844362
Ion Ratio Lower Upper
45 100
77 14.8 11.4 17.0
79 10.5 8.4 12.6









#16
Acetophenone
Concen: 20.101 ng/ul
RT: 8.799 min Scan# 988
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:105 Resp: 1018689
Ion Ratio Lower Upper
105 100
77 75.6 60.0 90.0
51 25.7 21.0 31.6



#17
N-Nitroso-di-n-propylamine
Concen: 20.421 ng/ul
RT: 8.787 min Scan# 986
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion: 70 Resp: 479680
Ion Ratio Lower Upper
70 100
42 55.0 46.3 69.5
101 10.2 8.1 12.1
130 22.7 17.3 25.9









#18
4-Methylphenol
Concen: 19.589 ng/ul
RT: 8.740 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:108 Resp: 666888
Ion Ratio Lower Upper
108 100
107 109.6 89.0 133.4



#19
Hexachloroethane
Concen: 19.436 ng/ul
RT: 9.063 min Scan# 1033
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:117 Resp: 257371
Ion Ratio Lower Upper
117 100
201 103.1 82.7 124.1
199 62.4 51.4 77.2









#22
Nitrobenzene
Concen: 20.952 ng/ul
RT: 9.181 min Scan# 1053
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion: 77 Resp: 655130
Ion Ratio Lower Upper
77 100
123 47.6 37.6 56.4
65 13.4 11.0 16.4



#23
Isophorone
Concen: 19.973 ng/ul
RT: 9.710 min Scan# 1143
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion: 82 Resp: 1344560
Ion Ratio Lower Upper
82 100
95 6.6 5.6 8.4
138 19.0 15.2 22.8









#25
2-Nitrophenol
Concen: 22.746 ng/ul
RT: 9.893 min Scan# 1174
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:139 Resp: 276536
Ion Ratio Lower Upper
139 100
65 49.2 38.5 57.7
109 29.3 22.3 33.5



#26
2,4-Dimethylphenol
Concen: 20.095 ng/ul
RT: 9.951 min Scan# 1184
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:107 Resp: 717660
Ion Ratio Lower Upper
107 100
121 52.6 42.9 64.3
122 89.7 71.8 107.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.342 ng/ul
 RT: 10.198 min Scan# 1226
 Delta R.T. 0.000 min
 Lab File: BP011602.D
 Acq: 01 Sep 2022 11:58




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020612


Tgt Ion: 93 Resp: 862002
 Ion Ratio Lower Upper
 93 100
 95 32.1 26.3 39.5
 123 11.5 8.7 13.1



#29
 2,4-Dichlorophenol
 Concen: 20.618 ng/ul
 RT: 10.428 min Scan# 1265
 Delta R.T. 0.000 min
 Lab File: BP011602.D
 Acq: 01 Sep 2022 11:58

Tgt Ion: 162 Resp: 618195
 Ion Ratio Lower Upper
 162 100
 164 65.0 53.4 80.2
 98 34.3 28.3 42.5









#30
Naphthalene
Concen: 19.900 ng/ul
RT: 10.840 min Scan# 1335
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:128 Resp: 2173802
Ion Ratio Lower Upper
128 100
129 10.9 8.8 13.2
127 13.1 10.2 15.4



#32
4-Chloroaniline
Concen: 19.552 ng/ul
RT: 10.946 min Scan# 1353
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:127 Resp: 928899
Ion Ratio Lower Upper
127 100
129 32.8 26.3 39.5









#33
Hexachlorobutadiene
Concen: 19.905 ng/ul
RT: 11.134 min Scan# 1113
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:225 Resp: 401479
Ion Ratio Lower Upper
225 100
223 61.6 50.5 75.7
227 62.5 51.1 76.7



#34
Caprolactam
Concen: 18.873 ng/ul
RT: 11.704 min Scan# 1482
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:113 Resp: 170603
Ion Ratio Lower Upper
113 100
55 175.9 137.1 205.7
56 131.6 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 20.564 ng/ul
RT: 12.057 min Scan# 1542
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:107 Resp: 621989
Ion Ratio Lower Upper
107 100
144 27.0 21.4 32.0
142 83.6 66.5 99.7



#36
2-Methylnaphthalene
Concen: 20.109 ng/ul
RT: 12.445 min Scan# 1608
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:142 Resp: 1467480
Ion Ratio Lower Upper
142 100
141 89.7 71.5 107.3









#37
1-Methylnaphthalene
Concen: 20.076 ng/ul
RT: 12.663 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:142 Resp: 1476003
Ion Ratio Lower Upper
142 100
141 91.7 74.2 111.4
116 3.7 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 20.109 ng/ul
RT: 12.810 min Scan# 1670
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:216 Resp: 781501
Ion Ratio Lower Upper
216 100
214 78.1 62.6 94.0
179 20.3 16.1 24.1
108 16.9 13.2 19.8









#40
Hexachlorocyclopentadiene
Concen: 18.778 ng/ul
RT: 12.792 min Scan# 1667
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:237 Resp: 414946
Ion Ratio Lower Upper
237 100
235 63.2 49.4 74.0
272 13.4 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 21.266 ng/ul
RT: 13.045 min Scan# 1710
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:196 Resp: 444699
Ion Ratio Lower Upper
196 100
198 95.9 75.1 112.7
200 31.0 24.9 37.3





#42

2,4,5-Trichlorophenol

Concen: 21.011 ng/ul

RT: 13.110 min Scan# 1721

Delta R.T. 0.000 min

Lab File: BP011602.D

Acq: 01 Sep 2022 11:58

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:196 Resp: 486063

Ion Ratio Lower Upper

196 100

198 95.8 77.0 115.6

200 30.9 24.1 36.1



#43

1,1'-Biphenyl

Concen: 20.233 ng/ul

RT: 13.451 min Scan# 1779

Delta R.T. 0.000 min

Lab File: BP011602.D

Acq: 01 Sep 2022 11:58




Tgt Ion:154 Resp: 1974316


Ion Ratio Lower Upper

154 100

153 40.0 31.4 47.0

76 12.1 9.8 14.8





#44

2-Chloronaphthalene

Concen: 20.117 ng/ul

RT: 13.492 min Scan# 1786

Delta R.T. 0.000 min

Lab File: BP011602.D

Acq: 01 Sep 2022 11:58

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:162 Resp: 1499295

Ion Ratio Lower Upper

162 100

164 33.2 26.5 39.7

127 37.9 30.3 45.5



#45

2-Nitroaniline

Concen: 22.539 ng/ul

RT: 13.692 min Scan# 1820

Delta R.T. 0.000 min

Lab File: BP011602.D

Acq: 01 Sep 2022 11:58




Tgt Ion: 65 Resp: 313189


Ion Ratio Lower Upper

65 100

92 69.8 55.4 83.2

138 115.3 89.1 133.7









#47
Dimethylphthalate
Concen: 20.154 ng/ul
RT: 14.075 min Scan# 1885
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:163 Resp: 1798199
Ion Ratio Lower Upper
163 100
194 4.1 3.3 4.9
164 10.0 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 22.007 ng/ul
RT: 14.187 min Scan# 1904
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:165 Resp: 283519
Ion Ratio Lower Upper
165 100
89 56.4 49.1 73.7
121 23.3 19.8 29.8









#50
Acenaphthylene
Concen: 20.261 ng/ul
RT: 14.339 min Scan# 1930
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:152 Resp: 2433072
Ion Ratio Lower Upper
152 100
151 20.1 15.8 23.6
153 13.2 10.5 15.7



#51
3-Nitroaniline
Concen: 20.286 ng/ul
RT: 14.510 min Scan# 1959
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:138 Resp: 323952
Ion Ratio Lower Upper
138 100
108 15.8 17.0 25.4#
92 112.0 88.2 132.4









#52
Acenaphthene
Concen: 19.995 ng/ul
RT: 14.681 min Scan# 1988
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:153 Resp: 1562598
Ion Ratio Lower Upper
153 100
152 48.3 39.0 58.6
154 89.7 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 16.530 ng/ul
RT: 14.716 min Scan# 1994
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:184 Resp: 90237
Ion Ratio Lower Upper
184 100
63 61.9 49.8 74.6
154 62.5 52.3 78.5









#55
4-Nitrophenol
Concen: 19.106 ng/ul
RT: 14.804 min Scan# 2009
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:109 Resp: 224614
Ion Ratio Lower Upper
109 100
139 141.4 108.4 162.6
65 133.5 104.7 157.1



#56
Dibenzofuran
Concen: 20.050 ng/ul
RT: 15.016 min Scan# 2045
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:168 Resp: 2221549
Ion Ratio Lower Upper
168 100
139 35.8 29.1 43.7









#57
2,4-Dinitrotoluene
Concen: 22.484 ng/ul
RT: 14.969 min Scan# 2037
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:165 Resp: 405101
Ion Ratio Lower Upper
165 100
63 42.3 37.3 55.9
182 2.7 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.617 ng/ul
RT: 15.234 min Scan# 2082
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:232 Resp: 369041
Ion Ratio Lower Upper
232 100
131 45.3 37.7 56.5
130 2.3 2.0 3.0
166 31.4 25.5 38.3









#59
Diethylphthalate
Concen: 20.025 ng/ul
RT: 15.439 min Scan# 2117
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:149 Resp: 1753905
Ion Ratio Lower Upper
149 100
177 22.5 17.7 26.5
150 12.3 9.9 14.9



#61
Fluorene
Concen: 20.055 ng/ul
RT: 15.663 min Scan# 2155
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:166 Resp: 1822775
Ion Ratio Lower Upper
166 100
165 99.2 77.7 116.5
167 13.6 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 20.129 ng/ul
RT: 15.663 min Scan# 2155
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:204 Resp: 909768
Ion Ratio Lower Upper
204 100
206 32.9 26.5 39.7
141 54.7 45.0 67.6



#63
4-Nitroaniline
Concen: 21.065 ng/ul
RT: 15.675 min Scan# 2157
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:138 Resp: 324843
Ion Ratio Lower Upper
138 100
92 47.2 39.1 58.7
108 75.2 65.5 98.3









#66
4,6-Dinitro-2-methylphenol
Concen: 17.432 ng/ul
RT: 15.728 min Scan# 2166
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:198 Resp: 162997
Ion Ratio Lower Upper
198 100
182 5.2 3.2 4.8#
77 26.1 19.4 29.0



#67
N-Nitrosodiphenylamine
Concen: 20.298 ng/ul
RT: 15.869 min Scan# 2190
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:169 Resp: 1520163
Ion Ratio Lower Upper
169 100
168 66.8 53.8 80.6
167 36.6 29.2 43.8









#68
4-Bromophenyl-phenylether
Concen: 20.177 ng/ul
RT: 16.557 min Scan# 2307
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:248 Resp: 531093
Ion Ratio Lower Upper
248 100
250 97.1 79.0 118.4
141 69.7 57.1 85.7



#69
Hexachlorobenzene
Concen: 19.982 ng/ul
RT: 16.669 min Scan# 2326
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:284 Resp: 631175
Ion Ratio Lower Upper
284 100
142 33.3 27.5 41.3
249 29.5 23.9 35.9





#70

Atrazine

Concen: 19.222 ng/ul

RT: 16.828 min Scan# 2353

Delta R.T. 0.000 min

Lab File: BP011602.D

Acq: 01 Sep 2022 11:58

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:200 Resp: 520912

Ion Ratio Lower Upper

200 100

173 25.9 21.4 32.2

215 49.2 40.0 60.0



#71

Pentachlorophenol

Concen: 18.311 ng/ul

RT: 17.010 min Scan# 2384

Delta R.T. 0.000 min

Lab File: BP011602.D

Acq: 01 Sep 2022 11:58




Tgt Ion:266 Resp: 294418


Ion Ratio Lower Upper

266 100

264 62.9 50.8 76.2

268 65.4 51.0 76.4





#72

Phenanthrene

Concen: 19.964 ng/ul

RT: 17.410 min Scan# 2452

Delta R.T. 0.000 min

Lab File: BP011602.D

Acq: 01 Sep 2022 11:58

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:178 Resp: 2764280

Ion Ratio Lower Upper

178 100

179 15.5 12.6 18.8

176 19.5 15.5 23.3



#74

Anthracene

Concen: 19.965 ng/ul

RT: 17.504 min Scan# 2468

Delta R.T. 0.000 min

Lab File: BP011602.D

Acq: 01 Sep 2022 11:58




Tgt Ion:178 Resp: 2777709


Ion Ratio Lower Upper

178 100

179 15.6 12.0 18.0

176 18.8 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.485 ng/uL
RT: 13.416 min Scan# 1773
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:216 Resp: 797769
Ion Ratio Lower Upper
216 100
214 78.3 61.7 92.5
218 47.7 38.6 58.0



#76
Pentachlorobenzene
Concen: 20.439 ng/uL
RT: 14.928 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:250 Resp: 753904
Ion Ratio Lower Upper
250 100
248 63.4 50.0 75.0
252 64.9 51.6 77.4









#77
 Carbazole
 Concen: 19.945 ng/ul
 RT: 17.769 min Scan# 2513
 Delta R.T. 0.000 min
 Lab File: BP011602.D
 Acq: 01 Sep 2022 11:58

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020612

Tgt Ion:167 Resp: 2484202




Ion	Ratio	Lower	Upper
167	100		
166	21.8	17.1	25.7
139	12.9	9.9	14.9




#78
 Di-n-butylphthalate
 Concen: 21.044 ng/ul
 RT: 18.327 min Scan# 2608
 Delta R.T. 0.000 min
 Lab File: BP011602.D
 Acq: 01 Sep 2022 11:58

Tgt Ion:149 Resp: 2842797

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.2	10.8
104	4.9	4.1	6.1









#80
Fluoranthene
Concen: 20.751 ng/ul
RT: 19.369 min Scan# 21
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:202 Resp: 3123662
Ion Ratio Lower Upper
202 100
101 11.4 9.4 14.2
100 8.6 7.3 10.9



#82
Pyrene
Concen: 20.649 ng/ul
RT: 19.722 min Scan# 2845
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:202 Resp: 3236279
Ion Ratio Lower Upper
202 100
101 13.0 10.6 16.0
100 10.5 8.6 12.8





#83

Butylbenzylphthalate

Concen: 23.615 ng/ul

RT: 20.604 min Scan# 2995

Delta R.T. 0.000 min

Lab File: BP011602.D

Acq: 01 Sep 2022 11:58

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:149 Resp: 1032342

Ion Ratio Lower Upper

149 100

91 65.9 54.4 81.6

206 20.9 16.7 25.1



#84

3,3'-Dichlorobenzidine

Concen: 17.759 ng/ul

RT: 21.363 min Scan# 3124

Delta R.T. 0.000 min

Lab File: BP011602.D

Acq: 01 Sep 2022 11:58




Tgt Ion:252 Resp: 792675


Ion Ratio Lower Upper

252 100

254 64.1 52.0 78.0

126 10.9 8.7 13.1









#85
Benzo(a)anthracene
Concen: 20.273 ng/ul
RT: 21.433 min Scan# 3136
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:228 Resp: 3069124
Ion Ratio Lower Upper
228 100
229 20.1 15.6 23.4
226 27.9 22.2 33.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 22.693 ng/ul
RT: 21.369 min Scan# 3125
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:149 Resp: 1684151
Ion Ratio Lower Upper
149 100
167 27.3 21.6 32.4
279 4.2 3.2 4.8





#87

Chrysene

Concen: 20.020 ng/ul

RT: 21.486 min Scan# 3145

Delta R.T. 0.000 min

Lab File: BP011602.D

Acq: 01 Sep 2022 11:58

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:228 Resp: 2939973

Ion Ratio Lower Upper

228 100

226 30.7 24.3 36.5

229 20.1 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 21.393 ng/ul




RT: 22.327 min Scan# 3288


Delta R.T. 0.000 min

Lab File: BP011602.D

Acq: 01 Sep 2022 11:58

Tgt Ion:149 Resp: 2677164









#90
Benzo(b)fluoranthene
Concen: 20.157 ng/ul
RT: 23.168 min Scan# 3431
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58




Instrument :
BNA_P
ClientSampleId :
SSTD020612


Tgt Ion:252 Resp: 3030637
Ion Ratio Lower Upper
252 100
253 21.6 17.8 26.8
125 10.1 8.2 12.4



#91
Benzo(k)fluoranthene
Concen: 20.247 ng/ul
RT: 23.215 min Scan# 3439
Delta R.T. 0.000 min
Lab File: BP011602.D
Acq: 01 Sep 2022 11:58

Tgt Ion:252 Resp: 2959316
Ion Ratio Lower Upper
252 100
253 22.0 18.0 27.0
125 10.2 8.2 12.4





#93

Benzo(a)pyrene

Concen: 20.376 ng/ul

RT: 23.815 min Scan# 3541

Delta R.T. 0.000 min

Lab File: BP011602.D

Acq: 01 Sep 2022 11:58

Instrument :

BNA_P

ClientSampleId :

SSTD020612





Tgt Ion:252 Resp: 3004402

Ion Ratio Lower Upper

252 100

253 22.0 17.8 26.6

125 11.1 8.9 13.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.865 ng/ul

RT: 26.527 min Scan# 4002

Delta R.T. 0.000 min

Lab File: BP011602.D

Acq: 01 Sep 2022 11:58




Tgt Ion:276 Resp: 3404977


Ion Ratio Lower Upper

276 100

138 23.5 17.2 25.8

277 24.5 19.4 29.2









#95
 Dibenzo(a,h)anthracene
 Concen: 19.971 ng/ul
 RT: 26.545 min Scan# 4005
 Delta R.T. 0.000 min
 Lab File: BP011602.D
 Acq: 01 Sep 2022 11:58

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020612

Tgt Ion:278 Resp: 2944802

Ion	Ratio	Lower	Upper
278	100		
139	16.0	12.4	18.6
279	23.4	19.3	28.9



#96
 Benzo(g,h,i)perylene
 Concen: 19.636 ng/ul
 RT: 27.315 min Scan# 4136
 Delta R.T. 0.000 min
 Lab File: BP011602.D
 Acq: 01 Sep 2022 11:58

Tgt Ion:276 Resp: 2831447

Ion	Ratio	Lower	Upper
276	100		
138	21.7	16.6	24.8
277	23.9	19.0	28.6

