

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090122\
 Data File : BP011603.D
 Acq On : 01 Sep 2022 12:32
 Operator : CG/JU
 Sample : SST04052
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST040613

Quant Time: Sep 01 13:46:43 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 01 13:42:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	510452	20.000	ng/u1	0.00
20) Naphthalene-d8	10.793	136	2170299	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.616	164	1288263	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.375	188	2601529	20.000	ng/u1	0.00
79) Chrysene-d12	21.451	240	2466721	20.000	ng/u1	0.00
88) Perylene-d12	23.933	264	2419319	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.364	96	191242	15.886	ng/uL	0.00
4) Pyridine-d5	3.781	84	1344781	41.575	ng/u1	0.00
7) Phenol-d5	7.122	99	1762954	41.905	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.305	67	1048549	41.358	ng/u1	0.00
11) 2-Chlorophenol-d4	7.505	132	1406996	44.274	ng/u1	0.00
15) 4-Methylphenol-d8	8.681	113	1355636	42.948	ng/u1	0.00
21) Nitrobenzene-d5	9.140	128	640211	50.082	ng/u1	0.00
24) 2-Nitrophenol-d4	9.869	143	572040	57.339	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.404	165	1354096	45.319	ng/u1	0.00
31) 4-Chloroaniline-d4	10.922	131	1940530	40.838	ng/u1	0.00
46) Dimethylphthalate-d6	14.028	166	3683958	42.288	ng/u1	0.00
49) Acenaphthylene-d8	14.310	160	4502337	43.138	ng/u1	0.00
54) 4-Nitrophenol-d4	14.798	143	642021	43.930	ng/u1	0.00
60) Fluorene-d10	15.610	176	3251874	41.682	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.722	200	383072	45.029	ng/u1	0.00
73) Anthracene-d10	17.475	188	4900843	42.071	ng/u1	0.00
81) Pyrene-d10	19.698	212	5395288	43.563	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.768	264	5194336	44.397	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.399	88	199031	15.666	ng/uL	100
5) Pyridine	3.799	79	1339989	41.938	ng/u1	93
6) Benzaldehyde	7.110	77	883993	45.112	ng/u1	99
8) Phenol	7.152	94	1792598	41.444	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.399	93	1414010	41.609	ng/u1	99
12) 2-Chlorophenol	7.534	128	1472796	43.898	ng/u1	99
13) 2-Methylphenol	8.416	108	1371647	41.776	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.516	45	1825189	40.391	ng/u1	98
16) Acetophenone	8.805	105	2161791	41.824	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.793	70	1044479	43.598	ng/u1	96
18) 4-Methylphenol	8.746	108	1467045	42.251	ng/u1	99
19) Hexachloroethane	9.063	117	575600	42.621	ng/u1	99
22) Nitrobenzene	9.187	77	1471791	45.947	ng/u1	97
23) Isophorone	9.716	82	2921837	42.368	ng/u1	99
25) 2-Nitrophenol	9.899	139	692839	55.629	ng/u1	98
26) 2,4-Dimethylphenol	9.957	107	1572489	42.980	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.199	93	1843667	42.470	ng/u1	99
29) 2,4-Dichlorophenol	10.434	162	1368708	44.559	ng/u1	98
30) Naphthalene	10.840	128	4668508	41.718	ng/u1	99
32) 4-Chloroaniline	10.951	127	1981492	40.713	ng/u1	99
33) Hexachlorobutadiene	11.134	225	882368	42.703	ng/u1	99
34) Caprolactam	11.710	113	216839	23.415	ng/u1	99
35) 4-Chloro-3-methylphenol	12.063	107	1358390	43.839	ng/u1	98
36) 2-Methylnaphthalene	12.451	142	3155307	42.206	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090122\
 Data File : BP011603.D
 Acq On : 01 Sep 2022 12:32
 Operator : CG/JU
 Sample : SST04052
 Misc :
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SST040613


Quant Time: Sep 01 13:46:43 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 01 13:42:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.669	142	3156486	41.908	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.810	216	1686514	43.481	ng/ul	99
40) Hexachlorocyclopentadiene	12.798	237	973946	44.162	ng/ul	98
41) 2,4,6-Trichlorophenol	13.051	196	999928	47.912	ng/ul	98
42) 2,4,5-Trichlorophenol	13.116	196	1097544	47.537	ng/ul	99
43) 1,1'-Biphenyl	13.457	154	4099583	42.096	ng/ul	98
44) 2-Chloronaphthalene	13.498	162	3151755	42.371	ng/ul	100
45) 2-Nitroaniline	13.692	65	733618	52.898	ng/ul	97
47) Dimethylphthalate	14.075	163	3742166	42.023	ng/ul	100
48) 2,6-Dinitrotoluene	14.192	165	686593	53.397	ng/ul	91
50) Acenaphthylene	14.339	152	5027177	41.945	ng/ul	99
51) 3-Nitroaniline	14.516	138	732908	45.984	ng/ul#	95
52) Acenaphthene	14.681	153	3246499	41.624	ng/ul	100
53) 2,4-Dinitrophenol	14.716	184	238404	43.756	ng/ul	96
55) 4-Nitrophenol	14.816	109	486277	41.444	ng/ul	94
56) Dibenzofuran	15.016	168	4603649	41.630	ng/ul	99
57) 2,4-Dinitrotoluene	14.975	165	950085	52.834	ng/ul	89
58) 2,3,4,6-Tetrachlorophenol	15.239	232	830593	48.747	ng/ul	96
59) Diethylphthalate	15.445	149	3692300	42.238	ng/ul	99
61) Fluorene	15.669	166	3744429	41.279	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.663	204	1899691	42.113	ng/ul	99
63) 4-Nitroaniline	15.681	138	718490	46.682	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.734	198	420206	45.247	ng/ul	98
67) N-Nitrosodiphenylamine	15.875	169	3150645	42.356	ng/ul	99
68) 4-Bromophenyl-phenylether	16.563	248	1143088	43.724	ng/ul	98
69) Hexachlorobenzene	16.669	284	1339350	42.691	ng/ul	99
70) Atrazine	16.833	200	1138797	42.311	ng/ul	98
71) Pentachlorophenol	17.016	266	697445	43.674	ng/ul	99
72) Phenanthrene	17.416	178	5669863	41.229	ng/ul	99
74) Anthracene	17.510	178	5754476	41.643	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.416	216	1723203	44.550	ng/uL	100
76) Pentachlorobenzene	14.934	250	1600019	43.674	ng/uL	99
77) Carbazole	17.769	167	5167407	41.772	ng/ul	99
78) Di-n-butylphthalate	18.328	149	6136797	45.738	ng/ul	99
80) Fluoranthene	19.375	202	6559781	43.832	ng/ul	98
82) Pyrene	19.727	202	6639732	42.612	ng/ul	99
83) Butylbenzylphthalate	20.604	149	2437131	56.075	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.369	252	1933554	43.572	ng/ul	99
85) Benzo(a)anthracene	21.439	228	6308639	41.913	ng/ul	98
86) Bis(2-ethylhexyl)phtha...	21.369	149	3880928	52.596	ng/ul	98
87) Chrysene	21.492	228	6086225	41.686	ng/ul	98
89) Di-n-octyl phthalate	22.327	149	6263016	50.619	ng/ul	100
90) Benzo(b)fluoranthene	23.174	252	6568531	44.187	ng/ul	99
91) Benzo(k)fluoranthene	23.227	252	6229046	43.105	ng/ul	99
93) Benzo(a)pyrene	23.821	252	6359933	43.626	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.533	276	7427267	43.827	ng/ul	99
95) Dibenzo(a,h)anthracene	26.557	278	6395401	43.869	ng/ul	100
96) Benzo(g,h,i)perylene	27.327	276	6093816	42.743	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 01 13:46:43 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090122.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Sep 01 13:42:41 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 15.666 ng/uL
 RT: 3.399 min Scan# 70
 Delta R.T. -0.006 min
 Lab File: BP011603.D
 Acq: 01 Sep 2022 12:32




Instrument :
 BNA_P
 ClientSampleId :
 SSTD040613


Tgt Ion: 88 Resp: 199031
 Ion Ratio Lower Upper
 88 100
 43 28.1 22.2 33.2
 58 76.7 61.4 92.0



#5
 Pyridine
 Concen: 41.938 ng/uL
 RT: 3.799 min Scan# 138
 Delta R.T. -0.006 min
 Lab File: BP011603.D
 Acq: 01 Sep 2022 12:32

Tgt Ion: 79 Resp: 1339989
 Ion Ratio Lower Upper
 79 100
 52 77.6 67.2 100.8
 51 34.8 30.6 45.8









#6
Benzaldehyde
Concen: 45.112 ng/ul
RT: 7.110 min Scan# 701
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion: 77 Resp: 883993
Ion Ratio Lower Upper
77 100
105 103.5 82.4 123.6
106 99.5 79.1 118.7



#8
Phenol
Concen: 41.444 ng/ul
RT: 7.152 min Scan# 708
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion: 94 Resp: 1792598
Ion Ratio Lower Upper
94 100
65 27.0 22.4 33.6
66 34.2 27.9 41.9









#10
Bis(2-Chloroethyl)ether
Concen: 41.609 ng/ul
RT: 7.399 min Scan# 749
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion: 93 Resp: 1414010
Ion Ratio Lower Upper
93 100
63 76.8 62.2 93.2
95 32.7 25.9 38.9



#12
2-Chlorophenol
Concen: 43.898 ng/ul
RT: 7.534 min Scan# 773
Delta R.T. 0.000 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:128 Resp: 1472796
Ion Ratio Lower Upper
128 100
64 44.5 36.4 54.6
130 32.7 25.7 38.5









#13
2-Methylphenol
Concen: 41.776 ng/ul
RT: 8.416 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:108 Resp: 1371647
Ion Ratio Lower Upper
108 100
107 89.3 70.3 105.5



#14
2,2'-oxybis(1-Chloropropane)
Concen: 40.391 ng/ul
RT: 8.516 min Scan# 940
Delta R.T. 0.012 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion: 45 Resp: 1825189
Ion Ratio Lower Upper
45 100
77 14.9 11.4 17.0
79 11.7 8.4 12.6









#16
Acetophenone
Concen: 41.824 ng/ul
RT: 8.805 min Scan# 988
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:105 Resp: 2161791
Ion Ratio Lower Upper
105 100
77 73.8 60.0 90.0
51 24.9 21.0 31.6



#17
N-Nitroso-di-n-propylamine
Concen: 43.598 ng/ul
RT: 8.793 min Scan# 987
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion: 70 Resp: 1044479
Ion Ratio Lower Upper
70 100
42 53.3 46.3 69.5
101 9.8 8.1 12.1
130 22.2 17.3 25.9









#18
4-Methylphenol
Concen: 42.251 ng/ul
RT: 8.746 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:108 Resp: 1467045
Ion Ratio Lower Upper
108 100
107 110.7 89.0 133.4



#19
Hexachloroethane
Concen: 42.621 ng/ul
RT: 9.063 min Scan# 1033
Delta R.T. 0.000 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:117 Resp: 575600
Ion Ratio Lower Upper
117 100
201 102.7 82.7 124.1
199 62.7 51.4 77.2









#22
Nitrobenzene
Concen: 45.947 ng/ul
RT: 9.187 min Scan# 1053
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion: 77 Resp: 1471791
Ion Ratio Lower Upper
77 100
123 49.6 37.6 56.4
65 13.5 11.0 16.4



#23
Isophorone
Concen: 42.368 ng/ul
RT: 9.716 min Scan# 1144
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion: 82 Resp: 2921837
Ion Ratio Lower Upper
82 100
95 6.9 5.6 8.4
138 19.5 15.2 22.8









#25
2-Nitrophenol
Concen: 55.629 ng/ul
RT: 9.899 min Scan# 1174
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:139 Resp: 692839
Ion Ratio Lower Upper
139 100
65 46.5 38.5 57.7
109 27.4 22.3 33.5



#26
2,4-Dimethylphenol
Concen: 42.980 ng/ul
RT: 9.957 min Scan# 1185
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:107 Resp: 1572489
Ion Ratio Lower Upper
107 100
121 53.3 42.9 64.3
122 89.1 71.8 107.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 42.470 ng/ul
 RT: 10.199 min Scan# 1226
 Delta R.T. 0.000 min
 Lab File: BP011603.D
 Acq: 01 Sep 2022 12:32

Instrument :
 BNA_P
 ClientSampleId :
 SSTD040613

Tgt Ion: 93 Resp: 1843667




Ion	Ratio	Lower	Upper
93	100		
95	32.5	26.3	39.5
123	11.8	8.7	13.1




#29
 2,4-Dichlorophenol
 Concen: 44.559 ng/ul
 RT: 10.434 min Scan# 1266
 Delta R.T. 0.006 min
 Lab File: BP011603.D
 Acq: 01 Sep 2022 12:32

Tgt Ion: 162 Resp: 1368708

Ion	Ratio	Lower	Upper
162	100		
164	65.7	53.4	80.2
98	32.9	28.3	42.5









#30
Naphthalene
Concen: 41.718 ng/ul
RT: 10.840 min Scan# 1335
Delta R.T. 0.000 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:128 Resp: 4668508
Ion Ratio Lower Upper
128 100
129 11.3 8.8 13.2
127 13.1 10.2 15.4



#32
4-Chloroaniline
Concen: 40.713 ng/ul
RT: 10.951 min Scan# 1354
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:127 Resp: 1981492
Ion Ratio Lower Upper
127 100
129 32.5 26.3 39.5









#33
Hexachlorobutadiene
Concen: 42.703 ng/ul
RT: 11.134 min Scan# 1385
Delta R.T. 0.000 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:225 Resp: 882368
Ion Ratio Lower Upper
225 100
223 63.1 50.5 75.7
227 64.9 51.1 76.7



#34
Caprolactam
Concen: 23.415 ng/ul
RT: 11.710 min Scan# 1483
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:113 Resp: 216839
Ion Ratio Lower Upper
113 100
55 170.2 137.1 205.7
56 129.0 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 43.839 ng/ul
RT: 12.063 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:107 Resp: 1358390
Ion Ratio Lower Upper
107 100
144 27.8 21.4 32.0
142 85.3 66.5 99.7



#36
2-Methylnaphthalene
Concen: 42.206 ng/ul
RT: 12.451 min Scan# 1609
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:142 Resp: 3155307
Ion Ratio Lower Upper
142 100
141 88.8 71.5 107.3









#37
1-Methylnaphthalene
Concen: 41.908 ng/ul
RT: 12.669 min Scan# 1645
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:142 Resp: 3156486
Ion Ratio Lower Upper
142 100
141 91.6 74.2 111.4
116 3.5 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 43.481 ng/ul
RT: 12.810 min Scan# 1670
Delta R.T. 0.000 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:216 Resp: 1686514
Ion Ratio Lower Upper
216 100
214 79.0 62.6 94.0
179 21.0 16.1 24.1
108 17.2 13.2 19.8









#40
Hexachlorocyclopentadiene
Concen: 44.162 ng/ul
RT: 12.798 min Scan# 1668
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:237 Resp: 973946
Ion Ratio Lower Upper
237 100
235 63.0 49.4 74.0
272 13.6 10.5 15.7



#41
2,4,6-Trichlorophenol
Concen: 47.912 ng/ul
RT: 13.051 min Scan# 1711
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:196 Resp: 999928
Ion Ratio Lower Upper
196 100
198 96.6 75.1 112.7
200 31.2 24.9 37.3









#42
2,4,5-Trichlorophenol
Concen: 47.537 ng/ul
RT: 13.116 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:196 Resp: 1097544
Ion Ratio Lower Upper
196 100
198 95.3 77.0 115.6
200 30.1 24.1 36.1



#43
1,1'-Biphenyl
Concen: 42.096 ng/ul
RT: 13.457 min Scan# 1780
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:154 Resp: 4099583
Ion Ratio Lower Upper
154 100
153 40.3 31.4 47.0
76 11.9 9.8 14.8





#44

2-Chloronaphthalene

Concen: 42.371 ng/ul

RT: 13.498 min Scan# 1786

Delta R.T. 0.006 min

Lab File: BP011603.D

Acq: 01 Sep 2022 12:32

Instrument :

BNA_P

ClientSampleId :

SSTD040613





Tgt Ion:162 Resp: 3151755

Ion Ratio Lower Upper

162 100

164 33.2 26.5 39.7

127 37.4 30.3 45.5



#45

2-Nitroaniline

Concen: 52.898 ng/ul

RT: 13.692 min Scan# 1820

Delta R.T. 0.000 min

Lab File: BP011603.D

Acq: 01 Sep 2022 12:32




Tgt Ion: 65 Resp: 733618


Ion Ratio Lower Upper

65 100

92 71.0 55.4 83.2

138 114.4 89.1 133.7









#47
Dimethylphthalate
Concen: 42.023 ng/ul
RT: 14.075 min Scan# 1885
Delta R.T. 0.000 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:163 Resp: 3742166
Ion Ratio Lower Upper
163 100
194 4.0 3.3 4.9
164 10.3 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 53.397 ng/ul
RT: 14.192 min Scan# 1905
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:165 Resp: 686593
Ion Ratio Lower Upper
165 100
89 54.2 49.1 73.7
121 21.1 19.8 29.8









#50
Acenaphthylene
Concen: 41.945 ng/ul
RT: 14.339 min Scan# 1930
Delta R.T. 0.000 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:152 Resp: 5027177
Ion Ratio Lower Upper
152 100
151 20.5 15.8 23.6
153 13.5 10.5 15.7



#51
3-Nitroaniline
Concen: 45.984 ng/ul
RT: 14.516 min Scan# 1960
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:138 Resp: 732908
Ion Ratio Lower Upper
138 100
108 12.9 17.0 25.4#
92 108.2 88.2 132.4









#52
Acenaphthene
Concen: 41.624 ng/ul
RT: 14.681 min Scan# 1988
Delta R.T. 0.000 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:153 Resp: 3246499
Ion Ratio Lower Upper
153 100
152 48.2 39.0 58.6
154 88.7 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 43.756 ng/ul
RT: 14.716 min Scan# 1994
Delta R.T. 0.000 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:184 Resp: 238404
Ion Ratio Lower Upper
184 100
63 63.8 49.8 74.6
154 61.4 52.3 78.5









#55
4-Nitrophenol
Concen: 41.444 ng/ul
RT: 14.816 min Scan# 2011
Delta R.T. 0.012 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:109 Resp: 486277
Ion Ratio Lower Upper
109 100
139 145.0 108.4 162.6
65 127.3 104.7 157.1



#56
Dibenzofuran
Concen: 41.630 ng/ul
RT: 15.016 min Scan# 2045
Delta R.T. 0.000 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:168 Resp: 4603649
Ion Ratio Lower Upper
168 100
139 37.1 29.1 43.7









#57
2,4-Dinitrotoluene
Concen: 52.834 ng/ul
RT: 14.975 min Scan# 2038
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:165 Resp: 950085
Ion Ratio Lower Upper
165 100
63 38.8 37.3 55.9
182 2.7 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 48.747 ng/ul
RT: 15.239 min Scan# 2083
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:232 Resp: 830593
Ion Ratio Lower Upper
232 100
131 43.8 37.7 56.5
130 2.3 2.0 3.0
166 30.7 25.5 38.3









#59
Diethylphthalate
Concen: 42.238 ng/ul
RT: 15.445 min Scan# 2117
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:149 Resp: 3692300
Ion Ratio Lower Upper
149 100
177 23.1 17.7 26.5
150 12.6 9.9 14.9



#61
Fluorene
Concen: 41.279 ng/ul
RT: 15.669 min Scan# 2156
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:166 Resp: 3744429
Ion Ratio Lower Upper
166 100
165 97.2 77.7 116.5
167 13.5 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 42.113 ng/ul
RT: 15.663 min Scan# 2155
Delta R.T. 0.000 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:204 Resp: 1899691
Ion Ratio Lower Upper
204 100
206 33.1 26.5 39.7
141 54.9 45.0 67.6



#63
4-Nitroaniline
Concen: 46.682 ng/ul
RT: 15.681 min Scan# 2158
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:138 Resp: 718490
Ion Ratio Lower Upper
138 100
92 47.2 39.1 58.7
108 73.8 65.5 98.3









#66
4,6-Dinitro-2-methylphenol
Concen: 45.247 ng/ul
RT: 15.734 min Scan# 2166
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:198 Resp: 420206
Ion Ratio Lower Upper
198 100
182 4.4 3.2 4.8
77 25.1 19.4 29.0



#67
N-Nitrosodiphenylamine
Concen: 42.356 ng/ul
RT: 15.875 min Scan# 2191
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:169 Resp: 3150645
Ion Ratio Lower Upper
169 100
168 66.6 53.8 80.6
167 36.9 29.2 43.8









#68
4-Bromophenyl-phenylether
Concen: 43.724 ng/ul
RT: 16.563 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:248 Resp: 1143088
Ion Ratio Lower Upper
248 100
250 98.5 79.0 118.4
141 66.7 57.1 85.7



#69
Hexachlorobenzene
Concen: 42.691 ng/ul
RT: 16.669 min Scan# 2326
Delta R.T. 0.000 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:284 Resp: 1339350
Ion Ratio Lower Upper
284 100
142 35.1 27.5 41.3
249 30.4 23.9 35.9





#70

Atrazine

Concen: 42.311 ng/ul

RT: 16.833 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP011603.D

Acq: 01 Sep 2022 12:32

Instrument :

BNA_P

ClientSampleId :

SSTD040613





Tgt Ion:200 Resp: 1138797

Ion Ratio Lower Upper

200 100

173 26.4 21.4 32.2

215 48.1 40.0 60.0



#71

Pentachlorophenol

Concen: 43.674 ng/ul

RT: 17.016 min Scan# 2385

Delta R.T. 0.006 min

Lab File: BP011603.D

Acq: 01 Sep 2022 12:32




Tgt Ion:266 Resp: 697445


Ion Ratio Lower Upper

266 100

264 63.5 50.8 76.2

268 64.9 51.0 76.4









#72
Phenanthrene
Concen: 41.229 ng/ul
RT: 17.416 min Scan# 2452
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:178 Resp: 5669863
Ion Ratio Lower Upper
178 100
179 15.8 12.6 18.8
176 19.9 15.5 23.3



#74
Anthracene
Concen: 41.643 ng/ul
RT: 17.510 min Scan# 2469
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:178 Resp: 5754476
Ion Ratio Lower Upper
178 100
179 15.7 12.0 18.0
176 19.3 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 44.550 ng/uL
RT: 13.416 min Scan# 1773
Delta R.T. 0.000 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:216 Resp: 1723203
Ion Ratio Lower Upper
216 100
214 77.3 61.7 92.5
218 48.2 38.6 58.0



#76
Pentachlorobenzene
Concen: 43.674 ng/uL
RT: 14.934 min Scan# 2031
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:250 Resp: 1600019
Ion Ratio Lower Upper
250 100
248 63.4 50.0 75.0
252 63.8 51.6 77.4









#77
 Carbazole
 Concen: 41.772 ng/ul
 RT: 17.769 min Scan# 2513
 Delta R.T. 0.000 min
 Lab File: BP011603.D
 Acq: 01 Sep 2022 12:32




Instrument :
 BNA_P
 ClientSampleId :
 SSTD040613


Tgt Ion:167 Resp: 5167407
 Ion Ratio Lower Upper
 167 100
 166 21.9 17.1 25.7
 139 13.0 9.9 14.9



#78
 Di-n-butylphthalate
 Concen: 45.738 ng/ul
 RT: 18.328 min Scan# 2608
 Delta R.T. 0.000 min
 Lab File: BP011603.D
 Acq: 01 Sep 2022 12:32

Tgt Ion:149 Resp: 6136797
 Ion Ratio Lower Upper
 149 100
 150 9.5 7.2 10.8
 104 5.0 4.1 6.1





#80

Fluoranthene

Concen: 43.832 ng/u1

RT: 19.375 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP011603.D

Acq: 01 Sep 2022 12:32

Instrument :

BNA_P

ClientSampleId :

SSTD040613





Tgt Ion:202 Resp: 6559781

Ion Ratio Lower Upper

202 100

101 11.0 9.4 14.2

100 8.5 7.3 10.9



#82

Pyrene

Concen: 42.612 ng/u1

RT: 19.727 min Scan# 2846

Delta R.T. 0.006 min

Lab File: BP011603.D

Acq: 01 Sep 2022 12:32




Tgt Ion:202 Resp: 6639732


Ion Ratio Lower Upper

202 100

101 12.7 10.6 16.0

100 10.3 8.6 12.8









#83
Butylbenzylphthalate
Concen: 56.075 ng/ul
RT: 20.604 min Scan# 2995
Delta R.T. 0.000 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:149 Resp: 2437131
Ion Ratio Lower Upper
149 100
91 66.4 54.4 81.6
206 21.6 16.7 25.1



#84
3,3'-Dichlorobenzidine
Concen: 43.572 ng/ul
RT: 21.369 min Scan# 3125
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:252 Resp: 1933554
Ion Ratio Lower Upper
252 100
254 64.4 52.0 78.0
126 9.7 8.7 13.1









#85
Benzo(a)anthracene
Concen: 41.913 ng/ul
RT: 21.439 min Scan# 3136
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:228 Resp: 6308639
Ion Ratio Lower Upper
228 100
229 20.4 15.6 23.4
226 28.8 22.2 33.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 52.596 ng/ul
RT: 21.369 min Scan# 3125
Delta R.T. 0.000 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:149 Resp: 3880928
Ion Ratio Lower Upper
149 100
167 28.0 21.6 32.4
279 4.3 3.2 4.8





#87

Chrysene

Concen: 41.686 ng/ul

RT: 21.492 min Scan# 3146

Delta R.T. 0.006 min

Lab File: BP011603.D

Acq: 01 Sep 2022 12:32

Instrument :

BNA_P

ClientSampleId :

SSTD040613





Tgt Ion:228 Resp: 6086225

Ion Ratio Lower Upper

228 100

226 31.5 24.3 36.5

229 20.4 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 50.619 ng/ul




RT: 22.327 min Scan# 3288


Delta R.T. 0.000 min

Lab File: BP011603.D

Acq: 01 Sep 2022 12:32

Tgt Ion:149 Resp: 6263016









#90
Benzo(b)fluoranthene
Concen: 44.187 ng/ul
RT: 23.174 min Scan# 3431
Delta R.T. 0.006 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32




Instrument :
BNA_P
ClientSampleId :
SSTD040613


Tgt Ion:252 Resp: 6568531
Ion Ratio Lower Upper
252 100
253 22.6 17.8 26.8
125 9.4 8.2 12.4



#91
Benzo(k)fluoranthene
Concen: 43.105 ng/ul
RT: 23.227 min Scan# 3441
Delta R.T. 0.012 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:252 Resp: 6229046
Ion Ratio Lower Upper
252 100
253 22.2 18.0 27.0
125 9.9 8.2 12.4





#93

Benzo(a)pyrene

Concen: 43.626 ng/ul

RT: 23.821 min Scan# 3

Delta R.T. 0.006 min

Lab File: BP011603.D

Acq: 01 Sep 2022 12:32

Instrument :

BNA_P

ClientSampleId :

SSTD040613





Tgt Ion:252 Resp: 6359933

Ion Ratio Lower Upper

252 100

253 22.4 17.8 26.6

125 10.7 8.9 13.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 43.827 ng/ul

RT: 26.533 min Scan# 4003

Delta R.T. 0.006 min

Lab File: BP011603.D

Acq: 01 Sep 2022 12:32




Tgt Ion:276 Resp: 7427267


Ion Ratio Lower Upper

276 100

138 22.7 17.2 25.8

277 24.4 19.4 29.2









#95
Dibenzo(a,h)anthracene
Concen: 43.869 ng/ul
RT: 26.557 min Scan# 4005
Delta R.T. 0.012 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Instrument :
BNA_P
ClientSampleId :
SSTD040613

Tgt Ion:278 Resp: 6395401
Ion Ratio Lower Upper
278 100
139 15.6 12.4 18.6
279 24.4 19.3 28.9



#96
Benzo(g,h,i)perylene
Concen: 42.743 ng/ul
RT: 27.327 min Scan# 4138
Delta R.T. 0.012 min
Lab File: BP011603.D
Acq: 01 Sep 2022 12:32

Tgt Ion:276 Resp: 6093816
Ion Ratio Lower Upper
276 100
138 20.7 16.6 24.8
277 24.3 19.0 28.6

