

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090122\
 Data File : BP011604.D
 Acq On : 01 Sep 2022 13:07
 Operator : CG/JU
 Sample : SSTD08053
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080614

Quant Time: Sep 01 14:19:54 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 01 13:42:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.975	152	498106	20.000	ng/u1	0.00
20) Naphthalene-d8	10.793	136	2093917	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.616	164	1239590	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.369	188	2477767	20.000	ng/u1	0.00
79) Chrysene-d12	21.451	240	2381623	20.000	ng/u1	0.00
88) Perylene-d12	23.933	264	2341666	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.364	96	371113	31.592	ng/uL	0.00
4) Pyridine-d5	3.781	84	2712904	85.951	ng/u1	0.00
7) Phenol-d5	7.128	99	3542091	86.282	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.305	67	2050349	82.877	ng/u1	0.00
11) 2-Chlorophenol-d4	7.505	132	2825671	91.119	ng/u1	0.00
15) 4-Methylphenol-d8	8.687	113	2691724	87.390	ng/u1	0.01
21) Nitrobenzene-d5	9.146	128	1317934	106.860	ng/u1	0.00
24) 2-Nitrophenol-d4	9.869	143	1277636	132.736	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.410	165	2737824	94.971	ng/u1	0.01
31) 4-Chloroaniline-d4	10.928	131	3777340	82.393	ng/u1	0.00
46) Dimethylphthalate-d6	14.034	166	7135615	85.126	ng/u1	0.00
49) Acenaphthylene-d8	14.316	160	8611931	85.753	ng/u1	0.00
54) 4-Nitrophenol-d4	14.810	143	1349891	95.994	ng/u1	0.02
60) Fluorene-d10	15.616	176	6291564	83.812	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.728	200	923089	113.926	ng/u1	0.01
73) Anthracene-d10	17.475	188	9379611	84.541	ng/u1	0.00
81) Pyrene-d10	19.698	212	10364649	86.677	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.774	264	10554757	93.205	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.399	88	395103	31.869	ng/uL	99
5) Pyridine	3.799	79	2686514	86.164	ng/u1	93
6) Benzaldehyde	7.111	77	1440971	75.358	ng/u1	99
8) Phenol	7.158	94	3570976	84.606	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.399	93	2774669	83.672	ng/u1	99
12) 2-Chlorophenol	7.534	128	2929628	89.485	ng/u1	99
13) 2-Methylphenol	8.416	108	2739123	85.494	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.516	45	3524727	79.935	ng/u1	98
16) Acetophenone	8.805	105	4146552	82.211	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.805	70	2005857	85.803	ng/u1	92
18) 4-Methylphenol	8.752	108	2888515	85.251	ng/u1	100
19) Hexachloroethane	9.063	117	1138404	86.383	ng/u1	100
22) Nitrobenzene	9.187	77	2879426	93.171	ng/u1	97
23) Isophorone	9.722	82	5712945	85.861	ng/u1	98
25) 2-Nitrophenol	9.899	139	1478237	123.020	ng/u1	98
26) 2,4-Dimethylphenol	9.957	107	3057820	86.627	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.204	93	3599071	85.930	ng/u1	99
29) 2,4-Dichlorophenol	10.434	162	2719914	91.778	ng/u1	98
30) Naphthalene	10.846	128	8926062	82.674	ng/u1	99
32) 4-Chloroaniline	10.952	127	3829581	81.555	ng/u1	100
33) Hexachlorobutadiene	11.134	225	1750426	87.804	ng/u1	99
34) Caprolactam	11.734	113	423475	47.397	ng/u1	95
35) 4-Chloro-3-methylphenol	12.069	107	2709011	90.615	ng/u1	97
36) 2-Methylnaphthalene	12.451	142	6097191	84.533	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090122\
 Data File : BP011604.D
 Acq On : 01 Sep 2022 13:07
 Operator : CG/JU
 Sample : SST08053
 Misc :
 ALS Vial : 6 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SST080614


Quant Time: Sep 01 14:19:54 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 01 13:42:41 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.669	142	6053100	83.297	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.816	216	3323036	89.037	ng/ul	99
40) Hexachlorocyclopentadiene	12.798	237	2065794	97.349	ng/ul	98
41) 2,4,6-Trichlorophenol	13.051	196	2064487	102.805	ng/ul	99
42) 2,4,5-Trichlorophenol	13.122	196	2224638	100.137	ng/ul	99
43) 1,1'-Biphenyl	13.457	154	7786276	83.091	ng/ul	98
44) 2-Chloronaphthalene	13.498	162	6030984	84.262	ng/ul	100
45) 2-Nitroaniline	13.698	65	1547299	115.950	ng/ul	93
47) Dimethylphthalate	14.081	163	7177369	83.763	ng/ul	99
48) 2,6-Dinitrotoluene	14.193	165	1456517	117.724	ng/ul	89
50) Acenaphthylene	14.340	152	9501693	82.392	ng/ul	98
51) 3-Nitroaniline	14.522	138	1488385	97.051	ng/ul#	95
52) Acenaphthene	14.687	153	6267468	83.511	ng/ul	99
53) 2,4-Dinitrophenol	14.722	184	590906	112.712	ng/ul	95
55) 4-Nitrophenol	14.822	109	967732	85.716	ng/ul	94
56) Dibenzofuran	15.022	168	8655837	81.346	ng/ul	99
57) 2,4-Dinitrotoluene	14.981	165	2015522	116.484	ng/ul	88
58) 2,3,4,6-Tetrachlorophenol	15.240	232	1762223	107.486	ng/ul	94
59) Diethylphthalate	15.445	149	7200247	85.601	ng/ul	98
61) Fluorene	15.669	166	6937143	79.479	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.663	204	3625055	83.518	ng/ul	98
63) 4-Nitroaniline	15.687	138	1353059	91.363	ng/ul	92
66) 4,6-Dinitro-2-methylph...	15.740	198	991271	112.071	ng/ul	100
67) N-Nitrosodiphenylamine	15.875	169	6099446	86.094	ng/ul	100
68) 4-Bromophenyl-phenylether	16.563	248	2331578	93.638	ng/ul	95
69) Hexachlorobenzene	16.675	284	2698956	90.324	ng/ul	96
70) Atrazine	16.834	200	2256097	88.010	ng/ul	99
71) Pentachlorophenol	17.016	266	1555707	102.285	ng/ul	99
72) Phenanthrene	17.416	178	10814102	82.564	ng/ul	98
74) Anthracene	17.510	178	10700978	81.308	ng/ul	96
75) 1,2,3,4-Tetrachloroben...	13.422	216	3362843	91.282	ng/ul	98
76) Pentachlorobenzene	14.934	250	3152412	90.347	ng/ul	99
77) Carbazole	17.775	167	9858688	83.676	ng/ul	97
78) Di-n-butylphthalate	18.328	149	11363347	88.923	ng/ul#	96
80) Fluoranthene	19.375	202	11956657	82.748	ng/ul	97
82) Pyrene	19.728	202	11890184	79.034	ng/ul	94
83) Butylbenzylphthalate	20.604	149	5187036	123.611	ng/ul	96
84) 3,3'-Dichlorobenzidine	21.363	252	3865955	90.230	ng/ul	99
85) Benzo(a)anthracene	21.439	228	12028425	82.769	ng/ul	96
86) Bis(2-ethylhexyl)phtha...	21.369	149	7573473	106.307	ng/ul#	94
87) Chrysene	21.492	228	11402956	80.892	ng/ul	94
89) Di-n-octyl phthalate	22.327	149	12623434	105.408	ng/ul	100
90) Benzo(b)fluoranthene	23.174	252	12789425	88.888	ng/ul	98
91) Benzo(k)fluoranthene	23.227	252	12371326	88.448	ng/ul	99
93) Benzo(a)pyrene	23.827	252	12569421	89.079	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.551	276	15284576	93.182	ng/ul	96
95) Dibenzo(a,h)anthracene	26.568	278	13025170	92.308	ng/ul	99
96) Benzo(g,h,i)perylene	27.345	276	12470151	90.368	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 01 14:19:54 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090122.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Sep 01 13:42:41 2022
Response via : Initial Calibration









#2
 1,4-Dioxane
 Concen: 31.869 ng/uL
 RT: 3.399 min Scan# 70
 Delta R.T. -0.006 min
 Lab File: BP011604.D
 Acq: 01 Sep 2022 13:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTD080614


Tgt Ion: 88 Resp: 395103
 Ion Ratio Lower Upper
 88 100
 43 27.5 22.2 33.2
 58 77.9 61.4 92.0



#5
 Pyridine
 Concen: 86.164 ng/uL
 RT: 3.799 min Scan# 138
 Delta R.T. -0.006 min
 Lab File: BP011604.D
 Acq: 01 Sep 2022 13:07

Tgt Ion: 79 Resp: 2686514
 Ion Ratio Lower Upper
 79 100
 52 77.1 67.2 100.8
 51 34.0 30.6 45.8









#6
Benzaldehyde
Concen: 75.358 ng/ul
RT: 7.111 min Scan# 701
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion: 77 Resp: 1440971
Ion Ratio Lower Upper
77 100
105 103.4 82.4 123.6
106 100.2 79.1 118.7



#8
Phenol
Concen: 84.606 ng/ul
RT: 7.158 min Scan# 709
Delta R.T. 0.012 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion: 94 Resp: 3570976
Ion Ratio Lower Upper
94 100
65 27.1 22.4 33.6
66 33.8 27.9 41.9









#10
Bis(2-Chloroethyl)ether
Concen: 83.672 ng/ul
RT: 7.399 min Scan# 749
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Instrument :
BNA_P
ClientSampleId :
SSTD080614

Tgt Ion: 93 Resp: 2774669




Ion	Ratio	Lower	Upper
93	100		
63	76.6	62.2	93.2
95	32.6	25.9	38.9




#12
2-Chlorophenol
Concen: 89.485 ng/ul
RT: 7.534 min Scan# 773
Delta R.T. 0.000 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:128 Resp: 2929628

Ion	Ratio	Lower	Upper
128	100		
64	44.9	36.4	54.6
130	33.1	25.7	38.5









#13
2-Methylphenol
Concen: 85.494 ng/ul
RT: 8.416 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:108 Resp: 2739123
Ion Ratio Lower Upper
108 100
107 88.8 70.3 105.5



#14
2,2'-oxybis(1-Chloropropane)
Concen: 79.935 ng/ul
RT: 8.516 min Scan# 940
Delta R.T. 0.012 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion: 45 Resp: 3524727
Ion Ratio Lower Upper
45 100
77 14.9 11.4 17.0
79 11.7 8.4 12.6









#16
Acetophenone
Concen: 82.211 ng/ul
RT: 8.805 min Scan# 989
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:105 Resp: 4146552
Ion Ratio Lower Upper
105 100
77 73.5 60.0 90.0
51 24.0 21.0 31.6



#17
N-Nitroso-di-n-propylamine
Concen: 85.803 ng/ul
RT: 8.805 min Scan# 989
Delta R.T. 0.018 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion: 70 Resp: 2005857
Ion Ratio Lower Upper
70 100
42 49.8 46.3 69.5
101 10.3 8.1 12.1
130 23.8 17.3 25.9









#18
4-Methylphenol
Concen: 85.251 ng/ul
RT: 8.752 min Scan# 91
Delta R.T. 0.012 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:108 Resp: 2888515
Ion Ratio Lower Upper
108 100
107 111.2 89.0 133.4



#19
Hexachloroethane
Concen: 86.383 ng/ul
RT: 9.063 min Scan# 1033
Delta R.T. 0.000 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:117 Resp: 1138404
Ion Ratio Lower Upper
117 100
201 103.6 82.7 124.1
199 64.0 51.4 77.2









#22
Nitrobenzene
Concen: 93.171 ng/ul
RT: 9.187 min Scan# 1053
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion: 77 Resp: 2879426
Ion Ratio Lower Upper
77 100
123 49.7 37.6 56.4
65 13.5 11.0 16.4



#23
Isophorone
Concen: 85.861 ng/ul
RT: 9.722 min Scan# 1145
Delta R.T. 0.012 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion: 82 Resp: 5712945
Ion Ratio Lower Upper
82 100
95 7.1 5.6 8.4
138 19.9 15.2 22.8









#25
2-Nitrophenol
Concen: 123.020 ng/ul
RT: 9.899 min Scan# 1174
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:139 Resp: 1478237
Ion Ratio Lower Upper
139 100
65 46.6 38.5 57.7
109 27.0 22.3 33.5



#26
2,4-Dimethylphenol
Concen: 86.627 ng/ul
RT: 9.957 min Scan# 1185
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:107 Resp: 3057820
Ion Ratio Lower Upper
107 100
121 52.2 42.9 64.3
122 89.9 71.8 107.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 85.930 ng/ul
 RT: 10.204 min Scan# 1227
 Delta R.T. 0.006 min
 Lab File: BP011604.D
 Acq: 01 Sep 2022 13:07




Instrument :
 BNA_P
 ClientSampleId :
 SSTD080614


Tgt Ion: 93 Resp: 3599071
 Ion Ratio Lower Upper
 93 100
 95 32.5 26.3 39.5
 123 11.9 8.7 13.1



#29
 2,4-Dichlorophenol
 Concen: 91.778 ng/ul
 RT: 10.434 min Scan# 1266
 Delta R.T. 0.006 min
 Lab File: BP011604.D
 Acq: 01 Sep 2022 13:07

Tgt Ion:162 Resp: 2719914
 Ion Ratio Lower Upper
 162 100
 164 65.7 53.4 80.2
 98 33.7 28.3 42.5









#30
Naphthalene
Concen: 82.674 ng/ul
RT: 10.846 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:128 Resp: 8926062
Ion Ratio Lower Upper
128 100
129 11.3 8.8 13.2
127 13.5 10.2 15.4



#32
4-Chloroaniline
Concen: 81.555 ng/ul
RT: 10.952 min Scan# 1354
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:127 Resp: 3829581
Ion Ratio Lower Upper
127 100
129 32.9 26.3 39.5









#33
Hexachlorobutadiene
Concen: 87.804 ng/ul
RT: 11.134 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:225 Resp: 1750426
Ion Ratio Lower Upper
225 100
223 63.6 50.5 75.7
227 64.9 51.1 76.7



#34
Caprolactam
Concen: 47.397 ng/ul
RT: 11.734 min Scan# 1487
Delta R.T. 0.030 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:113 Resp: 423475
Ion Ratio Lower Upper
113 100
55 163.9 137.1 205.7
56 122.3 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 90.615 ng/ul
RT: 12.069 min Scan# 11
Delta R.T. 0.012 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:107 Resp: 2709011
Ion Ratio Lower Upper
107 100
144 28.2 21.4 32.0
142 86.2 66.5 99.7



#36
2-Methylnaphthalene
Concen: 84.533 ng/ul
RT: 12.451 min Scan# 1609
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:142 Resp: 6097191
Ion Ratio Lower Upper
142 100
141 89.0 71.5 107.3









#37
1-Methylnaphthalene
Concen: 83.297 ng/ul
RT: 12.669 min Scan# 1645
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:142 Resp: 6053100
Ion Ratio Lower Upper
142 100
141 91.6 74.2 111.4
116 3.6 2.9 4.3



#39
1,2,4,5-Tetrachlorobenzene
Concen: 89.037 ng/ul
RT: 12.816 min Scan# 1671
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:216 Resp: 3323036
Ion Ratio Lower Upper
216 100
214 77.9 62.6 94.0
179 20.1 16.1 24.1
108 15.2 13.2 19.8





#40

Hexachlorocyclopentadiene

Concen: 97.349 ng/ul

RT: 12.798 min Scan# 1668

Delta R.T. 0.006 min

Lab File: BP011604.D

Acq: 01 Sep 2022 13:07

Instrument :

BNA_P

ClientSampleId :

SSTD080614





Tgt Ion:237 Resp: 2065794

Ion Ratio Lower Upper

237 100

235 63.2 49.4 74.0

272 13.7 10.5 15.7



#41

2,4,6-Trichlorophenol

Concen: 102.805 ng/ul

RT: 13.051 min Scan# 1711

Delta R.T. 0.006 min

Lab File: BP011604.D

Acq: 01 Sep 2022 13:07




Tgt Ion:196 Resp: 2064487


Ion Ratio Lower Upper

196 100

198 95.7 75.1 112.7

200 31.1 24.9 37.3









#42
2,4,5-Trichlorophenol
Concen: 100.137 ng/ul
RT: 13.122 min Scan# 1721
Delta R.T. 0.012 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:196 Resp: 2224638
Ion Ratio Lower Upper
196 100
198 96.1 77.0 115.6
200 31.5 24.1 36.1



#43
1,1'-Biphenyl
Concen: 83.091 ng/ul
RT: 13.457 min Scan# 1780
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:154 Resp: 7786276
Ion Ratio Lower Upper
154 100
153 41.0 31.4 47.0
76 12.4 9.8 14.8





#44

2-Chloronaphthalene

Concen: 84.262 ng/ul

RT: 13.498 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP011604.D

Acq: 01 Sep 2022 13:07

Instrument :

BNA_P

ClientSampleId :

SSTD080614





Tgt Ion:162 Resp: 6030984

Ion Ratio Lower Upper

162 100

164 33.5 26.5 39.7

127 38.0 30.3 45.5



#45

2-Nitroaniline

Concen: 115.950 ng/ul

RT: 13.698 min Scan# 1821

Delta R.T. 0.006 min

Lab File: BP011604.D

Acq: 01 Sep 2022 13:07




Tgt Ion: 65 Resp: 1547299


Ion Ratio Lower Upper

65 100

92 72.3 55.4 83.2

138 120.5 89.1 133.7









#47
Dimethylphthalate
Concen: 83.763 ng/ul
RT: 14.081 min Scan# 1885
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:163 Resp: 7177369
Ion Ratio Lower Upper
163 100
194 4.3 3.3 4.9
164 10.5 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 117.724 ng/ul
RT: 14.193 min Scan# 1905
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:165 Resp: 1456517
Ion Ratio Lower Upper
165 100
89 52.7 49.1 73.7
121 20.4 19.8 29.8









#50
Acenaphthylene
Concen: 82.392 ng/ul
RT: 14.340 min Scan# 1930
Delta R.T. 0.000 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:152 Resp: 9501693
Ion Ratio Lower Upper
152 100
151 20.8 15.8 23.6
153 14.0 10.5 15.7



#51
3-Nitroaniline
Concen: 97.051 ng/ul
RT: 14.522 min Scan# 1961
Delta R.T. 0.012 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:138 Resp: 1488385
Ion Ratio Lower Upper
138 100
108 11.7 17.0 25.4#
92 107.5 88.2 132.4









#52
Acenaphthene
Concen: 83.511 ng/ul
RT: 14.687 min Scan# 1988
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:153 Resp: 6267468
Ion Ratio Lower Upper
153 100
152 48.9 39.0 58.6
154 89.6 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 112.712 ng/ul
RT: 14.722 min Scan# 1995
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:184 Resp: 590906
Ion Ratio Lower Upper
184 100
63 60.0 49.8 74.6
154 60.0 52.3 78.5









#55
4-Nitrophenol
Concen: 85.716 ng/ul
RT: 14.822 min Scan# 2009
Delta R.T. 0.018 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:109 Resp: 967732
Ion Ratio Lower Upper
109 100
139 149.5 108.4 162.6
65 131.2 104.7 157.1



#56
Dibenzofuran
Concen: 81.346 ng/ul
RT: 15.022 min Scan# 2046
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:168 Resp: 8655837
Ion Ratio Lower Upper
168 100
139 37.0 29.1 43.7









#57
2,4-Dinitrotoluene
Concen: 116.484 ng/ul
RT: 14.981 min Scan# 2039
Delta R.T. 0.012 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:165 Resp: 2015522
Ion Ratio Lower Upper
165 100
63 38.3 37.3 55.9
182 2.8 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 107.486 ng/ul
RT: 15.240 min Scan# 2083
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:232 Resp: 1762223
Ion Ratio Lower Upper
232 100
131 42.4 37.7 56.5
130 2.2 2.0 3.0
166 29.8 25.5 38.3









#59
Diethylphthalate
Concen: 85.601 ng/ul
RT: 15.445 min Scan# 2117
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:149 Resp: 7200247
Ion Ratio Lower Upper
149 100
177 23.0 17.7 26.5
150 13.0 9.9 14.9



#61
Fluorene
Concen: 79.479 ng/ul
RT: 15.669 min Scan# 2156
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:166 Resp: 6937143
Ion Ratio Lower Upper
166 100
165 97.7 77.7 116.5
167 14.0 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 83.518 ng/ul
RT: 15.663 min Scan# 2155
Delta R.T. 0.000 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:204 Resp: 3625055
Ion Ratio Lower Upper
204 100
206 33.3 26.5 39.7
141 54.1 45.0 67.6



#63
4-Nitroaniline
Concen: 91.363 ng/ul
RT: 15.687 min Scan# 2159
Delta R.T. 0.012 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:138 Resp: 1353059
Ion Ratio Lower Upper
138 100
92 47.4 39.1 58.7
108 72.0 65.5 98.3









#66
4,6-Dinitro-2-methylphenol
Concen: 112.071 ng/ul
RT: 15.740 min Scan# 2166
Delta R.T. 0.012 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:198 Resp: 991271
Ion Ratio Lower Upper
198 100
182 3.9 3.2 4.8
77 24.3 19.4 29.0



#67
N-Nitrosodiphenylamine
Concen: 86.094 ng/ul
RT: 15.875 min Scan# 2191
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:169 Resp: 6099446
Ion Ratio Lower Upper
169 100
168 67.1 53.8 80.6
167 37.1 29.2 43.8









#68
4-Bromophenyl-phenylether
Concen: 93.638 ng/ul
RT: 16.563 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:248 Resp: 2331578
Ion Ratio Lower Upper
248 100
250 97.5 79.0 118.4
141 63.0 57.1 85.7



#69
Hexachlorobenzene
Concen: 90.324 ng/ul
RT: 16.675 min Scan# 2327
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:284 Resp: 2698956
Ion Ratio Lower Upper
284 100
142 30.1 27.5 41.3
249 30.1 23.9 35.9





#70

Atrazine

Concen: 88.010 ng/ul

RT: 16.834 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP011604.D

Acq: 01 Sep 2022 13:07

Instrument :

BNA_P

ClientSampleId :

SSTD080614





Tgt Ion:200 Resp: 2256097

Ion Ratio Lower Upper

200 100

173 27.0 21.4 32.2

215 49.2 40.0 60.0



#71

Pentachlorophenol

Concen: 102.285 ng/ul

RT: 17.016 min Scan# 2385

Delta R.T. 0.006 min

Lab File: BP011604.D

Acq: 01 Sep 2022 13:07




Tgt Ion:266 Resp: 1555707


Ion Ratio Lower Upper

266 100

264 64.1 50.8 76.2

268 64.3 51.0 76.4





#72

Phenanthrene

Concen: 82.564 ng/ul

RT: 17.416 min Scan# 2452

Delta R.T. 0.006 min

Lab File: BP011604.D

Acq: 01 Sep 2022 13:07

Instrument :

BNA_P

ClientSampleId :

SSTD080614





Tgt Ion:178 Resp:10814102

Ion Ratio Lower Upper

178 100

179 16.5 12.6 18.8

176 20.7 15.5 23.3



#74

Anthracene

Concen: 81.308 ng/ul

RT: 17.510 min Scan# 2469

Delta R.T. 0.006 min

Lab File: BP011604.D

Acq: 01 Sep 2022 13:07




Tgt Ion:178 Resp:10700978


Ion Ratio Lower Upper

178 100

179 16.7 12.0 18.0

176 20.3 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 91.282 ng/uL
RT: 13.422 min Scan# 1774
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:216 Resp: 3362843
Ion Ratio Lower Upper
216 100
214 78.9 61.7 92.5
218 49.8 38.6 58.0



#76
Pentachlorobenzene
Concen: 90.347 ng/uL
RT: 14.934 min Scan# 2031
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:250 Resp: 3152412
Ion Ratio Lower Upper
250 100
248 64.1 50.0 75.0
252 63.9 51.6 77.4





#77

Carbazole

Concen: 83.676 ng/ul

RT: 17.775 min Scan# 2513

Delta R.T. 0.006 min

Lab File: BP011604.D

Acq: 01 Sep 2022 13:07

Instrument :

BNA_P

ClientSampleId :

SSTD080614





Tgt Ion:167 Resp: 9858688

Ion Ratio Lower Upper

167 100

166 22.9 17.1 25.7

139 13.4 9.9 14.9



#78

Di-n-butylphthalate

Concen: 88.923 ng/ul

RT: 18.328 min Scan# 2608

Delta R.T. 0.000 min

Lab File: BP011604.D

Acq: 01 Sep 2022 13:07




Tgt Ion:149 Resp:11363347


Ion Ratio Lower Upper

149 100

150 10.9 7.2 10.8#

104 5.5 4.1 6.1









#80
Fluoranthene
Concen: 82.748 ng/ul
RT: 19.375 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:202 Resp:11956657
Ion Ratio Lower Upper
202 100
101 13.2 9.4 14.2
100 9.8 7.3 10.9



#82
Pyrene
Concen: 79.034 ng/ul
RT: 19.728 min Scan# 2846
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:202 Resp:11890184
Ion Ratio Lower Upper
202 100
101 15.8 10.6 16.0
100 12.8 8.6 12.8









#83
Butylbenzylphthalate
Concen: 123.611 ng/ul
RT: 20.604 min Scan# 2995
Delta R.T. 0.000 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:149 Resp: 5187036
Ion Ratio Lower Upper
149 100
91 65.1 54.4 81.6
206 23.3 16.7 25.1



#84
3,3'-Dichlorobenzidine
Concen: 90.230 ng/ul
RT: 21.363 min Scan# 3124
Delta R.T. 0.000 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:252 Resp: 3865955
Ion Ratio Lower Upper
252 100
254 65.3 52.0 78.0
126 10.4 8.7 13.1





#85

Benzo(a)anthracene

Concen: 82.769 ng/ul

RT: 21.439 min Scan# 3136

Delta R.T. 0.006 min

Lab File: BP011604.D

Acq: 01 Sep 2022 13:07

Instrument :

BNA_P

ClientSampleId :

SSTD080614





Tgt Ion:228 Resp:12028425

Ion Ratio Lower Upper

228 100

229 21.4 15.6 23.4

226 29.8 22.2 33.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 106.307 ng/ul

RT: 21.369 min Scan# 3125

Delta R.T. 0.000 min

Lab File: BP011604.D

Acq: 01 Sep 2022 13:07




Tgt Ion:149 Resp: 7573473


Ion Ratio Lower Upper

149 100

167 30.0 21.6 32.4

279 5.3 3.2 4.8#





#87

Chrysene

Concen: 80.892 ng/ul

RT: 21.492 min Scan# 3145

Delta R.T. 0.006 min

Lab File: BP011604.D

Acq: 01 Sep 2022 13:07

Instrument :

BNA_P

ClientSampleId :

SSTD080614





Tgt Ion:228 Resp:11402956

Ion Ratio Lower Upper

228 100

226 34.2 24.3 36.5

229 22.2 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 105.408 ng/ul




RT: 22.327 min Scan# 3288


Delta R.T. 0.000 min

Lab File: BP011604.D

Acq: 01 Sep 2022 13:07

Tgt Ion:149 Resp:12623434









#90
Benzo(b)fluoranthene
Concen: 88.888 ng/ul
RT: 23.174 min Scan# 3431
Delta R.T. 0.006 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07




Instrument :
BNA_P
ClientSampleId :
SSTD080614


Tgt Ion:252 Resp:12789425
Ion Ratio Lower Upper
252 100
253 23.3 17.8 26.8
125 10.1 8.2 12.4



#91
Benzo(k)fluoranthene
Concen: 88.448 ng/ul
RT: 23.227 min Scan# 3441
Delta R.T. 0.012 min
Lab File: BP011604.D
Acq: 01 Sep 2022 13:07

Tgt Ion:252 Resp:12371326
Ion Ratio Lower Upper
252 100
253 23.4 18.0 27.0
125 10.0 8.2 12.4





#93

Benzo(a)pyrene

Concen: 89.079 ng/ul

RT: 23.827 min Scan# 3

Delta R.T. 0.012 min

Lab File: BP011604.D

Acq: 01 Sep 2022 13:07

Instrument :

BNA_P

ClientSampleId :

SSTD080614





Tgt Ion:252 Resp:12569421

Ion Ratio Lower Upper

252 100

253 23.3 17.8 26.6

125 11.1 8.9 13.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 93.182 ng/ul

RT: 26.551 min Scan# 4006

Delta R.T. 0.024 min

Lab File: BP011604.D

Acq: 01 Sep 2022 13:07




Tgt Ion:276 Resp:15284576


Ion Ratio Lower Upper

276 100

138 23.6 17.2 25.8

277 25.8 19.4 29.2









#95
 Dibenzo(a,h)anthracene
 Concen: 92.308 ng/ul
 RT: 26.568 min Scan# 4009
 Delta R.T. 0.024 min
 Lab File: BP011604.D
 Acq: 01 Sep 2022 13:07

Instrument :
 BNA_P
 ClientSampleId :
 SSTD080614

Tgt Ion:278 Resp:13025170

Ion	Ratio	Lower	Upper
278	100		
139	15.4	12.4	18.6
279	24.9	19.3	28.9



#96
 Benzo(g,h,i)perylene
 Concen: 90.368 ng/ul
 RT: 27.345 min Scan# 4141
 Delta R.T. 0.030 min
 Lab File: BP011604.D
 Acq: 01 Sep 2022 13:07

Tgt Ion:276 Resp:12470151

Ion	Ratio	Lower	Upper
276	100		
138	20.7	16.6	24.8
277	24.4	19.0	28.6

