

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090122\
 Data File : BP011606.D
 Acq On : 01 Sep 2022 15:43
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SICV616

Quant Time: Sep 01 16:33:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 01 15:19:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.981	152	512227	20.000	ng/u1	0.00
20) Naphthalene-d8	10.799	136	2125232	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.622	164	1323699	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.381	188	2702427	20.000	ng/u1	0.01
79) Chrysene-d12	21.457	240	2626411	20.000	ng/u1	0.00
88) Perylene-d12	23.945	264	2603331	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.370	96	88662	7.639	ng/uL	0.00
4) Pyridine-d5	3.787	84	603849	18.712	ng/u1	0.00
7) Phenol-d5	7.128	99	718787	16.955	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.305	67	499405	19.721	ng/u1	0.00
11) 2-Chlorophenol-d4	7.505	132	623145	19.249	ng/u1	0.00
15) 4-Methylphenol-d8	8.681	113	589702	18.143	ng/u1	0.00
21) Nitrobenzene-d5	9.146	128	281010	20.289	ng/u1	0.00
24) 2-Nitrophenol-d4	9.869	143	242811	20.667	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.410	165	584587	19.437	ng/u1	0.01
31) 4-Chloroaniline-d4	10.928	131	867292	18.888	ng/u1	0.00
46) Dimethylphthalate-d6	14.034	166	1740934	19.539	ng/u1	0.00
49) Acenaphthylene-d8	14.316	160	2135964	20.072	ng/u1	0.00
54) 4-Nitrophenol-d4	14.798	143	292218	18.632	ng/u1	0.00
60) Fluorene-d10	15.616	176	1495692	18.911	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.728	200	167433	16.740	ng/u1	0.01
73) Anthracene-d10	17.481	188	2252156	18.814	ng/u1	0.01
81) Pyrene-d10	19.704	212	2549153	18.994	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.780	264	2364014	18.693	ng/u1	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	86564	7.070	ng/uL	99
5) Pyridine	3.805	79	549786	17.249	ng/u1	95
6) Benzaldehyde	7.110	77	524419	27.122	ng/u1	99
8) Phenol	7.152	94	768867	17.764	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.399	93	610487	17.872	ng/u1	98
12) 2-Chlorophenol	7.540	128	640041	18.669	ng/u1	96
13) 2-Methylphenol	8.416	108	579719	17.594	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.516	45	790977	18.001	ng/u1	98
16) Acetophenone	8.805	105	976487	18.744	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.793	70	452592	18.656	ng/u1	97
18) 4-Methylphenol	8.746	108	632237	17.918	ng/u1	100
19) Hexachloroethane	9.069	117	247847	18.288	ng/u1	99
22) Nitrobenzene	9.187	77	638255	19.517	ng/u1	99
23) Isophorone	9.716	82	1255513	18.949	ng/u1	100
25) 2-Nitrophenol	9.904	139	291439	20.560	ng/u1	97
26) 2,4-Dimethylphenol	9.957	107	660067	18.472	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.204	93	799247	18.614	ng/u1	99
29) 2,4-Dichlorophenol	10.434	162	593963	19.310	ng/u1	98
30) Naphthalene	10.846	128	2018595	18.459	ng/u1	99
32) 4-Chloroaniline	10.951	127	869913	18.543	ng/u1	100
33) Hexachlorobutadiene	11.140	225	388115	18.946	ng/u1	98
34) Caprolactam	11.710	113	196346	21.719	ng/u1	94
35) 4-Chloro-3-methylphenol	12.069	107	592979	19.391	ng/u1	96
36) 2-Methylnaphthalene	12.451	142	1404606	19.170	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090122\
 Data File : BP011606.D
 Acq On : 01 Sep 2022 15:43
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SICV616


Quant Time: Sep 01 16:33:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090122.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 01 15:19:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.675	142	1351169	18.398	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.816	216	756800	18.782	ng/ul	100
40) Hexachlorocyclopentadiene	12.804	237	597968	24.959	ng/ul	99
41) 2,4,6-Trichlorophenol	13.051	196	442628	19.424	ng/ul	97
42) 2,4,5-Trichlorophenol	13.122	196	490853	19.612	ng/ul	99
43) 1,1'-Biphenyl	13.457	154	1849965	18.541	ng/ul	99
44) 2-Chloronaphthalene	13.498	162	1414242	18.510	ng/ul	99
45) 2-Nitroaniline	13.698	65	324436	20.499	ng/ul	97
47) Dimethylphthalate	14.081	163	1709042	18.740	ng/ul	99
48) 2,6-Dinitrotoluene	14.192	165	338530	23.164	ng/ul	93
50) Acenaphthylene	14.345	152	2188705	18.017	ng/ul	100
51) 3-Nitroaniline	14.522	138	361156	20.658	ng/ul#	95
52) Acenaphthene	14.687	153	1488102	18.732	ng/ul	99
53) 2,4-Dinitrophenol	14.722	184	143914	22.492	ng/ul	96
55) 4-Nitrophenol	14.816	109	244684	20.517	ng/ul	95
56) Dibenzofuran	15.022	168	2094163	18.584	ng/ul	100
57) 2,4-Dinitrotoluene	14.981	165	426776	20.971	ng/ul	90
58) 2,3,4,6-Tetrachlorophenol	15.245	232	398316	21.134	ng/ul	95
59) Diethylphthalate	15.445	149	1672973	18.838	ng/ul	100
61) Fluorene	15.675	166	1726108	18.951	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.669	204	860020	18.630	ng/ul	99
63) 4-Nitroaniline	15.681	138	369003	22.364	ng/ul	92
66) 4,6-Dinitro-2-methylph...	15.739	198	182504	16.846	ng/ul	100
67) N-Nitrosodiphenylamine	15.881	169	1453089	18.808	ng/ul	99
68) 4-Bromophenyl-phenylether	16.569	248	523718	18.789	ng/ul	96
69) Hexachlorobenzene	16.675	284	616678	18.647	ng/ul	100
70) Atrazine	16.834	200	499126	17.803	ng/ul	99
71) Pentachlorophenol	17.022	266	386821	21.888	ng/ul	98
72) Phenanthrene	17.422	178	2620721	18.563	ng/ul	99
74) Anthracene	17.516	178	2655594	18.787	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.422	216	732539	17.645	ng/ul	99
76) Pentachlorobenzene	14.939	250	720378	18.375	ng/ul	98
77) Carbazole	17.781	167	2437271	19.566	ng/ul	100
78) Di-n-butylphthalate	18.333	149	2765418	19.559	ng/ul	99
80) Fluoranthene	19.380	202	2990464	18.552	ng/ul	97
82) Pyrene	19.733	202	3119498	18.760	ng/ul	97
83) Butylbenzylphthalate	20.610	149	1057788	19.619	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.374	252	1150037	25.171	ng/ul	99
85) Benzo(a)anthracene	21.445	228	3035665	19.096	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.374	149	1779953	20.695	ng/ul	99
87) Chrysene	21.498	228	2829029	18.429	ng/ul	99
89) Di-n-octyl phthalate	22.333	149	2650393	18.529	ng/ul	100
90) Benzo(b)fluoranthene	23.180	252	2965649	18.232	ng/ul	99
91) Benzo(k)fluoranthene	23.233	252	2989857	19.202	ng/ul	98
93) Benzo(a)pyrene	23.833	252	2623702	16.640	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.545	276	3279550	18.022	ng/ul	99
95) Dibenzo(a,h)anthracene	26.568	278	2890228	18.421	ng/ul	99
96) Benzo(g,h,i)perylene	27.339	276	2796096	18.419	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 01 16:33:03 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP090122.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Sep 01 15:19:54 2022
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.070 ng/uL
RT: 3.405 min Scan# 71
Delta R.T. -0.000 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion: 88	Resp: 86564
Ion Ratio	Lower Upper
88	100
43	27.9 22.2 33.2
58	76.0 61.4 92.0



#5
Pyridine
Concen: 17.249 ng/uL
RT: 3.805 min Scan# 139
Delta R.T. 0.000 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion: 79	Resp: 549786
Ion Ratio	Lower Upper
79	100
52	78.8 67.2 100.8
51	36.2 30.6 45.8









#6
Benzaldehyde
Concen: 27.122 ng/ul
RT: 7.110 min Scan# 701
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion: 77 Resp: 524419
Ion Ratio Lower Upper
77 100
105 102.0 82.4 123.6
106 97.6 79.1 118.7



#8
Phenol
Concen: 17.764 ng/ul
RT: 7.152 min Scan# 708
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion: 94 Resp: 768867
Ion Ratio Lower Upper
94 100
65 27.3 22.4 33.6
66 35.6 27.9 41.9









#10
Bis(2-Chloroethyl)ether
Concen: 17.872 ng/ul
RT: 7.399 min Scan# 749
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion: 93 Resp: 610487
Ion Ratio Lower Upper
93 100
63 75.9 62.2 93.2
95 31.4 25.9 38.9



#12
2-Chlorophenol
Concen: 18.669 ng/ul
RT: 7.540 min Scan# 774
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion:128 Resp: 640041
Ion Ratio Lower Upper
128 100
64 41.7 36.4 54.6
130 33.5 25.7 38.5









#13
2-Methylphenol
Concen: 17.594 ng/ul
RT: 8.416 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:108 Resp: 579719
Ion Ratio Lower Upper
108 100
107 90.5 70.3 105.5



#14
2,2'-oxybis(1-Chloropropane)
Concen: 18.001 ng/ul
RT: 8.516 min Scan# 940
Delta R.T. 0.012 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion: 45 Resp: 790977
Ion Ratio Lower Upper
45 100
77 14.7 11.4 17.0
79 11.5 8.4 12.6









#16
Acetophenone
Concen: 18.744 ng/ul
RT: 8.805 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:105 Resp: 976487
Ion Ratio Lower Upper
105 100
77 74.1 60.0 90.0
51 24.7 21.0 31.6



#17
N-Nitroso-di-n-propylamine
Concen: 18.656 ng/ul
RT: 8.793 min Scan# 987
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion: 70 Resp: 452592
Ion Ratio Lower Upper
70 100
42 55.6 46.3 69.5
101 9.9 8.1 12.1
130 23.3 17.3 25.9









#18
4-Methylphenol
Concen: 17.918 ng/ul
RT: 8.746 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:108 Resp: 632237
Ion Ratio Lower Upper
108 100
107 110.8 89.0 133.4



#19
Hexachloroethane
Concen: 18.288 ng/ul
RT: 9.069 min Scan# 1034
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion:117 Resp: 247847
Ion Ratio Lower Upper
117 100
201 102.2 82.7 124.1
199 63.7 51.4 77.2









#22
Nitrobenzene
Concen: 19.517 ng/ul
RT: 9.187 min Scan# 1053
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion: 77 Resp: 638255
Ion Ratio Lower Upper
77 100
123 47.4 37.6 56.4
65 13.6 11.0 16.4



#23
Isophorone
Concen: 18.949 ng/ul
RT: 9.716 min Scan# 1144
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion: 82 Resp: 1255513
Ion Ratio Lower Upper
82 100
95 7.0 5.6 8.4
138 19.2 15.2 22.8









#25
2-Nitrophenol
Concen: 20.560 ng/ul
RT: 9.904 min Scan# 1176
Delta R.T. 0.012 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:139	Resp:	291439
Ion Ratio	Lower	Upper
139	100	
65	45.2	38.5 57.7
109	27.4	22.3 33.5



#26
2,4-Dimethylphenol
Concen: 18.472 ng/ul
RT: 9.957 min Scan# 1185
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion:107	Resp:	660067
Ion Ratio	Lower	Upper
107	100	
121	52.8	42.9 64.3
122	90.2	71.8 107.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 18.614 ng/ul
 RT: 10.204 min Scan# 1227
 Delta R.T. 0.006 min
 Lab File: BP011606.D
 Acq: 01 Sep 2022 15:43




Instrument :
 BNA_P
 ClientSampleId :
 SICV616


Tgt Ion: 93 Resp: 799247
 Ion Ratio Lower Upper
 93 100
 95 32.8 26.3 39.5
 123 12.0 8.7 13.1



#29
 2,4-Dichlorophenol
 Concen: 19.310 ng/ul
 RT: 10.434 min Scan# 1266
 Delta R.T. 0.006 min
 Lab File: BP011606.D
 Acq: 01 Sep 2022 15:43

Tgt Ion: 162 Resp: 593963
 Ion Ratio Lower Upper
 162 100
 164 64.4 53.4 80.2
 98 34.6 28.3 42.5









#30
Naphthalene
Concen: 18.459 ng/ul
RT: 10.846 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:128 Resp: 2018595
Ion Ratio Lower Upper
128 100
129 11.2 8.8 13.2
127 13.1 10.2 15.4



#32
4-Chloroaniline
Concen: 18.543 ng/ul
RT: 10.951 min Scan# 1354
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion:127 Resp: 869913
Ion Ratio Lower Upper
127 100
129 32.8 26.3 39.5









#33
Hexachlorobutadiene
Concen: 18.946 ng/ul
RT: 11.140 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:225 Resp: 388115
Ion Ratio Lower Upper
225 100
223 61.3 50.5 75.7
227 63.0 51.1 76.7



#34
Caprolactam
Concen: 21.719 ng/ul
RT: 11.710 min Scan# 1483
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion:113 Resp: 196346
Ion Ratio Lower Upper
113 100
55 163.6 137.1 205.7
56 120.5 102.1 153.1









#35
4-Chloro-3-methylphenol
Concen: 19.391 ng/ul
RT: 12.069 min Scan# 1544
Delta R.T. 0.012 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:107 Resp: 592979
Ion Ratio Lower Upper
107 100
144 27.9 21.4 32.0
142 87.3 66.5 99.7



#36
2-Methylnaphthalene
Concen: 19.170 ng/ul
RT: 12.451 min Scan# 1609
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion:142 Resp: 1404606
Ion Ratio Lower Upper
142 100
141 89.1 71.5 107.3





#37

1-Methylnaphthalene

Concen: 18.398 ng/ul

RT: 12.675 min Scan# 1

Delta R.T. 0.012 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43

Instrument :

BNA_P

ClientSampleId :

SICV616





Tgt Ion:142 Resp: 1351169

Ion Ratio Lower Upper

142 100

141 91.4 74.2 111.4

116 3.7 2.9 4.3



#39

1,2,4,5-Tetrachlorobenzene

Concen: 18.782 ng/ul

RT: 12.816 min Scan# 1671

Delta R.T. 0.006 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43

Tgt Ion:216 Resp: 756800




Ion Ratio Lower Upper


216 100

214 78.4 62.6 94.0

179 20.6 16.1 24.1

108 16.7 13.2 19.8





#40

Hexachlorocyclopentadiene

Concen: 24.959 ng/ul

RT: 12.804 min Scan# 1

Delta R.T. 0.012 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43

Instrument :

BNA_P

ClientSampleId :

SICV616





Tgt Ion:237 Resp: 597968

Ion Ratio Lower Upper

237 100

235 62.2 49.4 74.0

272 13.6 10.5 15.7



#41

2,4,6-Trichlorophenol

Concen: 19.424 ng/ul

RT: 13.051 min Scan# 1711

Delta R.T. 0.006 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43




Tgt Ion:196 Resp: 442628


Ion Ratio Lower Upper

196 100

198 97.5 75.1 112.7

200 30.8 24.9 37.3





#42

2,4,5-Trichlorophenol

Concen: 19.612 ng/ul

RT: 13.122 min Scan# 1721

Delta R.T. 0.012 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43

Instrument :

BNA_P

ClientSampleId :

SICV616





Tgt Ion:196 Resp: 490853

Ion Ratio Lower Upper

196 100

198 96.5 77.0 115.6

200 30.9 24.1 36.1



#43

1,1'-Biphenyl

Concen: 18.541 ng/ul

RT: 13.457 min Scan# 1780

Delta R.T. 0.006 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43




Tgt Ion:154 Resp: 1849965


Ion Ratio Lower Upper

154 100

153 39.8 31.4 47.0

76 12.4 9.8 14.8





#44

2-Chloronaphthalene

Concen: 18.510 ng/ul

RT: 13.498 min Scan# 11

Delta R.T. 0.006 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43

Instrument :

BNA_P

ClientSampleId :

SICV616





Tgt Ion:162 Resp: 1414242

Ion Ratio Lower Upper

162 100

164 32.7 26.5 39.7

127 37.7 30.3 45.5



#45

2-Nitroaniline

Concen: 20.499 ng/ul

RT: 13.698 min Scan# 1821

Delta R.T. 0.006 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43




Tgt Ion: 65 Resp: 324436


Ion Ratio Lower Upper

65 100

92 71.1 55.4 83.2

138 115.0 89.1 133.7









#47
Dimethylphthalate
Concen: 18.740 ng/ul
RT: 14.081 min Scan# 1885
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:163 Resp: 1709042
Ion Ratio Lower Upper
163 100
194 4.0 3.3 4.9
164 10.4 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 23.164 ng/ul
RT: 14.192 min Scan# 1905
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion:165 Resp: 338530
Ion Ratio Lower Upper
165 100
89 55.9 49.1 73.7
121 21.2 19.8 29.8





#50

Acenaphthylene

Concen: 18.017 ng/ul

RT: 14.345 min Scan# 1930

Delta R.T. 0.006 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43

Instrument :

BNA_P

ClientSampleId :

SICV616





Tgt Ion:152 Resp: 2188705

Ion Ratio Lower Upper

152 100

151 19.8 15.8 23.6

153 13.2 10.5 15.7



#51

3-Nitroaniline

Concen: 20.658 ng/ul

RT: 14.522 min Scan# 1961

Delta R.T. 0.012 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43




Tgt Ion:138 Resp: 361156


Ion Ratio Lower Upper

138 100

108 12.6 17.0 25.4#

92 107.6 88.2 132.4









#52
Acenaphthene
Concen: 18.732 ng/ul
RT: 14.687 min Scan# 1988
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:153 Resp: 1488102
Ion Ratio Lower Upper
153 100
152 47.8 39.0 58.6
154 89.9 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 22.492 ng/ul
RT: 14.722 min Scan# 1995
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion:184 Resp: 143914
Ion Ratio Lower Upper
184 100
63 60.2 49.8 74.6
154 60.7 52.3 78.5









#55
4-Nitrophenol
Concen: 20.517 ng/ul
RT: 14.816 min Scan# 2009
Delta R.T. 0.012 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:109 Resp: 244684
Ion Ratio Lower Upper
109 100
139 148.0 108.4 162.6
65 130.6 104.7 157.1



#56
Dibenzofuran
Concen: 18.584 ng/ul
RT: 15.022 min Scan# 2046
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion:168 Resp: 2094163
Ion Ratio Lower Upper
168 100
139 36.3 29.1 43.7









#57
2,4-Dinitrotoluene
Concen: 20.971 ng/ul
RT: 14.981 min Scan# 2039
Delta R.T. 0.012 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:165 Resp: 426776
Ion Ratio Lower Upper
165 100
63 39.7 37.3 55.9
182 2.8 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 21.134 ng/ul
RT: 15.245 min Scan# 2084
Delta R.T. 0.012 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion:232 Resp: 398316
Ion Ratio Lower Upper
232 100
131 43.6 37.7 56.5
130 2.3 2.0 3.0
166 29.7 25.5 38.3









#59
Diethylphthalate
Concen: 18.838 ng/ul
RT: 15.445 min Scan# 2118
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:149 Resp: 1672973
Ion Ratio Lower Upper
149 100
177 22.2 17.7 26.5
150 12.6 9.9 14.9



#61
Fluorene
Concen: 18.951 ng/ul
RT: 15.675 min Scan# 2157
Delta R.T. 0.012 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion:166 Resp: 1726108
Ion Ratio Lower Upper
166 100
165 97.3 77.7 116.5
167 13.4 11.1 16.7









#62
4-Chlorophenyl-phenylether
Concen: 18.630 ng/ul
RT: 15.669 min Scan# 2155
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:204 Resp: 860020
Ion Ratio Lower Upper
204 100
206 33.4 26.5 39.7
141 55.8 45.0 67.6



#63
4-Nitroaniline
Concen: 22.364 ng/ul
RT: 15.681 min Scan# 2158
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion:138 Resp: 369003
Ion Ratio Lower Upper
138 100
92 47.6 39.1 58.7
108 71.1 65.5 98.3









#66
4,6-Dinitro-2-methylphenol
Concen: 16.846 ng/ul
RT: 15.739 min Scan# 2166
Delta R.T. 0.012 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:198 Resp: 182504
Ion Ratio Lower Upper
198 100
182 4.6 3.2 4.8
77 24.2 19.4 29.0



#67
N-Nitrosodiphenylamine
Concen: 18.808 ng/ul
RT: 15.881 min Scan# 2192
Delta R.T. 0.012 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion:169 Resp: 1453089
Ion Ratio Lower Upper
169 100
168 66.7 53.8 80.6
167 36.1 29.2 43.8









#68
4-Bromophenyl-phenylether
Concen: 18.789 ng/ul
RT: 16.569 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:248 Resp: 523718
Ion Ratio Lower Upper
248 100
250 97.7 79.0 118.4
141 65.5 57.1 85.7



#69
Hexachlorobenzene
Concen: 18.647 ng/ul
RT: 16.675 min Scan# 2327
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion:284 Resp: 616678
Ion Ratio Lower Upper
284 100
142 34.2 27.5 41.3
249 30.1 23.9 35.9









#70
Atrazine
Concen: 17.803 ng/ul
RT: 16.834 min Scan# 21
Delta R.T. 0.006 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:200 Resp: 499126
Ion Ratio Lower Upper
200 100
173 27.1 21.4 32.2
215 48.7 40.0 60.0



#71
Pentachlorophenol
Concen: 21.888 ng/ul
RT: 17.022 min Scan# 2386
Delta R.T. 0.012 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion:266 Resp: 386821
Ion Ratio Lower Upper
266 100
264 64.1 50.8 76.2
268 66.9 51.0 76.4









#72
Phenanthrene
Concen: 18.563 ng/ul
RT: 17.422 min Scan# 2452
Delta R.T. 0.012 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:178 Resp: 2620721
Ion Ratio Lower Upper
178 100
179 15.5 12.6 18.8
176 19.6 15.5 23.3



#74
Anthracene
Concen: 18.787 ng/ul
RT: 17.516 min Scan# 2470
Delta R.T. 0.012 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion:178 Resp: 2655594
Ion Ratio Lower Upper
178 100
179 15.2 12.0 18.0
176 19.0 15.0 22.6





#75

1,2,3,4-Tetrachlorobenzene

Concen: 17.645 ng/uL

RT: 13.422 min Scan# 1774

Delta R.T. 0.006 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43

Instrument :

BNA_P

ClientSampleId :

SICV616





Tgt Ion:216 Resp: 732539

Ion Ratio Lower Upper

216 100

214 78.4 61.7 92.5

218 48.5 38.6 58.0



#76

Pentachlorobenzene

Concen: 18.375 ng/uL

RT: 14.939 min Scan# 2032

Delta R.T. 0.012 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43




Tgt Ion:250 Resp: 720378


Ion Ratio Lower Upper

250 100

248 62.1 50.0 75.0

252 62.4 51.6 77.4





#77

Carbazole

Concen: 19.566 ng/ul

RT: 17.781 min Scan# 2515

Delta R.T. 0.012 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43

Instrument :

BNA_P

ClientSampleId :

SICV616





Tgt Ion:167 Resp: 2437271

Ion Ratio Lower Upper

167 100

166 21.3 17.1 25.7

139 12.5 9.9 14.9



#78

Di-n-butylphthalate

Concen: 19.559 ng/ul

RT: 18.333 min Scan# 2609

Delta R.T. 0.006 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43




Tgt Ion:149 Resp: 2765418


Ion Ratio Lower Upper

149 100

150 9.3 7.2 10.8

104 4.9 4.1 6.1









#80
Fluoranthene
Concen: 18.552 ng/ul
RT: 19.380 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:202 Resp: 2990464
Ion Ratio Lower Upper
202 100
101 10.7 9.4 14.2
100 8.0 7.3 10.9



#82
Pyrene
Concen: 18.760 ng/ul
RT: 19.733 min Scan# 2847
Delta R.T. 0.012 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion:202 Resp: 3119498
Ion Ratio Lower Upper
202 100
101 12.1 10.6 16.0
100 9.7 8.6 12.8





#83

Butylbenzylphthalate

Concen: 19.619 ng/ul

RT: 20.610 min Scan# 2995

Delta R.T. 0.006 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43

Instrument :

BNA_P

ClientSampleId :

SICV616





Tgt Ion:149 Resp: 1057788

Ion Ratio Lower Upper

149 100

91 68.9 54.4 81.6

206 20.9 16.7 25.1



#84

3,3'-Dichlorobenzidine

Concen: 25.171 ng/ul

RT: 21.374 min Scan# 3126

Delta R.T. 0.012 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43




Tgt Ion:252 Resp: 1150037


Ion Ratio Lower Upper

252 100

254 64.0 52.0 78.0

126 10.0 8.7 13.1





#85

Benzo(a)anthracene

Concen: 19.096 ng/ul

RT: 21.445 min Scan# 3136

Delta R.T. 0.012 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43

Instrument :

BNA_P

ClientSampleId :

SICV616





Tgt Ion:228 Resp: 3035665

Ion Ratio Lower Upper

228 100

229 19.9 15.6 23.4

226 27.4 22.2 33.2



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.695 ng/ul

RT: 21.374 min Scan# 3126

Delta R.T. 0.006 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43




Tgt Ion:149 Resp: 1779953


Ion Ratio Lower Upper

149 100

167 27.5 21.6 32.4

279 4.3 3.2 4.8





#87

Chrysene

Concen: 18.429 ng/ul

RT: 21.498 min Scan# 3145

Delta R.T. 0.012 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43

Instrument :

BNA_P

ClientSampleId :

SICV616





Tgt Ion:228 Resp: 2829029

Ion Ratio Lower Upper

228 100

226 30.8 24.3 36.5

229 20.0 15.8 23.6



#89

Di-n-octyl phthalate

Concen: 18.529 ng/ul




RT: 22.333 min Scan# 3289


Delta R.T. 0.006 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43

Tgt Ion:149 Resp: 2650393









#90
Benzo(b)fluoranthene
Concen: 18.232 ng/ul
RT: 23.180 min Scan# 3431
Delta R.T. 0.012 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43




Instrument :
BNA_P
ClientSampleId :
SICV616


Tgt Ion:252 Resp: 2965649
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.8
125 9.7 8.2 12.4



#91
Benzo(k)fluoranthene
Concen: 19.202 ng/ul
RT: 23.233 min Scan# 3442
Delta R.T. 0.018 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion:252 Resp: 2989857
Ion Ratio Lower Upper
252 100
253 21.9 18.0 27.0
125 9.2 8.2 12.4





#93

Benzo(a)pyrene

Concen: 16.640 ng/ul

RT: 23.833 min Scan# 3

Delta R.T. 0.018 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43

Instrument :

BNA_P

ClientSampleId :

SICV616





Tgt Ion:252 Resp: 2623702

Ion Ratio Lower Upper

252 100

253 21.9 17.8 26.6

125 10.3 8.9 13.3



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.022 ng/ul

RT: 26.545 min Scan# 4005

Delta R.T. 0.018 min

Lab File: BP011606.D

Acq: 01 Sep 2022 15:43




Tgt Ion:276 Resp: 3279550


Ion Ratio Lower Upper

276 100

138 22.4 17.2 25.8

277 24.6 19.4 29.2









#95
Dibenzo(a,h)anthracene
Concen: 18.421 ng/ul
RT: 26.568 min Scan# 4009
Delta R.T. 0.024 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Instrument :
BNA_P
ClientSampleId :
SICV616

Tgt Ion:278 Resp: 2890228
Ion Ratio Lower Upper
278 100
139 15.1 12.4 18.6
279 23.8 19.3 28.9



#96
Benzo(g,h,i)perylene
Concen: 18.419 ng/ul
RT: 27.339 min Scan# 4140
Delta R.T. 0.024 min
Lab File: BP011606.D
Acq: 01 Sep 2022 15:43

Tgt Ion:276 Resp: 2796096
Ion Ratio Lower Upper
276 100
138 21.3 16.6 24.8
277 24.6 19.0 28.6

