

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090324\
 Data File : BP021794.D
 Acq On : 03 Sep 2024 19:13
 Operator : RC/JU
 Sample : P3722-02MS
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DCYD5MS

Quant Time: Sep 04 04:48:01 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 29 23:56:59 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 09/04/2024
 Supervised By :mohammad ahmed 09/07/2024

Supervised By :mohammad
 ahmed
 09/07/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	217530	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.569	136	876792	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.422	164	545423	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.228	188	1198643	20.000	ng/u1	0.00
79) Chrysene-d12	21.680	240	1227059	20.000	ng/u1	0.01
88) Perylene-d12	25.080	264	1419838	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.276	96	32413	4.947	ng/uL	0.00
4) Pyridine-d5	3.687	84	479250	25.288	ng/u1	0.00
7) Phenol-d5	6.952	99	650512	28.021	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.111	67	354522	28.062	ng/u1	0.00
11) 2-Chlorophenol-d4	7.317	132	431695	28.803	ng/u1	0.00
15) 4-Methylphenol-d8	8.481	113	511090	28.491	ng/u1	0.00
21) Nitrobenzene-d5	8.928	128	212652	30.146	ng/u1	0.00
24) 2-Nitrophenol-d4	9.646	143	213243	29.781	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.187	165	430308	30.496	ng/u1	0.00
31) 4-Chloroaniline-d4	10.693	131	571610	27.845	ng/u1	0.00
46) Dimethylphthalate-d6	13.822	166	1303701	31.238	ng/u1	-0.02
49) Acenaphthylene-d8	14.110	160	1489598	31.791	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.622	143	253301	31.902	ng/u1	0.00
60) Fluorene-d10	15.434	176	1122720	31.998	ng/u1	-0.01
65) 4,6-Dinitro-2-methylph...	15.546	200	186045	28.220	ng/u1	-0.01
73) Anthracene-d10	17.328	188	1788278	31.436	ng/u1	-0.01
81) Pyrene-d10	19.698	212	2286960	32.604	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.851	264	2582254	34.107	ng/u1	0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.305	88	93658	13.450	ng/uL	97
5) Pyridine	3.705	79	634856	33.523	ng/u1	98
6) Benzaldehyde	6.922	77	461249	38.845	ng/u1	98
8) Phenol	6.975	94	841190	34.751	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.199	93	651546	34.628	ng/u1	99
12) 2-Chlorophenol	7.346	128	569967	36.232	ng/u1	98
13) 2-Methylphenol	8.216	108	600410	34.383	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.293	45	631833	34.791	ng/u1	98
16) Acetophenone	8.593	105	1001911	34.838	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.581	70	456695	33.650	ng/u1	97
18) 4-Methylphenol	8.546	108	668281	34.956	ng/u1	99
19) Hexachloroethane	8.858	117	252530	35.902	ng/u1	96
22) Nitrobenzene	8.969	77	731365	37.182	ng/u1	98
23) Isophorone	9.499	82	1342072	34.001	ng/u1	99
25) 2-Nitrophenol	9.675	139	298019	37.565	ng/u1	97
26) 2,4-Dimethylphenol	9.740	107	670719	34.446	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.975	93	860706	34.770	ng/u1	98
29) 2,4-Dichlorophenol	10.210	162	520042	36.677	ng/u1	97
30) Naphthalene	10.616	128	1781846	36.058	ng/u1	100
32) 4-Chloroaniline	10.716	127	681724	32.894	ng/u1	99
33) Hexachlorobutadiene	10.905	225	337463	35.472	ng/u1	98
34) Caprolactam	11.499	113	220494	36.774	ng/u1	97
35) 4-Chloro-3-methylphenol	11.852	107	694423	37.003	ng/u1	99
36) 2-Methylnaphthalene	12.228	142	1189668	35.593	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090324\
 Data File : BP021794.D
 Acq On : 03 Sep 2024 19:13
 Operator : RC/JU
 Sample : P3722-02MS
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DCYD5MS

Quant Time: Sep 04 04:48:01 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 29 23:56:59 2024
 Response via : Initial Calibration


Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 09/04/2024
 Supervised By :mohammad ahmed 09/07/2024

Supervised By :mohammad
 ahmed
 09/07/2024


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.446	142	1190928	35.424	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.599	216	628679	36.571	ng/ul	99
40) Hexachlorocyclopentadiene	12.581	237	304760	30.896	ng/ul	99
41) 2,4,6-Trichlorophenol	12.846	196	418575	37.818	ng/ul	99
42) 2,4,5-Trichlorophenol	12.910	196	453167	37.918	ng/ul	98
43) 1,1'-Biphenyl	13.240	154	1542238	37.374	ng/ul	99
44) 2-Chloronaphthalene	13.281	162	1198579	37.045	ng/ul	99
45) 2-Nitroaniline	13.487	65	406057	40.061	ng/ul	97
47) Dimethylphthalate	13.875	163	1512008	36.182	ng/ul	99
48) 2,6-Dinitrotoluene	13.987	165	298753	38.764	ng/ul	97
50) Acenaphthylene	14.140	152	1886266	37.431	ng/ul	100
51) 3-Nitroaniline	14.316	138	319446	37.930	ng/ul	94
52) Acenaphthene	14.487	153	1338805	37.778	ng/ul	99
53) 2,4-Dinitrophenol	14.528	184	154088	33.396	ng/ul	95
55) 4-Nitrophenol	14.640	109	303266	41.714	ng/ul	92
56) Dibenzofuran	14.828	168	1825559	37.235	ng/ul	98
57) 2,4-Dinitrotoluene	14.787	165	451341	38.971	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	15.057	232	376131	36.529	ng/ul	93
59) Diethylphthalate	15.263	149	1549775	36.931	ng/ul	99
61) Fluorene	15.487	166	1467641	37.009	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.487	204	718024	36.034	ng/ul	99
63) 4-Nitroaniline	15.498	138	333865	41.017	ng/ul	93
66) 4,6-Dinitro-2-methylph...	15.563	198	251203	34.064	ng/ul	98
67) N-Nitrosodiphenylamine	15.704	169	1278226	36.250	ng/ul	99
68) 4-Bromophenyl-phenylether	16.398	248	466966	35.290	ng/ul	96
69) Hexachlorobenzene	16.522	284	544048	34.633	ng/ul	98
70) Atrazine	16.681	200	481577	36.004	ng/ul	99
71) Pentachlorophenol	16.875	266	344542	34.270	ng/ul	98
72) Phenanthrene	17.269	178	2435370	37.008	ng/ul	99
74) Anthracene	17.369	178	2413384	36.200	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.210	216	624226	35.028	ng/ul	97
76) Pentachlorobenzene	14.746	250	623466	33.754	ng/ul	98
77) Carbazole	17.645	167	2364139	39.428	ng/ul	100
78) Di-n-butylphthalate	18.245	149	2749812	37.676	ng/ul	100
80) Fluoranthene	19.357	202	3033155	37.056	ng/ul	99
82) Pyrene	19.733	202	3251515	37.507	ng/ul	100
83) Butylbenzylphthalate	20.681	149	1331121	37.502	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.580	252	928099	30.013	ng/ul	99
85) Benzo(a)anthracene	21.663	228	3340731	38.273	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.604	149	1890553	36.087	ng/ul	99
87) Chrysene	21.733	228	3092115	37.641	ng/ul	99
89) Di-n-octyl phthalate	22.898	149	3326298	37.011	ng/ul	100
90) Benzo(b)fluoranthene	24.004	252	3467998	39.036	ng/ul	98
91) Benzo(k)fluoranthene	24.080	252	3378316	38.843	ng/ul	98
93) Benzo(a)pyrene	24.921	252	3190805	38.984	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	28.945	276	4080579m	38.666	ng/ul	
95) Dibenzo(a,h)anthracene	29.033	278	3303626	38.956	ng/ul	100
96) Benzo(g,h,i)perylene	30.133	276	3229837	39.610	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 13.450 ng/uL
RT: 3.305 min Scan# 54
Delta R.T. -0.006 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion: 88 Resp: 93658
Ion Ratio Lower Upper
88 100
43 22.0 15.4 23.2
58 61.3 48.2 72.2

Manual Integrations
APPROVED




Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024


Supervised By :mohammad
ahmed
09/07/2024



#5
Pyridine
Concen: 33.523 ng/uL
RT: 3.705 min Scan# 122
Delta R.T. 0.000 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13


Tgt Ion: 79 Resp: 634856
Ion Ratio Lower Upper
79 100
52 56.6 46.8 70.2
51 28.0 21.7 32.5





#6
Benzaldehyde
Concen: 38.845 ng/ul
RT: 6.922 min Scan# 669
Delta R.T. -0.006 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13



Instrument :
BNA_P
ClientSampleId :
DCYD5MS




Tgt Ion: 77 Resp: 461249
Ion Ratio Lower Upper
77 100
105 98.5 80.2 120.4
106 94.0 76.2 114.4

Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024





Supervised By :mohammad
ahmed
09/07/2024




#8
Phenol
Concen: 34.751 ng/ul
RT: 6.975 min Scan# 678
Delta R.T. -0.006 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13




Tgt Ion: 94 Resp: 841190
Ion Ratio Lower Upper
94 100
65 26.3 20.0 30.0
66 35.7 27.0 40.4





#10
Bis(2-Chloroethyl)ether
Concen: 34.628 ng/ul
RT: 7.199 min Scan# 716
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion: 93 Resp: 651540
Ion Ratio Lower Upper
93 100
63 69.1 54.9 82.3
95 32.1 26.0 39.0

Manual Integrations
APPROVED




Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024


Supervised By :mohammad
ahmed
09/07/2024



#12
2-Chlorophenol
Concen: 36.232 ng/ul
RT: 7.346 min Scan# 741
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13


Tgt Ion:128 Resp: 569967
Ion Ratio Lower Upper
128 100
64 50.6 39.5 59.3
130 31.6 25.9 38.9





#13
2-Methylphenol
Concen: 34.383 ng/ul
RT: 8.216 min Scan# 890
Delta R.T. -0.006 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion:108 Resp: 600410
Ion Ratio Lower Upper
108 100
107 91.8 73.2 109.8

Manual Integrations
APPROVED




Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024


Supervised By :mohammad
ahmed
09/07/2024



#14
2,2'-oxybis(1-Chloropropane)
Concen: 34.791 ng/ul
RT: 8.293 min Scan# 902
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13


Tgt Ion: 45 Resp: 631833
Ion Ratio Lower Upper
45 100
77 19.9 17.0 25.4
79 14.8 12.0 18.0





#16
Acetophenone
Concen: 34.838 ng/ul
RT: 8.593 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion: 105 Resp: 100191
Ion Ratio Lower Upper
105 100
77 75.2 59.4 89.2
51 20.4 16.4 24.6

Manual Integrations APPROVED




Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024


Supervised By :mohammad
ahmed
09/07/2024



#17
N-Nitroso-di-n-propylamine
Concen: 33.650 ng/ul
RT: 8.581 min Scan# 951
Delta R.T. -0.006 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13


Tgt Ion: 70 Resp: 456695
Ion Ratio Lower Upper
70 100
42 47.3 35.8 53.6
101 11.0 8.2 12.4
130 19.6 15.1 22.7





#18
4-Methylphenol
Concen: 34.956 ng/ul
RT: 8.546 min Scan# 945
Delta R.T. 0.000 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS





Tgt Ion:108 Resp: 66828
Ion Ratio Lower Upper
108 100
107 111.0 90.1 135.1

Manual Integrations
APPROVED




Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024


Supervised By :mohammad
ahmed
09/07/2024



#19
Hexachloroethane
Concen: 35.902 ng/ul
RT: 8.858 min Scan# 998
Delta R.T. -0.006 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13


Tgt Ion:117 Resp: 252530
Ion Ratio Lower Upper
117 100
201 91.1 75.3 112.9
199 55.8 47.4 71.0





#22
Nitrobenzene
Concen: 37.182 ng/ul
RT: 8.969 min Scan# 1017
Delta R.T. -0.006 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13



Instrument :
BNA_P
ClientSampleId :
DCYD5MS




Tgt Ion: 77 Resp: 731365
Ion Ratio Lower Upper
77 100
123 42.6 33.2 49.8
65 12.9 9.8 14.6

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024







Supervised By :mohammad
ahmed
09/07/2024



#23
Isophorone
Concen: 34.001 ng/ul
RT: 9.499 min Scan# 1107
Delta R.T. -0.006 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13


Tgt Ion: 82 Resp: 1342072
Ion Ratio Lower Upper
82 100
95 7.8 5.8 8.8
138 14.8 11.8 17.8





#25
2-Nitrophenol
Concen: 37.565 ng/ul
RT: 9.675 min Scan# 1137
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion:139 Resp: 298019
Ion Ratio Lower Upper
139 100
65 58.3 46.2 69.2
109 38.7 28.0 42.0

Manual Integrations
APPROVED




Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024


Supervised By :mohammad
ahmed
09/07/2024



#26
2,4-Dimethylphenol
Concen: 34.446 ng/ul
RT: 9.740 min Scan# 1148
Delta R.T. -0.006 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Tgt Ion:107 Resp: 670719
Ion Ratio Lower Upper
107 100
121 45.4 36.9 55.3
122 74.6 61.0 91.4





#27

Bis(2-Chloroethoxy)methane

Concen: 34.770 ng/ul

RT: 9.975 min Scan# 1188

Delta R.T. -0.006 min

Lab File: BP021794.D

Acq: 03 Sep 2024 19:13

Instrument :

BNA_P

ClientSampleId :

DCYD5MS

Tgt Ion: 93 Resp: 860700

Ion Ratio Lower Upper

93 100

95 33.0 25.3 37.9

123 9.8 7.4 11.0

Manual Integrations

APPROVED





Reviewed By :Jagrut Upadhyay 09/04/2024

Supervised By :mohammad ahmed 09/07/2024

Supervised By :mohammad

ahmed

09/07/2024



#29

2,4-Dichlorophenol

Concen: 36.677 ng/ul

RT: 10.210 min Scan# 1228

Delta R.T. -0.006 min

Lab File: BP021794.D

Acq: 03 Sep 2024 19:13




Tgt Ion:162 Resp: 520042


Ion Ratio Lower Upper

162 100

164 65.9 52.3 78.5


98 48.1 35.3 52.9





#30
Naphthalene
Concen: 36.058 ng/ul
RT: 10.616 min Scan# 1297
Delta R.T. -0.006 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion:128 Resp: 1781840
Ion Ratio Lower Upper
128 100
129 10.8 8.6 13.0
127 13.3 10.5 15.7

Manual Integrations
APPROVED




Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024


Supervised By :mohammad
ahmed
09/07/2024



#32
4-Chloroaniline
Concen: 32.894 ng/ul
RT: 10.716 min Scan# 1314
Delta R.T. -0.006 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13


Tgt Ion:127 Resp: 681724
Ion Ratio Lower Upper
127 100
129 31.8 25.8 38.8





#33
Hexachlorobutadiene
Concen: 35.472 ng/ul
RT: 10.905 min Scan# 1346
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion:225 Resp: 33746
Ion Ratio Lower Upper
225 100
223 62.0 51.5 77.3
227 63.7 51.3 76.9

Manual Integrations
APPROVED




Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024


Supervised By :mohammad
ahmed
09/07/2024



#34
Caprolactam
Concen: 36.774 ng/ul
RT: 11.499 min Scan# 1447
Delta R.T. -0.006 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13


Tgt Ion:113 Resp: 220494
Ion Ratio Lower Upper
113 100
55 141.5 110.2 165.2
56 109.2 84.9 127.3





#35
4-Chloro-3-methylphenol
Concen: 37.003 ng/ul
RT: 11.852 min Scan# 1507
Delta R.T. -0.006 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion:107 Resp: 69442
Ion Ratio Lower Upper
107 100
144 21.3 17.5 26.3
142 65.7 53.3 79.9


Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024



Supervised By :mohammad
ahmed
09/07/2024




#36
2-Methylnaphthalene
Concen: 35.593 ng/ul
RT: 12.228 min Scan# 1571
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13




Tgt Ion:142 Resp: 1189668
Ion Ratio Lower Upper
142 100
141 89.9 71.1 106.7





#37
1-Methylnaphthalene
Concen: 35.424 ng/ul
RT: 12.446 min Scan# 1608
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion:142 Resp: 1190928
Ion Ratio Lower Upper
142 100
141 92.3 74.6 111.8
116 4.5 3.5 5.3

Manual Integrations
APPROVED




Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024


Supervised By :mohammad
ahmed
09/07/2024



#39
1,2,4,5-Tetrachlorobenzene
Concen: 36.571 ng/ul
RT: 12.599 min Scan# 1634
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13


Tgt Ion:216 Resp: 628679
Ion Ratio Lower Upper
216 100
214 78.8 63.2 94.8
179 21.8 16.6 25.0
108 21.2 16.2 24.4





#40
Hexachlorocyclopentadiene
Concen: 30.896 ng/ul
RT: 12.581 min Scan# 1631
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion:237 Resp: 304760
Ion Ratio Lower Upper
237 100
235 63.2 50.7 76.1
272 12.6 11.0 16.6

Manual Integrations
APPROVED




Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024


Supervised By :mohammad
ahmed
09/07/2024



#41
2,4,6-Trichlorophenol
Concen: 37.818 ng/ul
RT: 12.846 min Scan# 1676
Delta R.T. -0.006 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13


Tgt Ion:196 Resp: 418575
Ion Ratio Lower Upper
196 100
198 96.7 76.5 114.7
200 31.6 25.3 37.9





#42
2,4,5-Trichlorophenol
Concen: 37.918 ng/ul
RT: 12.910 min Scan# 1687
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion:196 Resp: 45316

Ion	Ratio	Lower	Upper
196	100		
198	95.5	78.0	117.0
200	30.6	24.3	36.5

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024




Supervised By :mohammad
ahmed
09/07/2024




#43
1,1'-Biphenyl
Concen: 37.374 ng/ul
RT: 13.240 min Scan# 1743
Delta R.T. -0.018 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Tgt Ion:154 Resp: 1542238


Ion	Ratio	Lower	Upper
154	100		
153	39.9	31.6	47.4
76	15.2	12.6	18.8





#44
2-Chloronaphthalene
Concen: 37.045 ng/ul
RT: 13.281 min Scan# 1750
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion:162 Resp: 1198579
Ion Ratio Lower Upper
162 100
164 32.3 26.0 39.0
127 42.3 33.1 49.7

Manual Integrations
APPROVED




Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024


Supervised By :mohammad
ahmed
09/07/2024



#45
2-Nitroaniline
Concen: 40.061 ng/ul
RT: 13.487 min Scan# 1785
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13


Tgt Ion: 65 Resp: 406057
Ion Ratio Lower Upper
65 100
92 76.2 65.4 98.2
138 97.2 77.5 116.3





#47
Dimethylphthalate
Concen: 36.182 ng/ul
RT: 13.875 min Scan# 1851
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion:163 Resp: 1512008
Ion Ratio Lower Upper
163 100
194 3.8 3.5 5.3
164 10.3 7.8 11.8

Manual Integrations
APPROVED




Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024


Supervised By :mohammad
ahmed
09/07/2024



#48
2,6-Dinitrotoluene
Concen: 38.764 ng/ul
RT: 13.987 min Scan# 1870
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13


Tgt Ion:165 Resp: 298753
Ion Ratio Lower Upper
165 100
89 73.0 55.7 83.5
121 23.8 18.6 28.0





#50
Acenaphthylene
Concen: 37.431 ng/ul
RT: 14.140 min Scan# 1896
Delta R.T. -0.018 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion:152 Resp: 1886260
Ion Ratio Lower Upper
152 100
151 19.6 15.6 23.4
153 13.0 10.2 15.4


Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024



Supervised By :mohammad
ahmed
09/07/2024




#51
3-Nitroaniline
Concen: 37.930 ng/ul
RT: 14.316 min Scan# 1926
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13



Tgt Ion:138 Resp: 319446
Ion Ratio Lower Upper
138 100
108 14.2 12.0 18.0
92 150.4 114.1 171.1





#52

Acenaphthene

Concen: 37.778 ng/ul

RT: 14.487 min Scan# 1956

Delta R.T. -0.018 min

Lab File: BP021794.D

Acq: 03 Sep 2024 19:13

Instrument :

BNA_P

ClientSampleId :

DCYD5MS

Tgt Ion:153 Resp: 133880

Ion Ratio Lower Upper

153 100

152 47.1 38.0 57.0





154 88.4 70.2 105.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024

Supervised By :mohammad
ahmed
09/07/2024



#53

2,4-Dinitrophenol

Concen: 33.396 ng/ul

RT: 14.528 min Scan# 1962

Delta R.T. -0.012 min

Lab File: BP021794.D

Acq: 03 Sep 2024 19:13




Tgt Ion:184 Resp: 154088


Ion Ratio Lower Upper

184 100

63 75.9 55.5 83.3


154 65.6 51.3 76.9





#55
4-Nitrophenol
Concen: 41.714 ng/ul
RT: 14.640 min Scan# 1981
Delta R.T. -0.006 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion:109 Resp: 303260
Ion Ratio Lower Upper
109 100
139 109.6 96.2 144.2
65 116.3 98.5 147.7

Manual Integrations
APPROVED




Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024


Supervised By :mohammad
ahmed
09/07/2024



#56
Dibenzofuran
Concen: 37.235 ng/ul
RT: 14.828 min Scan# 2013
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13


Tgt Ion:168 Resp: 1825559
Ion Ratio Lower Upper
168 100
139 38.3 29.8 44.8





#57
2,4-Dinitrotoluene
Concen: 38.971 ng/ul
RT: 14.787 min Scan# 2006
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion:165 Resp: 45134
Ion Ratio Lower Upper
165 100
63 48.9 37.8 56.8
182 2.9 2.2 3.4


Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024



Supervised By :mohammad
ahmed
09/07/2024




#58
2,3,4,6-Tetrachlorophenol
Concen: 36.529 ng/ul
RT: 15.057 min Scan# 2052
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13




Tgt Ion:232 Resp: 376131
Ion Ratio Lower Upper
232 100
131 48.3 33.9 50.9
130 2.7 2.0 3.0
166 32.9 24.3 36.5





#59
Diethylphthalate
Concen: 36.931 ng/ul
RT: 15.263 min Scan# 2087
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion:149 Resp: 1549775
Ion Ratio Lower Upper
149 100
177 21.9 18.0 27.0
150 12.5 9.8 14.8


Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024



Supervised By :mohammad
ahmed
09/07/2024




#61
Fluorene
Concen: 37.009 ng/ul
RT: 15.487 min Scan# 2125
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13




Tgt Ion:166 Resp: 1467641
Ion Ratio Lower Upper
166 100
165 99.6 78.2 117.2
167 13.7 10.8 16.2





#62
4-Chlorophenyl-phenylether
Concen: 36.034 ng/ul
RT: 15.487 min Scan# 2126
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion:204 Resp: 718024
Ion Ratio Lower Upper
204 100
206 33.6 26.2 39.4
141 61.4 48.6 73.0

Manual Integrations
APPROVED




Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024


Supervised By :mohammad
ahmed
09/07/2024



#63
4-Nitroaniline
Concen: 41.017 ng/ul
RT: 15.498 min Scan# 2127
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13


Tgt Ion:138 Resp: 333865
Ion Ratio Lower Upper
138 100
92 65.5 50.7 76.1
108 99.5 72.2 108.2





#66
4,6-Dinitro-2-methylphenol
Concen: 34.064 ng/ul
RT: 15.563 min Scan# 2140
Delta R.T. -0.006 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion:198 Resp: 251200
Ion Ratio Lower Upper
198 100
182 4.6 3.3 4.9
77 29.9 23.3 34.9

Manual Integrations
APPROVED




Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024


Supervised By :mohammad
ahmed
09/07/2024



#67
N-Nitrosodiphenylamine
Concen: 36.250 ng/ul
RT: 15.704 min Scan# 2162
Delta R.T. -0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Tgt Ion:169 Resp: 1278226
Ion Ratio Lower Upper
169 100
168 66.7 52.6 79.0
167 36.2 28.2 42.4





#68

4-Bromophenyl-phenylether

Concen: 35.290 ng/ul

RT: 16.398 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP021794.D

Acq: 03 Sep 2024 19:13

Instrument :

BNA_P

ClientSampleId :

DCYD5MS

Tgt Ion:248 Resp: 466960

Ion Ratio Lower Upper

248 100

250 98.2 79.6 119.4





141 77.0 57.0 85.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024

Supervised By :mohammad
ahmed
09/07/2024



#69

Hexachlorobenzene

Concen: 34.633 ng/ul

RT: 16.522 min Scan# 2301

Delta R.T. 0.000 min

Lab File: BP021794.D

Acq: 03 Sep 2024 19:13




Tgt Ion:284 Resp: 544048


Ion Ratio Lower Upper

284 100

142 33.0 26.5 39.7

249 32.1 24.2 36.4





#70

Atrazine

Concen: 36.004 ng/ul

RT: 16.681 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP021794.D

Acq: 03 Sep 2024 19:13

Instrument :

BNA_P

ClientSampleId :

DCYD5MS





Tgt Ion:	200	Resp:	48157
Ion Ratio	Lower	Upper	
200	100		
173	27.3	21.4	32.0
215	48.4	39.0	58.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024

Supervised By :mohammad
ahmed
09/07/2024



#71

Pentachlorophenol

Concen: 34.270 ng/ul




RT: 16.875 min Scan# 2361


Delta R.T. 0.000 min

Lab File: BP021794.D

Acq: 03 Sep 2024 19:13


Tgt Ion:	266	Resp:	344542
Ion Ratio	Lower	Upper	
266	100		
264	65.3	51.4	77.0
268	65.7	50.6	75.8





#72
Phenanthrene
Concen: 37.008 ng/ul
RT: 17.269 min Scan# 2430
Delta R.T. -0.006 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion:178 Resp: 2435370
Ion Ratio Lower Upper
178 100
179 15.3 12.0 18.0
176 18.8 14.8 22.2

Manual Integrations
APPROVED




Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024


Supervised By :mohammad
ahmed
09/07/2024



#74
Anthracene
Concen: 36.200 ng/ul
RT: 17.369 min Scan# 2445
Delta R.T. 0.000 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Tgt Ion:178 Resp: 2413384
Ion Ratio Lower Upper
178 100
179 15.0 12.3 18.5
176 18.5 14.6 21.8









#75
 1,2,3,4-Tetrachlorobenzene
 Concen: 35.028 ng/uL
 RT: 13.210 min Scan# 11
 Delta R.T. -0.012 min
 Lab File: BP021794.D
 Acq: 03 Sep 2024 19:13

Instrument : BNA_P
 ClientSampleId : DCYD5MS




Tgt Ion:	216	Resp:	624220
Ion Ratio	Lower	Upper	
216	100		
214	76.5	63.9	95.9
218	48.4	39.5	59.3


Manual Integrations
APPROVED
 Reviewed By :Jagrut Upadhyay 09/04/2024
 Supervised By :mohammad ahmed 09/07/2024



#76
 Pentachlorobenzene
 Concen: 33.754 ng/uL
 RT: 14.746 min Scan# 1999
 Delta R.T. -0.012 min
 Lab File: BP021794.D
 Acq: 03 Sep 2024 19:13


Tgt Ion:	250	Resp:	623466
Ion Ratio	Lower	Upper	
250	100		
248	62.3	51.0	76.6
252	62.9	51.4	77.0





#77
 Carbazole
 Concen: 39.428 ng/ul
 RT: 17.645 min Scan# 2493
 Delta R.T. 0.006 min
 Lab File: BP021794.D
 Acq: 03 Sep 2024 19:13

Instrument :
 BNA_P
 ClientSampleId :
 DCYD5MS






Tgt Ion:167 Resp: 2364139
 Ion Ratio Lower Upper
 167 100
 166 21.1 17.1 25.7
 139 13.1 10.4 15.6

Manual Integrations
 APPROVED




Reviewed By :Jagrut Upadhyay 09/04/2024
 Supervised By :mohammad ahmed 09/07/2024


Supervised By :mohammad
 ahmed
 09/07/2024



#78
 Di-n-butylphthalate
 Concen: 37.676 ng/ul
 RT: 18.245 min Scan# 2594
 Delta R.T. 0.006 min
 Lab File: BP021794.D
 Acq: 03 Sep 2024 19:13


Tgt Ion:149 Resp: 2749812
 Ion Ratio Lower Upper
 149 100
 150 9.0 7.4 11.0
 104 6.1 4.8 7.2





#80
Fluoranthene
Concen: 37.056 ng/u1
RT: 19.357 min Scan# 21
Delta R.T. 0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion:202 Resp: 303315
Ion Ratio Lower Upper
202 100
101 13.2 11.2 16.8
100 10.4 8.5 12.7

Manual Integrations
APPROVED




Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024


Supervised By :mohammad
ahmed
09/07/2024



#82
Pyrene
Concen: 37.507 ng/u1
RT: 19.733 min Scan# 2847
Delta R.T. 0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Tgt Ion:202 Resp: 3251515
Ion Ratio Lower Upper
202 100
101 16.3 12.8 19.2
100 13.4 10.8 16.2





#83

Butylbenzylphthalate

Concen: 37.502 ng/ul

RT: 20.681 min Scan# 3009

Delta R.T. 0.006 min

Lab File: BP021794.D

Acq: 03 Sep 2024 19:13

Instrument :

BNA_P

ClientSampleId :

DCYD5MS

Tgt Ion:149 Resp: 133112

Ion Ratio Lower Upper

149 100

91 89.0 70.1 105.1





206 20.3 16.8 25.2

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024

Supervised By :mohammad
ahmed
09/07/2024



#84

3,3'-Dichlorobenzidine

Concen: 30.013 ng/ul

RT: 21.580 min Scan# 3161

Delta R.T. 0.012 min

Lab File: BP021794.D

Acq: 03 Sep 2024 19:13




Tgt Ion:252 Resp: 928099


Ion Ratio Lower Upper

252 100

254 64.8 52.2 78.4

126 12.1 8.9 13.3





#85

Benzo(a)anthracene

Concen: 38.273 ng/ul

RT: 21.663 min Scan# 3176

Delta R.T. 0.018 min

Lab File: BP021794.D

Acq: 03 Sep 2024 19:13

Instrument :

BNA_P

ClientSampleId :

DCYD5MS

Tgt Ion:228 Resp: 334073

Ion Ratio Lower Upper

228 100

229 19.7 15.7 23.5





226 26.6 21.7 32.5

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024

Supervised By :mohammad
ahmed
09/07/2024



#86

Bis(2-ethylhexyl)phthalate

Concen: 36.087 ng/ul

RT: 21.604 min Scan# 3165

Delta R.T. 0.012 min

Lab File: BP021794.D

Acq: 03 Sep 2024 19:13




Tgt Ion:149 Resp: 1890553


Ion Ratio Lower Upper

149 100

167 26.6 21.8 32.6

279 4.4 3.7 5.5





#87

Chrysene

Concen: 37.641 ng/ul

RT: 21.733 min Scan# 3187

Delta R.T. 0.018 min

Lab File: BP021794.D

Acq: 03 Sep 2024 19:13

Instrument :

BNA_P

ClientSampleId :

DCYD5MS

Tgt Ion:228 Resp: 309211

Ion Ratio Lower Upper

228 100

226 29.7 23.5 35.3





229 19.8 15.6 23.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024

Supervised By :mohammad
ahmed
09/07/2024



#89

Di-n-octyl phthalate

Concen: 37.011 ng/ul




RT: 22.898 min Scan# 3385


Delta R.T. 0.006 min

Lab File: BP021794.D

Acq: 03 Sep 2024 19:13


Tgt Ion:149 Resp: 3326298





#90
Benzo(b)fluoranthene
Concen: 39.036 ng/ul
RT: 24.004 min Scan# 3573
Delta R.T. 0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion:252 Resp: 3467998
Ion Ratio Lower Upper
252 100
253 21.6 18.1 27.1
125 10.6 8.6 12.8

Manual Integrations
APPROVED




Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024


Supervised By :mohammad
ahmed
09/07/2024



#91
Benzo(k)fluoranthene
Concen: 38.843 ng/ul
RT: 24.080 min Scan# 3586
Delta R.T. 0.024 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13


Tgt Ion:252 Resp: 3378316
Ion Ratio Lower Upper
252 100
253 21.1 17.8 26.8
125 11.2 8.6 13.0





#93
Benzo(a)pyrene
Concen: 38.984 ng/ul
RT: 24.921 min Scan# 31
Delta R.T. 0.012 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Instrument :
BNA_P
ClientSampleId :
DCYD5MS






Tgt Ion:252 Resp: 319080
Ion Ratio Lower Upper
252 100
253 21.6 17.4 26.0
125 12.4 9.6 14.4

Manual Integrations
APPROVED




Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024


Supervised By :mohammad
ahmed
09/07/2024



#94
Indeno(1,2,3-cd)pyrene
Concen: 38.666 ng/ul m
RT: 28.945 min Scan# 4413
Delta R.T. 0.018 min
Lab File: BP021794.D
Acq: 03 Sep 2024 19:13

Tgt Ion:276 Resp: 4080579
Ion Ratio Lower Upper
276 100
138 20.7 14.9 22.3
277 23.8 19.0 28.6





#95

Dibenzo(a,h)anthracene

Concen: 38.956 ng/ul

RT: 29.033 min Scan# 4427

Delta R.T. 0.018 min

Lab File: BP021794.D

Acq: 03 Sep 2024 19:13

Instrument :

BNA_P

ClientSampleId :

DCYD5MS

Tgt Ion:278 Resp: 330362

Ion Ratio Lower Upper

278 100

139 16.0 12.7 19.1





279 23.5 18.9 28.3

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 09/04/2024
Supervised By :mohammad ahmed 09/07/2024

Supervised By :mohammad
ahmed
09/07/2024



#96

Benzo(g,h,i)perylene

Concen: 39.610 ng/ul

RT: 30.133 min Scan# 4615

Delta R.T. 0.012 min

Lab File: BP021794.D

Acq: 03 Sep 2024 19:13

Tgt Ion:276 Resp: 3229837

Ion Ratio Lower Upper

276 100

138 22.1 17.3 25.9

277 23.6 19.4 29.0

