

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090424\
 Data File : BP021817.D
 Acq On : 04 Sep 2024 13:38
 Operator : RC/JU
 Sample : P3808-01DL 5X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 X0B24DL

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 09/05/2024
 Supervised By :mohammad ahmed 09/07/2024

Quant Time: Sep 04 14:09:05 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 29 23:56:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.752	152	277721	20.000	ng/u1	-0.04
20) Naphthalene-d8	10.540	136	1159429	20.000	ng/u1	-0.03
38) Acenaphthene-d10	14.410	164	765675	20.000	ng/u1	-0.02
64) Phenanthrene-d10	17.222	188	1728026	20.000	ng/u1	0.00
79) Chrysene-d12	21.686	240	1766858	20.000	ng/u1	0.02
88) Perylene-d12	25.098	264	1923336	20.000	ng/u1	0.04
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	9898	1.183	ng/uL	-0.01
4) Pyridine-d5	3.675	84	131576	5.438	ng/u1	-0.01
7) Phenol-d5	6.922	99	177002	5.972	ng/u1	-0.04
9) Bis-(2-Chloroethyl)eth...	7.081	67	100064	6.204	ng/u1	-0.04
11) 2-Chlorophenol-d4	7.287	132	117688	6.150	ng/u1	-0.04
15) 4-Methylphenol-d8	8.452	113	132744	5.796	ng/u1	-0.04
21) Nitrobenzene-d5	8.899	128	55046	5.901	ng/u1	-0.04
24) 2-Nitrophenol-d4	9.622	143	52877	5.584	ng/u1	-0.03
28) 2,4-Dichlorophenol-d3	10.169	165	103172	5.529	ng/u1	-0.02
31) 4-Chloroaniline-d4	10.663	131	162287	5.978	ng/u1	-0.04
46) Dimethylphthalate-d6	13.804	166	404925	6.911	ng/u1	-0.04
49) Acenaphthylene-d8	14.098	160	418418	6.361	ng/u1	-0.02
54) 4-Nitrophenol-d4	14.628	143	53973m	4.842	ng/u1	0.00
60) Fluorene-d10	15.422	176	342870	6.961	ng/u1	-0.02
65) 4,6-Dinitro-2-methylph...	15.534	200	35911	3.778	ng/u1	-0.02
73) Anthracene-d10	17.328	188	566061	6.902	ng/u1	-0.01
81) Pyrene-d10	19.704	212	702518	6.956	ng/u1	0.01
92) Benzo(a)pyrene-d12	24.857	264	702898	6.854	ng/u1	0.02
Target Compounds						
					Qvalue	
10) Bis(2-Chloroethyl)ether	7.175	93	334019	13.905	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.552	70	87466	5.048	ng/u1	97
19) Hexachloroethane	8.828	117	98279	10.944	ng/u1	93
22) Nitrobenzene	8.940	77	308579	11.864	ng/u1	100
23) Isophorone	9.469	82	1325550	25.396	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.951	93	894561	27.329	ng/u1	99
30) Naphthalene	10.593	128	1394627	21.342	ng/u1	100
33) Hexachlorobutadiene	10.887	225	323906	25.747	ng/u1	99
36) 2-Methylnaphthalene	12.204	142	495192	11.204	ng/u1	100
40) Hexachlorocyclopentadiene	12.563	237	186688	13.482	ng/u1	98
44) 2-Chloronaphthalene	13.269	162	759070	16.712	ng/u1	99
47) Dimethylphthalate	13.857	163	529668	9.029	ng/u1	99
48) 2,6-Dinitrotoluene	13.981	165	31220m	2.886	ng/u1	
50) Acenaphthylene	14.128	152	511378	7.229	ng/u1	99
59) Diethylphthalate	15.251	149	606559	10.296	ng/u1	99
61) Fluorene	15.481	166	433349	7.784	ng/u1	100
62) 4-Chlorophenyl-phenyle...	15.481	204	895656	32.019	ng/u1	98
68) 4-Bromophenyl-phenylether	16.392	248	306936	16.090	ng/u1	96
69) Hexachlorobenzene	16.510	284	327439	14.458	ng/u1	95
72) Phenanthrene	17.263	178	706149	7.443	ng/u1	98
74) Anthracene	17.363	178	845669	8.799	ng/u1	99
78) Di-n-butylphthalate	18.245	149	1229375	11.684	ng/u1	99
87) Chrysene	21.733	228	852415	7.207	ng/u1	99
89) Di-n-octyl phthalate	22.916	149	627428	5.154	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090424\
 Data File : BP021817.D
 Acq On : 04 Sep 2024 13:38
 Operator : RC/JU
 Sample : P3808-01DL 5X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 X0B24DL

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 09/05/2024
 Supervised By :mohammad ahmed 09/07/2024

Quant Time: Sep 04 14:09:05 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 29 23:56:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
90) Benzo(b)fluoranthene	24.015	252	733249	6.093	ng/ul	98
91) Benzo(k)fluoranthene	24.080	252	1342587	11.396	ng/ul	99
93) Benzo(a)pyrene	24.933	252	792050	7.144	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	28.962	276	1291919	9.037	ng/ul	98
95) Dibenzo(a,h)anthracene	29.039	278	522776	4.551	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090424\
 Data File : BP021817.D
 Acq On : 04 Sep 2024 13:38
 Operator : RC/JU
 Sample : P3808-01DL 5X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 X0B24DL

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 09/05/2024
 Supervised By :mohammad ahmed 09/07/2024


Quant Time: Sep 04 14:09:05 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Aug 29 23:56:59 2024
 Response via : Initial Calibration





#10
Bis(2-Chloroethyl)ether
Concen: 13.905 ng/ul
RT: 7.175 min Scan# 716
Delta R.T. -0.035 min
Lab File: BP021817.D
Acq: 04 Sep 2024 13:38




Instrument :
BNA_P
ClientSampleId :
X0B24DL



Tgt Ion: 93 Resp: 334019
Ion Ratio Lower Upper
93 100
63 68.0 54.9 82.3
95 32.5 26.0 39.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 09/05/2024
Supervised By :mohammad ahmed 09/07/2024



#17
N-Nitroso-di-n-propylamine
Concen: 5.048 ng/ul
RT: 8.552 min Scan# 946
Delta R.T. -0.035 min
Lab File: BP021817.D
Acq: 04 Sep 2024 13:38


Tgt Ion: 70 Resp: 87466
Ion Ratio Lower Upper
70 100
42 42.5 35.8 53.6
101 9.4 8.2 12.4
130 18.1 15.1 22.7





#19
Hexachloroethane
Concen: 10.944 ng/ul
RT: 8.828 min Scan# 993
Delta R.T. -0.035 min
Lab File: BP021817.D
Acq: 04 Sep 2024 13:38




Instrument :
BNA_P
ClientSampleId :
X0B24DL




Tgt Ion:117 Resp: 98279
Ion Ratio Lower Upper
117 100
201 87.3 75.3 112.9
199 54.8 47.4 71.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 09/05/2024
Supervised By :mohammad ahmed 09/07/2024




#22
Nitrobenzene
Concen: 11.864 ng/ul
RT: 8.940 min Scan# 1012
Delta R.T. -0.035 min
Lab File: BP021817.D
Acq: 04 Sep 2024 13:38




Tgt Ion: 77 Resp: 308579
Ion Ratio Lower Upper
77 100
123 41.4 33.2 49.8
65 12.2 9.8 14.6





#23
Isophorone
Concen: 25.396 ng/ul
RT: 9.469 min Scan# 1106
Delta R.T. -0.035 min
Lab File: BP021817.D
Acq: 04 Sep 2024 13:38




Instrument :
BNA_P
ClientSampleId :
X0B24DL



Tgt Ion: 82 Resp: 1325550
Ion Ratio Lower Upper
82 100
95 7.5 5.8 8.8
138 15.2 11.8 17.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 09/05/2024
Supervised By :mohammad ahmed 09/07/2024



#27
Bis(2-Chloroethoxy)methane
Concen: 27.329 ng/ul
RT: 9.951 min Scan# 1184
Delta R.T. -0.030 min
Lab File: BP021817.D
Acq: 04 Sep 2024 13:38

Tgt Ion: 93 Resp: 894561
Ion Ratio Lower Upper
93 100
95 32.3 25.3 37.9
123 9.8 7.4 11.0





#30

Naphthalene

Concen: 21.342 ng/ul

RT: 10.593 min Scan# 1295

Delta R.T. -0.030 min

Lab File: BP021817.D

Acq: 04 Sep 2024 13:38

Instrument :

BNA_P

ClientSampleId :

X0B24DL

Tgt Ion:128 Resp: 139462

Ion Ratio Lower Upper

128 100





129 11.0 8.6 13.0

127 13.2 10.5 15.7

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 09/05/2024
Supervised By :mohammad ahmed 09/07/2024



#33

Hexachlorobutadiene

Concen: 25.747 ng/ul

RT: 10.887 min Scan# 1343

Delta R.T. -0.030 min

Lab File: BP021817.D

Acq: 04 Sep 2024 13:38




Tgt Ion:225 Resp: 323906


Ion Ratio Lower Upper

225 100

223 62.8 51.5 77.3

227 63.5 51.3 76.9









#36
2-Methylnaphthalene
Concen: 11.204 ng/ul
RT: 12.204 min Scan# 1567
Delta R.T. -0.035 min
Lab File: BP021817.D
Acq: 04 Sep 2024 13:38

Instrument :
BNA_P
ClientSampleId :
X0B24DL

Tgt Ion:142 Resp: 49519
Ion Ratio Lower Upper
142 100
141 89.4 71.1 106.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 09/05/2024
Supervised By :mohammad ahmed 09/07/2024



#40
Hexachlorocyclopentadiene
Concen: 13.482 ng/ul
RT: 12.563 min Scan# 1628
Delta R.T. -0.030 min
Lab File: BP021817.D
Acq: 04 Sep 2024 13:38


Tgt Ion:237 Resp: 186688
Ion Ratio Lower Upper
237 100
235 62.4 50.7 76.1
272 12.3 11.0 16.6





#44
2-Chloronaphthalene
Concen: 16.712 ng/ul
RT: 13.269 min Scan# 11
Delta R.T. -0.024 min
Lab File: BP021817.D
Acq: 04 Sep 2024 13:38




Instrument :
BNA_P
ClientSampleId :
X0B24DL




Tgt Ion:162 Resp: 759070
Ion Ratio Lower Upper
162 100
164 32.7 26.0 39.0
127 42.8 33.1 49.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 09/05/2024
Supervised By :mohammad ahmed 09/07/2024




#47
Dimethylphthalate
Concen: 9.029 ng/ul
RT: 13.857 min Scan# 1848
Delta R.T. -0.030 min
Lab File: BP021817.D
Acq: 04 Sep 2024 13:38



Tgt Ion:163 Resp: 529668
Ion Ratio Lower Upper
163 100
194 3.7 3.5 5.3
164 10.1 7.8 11.8









#48
2,6-Dinitrotoluene
Concen: 2.886 ng/ul m
RT: 13.981 min Scan# 1869
Delta R.T. -0.018 min
Lab File: BP021817.D
Acq: 04 Sep 2024 13:38

Instrument :
BNA_P
ClientSampleId :
X0B24DL

Tgt Ion:165 Resp: 31220
Ion Ratio Lower Upper
165 100
89 75.4 55.7 83.5
121 29.5 18.6 28.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 09/05/2024
Supervised By :mohammad ahmed 09/07/2024



#50
Acenaphthylene
Concen: 7.229 ng/ul
RT: 14.128 min Scan# 1894
Delta R.T. -0.030 min
Lab File: BP021817.D
Acq: 04 Sep 2024 13:38

Tgt Ion:152 Resp: 511378
Ion Ratio Lower Upper
152 100
151 19.7 15.6 23.4
153 13.7 10.2 15.4









#59
Diethylphthalate
Concen: 10.296 ng/ul
RT: 15.251 min Scan# 2085
Delta R.T. -0.024 min
Lab File: BP021817.D
Acq: 04 Sep 2024 13:38

Instrument :
BNA_P
ClientSampleId :
X0B24DL

Tgt Ion:149 Resp: 606559
Ion Ratio Lower Upper
149 100
177 21.7 18.0 27.0
150 12.7 9.8 14.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 09/05/2024
Supervised By :mohammad ahmed 09/07/2024



#61
Fluorene
Concen: 7.784 ng/ul
RT: 15.481 min Scan# 2124
Delta R.T. -0.018 min
Lab File: BP021817.D
Acq: 04 Sep 2024 13:38

Tgt Ion:166 Resp: 433349
Ion Ratio Lower Upper
166 100
165 97.3 78.2 117.2
167 13.7 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 32.019 ng/ul
RT: 15.481 min Scan# 2125
Delta R.T. -0.018 min
Lab File: BP021817.D
Acq: 04 Sep 2024 13:38

Instrument :
BNA_P
ClientSampleId :
X0B24DL

Tgt Ion:204 Resp: 895650
Ion Ratio Lower Upper
204 100
206 33.2 26.2 39.4
141 63.3 48.6 73.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 09/05/2024
Supervised By :mohammad ahmed 09/07/2024



#68
4-Bromophenyl-phenylether
Concen: 16.090 ng/ul
RT: 16.392 min Scan# 2279
Delta R.T. -0.012 min
Lab File: BP021817.D
Acq: 04 Sep 2024 13:38

Tgt Ion:248 Resp: 306936
Ion Ratio Lower Upper
248 100
250 96.8 79.6 119.4
141 76.3 57.0 85.4





#69

Hexachlorobenzene

Concen: 14.458 ng/ul

RT: 16.510 min Scan# 21

Delta R.T. -0.012 min

Lab File: BP021817.D

Acq: 04 Sep 2024 13:38

Instrument :

BNA_P

ClientSampleId :

X0B24DL

Tgt Ion:284 Resp: 327439

Ion Ratio Lower Upper

284 100





142 36.3 26.5 39.7

249 32.4 24.2 36.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 09/05/2024
Supervised By :mohammad ahmed 09/07/2024



#72

Phenanthrene

Concen: 7.443 ng/ul

RT: 17.263 min Scan# 2427

Delta R.T. -0.012 min

Lab File: BP021817.D

Acq: 04 Sep 2024 13:38




Tgt Ion:178 Resp: 706149


Ion Ratio Lower Upper

178 100

179 15.9 12.0 18.0

176 19.5 14.8 22.2





#74

Anthracene

Concen: 8.799 ng/ul

RT: 17.363 min Scan# 2443

Delta R.T. -0.006 min

Lab File: BP021817.D

Acq: 04 Sep 2024 13:38

Instrument :

BNA_P

ClientSampleId :





X0B24DL

Tgt Ion:178	Resp: 845669
Ion Ratio	Lower Upper
178 100	
179 15.7	12.3 18.5
176 18.3	14.6 21.8

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 09/05/2024
Supervised By :mohammad ahmed 09/07/2024



#78

Di-n-butylphthalate

Concen: 11.684 ng/ul




RT: 18.245 min Scan# 2594


Delta R.T. 0.006 min

Lab File: BP021817.D

Acq: 04 Sep 2024 13:38

Tgt Ion:149	Resp: 1229375
Ion Ratio	Lower Upper
149 100	
150 8.9	7.4 11.0
104 6.1	4.8 7.2





#87

Chrysene

Concen: 7.207 ng/ul

RT: 21.733 min Scan# 3186

Delta R.T. 0.018 min

Lab File: BP021817.D

Acq: 04 Sep 2024 13:38

Instrument :

BNA_P

ClientSampleId :





X0B24DL

Tgt Ion:228 Resp: 852415
 Ion Ratio Lower Upper
 228 100
 226 29.8 23.5 35.3
 229 19.3 15.6 23.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 09/05/2024
 Supervised By :mohammad ahmed 09/07/2024



#89

Di-n-octyl phthalate

Concen: 5.154 ng/ul




RT: 22.916 min Scan# 3388


Delta R.T. 0.023 min

Lab File: BP021817.D

Acq: 04 Sep 2024 13:38

Tgt Ion:149 Resp: 627428









#90
Benzo(b)fluoranthene
Concen: 6.093 ng/u1
RT: 24.015 min Scan# 3575
Delta R.T. 0.023 min
Lab File: BP021817.D
Acq: 04 Sep 2024 13:38

Instrument :
BNA_P
ClientSampleId :
X0B24DL

Tgt Ion:252 Resp: 733249
Ion Ratio Lower Upper
252 100
253 21.7 18.1 27.1
125 10.1 8.6 12.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 09/05/2024
Supervised By :mohammad ahmed 09/07/2024



#91
Benzo(k)fluoranthene
Concen: 11.396 ng/u1
RT: 24.080 min Scan# 3586
Delta R.T. 0.023 min
Lab File: BP021817.D
Acq: 04 Sep 2024 13:38

Tgt Ion:252 Resp: 1342587
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.8
125 11.0 8.6 13.0





#93

Benzo(a)pyrene

Concen: 7.144 ng/ul

RT: 24.933 min Scan# 31

Delta R.T. 0.023 min

Lab File: BP021817.D

Acq: 04 Sep 2024 13:38

Instrument :

BNA_P

ClientSampleId :





X0B24DL

Tgt Ion:	252	Resp:	792050
Ion Ratio	Lower	Upper	
252	100		
253	21.4	17.4	26.0
125	12.3	9.6	14.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 09/05/2024
Supervised By :mohammad ahmed 09/07/2024



#94

Indeno(1,2,3-cd)pyrene

Concen: 9.037 ng/ul




RT: 28.962 min Scan# 4416


Delta R.T. 0.035 min

Lab File: BP021817.D

Acq: 04 Sep 2024 13:38


Tgt Ion:	276	Resp:	1291919
Ion Ratio	Lower	Upper	
276	100		
138	20.4	14.9	22.3
277	23.3	19.0	28.6





#95
 Dibenzo(a,h)anthracene
 Concen: 4.551 ng/ul
 RT: 29.039 min Scan# 44
 Delta R.T. 0.023 min
 Lab File: BP021817.D
 Acq: 04 Sep 2024 13:38

Instrument :
 BNA_P
 ClientSampleId :
 X0B24DL



Tgt Ion:	278	Resp:	522770
Ion Ratio	Lower	Upper	
278	100		
139	15.5	12.7	19.1
279	23.4	18.9	28.3

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 09/05/2024
 Supervised By :mohammad ahmed 09/07/2024

