

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090424\
 Data File : BP021825.D
 Acq On : 04 Sep 2024 20:40
 Operator : RC/JU
 Sample : P3809-09MS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 YE3E3MS

Quant Time: Sep 05 05:47:53 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 05 05:03:57 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Jagrut
 Upadhyay

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.775	152	326350	20.000	ng/u1	0.01
20) Naphthalene-d8	10.557	136	1339723	20.000	ng/u1	0.01
38) Acenaphthene-d10	14.410	164	856668	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.222	188	1886367	20.000	ng/u1	0.01
79) Chrysene-d12	21.680	240	1877637	20.000	ng/u1	0.00
88) Perylene-d12	25.092	264	2123628	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.270	96	36949	3.759	ng/uL	0.00
4) Pyridine-d5	3.687	84	240346	8.453	ng/u1	0.01
7) Phenol-d5	6.946	99	209357	6.011	ng/u1	0.01
9) Bis-(2-Chloroethyl)eth...	7.105	67	444804	23.468	ng/u1	0.01
11) 2-Chlorophenol-d4	7.311	132	484206	21.534	ng/u1	0.00
15) 4-Methylphenol-d8	8.475	113	385230	14.314	ng/u1	0.01
21) Nitrobenzene-d5	8.922	128	267451	24.813	ng/u1	0.01
24) 2-Nitrophenol-d4	9.640	143	282494	25.820	ng/u1	0.01
28) 2,4-Dichlorophenol-d3	10.181	165	534990	24.814	ng/u1	0.01
31) 4-Chloroaniline-d4	10.687	131	662658	21.126	ng/u1	0.00
46) Dimethylphthalate-d6	13.816	166	1854674	28.294	ng/u1	0.00
49) Acenaphthylene-d8	14.098	160	1969490	26.761	ng/u1	0.00
54) 4-Nitrophenol-d4	14.622	143	93682	7.512	ng/u1	0.01
60) Fluorene-d10	15.422	176	1548479	28.098	ng/u1	0.01
65) 4,6-Dinitro-2-methylph...	15.551	200	302648	29.170	ng/u1	0.02
73) Anthracene-d10	17.322	188	2480772	27.710	ng/u1	0.00
81) Pyrene-d10	19.698	212	3183305	29.658	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.857	264	3380841	29.856	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.305	88	120582	11.542	ng/uL	96
5) Pyridine	3.705	79	244916	8.620	ng/u1	96
6) Benzaldehyde	6.922	77	524349	29.434	ng/u1	99
8) Phenol	6.969	94	229660	6.324	ng/u1	95
10) Bis(2-Chloroethyl)ether	7.199	93	639971	22.671	ng/u1	99
12) 2-Chlorophenol	7.346	128	503863	21.350	ng/u1	100
13) 2-Methylphenol	8.216	108	424882	16.218	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.287	45	613143	22.504	ng/u1	99
16) Acetophenone	8.587	105	1004565	23.283	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.569	70	461319	22.657	ng/u1	96
18) 4-Methylphenol	8.534	108	415360	14.482	ng/u1	95
19) Hexachloroethane	8.846	117	182041	17.251	ng/u1	92
22) Nitrobenzene	8.963	77	730704	24.312	ng/u1	99
23) Isophorone	9.493	82	1381603	22.908	ng/u1	99
25) 2-Nitrophenol	9.675	139	302607	24.963	ng/u1	97
26) 2,4-Dimethylphenol	9.734	107	583036	19.596	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.963	93	864863	22.866	ng/u1	98
29) 2,4-Dichlorophenol	10.210	162	522837	24.132	ng/u1	99
30) Naphthalene	10.604	128	1662915	22.023	ng/u1	100
32) 4-Chloroaniline	10.710	127	646496	20.415	ng/u1	99
33) Hexachlorobutadiene	10.899	225	280677	19.308	ng/u1	98
34) Caprolactam	11.499	113	37188m	4.059	ng/u1	
35) 4-Chloro-3-methylphenol	11.840	107	704041	24.552	ng/u1	98
36) 2-Methylnaphthalene	12.222	142	1175333	23.013	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090424\
 Data File : BP021825.D
 Acq On : 04 Sep 2024 20:40
 Operator : RC/JU
 Sample : P3809-09MS
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 YE3E3MS


Quant Time: Sep 05 05:47:53 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP082324.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 05 05:03:57 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Jagrut
 Upadhyay


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
37) 1-Methylnaphthalene	12.434	142	1174310	22.860	ng/ul	99	09/05/2024 Supervised By :mohammad ahmed
39) 1,2,4,5-Tetrachloroben...	12.587	216	636268	23.565	ng/ul	99	
40) Hexachlorocyclopentadiene	12.569	237	251664	16.244	ng/ul	97	
41) 2,4,6-Trichlorophenol	12.834	196	471382	27.116	ng/ul	99	
42) 2,4,5-Trichlorophenol	12.904	196	514422	27.405	ng/ul	98	
43) 1,1'-Biphenyl	13.234	154	1620762	25.007	ng/ul	99	09/07/2024
44) 2-Chloronaphthalene	13.275	162	1235858	24.319	ng/ul	98	
45) 2-Nitroaniline	13.475	65	478644	30.065	ng/ul	96	
47) Dimethylphthalate	13.869	163	1777499	27.081	ng/ul	99	
48) 2,6-Dinitrotoluene	13.975	165	363766	30.051	ng/ul	95	
50) Acenaphthylene	14.134	152	2052103	25.926	ng/ul	99	
51) 3-Nitroaniline	14.310	138	381290	28.825	ng/ul	94	
52) Acenaphthene	14.481	153	1457819	26.191	ng/ul	99	
53) 2,4-Dinitrophenol	14.528	184	194523	26.842	ng/ul	93	
55) 4-Nitrophenol	14.634	109	95576	8.370	ng/ul	95	
56) Dibenzofuran	14.816	168	2052008	26.647	ng/ul	98	
57) 2,4-Dinitrotoluene	14.781	165	547667	30.108	ng/ul	96	
58) 2,3,4,6-Tetrachlorophenol	15.045	232	458025	28.321	ng/ul	91	
59) Diethylphthalate	15.257	149	1865619	28.305	ng/ul	99	
61) Fluorene	15.487	166	1686049	27.070	ng/ul	100	
62) 4-Chlorophenyl-phenyle...	15.481	204	827052	26.426	ng/ul	98	
63) 4-Nitroaniline	15.498	138	425598	33.290	ng/ul	93	
66) 4,6-Dinitro-2-methylph...	15.563	198	328382	28.295	ng/ul	100	
67) N-Nitrosodiphenylamine	15.692	169	1422411	25.633	ng/ul	98	
68) 4-Bromophenyl-phenylether	16.387	248	549897	26.406	ng/ul	95	
69) Hexachlorobenzene	16.510	284	648536	26.233	ng/ul	94	
70) Atrazine	16.675	200	584452	27.765	ng/ul	98	
71) Pentachlorophenol	16.863	266	425649	26.902	ng/ul	98	
72) Phenanthrene	17.269	178	2927411	28.267	ng/ul	99	
74) Anthracene	17.357	178	2871237	27.367	ng/ul	100	
75) 1,2,3,4-Tetrachloroben...	13.198	216	638671	22.773	ng/ul	97	
76) Pentachlorobenzene	14.734	250	703019	24.185	ng/ul	100	
77) Carbazole	17.639	167	2831097	30.002	ng/ul	99	
78) Di-n-butylphthalate	18.233	149	3419676	29.772	ng/ul	100	
80) Fluoranthene	19.351	202	3614172	28.855	ng/ul	100	
82) Pyrene	19.728	202	3803164	28.670	ng/ul	98	
83) Butylbenzylphthalate	20.675	149	1675786	30.854	ng/ul	98	
84) 3,3'-Dichlorobenzidine	21.574	252	624835	13.205	ng/ul	99	
85) Benzo(a)anthracene	21.657	228	3797226	28.430	ng/ul	99	
86) Bis(2-ethylhexyl)phtha...	21.598	149	2382616	29.722	ng/ul	100	
87) Chrysene	21.727	228	3573207	28.426	ng/ul	100	
89) Di-n-octyl phthalate	22.898	149	3982393	29.626	ng/ul	100	
90) Benzo(b)fluoranthene	23.998	252	3922704	29.521	ng/ul	99	
91) Benzo(k)fluoranthene	24.080	252	3769136	28.974	ng/ul	99	
93) Benzo(a)pyrene	24.927	252	3526662	28.808	ng/ul	99	
94) Indeno(1,2,3-cd)pyrene	28.945	276	4397749m	27.861	ng/ul		
95) Dibenzo(a,h)anthracene	29.051	278	3522145	27.769	ng/ul	99	
96) Benzo(g,h,i)perylene	30.156	276	3459908m	28.370	ng/ul		

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 11.542 ng/uL
RT: 3.305 min Scan# 54
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40



Instrument :
BNA_P
ClientSampleId :
YE3E3MS



Tgt Ion: 88 Resp: 12058
Ion Ratio Lower Upper
88 100
43 21.0 15.4 23.2
58 62.9 48.2 72.2


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






09/05/2024
Supervised By :mohammad
ahmed


09/07/2024



#5
Pyridine
Concen: 8.620 ng/uL
RT: 3.705 min Scan# 122
Delta R.T. 0.012 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40

Tgt Ion: 79 Resp: 244916
Ion Ratio Lower Upper
79 100
52 61.6 46.8 70.2
51 29.3 21.7 32.5







#6
Benzaldehyde
Concen: 29.434 ng/ul
RT: 6.922 min Scan# 669
Delta R.T. 0.012 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40

Instrument :
BNA_P
ClientSampleId :
YE3E3MS

Tgt Ion: 77 Resp: 524349
Ion Ratio Lower Upper
77 100
105 101.7 80.2 120.4
106 96.6 76.2 114.4



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






09/05/2024
Supervised By :mohammad
ahmed


09/07/2024



#8
Phenol
Concen: 6.324 ng/ul
RT: 6.969 min Scan# 677
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40


Tgt Ion: 94 Resp: 229660
Ion Ratio Lower Upper
94 100
65 27.3 20.0 30.0
66 36.7 27.0 40.4





#10
Bis(2-Chloroethyl)ether
Concen: 22.671 ng/ul
RT: 7.199 min Scan# 714
Delta R.T. 0.012 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40



Instrument :
BNA_P
ClientSampleId :
YE3E3MS



Tgt Ion: 93 Resp: 63997
Ion Ratio Lower Upper
93 100
63 68.2 54.9 82.3
95 31.5 26.0 39.0


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






09/05/2024
Supervised By :mohammad
ahmed


09/07/2024



#12
2-Chlorophenol
Concen: 21.350 ng/ul
RT: 7.346 min Scan# 741
Delta R.T. 0.012 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40

Tgt Ion:128 Resp: 503863
Ion Ratio Lower Upper
128 100
64 49.3 39.5 59.3
130 32.6 25.9 38.9







#13
2-Methylphenol
Concen: 16.218 ng/ul
RT: 8.216 min Scan# 81
Delta R.T. 0.012 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40

Instrument :
BNA_P
ClientSampleId :
YE3E3MS

Tgt Ion:108 Resp: 42488
Ion Ratio Lower Upper
108 100
107 91.3 73.2 109.8



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






09/05/2024
Supervised By :mohammad
ahmed


09/07/2024



#14
2,2'-oxybis(1-Chloropropane)
Concen: 22.504 ng/ul
RT: 8.287 min Scan# 901
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40


Tgt Ion: 45 Resp: 613143
Ion Ratio Lower Upper
45 100
77 20.5 17.0 25.4
79 15.1 12.0 18.0





#16
Acetophenone
Concen: 23.283 ng/ul
RT: 8.587 min Scan# 91
Delta R.T. 0.012 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40



Instrument :
BNA_P
ClientSampleId :
YE3E3MS



Tgt Ion: 105 Resp: 100456
Ion Ratio Lower Upper
105 100
77 75.9 59.4 89.2
51 21.4 16.4 24.6


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






09/05/2024
Supervised By :mohammad
ahmed


09/07/2024



#17
N-Nitroso-di-n-propylamine
Concen: 22.657 ng/ul
RT: 8.569 min Scan# 949
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40

Tgt Ion: 70 Resp: 461319
Ion Ratio Lower Upper
70 100
42 48.3 35.8 53.6
101 11.0 8.2 12.4
130 18.8 15.1 22.7







#18
4-Methylphenol
Concen: 14.482 ng/ul
RT: 8.534 min Scan# 942
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40

Instrument :
BNA_P
ClientSampleId :
YE3E3MS

Tgt Ion:108 Resp: 415360
Ion Ratio Lower Upper
108 100
107 118.4 90.1 135.1



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






09/05/2024
Supervised By :mohammad
ahmed


09/07/2024



#19
Hexachloroethane
Concen: 17.251 ng/ul
RT: 8.846 min Scan# 996
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40


Tgt Ion:117 Resp: 182041
Ion Ratio Lower Upper
117 100
201 85.2 75.3 112.9
199 55.4 47.4 71.0





#22
Nitrobenzene
Concen: 24.312 ng/ul
RT: 8.963 min Scan# 1106
Delta R.T. 0.012 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40



Instrument :
BNA_P
ClientSampleId :
YE3E3MS



Tgt Ion: 77 Resp: 730704
Ion Ratio Lower Upper
77 100
123 41.8 33.2 49.8
65 12.9 9.8 14.6


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




09/05/2024
Supervised By :mohammad
ahmed



09/07/2024




#23
Isophorone
Concen: 22.908 ng/ul
RT: 9.493 min Scan# 1106
Delta R.T. 0.012 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40



Tgt Ion: 82 Resp: 1381603
Ion Ratio Lower Upper
82 100
95 7.6 5.8 8.8
138 15.3 11.8 17.8







#25
2-Nitrophenol
Concen: 24.963 ng/ul
RT: 9.675 min Scan# 1134
Delta R.T. 0.018 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40

Instrument :
BNA_P
ClientSampleId :
YE3E3MS

Tgt Ion:139 Resp: 30260
Ion Ratio Lower Upper
139 100
65 58.7 46.2 69.2
109 37.9 28.0 42.0



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






09/05/2024
Supervised By :mohammad
ahmed


09/07/2024



#26
2,4-Dimethylphenol
Concen: 19.596 ng/ul
RT: 9.734 min Scan# 1147
Delta R.T. 0.012 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40


Tgt Ion:107 Resp: 583036
Ion Ratio Lower Upper
107 100
121 46.8 36.9 55.3
122 75.8 61.0 91.4





#27
Bis(2-Chloroethoxy)methane
Concen: 22.866 ng/ul
RT: 9.963 min Scan# 1186
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40


Instrument :
BNA_P
ClientSampleId :
YE3E3MS



Tgt Ion: 93 Resp: 86486
Ion Ratio Lower Upper
93 100
95 32.6 25.3 37.9
123 9.9 7.4 11.0


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






09/05/2024
Supervised By :mohammad
ahmed


09/07/2024



#29
2,4-Dichlorophenol
Concen: 24.132 ng/ul
RT: 10.210 min Scan# 1228
Delta R.T. 0.012 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40

Tgt Ion:162 Resp: 522837
Ion Ratio Lower Upper
162 100
164 65.2 52.3 78.5
98 46.0 35.3 52.9







#30
Naphthalene
Concen: 22.023 ng/ul
RT: 10.604 min Scan# 1295
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40

Instrument :
BNA_P
ClientSampleId :
YE3E3MS

Tgt Ion:128 Resp: 1662915
Ion Ratio Lower Upper
128 100
129 10.9 8.6 13.0
127 13.0 10.5 15.7



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






09/05/2024
Supervised By :mohammad
ahmed


09/07/2024



#32
4-Chloroaniline
Concen: 20.415 ng/ul
RT: 10.710 min Scan# 1313
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40


Tgt Ion:127 Resp: 646496
Ion Ratio Lower Upper
127 100
129 33.0 25.8 38.8





#33
Hexachlorobutadiene
Concen: 19.308 ng/ul
RT: 10.899 min Scan# 11
Delta R.T. 0.012 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40



Instrument :
BNA_P
ClientSampleId :
YE3E3MS



Tgt Ion:225 Resp: 28067
Ion Ratio Lower Upper
225 100
223 62.0 51.5 77.3
227 63.6 51.3 76.9


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






09/05/2024
Supervised By :mohammad
ahmed


09/07/2024



#34
Caprolactam
Concen: 4.059 ng/ul m
RT: 11.499 min Scan# 1447
Delta R.T. 0.018 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40


Tgt Ion:113 Resp: 37188
Ion Ratio Lower Upper
113 100
55 142.8 110.2 165.2
56 102.7 84.9 127.3





#35
4-Chloro-3-methylphenol
Concen: 24.552 ng/ul
RT: 11.840 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40



Instrument :
BNA_P
ClientSampleId :
YE3E3MS



Tgt Ion:107 Resp: 70404
Ion Ratio Lower Upper
107 100
144 20.8 17.5 26.3
142 65.3 53.3 79.9


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






09/05/2024
Supervised By :mohammad
ahmed


09/07/2024



#36
2-Methylnaphthalene
Concen: 23.013 ng/ul
RT: 12.222 min Scan# 1570
Delta R.T. 0.012 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40

Tgt Ion:142 Resp: 1175333
Ion Ratio Lower Upper
142 100
141 89.6 71.1 106.7






#37
1-Methylnaphthalene
Concen: 22.860 ng/ul
RT: 12.434 min Scan# 1606
Delta R.T. -0.000 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40

Instrument :

BNA_P

ClientSampleId :



YE3E3MS



Tgt Ion:142 Resp: 1174310
Ion Ratio Lower Upper
142 100
141 94.1 74.6 111.8
116 4.7 3.5 5.3


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






09/05/2024
Supervised By :mohammad
ahmed


09/07/2024



#39
1,2,4,5-Tetrachlorobenzene
Concen: 23.565 ng/ul
RT: 12.587 min Scan# 1632
Delta R.T. -0.000 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40


Tgt Ion:216 Resp: 636268
Ion Ratio Lower Upper
216 100
214 79.4 63.2 94.8
179 21.8 16.6 25.0
108 21.6 16.2 24.4





#40
Hexachlorocyclopentadiene
Concen: 16.244 ng/ul
RT: 12.569 min Scan# 1629
Delta R.T. -0.000 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40



Instrument :
BNA_P
ClientSampleId :
YE3E3MS



Tgt Ion: 237 Resp: 251664
Ion Ratio Lower Upper
237 100
235 60.8 50.7 76.1
272 12.2 11.0 16.6


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




09/05/2024
Supervised By :mohammad
ahmed



09/07/2024




#41
2,4,6-Trichlorophenol
Concen: 27.116 ng/ul
RT: 12.834 min Scan# 1674
Delta R.T. 0.012 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40




Tgt Ion: 196 Resp: 471382
Ion Ratio Lower Upper
196 100
198 96.2 76.5 114.7
200 31.3 25.3 37.9





#42
2,4,5-Trichlorophenol
Concen: 27.405 ng/ul
RT: 12.904 min Scan# 1685
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40



Instrument :
BNA_P
ClientSampleId :
YE3E3MS



Tgt Ion:196 Resp: 51442
Ion Ratio Lower Upper
196 100
198 95.4 78.0 117.0
200 31.2 24.3 36.5


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




09/05/2024
Supervised By :mohammad
ahmed



09/07/2024




#43
1,1'-Biphenyl
Concen: 25.007 ng/ul
RT: 13.234 min Scan# 1742
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40




Tgt Ion:154 Resp: 1620762
Ion Ratio Lower Upper
154 100
153 39.7 31.6 47.4
76 14.9 12.6 18.8





#44
2-Chloronaphthalene
Concen: 24.319 ng/ul
RT: 13.275 min Scan# 1748
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40



Instrument :
BNA_P
ClientSampleId :
YE3E3MS



Tgt Ion:162 Resp: 123585
Ion Ratio Lower Upper
162 100
164 32.4 26.0 39.0
127 43.4 33.1 49.7


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




09/05/2024
Supervised By :mohammad
ahmed



09/07/2024




#45
2-Nitroaniline
Concen: 30.065 ng/ul
RT: 13.475 min Scan# 1783
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40




Tgt Ion: 65 Resp: 478644
Ion Ratio Lower Upper
65 100
92 76.1 65.4 98.2
138 94.1 77.5 116.3





#47
Dimethylphthalate
Concen: 27.081 ng/ul
RT: 13.869 min Scan# 1848
Delta R.T. 0.012 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40



Instrument :
BNA_P
ClientSampleId :
YE3E3MS



Tgt Ion:163 Resp: 1777499
Ion Ratio Lower Upper
163 100
194 4.0 3.5 5.3
164 10.2 7.8 11.8


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






09/05/2024
Supervised By :mohammad
ahmed


09/07/2024



#48
2,6-Dinitrotoluene
Concen: 30.051 ng/ul
RT: 13.975 min Scan# 1868
Delta R.T. -0.000 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40


Tgt Ion:165 Resp: 363766
Ion Ratio Lower Upper
165 100
89 74.7 55.7 83.5
121 23.9 18.6 28.0





#50
Acenaphthylene
Concen: 25.926 ng/ul
RT: 14.134 min Scan# 1895
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40



Instrument :
BNA_P
ClientSampleId :
YE3E3MS



Tgt Ion:152 Resp: 205210
Ion Ratio Lower Upper
152 100
151 19.1 15.6 23.4
153 13.0 10.2 15.4


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




09/05/2024
Supervised By :mohammad
ahmed



09/07/2024




#51
3-Nitroaniline
Concen: 28.825 ng/ul
RT: 14.310 min Scan# 1925
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40



Tgt Ion:138 Resp: 381290
Ion Ratio Lower Upper
138 100
108 14.2 12.0 18.0
92 149.9 114.1 171.1







#52
 Acenaphthene
 Concen: 26.191 ng/ul
 RT: 14.481 min Scan# 1954
 Delta R.T. 0.012 min
 Lab File: BP021825.D
 Acq: 04 Sep 2024 20:40

Instrument : BNA_P
 ClientSampleId : YE3E3MS

Tgt Ion:153 Resp: 1457819



Ion	Ratio	Lower	Upper
153	100		
152	46.7	38.0	57.0
154	88.1	70.2	105.4

Manual Integrations
APPROVED
 Reviewed By :Jagrut Upadhyay



09/05/2024
 Supervised By :mohammad ahmed




09/07/2024




#53
 2,4-Dinitrophenol
 Concen: 26.842 ng/ul
 RT: 14.528 min Scan# 1962
 Delta R.T. 0.012 min
 Lab File: BP021825.D
 Acq: 04 Sep 2024 20:40

Tgt Ion:184 Resp: 194523

Ion	Ratio	Lower	Upper
184	100		
63	77.1	55.5	83.3
154	68.3	51.3	76.9







#55
4-Nitrophenol
Concen: 8.370 ng/ul
RT: 14.634 min Scan# 1979
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40

Instrument :
BNA_P
ClientSampleId :
YE3E3MS

Tgt Ion:109 Resp: 95570
Ion Ratio Lower Upper
109 100
139 109.4 96.2 144.2
65 122.1 98.5 147.7



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






09/05/2024
Supervised By :mohammad
ahmed


09/07/2024



#56
Dibenzofuran
Concen: 26.647 ng/ul
RT: 14.816 min Scan# 2011
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40

Tgt Ion:168 Resp: 2052008
Ion Ratio Lower Upper
168 100
139 38.7 29.8 44.8







#57
2,4-Dinitrotoluene
Concen: 30.108 ng/ul
RT: 14.781 min Scan# 2005
Delta R.T. 0.012 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40

Instrument :
BNA_P
ClientSampleId :
YE3E3MS

Tgt Ion:165 Resp: 54766
Ion Ratio Lower Upper
165 100
63 49.9 37.8 56.8
182 2.9 2.2 3.4



Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






09/05/2024
Supervised By :mohammad
ahmed


09/07/2024



#58
2,3,4,6-Tetrachlorophenol
Concen: 28.321 ng/ul
RT: 15.045 min Scan# 2050
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40


Tgt Ion:232 Resp: 458025
Ion Ratio Lower Upper
232 100
131 50.1 33.9 50.9
130 2.6 2.0 3.0
166 33.6 24.3 36.5





#59
Diethylphthalate
Concen: 28.305 ng/ul
RT: 15.257 min Scan# 2086
Delta R.T. 0.012 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40



Instrument :
BNA_P
ClientSampleId :
YE3E3MS



Tgt Ion:149 Resp: 1865619
Ion Ratio Lower Upper
149 100
177 21.9 18.0 27.0
150 12.9 9.8 14.8


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






09/05/2024
Supervised By :mohammad
ahmed


09/07/2024



#61
Fluorene
Concen: 27.070 ng/ul
RT: 15.487 min Scan# 2125
Delta R.T. 0.018 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40


Tgt Ion:166 Resp: 1686049
Ion Ratio Lower Upper
166 100
165 98.1 78.2 117.2
167 13.2 10.8 16.2





#62
4-Chlorophenyl-phenylether
Concen: 26.426 ng/ul
RT: 15.481 min Scan# 2124
Delta R.T. 0.012 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40



Instrument :
BNA_P
ClientSampleId :
YE3E3MS



Tgt Ion:204 Resp: 827052
Ion Ratio Lower Upper
204 100
206 32.6 26.2 39.4
141 62.5 48.6 73.0


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay




09/05/2024
Supervised By :mohammad
ahmed



09/07/2024




#63
4-Nitroaniline
Concen: 33.290 ng/ul
RT: 15.498 min Scan# 2127
Delta R.T. 0.012 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40




Tgt Ion:138 Resp: 425598
Ion Ratio Lower Upper
138 100
92 65.7 50.7 76.1
108 99.3 72.2 108.2





#66
4,6-Dinitro-2-methylphenol
Concen: 28.295 ng/ul
RT: 15.563 min Scan# 2136
Delta R.T. 0.012 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40



Instrument :
BNA_P
ClientSampleId :
YE3E3MS



Tgt Ion:198 Resp: 328381
Ion Ratio Lower Upper
198 100
182 4.0 3.3 4.9
77 28.9 23.3 34.9


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






09/05/2024
Supervised By :mohammad
ahmed


09/07/2024



#67
N-Nitrosodiphenylamine
Concen: 25.633 ng/ul
RT: 15.692 min Scan# 2160
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40


Tgt Ion:169 Resp: 1422411
Ion Ratio Lower Upper
169 100
168 67.5 52.6 79.0
167 36.4 28.2 42.4





#68
4-Bromophenyl-phenylether
Concen: 26.406 ng/ul
RT: 16.387 min Scan# 2277
Delta R.T. 0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40



Instrument :
BNA_P
ClientSampleId :
YE3E3MS



Tgt Ion:248 Resp: 54989
Ion Ratio Lower Upper
248 100
250 96.5 79.6 119.4
141 77.1 57.0 85.4


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






09/05/2024
Supervised By :mohammad
ahmed


09/07/2024



#69
Hexachlorobenzene
Concen: 26.233 ng/ul
RT: 16.510 min Scan# 2299
Delta R.T. 0.012 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40

Tgt Ion:284 Resp: 648536
Ion Ratio Lower Upper
284 100
142 37.7 26.5 39.7
249 32.0 24.2 36.4





#70

Atrazine

Concen: 27.765 ng/ul

RT: 16.675 min Scan# 21

Delta R.T. 0.006 min

Lab File: BP021825.D

Acq: 04 Sep 2024 20:40

Instrument :

BNA_P

ClientSampleId :

YE3E3MS

Tgt Ion:200 Resp: 58445

Ion Ratio Lower Upper

200 100

173 26.1 21.4 32.0



215 46.9 39.0 58.6

Manual Integrations

APPROVED

Reviewed By :Jagrut



Upadhyay



09/05/2024

Supervised By :mohammad
ahmed

09/07/2024



#71

Pentachlorophenol

Concen: 26.902 ng/ul

RT: 16.863 min Scan# 2359

Delta R.T. 0.006 min

Lab File: BP021825.D

Acq: 04 Sep 2024 20:40




Tgt Ion:266 Resp: 425649


Ion Ratio Lower Upper

266 100

264 61.6 51.4 77.0

268 63.3 50.6 75.8





#72

Phenanthrene

Concen: 28.267 ng/ul

RT: 17.269 min Scan# 2426

Delta R.T. 0.012 min

Lab File: BP021825.D

Acq: 04 Sep 2024 20:40

Instrument :

BNP_P

ClientSampleId :

YE3E3MS

Tgt Ion:178 Resp: 292741

Ion Ratio Lower Upper

178 100

179 15.4 12.0 18.0



176 19.0 14.8 22.2

Manual Integrations

APPROVED

Reviewed By :Jagrut



Upadhyay



09/05/2024

Supervised By :mohammad
ahmed

09/07/2024



#74

Anthracene

Concen: 27.367 ng/ul

RT: 17.357 min Scan# 2443

Delta R.T. 0.006 min

Lab File: BP021825.D

Acq: 04 Sep 2024 20:40




Tgt Ion:178 Resp: 2871237


Ion Ratio Lower Upper

178 100

179 15.4 12.3 18.5

176 18.5 14.6 21.8





#75

1,2,3,4-Tetrachlorobenzene

Concen: 22.773 ng/uL

RT: 13.198 min Scan# 1

Delta R.T. 0.012 min

Lab File: BP021825.D

Acq: 04 Sep 2024 20:40

Instrument :

BNA_P

ClientSampleId :

YE3E3MS

Tgt Ion:216 Resp: 63867

Ion Ratio Lower Upper

216 100

214 76.9 63.9 95.9



218 48.3 39.5 59.3

Manual Integrations

APPROVED

Reviewed By :Jagrut



Upadhyay



09/05/2024

Supervised By :mohammad
ahmed

09/07/2024



#76

Pentachlorobenzene

Concen: 24.185 ng/uL

RT: 14.734 min Scan# 1997

Delta R.T. -0.000 min

Lab File: BP021825.D

Acq: 04 Sep 2024 20:40




Tgt Ion:250 Resp: 703019


Ion Ratio Lower Upper

250 100

248 64.1 51.0 76.6


252 64.4 51.4 77.0





#77
 Carbazole
 Concen: 30.002 ng/ul
 RT: 17.639 min Scan# 2489
 Delta R.T. 0.012 min
 Lab File: BP021825.D
 Acq: 04 Sep 2024 20:40



Instrument :
 BNA_P
 ClientSampleId :
 YE3E3MS



Tgt Ion:167 Resp: 283109
 Ion Ratio Lower Upper
 167 100
 166 21.1 17.1 25.7
 139 13.4 10.4 15.6


Manual Integrations
 APPROVED

Reviewed By :Jagrut
 Upadhyay




09/05/2024
 Supervised By :mohammad
 ahmed



09/07/2024




#78
 Di-n-butylphthalate
 Concen: 29.772 ng/ul
 RT: 18.233 min Scan# 2592
 Delta R.T. -0.000 min
 Lab File: BP021825.D
 Acq: 04 Sep 2024 20:40




Tgt Ion:149 Resp: 3419676
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.4 11.0
 104 6.3 4.8 7.2





#80
Fluoranthene
Concen: 28.855 ng/ul
RT: 19.351 min Scan# 21
Delta R.T. -0.000 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40



Instrument :
BNA_P
ClientSampleId :
YE3E3MS



Tgt Ion:202 Resp: 3614172
Ion Ratio Lower Upper
202 100
101 14.0 11.2 16.8
100 10.8 8.5 12.7


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






09/05/2024
Supervised By :mohammad
ahmed


09/07/2024



#82
Pyrene
Concen: 28.670 ng/ul
RT: 19.728 min Scan# 2846
Delta R.T. -0.000 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40

Tgt Ion:202 Resp: 3803164
Ion Ratio Lower Upper
202 100
101 16.7 12.8 19.2
100 14.2 10.8 16.2





#83

Butylbenzylphthalate

Concen: 30.854 ng/ul

RT: 20.675 min Scan# 3008

Delta R.T. -0.006 min

Lab File: BP021825.D

Acq: 04 Sep 2024 20:40

Instrument :

BNA_P

ClientSampleId :

YE3E3MS

Tgt Ion:149 Resp: 167578

Ion Ratio Lower Upper

149 100

91 85.7 70.1 105.1



206 20.0 16.8 25.2

Manual Integrations

APPROVED

Reviewed By :Jagrut



Upadhyay



09/05/2024

Supervised By :mohammad
ahmed

09/07/2024



#84

3,3'-Dichlorobenzidine

Concen: 13.205 ng/ul

RT: 21.574 min Scan# 3160

Delta R.T. -0.000 min

Lab File: BP021825.D

Acq: 04 Sep 2024 20:40




Tgt Ion:252 Resp: 624835


Ion Ratio Lower Upper

252 100

254 65.1 52.2 78.4

126 11.9 8.9 13.3





#85

Benzo(a)anthracene

Concen: 28.430 ng/ul

RT: 21.657 min Scan# 3174

Delta R.T. -0.000 min

Lab File: BP021825.D

Acq: 04 Sep 2024 20:40

Instrument :

BNA_P

ClientSampleId :

YE3E3MS

Tgt Ion:228 Resp: 3797220

Ion Ratio Lower Upper

228 100

229 19.9 15.7 23.5



226 26.8 21.7 32.5

Manual Integrations

APPROVED

Reviewed By :Jagrut



Upadhyay



09/05/2024

Supervised By :mohammad
ahmed

09/07/2024



#86

Bis(2-ethylhexyl)phthalate

Concen: 29.722 ng/ul

RT: 21.598 min Scan# 3164

Delta R.T. -0.000 min

Lab File: BP021825.D

Acq: 04 Sep 2024 20:40




Tgt Ion:149 Resp: 2382616


Ion Ratio Lower Upper

149 100

167 27.3 21.8 32.6

279 4.2 3.7 5.5





#87

Chrysene

Concen: 28.426 ng/ul

RT: 21.727 min Scan# 3187

Delta R.T. -0.006 min

Lab File: BP021825.D

Acq: 04 Sep 2024 20:40

Instrument :

BNA_P

ClientSampleId :

YE3E3MS

Tgt Ion:228 Resp: 357320

Ion Ratio Lower Upper

228 100

226 29.7 23.5 35.3



229 19.6 15.6 23.4

Manual Integrations

APPROVED

Reviewed By :Jagrut



Upadhyay



09/05/2024

Supervised By :mohammad
ahmed

09/07/2024



#89

Di-n-octyl phthalate

Concen: 29.626 ng/ul




RT: 22.898 min Scan# 3385


Delta R.T. -0.006 min

Lab File: BP021825.D

Acq: 04 Sep 2024 20:40


Tgt Ion:149 Resp: 3982393





#90
Benzo(b)fluoranthene
Concen: 29.521 ng/ul
RT: 23.998 min Scan# 3572
Delta R.T. -0.006 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40



Instrument :
BNA_P
ClientSampleId :
YE3E3MS



Tgt Ion:252 Resp: 3922704
Ion Ratio Lower Upper
252 100
253 22.1 18.1 27.1
125 10.6 8.6 12.8


Manual Integrations
APPROVED

Reviewed By :Jagrut
Upadhyay






09/05/2024
Supervised By :mohammad
ahmed


09/07/2024



#91
Benzo(k)fluoranthene
Concen: 28.974 ng/ul
RT: 24.080 min Scan# 3586
Delta R.T. -0.000 min
Lab File: BP021825.D
Acq: 04 Sep 2024 20:40

Tgt Ion:252 Resp: 3769136
Ion Ratio Lower Upper
252 100
253 21.5 17.8 26.8
125 10.9 8.6 13.0





#93

Benzo(a)pyrene

Concen: 28.808 ng/ul

RT: 24.927 min Scan# 31

Delta R.T. -0.006 min

Lab File: BP021825.D

Acq: 04 Sep 2024 20:40

Instrument :

BNA_P

ClientSampleId :

YE3E3MS

Tgt Ion:252 Resp: 352666

Ion Ratio Lower Upper

252 100

253 22.0 17.4 26.0



125 12.2 9.6 14.4

Manual Integrations

APPROVED

Reviewed By :Jagrut



Upadhyay



09/05/2024

Supervised By :mohammad
ahmed

09/07/2024



#94

Indeno(1,2,3-cd)pyrene

Concen: 27.861 ng/ul m

RT: 28.945 min Scan# 4413

Delta R.T. -0.018 min

Lab File: BP021825.D

Acq: 04 Sep 2024 20:40




Tgt Ion:276 Resp: 4397749


Ion Ratio Lower Upper

276 100

138 19.4 14.9 22.3

277 23.8 19.0 28.6





#95

Dibenzo(a,h)anthracene

Concen: 27.769 ng/ul

RT: 29.051 min Scan# 4429

Delta R.T. 0.012 min

Lab File: BP021825.D

Acq: 04 Sep 2024 20:40

Instrument :

BNA_P

ClientSampleId :

YE3E3MS

Tgt Ion:278 Resp: 352214

Ion Ratio Lower Upper

278 100

139 16.4 12.7 19.1



279 23.9 18.9 28.3

Manual Integrations

APPROVED

Reviewed By :Jagrut



Upadhyay



09/05/2024

Supervised By :mohammad
ahmed

09/07/2024



#96

Benzo(g,h,i)perylene

Concen: 28.370 ng/ul m

RT: 30.156 min Scan# 4619

Delta R.T. 0.018 min

Lab File: BP021825.D

Acq: 04 Sep 2024 20:40

Tgt Ion:276 Resp: 3459908

Ion Ratio Lower Upper

276 100

138 20.8 17.3 25.9

277 23.3 19.4 29.0

