

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090425\
 Data File : BP025668.D
 Acq On : 04 Sep 2025 20:19
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 05 01:12:32 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 14 18:14:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	101	-0.02
2	1,4-Dioxane	40.000	37.738	5.7	101	-0.02
3	Pyridine	40.000	37.533	6.2	100	-0.01
4	n-Nitrosodimethylamine	40.000	44.240	-10.6	116	-0.01
5 S	2-Fluorophenol	80.000	77.102	3.6	101	-0.01
6	Aniline	40.000	37.613	6.0	98	-0.02
7 S	Phenol-d6	80.000	76.921	3.8	99	-0.01
8	2-Chlorophenol	40.000	38.355	4.1	100	-0.02
9	Benzaldehyde	40.000	47.047	-17.6	94	-0.02
10 C	Phenol	40.000	37.814	5.5	98	-0.01
11	bis(2-Chloroethyl)ether	40.000	37.828	5.4	99	-0.02
12	1,3-Dichlorobenzene	40.000	37.692	5.8	99	-0.02
13 C	1,4-Dichlorobenzene	40.000	37.988	5.0	100	-0.02
14	1,2-Dichlorobenzene	40.000	37.549	6.1	98	-0.02
15	Benzyl Alcohol	40.000	39.558	1.1	103	-0.02
16	2,2'-oxybis(1-Chloropropane	40.000	40.641	-1.6	107	-0.04
17	2-Methylphenol	40.000	38.194	4.5	98	-0.01
18	Hexachloroethane	40.000	39.366	1.6	103	-0.03
19 P	n-Nitroso-di-n-propylamine	40.000	38.904	2.7	101	-0.02
20	3+4-Methylphenols	40.000	37.041	7.4	95	-0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	99	-0.02
22	Acetophenone	40.000	38.775	3.1	99	-0.02
23 S	Nitrobenzene-d5	80.000	79.853	0.2	102	-0.02
24	Nitrobenzene	40.000	40.250	-0.6	102	-0.02
25	Isophorone	40.000	38.310	4.2	99	-0.02
26 C	2-Nitrophenol	40.000	38.686	3.3	96	-0.02
27	2,4-Dimethylphenol	40.000	37.628	5.9	94	-0.02
28	bis(2-Chloroethoxy)methane	40.000	37.566	6.1	97	-0.02
29 C	2,4-Dichlorophenol	40.000	37.581	6.0	94	-0.02
30	1,2,4-Trichlorobenzene	40.000	37.474	6.3	96	-0.02
31	Naphthalene	40.000	37.936	5.2	98	-0.02
32	Benzoic acid	40.000	35.949	10.1	90	0.01
33	4-Chloroaniline	40.000	37.349	6.6	94	-0.01
34 C	Hexachlorobutadiene	40.000	38.629	3.4	99	-0.03
35	Caprolactam	40.000	36.753	8.1	95	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.203	2.0	100	0.00
37	2-Methylnaphthalene	40.000	37.185	7.0	96	-0.02
38	1-Methylnaphthalene	40.000	36.766	8.1	96	-0.02
39 I	Acenaphthene-d10	20.000	20.000	0.0	97	-0.02
40	1,2,4,5-Tetrachlorobenzene	40.000	38.564	3.6	97	-0.02
41 P	Hexachlorocyclopentadiene	40.000	39.573	1.1	97	-0.02
42 S	2,4,6-Tribromophenol	80.000	80.310	-0.4	100	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.313	4.2	95	-0.02
44	2,4,5-Trichlorophenol	40.000	37.894	5.3	93	0.00
45 S	2-Fluorobiphenyl	80.000	74.401	7.0	95	-0.02
46	1,1'-Biphenyl	40.000	37.565	6.1	94	-0.02
47	2-Chloronaphthalene	40.000	37.268	6.8	93	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090425\
 Data File : BP025668.D
 Acq On : 04 Sep 2025 20:19
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 05 01:12:32 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 14 18:14:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.114	-5.3	102	-0.02
49	Acenaphthylene	40.000	37.324	6.7	94	-0.02
50	Dimethylphthalate	40.000	37.689	5.8	96	-0.02
51	2,6-Dinitrotoluene	40.000	38.454	3.9	94	0.00
52 C	Acenaphthene	40.000	36.699	8.3	94	-0.01
53	3-Nitroaniline	40.000	37.624	5.9	93	0.00
54 P	2,4-Dinitrophenol	40.000	45.606	-14.0	112	0.00
55	Dibenzofuran	40.000	37.238	6.9	94	-0.01
56 P	4-Nitrophenol	40.000	33.247	16.9	83	0.02
57	2,4-Dinitrotoluene	40.000	39.619	1.0	97	0.00
58	Fluorene	40.000	37.004	7.5	95	-0.01
59	2,3,4,6-Tetrachlorophenol	40.000	37.302	6.7	93	0.00
60	Diethylphthalate	40.000	37.612	6.0	95	-0.01
61	4-Chlorophenyl-phenylether	40.000	37.210	7.0	95	-0.02
62	4-Nitroaniline	40.000	38.171	4.6	94	0.00
63	Azobenzene	40.000	40.139	-0.3	101	-0.01
64 I	Phenanthrene-d10	20.000	20.000	0.0	95	-0.01
65	4,6-Dinitro-2-methylphenol	40.000	43.549	-8.9	106	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.293	4.3	94	-0.01
67	4-Bromophenyl-phenylether	40.000	38.620	3.5	95	-0.01
68	Hexachlorobenzene	40.000	38.475	3.8	97	-0.01
69	Atrazine	40.000	37.722	5.7	94	-0.01
70 C	Pentachlorophenol	40.000	37.826	5.4	92	0.00
71	Phenanthrene	40.000	37.482	6.3	95	-0.01
72	Anthracene	40.000	37.715	5.7	94	-0.01
73	Carbazole	40.000	36.909	7.7	92	0.00
74	Di-n-butylphthalate	40.000	38.544	3.6	96	-0.02
75 C	Fluoranthene	40.000	36.453	8.9	93	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	-0.01
77	Benzydine	40.000	32.582	18.5	58	0.00
78	Pyrene	40.000	36.069	9.8	92	0.00
79 S	Terphenyl-d14	80.000	73.783	7.8	97	0.00
80	Butylbenzylphthalate	40.000	37.832	5.4	96	0.00
81	Benzo(a)anthracene	40.000	36.805	8.0	96	-0.02
82	3,3'-Dichlorobenzidine	40.000	35.269	11.8	90	-0.01
83	Chrysene	40.000	37.283	6.8	98	-0.01
84	Bis(2-ethylhexyl)phthalate	40.000	37.368	6.6	98	-0.03
85 c	Di-n-octyl phthalate	40.000	37.446	6.4	97	-0.04
86 I	Perylene-d12	20.000	20.000	0.0	98	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.727	5.7	96	-0.02
88	Benzo(b)fluoranthene	40.000	37.731	5.7	96	-0.02
89	Benzo(k)fluoranthene	40.000	38.097	4.8	98	-0.01
90 C	Benzo(a)pyrene	40.000	37.596	6.0	96	0.00
91	Dibenzo(a,h)anthracene	40.000	37.949	5.1	97	-0.02
92	Benzo(g,h,i)perylene	40.000	37.593	6.0	95	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090425\
Data File : BP025668.D
Acq On : 04 Sep 2025 20:19
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Sep 05 01:12:32 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 14 18:14:31 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0