

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090525\
 Data File : BP025682.D
 Acq On : 05 Sep 2025 18:49
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 05 19:37:29 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 14 18:14:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	110	-0.03
2	1,4-Dioxane	40.000	37.068	7.3	108	-0.02
3	Pyridine	40.000	36.753	8.1	107	-0.02
4	n-Nitrosodimethylamine	40.000	42.426	-6.1	121	-0.02
5 S	2-Fluorophenol	80.000	76.043	4.9	108	-0.02
6	Aniline	40.000	36.963	7.6	105	-0.02
7 S	Phenol-d6	80.000	75.575	5.5	106	-0.01
8	2-Chlorophenol	40.000	38.003	5.0	108	-0.02
9	Benzaldehyde	40.000	46.955	-17.4	102	-0.02
10 C	Phenol	40.000	37.492	6.3	106	-0.01
11	bis(2-Chloroethyl)ether	40.000	37.668	5.8	107	-0.02
12	1,3-Dichlorobenzene	40.000	37.847	5.4	108	-0.02
13 C	1,4-Dichlorobenzene	40.000	37.308	6.7	107	-0.03
14	1,2-Dichlorobenzene	40.000	37.056	7.4	106	-0.02
15	Benzyl Alcohol	40.000	39.010	2.5	110	-0.02
16	2,2'-oxybis(1-Chloropropane	40.000	40.814	-2.0	117	-0.03
17	2-Methylphenol	40.000	37.875	5.3	106	-0.01
18	Hexachloroethane	40.000	39.504	1.2	112	-0.03
19 P	n-Nitroso-di-n-propylamine	40.000	39.200	2.0	111	-0.02
20	3+4-Methylphenols	40.000	37.452	6.4	105	-0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	108	-0.02
22	Acetophenone	40.000	38.423	3.9	107	-0.02
23 S	Nitrobenzene-d5	80.000	79.907	0.1	112	-0.02
24	Nitrobenzene	40.000	39.197	2.0	109	-0.02
25	Isophorone	40.000	39.466	1.3	111	-0.02
26 C	2-Nitrophenol	40.000	39.166	2.1	106	-0.02
27	2,4-Dimethylphenol	40.000	37.252	6.9	102	-0.02
28	bis(2-Chloroethoxy)methane	40.000	37.229	6.9	105	-0.02
29 C	2,4-Dichlorophenol	40.000	38.257	4.4	105	-0.01
30	1,2,4-Trichlorobenzene	40.000	36.857	7.9	104	-0.02
31	Naphthalene	40.000	36.866	7.8	104	-0.02
32	Benzoic acid	40.000	37.988	5.0	104	0.02
33	4-Chloroaniline	40.000	36.868	7.8	101	-0.01
34 C	Hexachlorobutadiene	40.000	38.813	3.0	109	-0.02
35	Caprolactam	40.000	37.668	5.8	106	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.466	1.3	110	0.00
37	2-Methylnaphthalene	40.000	36.926	7.7	104	-0.01
38	1-Methylnaphthalene	40.000	36.626	8.4	104	-0.02
39 I	Acenaphthene-d10	20.000	20.000	0.0	105	-0.01
40	1,2,4,5-Tetrachlorobenzene	40.000	39.187	2.0	106	-0.02
41 P	Hexachlorocyclopentadiene	40.000	39.057	2.4	103	-0.02
42 S	2,4,6-Tribromophenol	80.000	82.063	-2.6	111	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.412	1.5	106	0.00
44	2,4,5-Trichlorophenol	40.000	39.142	2.1	104	0.00
45 S	2-Fluorobiphenyl	80.000	74.767	6.5	103	-0.02
46	1,1'-Biphenyl	40.000	38.124	4.7	104	-0.01
47	2-Chloronaphthalene	40.000	37.711	5.7	102	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090525\
 Data File : BP025682.D
 Acq On : 05 Sep 2025 18:49
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 05 19:37:29 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081425.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Aug 14 18:14:31 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.216	-8.0	114	-0.01
49	Acenaphthylene	40.000	37.697	5.8	103	-0.01
50	Dimethylphthalate	40.000	38.246	4.4	105	-0.01
51	2,6-Dinitrotoluene	40.000	38.803	3.0	102	0.00
52 C	Acenaphthene	40.000	36.500	8.8	101	0.00
53	3-Nitroaniline	40.000	37.907	5.2	101	0.00
54 P	2,4-Dinitrophenol	40.000	47.052	-17.6	125	0.01
55	Dibenzofuran	40.000	36.815	8.0	100	-0.01
56 P	4-Nitrophenol	40.000	35.447	11.4	95	0.03
57	2,4-Dinitrotoluene	40.000	39.780	0.5	105	0.00
58	Fluorene	40.000	37.450	6.4	103	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.273	1.8	106	0.00
60	Diethylphthalate	40.000	39.328	1.7	107	0.00
61	4-Chlorophenyl-phenylether	40.000	37.895	5.3	104	0.00
62	4-Nitroaniline	40.000	37.595	6.0	100	0.00
63	Azobenzene	40.000	40.576	-1.4	110	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	104	-0.01
65	4,6-Dinitro-2-methylphenol	40.000	43.897	-9.7	116	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.174	4.6	102	0.00
67	4-Bromophenyl-phenylether	40.000	38.402	4.0	103	0.00
68	Hexachlorobenzene	40.000	39.075	2.3	107	0.00
69	Atrazine	40.000	39.232	1.9	106	0.00
70 C	Pentachlorophenol	40.000	39.493	1.3	105	0.00
71	Phenanthrene	40.000	37.761	5.6	103	0.00
72	Anthracene	40.000	38.280	4.3	103	0.00
73	Carbazole	40.000	37.141	7.1	101	0.00
74	Di-n-butylphthalate	40.000	40.316	-0.8	109	0.00
75 C	Fluoranthene	40.000	36.981	7.5	102	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	104	0.00
77	Benzydine	40.000	32.046	19.9	59	0.00
78	Pyrene	40.000	37.633	5.9	100	0.00
79 S	Terphenyl-d14	80.000	77.158	3.6	105	0.00
80	Butylbenzylphthalate	40.000	39.983	0.0	106	0.00
81	Benzo(a)anthracene	40.000	37.071	7.3	101	-0.01
82	3,3'-Dichlorobenzidine	40.000	36.103	9.7	96	0.01
83	Chrysene	40.000	37.280	6.8	102	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	39.381	1.5	107	-0.02
85 c	Di-n-octyl phthalate	40.000	40.980	-2.4	110	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	106	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	37.498	6.3	103	0.02
88	Benzo(b)fluoranthene	40.000	36.537	8.7	101	0.00
89	Benzo(k)fluoranthene	40.000	37.096	7.3	103	0.00
90 C	Benzo(a)pyrene	40.000	37.242	6.9	102	0.01
91	Dibenzo(a,h)anthracene	40.000	38.183	4.5	105	0.00
92	Benzo(g,h,i)perylene	40.000	37.914	5.2	104	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090525\
Data File : BP025682.D
Acq On : 05 Sep 2025 18:49
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Sep 05 19:37:29 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081425.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Aug 14 18:14:31 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0