

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090624\
 Data File : BP021856.D
 Acq On : 06 Sep 2024 18:17
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP090624

Quant Time: Sep 06 23:55:51 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP090624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 06 23:52:54 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	99	0.00
2	1,4-Dioxane	40.000	36.905	7.7	96	0.00
3	Pyridine	40.000	37.232	6.9	99	0.00
4	n-Nitrosodimethylamine	40.000	37.698	5.8	100	0.00
5 S	2-Fluorophenol	80.000	77.149	3.6	98	0.00
6	Aniline	40.000	40.861	-2.2	101	0.00
7 S	Phenol-d6	80.000	81.470	-1.8	100	-0.01
8	2-Chlorophenol	40.000	39.048	2.4	100	0.00
9	Benzaldehyde	40.000	42.450	-6.1	98	-0.01
10 C	Phenol	40.000	40.662	-1.7	101	0.00
11	bis(2-Chloroethyl)ether	40.000	39.987	0.0	99	0.00
12	1,3-Dichlorobenzene	40.000	39.044	2.4	99	0.00
13 C	1,4-Dichlorobenzene	40.000	38.744	3.1	98	0.00
14	1,2-Dichlorobenzene	40.000	38.594	3.5	98	0.00
15	Benzyl Alcohol	40.000	40.404	-1.0	99	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.897	0.3	97	-0.01
17	2-Methylphenol	40.000	40.755	-1.9	101	0.00
18	Hexachloroethane	40.000	39.626	0.9	99	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.855	-2.1	99	0.00
20	3+4-Methylphenols	40.000	41.208	-3.0	101	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	100	0.00
22	Acetophenone	40.000	39.172	2.1	100	0.00
23 S	Nitrobenzene-d5	80.000	80.420	-0.5	100	0.00
24	Nitrobenzene	40.000	39.656	0.9	100	0.00
25	Isophorone	40.000	41.759	-4.4	100	0.00
26 C	2-Nitrophenol	40.000	37.835	5.4	91	0.00
27	2,4-Dimethylphenol	40.000	36.831	7.9	89	-0.01
28	bis(2-Chloroethoxy)methane	40.000	35.271	11.8	86	0.00
29 C	2,4-Dichlorophenol	40.000	36.219	9.5	90	0.00
30	1,2,4-Trichlorobenzene	40.000	39.081	2.3	100	-0.01
31	Naphthalene	40.000	38.761	3.1	99	0.00
32	Benzoic acid	40.000	44.136	-10.3	109	0.00
33	4-Chloroaniline	40.000	40.150	-0.4	100	0.00
34 C	Hexachlorobutadiene	40.000	38.774	3.1	98	0.00
35	Caprolactam	40.000	40.811	-2.0	102	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.592	-1.5	100	0.00
37	2-Methylnaphthalene	40.000	40.004	-0.0	109	0.00
38	1-Methylnaphthalene	40.000	38.448	3.9	101	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	99	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.438	1.4	98	0.00
41 P	Hexachlorocyclopentadiene	40.000	42.289	-5.7	96	0.00
42 S	2,4,6-Tribromophenol	80.000	80.034	-0.0	98	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.334	-0.8	99	0.00
44	2,4,5-Trichlorophenol	40.000	39.971	0.1	99	0.00
45 S	2-Fluorobiphenyl	80.000	78.724	1.6	100	0.00
46	1,1'-Biphenyl	40.000	39.193	2.0	98	0.00
47	2-Chloronaphthalene	40.000	39.187	2.0	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090624\
 Data File : BP021856.D
 Acq On : 06 Sep 2024 18:17
 Operator : RC/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP090624

Quant Time: Sep 06 23:55:51 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP090624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Sep 06 23:52:54 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.236	-5.6	101	0.00
49	Acenaphthylene	40.000	40.279	-0.7	100	0.00
50	Dimethylphthalate	40.000	39.089	2.3	98	0.00
51	2,6-Dinitrotoluene	40.000	41.317	-3.3	100	0.00
52 C	Acenaphthene	40.000	39.018	2.5	98	0.00
53	3-Nitroaniline	40.000	41.286	-3.2	101	0.00
54 P	2,4-Dinitrophenol	40.000	41.133	-2.8	111	0.00
55	Dibenzofuran	40.000	39.374	1.6	99	0.00
56 P	4-Nitrophenol	40.000	42.825	-7.1	103	0.00
57	2,4-Dinitrotoluene	40.000	42.022	-5.1	99	0.00
58	Fluorene	40.000	39.560	1.1	99	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	39.328	1.7	99	0.00
60	Diethylphthalate	40.000	40.164	-0.4	98	-0.01
61	4-Chlorophenyl-phenylether	40.000	39.693	0.8	98	0.00
62	4-Nitroaniline	40.000	42.122	-5.3	99	-0.01
63	Azobenzene	40.000	40.387	-1.0	98	-0.01
64 I	Phenanthrene-d10	20.000	20.000	0.0	81	0.00
65	4,6-Dinitro-2-methylphenol	40.000	49.267	-23.2	102	-0.01
66 c	n-Nitrosodiphenylamine	40.000	48.569	-21.4#	100	0.00
67	4-Bromophenyl-phenylether	40.000	49.368	-23.4	99	0.00
68	Hexachlorobenzene	40.000	48.084	-20.2	99	0.00
69	Atrazine	40.000	56.294	-40.7#	97	0.00
70 C	Pentachlorophenol	40.000	50.727	-26.8#	100	0.00
71	Phenanthrene	40.000	38.242	4.4	79	0.00
72	Anthracene	40.000	38.700	3.2	79	0.00
73	Carbazole	40.000	39.846	0.4	83	0.00
74	Di-n-butylphthalate	40.000	49.525	-23.8	96	0.00
75 C	Fluoranthene	40.000	48.356	-20.9#	100	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzydine	40.000	39.273	1.8	84	-0.01
78	Pyrene	40.000	39.593	1.0	100	0.00
79 S	Terphenyl-d14	80.000	77.804	2.7	96	0.00
80	Butylbenzylphthalate	40.000	40.685	-1.7	97	-0.01
81	Benzo(a)anthracene	40.000	39.678	0.8	101	0.00
82	3,3'-Dichlorobenzidine	40.000	41.111	-2.8	99	-0.01
83	Chrysene	40.000	39.502	1.2	101	-0.01
84	Bis(2-ethylhexyl)phthalate	40.000	40.813	-2.0	97	-0.01
85 c	Di-n-octyl phthalate	40.000	41.268	-3.2	97	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	100	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	38.646	3.4	100	-0.01
88	Benzo(b)fluoranthene	40.000	40.542	-1.4	102	-0.01
89	Benzo(k)fluoranthene	40.000	39.393	1.5	99	-0.02
90 C	Benzo(a)pyrene	40.000	40.589	-1.5	109	-0.01
91	Dibenzo(a,h)anthracene	40.000	38.357	4.1	101	-0.02
92	Benzo(g,h,i)perylene	40.000	38.904	2.7	101	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090624\
Data File : BP021856.D
Acq On : 06 Sep 2024 18:17
Operator : RC/JU
Sample : SSTDICV040
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP090624

Quant Time: Sep 06 23:55:51 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP090624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Sep 06 23:52:54 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 3