

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090722\
 Data File : BP011650.D
 Acq On : 07 Sep 2022 16:22
 Operator : CG/JU
 Sample : N4501-10
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SB-12-COMP

Quant Time: Sep 08 04:06:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP090222.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 03 04:07:09 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.969	152	703377	20.000	ng	0.00
21) Naphthalene-d8	10.781	136	2850811	20.000	ng	#-0.01
39) Acenaphthene-d10	14.610	164	1575810	20.000	ng	0.00
64) Phenanthrene-d10	17.363	188	2848792	20.000	ng	# 0.00
76) Chrysene-d12	21.445	240	1889345	20.000	ng	0.00
86) Perylene-d12	23.915	264	1482007	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.528	112	7862620	197.701	ng	0.00
7) Phenol-d6	7.134	99	9980766	189.660	ng	0.00
23) Nitrobenzene-d5	9.140	82	5867710	126.344	ng	0.00
42) 2,4,6-Tribromophenol	16.104	330	2903588	163.084	ng	0.00
45) 2-Fluorobiphenyl	13.239	172	12216186	110.858	ng	0.00
79) Terphenyl-d14	19.916	244	12099591	134.134	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090722\
Data File : BP011650.D
Acq On : 07 Sep 2022 16:22
Operator : CG/JU
Sample : N4501-10
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SB-12-COMP

Quant Time: Sep 08 04:06:18 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP090222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 03 04:07:09 2022
Response via : Initial Calibration

