

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090821\
 Data File : BP006885.D
 Acq On : 08 Sep 2021 16:24
 Operator : CG/JU
 Sample : SSTD01016
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD010616

Quant Time: Sep 09 01:59:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP090821.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 09 01:57:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.940	152	246091	20.000	ng/u1	0.00
20) Naphthalene-d8	10.740	136	989046	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.563	164	626553	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.316	188	1322610	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.398	240	1372593	20.000	ng/u1	0.00
88) Perylene-d12	23.833	264	1436192	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.370	96	25217	3.466	ng/uL	0.00
4) Pyridine-d5	3.787	84	149179	7.133	ng/u1	0.00
7) Phenol-d5	7.093	99	166954	6.749	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.269	67	102026	6.711	ng/u1	0.00
11) 2-Chlorophenol-d4	7.469	132	137476	7.461	ng/u1	0.00
15) 4-Methylphenol-d8	8.640	113	128846	6.519	ng/u1	0.00
21) Nitrobenzene-d5	9.099	128	53446	6.366	ng/u1	0.00
24) 2-Nitrophenol-d4	9.822	143	49182	5.598	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.358	165	129340	8.714	ng/u1	0.00
31) 4-Chloroaniline-d4	10.875	131	180248	7.472	ng/u1	0.00
46) Dimethylphthalate-d6	13.975	166	393744	8.194	ng/u1	0.00
49) Acenaphthylene-d8	14.257	160	493498	8.373	ng/u1	0.00
54) 4-Nitrophenol-d4	14.746	143	53186	5.299	ng/u1	0.00
60) Fluorene-d10	15.557	176	353778	9.053	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.669	200	38122	4.361	ng/u1	0.00
73) Anthracene-d10	17.410	188	540809	8.642	ng/u1	0.00
81) Pyrene-d10	19.645	212	628803	8.724	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.674	264	668969	8.638	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.405	88	28111	3.803	ng/uL	95
5) Pyridine	3.811	79	153504	7.090	ng/u1	91
6) Benzaldehyde	7.075	77	113670	9.345	ng/u1	90
8) Phenol	7.122	94	176261	6.976	ng/u1	92
10) Bis(2-Chloroethyl)ether	7.364	93	146948	7.546	ng/u1	96
12) 2-Chlorophenol	7.499	128	144518	7.778	ng/u1	93
13) 2-Methylphenol	8.375	108	131441	7.071	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.469	45	172202	7.568	ng/u1	97
16) Acetophenone	8.763	105	215624	6.930	ng/u1	89
17) N-Nitroso-di-n-propyla...	8.746	70	97760	5.841	ng/u1	97
18) 4-Methylphenol	8.699	108	139839	6.919	ng/u1	100
19) Hexachloroethane	9.022	117	58985	7.296	ng/u1#	81
22) Nitrobenzene	9.140	77	134766	6.131	ng/u1	91
23) Isophorone	9.669	82	270512	6.600	ng/u1	95
25) 2-Nitrophenol	9.852	139	55281	5.966	ng/u1#	85
26) 2,4-Dimethylphenol	9.910	107	155747	7.824	ng/u1	94
27) Bis(2-Chloroethoxy)met...	10.152	93	185665	7.814	ng/u1	99
29) 2,4-Dichlorophenol	10.381	162	131467	9.130	ng/u1	93
30) Naphthalene	10.793	128	480550	8.953	ng/u1	99
32) 4-Chloroaniline	10.899	127	184133	7.778	ng/u1	100
33) Hexachlorobutadiene	11.081	225	92182	10.858	ng/u1	95
34) Caprolactam	11.675	113	24993	4.335	ng/u1	97
35) 4-Chloro-3-methylphenol	12.010	107	130899	7.106	ng/u1	90
36) 2-Methylnaphthalene	12.399	142	319709	8.668	ng/u1	96

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP090821\
 Data File : BP006885.D
 Acq On : 08 Sep 2021 16:24
 Operator : CG/JU
 Sample : SST01016
 Misc :
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SST010616


Quant Time: Sep 09 01:59:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP090821.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Sep 09 01:57:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.616	142	324074	8.829	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.763	216	176000	10.423	ng/ul#	95
40) Hexachlorocyclopentadiene	12.746	237	101524	9.165	ng/ul	98
41) 2,4,6-Trichlorophenol	12.999	196	94810	8.176	ng/ul	98
42) 2,4,5-Trichlorophenol	13.063	196	108712	8.806	ng/ul	95
43) 1,1'-Biphenyl	13.404	154	438628	9.005	ng/ul#	96
44) 2-Chloronaphthalene	13.440	162	333647	9.029	ng/ul	93
45) 2-Nitroaniline	13.640	65	57154	4.289	ng/ul#	85
47) Dimethylphthalate	14.022	163	400548	8.428	ng/ul	99
48) 2,6-Dinitrotoluene	14.140	165	55837	6.087	ng/ul#	79
50) Acenaphthylene	14.287	152	526275	9.134	ng/ul	99
51) 3-Nitroaniline	14.463	138	56545	5.383	ng/ul#	81
52) Acenaphthene	14.628	153	350675	8.536	ng/ul	98
53) 2,4-Dinitrophenol	14.669	184	21296	3.589	ng/ul#	75
55) 4-Nitrophenol	14.757	109	41757	4.640	ng/ul#	83
56) Dibenzofuran	14.963	168	510445	9.293	ng/ul	87
57) 2,4-Dinitrotoluene	14.922	165	82137	6.086	ng/ul#	78
58) 2,3,4,6-Tetrachlorophenol	15.187	232	84154	8.211	ng/ul#	77
59) Diethylphthalate	15.387	149	386211	7.645	ng/ul	95
61) Fluorene	15.610	166	411795	8.991	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.610	204	207271	9.997	ng/ul#	80
63) 4-Nitroaniline	15.622	138	56898	5.784	ng/ul	86
66) 4,6-Dinitro-2-methylph...	15.681	198	38564	4.413	ng/ul#	80
67) N-Nitrosodiphenylamine	15.822	169	347995	8.566	ng/ul	97
68) 4-Bromophenyl-phenylether	16.504	248	122665	9.404	ng/ul#	83
69) Hexachlorobenzene	16.616	284	142774	9.685	ng/ul#	87
70) Atrazine	16.775	200	120736	8.377	ng/ul	95
71) Pentachlorophenol	16.963	266	59690	6.182	ng/ul	96
72) Phenanthrene	17.357	178	667692	8.936	ng/ul	98
74) Anthracene	17.445	178	658610	8.680	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.369	216	181757	10.106	ng/uL	99
76) Pentachlorobenzene	14.881	250	176052	9.935	ng/uL	97
77) Carbazole	17.716	167	580717	8.194	ng/ul	96
78) Di-n-butylphthalate	18.275	149	627561	6.849	ng/ul	98
80) Fluoranthene	19.322	202	794467	8.769	ng/ul#	94
82) Pyrene	19.669	202	816912	8.730	ng/ul	96
83) Butylbenzylphthalate	20.545	149	254412	5.499	ng/ul	91
84) 3,3'-Dichlorobenzidine	21.316	252	244432	7.804	ng/ul	98
85) Benzo(a)anthracene	21.380	228	788875	8.824	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.310	149	391882	5.497	ng/ul	94
87) Chrysene	21.433	228	789227	9.156	ng/ul	98
89) Di-n-octyl phthalate	22.245	149	652063	5.252	ng/ul	100
90) Benzo(b)fluoranthene	23.086	252	812303	8.707	ng/ul	98
91) Benzo(k)fluoranthene	23.133	252	792271	8.771	ng/ul	98
93) Benzo(a)pyrene	23.721	252	795322	9.410	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.380	276	912828	8.260	ng/ul	97
95) Dibenzo(a,h)anthracene	26.410	278	801820	8.638	ng/ul	96
96) Benzo(g,h,i)perylene	27.162	276	798880	8.707	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 09 01:59:41 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP090821.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Sep 09 01:57:53 2021
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 3.803 ng/uL
RT: 3.405 min Scan# 54
Delta R.T. 0.006 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion: 88 Resp: 28111
Ion Ratio Lower Upper
88 100
43 25.0 20.1 30.1
58 66.0 48.4 72.6



#5
Pyridine
Concen: 7.090 ng/uL
RT: 3.811 min Scan# 123
Delta R.T. 0.006 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion: 79 Resp: 153504
Ion Ratio Lower Upper
79 100
52 69.0 49.2 73.8
51 32.2 22.7 34.1









#6
Benzaldehyde
Concen: 9.345 ng/ul
RT: 7.075 min Scan# 678
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion: 77 Resp: 113670
Ion Ratio Lower Upper
77 100
105 102.9 74.4 111.6
106 96.6 70.3 105.5



#8
Phenol
Concen: 6.976 ng/ul
RT: 7.122 min Scan# 686
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion: 94 Resp: 176261
Ion Ratio Lower Upper
94 100
65 26.1 24.4 36.6
66 34.8 31.8 47.6









#10
 Bis(2-Chloroethyl)ether
 Concen: 7.546 ng/ul
 RT: 7.364 min Scan# 727
 Delta R.T. 0.000 min
 Lab File: BP006885.D
 Acq: 08 Sep 2021 16:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTD010616


Tgt Ion:	93	Resp:	146948
Ion	Ratio	Lower	Upper
93	100		
63	71.6	59.7	89.5
95	33.9	25.3	37.9



#12
 2-Chlorophenol
 Concen: 7.778 ng/ul
 RT: 7.499 min Scan# 750
 Delta R.T. 0.000 min
 Lab File: BP006885.D
 Acq: 08 Sep 2021 16:24

Tgt Ion:	128	Resp:	144518
Ion	Ratio	Lower	Upper
128	100		
64	43.2	40.6	60.8
130	32.7	26.4	39.6









#13
2-Methylphenol
Concen: 7.071 ng/ul
RT: 8.375 min Scan# 899
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:108 Resp: 131441
Ion Ratio Lower Upper
108 100
107 87.7 72.2 108.2



#14
2,2'-oxybis(1-Chloropropane)
Concen: 7.568 ng/ul
RT: 8.469 min Scan# 915
Delta R.T. -0.006 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion: 45 Resp: 172202
Ion Ratio Lower Upper
45 100
77 17.7 13.8 20.8
79 11.7 10.9 16.3









#16
Acetophenone
Concen: 6.930 ng/ul
RT: 8.763 min Scan# 965
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:	105	Resp:	215624
Ion	Ratio	Lower	Upper
105	100		
77	71.5	67.0	100.4
51	23.0	19.0	28.6



#17
N-Nitroso-di-n-propylamine
Concen: 5.841 ng/ul
RT: 8.746 min Scan# 962
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:	70	Resp:	97760
Ion	Ratio	Lower	Upper
70	100		
42	48.8	40.4	60.6
101	9.7	7.2	10.8
130	21.5	14.9	22.3









#18
4-Methylphenol
Concen: 6.919 ng/ul
RT: 8.699 min Scan# 954
Delta R.T. -0.006 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:108 Resp: 139839
Ion Ratio Lower Upper
108 100
107 110.3 88.6 133.0



#19
Hexachloroethane
Concen: 7.296 ng/ul
RT: 9.022 min Scan# 1009
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:117 Resp: 58985
Ion Ratio Lower Upper
117 100
201 104.6 68.2 102.2#
199 64.8 43.5 65.3









#22
Nitrobenzene
Concen: 6.131 ng/ul
RT: 9.140 min Scan# 1029
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion: 77 Resp: 134766
Ion Ratio Lower Upper
77 100
123 45.5 31.1 46.7
65 12.0 10.6 16.0



#23
Isophorone
Concen: 6.600 ng/ul
RT: 9.669 min Scan# 1119
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion: 82 Resp: 270512
Ion Ratio Lower Upper
82 100
95 6.8 5.8 8.8
138 18.2 12.5 18.7









#25
2-Nitrophenol
Concen: 5.966 ng/ul
RT: 9.852 min Scan# 1150
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:139 Resp: 55281
Ion Ratio Lower Upper
139 100
65 48.1 48.4 72.6#
109 28.3 28.2 42.2



#26
2,4-Dimethylphenol
Concen: 7.824 ng/ul
RT: 9.910 min Scan# 1160
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:107 Resp: 155747
Ion Ratio Lower Upper
107 100
121 53.1 38.1 57.1
122 84.4 64.2 96.4









#27
Bis(2-Chloroethoxy)methane
Concen: 7.814 ng/ul
RT: 10.152 min Scan# 1201
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion: 93 Resp: 185665
Ion Ratio Lower Upper
93 100
95 32.1 25.6 38.4
123 11.7 8.2 12.4



#29
2,4-Dichlorophenol
Concen: 9.130 ng/ul
RT: 10.381 min Scan# 1240
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion: 162 Resp: 131467
Ion Ratio Lower Upper
162 100
164 67.7 52.0 78.0
98 35.4 34.3 51.5









#30
Naphthalene
Concen: 8.953 ng/ul
RT: 10.793 min Scan# 1310
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:128 Resp: 480550
Ion Ratio Lower Upper
128 100
129 10.6 8.8 13.2
127 13.4 10.6 15.8



#32
4-Chloroaniline
Concen: 7.778 ng/ul
RT: 10.899 min Scan# 1328
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:127 Resp: 184133
Ion Ratio Lower Upper
127 100
129 32.4 25.7 38.5









#33
Hexachlorobutadiene
Concen: 10.858 ng/ul
RT: 11.081 min Scan# 1359
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:225 Resp: 92182
Ion Ratio Lower Upper
225 100
223 65.8 48.2 72.2
227 64.1 49.6 74.4



#34
Caprolactam
Concen: 4.335 ng/ul
RT: 11.675 min Scan# 1460
Delta R.T. 0.006 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:113 Resp: 24993
Ion Ratio Lower Upper
113 100
55 157.5 125.8 188.6
56 116.4 98.2 147.2









#35
4-Chloro-3-methylphenol
Concen: 7.106 ng/ul
RT: 12.010 min Scan# 1517
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:	107	Resp:	130899
Ion	Ratio	Lower	Upper
107	100		
144	24.9	18.8	28.2
142	83.0	58.4	87.6



#36
2-Methylnaphthalene
Concen: 8.668 ng/ul
RT: 12.399 min Scan# 1583
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:	142	Resp:	319709
Ion	Ratio	Lower	Upper
142	100		
141	86.9	72.3	108.5









#37
1-Methylnaphthalene
Concen: 8.829 ng/ul
RT: 12.616 min Scan# 1620
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:	142	Resp:	324074
Ion Ratio	Lower	Upper	
142	100		
141	92.8	74.9	112.3
116	3.8	3.3	4.9



#39
1,2,4,5-Tetrachlorobenzene
Concen: 10.423 ng/ul
RT: 12.763 min Scan# 1645
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:	216	Resp:	176000
Ion Ratio	Lower	Upper	
216	100		
214	79.5	64.3	96.5
179	20.0	19.3	28.9
108	16.3	19.8	29.6









#40
Hexachlorocyclopentadiene
Concen: 9.165 ng/ul
RT: 12.746 min Scan# 1642
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:237 Resp: 101524
Ion Ratio Lower Upper
237 100
235 61.6 50.4 75.6
272 12.9 9.7 14.5



#41
2,4,6-Trichlorophenol
Concen: 8.176 ng/ul
RT: 12.999 min Scan# 1685
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:196 Resp: 94810
Ion Ratio Lower Upper
196 100
198 93.4 76.2 114.4
200 30.0 24.6 36.8









#42
2,4,5-Trichlorophenol
Concen: 8.806 ng/ul
RT: 13.063 min Scan# 1696
Delta R.T. -0.006 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:196 Resp: 108712
Ion Ratio Lower Upper
196 100
198 92.6 79.3 118.9
200 31.1 25.8 38.8



#43
1,1'-Biphenyl
Concen: 9.005 ng/ul
RT: 13.404 min Scan# 1754
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:154 Resp: 438628
Ion Ratio Lower Upper
154 100
153 39.0 32.4 48.6
76 12.4 12.9 19.3#









#44
2-Chloronaphthalene
Concen: 9.029 ng/ul
RT: 13.440 min Scan# 1760
Delta R.T. -0.006 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:162 Resp: 333647
Ion Ratio Lower Upper
162 100
164 32.6 25.6 38.4
127 38.1 36.2 54.2



#45
2-Nitroaniline
Concen: 4.289 ng/ul
RT: 13.640 min Scan# 1794
Delta R.T. -0.006 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion: 65 Resp: 57154
Ion Ratio Lower Upper
65 100
92 72.1 53.7 80.5
138 109.6 71.4 107.2#









#47
Dimethylphthalate
Concen: 8.428 ng/ul
RT: 14.022 min Scan# 1859
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:163 Resp: 400548
Ion Ratio Lower Upper
163 100
194 4.3 3.0 4.4
164 9.8 7.8 11.8



#48
2,6-Dinitrotoluene
Concen: 6.087 ng/ul
RT: 14.140 min Scan# 1879
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:165 Resp: 55837
Ion Ratio Lower Upper
165 100
89 55.6 62.2 93.2#
121 23.0 22.1 33.1









#50
Acenaphthylene
Concen: 9.134 ng/ul
RT: 14.287 min Scan# 1904
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:	152	Resp:	526275
Ion Ratio	Lower	Upper	
152	100		
151	19.5	15.9	23.9
153	13.2	10.4	15.6



#51
3-Nitroaniline
Concen: 5.383 ng/ul
RT: 14.463 min Scan# 1934
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:	138	Resp:	56545
Ion Ratio	Lower	Upper	
138	100		
108	17.7	9.7	14.5#
92	110.3	106.7	160.1









#52
Acenaphthene
Concen: 8.536 ng/ul
RT: 14.628 min Scan# 1962
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:153 Resp: 350675
Ion Ratio Lower Upper
153 100
152 48.5 38.6 57.8
154 89.2 69.0 103.6



#53
2,4-Dinitrophenol
Concen: 3.589 ng/ul
RT: 14.669 min Scan# 1969
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:184 Resp: 21296
Ion Ratio Lower Upper
184 100
63 59.0 73.4 110.2#
154 62.1 59.4 89.0









#55
4-Nitrophenol
Concen: 4.640 ng/ul
RT: 14.757 min Scan# 1984
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:109 Resp: 41757
Ion Ratio Lower Upper
109 100
139 148.3 90.3 135.5#
65 131.3 102.2 153.2



#56
Dibenzofuran
Concen: 9.293 ng/ul
RT: 14.963 min Scan# 2019
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:168 Resp: 510445
Ion Ratio Lower Upper
168 100
139 35.9 35.3 52.9









#57
2,4-Dinitrotoluene
Concen: 6.086 ng/ul
RT: 14.922 min Scan# 2012
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:165 Resp: 82137
Ion Ratio Lower Upper
165 100
63 39.4 45.2 67.8#
182 2.5 1.9 2.9



#58
2,3,4,6-Tetrachlorophenol
Concen: 8.211 ng/ul
RT: 15.187 min Scan# 2057
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:232 Resp: 84154
Ion Ratio Lower Upper
232 100
131 42.9 51.8 77.6#
130 2.4 3.1 4.7#
166 29.0 31.0 46.4#









#59
Diethylphthalate
Concen: 7.645 ng/ul
RT: 15.387 min Scan# 2091
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:	149	Resp:	386211
Ion Ratio	Lower	Upper	
149	100		
177	22.7	15.5	23.3
150	11.9	9.9	14.9



#61
Fluorene
Concen: 8.991 ng/ul
RT: 15.610 min Scan# 2129
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:	166	Resp:	411795
Ion Ratio	Lower	Upper	
166	100		
165	96.5	79.6	119.4
167	14.0	10.7	16.1









#62
4-Chlorophenyl-phenylether
Concen: 9.997 ng/ul
RT: 15.610 min Scan# 2129
Delta R.T. 0.006 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:204 Resp: 207271
Ion Ratio Lower Upper
204 100
206 33.7 26.3 39.5
141 53.2 62.0 93.0#



#63
4-Nitroaniline
Concen: 5.784 ng/ul
RT: 15.622 min Scan# 2131
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:138 Resp: 56898
Ion Ratio Lower Upper
138 100
92 52.5 47.1 70.7
108 75.0 73.8 110.8









#66
4,6-Dinitro-2-methylphenol
Concen: 4.413 ng/ul
RT: 15.681 min Scan# 2141
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:198 Resp: 38564
Ion Ratio Lower Upper
198 100
182 6.9 4.5 6.7#
77 26.1 31.7 47.5#



#67
N-Nitrosodiphenylamine
Concen: 8.566 ng/ul
RT: 15.822 min Scan# 2165
Delta R.T. 0.006 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:169 Resp: 347995
Ion Ratio Lower Upper
169 100
168 65.2 53.9 80.9
167 34.9 30.0 45.0









#68
4-Bromophenyl-phenylether
Concen: 9.404 ng/ul
RT: 16.504 min Scan# 2281
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:248 Resp: 122665
Ion Ratio Lower Upper
248 100
250 95.1 78.4 117.6
141 69.4 80.6 121.0#



#69
Hexachlorobenzene
Concen: 9.685 ng/ul
RT: 16.616 min Scan# 2300
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:284 Resp: 142774
Ion Ratio Lower Upper
284 100
142 34.2 37.8 56.6#
249 31.9 27.1 40.7









#70
Atrazine
Concen: 8.377 ng/ul
RT: 16.775 min Scan# 2327
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:200 Resp: 120736
Ion Ratio Lower Upper
200 100
173 26.1 22.2 33.4
215 47.9 35.4 53.0



#71
Pentachlorophenol
Concen: 6.182 ng/ul
RT: 16.963 min Scan# 2359
Delta R.T. 0.006 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:266 Resp: 59690
Ion Ratio Lower Upper
266 100
264 62.0 51.0 76.4
268 66.3 49.4 74.2









#72
Phenanthrene
Concen: 8.936 ng/ul
RT: 17.357 min Scan# 2426
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:178 Resp: 667692
Ion Ratio Lower Upper
178 100
179 15.4 12.0 18.0
176 18.6 15.6 23.4



#74
Anthracene
Concen: 8.680 ng/ul
RT: 17.445 min Scan# 2441
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:178 Resp: 658610
Ion Ratio Lower Upper
178 100
179 15.4 12.1 18.1
176 17.9 14.9 22.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 10.106 ng/uL
RT: 13.369 min Scan# 1748
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:216 Resp: 181757
Ion Ratio Lower Upper
216 100
214 77.9 62.9 94.3
218 48.1 37.8 56.8



#76
Pentachlorobenzene
Concen: 9.935 ng/uL
RT: 14.881 min Scan# 2005
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:250 Resp: 176052
Ion Ratio Lower Upper
250 100
248 65.0 51.0 76.6
252 67.0 51.0 76.4









#77
Carbazole
Concen: 8.194 ng/ul
RT: 17.716 min Scan# 2487
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:	167	Resp:	580717
Ion Ratio	Lower	Upper	
167	100		
166	20.7	17.6	26.4
139	12.9	11.9	17.9



#78
Di-n-butylphthalate
Concen: 6.849 ng/ul
RT: 18.275 min Scan# 2582
Delta R.T. 0.006 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:	149	Resp:	627561
Ion Ratio	Lower	Upper	
149	100		
150	9.0	7.3	10.9
104	4.9	4.8	7.2









#80
Fluoranthene
Concen: 8.769 ng/ul
RT: 19.322 min Scan# 2760
Delta R.T. 0.006 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:202 Resp: 794467
Ion Ratio Lower Upper
202 100
101 10.8 11.0 16.4#
100 8.7 8.3 12.5



#82
Pyrene
Concen: 8.730 ng/ul
RT: 19.669 min Scan# 2819
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:202 Resp: 816912
Ion Ratio Lower Upper
202 100
101 14.5 12.6 19.0
100 11.4 10.6 15.8









#83
Butylbenzylphthalate
Concen: 5.499 ng/ul
RT: 20.545 min Scan# 2968
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:149 Resp: 254412
Ion Ratio Lower Upper
149 100
91 70.6 63.0 94.6
206 21.0 14.1 21.1



#84
3,3'-Dichlorobenzidine
Concen: 7.804 ng/ul
RT: 21.316 min Scan# 3099
Delta R.T. 0.006 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:252 Resp: 244432
Ion Ratio Lower Upper
252 100
254 63.9 51.6 77.4
126 10.8 10.7 16.1









#85
Benzo(a)anthracene
Concen: 8.824 ng/ul
RT: 21.380 min Scan# 3110
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:228 Resp: 788875
Ion Ratio Lower Upper
228 100
229 19.7 15.8 23.8
226 27.0 21.6 32.4



#86
Bis(2-ethylhexyl)phthalate
Concen: 5.497 ng/ul
RT: 21.310 min Scan# 3098
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:149 Resp: 391882
Ion Ratio Lower Upper
149 100
167 28.6 20.2 30.2
279 3.9 2.7 4.1









#87
Chrysene
Concen: 9.156 ng/ul
RT: 21.433 min Scan# 3119
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:228 Resp: 789227
Ion Ratio Lower Upper
228 100
226 29.3 24.5 36.7
229 19.6 15.9 23.9



#89
Di-n-octyl phthalate
Concen: 5.252 ng/ul
RT: 22.245 min Scan# 3257
Delta R.T. -0.006 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:149 Resp: 652063









#90
Benzo(b)fluoranthene
Concen: 8.707 ng/ul
RT: 23.086 min Scan# 3400
Delta R.T. 0.000 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:252 Resp: 812303
Ion Ratio Lower Upper
252 100
253 22.5 17.7 26.5
125 10.6 10.1 15.1



#91
Benzo(k)fluoranthene
Concen: 8.771 ng/ul
RT: 23.133 min Scan# 3408
Delta R.T. -0.006 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:252 Resp: 792271
Ion Ratio Lower Upper
252 100
253 22.0 17.2 25.8
125 10.3 9.4 14.0









#93
Benzo(a)pyrene
Concen: 9.410 ng/ul
RT: 23.721 min Scan# 3508
Delta R.T. -0.006 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24




Instrument :
BNA_P
ClientSampleId :
SSTD010616


Tgt Ion:252 Resp: 795322
Ion Ratio Lower Upper
252 100
253 22.1 17.6 26.4
125 12.0 11.3 16.9



#94
Indeno(1,2,3-cd)pyrene
Concen: 8.260 ng/ul
RT: 26.380 min Scan# 3960
Delta R.T. -0.006 min
Lab File: BP006885.D
Acq: 08 Sep 2021 16:24

Tgt Ion:276 Resp: 912828
Ion Ratio Lower Upper
276 100
138 25.0 21.8 32.8
277 24.3 19.9 29.9





#95

Dibenzo(a,h)anthracene

Concen: 8.638 ng/ul

RT: 26.410 min Scan# 3965

Delta R.T. 0.006 min

Lab File: BP006885.D

Acq: 08 Sep 2021 16:24

Instrument :

BNA_P

ClientSampleId :

SSTD010616





Tgt Ion:278 Resp: 801820

Ion Ratio Lower Upper

278 100

139 15.7 15.4 23.2

279 23.7 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 8.707 ng/ul

RT: 27.162 min Scan# 4093

Delta R.T. -0.006 min

Lab File: BP006885.D

Acq: 08 Sep 2021 16:24

Tgt Ion:276 Resp: 798880

Ion Ratio Lower Upper

276 100

138 22.4 20.7 31.1

277 23.1 19.0 28.4

